



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture
Project: Amtrak Sawtooth Bridges 2025

SDG No.: Q1194
RunNo.: LB134432

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	01/27/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	01/27/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	01/27/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q1194

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB134452

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	01/28/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	01/28/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	01/28/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	01/28/2025