

Laboratory Control Sample Summary

Client: Portal Partners Tri-Venture

SDG No.: Q1194

Project: Amtrak Sawtooth Bridges 2025

Run No.: LB134432

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134432BS							
Hexavalent Chromium	mg/L	0.5	0.50		100	1	90-111	01/27/2025

Laboratory Control Sample Summary

Client: Portal Partners Tri-Venture

SDG No.: Q1194

Project: Amtrak Sawtooth Bridges 2025

Run No.: LB134452

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166262BS							
Hexavalent Chromium	mg/Kg	20	20.0		100	1	84-110	01/28/2025