


LB134437

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	-0.007	-0.007	52	0.000	52					24
0.2	0.200	0.258	0.058	1288	0.0 %	1288					3
2.5	2.500	2.415	-0.085	11351	0.0 %	11351					2
5.0	5.000	4.921	-0.079	23038	0.0 %	23038					3
7.5	7.500	7.757	0.257	36268	0.0 %	36268					1
10.0	10.000	9.857	-0.143	46064	0.0 %	46064					

LB134437 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Jan 2025 16:41:16

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.0000	7471B	PPB	27 Jan 2025 17:09:29	S	Std
0.2 - 1	50.2	1288	-	0.2000	7471B	PPB	27 Jan 2025 17:11:48	S	Std
2.5 - 1	52.5	11351	-	2.5000	7471B	PPB	27 Jan 2025 17:14:04	S	Std
5.0 - 1	55	23038	-	5.0000	7471B	PPB	27 Jan 2025 17:18:57	S	Std
7.5 - 1	57.5	36268	-	7.5000	7471B	PPB	27 Jan 2025 17:23:53	S	Std
10.0 - 1	57.5	46064	-	10.0000	7471B	PPB	27 Jan 2025 17:35:17	S	Std
ICV70 - 1	ICV70	18225	3.8888	-7471B	PPB	27 Jan 2025 17:42:03	U	SMPL	
ICB70 - 1	ICB70	-233	-0.0682	-7471B	PPB	27 Jan 2025 17:44:18	U	SMPL	
CCV74 - 1	CCV74	24468	5.2271	-7471B	PPB	27 Jan 2025 17:46:36	U	SMPL	
CCB74 - 1	CCB74	-334	-0.0899	-7471B	PPB	27 Jan 2025 17:48:51	U	SMPL	
CRA - 1	CRA	1266	0.2531	-7471B	PPB	27 Jan 2025 17:51:10	U	SMPL	
HighStd - 1	HighStd	45773	9.7945	-7471B	PPB	27 Jan 2025 17:53:25	U	SMPL	
ChkStd - 1	ChkStd	33704	7.2071	-7471B	PPB	27 Jan 2025 17:55:41	U	SMPL	
PB166289BL - 1	PBS	-179	-0.0567	-7471B	PPB	27 Jan 2025 17:58:03	U	SMPL	
PB166289BS - 1	LCSS	17900	3.8191	-7471B	PPB	27 Jan 2025 18:03:22	U	SMPL	
Q1169-01 - 1	ARS20-0006	3230	0.6741	-7471B	PPB	27 Jan 2025 18:05:38	U	SMPL	
Q1169-03 - 1	REGULATOR-BUILDING-GRAVEL	-154	-0.0513	-7471B	PPB	27 Jan 2025 18:07:57	U	SMPL	
Q1169-05 - 1	SPILL-SITE	3270	0.6827	-7471B	PPB	27 Jan 2025 18:10:16	U	SMPL	
Q1169-05DUP - 1	SPILL-SITEDUP	3341	0.6979	-7471B	PPB	27 Jan 2025 18:12:33	U	SMPL	
Q1169-05MS - 1	SPILL-SITEMS	20710	4.4215	-7471B	PPB	27 Jan 2025 18:14:50	U	SMPL	
CCV75 - 1	CCV75	24064	5.1405	-7471B	PPB	27 Jan 2025 18:17:06	U	SMPL	
CCB75 - 1	CCB75	-213	-0.0640	-7471B	PPB	27 Jan 2025 18:19:24	U	SMPL	
Q1169-05MSD - 1	SPILL-SITEMSD	20553	4.3878	-7471B	PPB	27 Jan 2025 18:21:42	U	SMPL	
Q1170-01 - 1	OR-03-01232025	3073	0.6405	-7471B	PPB	27 Jan 2025 18:23:57	U	SMPL	
Q1183-01 - 1	VNJ-237	197	0.0239	-7471B	PPB	27 Jan 2025 18:26:15	U	SMPL	
Q1187-01 - 1	SP-COMP	706	0.1331	-7471B	PPB	27 Jan 2025 18:28:32	U	SMPL	
Q1194-01 - 1	B-110-SB01	1779	0.3631	-7471B	PPB	27 Jan 2025 18:30:49	U	SMPL	
Q1194-02 - 1	B-110-SB02	277153	59.3975	-7471B	PPB	27 Jan 2025 18:33:05	U	SMPL	
Q1194-03 - 1	B-113-SB01	320	0.0503	-7471B	PPB	27 Jan 2025 18:35:22	U	SMPL	
Q1194-04 - 1	B-113-SB02	1679	0.3416	-7471B	PPB	27 Jan 2025 18:38:08	U	SMPL	
Q1169-05LX5 - 1		740	0.1403	-7471B	PPB	27 Jan 2025 18:42:00	U	SMPL	
Q1169-05A - 1		20662	4.4112	-7471B	PPB	27 Jan 2025 18:44:15	U	SMPL	
CCV76 - 1	CCV76	24054	5.1384	-7471B	PPB	27 Jan 2025 18:46:31	U	SMPL	
CCB76 - 1	CCB76	-185	-0.0580	-7471B	PPB	27 Jan 2025 18:48:48	U	SMPL	
Q1194-02DLX10 - 1		32074	6.8577	-7471B	PPB	27 Jan 2025 18:52:10	U	SMPL	
CCV77 - 1	CCV77	24527	5.2398	-7471B	PPB	27 Jan 2025 18:54:25	U	SMPL	
CCB77 - 1	CCB77	-263	-0.0747	-7471B	PPB	27 Jan 2025 18:56:45	U	SMPL	