



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Portal Partners Tri-Venture

**SDG No.:** Q1194

**Contract:** PORT06

**Lab Code:** CHEM

**Case No.:** Q1194

**SAS No.:** Q1194

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| ICB70     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/27/2025       | 17:44            | LB134437      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Portal Partners Tri-Venture

**SDG No.:** Q1194

**Contract:** PORT06

**Lab Code:** CHEM

**Case No.:** Q1194

**SAS No.:** Q1194

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| CCB74     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/27/2025       | 17:48            | LB134437      |
| CCB75     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/27/2025       | 18:19            | LB134437      |
| CCB76     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/27/2025       | 18:48            | LB134437      |
| CCB77     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/27/2025       | 18:56            | LB134437      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b>   | Portal Partners Tri-Venture |                  |                     |              | <b>SDG No.:</b>  | Q1194 |                  |                  |               |
|------------------|-----------------------------|------------------|---------------------|--------------|------------------|-------|------------------|------------------|---------------|
| <b>Contract:</b> | PORT06                      | <b>Lab Code:</b> | CHEM                |              | <b>Case No.:</b> | Q1194 | <b>SAS No.:</b>  | Q1194            |               |
| Sample ID        | Analyte                     | Result<br>ug/L   | Acceptance<br>Limit | Conc<br>Qual | CRQL             | M     | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| ICB72            | Mercury                     | 0.20             | +/-0.20             | U            | 0.20             | CV    | 01/28/2025       | 12:42            | LB134454      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Portal Partners Tri-Venture

**SDG No.:** Q1194

**Contract:** PORT06

**Lab Code:** CHEM

**Case No.:** Q1194

**SAS No.:** Q1194

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| CCB79     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/28/2025       | 12:46            | LB134454      |
| CCB80     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/28/2025       | 13:14            | LB134454      |
| CCB81     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/28/2025       | 13:48            | LB134454      |
| CCB82     | Mercury | 0.20           | +/-0.20             | U            | 0.20 | CV | 01/28/2025       | 14:09            | LB134454      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| ICB01                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 12:21            | LB134457      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 12:21            | LB134457      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB01                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 12:43            | LB134457      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 12:43            | LB134457      |
| CCB02                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 13:46            | LB134457      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB02                                      | Silver                | 1.29                   | +/-10.0               | J            | 10.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 13:46            | LB134457      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 13:46            | LB134457      |
| CCB03                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 14:37            | LB134457      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 14:37            | LB134457      |
| CCB04                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/28/2025       | 15:36            | LB134457      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB04                                      | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Silver                | 1.18                   | +/-10.0               | J            | 10.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 15:36            | LB134457      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 15:36            | LB134457      |
| CCB05                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 15:56            | LB134457      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/28/2025       | 15:56            | LB134457      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| ICB01                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 11:37            | LB134477      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 11:37            | LB134477      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB01                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 12:10            | LB134477      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 12:10            | LB134477      |
| CCB02                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 13:16            | LB134477      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB02                                      | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 13:16            | LB134477      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 13:16            | LB134477      |
| CCB03                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Cadmium               | 0.19                   | +/-6.00               | J            | 6.00 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 14:06            | LB134477      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/29/2025       | 14:06            | LB134477      |
| CCB04                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/29/2025       | 14:28            | LB134477      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/29/2025       | 14:28            | LB134477      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Portal Partners Tri-Venture **SDG No.:** Q1194  
**Contract:** PORT06 **Lab Code:** CHEM **Case No.:** Q1194 **SAS No.:** Q1194

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB04     | Magnesium | 2000           | +/-2000             | U            | 2000 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Manganese | 20.0           | +/-20.0             | U            | 20.0 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Nickel    | 40.0           | +/-40.0             | U            | 40.0 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Potassium | 2000           | +/-2000             | U            | 2000 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Selenium  | 20.0           | +/-20.0             | U            | 20.0 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Silver    | 10.0           | +/-10.0             | U            | 10.0 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Sodium    | 2000           | +/-2000             | U            | 2000 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Thallium  | 40.0           | +/-40.0             | U            | 40.0 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Vanadium  | 40.0           | +/-40.0             | U            | 40.0 | P | 01/29/2025       | 14:28            | LB134477      |
|           | Zinc      | 40.0           | +/-40.0             | U            | 40.0 | P | 01/29/2025       | 14:28            | LB134477      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| ICB01                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 12:00            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 12:00            | LB134493      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB01                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 12:47            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 12:47            | LB134493      |
| CCB02                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 13:37            | LB134493      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB02                                      | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 13:37            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 13:37            | LB134493      |
| CCB03                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 14:38            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 14:38            | LB134493      |
| CCB04                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 15:29            | LB134493      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB04                                      | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 15:29            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 15:29            | LB134493      |
| CCB05                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 16:28            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 16:28            | LB134493      |
| CCB06                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 17:24            | LB134493      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB06                                      | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 17:24            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 17:24            | LB134493      |
| CCB07                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 18:15            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 18:15            | LB134493      |
| CCB08                                      | Aluminum              | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 01/30/2025       | 18:51            | LB134493      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> Portal Partners Tri-Venture |                       | <b>SDG No.:</b> Q1194  |                       |              |      |   |                  |                  |               |
|--------------------------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q1194 | <b>SAS No.:</b> Q1194 |              |      |   |                  |                  |               |
| Sample ID                                  | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB08                                      | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Barium                | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Iron                  | 100                    | +/-100                | U            | 100  | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 18:51            | LB134493      |
|                                            | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 01/30/2025       | 18:51            | LB134493      |