

Continuing Calibration Report for SequenceID : FC012825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068141.D

Aliphatic EPH	42381773.000	360.000	3.129	21.605	117727.147	123700.655	4.829
Aliphatic C9-C12	7788410.000	60.000	3.129	6.397	129806.833	138488.174	6.269
Aliphatic C12-C16	5151715.000	40.000	6.398	9.782	128792.875	137713.071	6.477
Aliphatic C16-C21	7539228.000	60.000	9.783	13.137	125653.800	135019.289	6.936
Aliphatic C21-C28	9526036.000	80.000	13.138	16.791	119075.450	127207.526	6.393
Aliphatic C28-C40	12376384.000	120.000	16.792	21.605	103136.533	103638.860	0.485