

Continuing Calibration Report for SequenceID : FD012825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048966.D

Aromatic C12-C16	9924062.000	60.000	5.795	8.400	165401.033	162211.796	-1.966
Aromatic C16-C21	12250818.000	80.000	8.401	12.666	153135.225	154648.237	0.978
Aromatic C21-C36	26642499.000	180.000	12.667	18.077	148013.883	142502.922	-3.867
Aromatic EPH	55294847.000	360.000	4.075	18.077	153596.797	149900.883	-2.466
Aromatic C10-C12	6477468.000	40.000	4.075	5.794	161936.700	155230.631	-4.320