

Continuing Calibration Report for SequenceID : FD012825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048975.D

Aromatic C12-C16	9985859.000	60.000	5.795	8.400	166430.983	162211.796	-2.601
Aromatic C16-C21	12307844.000	80.000	8.401	12.666	153848.050	154648.237	0.517
Aromatic C21-C36	26576852.000	180.000	12.667	18.077	147649.178	142502.922	-3.611
Aromatic EPH	55347488.000	360.000	4.075	18.077	153743.022	149900.883	-2.563
Aromatic C10-C12	6476933.000	40.000	4.075	5.794	161923.325	155230.631	-4.311