

Continuing Calibration Report for SequenceID : FF012825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015325.D

Aromatic C10-C12	6827875.000	40.000	4.485	6.355	170696.875	160369.151	-6.440
Aromatic C12-C16	9997025.000	60.000	6.356	9.053	166617.083	160326.403	-3.924
Aromatic C16-C21	12162785.000	80.000	9.054	13.372	152034.813	156685.093	2.968
Aromatic C21-C36	25021267.000	180.000	13.373	18.824	139007.039	145594.493	4.525
Aromatic EPH	54008952.000	360.000	4.485	18.824	150024.867	152156.018	1.401