

Continuing Calibration Report for SequenceID : FF012825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015340.D

Aromatic C10-C12	6826376.000	40.000	4.485	6.355	170659.400	160369.151	-6.417
Aromatic C12-C16	9928652.000	60.000	6.356	9.053	165477.533	160326.403	-3.213
Aromatic C16-C21	12128232.000	80.000	9.054	13.372	151602.900	156685.093	3.244
Aromatic C21-C36	25569173.000	180.000	13.373	18.824	142050.961	145594.493	2.434
Aromatic EPH	54452433.000	360.000	4.485	18.824	151256.758	152156.018	0.591