

Continuing Calibration Report for SequenceID : FF012925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015352.D

Aromatic C10-C12	7060922.000	40.000	4.500	6.368	176523.050	160369.151	-10.073
Aromatic C12-C16	10213821.000	60.000	6.369	9.062	170230.350	160326.403	-6.177
Aromatic C16-C21	12400069.000	80.000	9.063	13.376	155000.863	156685.093	1.075
Aromatic C21-C36	26284946.000	180.000	13.377	18.826	146027.478	145594.493	-0.297
Aromatic EPH	55959758.000	360.000	4.500	18.826	155443.772	152156.018	-2.161