

Continuing Calibration Report for SequenceID : FF012925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015359.D

Aromatic C10-C12	6826723.000	40.000	4.500	6.368	170668.075	160369.151	-6.422
Aromatic C12-C16	9825201.000	60.000	6.369	9.062	163753.350	160326.403	-2.137
Aromatic C16-C21	11801724.000	80.000	9.063	13.376	147521.550	156685.093	5.848
Aromatic C21-C36	25508347.000	180.000	13.377	18.826	141713.039	145594.493	2.666
Aromatic EPH	53961995.000	360.000	4.500	18.826	149894.431	152156.018	1.486