

Continuing Calibration Report for SequenceID : FG012825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015182.D

Aliphatic C9-C12	7819920.000	60.000	3.178	6.820	130332.000	132888.697	1.924
Aliphatic C12-C16	5548295.000	40.000	6.821	10.273	138707.375	141209.859	1.772
Aliphatic C16-C21	8448943.000	60.000	10.274	13.656	140815.717	144518.745	2.562
Aliphatic C21-C28	10685002.000	80.000	13.657	17.330	133562.525	136707.720	2.301
Aliphatic C28-C40	12783559.000	120.000	17.331	22.272	106529.658	105590.394	-0.890
Aliphatic EPH	45285719.000	360.000	3.178	22.272	125793.664	127500.849	1.339