

Continuing Calibration Report for SequenceID : FG012825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015197.D

Aliphatic C9-C12	7701724.000	60.000	3.178	6.820	128362.067	132888.697	3.406
Aliphatic C12-C16	5462110.000	40.000	6.821	10.273	136552.750	141209.859	3.298
Aliphatic C16-C21	8325655.000	60.000	10.274	13.656	138760.917	144518.745	3.984
Aliphatic C21-C28	10490876.000	80.000	13.657	17.330	131135.950	136707.720	4.076
Aliphatic C28-C40	12721706.000	120.000	17.331	22.272	106014.217	105590.394	-0.401
Aliphatic EPH	44702071.000	360.000	3.178	22.272	124172.419	127500.849	2.611