

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1194 **SAS No :** Q1194 **SDG No:** Q1194
Sample No : PB166219BS **Datafile:** FC068134.D
Client ID : PB166219BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	194	65		(40-140)
Aliphatic C12-C16	200	175	88		(40-140)
Aliphatic C16-C21	300	292	97		(40-140)
Aliphatic C21-C28	400	391	98		(40-140)
Aliphatic C28-C40	600	644	107		(40-140)