

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG NO.: Q1194

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 01/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO109302.D Time Analyzed: 22:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.793	4.695	4.895	490.700	500.000	-1.9
Aroclor-1016-2	4.812	4.714	4.914	492.170	500.000	-1.6
Aroclor-1016-3	4.869	4.770	4.970	483.870	500.000	-3.2
Aroclor-1016-4	4.989	4.891	5.091	488.000	500.000	-2.4
Aroclor-1016-5	5.247	5.149	5.349	496.010	500.000	-0.8
Aroclor-1260-1	6.289	6.190	6.390	430.090	500.000	-14.0
Aroclor-1260-2	6.477	6.379	6.579	440.410	500.000	-11.9
Aroclor-1260-3	6.846	6.747	6.947	433.480	500.000	-13.3
Aroclor-1260-4	7.107	7.008	7.208	394.520	500.000	-21.1
Aroclor-1260-5	7.348	7.249	7.449	381.150	500.000	-23.8
Decachlorobiphenyl	8.756	8.658	8.858	36.570	50.000	-26.9
Tetrachloro-m-xylene	3.698	3.600	3.800	48.880	50.000	-2.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG NO.: Q1194

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 01/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO109302.D Time Analyzed: 22:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.779	4.703	4.903	523.440	500.000	4.7
Aroclor-1016-2	4.798	4.722	4.922	529.350	500.000	5.9
Aroclor-1016-3	4.974	4.898	5.098	506.610	500.000	1.3
Aroclor-1016-4	5.016	4.940	5.140	485.490	500.000	-2.9
Aroclor-1016-5	5.229	5.153	5.353	525.720	500.000	5.1
Aroclor-1260-1	6.263	6.186	6.386	466.410	500.000	-6.7
Aroclor-1260-2	6.450	6.373	6.573	457.160	500.000	-8.6
Aroclor-1260-3	6.604	6.527	6.727	451.860	500.000	-9.6
Aroclor-1260-4	7.075	6.999	7.199	440.120	500.000	-12.0
Aroclor-1260-5	7.315	7.239	7.439	436.020	500.000	-12.8
Decachlorobiphenyl	8.708	8.633	8.833	40.200	50.000	-19.6
Tetrachloro-m-xylene	3.696	3.618	3.818	52.770	50.000	5.5