

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG No.: Q1194

Instrument ID: BNA_F Calibration Date/Time: 01/30/2025 08:54

Lab File ID: BF141301.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.303	1.194		-8.4	
Benzaldehyde	0.960	0.852		-11.3	
Phenol-d6	1.657	1.545		-6.8	
Phenol	1.757	1.657		-5.7	20.0
bis(2-Chloroethyl)ether	1.347	1.242		-7.8	
2-Chlorophenol	1.396	1.317		-5.7	
2-Methylphenol	1.135	1.057		-6.9	
2,2-oxybis(1-Chloropropane)	2.342	2.216		-5.4	
Acetophenone	0.475	0.435		-8.4	
3+4-Methylphenols	1.393	1.298		-6.8	
n-Nitroso-di-n-propylamine	0.985	0.911	0.050	-7.5	
Nitrobenzene-d5	0.379	0.356		-6.1	
Hexachloroethane	0.571	0.533		-6.7	
Nitrobenzene	0.393	0.370		-5.9	
Isophorone	0.656	0.621		-5.3	
2-Nitrophenol	0.185	0.177		-4.3	20.0
2,4-Dimethylphenol	0.236	0.227		-3.8	
bis(2-Chloroethoxy)methane	0.418	0.394		-5.7	
2,4-Dichlorophenol	0.289	0.273		-5.5	20.0
Naphthalene	1.057	0.980		-7.3	
4-Chloroaniline	0.358	0.341		-4.7	
Hexachlorobutadiene	0.194	0.181		-6.7	20.0
Caprolactam	0.094	0.091		-3.2	
4-Chloro-3-methylphenol	0.310	0.295		-4.8	20.0
2-Methylnaphthalene	0.672	0.622		-7.4	
Hexachlorocyclopentadiene	0.168	0.172	0.050	2.4	
2,4,6-Trichlorophenol	0.372	0.361		-3.0	20.0
2-Fluorobiphenyl	1.299	1.166		-10.2	
2,4,5-Trichlorophenol	0.405	0.374		-7.7	
1,1-Biphenyl	1.560	1.444		-7.4	
2-Chloronaphthalene	1.216	1.115		-8.3	
2-Nitroaniline	0.382	0.364		-4.7	
Dimethylphthalate	1.351	1.260		-6.7	
Acenaphthylene	1.702	1.594		-6.3	
2,6-Dinitrotoluene	0.314	0.293		-6.7	
3-Nitroaniline	0.307	0.296		-3.6	
Acenaphthene	1.216	1.130		-7.1	20.0
2,4-Dinitrophenol	0.146	0.139	0.050	-4.8	
4-Nitrophenol	0.225	0.221	0.050	-1.8	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG No.: Q1194

Instrument ID: BNA_F Calibration Date/Time: 01/30/2025 08:54

Lab File ID: BF141301.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.627	1.502		-7.7	
2,4-Dinitrotoluene	0.406	0.387		-4.7	
Diethylphthalate	1.310	1.217		-7.1	
4-Chlorophenyl-phenylether	0.617	0.568		-7.9	
Fluorene	1.262	1.151		-8.8	
4-Nitroaniline	0.302	0.294		-2.6	
4,6-Dinitro-2-methylphenol	0.118	0.123		4.2	
n-Nitrosodiphenylamine	0.647	0.610		-5.7	20.0
2,4,6-Tribromophenol	0.183	0.174		-4.9	
4-Bromophenyl-phenylether	0.224	0.216		-3.6	
Hexachlorobenzene	0.238	0.229		-3.8	
Atrazine	0.216	0.194		-10.2	
Pentachlorophenol	0.139	0.142		2.2	20.0
Phenanthrene	1.075	1.008		-6.2	
Anthracene	1.053	0.988		-6.2	
Carbazole	0.934	0.889		-4.8	
Di-n-butylphthalate	1.132	1.094		-3.4	
Fluoranthene	1.062	1.005		-5.4	20.0
Pyrene	1.920	1.777		-7.4	
Terphenyl-d14	1.297	1.185		-8.6	
Butylbenzylphthalate	0.671	0.641		-4.5	
3,3-Dichlorobenzidine	0.439	0.407		-7.3	
Benzo (a) anthracene	1.380	1.214		-12.0	
Chrysene	1.265	1.188		-6.1	
Bis(2-ethylhexyl)phthalate	0.851	0.798		-6.2	
Di-n-octyl phthalate	1.321	1.218		-7.8	20.0
Benzo (b) fluoranthene	1.258	1.162		-7.6	
Benzo (k) fluoranthene	1.114	1.056		-5.2	
Benzo (a) pyrene	1.063	1.010		-5.0	20.0
Indeno (1,2,3-cd) pyrene	1.291	1.200		-7.0	
Dibenzo (a,h) anthracene	1.039	0.965		-7.1	
Benzo (g,h,i) perylene	1.080	0.995		-7.9	
1,2,4,5-Tetrachlorobenzene	0.569	0.533		-6.3	
1,4-Dioxane	0.573	0.542		-5.4	20.0
2,3,4,6-Tetrachlorophenol	0.319	0.312		-2.2	

All other compounds must meet a minimum RRF of 0.010.