

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG No.: Q1194

Instrument ID: BNA_F Calibration Date/Time: 02/04/2025 15:58

Lab File ID: BF141410.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.303	1.284		-1.5	
Benzaldehyde	0.960	1.013		5.5	
Phenol-d6	1.657	1.572		-5.1	
Phenol	1.757	1.636		-6.9	20.0
bis(2-Chloroethyl)ether	1.347	1.259		-6.5	
2-Chlorophenol	1.396	1.368		-2.0	
2-Methylphenol	1.135	1.080		-4.8	
2,2-oxybis(1-Chloropropane)	2.342	2.088		-10.8	
Acetophenone	0.475	0.494		4.0	
3+4-Methylphenols	1.393	1.396		0.2	
n-Nitroso-di-n-propylamine	0.985	0.944	0.050	-4.2	
Nitrobenzene-d5	0.379	0.388		2.4	
Hexachloroethane	0.571	0.577		1.1	
Nitrobenzene	0.393	0.393		0.0	
Isophorone	0.656	0.640		-2.4	
2-Nitrophenol	0.185	0.192		3.8	20.0
2,4-Dimethylphenol	0.236	0.228		-3.4	
bis(2-Chloroethoxy)methane	0.418	0.398		-4.8	
2,4-Dichlorophenol	0.289	0.292		1.0	20.0
Naphthalene	1.057	1.049		-0.8	
4-Chloroaniline	0.358	0.319		-10.9	
Hexachlorobutadiene	0.194	0.208		7.2	20.0
Caprolactam	0.094	0.093		-1.1	
4-Chloro-3-methylphenol	0.310	0.319		2.9	20.0
2-Methylnaphthalene	0.672	0.673		0.1	
Hexachlorocyclopentadiene	0.168	0.185	0.050	10.1	
2,4,6-Trichlorophenol	0.372	0.385		3.5	20.0
2-Fluorobiphenyl	1.299	1.327		2.2	
2,4,5-Trichlorophenol	0.405	0.420		3.7	
1,1-Biphenyl	1.560	1.554		-0.4	
2-Chloronaphthalene	1.216	1.203		-1.1	
2-Nitroaniline	0.382	0.396		3.7	
Dimethylphthalate	1.351	1.359		0.6	
Acenaphthylene	1.702	1.720		1.1	
2,6-Dinitrotoluene	0.314	0.322		2.5	
3-Nitroaniline	0.307	0.312		1.6	
Acenaphthene	1.216	1.251		2.9	20.0
2,4-Dinitrophenol	0.146	0.175	0.050	19.9	
4-Nitrophenol	0.225	0.248	0.050	10.2	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG No.: Q1194

Instrument ID: BNA_F Calibration Date/Time: 02/04/2025 15:58

Lab File ID: BF141410.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.627	1.657		1.8	
2,4-Dinitrotoluene	0.406	0.439		8.1	
Diethylphthalate	1.310	1.369		4.5	
4-Chlorophenyl-phenylether	0.617	0.644		4.4	
Fluorene	1.262	1.315		4.2	
4-Nitroaniline	0.302	0.342		13.2	
4,6-Dinitro-2-methylphenol	0.118	0.142		20.3	
n-Nitrosodiphenylamine	0.647	0.643		-0.6	20.0
2,4,6-Tribromophenol	0.183	0.206		12.6	
4-Bromophenyl-phenylether	0.224	0.229		2.2	
Hexachlorobenzene	0.238	0.248		4.2	
Atrazine	0.216	0.211		-2.3	
Pentachlorophenol	0.139	0.158		13.7	20.0
Phenanthrene	1.075	1.113		3.5	
Anthracene	1.053	1.102		4.7	
Carbazole	0.934	1.020		9.2	
Di-n-butylphthalate	1.132	1.231		8.7	
Fluoranthene	1.062	1.232		16.0	20.0
Pyrene	1.920	1.732		-9.8	
Terphenyl-d14	1.297	1.228		-5.3	
Butylbenzylphthalate	0.671	0.671		0.0	
3,3-Dichlorobenzidine	0.439	0.410		-6.6	
Benzo (a) anthracene	1.380	1.278		-7.4	
Chrysene	1.265	1.238		-2.1	
Bis(2-ethylhexyl)phthalate	0.851	0.877		3.1	
Di-n-octyl phthalate	1.321	1.327		0.5	20.0
Benzo (b) fluoranthene	1.258	1.276		1.4	
Benzo (k) fluoranthene	1.114	1.215		9.1	
Benzo (a) pyrene	1.063	1.092		2.7	20.0
Indeno (1,2,3-cd) pyrene	1.291	1.304		1.0	
Dibenzo (a,h) anthracene	1.039	1.055		1.5	
Benzo (g,h,i) perylene	1.080	1.076		-0.4	
1,2,4,5-Tetrachlorobenzene	0.569	0.587		3.2	
1,4-Dioxane	0.573	0.535		-6.6	20.0
2,3,4,6-Tetrachlorophenol	0.319	0.336		5.3	

All other compounds must meet a minimum RRF of 0.010.