

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG NO.: Q1194

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/30/2025

Lab File ID: BF141301.D Time Analyzed: 08:54

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	146467	6.798	575983	8.08	308835	9.84
UPPER LIMIT	292934	7.298	1151970	8.581	617670	10.339
LOWER LIMIT	73233.5	6.298	287992	7.581	154418	9.339
EPA SAMPLE NO.						
01 PB166339BL	146115	6.80	599878	8.08	324432	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.