

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG NO.: Q1194

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/03/2025

Lab File ID: BF141361.D Time Analyzed: 11:24

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	93803	6.798	362783	8.08	198566	9.83
UPPER LIMIT	187606	7.298	725566	8.581	397132	10.333
LOWER LIMIT	46901.5	6.298	181392	7.581	99283	9.333
EPA SAMPLE NO.						
01 EB	101266	6.79	409757	8.08	228085	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.