

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG NO.: Q1194

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/05/2025

Lab File ID: BG063996.D Time Analyzed: 10:26

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	63436	7.871	270173	10.66	179556	14.50
UPPER LIMIT	126872	8.371	540346	11.156	359112	14.998
LOWER LIMIT	31718	7.371	135087	10.156	89778	13.998
EPA SAMPLE NO.						
01 PB166339BL	63484	7.87	259668	10.66	167864	14.50

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.