

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1194      SDG NO.: Q1194

Lab File ID: BF141360.D

DFTPP Injection Date: 02/03/2025

Instrument ID: BNA\_F

DFTPP Injection Time: 09:53

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	43.2
68	Less than 2.0% of mass 69	0.6 ( 1.7 ) 1
69	Mass 69 relative abundance	35.5
70	Less than 2.0% of mass 69	0.2 ( 0.7 ) 1
127	10.0 - 80.0% of mass 198	47.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	25.7
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	12.3
442	Greater than 50% of mass 198	77.3
443	15.0 - 24.0% of mass 442	14.4 ( 18.6 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141361.D	02/03/2025	11:24
EB	Q1194-08	BF141367.D	02/03/2025	14:08