

Surrogate Summary

SW-846

SDG No.: Q1194

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166294BL	PB166294BL	2-Fluorophenol	150	132	88		30 (18)	130 (112)
		Phenol-d6	150	126	84		30 (15)	130 (107)
		Nitrobenzene-d5	100	91.8	92		30 (18)	130 (107)
		2-Fluorobiphenyl	100	92.9	93		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	142	95		30 (10)	130 (116)
		Terphenyl-d14	100	85.6	86		30 (10)	130 (105)
PB166294BS	PB166294BS	2-Fluorophenol	150	133	88		30 (18)	130 (112)
		Phenol-d6	150	128	85		30 (15)	130 (107)
		Nitrobenzene-d5	100	95.1	95		30 (18)	130 (107)
		2-Fluorobiphenyl	100	93.5	94		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	155	103		30 (10)	130 (116)
		Terphenyl-d14	100	95.8	96		30 (10)	130 (105)
Q1194-01	B-110-SB01	2-Fluorophenol	150	87.7	58		30 (18)	130 (112)
		Phenol-d6	150	84.9	57		30 (15)	130 (107)
		Nitrobenzene-d5	100	60.0	60		30 (18)	130 (107)
		2-Fluorobiphenyl	100	59.4	59		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	81.0	54		30 (10)	130 (116)
		Terphenyl-d14	100	46.2	46		30 (10)	130 (105)
Q1194-02	B-110-SB02	2-Fluorophenol	150	73.0	49		30 (18)	130 (112)
		Phenol-d6	150	69.9	47		30 (15)	130 (107)
		Nitrobenzene-d5	100	47.9	48		30 (18)	130 (107)
		2-Fluorobiphenyl	100	46.0	46		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	61.9	41		30 (10)	130 (116)
		Terphenyl-d14	100	35.0	35		30 (10)	130 (105)
Q1194-03	B-113-SB01	2-Fluorophenol	150	110	74		30 (18)	130 (112)
		Phenol-d6	150	106	70		30 (15)	130 (107)
		Nitrobenzene-d5	100	76.3	76		30 (18)	130 (107)
		2-Fluorobiphenyl	100	74.1	74		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	107	71		30 (10)	130 (116)
		Terphenyl-d14	100	60.0	60		30 (10)	130 (105)
Q1194-04	B-113-SB02	2-Fluorophenol	150	108	72		30 (18)	130 (112)
		Phenol-d6	150	104	69		30 (15)	130 (107)
		Nitrobenzene-d5	100	72.9	73		30 (18)	130 (107)
		2-Fluorobiphenyl	100	68.9	69		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	102	68		30 (10)	130 (116)
		Terphenyl-d14	100	56.4	56		30 (10)	130 (105)
Q1194-04MS	B-113-SB02MS	2-Fluorophenol	150	93.0	62		30 (18)	130 (112)
		Phenol-d6	150	91.1	61		30 (15)	130 (107)
		Nitrobenzene-d5	100	65.2	65		30 (18)	130 (107)
		2-Fluorobiphenyl	100	61.8	62		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	92.2	61		30 (10)	130 (116)
		Terphenyl-d14	100	52.5	53		30 (10)	130 (105)
Q1194-04MSD	B-113-SB02MSD	2-Fluorophenol	150	92.8	62		30 (18)	130 (112)
		Phenol-d6	150	89.2	59		30 (15)	130 (107)
		Nitrobenzene-d5	100	64.3	64		30 (18)	130 (107)
		2-Fluorobiphenyl	100	60.9	61		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	88.9	59		30 (10)	130 (116)
		Terphenyl-d14	100	49.9	50		30 (10)	130 (105)

Surrogate Summary

SW-846

SDG No.: Q1194

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166339BL	PB166339BL	2-Fluorophenol	150	120	80		15 (10)	110 (139)
		Phenol-d6	150	117	78		15 (10)	110 (134)
		Nitrobenzene-d5	100	81.4	81		30 (49)	130 (133)
		2-Fluorobiphenyl	100	78.3	78		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	120	80		15 (44)	110 (137)
		Terphenyl-d14	100	81.6	82		30 (48)	130 (125)
PB166339BS	PB166339BS	2-Fluorophenol	150	136	90		15 (10)	110 (139)
		Phenol-d6	150	130	86		15 (10)	110 (134)
		Nitrobenzene-d5	100	96.4	96		30 (49)	130 (133)
		2-Fluorobiphenyl	100	94.6	95		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	152	102		15 (44)	110 (137)
		Terphenyl-d14	100	95.4	95		30 (48)	130 (125)
PB166339BSD	PB166339BSD	2-Fluorophenol	150	131	87		15 (10)	110 (139)
		Phenol-d6	150	124	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	93.1	93		30 (49)	130 (133)
		2-Fluorobiphenyl	100	93.4	93		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	151	101		15 (44)	110 (137)
		Terphenyl-d14	100	96.5	96		30 (48)	130 (125)
Q1194-08	EB	2-Fluorophenol	75	25.5	34		15 (10)	110 (139)
		Phenol-d6	75	16.6	22		15 (10)	110 (134)
		Nitrobenzene-d5	50	32.2	64		30 (49)	130 (133)
		2-Fluorobiphenyl	50	32.8	66		30 (52)	130 (132)
		2,4,6-Tribromophenol	75	52.9	71		15 (44)	110 (137)
		Terphenyl-d14	50	35.1	70		30 (48)	130 (125)