

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG NO.: Q1194
Lab File ID: VN085516.D Date Analyzed: 01/27/2025
Instrument ID: MSVOA_N Time Analyzed: 14:50
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	179985	8.22	325755	9.09	284425	11.87
UPPER LIMIT	359970	8.724	651510	9.594	568850	12.365
LOWER LIMIT	89992.5	7.724	162878	8.594	142213	11.365
EPA SAMPLE NO.						
TB	180339	8.22	326143	9.10	288357	11.87
VN0127WBL01	197682	8.22	347425	9.10	297114	11.87
VN0127WBS01	208674	8.22	353856	9.10	309822	11.87
VN0127WBSD01	201491	8.22	340191	9.10	297834	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.