

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG NO.: Q1194
 Lab File ID: VN085526.D Date Analyzed: 01/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	203638	8.22	333806	9.09	302551	11.87
UPPER LIMIT	407276	8.718	667612	9.594	605102	12.365
LOWER LIMIT	101819	7.718	166903	8.594	151276	11.365
EPA SAMPLE NO.						
FB	169030	8.22	313197	9.10	274340	11.87
VN0128WBL01	183717	8.22	332601	9.10	291066	11.87
VN0128WBS01	185562	8.22	307569	9.10	271043	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.