

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1194 SAS No.: Q1194 SDG NO.: Q1194
 Lab File ID: VN085516.D Date Analyzed: 01/27/2025
 Instrument ID: MSVOA_N Time Analyzed: 14:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	139987	13.788				
UPPER LIMIT	279974	14.288				
LOWER LIMIT	69993.5	13.288				
EPA SAMPLE NO.						
TB	112650	13.79				
VN0127WBL01	116627	13.79				
VN0127WBS01	147310	13.79				
VN0127WBSD01	142141	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.