
From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Wednesday, February 5, 2025 1:17 PM
To: Sohil Jodhani; Nimisha.Pandya@alliancetg.com; Mohammad Ahmed
Cc: Moura, Kenneth S; R5 - R5RSCC@epa.gov (R5RSCC@epa.gov)
Subject: TASK AREA DAS | REGION 5 | CASE 51900 | LAB ACE | SDG E2955 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon ACE,

Regarding the laboratory's reported issue for Case 51900/SDG E2955, please see the Region's input below and let me know if you have any questions.

LAB: "The Lab has received water samples for TVOA analysis where samples E2965 and E2966 were found with high concentration of target analytes and required dilutions. Please see attached Raw data for your reference. Due to possible matrix interference and high concentration, more than three surrogates are outside of QC limits. In this Case, the lab would like to confirm that the lab will report the sample E2965 and E2966 with surrogate failure, further dilution of the samples in final electronic deliverables.

The Lab performed the analysis for the samples E2965 AND E2966 in continuous analytical sequence. where samples have high concentrations of target analytes and require dilution as well as you can review attached quant reports for the samples. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples. There is no other lab QC failure for the associate samples. Therefore, the lab would like to confirm that the lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis for final electronic deliverables."

REGION: *"...report the sample E2965 and E2966 with surrogate failure, further dilution of the samples in final electronic deliverables."*

"...report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis for final electronic deliverables."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>
Sent: Wednesday, February 5, 2025 1:01 PM
To: Irving, Madison <Madison.Irving@gdit.com>
Subject: FW: TASK AREA DAS | REGION 5 | CASE 51900 | LAB ACE | SDG E2955 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,
Please see the client's response below.

Thanks,

Darcie D. Tone-Pah-Hote

Regional Sample Control Coordinator
LSASD-SQAB
U.S. EPA Region 5
77 W Jackson Blvd
Chicago, IL 60604
Cubicle #: 06078
Phone #: 312-886-4298

From: Ifkovich, Timothy W CIV USARMY CELRB (USA) <Timothy.W.Ifkovich@usace.army.mil>
Sent: Wednesday, February 5, 2025 11:27 AM
To: R5RSCC <R5RSCC@epa.gov>; Hares, Sarah M CIV USARMY CELRB (USA) <Sarah.M.Hares@usace.army.mil>; Kowal, Shawn F CIV USARMY CELRB (USA) <Shawn.F.Kowal@usace.army.mil>
Cc: Green, Nilia <Green.Nilia@epa.gov>
Subject: RE: TASK AREA DAS | REGION 5 | CASE 51900 | LAB ACE | SDG E2955 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

USACE confirms the following:

"...report the sample E2965 and E2966 with surrogate failure, further dilution of the samples in final electronic deliverables."

"...report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis for final electronic deliverables."

Timothy Ifkovich
Environmental Engineer
Cell: (716) 698-8311
timothy.w.ifkovich@usace.army.mil

From: R5RSCC <R5RSCC@epa.gov>
Sent: Wednesday, February 5, 2025 11:06 AM
To: Hares, Sarah M CIV USARMY CELRB (USA) <Sarah.M.Hares@usace.army.mil>; Kowal, Shawn F CIV USARMY CELRB (USA) <Shawn.F.Kowal@usace.army.mil>
Cc: Ifkovich, Timothy W CIV USARMY CELRB (USA) <Timothy.W.Ifkovich@usace.army.mil>; Green, Nilia <Green.Nilia@epa.gov>
Subject: [Non-DoD Source] FW: TASK AREA DAS | REGION 5 | CASE 51900 | LAB ACE | SDG E2955 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Good morning,

Please see the email below and let me know how you would like the lab to proceed.

Thank you,

Darcie D. Tone-Pah-Hote

Regional Sample Control Coordinator
LSASD-SQAB
U.S. EPA Region 5
77 W Jackson Blvd
Chicago, IL 60604
Cubicle #: 06078
Phone #: 312-886-4298

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Wednesday, February 5, 2025 9:47 AM
To: R5RSCC <R5RSCC@epa.gov>
Subject: TASK AREA DAS | REGION 5 | CASE 51900 | LAB ACE | SDG E2955 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Region 5,

Please advise on the laboratory's reported issue below for Case 51900/SDG E2955.

LAB: "The Lab has received water samples for TVOA analysis where samples E2965 and E2966 were found with high concentration of target analytes and required dilutions. Please see attached Raw data for your reference. Due to possible matrix interference and high concentration, more than three surrogates are outside of QC limits. In this Case, the lab would like to confirm that the lab will report the sample E2965 and E2966 with surrogate failure, further dilution of the samples in final electronic deliverables.

The Lab performed the analysis for the samples E2965 AND E2966 in continuous analytical sequence. where samples have high concentrations of target analytes and require dilution as well as you can review attached quant reports for the samples. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples. There is no other lab QC failure for the associate samples. Therefore, the lab would like to confirm that the lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis for final electronic deliverables."

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport
Sent: Wednesday, February 5, 2025 10:04 AM
To: 'Nimisha Pandya' <Nimisha.Pandya@alliancetg.com>
Cc: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Subject: REGION 5 | CASE 51900 | LAB ACE | SDG E2955 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team
DASsupport@gdit.com

GDIT Federal Civilian Division

Under contract to the EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Nimisha Pandya <Nimisha.Pandya@alliancetg.com>
Sent: Wednesday, February 5, 2025 9:49 AM
To: DASsupport <DASsupport@gdit.com>
Cc: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Subject: REGION 5 | CASE 51900 | LAB ACE | SDG E2955 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,

The Lab is sending this email regarding Case 51900 & SDG E2955

Th Lab has received water samples for TVOA analysis where samples E2965 and E2966 were found with high concentration of target analytes and required dilutions. Please see attached Raw data for your reference. Due to possible matrix interference and high concentration, more than three surrogates are outside of QC limits. In this Case, the lab would like to confirm that the lab will report the sample E2965 and E2966 with surrogate failure, further dilution of the samples in final electronic deliverables.

The Lab performed the analysis for the samples E2965 AND E2966 in continuous analytical sequence. where samples have high concentrations of target analytes and require dilution as well as you can review attached quant reports for the samples. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples. There is no other lab QC failure for the associate samples. Therefore, the lab would like to confirm that the lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis for final electronic deliverables.

Thanks & Regards,

Nimisha N Pandya
QA/QC supervisor
An Alliance Technical Group Company
Main: 908-789-8900

Direct: 908-728-3153
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com    
