

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199

Instrument ID: BNA_N Calibration Date/Time: 01/30/2025 03:43

Lab File ID: BN036128.D Init. Calib. Date(s): 01/22/2025 01/22/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:02 14:36

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.601		10.5	50.0
Fluoranthene-d10	1.036	1.159		11.9	50.0
2-Fluorophenol	1.040	1.189		14.3	50.0
Phenol-d6	1.222	1.499		22.7	50.0
Nitrobenzene-d5	0.378	0.413		9.3	50.0
2-Fluorobiphenyl	1.786	1.729		-3.2	50.0
2,4,6-Tribromophenol	0.257	0.251		-2.3	50.0
Terphenyl-d14	0.831	1.004		20.8	50.0
1,4-Dioxane	0.447	0.408		-8.7	50.0

All other compounds must meet a minimum RRF of 0.010.