

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199

Instrument ID: MSVOA_N Calibration Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Calibration Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085438.D		RRF050 = VN085439.D		RRF020 = VN085440.D			
		RRF010 = VN085441.D		RRF005 = VN085442.D		RRF001 = VN085443.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Chloromethane	0.680	0.680	0.727	0.693	0.779	0.839	0.733	8.8	
Vinyl Chloride	0.697	0.686	0.727	0.711	0.781	0.819	0.737	7.1	
Bromomethane	0.392	0.417	0.454	0.437	0.525		0.445	11.3	
Chloroethane	0.435	0.424	0.468	0.429	0.505	0.542	0.467	10.3	
Trichlorofluoromethane	1.046	0.997	1.097	1.040	1.077	1.157	1.069	5.1	
1,1,2-Trichlorotrifluoroethane	0.590	0.542	0.609	0.587	0.639	0.646	0.602	6.4	
1,1-Dichloroethene	0.548	0.533	0.556	0.526	0.559	0.497	0.537	4.3	
Acetone	0.238	0.252	0.252	0.247	0.269	0.306	0.261	9.3	
Carbon Disulfide	1.555	1.477	1.647	1.537	1.719	1.978	1.652	11	
Methyl tert-butyl Ether	1.834	1.873	1.853	1.664	1.685	1.545	1.742	7.5	
Methylene Chloride	0.629	0.629	0.658	0.606	0.696	0.656	0.646	4.9	
trans-1,2-Dichloroethene	0.571	0.555	0.574	0.539	0.569	0.632	0.573	5.5	
1,1-Dichloroethane	1.164	1.170	1.206	1.100	1.226	1.204	1.178	3.8	
2-Butanone	0.378	0.390	0.398	0.363	0.387	0.386	0.384	3.1	
Carbon Tetrachloride	0.574	0.530	0.579	0.529	0.565	0.567	0.557	4	
cis-1,2-Dichloroethene	0.691	0.683	0.715	0.639	0.669	0.655	0.675	4	
Chloroform	1.197	1.175	1.241	1.169	1.253	1.273	1.218	3.6	
1,1,1-Trichloroethane	1.053	1.016	1.091	1.000	1.148	1.102	1.068	5.2	
Methylcyclohexane	0.564	0.463	0.477	0.407	0.437	0.397	0.457	13.3	
Benzene	1.551	1.449	1.527	1.376	1.474	1.400	1.463	4.7	
1,2-Dichloroethane	0.569	0.547	0.575	0.522	0.574	0.517	0.551	4.8	
Trichloroethene	0.362	0.324	0.352	0.310	0.343	0.352	0.341	5.8	
1,2-Dichloropropane	0.390	0.371	0.388	0.334	0.388	0.371	0.374	5.7	
Bromodichloromethane	0.590	0.559	0.579	0.514	0.569	0.484	0.549	7.5	
4-Methyl-2-Pentanone	0.499	0.492	0.495	0.432	0.443	0.380	0.457	10.3	
Toluene	0.964	0.870	0.919	0.808	0.835	0.690	0.848	11.3	
t-1,3-Dichloropropene	0.594	0.551	0.544	0.481	0.527	0.416	0.519	12	
cis-1,3-Dichloropropene	0.623	0.588	0.601	0.527	0.538	0.450	0.554	11.4	
1,1,2-Trichloroethane	0.348	0.340	0.353	0.309	0.349	0.314	0.335	5.7	
2-Hexanone	0.358	0.357	0.353	0.298	0.302	0.261	0.321	12.6	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199
 Instrument ID: MSVOA_N Calibration Date(s): 01/14/2025 01/14/2025
 Heated Purge: (Y/N) N Calibration Time(s): 14:56 17:19
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085438.D		RRF050 = VN085439.D		RRF020 = VN085440.D			
		RRF010 = VN085441.D		RRF005 = VN085442.D		RRF001 = VN085443.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.430	0.414	0.412	0.368	0.420	0.386	0.405	5.8	
Tetrachloroethene	0.351	0.322	0.365	0.338	0.346	0.323	0.341	4.9	
Chlorobenzene	1.133	1.076	1.154	1.047	1.110	1.051	1.095	4	
Ethyl Benzene	2.072	1.867	1.940	1.685	1.709	1.430	1.784	12.7	
m/p-Xylenes	0.775	0.707	0.750	0.615	0.616	0.492	0.659	16	
o-Xylene	0.738	0.681	0.713	0.584	0.582	0.482	0.630	15.5	
Styrene	1.271	1.173	1.186	0.956	0.929	0.742	1.043	19.2	
Bromoform	0.311	0.311	0.312	0.273	0.284	0.235	0.288	10.6	
Isopropylbenzene	3.922	3.448	3.681	3.272	3.157	2.766	3.375	12.1	
1,1,2,2-Tetrachloroethane	1.121	1.145	1.187	1.157	1.228	1.314	1.192	5.9	
1,3-Dichlorobenzene	1.720	1.565	1.701	1.574	1.656	1.526	1.624	4.9	
1,4-Dichlorobenzene	1.706	1.562	1.713	1.607	1.743	1.767	1.683	4.8	
1,2-Dichlorobenzene	1.611	1.555	1.654	1.532	1.600	1.766	1.620	5.2	
1,2-Dichloroethane-d4	0.774	0.831	0.754	0.762	0.914		0.807	8.3	
Dibromofluoromethane	0.359	0.358	0.335	0.310	0.373		0.347	7.1	
Toluene-d8	1.339	1.267	1.207	1.076	1.274		1.232	8.1	
4-Bromofluorobenzene	0.475	0.449	0.410	0.357	0.417		0.422	10.6	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.