

Cover Page

Order ID : Q1199

Project ID : CTO WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q1199-01
Q1199-02
Q1199-03
Q1199-04
Q1199-05
Q1199-06
Q1199-07

Client Sample Number

BP-VPB-192-TB-20250123
BP-VPB-192-EB-20250124
BP-VPB-192-GW-280-282
BP-VPB-192-GW-260-262
BP-VPB-192-GW-240-242
BP-VPB-192-GW-220-222
BP-VPB-192-DUP-20250123

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 2/26/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: CTO WE13

Project Manager: Ernie Wu

Chemtech Project # Q1199

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

7 Water samples were received on 01/27/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID VN085526.D met the requirements except for 2-Hexanone and Styrene failing high but associated sample having hit below CRQL therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

This data package has been revised due to client ID changed for sample#07 as per client request.

The Sample # BP-VPB-192-GW-220-222 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1199

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration File ID VN085526.D met the requirements except for 2- Hexanone and Styrene failing high but associated sample having hit below CRQL therefore no corrective action taken.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

This data package has been revised due to client ID changed for sample#07 as per client request.

The Sample #BP-VPB-192-GW-220-222 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).”
Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1199

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 02/26/2025

LAB CHRONICLE

OrderID:	Q1199	OrderDate:	1/27/2025 3:35:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	CTO WE13
Contact:	Ernie Wu	Location:	N21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1199-01	BP-VPB-192-TB-2025 0123	Water			01/23/25			01/27/25
			VOCMS Group1	8260-Low			01/28/25	
Q1199-02	BP-VPB-192-EB-2025 0124	Water			01/24/25			01/27/25
			VOCMS Group1	8260-Low			01/28/25	
Q1199-03	BP-VPB-192-GW-280- 282	Water			01/24/25			01/27/25
			VOCMS Group1	8260-Low			01/28/25	
Q1199-04	BP-VPB-192-GW-260- 262	Water			01/23/25			01/27/25
			VOCMS Group1	8260-Low			01/28/25	
Q1199-05	BP-VPB-192-GW-240- 242	Water			01/23/25			01/27/25
			VOCMS Group1	8260-Low			01/28/25	
Q1199-06	BP-VPB-192-GW-220- 222	Water			01/23/25			01/27/25
			VOCMS Group1	8260-Low			01/28/25	
Q1199-07	BP-VPB-192-DUP-202 50123	Water			01/23/25			01/27/25
			VOCMS Group1	8260-Low			01/28/25	

Hit Summary Sheet
SW-846

SDG No.: Q1199
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID: Q1199-03	BP-VPB-192-GW-280-282 BP-VPB-192-GW-2 Water	Acetone		3.80	J	1.40	3.80	5.00	ug/L
		Total Voc :		3.80					
		Total Concentration:		3.80					
Client ID: Q1199-04	BP-VPB-192-GW-260-262 BP-VPB-192-GW-2 Water	Acetone		6.30		1.40	3.80	5.00	ug/L
		Total Voc :		6.30					
		Total Concentration:		6.30					
Client ID: Q1199-05	BP-VPB-192-GW-240-242 BP-VPB-192-GW-2 Water	Acetone		3.90	J	1.40	3.80	5.00	ug/L
Q1199-05	BP-VPB-192-GW-2 Water	Trichloroethene		12.9		0.32	0.75	1.00	ug/L
Q1199-05	BP-VPB-192-GW-2 Water	Tetrachloroethene		0.95	J	0.25	0.50	1.00	ug/L
		Total Voc :		17.8					
		Total Concentration:		17.8					
Client ID: Q1199-06	BP-VPB-192-GW-220-222 BP-VPB-192-GW-2 Water	Chloromethane		0.70	J	0.35	0.50	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	1,1-Dichloroethene		0.48	J	0.26	0.75	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	Acetone		6.50		1.40	3.80	5.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	Methyl tert-butyl Ether		0.56	J	0.16	0.50	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	Chloroform		0.83	J	0.26	0.50	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	Trichloroethene		1.40		0.32	0.75	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	Toluene		0.52	J	0.18	0.50	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	Ethyl Benzene		0.34	J	0.16	0.50	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	o-Xylene		0.30	J	0.14	0.50	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	Styrene		0.31	J	0.16	0.50	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	Isopropylbenzene		0.27	J	0.13	0.50	1.00	ug/L
Q1199-06	BP-VPB-192-GW-2 Water	1,1,2,2-Tetrachloroethane		0.61	J	0.27	0.50	1.00	ug/L
		Total Voc :		12.8					
		Total Concentration:		12.8					
Client ID: Q1199-07	BP-VPB-192-DUP-20250123 BP-VPB-192-DUP- Water	Acetone		5.30		1.40	3.80	5.00	ug/L
Q1199-07	BP-VPB-192-DUP- Water	Trichloroethene		1.10		0.32	0.75	1.00	ug/L
		Total Voc :		6.40					
		Total Concentration:		6.40					



QC SUMMARY

Surrogate Summary

SDG No.: **Q1199**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1199-01	BP-VPB-192-TB-20250123	1,2-Dichloroethane-d4	50	55.4	111	81	118
		Dibromofluoromethane	50	52.1	104	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	45.5	91	85	114
Q1199-02	BP-VPB-192-EB-20250124	1,2-Dichloroethane-d4	50	56.7	113	81	118
		Dibromofluoromethane	50	53.7	107	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	43.3	87	85	114
Q1199-03	BP-VPB-192-GW-280-282	1,2-Dichloroethane-d4	50	57.0	114	81	118
		Dibromofluoromethane	50	54.0	108	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	47.0	94	85	114
Q1199-04	BP-VPB-192-GW-260-262	1,2-Dichloroethane-d4	50	57.2	114	81	118
		Dibromofluoromethane	50	53.6	107	80	119
		Toluene-d8	50	48.8	98	89	112
		4-Bromofluorobenzene	50	43.5	87	85	114
Q1199-05	BP-VPB-192-GW-240-242	1,2-Dichloroethane-d4	50	57.1	114	81	118
		Dibromofluoromethane	50	52.6	105	80	119
		Toluene-d8	50	49.4	99	89	112
		4-Bromofluorobenzene	50	45.4	91	85	114
Q1199-06	BP-VPB-192-GW-220-222	1,2-Dichloroethane-d4	50	58.5	117	81	118
		Dibromofluoromethane	50	53.6	107	80	119
		Toluene-d8	50	46.9	94	89	112
		4-Bromofluorobenzene	50	44.5	89	85	114
Q1199-07	BP-VPB-192-DUP-20250123	1,2-Dichloroethane-d4	50	55.6	111	81	118
		Dibromofluoromethane	50	52.5	105	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	46.6	93	85	114
VN0128WBL01	VN0128WBL01	1,2-Dichloroethane-d4	50	53.7	107	81	118
		Dibromofluoromethane	50	52.2	104	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	44.2	88	85	114
VN0128WBS01	VN0128WBS01	1,2-Dichloroethane-d4	50	51.1	102	81	118
		Dibromofluoromethane	50	53.2	106	80	119
		Toluene-d8	50	54.1	108	89	112
		4-Bromofluorobenzene	50	52.0	104	85	114
VN0128WBSD0	VN0128WBSD01	1,2-Dichloroethane-d4	50	50.3	101	81	118
		Dibromofluoromethane	50	50.2	100	80	119
		Toluene-d8	50	51.1	102	89	112
		4-Bromofluorobenzene	50	51.0	102	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1199

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN085529.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0128WBS01	Chloromethane	20	17.3	ug/L	86			50	139	
	Vinyl chloride	20	18.2	ug/L	91			58	137	
	Bromomethane	20	19.5	ug/L	98			53	141	
	Chloroethane	20	19.8	ug/L	99			60	138	
	Trichlorofluoromethane	20	20.7	ug/L	104			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.5	ug/L	103			70	136	
	1,1-Dichloroethene	20	19.1	ug/L	96			71	131	
	Acetone	100	100	ug/L	100			39	160	
	Carbon disulfide	20	16.4	ug/L	82			64	133	
	Methyl tert-butyl Ether	20	20.9	ug/L	104			71	124	
	Methylene Chloride	20	19.8	ug/L	99			74	124	
	trans-1,2-Dichloroethene	20	18.9	ug/L	95			75	124	
	1,1-Dichloroethane	20	20.1	ug/L	101			77	125	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	20.5	ug/L	103			72	136	
	cis-1,2-Dichloroethene	20	19.9	ug/L	100			78	123	
	Chloroform	20	20.6	ug/L	103			79	124	
	1,1,1-Trichloroethane	20	20.4	ug/L	102			74	131	
	Methylcyclohexane	20	18.9	ug/L	95			72	132	
	Benzene	20	20.7	ug/L	104			79	120	
	1,2-Dichloroethane	20	21.0	ug/L	105			73	128	
	Trichloroethene	20	20.4	ug/L	102			79	123	
	1,2-Dichloropropane	20	20.7	ug/L	104			78	122	
	Bromodichloromethane	20	21.7	ug/L	109			79	125	
	4-Methyl-2-Pentanone	100	120	ug/L	120			67	130	
	Toluene	20	21.9	ug/L	110			80	121	
	t-1,3-Dichloropropene	20	21.4	ug/L	107			73	127	
	cis-1,3-Dichloropropene	20	21.5	ug/L	108			75	124	
	1,1,2-Trichloroethane	20	21.9	ug/L	110			80	119	
	2-Hexanone	100	120	ug/L	120			57	139	
	Dibromochloromethane	20	21.2	ug/L	106			74	126	
	Tetrachloroethene	20	21.7	ug/L	109			74	129	
	Chlorobenzene	20	20.6	ug/L	103			82	118	
	Ethyl Benzene	20	20.8	ug/L	104			79	121	
	m/p-Xylenes	40	43.1	ug/L	108			80	121	
	o-Xylene	20	20.9	ug/L	104			78	122	
	Styrene	20	21.3	ug/L	106			78	123	
	Bromoform	20	22.5	ug/L	113			66	130	
	Isopropylbenzene	20	20.9	ug/L	104			72	131	
	1,1,2,2-Tetrachloroethane	20	21.6	ug/L	108			71	121	
	1,3-Dichlorobenzene	20	20.8	ug/L	104			80	119	
	1,4-Dichlorobenzene	20	19.8	ug/L	99			79	118	
	1,2-Dichlorobenzene	20	20.4	ug/L	102			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1199

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN085530.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0128WBSD01	Chloromethane	20	17.7	ug/L	89	3		50	139	20
	Vinyl chloride	20	17.8	ug/L	89	2		58	137	20
	Bromomethane	20	18.0	ug/L	90	9		53	141	20
	Chloroethane	20	19.6	ug/L	98	1		60	138	20
	Trichlorofluoromethane	20	19.5	ug/L	98	6		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	20.3	ug/L	102	1		70	136	20
	1,1-Dichloroethene	20	19.5	ug/L	98	2		71	131	20
	Acetone	100	110	ug/L	110	10		39	160	20
	Carbon disulfide	20	16.2	ug/L	81	1		64	133	20
	Methyl tert-butyl Ether	20	21.9	ug/L	110	6		71	124	20
	Methylene Chloride	20	19.7	ug/L	99	0		74	124	20
	trans-1,2-Dichloroethene	20	18.9	ug/L	95	0		75	124	20
	1,1-Dichloroethane	20	20.4	ug/L	102	1		77	125	20
	2-Butanone	100	110	ug/L	110	0		56	143	20
	Carbon Tetrachloride	20	20.0	ug/L	100	3		72	136	20
	cis-1,2-Dichloroethene	20	20.2	ug/L	101	1		78	123	20
	Chloroform	20	20.6	ug/L	103	0		79	124	20
	1,1,1-Trichloroethane	20	20.1	ug/L	101	1		74	131	20
	Methylcyclohexane	20	18.7	ug/L	94	1		72	132	20
	Benzene	20	20.2	ug/L	101	3		79	120	20
	1,2-Dichloroethane	20	21.5	ug/L	108	3		73	128	20
	Trichloroethene	20	19.4	ug/L	97	5		79	123	20
	1,2-Dichloropropane	20	20.7	ug/L	104	0		78	122	20
	Bromodichloromethane	20	21.7	ug/L	109	0		79	125	20
	4-Methyl-2-Pentanone	100	120	ug/L	120	0		67	130	20
	Toluene	20	21.5	ug/L	108	2		80	121	20
	t-1,3-Dichloropropene	20	21.5	ug/L	108	1		73	127	20
	cis-1,3-Dichloropropene	20	21.8	ug/L	109	1		75	124	20
	1,1,2-Trichloroethane	20	21.5	ug/L	108	2		80	119	20
	2-Hexanone	100	120	ug/L	120	0		57	139	20
	Dibromochloromethane	20	21.9	ug/L	110	4		74	126	20
	Tetrachloroethene	20	21.0	ug/L	105	4		74	129	20
	Chlorobenzene	20	20.6	ug/L	103	0		82	118	20
	Ethyl Benzene	20	20.6	ug/L	103	1		79	121	20
	m/p-Xylenes	40	43.4	ug/L	109	1		80	121	20
	o-Xylene	20	20.8	ug/L	104	0		78	122	20
	Styrene	20	21.3	ug/L	106	0		78	123	20
	Bromoform	20	22.5	ug/L	113	0		66	130	20
	Isopropylbenzene	20	21.2	ug/L	106	2		72	131	20
	1,1,2,2-Tetrachloroethane	20	21.8	ug/L	109	1		71	121	20
	1,3-Dichlorobenzene	20	21.0	ug/L	105	1		80	119	20
	1,4-Dichlorobenzene	20	20.4	ug/L	102	3		79	118	20
	1,2-Dichlorobenzene	20	20.4	ug/L	102	0		80	119	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0128WBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1199

SAS No.: Q1199 SDG NO.: Q1199

Lab File ID: VN085528.D

Lab Sample ID: VN0128WBL01

Date Analyzed: 01/28/2025

Time Analyzed: 10:51

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0128WBS01	VN0128WBS01	VN085529.D	01/28/2025
VN0128WBSD01	VN0128WBSD01	VN085530.D	01/28/2025
BP-VPB-192-TB-20250123	Q1199-01	VN085534.D	01/28/2025
BP-VPB-192-EB-20250124	Q1199-02	VN085535.D	01/28/2025
BP-VPB-192-GW-280-282	Q1199-03	VN085536.D	01/28/2025
BP-VPB-192-GW-260-262	Q1199-04	VN085537.D	01/28/2025
BP-VPB-192-GW-240-242	Q1199-05	VN085538.D	01/28/2025
BP-VPB-192-DUP-20250123	Q1199-07	VN085540.D	01/28/2025
BP-VPB-192-GW-220-222	Q1199-06	VN085545.D	01/28/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG NO.: Q1199
Lab File ID: VN085437.D BFB Injection Date: 01/14/2025
Instrument ID: MSVOA_N BFB Injection Time: 14:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.3
75	30.0 - 60.0% of mass 95	58
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.8
173	Less than 2.0% of mass 174	1.4 (1.8) 1
174	50.0 - 100.0% of mass 95	76
175	5.0 - 9.0% of mass 174	5.4 (7.1) 1
176	95.0 - 101.0% of mass 174	74.1 (97.4) 1
177	5.0 - 9.0% of mass 176	4.9 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN085438.D	01/14/2025	14:56
VSTDICCC050	VSTDICCC050	VN085439.D	01/14/2025	15:19
VSTDICC020	VSTDICC020	VN085440.D	01/14/2025	15:43
VSTDICC010	VSTDICC010	VN085441.D	01/14/2025	16:07
VSTDICC005	VSTDICC005	VN085442.D	01/14/2025	16:31
VSTDICC001	VSTDICC001	VN085443.D	01/14/2025	17:19



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG NO.: Q1199
Lab File ID: VN085525.D BFB Injection Date: 01/28/2025
Instrument ID: MSVOA_N BFB Injection Time: 09:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.1
75	30.0 - 60.0% of mass 95	55.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.4
173	Less than 2.0% of mass 174	1.5 (1.9) 1
174	50.0 - 100.0% of mass 95	75.9
175	5.0 - 9.0% of mass 174	5.5 (7.2) 1
176	95.0 - 101.0% of mass 174	73.7 (97.2) 1
177	5.0 - 9.0% of mass 176	5 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN085526.D	01/28/2025	09:53
VN0128WBL01	VN0128WBL01	VN085528.D	01/28/2025	10:51
VN0128WBS01	VN0128WBS01	VN085529.D	01/28/2025	11:15
VN0128WBSD01	VN0128WBSD01	VN085530.D	01/28/2025	11:48
BP-VPB-192-TB-20250123	Q1199-01	VN085534.D	01/28/2025	13:24
BP-VPB-192-EB-20250124	Q1199-02	VN085535.D	01/28/2025	13:49
BP-VPB-192-GW-280-282	Q1199-03	VN085536.D	01/28/2025	14:13
BP-VPB-192-GW-260-262	Q1199-04	VN085537.D	01/28/2025	14:37
BP-VPB-192-GW-240-242	Q1199-05	VN085538.D	01/28/2025	15:01
BP-VPB-192-DUP-20250123	Q1199-07	VN085540.D	01/28/2025	15:49
BP-VPB-192-GW-220-222	Q1199-06	VN085545.D	01/28/2025	17:50
VSTDCCC050EC	VSTDCCC050	VN085546.D	01/28/2025	18:14

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG NO.: Q1199
Lab File ID: VN085526.D Date Analyzed: 01/28/2025
Instrument ID: MSVOA_N Time Analyzed: 09:53
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	203638	8.22	333806	9.09	302551	11.87
UPPER LIMIT	407276	8.718	667612	9.594	605102	12.365
LOWER LIMIT	101819	7.718	166903	8.594	151276	11.365
EPA SAMPLE NO.						
BP-VPB-192-TB-20250123	179083	8.22	333621	9.10	293971	11.87
BP-VPB-192-EB-20250124	158329	8.22	293573	9.10	250987	11.87
BP-VPB-192-GW-280-282	162876	8.22	301319	9.10	270559	11.87
BP-VPB-192-GW-260-262	153446	8.22	287847	9.10	248805	11.87
BP-VPB-192-GW-240-242	162553	8.22	306518	9.10	270799	11.87
BP-VPB-192-GW-220-222	169137	8.22	327198	9.10	270505	11.87
BP-VPB-192-DUP-20250123	162005	8.22	300571	9.10	271698	11.87
VN0128WBL01	183717	8.22	332601	9.10	291066	11.87
VN0128WBS01	185562	8.22	307569	9.10	271043	11.87
VN0128WBSD01	182154	8.22	309854	9.10	273221	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG NO.: Q1199
 Lab File ID: VN085526.D Date Analyzed: 01/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	155713	13.788				
UPPER LIMIT	311426	14.288				
LOWER LIMIT	77856.5	13.288				
EPA SAMPLE NO.						
BP-VPB-192-TB-20250123	113614	13.79				
BP-VPB-192-EB-20250124	92670	13.79				
BP-VPB-192-GW-280-282	109450	13.79				
BP-VPB-192-GW-260-262	90634	13.79				
BP-VPB-192-GW-240-242	108170	13.79				
BP-VPB-192-GW-220-222	106728	13.79				
BP-VPB-192-DUP-20250123	107447	13.79				
VN0128WBL01	116250	13.79				
VN0128WBS01	126721	13.79				
VN0128WBSD01	127977	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	01/23/25		
Project:	CTO WE13			Date Received:	01/27/25		
Client Sample ID:	BP-VPB-192-TB-20250123			SDG No.:	Q1199		
Lab Sample ID:	Q1199-01			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085534.D	1		01/28/25 13:24	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/23/25
Project:	CTO WE13	Date Received:	01/27/25
Client Sample ID:	BP-VPB-192-TB-20250123	SDG No.:	Q1199
Lab Sample ID:	Q1199-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085534.D	1		01/28/25 13:24	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.4		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.1		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	49.5		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.5		85 - 114		91%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	179000	8.224				
540-36-3	1,4-Difluorobenzene	334000	9.1				
3114-55-4	Chlorobenzene-d5	294000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	114000	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/23/25	
Project:	CTO WE13		Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-TB-20250123		SDG No.:	Q1199	
Lab Sample ID:	Q1199-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085534.D	1		01/28/25 13:24	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085534.D
 Acq On : 28 Jan 2025 13:24
 Operator : JC\MD
 Sample : Q1199-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-TB-20250123

Quant Time: Jan 29 00:38:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	179083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	333621	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	293971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	113614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160235	55.430	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.860%		
35) Dibromofluoromethane	8.165	113	120678	52.140	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.280%		
50) Toluene-d8	10.565	98	407027	49.496	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.000%		
62) 4-Bromofluorobenzene	12.847	95	127940	45.481	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 90.960%		


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085534.D
Acq On : 28 Jan 2025 13:24
Operator : JC\MD
Sample : Q1199-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-TB-20250123


Quant Time: Jan 29 00:38:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085534.D
Acq: 28 Jan 2025 13:24

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-TB-20250123







Tgt Ion:168 Resp: 179083
Ion Ratio Lower Upper
168 100
99 64.7 53.6 80.4



#33
1,2-Dichloroethane-d4
Concen: 55.430 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN085534.D
Acq: 28 Jan 2025 13:24

Tgt Ion: 65 Resp: 160235
Ion Ratio Lower Upper
65 100
67 50.0 0.0 101.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 12

Delta R.T. 0.006 min

Lab File: VN085534.D

Acq: 28 Jan 2025 13:24





Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-TB-20250123

Tgt Ion:114	Resp: 333621
Ion Ratio	Lower Upper
114 100	
63 21.2	0.0 47.6
88 15.6	0.0 32.6



#35

Dibromofluoromethane

Concen: 52.140 ug/l




RT: 8.165 min Scan# 1056


Delta R.T. 0.006 min

Lab File: VN085534.D

Acq: 28 Jan 2025 13:24

Tgt Ion:113	Resp: 120678
Ion Ratio	Lower Upper
113 100	
111 102.8	82.7 124.1
192 17.6	14.3 21.5





#50

Toluene-d8

Concen: 49.496 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085534.D


Acq: 28 Jan 2025 13:24

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-TB-20250123






Tgt Ion: 98 Resp: 407027

Ion Ratio Lower Upper

98 100

100 64.4 52.2 78.4



#62

4-Bromofluorobenzene

Concen: 45.481 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min

Lab File: VN085534.D

Acq: 28 Jan 2025 13:24




Tgt Ion: 95 Resp: 127940


Ion Ratio Lower Upper

95 100

174 74.9 0.0 145.0

176 71.5 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085534.D

Acq: 28 Jan 2025 13:24

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-TB-20250123





Tgt Ion:117 Resp: 293971

Ion Ratio Lower Upper

117 100

82 56.7 48.6 72.8

119 32.0 26.6 39.8



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085534.D

Acq: 28 Jan 2025 13:24




Tgt Ion:152 Resp: 113614

Ion Ratio Lower Upper

152 100

115 62.6 31.1 93.3

150 159.2 0.0 343.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085534.D
Acq On : 28 Jan 2025 13:24
Operator : JC\MD
Sample : Q1199-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-TB-20250123

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Title : SW846 8260

Signal : TIC: VN085534.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	8.165	1046	1056	1060	rBV	171155	395569	35.77%	6.933%
2	8.224	1060	1066	1078	rVB	267035	597722	54.05%	10.476%
3	8.577	1117	1126	1139	rBV	192146	427450	38.65%	7.492%
4	9.100	1206	1215	1226	rBV	409612	822324	74.36%	14.413%
5	10.565	1456	1464	1477	rBV	631405	1105941	100.00%	19.383%
6	11.865	1677	1685	1698	rBV	547850	956941	86.53%	16.772%
7	12.847	1845	1852	1862	rBV	391610	657078	59.41%	11.516%
8	13.788	2005	2012	2023	rBV	441663	742576	67.14%	13.015%




Sum of corrected areas: 5705601

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085534.D
Acq On : 28 Jan 2025 13:24
Operator : JC\MD
Sample : Q1199-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-TB-20250123

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085534.D
Acq On : 28 Jan 2025 13:24
Operator : JC\MD
Sample : Q1199-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-TB-20250123

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085534.D
Acq On : 28 Jan 2025 13:24
Operator : JC\MD
Sample : Q1199-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-TB-20250123

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	01/24/25		
Project:	CTO WE13			Date Received:	01/27/25		
Client Sample ID:	BP-VPB-192-EB-20250124			SDG No.:	Q1199		
Lab Sample ID:	Q1199-02			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085535.D	1		01/28/25 13:49	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/24/25	
Project:	CTO WE13		Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-EB-20250124		SDG No.:	Q1199	
Lab Sample ID:	Q1199-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085535.D	1		01/28/25 13:49	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.7		81 - 118		113%	SPK: 50
1868-53-7	Dibromofluoromethane	53.7		80 - 119		107%	SPK: 50
2037-26-5	Toluene-d8	49.5		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.3		85 - 114		87%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	158000	8.224				
540-36-3	1,4-Difluorobenzene	294000	9.1				
3114-55-4	Chlorobenzene-d5	251000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	92700	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/24/25	
Project:	CTO WE13		Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-EB-20250124		SDG No.:	Q1199	
Lab Sample ID:	Q1199-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085535.D	1		01/28/25 13:49	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085535.D
 Acq On : 28 Jan 2025 13:49
 Operator : JC\MD
 Sample : Q1199-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-EB-20250124

Quant Time: Jan 29 00:38:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	293573	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	250987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	144816	56.663	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.320%
35) Dibromofluoromethane	8.165	113	109419	53.724	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.440%
50) Toluene-d8	10.565	98	357973	49.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.940%
62) 4-Bromofluorobenzene	12.847	95	107196	43.305	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.620%


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085535.D
Acq On : 28 Jan 2025 13:49
Operator : JC\MD
Sample : Q1199-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-EB-20250124


Quant Time: Jan 29 00:38:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085535.D
Acq: 28 Jan 2025 13:49

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-EB-20250124







Tgt Ion:168 Resp: 158329
Ion Ratio Lower Upper
168 100
99 65.4 53.6 80.4



#33
1,2-Dichloroethane-d4
Concen: 56.663 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN085535.D
Acq: 28 Jan 2025 13:49

Tgt Ion: 65 Resp: 144816
Ion Ratio Lower Upper
65 100
67 50.3 0.0 101.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 12

Delta R.T. 0.006 min

Lab File: VN085535.D

Acq: 28 Jan 2025 13:49





Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-EB-20250124

Tgt Ion:	114	Resp:	293573
Ion Ratio	Lower	Upper	
114	100		
63	22.8	0.0	47.6
88	17.0	0.0	32.6



#35

Dibromofluoromethane

Concen: 53.724 ug/l




RT: 8.165 min Scan# 1056


Delta R.T. 0.006 min

Lab File: VN085535.D

Acq: 28 Jan 2025 13:49

Tgt Ion:	113	Resp:	109419
Ion Ratio	Lower	Upper	
113	100		
111	102.6	82.7	124.1
192	16.8	14.3	21.5





#50

Toluene-d8

Concen: 49.469 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085535.D


Acq: 28 Jan 2025 13:49

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-EB-20250124






Tgt Ion: 98 Resp: 357973

Ion Ratio Lower Upper

98 100

100 63.3 52.2 78.4



#62

4-Bromofluorobenzene

Concen: 43.305 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. 0.000 min

Lab File: VN085535.D

Acq: 28 Jan 2025 13:49




Tgt Ion: 95 Resp: 107196


Ion Ratio Lower Upper

95 100

174 74.1 0.0 145.0

176 69.7 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085535.D

Acq: 28 Jan 2025 13:49





Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-EB-20250124

Tgt Ion:	117	Resp:	250987
Ion	Ratio	Lower	Upper
117	100		
82	59.3	48.6	72.8
119	33.4	26.6	39.8



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085535.D

Acq: 28 Jan 2025 13:49

Tgt Ion:	152	Resp:	92670
Ion	Ratio	Lower	Upper
152	100		
115	61.4	31.1	93.3
150	158.3	0.0	343.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085535.D
Acq On : 28 Jan 2025 13:49
Operator : JC\MD
Sample : Q1199-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-EB-20250124

Integration Parameters: RTEINT.P
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Title : SW846 8260

Signal : TIC: VN085535.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.318	50	62	67	rBV	53437	81473	8.36%	1.619%
2	8.165	1046	1056	1060	rBV	153066	358050	36.75%	7.115%
3	8.224	1060	1066	1077	rVB	235036	527888	54.18%	10.489%
4	8.577	1116	1126	1140	rBV	174352	388867	39.91%	7.727%
5	9.100	1206	1215	1227	rBV	362094	736413	75.59%	14.633%
6	10.565	1456	1464	1479	rBV	539422	974245	100.00%	19.359%
7	11.865	1676	1685	1697	rBV	463575	820833	84.25%	16.310%
8	12.847	1844	1852	1862	rBV	316744	544778	55.92%	10.825%
9	13.788	2004	2012	2023	rBV	358369	600081	61.59%	11.924%




Sum of corrected areas: 5032628

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085535.D
Acq On : 28 Jan 2025 13:49
Operator : JC\MD
Sample : Q1199-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-EB-20250124

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085535.D
Acq On : 28 Jan 2025 13:49
Operator : JC\MD
Sample : Q1199-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-EB-20250124

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085535.D
Acq On : 28 Jan 2025 13:49
Operator : JC\MD
Sample : Q1199-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-EB-20250124

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	01/24/25		
Project:	CTO WE13			Date Received:	01/27/25		
Client Sample ID:	BP-VPB-192-GW-280-282			SDG No.:	Q1199		
Lab Sample ID:	Q1199-03			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085536.D	1		01/28/25 14:13	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	01/24/25	
Project:	CTO WE13			Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-GW-280-282			SDG No.:	Q1199	
Lab Sample ID:	Q1199-03			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085536.D	1		01/28/25 14:13	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	57.0		81 - 118		114%	SPK: 50
1868-53-7	Dibromofluoromethane	54.0		80 - 119		108%	SPK: 50
2037-26-5	Toluene-d8	50.0		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.0		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	163000	8.224				
540-36-3	1,4-Difluorobenzene	301000	9.1				
3114-55-4	Chlorobenzene-d5	271000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	109000	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/24/25	
Project:	CTO WE13		Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-GW-280-282		SDG No.:	Q1199	
Lab Sample ID:	Q1199-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085536.D	1		01/28/25 14:13	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085536.D
 Acq On : 28 Jan 2025 14:13
 Operator : JC\MD
 Sample : Q1199-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-280-282

Quant Time: Jan 29 00:39:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	162876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301319	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	109450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	149778	56.968	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.940%			
35) Dibromofluoromethane	8.165	113	112960	54.037	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.080%			
50) Toluene-d8	10.565	98	371723	50.049	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.100%			
62) 4-Bromofluorobenzene	12.847	95	119312	46.961	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.920%			
Target Compounds						
16) Acetone	4.442	43	3210	3.779	ug/l	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085536.D
Acq On : 28 Jan 2025 14:13
Operator : JC\MD
Sample : Q1199-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-280-282


Quant Time: Jan 29 00:39:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085536.D
Acq: 28 Jan 2025 14:13

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-280-282







Tgt Ion:168 Resp: 162876
Ion Ratio Lower Upper
168 100
99 66.2 53.6 80.4



#16
Acetone
Concen: 3.779 ug/l
RT: 4.442 min Scan# 423
Delta R.T. 0.024 min
Lab File: VN085536.D
Acq: 28 Jan 2025 14:13

Tgt Ion: 43 Resp: 3210
Ion Ratio Lower Upper
43 100
58 29.0 23.8 35.6





#33

1,2-Dichloroethane-d4

Concen: 56.968 ug/l

RT: 8.577 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085536.D

Acq: 28 Jan 2025 14:13

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-GW-280-282

Tgt Ion: 65 Resp: 149778

Ion Ratio Lower Upper

65 100

67 50.5 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085536.D

Acq: 28 Jan 2025 14:13




Tgt Ion: 114 Resp: 301319


Ion Ratio Lower Upper

114 100

63 21.8 0.0 47.6

88 16.5 0.0 32.6





#35

Dibromofluoromethane

Concen: 54.037 ug/l

RT: 8.165 min Scan# 1055

Delta R.T. 0.006 min

Lab File: VN085536.D


Acq: 28 Jan 2025 14:13

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-280-282






Tgt Ion: 113 Resp: 112960

Ion Ratio Lower Upper

113 100

111 103.0 82.7 124.1

192 17.7 14.3 21.5



#50

Toluene-d8

Concen: 50.049 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085536.D




Acq: 28 Jan 2025 14:13


Tgt Ion: 98 Resp: 371723

Ion Ratio Lower Upper

98 100

100 64.3 52.2 78.4





#62

4-Bromofluorobenzene

Concen: 46.961 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. 0.000 min

Lab File: VN085536.D

Acq: 28 Jan 2025 14:13

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-280-282





Tgt Ion: 95 Resp: 119312

Ion Ratio Lower Upper

95 100

174 75.7 0.0 145.0

176 71.4 0.0 142.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085536.D

Acq: 28 Jan 2025 14:13




Tgt Ion: 117 Resp: 270559


Ion Ratio Lower Upper

117 100

82 58.4 48.6 72.8


119 32.2 26.6 39.8







#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085536.D
Acq: 28 Jan 2025 14:13

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-280-282



Tgt Ion	Ratio	Lower	Upper
152	100		
115	62.6	31.1	93.3
150	159.5	0.0	343.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085536.D
 Acq On : 28 Jan 2025 14:13
 Operator : JC\MD
 Sample : Q1199-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-280-282

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M

Title : SW846 8260

Signal : TIC: VN085536.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.718	461	470	480	rBV4	4037	11562	1.14%	0.214%
2	8.165	1043	1056	1061	rBV2	158417	391775	38.48%	7.255%
3	8.224	1061	1066	1076	rVB2	245965	522468	51.32%	9.675%
4	8.577	1116	1126	1140	rBV2	179747	400957	39.38%	7.425%
5	9.100	1206	1215	1223	rBV	370767	755027	74.16%	13.981%
6	10.565	1456	1464	1473	rBV	565147	1018115	100.00%	18.853%
7	11.300	1583	1589	1597	rVB	23159	39938	3.92%	0.740%
8	11.865	1676	1685	1695	rBV	498106	882446	86.67%	16.341%
9	12.847	1845	1852	1863	rVB	367128	612540	60.16%	11.343%
10	13.618	1978	1983	1989	rVB3	8547	13309	1.31%	0.246%
11	13.788	2005	2012	2024	rBV	423608	710833	69.82%	13.163%
12	14.553	2137	2142	2151	rBV4	26449	41259	4.05%	0.764%




Sum of corrected areas: 5400229

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085536.D
Acq On : 28 Jan 2025 14:13
Operator : JC\MD
Sample : Q1199-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-280-282

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085536.D
Acq On : 28 Jan 2025 14:13
Operator : JC\MD
Sample : Q1199-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-280-282

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085536.D
Acq On : 28 Jan 2025 14:13
Operator : JC\MD
Sample : Q1199-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-280-282

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	01/23/25		
Project:	CTO WE13			Date Received:	01/27/25		
Client Sample ID:	BP-VPB-192-GW-260-262			SDG No.:	Q1199		
Lab Sample ID:	Q1199-04			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085537.D	1		01/28/25 14:37	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	6.30		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/23/25
Project:	CTO WE13	Date Received:	01/27/25
Client Sample ID:	BP-VPB-192-GW-260-262	SDG No.:	Q1199
Lab Sample ID:	Q1199-04	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085537.D	1		01/28/25 14:37	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	57.2		81 - 118		114%	SPK: 50
1868-53-7	Dibromofluoromethane	53.6		80 - 119		107%	SPK: 50
2037-26-5	Toluene-d8	48.8		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.5		85 - 114		87%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	153000	8.224				
540-36-3	1,4-Difluorobenzene	288000	9.1				
3114-55-4	Chlorobenzene-d5	249000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	90600	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/23/25	
Project:	CTO WE13		Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-GW-260-262		SDG No.:	Q1199	
Lab Sample ID:	Q1199-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085537.D	1		01/28/25 14:37	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085537.D
 Acq On : 28 Jan 2025 14:37
 Operator : JC\MD
 Sample : Q1199-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-260-262

Quant Time: Jan 29 00:39:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	153446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	287847	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	90634	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	141633	57.181	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.360%			
35) Dibromofluoromethane	8.165	113	107001	53.582	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.160%			
50) Toluene-d8	10.565	98	346432	48.827	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.660%			
62) 4-Bromofluorobenzene	12.847	95	105662	43.535	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 87.060%			
Target Compounds						
16) Acetone	4.447	43	5039	6.296	ug/l	Qvalue 93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085537.D
Acq On : 28 Jan 2025 14:37
Operator : JC\MD
Sample : Q1199-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-260-262


Quant Time: Jan 29 00:39:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085537.D
Acq: 28 Jan 2025 14:37

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-260-262







Tgt Ion:168 Resp: 153446
Ion Ratio Lower Upper
168 100
99 65.1 53.6 80.4



#16
Acetone
Concen: 6.296 ug/l
RT: 4.447 min Scan# 424
Delta R.T. 0.029 min
Lab File: VN085537.D
Acq: 28 Jan 2025 14:37

Tgt Ion: 43 Resp: 5039
Ion Ratio Lower Upper
43 100
58 25.9 23.8 35.6





#33

1,2-Dichloroethane-d4

Concen: 57.181 ug/l

RT: 8.577 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085537.D

Acq: 28 Jan 2025 14:37

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-GW-260-262

Tgt Ion: 65 Resp: 141633

Ion Ratio Lower Upper

65 100

67 49.5 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085537.D

Acq: 28 Jan 2025 14:37




Tgt Ion: 114 Resp: 287847


Ion Ratio Lower Upper

114 100

63 22.4 0.0 47.6

88 14.8 0.0 32.6





#35

Dibromofluoromethane

Concen: 53.582 ug/l

RT: 8.165 min Scan# 1055

Delta R.T. 0.006 min

Lab File: VN085537.D

Acq: 28 Jan 2025 14:37

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-260-262





Tgt Ion:113 Resp: 107001

Ion Ratio Lower Upper

113 100

111 102.5 82.7 124.1

192 17.7 14.3 21.5



#50

Toluene-d8

Concen: 48.827 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085537.D




Acq: 28 Jan 2025 14:37


Tgt Ion: 98 Resp: 346432

Ion Ratio Lower Upper

98 100

100 64.1 52.2 78.4





#62

4-Bromofluorobenzene

Concen: 43.535 ug/l

RT: 12.847 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN085537.D


Acq: 28 Jan 2025 14:37

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-260-262






Tgt Ion: 95 Resp: 105662

Ion Ratio Lower Upper

95 100

174 76.3 0.0 145.0

176 69.3 0.0 142.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN085537.D

Acq: 28 Jan 2025 14:37




Tgt Ion: 117 Resp: 248805


Ion Ratio Lower Upper

117 100

82 59.5 48.6 72.8

119 32.3 26.6 39.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 20

Delta R.T. -0.000 min

Lab File: VN085537.D


Acq: 28 Jan 2025 14:37

Instrument :



MSVOA_N

ClientSampleId :

BP-VPB-192-GW-260-262



Tgt Ion:	152	Resp:	90634
Ion	Ratio	Lower	Upper
152	100		
115	62.2	31.1	93.3
150	161.9	0.0	343.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085537.D
Acq On : 28 Jan 2025 14:37
Operator : JC\MD
Sample : Q1199-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-260-262

Integration Parameters: RTEINT.P
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Title : SW846 8260

Signal : TIC: VN085537.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.665	115	121	126	rBV4	5647	11275	1.19%	0.231%
2	8.165	1045	1056	1061	rBV2	152611	371887	39.25%	7.623%
3	8.224	1061	1066	1079	rVB	228846	493592	52.10%	10.118%
4	8.577	1115	1126	1140	rBV	167156	380108	40.12%	7.792%
5	9.100	1206	1215	1230	rBV	358697	714152	75.38%	14.639%
6	10.565	1456	1464	1475	rBV	525977	947450	100.00%	19.422%
7	11.294	1584	1588	1596	rBB3	8370	13899	1.47%	0.285%
8	11.865	1677	1685	1696	rBV	460140	807919	85.27%	16.562%
9	12.847	1844	1852	1864	rBV	326230	541777	57.18%	11.106%
10	13.788	2004	2012	2021	rBV	357706	596216	62.93%	12.222%




Sum of corrected areas: 4878275

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085537.D
Acq On : 28 Jan 2025 14:37
Operator : JC\MD
Sample : Q1199-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-260-262

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085537.D
Acq On : 28 Jan 2025 14:37
Operator : JC\MD
Sample : Q1199-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-260-262

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085537.D
Acq On : 28 Jan 2025 14:37
Operator : JC\MD
Sample : Q1199-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-260-262

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	01/23/25		
Project:	CTO WE13			Date Received:	01/27/25		
Client Sample ID:	BP-VPB-192-GW-240-242			SDG No.:	Q1199		
Lab Sample ID:	Q1199-05			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085538.D	1		01/28/25 15:01	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.90	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	12.9		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/23/25
Project:	CTO WE13	Date Received:	01/27/25
Client Sample ID:	BP-VPB-192-GW-240-242	SDG No.:	Q1199
Lab Sample ID:	Q1199-05	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085538.D	1		01/28/25 15:01	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.95	J	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	57.1		81 - 118		114%	SPK: 50
1868-53-7	Dibromofluoromethane	52.6		80 - 119		105%	SPK: 50
2037-26-5	Toluene-d8	49.4		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.4		85 - 114		91%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	163000	8.224				
540-36-3	1,4-Difluorobenzene	307000	9.1				
3114-55-4	Chlorobenzene-d5	271000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	108000	13.794				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/23/25	
Project:	CTO WE13		Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-GW-240-242		SDG No.:	Q1199	
Lab Sample ID:	Q1199-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085538.D	1		01/28/25 15:01	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085538.D
 Acq On : 28 Jan 2025 15:01
 Operator : JC\MD
 Sample : Q1199-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-240-242

Quant Time: Jan 29 00:39:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	162553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306518	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	149816	57.096	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.200%
35) Dibromofluoromethane	8.165	113	111820	52.585	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.160%
50) Toluene-d8	10.565	98	373062	49.377	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.760%
62) 4-Bromofluorobenzene	12.847	95	117423	45.434	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.860%
Target Compounds						
16) Acetone	4.441	43	3317	3.912	ug/l	95
44) Trichloroethene	9.353	130	27023	12.946	ug/l	89
64) Tetrachloroethene	11.100	164	1751	0.948	ug/l #	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085538.D
Acq On : 28 Jan 2025 15:01
Operator : JC\MD
Sample : Q1199-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-240-242


Quant Time: Jan 29 00:39:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085538.D
Acq: 28 Jan 2025 15:01

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-240-242







Tgt Ion:168 Resp: 162553
Ion Ratio Lower Upper
168 100
99 65.9 53.6 80.4



#16
Acetone
Concen: 3.912 ug/l
RT: 4.441 min Scan# 423
Delta R.T. 0.023 min
Lab File: VN085538.D
Acq: 28 Jan 2025 15:01

Tgt Ion: 43 Resp: 3317
Ion Ratio Lower Upper
43 100
58 26.9 23.8 35.6





#33

1,2-Dichloroethane-d4

Concen: 57.096 ug/l

RT: 8.576 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085538.D

Acq: 28 Jan 2025 15:01

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-GW-240-242

Tgt Ion: 65 Resp: 149816

Ion Ratio Lower Upper

65 100

67 51.0 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085538.D

Acq: 28 Jan 2025 15:01




Tgt Ion: 114 Resp: 306518


Ion Ratio Lower Upper

114 100

63 22.0 0.0 47.6

88 16.4 0.0 32.6





#35

Dibromofluoromethane

Concen: 52.585 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085538.D


Acq: 28 Jan 2025 15:01

Instrument :




MSVOA_N

ClientSampleId :

BP-VPB-192-GW-240-242



Tgt Ion:	113	Resp:	111820
Ion	Ratio	Lower	Upper
113	100		
111	103.5	82.7	124.1
192	17.5	14.3	21.5



#44

Trichloroethene

Concen: 12.946 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. 0.006 min

Lab File: VN085538.D

Acq: 28 Jan 2025 15:01

Tgt Ion:	130	Resp:	27023
Ion	Ratio	Lower	Upper
130	100		
95	109.0	0.0	195.8





#50

Toluene-d8

Concen: 49.377 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085538.D


Acq: 28 Jan 2025 15:01

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-240-242






Tgt Ion: 98 Resp: 373062

Ion Ratio Lower Upper

98 100

100 64.2 52.2 78.4



#62

4-Bromofluorobenzene

Concen: 45.434 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min

Lab File: VN085538.D

Acq: 28 Jan 2025 15:01




Tgt Ion: 95 Resp: 117423


Ion Ratio Lower Upper

95 100

174 75.0 0.0 145.0

176 71.6 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN085538.D


Acq: 28 Jan 2025 15:01

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-240-242






Tgt Ion:117 Resp: 270799

Ion Ratio Lower Upper

117 100

82 57.1 48.6 72.8

119 31.6 26.6 39.8



#64

Tetrachloroethene

Concen: 0.948 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN085538.D

Acq: 28 Jan 2025 15:01

Tgt Ion:164 Resp: 1751




Ion Ratio Lower Upper


164 100

166 109.0 103.4 155.2

129 78.4 79.2 118.8#


131 78.3 77.1 115.7







#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN085538.D
Acq: 28 Jan 2025 15:01

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-240-242



Tgt Ion:	152	Resp:	108170
Ion	Ratio	Lower	Upper
152	100		
115	60.7	31.1	93.3
150	152.8	0.0	343.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085538.D
Acq On : 28 Jan 2025 15:01
Operator : JC\MD
Sample : Q1199-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-240-242

Integration Parameters: RTEINT.P
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Title : SW846 8260

Signal : TIC: VN085538.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.659	116	120	127	rVB	6663	12159	1.19%	0.223%
2	8.165	1047	1056	1061	rBV	160191	385326	37.64%	7.072%
3	8.224	1061	1066	1077	rVB2	249439	528183	51.59%	9.694%
4	8.576	1117	1126	1136	rBV	179179	398978	38.97%	7.323%
5	9.100	1206	1215	1228	rBV	381855	762173	74.45%	13.989%
6	9.353	1250	1258	1271	rBV	73625	154770	15.12%	2.841%
7	10.565	1454	1464	1478	rBV	567041	1023760	100.00%	18.790%
8	11.100	1550	1555	1561	rVB3	6887	11160	1.09%	0.205%
9	11.300	1584	1589	1596	rVB3	7114	11315	1.11%	0.208%
10	11.865	1677	1685	1698	rBV	501371	875267	85.50%	16.064%
11	12.847	1845	1852	1864	rBV	355355	599915	58.60%	11.011%
12	13.788	2005	2012	2022	rBV	402249	685547	66.96%	12.582%




Sum of corrected areas: 5448553

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085538.D
Acq On : 28 Jan 2025 15:01
Operator : JC\MD
Sample : Q1199-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-240-242

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085538.D
Acq On : 28 Jan 2025 15:01
Operator : JC\MD
Sample : Q1199-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-240-242

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085538.D
Acq On : 28 Jan 2025 15:01
Operator : JC\MD
Sample : Q1199-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-240-242

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	01/23/25	
Project:	CTO WE13			Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-GW-220-222			SDG No.:	Q1199	
Lab Sample ID:	Q1199-06			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085545.D	1		01/28/25 17:50	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.70	J	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.48	J	0.26	0.75	1.00	ug/L
67-64-1	Acetone	6.50		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.56	J	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.83	J	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	1.40		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.52	J	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/23/25
Project:	CTO WE13	Date Received:	01/27/25
Client Sample ID:	BP-VPB-192-GW-220-222	SDG No.:	Q1199
Lab Sample ID:	Q1199-06	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085545.D	1		01/28/25 17:50	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.34	J	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.30	J	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.31	J	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.27	J	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.61	J	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	58.5		81 - 118		117%	SPK: 50
1868-53-7	Dibromofluoromethane	53.6		80 - 119		107%	SPK: 50
2037-26-5	Toluene-d8	46.9		89 - 112		94%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.5		85 - 114		89%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	169000	8.224				
540-36-3	1,4-Difluorobenzene	327000	9.1				
3114-55-4	Chlorobenzene-d5	271000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	107000	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/23/25	
Project:	CTO WE13		Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-GW-220-222		SDG No.:	Q1199	
Lab Sample ID:	Q1199-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085545.D	1		01/28/25 17:50	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085545.D
 Acq On : 28 Jan 2025 17:50
 Operator : JC\MD
 Sample : Q1199-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-220-222

Quant Time: Jan 29 00:42:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	169137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	327198	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	106728	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159806	58.532	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.060%			
35) Dibromofluoromethane	8.165	113	121606	53.572	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.140%			
50) Toluene-d8	10.565	98	378279	46.903	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 93.800%			
62) 4-Bromofluorobenzene	12.847	95	122753	44.494	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 88.980%			
Target Compounds						
					Qvalue	
3) Chloromethane	2.359	50	1726	0.696	ug/l #	88
5) Bromomethane	2.971	94	960	0.638	ug/l #	68
12) 1,1-Dichloroethene	4.353	96	867	0.478	ug/l #	64
16) Acetone	4.442	43	5696	6.457	ug/l #	87
19) Methyl tert-butyl Ether	5.789	73	3301	0.560	ug/l #	83
30) Chloroform	7.965	83	3416	0.829	ug/l	96
44) Trichloroethene	9.347	130	3125	1.402	ug/l	87
52) Toluene	10.629	92	2909	0.524	ug/l	90
67) Ethyl Benzene	11.959	91	3265	0.338	ug/l	91
68) m/p-Xylenes	12.070	106	1081	0.303	ug/l	97
69) o-Xylene	12.394	106	1009	0.296	ug/l	98
70) Styrene	12.406	104	1725	0.306	ug/l #	89
73) Isopropylbenzene	12.694	105	1965	0.273	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	1551	0.610	ug/l #	95
80) 1,3,5-Trimethylbenzene	13.170	105	1791	0.301	ug/l	99
84) 1,2,4-Trimethylbenzene	13.476	105	1506	0.254	ug/l	84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085545.D
Acq On : 28 Jan 2025 17:50
Operator : JC\MD
Sample : Q1199-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-220-222


Quant Time: Jan 29 00:42:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085545.D
Acq: 28 Jan 2025 17:50

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-220-222







Tgt Ion:168 Resp: 169137
Ion Ratio Lower Upper
168 100
99 67.2 53.6 80.4



#3
Chloromethane
Concen: 0.696 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085545.D
Acq: 28 Jan 2025 17:50

Tgt Ion: 50 Resp: 1726
Ion Ratio Lower Upper
50 100
52 40.6 27.0 40.6#



#5
Bromomethane

Concen: 0.638 ug/l

RT: 2.971 min Scan# 17

Delta R.T. 0.018 min

Lab File: VN085545.D


Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-220-222






Tgt Ion: 94 Resp: 960

Ion Ratio Lower Upper

94 100

96 59.8 71.8 107.6#



#12

1,1-Dichloroethene


Concen: 0.478 ug/l

RT: 4.353 min Scan# 408

Delta R.T. 0.012 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50





Tgt Ion: 96 Resp: 867


Ion Ratio Lower Upper

96 100

61 109.4 141.0 211.4#

98 59.4 49.0 73.6





#16

Acetone

Concen: 6.457 ug/l

RT: 4.442 min Scan# 423

Delta R.T. 0.023 min

Lab File: VN085545.D


Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-220-222






Tgt Ion: 43 Resp: 5696

Ion Ratio Lower Upper

43 100

58 36.8 23.8 35.6#



#19

Methyl tert-butyl Ether

Concen: 0.560 ug/l

RT: 5.789 min Scan# 652

Delta R.T. -0.000 min

Lab File: VN085545.D




Acq: 28 Jan 2025 17:50


Tgt Ion: 73 Resp: 3301

Ion Ratio Lower Upper

73 100

57 14.0 17.8 26.8#





#30

Chloroform

Concen: 0.829 ug/l

RT: 7.965 min Scan# 10

Delta R.T. 0.006 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-GW-220-222

Tgt Ion: 83 Resp: 3416

Ion Ratio Lower Upper

83 100

85 61.3 51.8 77.6



#33

1,2-Dichloroethane-d4

Concen: 58.532 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.006 min

Lab File: VN085545.D




Acq: 28 Jan 2025 17:50


Tgt Ion: 65 Resp: 159806

Ion Ratio Lower Upper

65 100

67 49.8 0.0 101.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1214

Delta R.T. 0.006 min

Lab File: VN085545.D


Acq: 28 Jan 2025 17:50

Instrument :




MSVOA_N

ClientSampleId :

BP-VPB-192-GW-220-222



Tgt Ion:	114	Resp:	327198
Ion	Ratio	Lower	Upper
114	100		
63	22.5	0.0	47.6
88	16.4	0.0	32.6



#35

Dibromofluoromethane


Concen: 53.572 ug/l

RT: 8.165 min Scan# 1056



Delta R.T. 0.006 min


Lab File: VN085545.D

Acq: 28 Jan 2025 17:50



Tgt Ion:	113	Resp:	121606
Ion	Ratio	Lower	Upper
113	100		
111	101.6	82.7	124.1
192	17.3	14.3	21.5





#44

Trichloroethene

Concen: 1.402 ug/l

RT: 9.347 min Scan# 12

Delta R.T. -0.000 min

Lab File: VN085545.D


Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-220-222






Tgt Ion:130 Resp: 3125

Ion Ratio Lower Upper

130 100

95 85.2 0.0 195.8



#50

Toluene-d8

Concen: 46.903 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085545.D




Acq: 28 Jan 2025 17:50


Tgt Ion: 98 Resp: 378279

Ion Ratio Lower Upper

98 100

100 63.6 52.2 78.4





#52

Toluene

Concen: 0.524 ug/l

RT: 10.629 min Scan# 1474

Delta R.T. 0.005 min

Lab File: VN085545.D


Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-220-222






Tgt Ion: 92 Resp: 2909

Ion Ratio Lower Upper

92 100

91 159.9 139.2 208.8



#62

4-Bromofluorobenzene

Concen: 44.494 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50




Tgt Ion: 95 Resp: 122753


Ion Ratio Lower Upper

95 100

174 74.2 0.0 145.0

176 71.2 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-220-222





Tgt Ion:117 Resp: 270505

Ion Ratio Lower Upper

117 100

82 56.8 48.6 72.8

119 31.1 26.6 39.8



#67

Ethyl Benzene

Concen: 0.338 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN085545.D




Acq: 28 Jan 2025 17:50


Tgt Ion: 91 Resp: 3265

Ion Ratio Lower Upper

91 100

106 24.8 23.8 35.8





#68

m/p-Xylenes

Concen: 0.303 ug/l

RT: 12.070 min Scan# 1719

Delta R.T. 0.006 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-GW-220-222

Tgt Ion:106 Resp: 1081

Ion Ratio Lower Upper

106 100

91 214.1 167.7 251.5



#69

o-Xylene

Concen: 0.296 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. -0.000 min

Lab File: VN085545.D




Acq: 28 Jan 2025 17:50


Tgt Ion:106 Resp: 1009

Ion Ratio Lower Upper

106 100

91 216.8 110.4 331.2





#70

Styrene

Concen: 0.306 ug/l

RT: 12.406 min Scan# 17

Delta R.T. -0.000 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50





Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-GW-220-222

Tgt Ion:	104	Resp:	1725
Ion Ratio	Lower	Upper	
104	100		
78	38.2	42.5	63.7#
103	54.2	43.8	65.8



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. -0.000 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50

Tgt Ion:	152	Resp:	106728
Ion Ratio	Lower	Upper	
152	100		
115	62.4	31.1	93.3
150	160.4	0.0	343.6





#73

Isopropylbenzene

Concen: 0.273 ug/l

RT: 12.694 min Scan# 18

Delta R.T. 0.006 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-GW-220-222

Tgt Ion:105 Resp: 1965

Ion Ratio Lower Upper

105 100

120 26.1 12.8 38.3



#75

1,1,2,2-Tetrachloroethane

Concen: 0.610 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50




Tgt Ion: 83 Resp: 1551


Ion Ratio Lower Upper

83 100

131 0.0 4.8 14.4#

85 62.9 32.2 96.6





#80

1,3,5-Trimethylbenzene

Concen: 0.301 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN085545.D

Acq: 28 Jan 2025 17:50

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-GW-220-222

Tgt Ion:105 Resp: 1791

Ion Ratio Lower Upper

105 100

120 47.1 23.9 71.7



#84

1,2,4-Trimethylbenzene

Concen: 0.254 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. -0.000 min

Lab File: VN085545.D




Acq: 28 Jan 2025 17:50

Tgt Ion:105 Resp: 1506

Ion Ratio Lower Upper

105 100

120 33.3 21.6 65.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085545.D
 Acq On : 28 Jan 2025 17:50
 Operator : JC\MD
 Sample : Q1199-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-220-222

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Title : SW846 8260

Signal : TIC: VN085545.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.300	52	59	66	rBV4	4605	13190	1.28%	0.231%
2	2.659	115	120	126	rBV2	6095	11147	1.08%	0.195%
3	4.730	463	472	481	rVB4	3840	11268	1.09%	0.197%
4	8.165	1045	1056	1061	rBV	173837	417034	40.49%	7.288%
5	8.224	1061	1066	1078	rVB	252155	555963	53.98%	9.716%
6	8.577	1115	1126	1136	rBV	194360	436388	42.37%	7.626%
7	8.677	1136	1143	1146	rBV5	5979	13917	1.35%	0.243%
8	8.777	1153	1160	1180	rVB2	30907	118370	11.49%	2.069%
9	9.100	1205	1215	1230	rVB	412877	822253	79.84%	14.370%
10	9.353	1252	1258	1263	rBV4	9141	17070	1.66%	0.298%
11	10.565	1456	1464	1472	rBV	576878	1029901	100.00%	17.998%
12	10.623	1472	1474	1481	rVB3	6722	10992	1.07%	0.192%
13	11.300	1583	1589	1595	rBV4	10443	18874	1.83%	0.330%
14	11.865	1676	1685	1696	rBV	501934	890322	86.45%	15.559%
15	12.847	1845	1852	1864	rBV	375912	632785	61.44%	11.058%
16	13.788	2005	2012	2022	rVB	408690	704526	68.41%	12.312%
17	14.559	2137	2143	2147	rBV3	12269	18162	1.76%	0.317%




Sum of corrected areas: 5722162

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085545.D
Acq On : 28 Jan 2025 17:50
Operator : JC\MD
Sample : Q1199-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-220-222

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085545.D
Acq On : 28 Jan 2025 17:50
Operator : JC\MD
Sample : Q1199-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-220-222

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085545.D
Acq On : 28 Jan 2025 17:50
Operator : JC\MD
Sample : Q1199-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-220-222

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	01/23/25		
Project:	CTO WE13			Date Received:	01/27/25		
Client Sample ID:	BP-VPB-192-DUP-20250123			SDG No.:	Q1199		
Lab Sample ID:	Q1199-07			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085540.D	1		01/28/25 15:49	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	5.30		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	1.10		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/23/25
Project:	CTO WE13	Date Received:	01/27/25
Client Sample ID:	BP-VPB-192-DUP-20250123	SDG No.:	Q1199
Lab Sample ID:	Q1199-07	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085540.D	1		01/28/25 15:49	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.6		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.5		80 - 119		105%	SPK: 50
2037-26-5	Toluene-d8	50.1		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.6		85 - 114		93%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	162000	8.224				
540-36-3	1,4-Difluorobenzene	301000	9.1				
3114-55-4	Chlorobenzene-d5	272000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	107000	13.794				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/23/25	
Project:	CTO WE13		Date Received:	01/27/25	
Client Sample ID:	BP-VPB-192-DUP-20250123		SDG No.:	Q1199	
Lab Sample ID:	Q1199-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085540.D	1		01/28/25 15:49	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085540.D
Acq On : 28 Jan 2025 15:49
Operator : JC\MD
Sample : Q1199-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-DUP-20250123

Quant Time: Jan 29 00:40:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	162005	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	300571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	271698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	107447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	145516	55.645	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	111.280%	
35) Dibromofluoromethane	8.165	113	109371	52.450	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.900%	
50) Toluene-d8	10.565	98	370808	50.050	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.100%	
62) 4-Bromofluorobenzene	12.847	95	118170	46.627	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	93.260%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	4495	5.320	ug/l	94
44) Trichloroethene	9.359	130	2214	1.082	ug/l	95
95) Naphthalene	15.641	128	3659	0.758	ug/l #	83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085540.D
Acq On : 28 Jan 2025 15:49
Operator : JC\MD
Sample : Q1199-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-DUP-20250123


Quant Time: Jan 29 00:40:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085540.D
Acq: 28 Jan 2025 15:49

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-DUP-20250123







Tgt Ion:168 Resp: 162005
Ion Ratio Lower Upper
168 100
99 63.9 53.6 80.4



#16
Acetone
Concen: 5.320 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.017 min
Lab File: VN085540.D
Acq: 28 Jan 2025 15:49

Tgt Ion: 43 Resp: 4495
Ion Ratio Lower Upper
43 100
58 32.8 23.8 35.6





#33

1,2-Dichloroethane-d4

Concen: 55.645 ug/l

RT: 8.576 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085540.D

Acq: 28 Jan 2025 15:49

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-DUP-20250123

Tgt Ion: 65 Resp: 145516

Ion Ratio Lower Upper

65 100

67 50.6 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085540.D

Acq: 28 Jan 2025 15:49




Tgt Ion: 114 Resp: 300571


Ion Ratio Lower Upper

114 100

63 22.2 0.0 47.6

88 15.7 0.0 32.6





#35

Dibromofluoromethane

Concen: 52.450 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085540.D


Acq: 28 Jan 2025 15:49

Instrument :




MSVOA_N

ClientSampleId :

BP-VPB-192-DUP-20250123



Tgt Ion:	113	Resp:	109371
Ion	Ratio	Lower	Upper
113	100		
111	103.1	82.7	124.1
192	18.7	14.3	21.5



#44

Trichloroethene

Concen: 1.082 ug/l




RT: 9.359 min Scan# 1259


Delta R.T. 0.012 min

Lab File: VN085540.D

Acq: 28 Jan 2025 15:49

Tgt Ion:	130	Resp:	2214
Ion	Ratio	Lower	Upper
130	100		
95	92.9	0.0	195.8





#50

Toluene-d8

Concen: 50.050 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085540.D

Acq: 28 Jan 2025 15:49

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-192-DUP-20250123

Tgt Ion: 98 Resp: 370808

Ion Ratio Lower Upper

98 100

100 63.4 52.2 78.4



#62

4-Bromofluorobenzene

Concen: 46.627 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min

Lab File: VN085540.D

Acq: 28 Jan 2025 15:49




Tgt Ion: 95 Resp: 118170


Ion Ratio Lower Upper

95 100

174 74.9 0.0 145.0

176 73.0 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 16

Delta R.T. -0.000 min

Lab File: VN085540.D

Acq: 28 Jan 2025 15:49

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-192-DUP-20250123





Tgt Ion:117 Resp: 271698

Ion Ratio Lower Upper

117 100

82 56.7 48.6 72.8

119 31.5 26.6 39.8



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.006 min

Lab File: VN085540.D

Acq: 28 Jan 2025 15:49




Tgt Ion:152 Resp: 107447


Ion Ratio Lower Upper

152 100

115 61.4 31.1 93.3

150 160.0 0.0 343.6





#95

Naphthalene

Concen: 0.758 ug/l

RT: 15.641 min Scan# 23

Delta R.T. 0.006 min

Lab File: VN085540.D


Acq: 28 Jan 2025 15:49

Instrument :



MSVOA_N

ClientSampleId :

BP-VPB-192-DUP-20250123



Tgt Ion:	128	Resp:	3659
Ion	Ratio	Lower	Upper
128	100		
127	5.5	10.6	16.0#
129	5.7	8.8	13.2#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085540.D
 Acq On : 28 Jan 2025 15:49
 Operator : JC\MD
 Sample : Q1199-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-DUP-20250123

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Title : SW846 8260

Signal : TIC: VN085540.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.295	50	58	65	rBV5	4564	12896	1.28%	0.243%
2	8.165	1046	1056	1061	rBV2	160392	382091	37.99%	7.206%
3	8.224	1061	1066	1080	rVB	247782	520168	51.72%	9.810%
4	8.576	1117	1126	1139	rBV	175592	397042	39.48%	7.488%
5	9.100	1204	1215	1227	rBV	370302	749008	74.47%	14.126%
6	9.347	1251	1257	1263	rBV4	5500	10997	1.09%	0.207%
7	10.565	1456	1464	1473	rBV	554585	1005766	100.00%	18.969%
8	11.300	1584	1589	1594	rVB3	9816	16342	1.62%	0.308%
9	11.865	1677	1685	1698	rVB	505506	880668	87.56%	16.609%
10	12.847	1844	1852	1863	rVB	357308	608936	60.54%	11.485%
11	13.788	2005	2012	2023	rBV	408600	701074	69.71%	13.222%
12	14.553	2136	2142	2147	rBV4	11268	17228	1.71%	0.325%




Sum of corrected areas: 5302216

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085540.D
Acq On : 28 Jan 2025 15:49
Operator : JC\MD
Sample : Q1199-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-DUP-20250123

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085540.D
Acq On : 28 Jan 2025 15:49
Operator : JC\MD
Sample : Q1199-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-DUP-20250123

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085540.D
Acq On : 28 Jan 2025 15:49
Operator : JC\MD
Sample : Q1199-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-DUP-20250123

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc



CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199
 Instrument ID: MSVOA_N Calibration Date(s): 01/14/2025 01/14/2025
 Heated Purge: (Y/N) N Calibration Time(s): 14:56 17:19
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF100 = VN085438.D			RRF050 = VN085439.D		RRF020 = VN085440.D		
	RRF010 = VN085441.D			RRF005 = VN085442.D		RRF001 = VN085443.D		
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD
Chloromethane	0.680	0.680	0.727	0.693	0.779	0.839	0.733	8.8
Vinyl Chloride	0.697	0.686	0.727	0.711	0.781	0.819	0.737	7.1
Bromomethane	0.392	0.417	0.454	0.437	0.525		0.445	11.3
Chloroethane	0.435	0.424	0.468	0.429	0.505	0.542	0.467	10.3
Trichlorofluoromethane	1.046	0.997	1.097	1.040	1.077	1.157	1.069	5.1
1,1,2-Trichlorotrifluoroethane	0.590	0.542	0.609	0.587	0.639	0.646	0.602	6.4
1,1-Dichloroethene	0.548	0.533	0.556	0.526	0.559	0.497	0.537	4.3
Acetone	0.238	0.252	0.252	0.247	0.269	0.306	0.261	9.3
Carbon Disulfide	1.555	1.477	1.647	1.537	1.719	1.978	1.652	11
Methyl tert-butyl Ether	1.834	1.873	1.853	1.664	1.685	1.545	1.742	7.5
Methylene Chloride	0.629	0.629	0.658	0.606	0.696	0.656	0.646	4.9
trans-1,2-Dichloroethene	0.571	0.555	0.574	0.539	0.569	0.632	0.573	5.5
1,1-Dichloroethane	1.164	1.170	1.206	1.100	1.226	1.204	1.178	3.8
2-Butanone	0.378	0.390	0.398	0.363	0.387	0.386	0.384	3.1
Carbon Tetrachloride	0.574	0.530	0.579	0.529	0.565	0.567	0.557	4
cis-1,2-Dichloroethene	0.691	0.683	0.715	0.639	0.669	0.655	0.675	4
Chloroform	1.197	1.175	1.241	1.169	1.253	1.273	1.218	3.6
1,1,1-Trichloroethane	1.053	1.016	1.091	1.000	1.148	1.102	1.068	5.2
Methylcyclohexane	0.564	0.463	0.477	0.407	0.437	0.397	0.457	13.3
Benzene	1.551	1.449	1.527	1.376	1.474	1.400	1.463	4.7
1,2-Dichloroethane	0.569	0.547	0.575	0.522	0.574	0.517	0.551	4.8
Trichloroethene	0.362	0.324	0.352	0.310	0.343	0.352	0.341	5.8
1,2-Dichloropropane	0.390	0.371	0.388	0.334	0.388	0.371	0.374	5.7
Bromodichloromethane	0.590	0.559	0.579	0.514	0.569	0.484	0.549	7.5
4-Methyl-2-Pentanone	0.499	0.492	0.495	0.432	0.443	0.380	0.457	10.3
Toluene	0.964	0.870	0.919	0.808	0.835	0.690	0.848	11.3
t-1,3-Dichloropropene	0.594	0.551	0.544	0.481	0.527	0.416	0.519	12
cis-1,3-Dichloropropene	0.623	0.588	0.601	0.527	0.538	0.450	0.554	11.4
1,1,2-Trichloroethane	0.348	0.340	0.353	0.309	0.349	0.314	0.335	5.7
2-Hexanone	0.358	0.357	0.353	0.298	0.302	0.261	0.321	12.6

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199
 Instrument ID: MSVOA_N Calibration Date(s): 01/14/2025 01/14/2025
 Heated Purge: (Y/N) N Calibration Time(s): 14:56 17:19
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085438.D		RRF050 = VN085439.D		RRF020 = VN085440.D			
		RRF010 = VN085441.D		RRF005 = VN085442.D		RRF001 = VN085443.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.430	0.414	0.412	0.368	0.420	0.386	0.405	5.8	
Tetrachloroethene	0.351	0.322	0.365	0.338	0.346	0.323	0.341	4.9	
Chlorobenzene	1.133	1.076	1.154	1.047	1.110	1.051	1.095	4	
Ethyl Benzene	2.072	1.867	1.940	1.685	1.709	1.430	1.784	12.7	
m/p-Xylenes	0.775	0.707	0.750	0.615	0.616	0.492	0.659	16	
o-Xylene	0.738	0.681	0.713	0.584	0.582	0.482	0.630	15.5	
Styrene	1.271	1.173	1.186	0.956	0.929	0.742	1.043	19.2	
Bromoform	0.311	0.311	0.312	0.273	0.284	0.235	0.288	10.6	
Isopropylbenzene	3.922	3.448	3.681	3.272	3.157	2.766	3.375	12.1	
1,1,2,2-Tetrachloroethane	1.121	1.145	1.187	1.157	1.228	1.314	1.192	5.9	
1,3-Dichlorobenzene	1.720	1.565	1.701	1.574	1.656	1.526	1.624	4.9	
1,4-Dichlorobenzene	1.706	1.562	1.713	1.607	1.743	1.767	1.683	4.8	
1,2-Dichlorobenzene	1.611	1.555	1.654	1.532	1.600	1.766	1.620	5.2	
1,2-Dichloroethane-d4	0.774	0.831	0.754	0.762	0.914		0.807	8.3	
Dibromofluoromethane	0.359	0.358	0.335	0.310	0.373		0.347	7.1	
Toluene-d8	1.339	1.267	1.207	1.076	1.274		1.232	8.1	
4-Bromofluorobenzene	0.475	0.449	0.410	0.357	0.417		0.422	10.6	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N011425W.M
 Title : SW846 8260
 Last Update : Wed Jan 15 02:16:08 2025
 Response Via : Initial Calibration

Calibration Files

1 =VN085443.D 5 =VN085442.D 10 =VN085441.D 20 =VN085440.D 50 =VN085439.D 100 =VN085438.

D

	Compound	1	5	10	20	50	100	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.714	0.708	0.667	0.681	0.629	0.664	0.677	4.59
3) P	Chloromethane	0.839	0.779	0.693	0.727	0.680	0.680	0.733	8.77
4) C	Vinyl Chloride	0.819	0.781	0.711	0.727	0.686	0.697	0.737	7.06#
5) T	Bromomethane	0.525	0.437	0.454	0.417	0.392	0.445		11.26
6) T	Chloroethane	0.542	0.505	0.429	0.468	0.424	0.435	0.467	10.27
7) T	Trichlorofluor...	1.157	1.077	1.040	1.097	0.997	1.046	1.069	5.14
8) T	Diethyl Ether	0.457	0.357	0.317	0.363	0.360	0.362	0.369	12.59
9) T	1,1,2-Trichlor...	0.646	0.639	0.587	0.609	0.542	0.590	0.602	6.35
10) T	Methyl Iodide	0.661	0.640	0.704	0.719	0.723	0.690		5.35
11) T	Tert butyl alc...	0.097	0.092	0.095	0.093	0.085	0.092		4.80
12) CM	1,1-Dichloroet...	0.497	0.559	0.526	0.556	0.533	0.548	0.537	4.31#
13) T	Acrolein	0.101	0.131	0.124	0.140	0.136	0.126		12.09
14) T	Allyl chloride	0.952	0.897	0.823	0.843	0.846	0.863	0.871	5.42
15) T	Acrylonitrile	0.298	0.299	0.275	0.303	0.298	0.289	0.294	3.39
16) T	Acetone	0.306	0.269	0.247	0.252	0.252	0.238	0.261	9.28
17) T	Carbon Disulfide	1.978	1.719	1.537	1.647	1.477	1.555	1.652	10.95
18) T	Methyl Acetate	0.871	0.810	0.779	0.758	0.790	0.751	0.793	5.54
19) T	Methyl tert-bu...	1.545	1.685	1.664	1.853	1.873	1.834	1.742	7.53
20) T	Methylene Chlo...	0.656	0.696	0.606	0.658	0.629	0.629	0.646	4.86
21) T	trans-1,2-Dich...	0.632	0.569	0.539	0.574	0.555	0.571	0.573	5.50
22) T	Diisopropyl ether	1.705	1.918	1.834	2.076	2.037	2.026	1.933	7.38
23) T	Vinyl Acetate	1.110	1.270	1.279	1.466	1.495	1.495	1.353	11.67
24) P	1,1-Dichloroet...	1.204	1.226	1.100	1.206	1.170	1.164	1.178	3.81
25) T	2-Butanone	0.386	0.387	0.363	0.398	0.390	0.378	0.384	3.12
26) T	2,2-Dichloropr...	0.930	0.986	0.922	0.985	0.936	0.958	0.953	2.96
27) T	cis-1,2-Dichlo...	0.655	0.669	0.639	0.715	0.683	0.691	0.675	4.02
28) T	Bromochloromet...	0.624	0.595	0.486	0.513	0.542	0.530	0.548	9.44
29) T	Tetrahydrofuran	0.221	0.236	0.233	0.261	0.260	0.248	0.243	6.58
30) C	Chloroform	1.273	1.253	1.169	1.241	1.175	1.197	1.218	3.59#
31) T	Cyclohexane	1.198	1.026	1.033	0.881	0.984	1.024		11.18
32) T	1,1,1-Trichlor...	1.102	1.148	1.000	1.091	1.016	1.053	1.068	5.24
33) S	1,2-Dichloroet...	0.914	0.762	0.754	0.831	0.774	0.807		8.28
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...	0.373	0.310	0.335	0.358	0.359	0.347		7.10
36) T	1,1-Dichloropr...	0.509	0.490	0.441	0.499	0.462	0.521	0.487	6.17
37) T	Ethyl Acetate	0.504	0.520	0.452	0.488	0.488	0.496	0.491	4.66
38) T	Carbon Tetrach...	0.567	0.565	0.529	0.579	0.530	0.574	0.557	3.96
39) T	Methylcyclohexane	0.397	0.437	0.407	0.477	0.463	0.564	0.457	13.29
40) TM	Benzene	1.400	1.474	1.376	1.527	1.449	1.551	1.463	4.70
41) T	Methacrylonitrile	0.241	0.227	0.248	0.268	0.270	0.280	0.256	7.89
42) TM	1,2-Dichloroet...	0.517	0.574	0.522	0.575	0.547	0.569	0.551	4.78
43) T	Isopropyl Acetate	0.793	0.764	0.719	0.801	0.813	0.835	0.787	5.20
44) TM	Trichloroethene	0.352	0.343	0.310	0.352	0.324	0.362	0.341	5.79
45) C	1,2-Dichloropr...	0.371	0.388	0.334	0.388	0.371	0.390	0.374	5.69#
46) T	Dibromomethane	0.292	0.254	0.254	0.277	0.267	0.275	0.270	5.50
47) T	Bromodichlorom...	0.484	0.569	0.514	0.579	0.559	0.590	0.549	7.48
48) T	Methyl methacr...	0.311	0.326	0.321	0.369	0.392	0.406	0.354	11.37
49) T	1,4-Dioxane	0.006	0.006	0.006	0.006	0.006	0.006	0.006	6.72
50) S	Toluene-d8	1.274	1.076	1.207	1.267	1.339	1.232		8.07
51) T	4-Methyl-2-Pen...	0.380	0.443	0.432	0.495	0.492	0.499	0.457	10.33
52) CM	Toluene	0.690	0.835	0.808	0.919	0.870	0.964	0.848	11.27#
53) T	t-1,3-Dichloro...	0.416	0.527	0.481	0.544	0.551	0.594	0.519	11.98
54) T	cis-1,3-Dichlo...	0.450	0.538	0.527	0.601	0.588	0.623	0.554	11.40
55) T	1,1,2-Trichlor...	0.314	0.349	0.309	0.353	0.340	0.348	0.335	5.67

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\										
Method File : 82N011425W.M										
56)	T	Ethyl methacry...	0.317	0.434	0.418	0.521	0.544	0.582	0.469	20.85
57)	T	1,3-Dichloropr...	0.514	0.584	0.555	0.618	0.602	0.627	0.583	7.31
58)	T	2-Chloroethyl ...	0.163	0.205	0.185	0.226	0.241	0.258	0.213	16.67
59)	T	2-Hexanone	0.261	0.302	0.298	0.353	0.357	0.358	0.321	12.60
60)	T	Dibromochlorom...	0.386	0.420	0.368	0.412	0.414	0.430	0.405	5.79
61)	T	1,2-Dibromoethane	0.334	0.321	0.309	0.356	0.334	0.349	0.334	5.16
62)	S	4-Bromofluorob...		0.417	0.357	0.410	0.449	0.475	0.422	10.57
-----ISTD-----										
63)	I	Chlorobenzene-d5								
64)	T	Tetrachloroethene	0.323	0.346	0.338	0.365	0.322	0.351	0.341	4.87
65)	PM	Chlorobenzene	1.051	1.110	1.047	1.154	1.076	1.133	1.095	4.03
66)	T	1,1,1,2-Tetrac...	0.425	0.412	0.371	0.408	0.392	0.404	0.402	4.65
67)	C	Ethyl Benzene	1.430	1.709	1.685	1.940	1.867	2.072	1.784	12.67#
68)	T	m/p-Xylenes	0.492	0.616	0.615	0.750	0.707	0.775	0.659	16.01
69)	T	o-Xylene	0.482	0.582	0.584	0.713	0.681	0.738	0.630	15.47
70)	T	Styrene	0.742	0.929	0.956	1.186	1.173	1.271	1.043	19.21
71)	P	Bromoform	0.235	0.284	0.273	0.312	0.311	0.311	0.288	10.61
-----ISTD-----										
72)	I	1,4-Dichlorobenzen...								
73)	T	Isopropylbenzene	2.766	3.157	3.272	3.681	3.448	3.922	3.375	12.06
74)	T	N-amyl acetate	1.364	1.394	1.405	1.642	1.622	1.673	1.517	9.43
75)	P	1,1,2,2-Tetrac...	1.314	1.228	1.157	1.187	1.145	1.121	1.192	5.89
76)	T	1,2,3-Trichlor...	1.104	0.965	0.957	1.118	0.893	1.061	1.016	8.96
77)	T	Bromobenzene	0.871	0.880	0.857	0.906	0.858	0.919	0.882	2.90
78)	T	n-propylbenzene	3.227	3.605	3.846	4.400	4.154	4.738	3.995	13.74
79)	T	2-Chlorotoluene	2.252	2.533	2.571	2.761	2.585	2.812	2.586	7.67
80)	T	1,3,5-Trimethy...	2.160	2.529	2.695	3.140	2.940	3.260	2.787	14.69
81)	T	trans-1,4-Dich...		0.335	0.347	0.386	0.390	0.419	0.376	9.04
82)	T	4-Chlorotoluene	2.144	2.536	2.512	2.820	2.588	2.847	2.574	9.91
83)	T	tert-Butylbenzene	1.885	2.189	2.236	2.570	2.439	2.726	2.341	12.85
84)	T	1,2,4-Trimethy...	1.888	2.542	2.710	3.190	3.016	3.323	2.778	18.89
85)	T	sec-Butylbenzene	2.318	3.026	3.147	3.643	3.421	3.912	3.244	17.17
86)	T	p-Isopropyltol...	1.777	2.368	2.541	2.942	2.872	3.286	2.631	20.04
87)	T	1,3-Dichlorobe...	1.526	1.656	1.574	1.701	1.565	1.720	1.624	4.91
88)	T	1,4-Dichlorobe...	1.767	1.743	1.607	1.713	1.562	1.706	1.683	4.78
89)	T	n-Butylbenzene	1.772	2.176	2.168	2.485	2.483	2.881	2.327	16.22
90)	T	Hexachloroethane	0.681	0.616	0.574	0.613	0.577	0.645	0.618	6.59
91)	T	1,2-Dichlorobe...	1.766	1.600	1.532	1.654	1.555	1.611	1.620	5.15
92)	T	1,2-Dibromo-3-...	0.202	0.228	0.212	0.224	0.222	0.218	0.218	4.31
93)	T	1,2,4-Trichlor...	0.658	0.717	0.704	0.799	0.781	0.858	0.753	9.69
94)	T	Hexachlorobuta...	0.413	0.441	0.395	0.396	0.367	0.386	0.400	6.31
95)	T	Naphthalene	2.115	1.987	1.958	2.348	2.482	2.588	2.246	11.79
96)	T	1,2,3-Trichlor...	0.750	0.693	0.732	0.792	0.786	0.817	0.762	5.96

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085438.D
 Acq On : 14 Jan 2025 14:56
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:41:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	202613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	329363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146196	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	313645	95.899	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	191.800%#
35) Dibromofluoromethane	8.159	113	236159	103.353	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	206.700%#
50) Toluene-d8	10.565	98	881956	108.636	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	217.280%#
62) 4-Bromofluorobenzene	12.847	95	313161	112.765	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	225.520%#

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	269023	98.065	ug/l	99
3) Chloromethane	2.359	50	275415	92.724	ug/l	98
4) Vinyl Chloride	2.512	62	282361	94.576	ug/l	99
5) Bromomethane	2.942	94	158888	88.105	ug/l	98
6) Chloroethane	3.112	64	176111	93.043	ug/l	99
7) Trichlorofluoromethane	3.494	101	423967	97.870	ug/l	100
8) Diethyl Ether	3.953	74	146889	98.154	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	239104	97.997	ug/l	99
10) Methyl Iodide	4.589	142	292983	104.856	ug/l	95
11) Tert butyl alcohol	5.512	59	173044	462.061	ug/l	99
12) 1,1-Dichloroethene	4.336	96	222117	102.146	ug/l	99
13) Acrolein	4.171	56	275350	538.600	ug/l	99
14) Allyl chloride	5.018	41	349717	99.111	ug/l	98
15) Acrylonitrile	5.712	53	585370	492.095	ug/l	99
16) Acetone	4.418	43	483221	457.268	ug/l	96
17) Carbon Disulfide	4.712	76	630294	94.140	ug/l	100
18) Methyl Acetate	5.012	43	304226	94.674	ug/l	99
19) Methyl tert-butyl Ether	5.788	73	743008	105.230	ug/l	100
20) Methylene Chloride	5.271	84	254912	97.440	ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	231503	99.618	ug/l	97
22) Diisopropyl ether	6.665	45	820871	104.812	ug/l	99
23) Vinyl Acetate	6.594	43	3029568	550.850	ug/l	99
24) 1,1-Dichloroethane	6.565	63	471831	98.803	ug/l	99
25) 2-Butanone	7.477	43	764979	491.912	ug/l	98
26) 2,2-Dichloropropane	7.482	77	388140	100.530	ug/l	100
27) cis-1,2-Dichloroethene	7.482	96	280197	102.367	ug/l	99
28) Bromochloromethane	7.806	49	214796	96.679	ug/l	100
29) Tetrahydrofuran	7.830	42	503459	510.635	ug/l	99
30) Chloroform	7.959	83	484941	98.253	ug/l	100
31) Cyclohexane	8.253	56	398641	96.041	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	426782	98.578	ug/l	98
36) 1,1-Dichloropropene	8.365	75	342961	106.945	ug/l	100
37) Ethyl Acetate	7.553	43	326802	100.961	ug/l	97
38) Carbon Tetrachloride	8.359	117	378390	103.065	ug/l	97
39) Methylcyclohexane	9.594	83	371734	123.354	ug/l	99
40) Benzene	8.600	78	1021393	106.001	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085438.D
 Acq On : 14 Jan 2025 14:56
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:41:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	184326	109.454	ug/l	96
42) 1,2-Dichloroethane	8.665	62	374938	103.314	ug/l	100
43) Isopropyl Acetate	8.682	43	549969	106.035	ug/l	100
44) Trichloroethene	9.347	130	238296	106.242	ug/l	96
45) 1,2-Dichloropropane	9.618	63	256964	104.405	ug/l	97
46) Dibromomethane	9.706	93	180990	101.913	ug/l	99
47) Bromodichloromethane	9.882	83	388499	107.377	ug/l	98
48) Methyl methacrylate	9.676	41	267689	114.685	ug/l	99
49) 1,4-Dioxane	9.688	88	83879	2133.992	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	1643800	546.159	ug/l	100
52) Toluene	10.623	92	634743	113.689	ug/l	99
53) t-1,3-Dichloropropene	10.829	75	391238	114.424	ug/l	97
54) cis-1,3-Dichloropropene	10.306	75	410403	112.369	ug/l	99
55) 1,1,2-Trichloroethane	11.012	97	229522	103.879	ug/l	98
56) Ethyl methacrylate	10.870	69	383663	101.029	ug/l	99
57) 1,3-Dichloropropane	11.159	76	412886	107.469	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.153	63	848636	605.243	ug/l	100
59) 2-Hexanone	11.194	43	1178831	556.643	ug/l	99
60) Dibromochloromethane	11.353	129	283361	106.227	ug/l	99
61) 1,2-Dibromoethane	11.465	107	230104	104.623	ug/l	98
64) Tetrachloroethene	11.100	164	209870	102.980	ug/l	98
65) Chlorobenzene	11.888	112	677679	103.482	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	241593	100.517	ug/l	99
67) Ethyl Benzene	11.959	91	1239080	116.169	ug/l	99
68) m/p-Xylenes	12.070	106	927244	235.219	ug/l	99
69) o-Xylene	12.394	106	441279	117.135	ug/l	100
70) Styrene	12.406	104	760136	121.919	ug/l	99
71) Bromoform	12.576	173	186124	108.227	ug/l #	99
73) Isopropylbenzene	12.694	105	1146752	116.221	ug/l	100
74) N-amyl acetate	12.488	43	489217	110.308	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	327767	94.039	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	310115m	104.530	ug/l	
77) Bromobenzene	12.976	156	268616	104.200	ug/l	97
78) n-propylbenzene	13.029	91	1385353	118.600	ug/l	100
79) 2-Chlorotoluene	13.123	91	822258	108.767	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	953066	116.938	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	122508	111.544	ug/l	87
82) 4-Chlorotoluene	13.217	91	832317	110.574	ug/l	99
83) tert-Butylbenzene	13.435	119	797069	116.457	ug/l	98
84) 1,2,4-Trimethylbenzene	13.476	105	971729	119.625	ug/l	100
85) sec-Butylbenzene	13.611	105	1143885	120.587	ug/l	100
86) p-Isopropyltoluene	13.723	119	960747	100.111	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	502960	105.943	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	498709	101.341	ug/l	97
89) n-Butylbenzene	14.053	91	842368	123.781	ug/l	98
90) Hexachloroethane	14.329	117	188541	104.405	ug/l	100
91) 1,2-Dichlorobenzene	14.100	146	470919	99.445	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	63846	100.313	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	250787	113.946	ug/l	99
94) Hexachlorobutadiene	15.500	225	112899	96.619	ug/l	99
95) Naphthalene	15.635	128	756816	115.234	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	238860	107.256	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085438.D
Acq On : 14 Jan 2025 14:56
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085438.D
 Acq On : 14 Jan 2025 14:56
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Manual Integrations

APPROVED

Reviewed By : John Carlone 01/15/2025

Supervised By : Mahesh Dadoda 01/15/2025


Quant Time: Jan 15 01:41:17 2025


Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M

Quant Title : SW846 8260

QLast Update : Wed Jan 15 01:28:51 2025

Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:168 Resp: 202613
Ion Ratio Lower Upper
168 100
99 64.9 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#2
Dichlorodifluoromethane
Concen: 98.065 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 85 Resp: 269023
Ion Ratio Lower Upper
85 100
87 32.7 16.7 50.1





#3

Chloromethane

Concen: 92.724 ug/l

RT: 2.359 min Scan# 69

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 50 Resp: 275415

Ion Ratio Lower Upper

50 100





52 32.4 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#4

Vinyl Chloride

Concen: 94.576 ug/l

RT: 2.512 min Scan# 95

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 62 Resp: 282361

Ion Ratio Lower Upper

62 100

64 31.6 24.8 37.2





#5

Bromomethane

Concen: 88.105 ug/l

RT: 2.942 min Scan# 16

Delta R.T. -0.012 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 94 Resp: 158888

Ion Ratio Lower Upper

94 100





96 91.9 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#6

Chloroethane

Concen: 93.043 ug/l

RT: 3.112 min Scan# 197

Delta R.T. -0.006 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 64 Resp: 176111

Ion Ratio Lower Upper

64 100

66 31.1 24.6 36.8





#7

Trichlorofluoromethane

Concen: 97.870 ug/l

RT: 3.494 min Scan# 26

Delta R.T. -0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:101 Resp: 423967

Ion Ratio Lower Upper

101 100

103 64.4 51.4 77.2

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#8

Diethyl Ether

Concen: 98.154 ug/l

RT: 3.953 min Scan# 340

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 74 Resp: 146889

Ion Ratio Lower Upper

74 100


45 97.4 49.7 149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 97.997 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:101 Resp: 239104
Ion Ratio Lower Upper
101 100
85 47.1 37.8 56.8
151 71.8 58.8 88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#10
Methyl Iodide
Concen: 104.856 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:142 Resp: 292983
Ion Ratio Lower Upper
142 100
127 47.1 40.3 60.5
141 13.1 12.8 19.2





#11

Tert butyl alcohol

Concen: 462.061 ug/l

RT: 5.512 min Scan# 605

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 59 Resp: 173044

Ion Ratio Lower Upper

59 100





57 11.1 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#12

1,1-Dichloroethene

Concen: 102.146 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 96 Resp: 222117


Ion Ratio Lower Upper

96 100

61 175.3 141.0 211.4

98 62.7 49.0 73.6





#13

Acrolein

Concen: 538.600 ug/l

RT: 4.171 min Scan# 377

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 275350

Ion Ratio Lower Upper

56 100





55 71.1 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#14

Allyl chloride

Concen: 99.111 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 41 Resp: 349717


Ion Ratio Lower Upper

41 100

39 83.3 64.4 96.6

76 37.9 30.5 45.7





#15

Acrylonitrile

Concen: 492.095 ug/l

RT: 5.712 min Scan# 639

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC100

Tgt Ion:	53	Resp:	585370
Ion	Ratio	Lower	Upper
53	100		
52	83.0	65.5	98.3
51	37.1	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#16

Acetone

Concen: 457.268 ug/l




RT: 4.418 min Scan# 419


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	43	Resp:	483221
Ion	Ratio	Lower	Upper
43	100		
58	27.3	23.8	35.6





#17

Carbon Disulfide

Concen: 94.140 ug/l

RT: 4.712 min Scan# 46

Delta R.T. 0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 630294

Ion Ratio Lower Upper

76 100





78 8.7 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#18

Methyl Acetate

Concen: 94.674 ug/l

RT: 5.012 min Scan# 520

Delta R.T. -0.006 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 43 Resp: 304226

Ion Ratio Lower Upper

43 100

74 23.6 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 105.230 ug/l

RT: 5.788 min Scan# 65

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 73 Resp: 743008

Ion Ratio Lower Upper

73 100





57 22.3 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#20

Methylene Chloride

Concen: 97.440 ug/l

RT: 5.271 min Scan# 564

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion: 84 Resp: 254912




Ion Ratio Lower Upper


84 100

49 128.5 104.1 156.1

51 37.8 31.0 46.4

86 61.1 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 99.618 ug/l

RT: 5.783 min Scan# 651

Delta R.T. -0.000 min

Lab File: VN085438.D


Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100



Tgt Ion: 96 Resp: 231503

Ion Ratio Lower Upper

96 100

61 148.3 115.7 173.5




98 65.0 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether

Concen: 104.812 ug/l

RT: 6.665 min Scan# 801

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion: 45 Resp: 820871




Ion Ratio Lower Upper


45 100

43 55.7 45.0 67.6

87 27.2 21.8 32.8

59 11.6 10.2 15.2





#23

Vinyl Acetate

Concen: 550.850 ug/l

RT: 6.594 min Scan# 789

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 3029568

Ion Ratio Lower Upper

43 100





86 10.1 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 98.803 ug/l

RT: 6.565 min Scan# 784

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 63 Resp: 471831


Ion Ratio Lower Upper

63 100

98 6.4 3.3 9.8

100 4.6 2.0 6.0





#25

2-Butanone

Concen: 491.912 ug/l

RT: 7.477 min Scan# 939

Delta R.T. -0.000 min

Lab File: VN085438.D


Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100



Tgt Ion: 43 Resp: 764979

Ion Ratio Lower Upper

43 100




72 26.3 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#26

2,2-Dichloropropane


Concen: 100.530 ug/l

RT: 7.482 min Scan# 940

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56






Tgt Ion: 77 Resp: 388140

Ion Ratio Lower Upper

77 100


97 21.4 10.7 32.1





#27
cis-1,2-Dichloroethene
Concen: 102.367 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 280197
Ion Ratio Lower Upper
96 100
61 156.8 0.0 311.8
98 65.0 0.0 126.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#28
Bromochloromethane
Concen: 96.679 ug/l
RT: 7.806 min Scan# 995
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 49 Resp: 214796
Ion Ratio Lower Upper
49 100
129 1.9 0.0 4.4
130 68.4 55.0 82.4





#29

Tetrahydrofuran

Concen: 510.635 ug/l

RT: 7.830 min Scan# 999

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 42 Resp: 503459

Ion Ratio Lower Upper

42 100

72 43.1 34.2 51.2





71 40.9 32.5 48.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#30

Chloroform

Concen: 98.253 ug/l

RT: 7.959 min Scan# 1021

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 83 Resp: 484941

Ion Ratio Lower Upper

83 100

85 65.0 51.8 77.6





#31

Cyclohexane

Concen: 96.041 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 56 Resp: 398641

Ion Ratio Lower Upper

56 100

69 29.2 22.2 33.4





84 86.4 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#32

1,1,1-Trichloroethane

Concen: 98.578 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 97 Resp: 426782


Ion Ratio Lower Upper

97 100

99 64.2 49.8 74.6

61 50.6 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 95.899 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 65 Resp: 313645

Ion Ratio Lower Upper

65 100





67 51.1 0.0 101.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:114 Resp: 329363


Ion Ratio Lower Upper

114 100

63 21.2 0.0 47.6

88 16.2 0.0 32.6





#35

Dibromofluoromethane

Concen: 103.353 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:113 Resp: 236159

Ion Ratio Lower Upper

113 100

111 102.3 82.7 124.1





192 17.4 14.3 21.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#36

1,1-Dichloropropene

Concen: 106.945 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 75 Resp: 342961


Ion Ratio Lower Upper

75 100

110 32.6 16.5 49.5

77 30.6 24.4 36.6





#37

Ethyl Acetate

Concen: 100.961 ug/l

RT: 7.553 min Scan# 95

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 326802

Ion Ratio Lower Upper

43 100

61 12.0 10.9 16.3





70 9.6 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#38

Carbon Tetrachloride

Concen: 103.065 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:117 Resp: 378390


Ion Ratio Lower Upper

117 100

119 98.1 75.4 113.2

121 30.2 24.6 37.0





#39

Methylcyclohexane

Concen: 123.354 ug/l

RT: 9.594 min Scan# 12

Delta R.T. -0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC100

Tgt Ion:	83	Resp:	371734
Ion	Ratio	Lower	Upper
83	100		
55	79.3	62.6	94.0
98	46.7	37.7	56.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#40

Benzene

Concen: 106.001 ug/l




RT: 8.600 min Scan# 1130


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	78	Resp:	1021393
Ion	Ratio	Lower	Upper
78	100		
77	23.9	19.0	28.6





#41

Methacrylonitrile

Concen: 109.454 ug/l

RT: 7.777 min Scan# 99

Delta R.T. 0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 41 Resp: 184326

Ion Ratio Lower Upper

41 100

39 52.8 46.0 69.0

67 68.8 57.4 86.2





52 32.3 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#42

1,2-Dichloroethane

Concen: 103.314 ug/l

RT: 8.665 min Scan# 1141

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 62 Resp: 374938

Ion Ratio Lower Upper

62 100

98 8.4 0.0 17.0





#43

Isopropyl Acetate

Concen: 106.035 ug/l

RT: 8.682 min Scan# 1144

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 549969

Ion Ratio Lower Upper

43 100

61 25.7 20.7 31.1





87 12.2 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#44

Trichloroethene

Concen: 106.242 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion:130 Resp: 238296

Ion Ratio Lower Upper

130 100

95 101.5 0.0 195.8





#45

1,2-Dichloropropane

Concen: 104.405 ug/l

RT: 9.618 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 256964

Ion Ratio Lower Upper

63 100





65 30.3 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#46

Dibromomethane

Concen: 101.913 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 93 Resp: 180990


Ion Ratio Lower Upper

93 100

95 81.6 64.7 97.1

174 87.5 69.0 103.4





#47

Bromodichloromethane

Concen: 107.377 ug/l

RT: 9.882 min Scan# 1348

Delta R.T. 0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC100

Tgt Ion:	83	Resp:	388499
Ion	Ratio	Lower	Upper
83	100		
85	65.9	51.2	76.8
127	8.4	6.5	9.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#48

Methyl methacrylate

Concen: 114.685 ug/l




RT: 9.676 min Scan# 1313


Delta R.T. -0.001 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	41	Resp:	267689
Ion	Ratio	Lower	Upper
41	100		
69	80.9	64.7	97.1
39	63.1	49.0	73.6





#49

1,4-Dioxane

Concen: 2133.992 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 88 Resp: 83879

Ion Ratio Lower Upper

88 100

43 31.3 26.6 39.8





58 72.1 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#50

Toluene-d8

Concen: 108.636 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 98 Resp: 881956

Ion Ratio Lower Upper

98 100

100 64.5 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 546.159 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1643800
Ion Ratio Lower Upper
43 100
58 38.4 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 113.689 ug/l
RT: 10.623 min Scan# 1474
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 92 Resp: 634743
Ion Ratio Lower Upper
92 100
91 173.1 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 114.424 ug/l

RT: 10.829 min Scan# 1510

Delta R.T. -0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 75 Resp: 391238

Ion Ratio Lower Upper

75 100





77 31.0 23.5 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#54

cis-1,3-Dichloropropene

Concen: 112.369 ug/l

RT: 10.306 min Scan# 1420

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 75 Resp: 410403


Ion Ratio Lower Upper

75 100

77 31.9 25.0 37.4

39 54.3 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 103.879 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 229522

Ion Ratio Lower Upper

97 100

83 88.4 69.0 103.6

85 56.3 44.6 66.8





99 62.9 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#56

Ethyl methacrylate

Concen: 101.029 ug/l

RT: 10.870 min Scan# 1516

Delta R.T. -0.001 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 69 Resp: 383663


Ion Ratio Lower Upper

69 100

41 69.1 54.6 82.0

39 40.1 32.4 48.6





#57

1,3-Dichloropropane

Concen: 107.469 ug/l

RT: 11.159 min Scan# 15

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 412886

Ion Ratio Lower Upper

76 100





78 32.2 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#58

2-Chloroethyl Vinyl ether

Concen: 605.243 ug/l

RT: 10.153 min Scan# 1394

Delta R.T. -0.006 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 63 Resp: 848636

Ion Ratio Lower Upper

63 100

106 26.8 21.6 32.4





#59

2-Hexanone

Concen: 556.643 ug/l

RT: 11.194 min Scan# 1570

Delta R.T. 0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 1178831

Ion Ratio Lower Upper

43 100





58 53.2 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#60

Dibromochloromethane

Concen: 106.227 ug/l

RT: 11.353 min Scan# 1598

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion:129 Resp: 283361

Ion Ratio Lower Upper

129 100

127 78.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 104.623 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 107 Resp: 230104
Ion Ratio Lower Upper
107 100
109 93.1 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#62
4-Bromofluorobenzene
Concen: 112.765 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 95 Resp: 313161
Ion Ratio Lower Upper
95 100
174 73.4 0.0 145.0
176 70.5 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:117 Resp: 298938

Ion Ratio Lower Upper

117 100

82 59.3 48.6 72.8





119 31.6 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#64

Tetrachloroethene

Concen: 102.980 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:164 Resp: 209870




Ion Ratio Lower Upper


164 100

166 126.8 103.4 155.2

129 98.3 79.2 118.8

131 94.7 77.1 115.7





#65

Chlorobenzene

Concen: 103.482 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:112 Resp: 677679

Ion Ratio Lower Upper

112 100





114 32.5 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 100.517 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:131 Resp: 241593


Ion Ratio Lower Upper

131 100

133 94.3 47.4 142.3

119 64.9 33.1 99.5





#67

Ethyl Benzene

Concen: 116.169 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 1239080

Ion Ratio Lower Upper

91 100





106 30.3 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#68

m/p-Xylenes

Concen: 235.219 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. 0.006 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion:106 Resp: 927244

Ion Ratio Lower Upper

106 100

91 208.0 167.7 251.5





#69

o-Xylene

Concen: 117.135 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:106 Resp: 441279

Ion Ratio Lower Upper

106 100





91 221.3 110.4 331.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#70

Styrene

Concen: 121.919 ug/l

RT: 12.406 min Scan# 1777

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:104 Resp: 760136


Ion Ratio Lower Upper

104 100

78 52.8 42.5 63.7

103 55.3 43.8 65.8





#71

Bromoform

Concen: 108.227 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion:	173	Resp:	186124
Ion Ratio	Lower	Upper	
173	100		
175	48.4	24.4	73.2
254	0.1	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	152	Resp:	146196
Ion Ratio	Lower	Upper	
152	100		
115	62.7	31.1	93.3
150	193.1	0.0	343.6





#73

Isopropylbenzene

Concen: 116.221 ug/l

RT: 12.694 min Scan# 1825

Delta R.T. 0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:105 Resp: 1146752

Ion Ratio Lower Upper

105 100





120 25.7 12.8 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#74

N-ethyl acetate

Concen: 110.308 ug/l

RT: 12.488 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 489217




Ion Ratio Lower Upper


43 100

70 41.9 34.0 51.0

55 25.7 21.4 32.2

61 23.5 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 94.039 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion:	83	Resp:	327767
Ion	Ratio	Lower	Upper
83	100		
131	9.8	4.8	14.4
85	64.1	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#76

1,2,3-Trichloropropane

Concen: 104.530 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	75	Resp:	310115
Ion	Ratio	Lower	Upper
75	100		
77	192.6	109.7	329.2





#77

Bromobenzene

Concen: 104.200 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:156 Resp: 268616

Ion Ratio Lower Upper

156 100

77 222.4 114.1 342.4





158 96.3 48.9 146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#78

n-propylbenzene

Concen: 118.600 ug/l

RT: 13.029 min Scan# 1883

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 91 Resp: 1385353

Ion Ratio Lower Upper

91 100

120 21.8 10.9 32.6





#79

2-Chlorotoluene

Concen: 108.767 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 822258

Ion Ratio Lower Upper

91 100





126 31.8 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#80

1,3,5-Trimethylbenzene

Concen: 116.938 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion:105 Resp: 953066

Ion Ratio Lower Upper

105 100

120 47.8 23.9 71.7





#81

trans-1,4-Dichloro-2-butene

Concen: 111.544 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 75 Resp: 122508

Ion Ratio Lower Upper

75 100

53 100.9 95.6 143.4





89 44.4 37.0 55.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#82

4-Chlorotoluene

Concen: 110.574 ug/l

RT: 13.217 min Scan# 1915

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 91 Resp: 832317

Ion Ratio Lower Upper

91 100

126 32.5 15.9 47.7





#83

tert-Butylbenzene

Concen: 116.457 ug/l

RT: 13.435 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:119 Resp: 797069

Ion Ratio Lower Upper

119 100

91 70.7 35.1 105.3





134 26.6 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#84

1,2,4-Trimethylbenzene

Concen: 119.625 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion:105 Resp: 971729

Ion Ratio Lower Upper

105 100

120 43.4 21.6 65.0





#85

sec-Butylbenzene

Concen: 120.587 ug/l

RT: 13.611 min Scan# 19

Delta R.T. -0.001 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:105 Resp: 1143885

Ion Ratio Lower Upper

105 100





134 19.2 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#86

p-Isopropyltoluene

Concen: 100.111 ug/l

RT: 13.723 min Scan# 2001

Delta R.T. 0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:119 Resp: 960747


Ion Ratio Lower Upper

119 100

134 25.3 12.7 38.0

91 25.0 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 105.943 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:146 Resp: 502960

Ion Ratio Lower Upper

146 100

111 42.8 21.4 64.3





148 63.6 32.3 96.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#88

1,4-Dichlorobenzene

Concen: 101.341 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:146 Resp: 498709


Ion Ratio Lower Upper

146 100

111 40.1 21.3 63.7

148 63.1 32.4 97.0





#89

n-Butylbenzene

Concen: 123.781 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC100

Tgt Ion:	91	Resp:	842368
Ion	Ratio	Lower	Upper
91	100		
92	52.8	25.8	77.3
134	24.3	11.7	35.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90

Hexachloroethane

Concen: 104.405 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	117	Resp:	188541
Ion	Ratio	Lower	Upper
117	100		
201	67.2	33.7	101.0









#91
1,2-Dichlorobenzene
Concen: 99.445 ug/l
RT: 14.100 min Scan# 2065
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:146 Resp: 470919
Ion Ratio Lower Upper
146 100
111 43.9 21.7 65.1
148 64.0 31.4 94.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 100.313 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 63846
Ion Ratio Lower Upper
75 100
155 73.5 36.4 109.2
157 93.0 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 113.946 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:180 Resp: 250787

Ion Ratio Lower Upper

180 100

182 94.9 46.8 140.3





145 32.6 16.0 48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene

Concen: 96.619 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:225 Resp: 112899


Ion Ratio Lower Upper

225 100

223 61.8 30.7 92.1

227 63.6 30.9 92.5





#95

Naphthalene

Concen: 115.234 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:128 Resp: 756816

Ion Ratio Lower Upper

128 100

127 13.1 10.6 16.0





129 10.8 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#96

1,2,3-Trichlorobenzene

Concen: 107.256 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:180 Resp: 238860

Ion Ratio Lower Upper

180 100

182 95.0 47.4 142.2

145 34.3 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085439.D
 Acq On : 14 Jan 2025 15:19
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:42:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	199403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	339872	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146624	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	165768	51.500	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.000%
35) Dibromofluoromethane	8.159	113	121554	51.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.100%
50) Toluene-d8	10.565	98	430698	51.411	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.820%
62) 4-Bromofluorobenzene	12.847	95	152568	53.239	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.480%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	125504	46.486	ug/l	100
3) Chloromethane	2.359	50	135631	46.398	ug/l	100
4) Vinyl Chloride	2.512	62	136831	46.569	ug/l	100
5) Bromomethane	2.953	94	83199	46.877	ug/l	100
6) Chloroethane	3.118	64	84517	45.371	ug/l	100
7) Trichlorofluoromethane	3.501	101	198726	46.613	ug/l	100
8) Diethyl Ether	3.953	74	71764	48.726	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	108081	45.010	ug/l	100
10) Methyl Iodide	4.589	142	143383	52.142	ug/l	100
11) Tert butyl alcohol	5.512	59	92841	251.894	ug/l	100
12) 1,1-Dichloroethene	4.342	96	106247	49.647	ug/l	100
13) Acrolein	4.171	56	139234	276.734	ug/l	100
14) Allyl chloride	5.018	41	168654	48.567	ug/l	100
15) Acrylonitrile	5.712	53	296781	253.507	ug/l	100
16) Acetone	4.418	43	251271	241.603	ug/l	100
17) Carbon Disulfide	4.706	76	294604	44.710	ug/l	100
18) Methyl Acetate	5.018	43	157485	49.798	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	373561	53.758	ug/l	100
20) Methylene Chloride	5.271	84	125508	48.748	ug/l	100
21) trans-1,2-Dichloroethene	5.783	96	110684	48.395	ug/l	100
22) Diisopropyl ether	6.665	45	406253	52.707	ug/l	100
23) Vinyl Acetate	6.594	43	1491031	275.471	ug/l	100
24) 1,1-Dichloroethane	6.565	63	233336	49.648	ug/l	100
25) 2-Butanone	7.477	43	388457	253.815	ug/l	100
26) 2,2-Dichloropropane	7.483	77	186591	49.106	ug/l	100
27) cis-1,2-Dichloroethene	7.483	96	136229	50.571	ug/l	100
28) Bromochloromethane	7.806	49	108148	49.461	ug/l	100
29) Tetrahydrofuran	7.830	42	259683	267.624	ug/l	100
30) Chloroform	7.959	83	234271	48.229	ug/l	100
31) Cyclohexane	8.253	56	175588	42.984	ug/l	100
32) 1,1,1-Trichloroethane	8.165	97	202590	47.547	ug/l	100
36) 1,1-Dichloropropene	8.365	75	157031	47.453	ug/l	100
37) Ethyl Acetate	7.553	43	165866	49.658	ug/l	100
38) Carbon Tetrachloride	8.359	117	180199	47.564	ug/l	100
39) Methylcyclohexane	9.600	83	157234	50.562	ug/l	100
40) Benzene	8.600	78	492606	49.542	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085439.D
 Acq On : 14 Jan 2025 15:19
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:42:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	91711	52.775	ug/l	100
42) 1,2-Dichloroethane	8.665	62	186065	49.685	ug/l	100
43) Isopropyl Acetate	8.683	43	276228	51.611	ug/l	100
44) Trichloroethene	9.347	130	109953	47.506	ug/l	100
45) 1,2-Dichloropropane	9.618	63	126084	49.644	ug/l	100
46) Dibromomethane	9.706	93	90670	49.476	ug/l	100
47) Bromodichloromethane	9.882	83	189915	50.867	ug/l	100
48) Methyl methacrylate	9.677	41	133399	55.385	ug/l	100
49) 1,4-Dioxane	9.688	88	43048	1061.334	ug/l	100
51) 4-Methyl-2-Pentanone	10.441	43	836707	269.403	ug/l	100
52) Toluene	10.624	92	295792	51.341	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	187439	53.125	ug/l	100
54) cis-1,3-Dichloropropene	10.306	75	199771	53.006	ug/l	100
55) 1,1,2-Trichloroethane	11.012	97	115462	50.641	ug/l	100
56) Ethyl methacrylate	10.871	69	184775	48.182	ug/l	100
57) 1,3-Dichloropropane	11.159	76	204590	51.606	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	409594	283.088	ug/l	100
59) 2-Hexanone	11.188	43	606740	277.643	ug/l	100
60) Dibromochloromethane	11.353	129	140704	51.116	ug/l	100
61) 1,2-Dibromoethane	11.465	107	113541	50.028	ug/l	100
64) Tetrachloroethene	11.100	164	95879	47.295	ug/l	100
65) Chlorobenzene	11.888	112	319992	49.121	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	116488	48.722	ug/l	100
67) Ethyl Benzene	11.959	91	555322	52.339	ug/l	100
68) m/p-Xylenes	12.065	106	420702	107.286	ug/l	100
69) o-Xylene	12.394	106	202597	54.062	ug/l	100
70) Styrene	12.406	104	348720	56.227	ug/l	100
71) Bromoform	12.576	173	92366	53.993	ug/l #	100
73) Isopropylbenzene	12.688	105	505619	51.094	ug/l	100
74) N-amyl acetate	12.488	43	237870	53.478	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	167856	48.019	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	130866m	43.982	ug/l	
77) Bromobenzene	12.976	156	125800	48.657	ug/l	100
78) n-propylbenzene	13.029	91	609022	51.986	ug/l	100
79) 2-Chlorotoluene	13.123	91	378981	49.985	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	431021	52.730	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	57184	51.914	ug/l	100
82) 4-Chlorotoluene	13.218	91	379524	50.273	ug/l	100
83) tert-Butylbenzene	13.435	119	357563	52.090	ug/l	100
84) 1,2,4-Trimethylbenzene	13.476	105	442175	54.275	ug/l	100
85) sec-Butylbenzene	13.612	105	501561	52.719	ug/l	100
86) p-Isopropyltoluene	13.723	119	421052	49.240	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	229518	48.204	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	229084	46.416	ug/l	100
89) n-Butylbenzene	14.053	91	364111	53.348	ug/l	100
90) Hexachloroethane	14.329	117	84602	46.712	ug/l	100
91) 1,2-Dichlorobenzene	14.100	146	228024	48.012	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	32536	50.971	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	114468	51.857	ug/l	100
94) Hexachlorobutadiene	15.500	225	53846	45.947	ug/l	100
95) Naphthalene	15.635	128	363882	55.243	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	115251	51.601	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085439.D
Acq On : 14 Jan 2025 15:19
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:42:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085439.D
Acq On : 14 Jan 2025 15:19
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Quant Time: Jan 15 01:42:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:168 Resp: 199403
Ion Ratio Lower Upper
168 100
99 67.0 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#2
Dichlorodifluoromethane
Concen: 46.486 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 85 Resp: 125504
Ion Ratio Lower Upper
85 100
87 33.4 16.7 50.1





#3

Chloromethane

Concen: 46.398 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 50 Resp: 135631

Ion Ratio Lower Upper

50 100





52 33.8 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#4

Vinyl Chloride

Concen: 46.569 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 62 Resp: 136831

Ion Ratio Lower Upper

62 100

64 31.0 24.8 37.2





#5

Bromomethane

Concen: 46.877 ug/l

RT: 2.953 min Scan# 170

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 94 Resp: 83199

Ion Ratio Lower Upper

94 100





96 89.7 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#6

Chloroethane

Concen: 45.371 ug/l

RT: 3.118 min Scan# 198

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 64 Resp: 84517

Ion Ratio Lower Upper

64 100

66 30.7 24.6 36.8





#7

Trichlorofluoromethane

Concen: 46.613 ug/l

RT: 3.501 min Scan# 263

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:101 Resp: 198726

Ion Ratio Lower Upper

101 100





103 64.3 51.4 77.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#8

Diethyl Ether

Concen: 48.726 ug/l

RT: 3.953 min Scan# 340

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 74 Resp: 71764

Ion Ratio Lower Upper

74 100

45 99.4 49.7 149.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 45.010 ug/l

RT: 4.371 min Scan# 411

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	101	Resp:	108081
Ion Ratio	Lower	Upper	
101	100		
85	47.3	37.8	56.8
151	73.5	58.8	88.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#10

Methyl Iodide

Concen: 52.142 ug/l




RT: 4.589 min Scan# 448


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	142	Resp:	143383
Ion Ratio	Lower	Upper	
142	100		
127	50.4	40.3	60.5
141	16.0	12.8	19.2





#11

Tert butyl alcohol

Concen: 251.894 ug/l

RT: 5.512 min Scan# 605

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 92841

Ion Ratio Lower Upper

59 100





57 10.8 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#12

1,1-Dichloroethene

Concen: 49.647 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 96 Resp: 106247


Ion Ratio Lower Upper

96 100

61 176.2 141.0 211.4

98 61.3 49.0 73.6





#13

Acrolein

Concen: 276.734 ug/l

RT: 4.171 min Scan# 377

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 139234

Ion Ratio Lower Upper

56 100





55 70.4 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#14

Allyl chloride

Concen: 48.567 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 41 Resp: 168654


Ion Ratio Lower Upper

41 100

39 80.5 64.4 96.6

76 38.1 30.5 45.7





#15

Acrylonitrile

Concen: 253.507 ug/l

RT: 5.712 min Scan# 639

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	53	Resp:	296781
Ion	Ratio	Lower	Upper
53	100		
52	81.9	65.5	98.3
51	37.3	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#16

Acetone

Concen: 241.603 ug/l




RT: 4.418 min Scan# 419


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	43	Resp:	251271
Ion	Ratio	Lower	Upper
43	100		
58	29.7	23.8	35.6





#17

Carbon Disulfide

Concen: 44.710 ug/l

RT: 4.706 min Scan# 468

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 294604

Ion Ratio Lower Upper

76 100





78 8.6 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#18

Methyl Acetate

Concen: 49.798 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 43 Resp: 157485

Ion Ratio Lower Upper

43 100

74 23.3 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 53.758 ug/l

RT: 5.789 min Scan# 65

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 373561

Ion Ratio Lower Upper

73 100





57 22.3 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#20

Methylene Chloride

Concen: 48.748 ug/l

RT: 5.271 min Scan# 564

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion: 84 Resp: 125508




Ion Ratio Lower Upper


84 100

49 130.1 104.1 156.1

51 38.7 31.0 46.4

86 64.9 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 48.395 ug/l

RT: 5.783 min Scan# 651

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 110684

Ion Ratio Lower Upper

96 100

61 144.6 115.7 173.5





98 62.3 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether

Concen: 52.707 ug/l

RT: 6.665 min Scan# 801

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion: 45 Resp: 406253




Ion Ratio Lower Upper


45 100

43 56.3 45.0 67.6

87 27.3 21.8 32.8

59 12.7 10.2 15.2





#23

Vinyl Acetate

Concen: 275.471 ug/l

RT: 6.594 min Scan# 789

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 1491031

Ion Ratio Lower Upper

43 100





86 10.5 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 49.648 ug/l

RT: 6.565 min Scan# 784

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 63 Resp: 233336


Ion Ratio Lower Upper

63 100

98 6.5 3.3 9.8

100 4.0 2.0 6.0





#25

2-Butanone

Concen: 253.815 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 388457

Ion Ratio Lower Upper

43 100

72 25.3 20.2 30.4

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#26

2,2-Dichloropropane

Concen: 49.106 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 77 Resp: 186591

Ion Ratio Lower Upper

77 100

97 21.4 10.7 32.1





#27

cis-1,2-Dichloroethene

Concen: 50.571 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 136229

Ion Ratio Lower Upper

96 100

61 155.9 0.0 311.8





98 63.0 0.0 126.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#28

Bromochloromethane

Concen: 49.461 ug/l

RT: 7.806 min Scan# 995

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 49 Resp: 108148


Ion Ratio Lower Upper

49 100

129 2.2 0.0 4.4

130 68.7 55.0 82.4









#29
Tetrahydrofuran
Concen: 267.624 ug/l
RT: 7.830 min Scan# 999
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 42 Resp: 259683
Ion Ratio Lower Upper
42 100
72 42.7 34.2 51.2
71 40.6 32.5 48.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#30
Chloroform
Concen: 48.229 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 83 Resp: 234271
Ion Ratio Lower Upper
83 100
85 64.7 51.8 77.6





#31

Cyclohexane

Concen: 42.984 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion: 56 Resp: 175588

Ion Ratio Lower Upper

56 100

69 27.8 22.2 33.4

84 83.0 66.4 99.6

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#32

1,1,1-Trichloroethane

Concen: 47.547 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 97 Resp: 202590


Ion Ratio Lower Upper

97 100

99 62.2 49.8 74.6

61 51.8 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 51.500 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 165768

Ion Ratio Lower Upper

65 100





67 50.8 0.0 101.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion:114 Resp: 339872


Ion Ratio Lower Upper

114 100

63 23.8 0.0 47.6

88 16.3 0.0 32.6





#35

Dibromofluoromethane

Concen: 51.552 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 113 Resp: 121554

Ion Ratio Lower Upper

113 100

111 103.4 82.7 124.1





192 17.9 14.3 21.5

Manual Integrations

APPROVED

Reviewed By : John Carlone 01/15/2025

Supervised By : Mahesh Dadoda 01/15/2025



#36

1,1-Dichloropropene

Concen: 47.453 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 75 Resp: 157031


Ion Ratio Lower Upper

75 100

110 33.0 16.5 49.5

77 30.5 24.4 36.6





#37

Ethyl Acetate

Concen: 49.658 ug/l

RT: 7.553 min Scan# 952

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 165866

Ion Ratio Lower Upper

43 100

61 13.6 10.9 16.3





70 9.4 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#38

Carbon Tetrachloride

Concen: 47.564 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion:117 Resp: 180199


Ion Ratio Lower Upper

117 100

119 94.3 75.4 113.2

121 30.8 24.6 37.0





#39

Methylcyclohexane

Concen: 50.562 ug/l

RT: 9.600 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	83	Resp:	157234
Ion	Ratio	Lower	Upper
83	100		
55	78.3	62.6	94.0
98	47.1	37.7	56.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#40

Benzene

Concen: 49.542 ug/l




RT: 8.600 min Scan# 1130


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	78	Resp:	492606
Ion	Ratio	Lower	Upper
78	100		
77	23.8	19.0	28.6





#41

Methacrylonitrile

Concen: 52.775 ug/l

RT: 7.771 min Scan# 989

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 41 Resp: 91711

Ion Ratio Lower Upper

41 100

39 57.5 46.0 69.0

67 71.8 57.4 86.2





52 31.9 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#42

1,2-Dichloroethane

Concen: 49.685 ug/l

RT: 8.665 min Scan# 1141

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 62 Resp: 186065

Ion Ratio Lower Upper

62 100

98 8.5 0.0 17.0





#43

Isopropyl Acetate

Concen: 51.611 ug/l

RT: 8.683 min Scan# 1144

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 276228

Ion Ratio Lower Upper

43 100

61 25.9 20.7 31.1





87 12.3 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#44

Trichloroethene

Concen: 47.506 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion:130 Resp: 109953

Ion Ratio Lower Upper

130 100

95 97.9 0.0 195.8





#45

1,2-Dichloropropane

Concen: 49.644 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 126084

Ion Ratio Lower Upper

63 100





65 32.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#46

Dibromomethane

Concen: 49.476 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 93 Resp: 90670


Ion Ratio Lower Upper

93 100

95 80.9 64.7 97.1

174 86.2 69.0 103.4





#47

Bromodichloromethane

Concen: 50.867 ug/l

RT: 9.882 min Scan# 1348

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	83	Resp:	189915
Ion	Ratio	Lower	Upper
83	100		
85	64.0	51.2	76.8
127	8.1	6.5	9.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#48

Methyl methacrylate

Concen: 55.385 ug/l




RT: 9.677 min Scan# 1313


Delta R.T. -0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19


Tgt Ion:	41	Resp:	133399
Ion	Ratio	Lower	Upper
41	100		
69	80.9	64.7	97.1
39	61.3	49.0	73.6





#49
1,4-Dioxane
Concen: 1061.334 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 88 Resp: 43048
Ion Ratio Lower Upper
88 100
43 33.2 26.6 39.8
58 74.4 59.5 89.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#50
Toluene-d8
Concen: 51.411 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 98 Resp: 430698
Ion Ratio Lower Upper
98 100
100 65.3 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 269.403 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 836707
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 51.341 ug/l
RT: 10.624 min Scan# 1474
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 92 Resp: 295792
Ion Ratio Lower Upper
92 100
91 174.0 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 53.125 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 75 Resp: 187439

Ion Ratio Lower Upper

75 100





77 29.4 23.5 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#54

cis-1,3-Dichloropropene

Concen: 53.006 ug/l

RT: 10.306 min Scan# 1420

Delta R.T. -0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 75 Resp: 199771


Ion Ratio Lower Upper

75 100

77 31.2 25.0 37.4

39 53.9 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 50.641 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 115462

Ion Ratio Lower Upper

97 100

83 86.3 69.0 103.6

85 55.7 44.6 66.8





99 60.7 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#56

Ethyl methacrylate

Concen: 48.182 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 69 Resp: 184775


Ion Ratio Lower Upper

69 100

41 68.3 54.6 82.0

39 40.5 32.4 48.6





#57

1,3-Dichloropropane

Concen: 51.606 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 204596

Ion Ratio Lower Upper

76 100





78 32.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#58

2-Chloroethyl Vinyl ether

Concen: 283.088 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 63 Resp: 409594

Ion Ratio Lower Upper

63 100

106 27.0 21.6 32.4





#59

2-Hexanone

Concen: 277.643 ug/l

RT: 11.188 min Scan# 1570

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 606740

Ion Ratio Lower Upper

43 100





58 52.4 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#60

Dibromochloromethane

Concen: 51.116 ug/l

RT: 11.353 min Scan# 1598

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion:129 Resp: 140704

Ion Ratio Lower Upper

129 100

127 77.2 38.6 115.8









#61
1,2-Dibromoethane
Concen: 50.028 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 107 Resp: 113541
Ion Ratio Lower Upper
107 100
109 94.9 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#62
4-Bromofluorobenzene
Concen: 53.239 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 95 Resp: 152568
Ion Ratio Lower Upper
95 100
174 72.5 0.0 145.0
176 71.2 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:117 Resp: 297366

Ion Ratio Lower Upper

117 100

82 60.7 48.6 72.8





119 33.2 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#64

Tetrachloroethene

Concen: 47.295 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:164 Resp: 95879




Ion Ratio Lower Upper


164 100

166 129.3 103.4 155.2

129 99.0 79.2 118.8

131 96.4 77.1 115.7





#65

Chlorobenzene

Concen: 49.121 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:112 Resp: 319992

Ion Ratio Lower Upper

112 100





114 31.6 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 48.722 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion:131 Resp: 116488


Ion Ratio Lower Upper

131 100

133 94.9 47.4 142.3

119 66.3 33.1 99.5





#67

Ethyl Benzene

Concen: 52.339 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 555322

Ion Ratio Lower Upper

91 100





106 29.8 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#68

m/p-Xylenes

Concen: 107.286 ug/l

RT: 12.065 min Scan# 1719

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion:106 Resp: 420702

Ion Ratio Lower Upper

106 100

91 209.6 167.7 251.5





#69

o-Xylene

Concen: 54.062 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:106 Resp: 202597

Ion Ratio Lower Upper

106 100





91 220.8 110.4 331.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#70

Styrene

Concen: 56.227 ug/l

RT: 12.406 min Scan# 1777

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion:104 Resp: 348720


Ion Ratio Lower Upper

104 100

78 53.1 42.5 63.7

103 54.8 43.8 65.8





#71

Bromoform

Concen: 53.993 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	173	Resp:	92366
Ion Ratio	Lower	Upper	
173	100		
175	48.8	24.4	73.2
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	152	Resp:	146624
Ion Ratio	Lower	Upper	
152	100		
115	62.2	31.1	93.3
150	171.8	0.0	343.6





#73

Isopropylbenzene

Concen: 51.094 ug/l

RT: 12.688 min Scan# 1825

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:105 Resp: 505619

Ion Ratio Lower Upper

105 100





120 25.5 12.8 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#74

N-ethyl acetate

Concen: 53.478 ug/l

RT: 12.488 min Scan# 1791

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 237870




Ion Ratio Lower Upper


43 100

70 42.5 34.0 51.0

55 26.8 21.4 32.2

61 23.7 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 48.019 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:	83	Resp:	167856
Ion	Ratio	Lower	Upper
83	100		
131	9.6	4.8	14.4
85	64.4	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#76

1,2,3-Trichloropropane

Concen: 43.982 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	75	Resp:	130866
Ion	Ratio	Lower	Upper
75	100		
77	219.5	109.7	329.2





#77

Bromobenzene

Concen: 48.657 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:156 Resp: 125800

Ion Ratio Lower Upper

156 100

77 228.3 114.1 342.4





158 97.9 48.9 146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#78

n-propylbenzene

Concen: 51.986 ug/l

RT: 13.029 min Scan# 1883

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 91 Resp: 609022

Ion Ratio Lower Upper

91 100

120 21.7 10.9 32.6





#79

2-Chlorotoluene

Concen: 49.985 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 378981

Ion Ratio Lower Upper

91 100





126 31.4 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#80

1,3,5-Trimethylbenzene

Concen: 52.730 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion:105 Resp: 431021

Ion Ratio Lower Upper

105 100

120 47.8 23.9 71.7





#81

trans-1,4-Dichloro-2-butene

Concen: 51.914 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:	75	Resp:	57184
Ion	Ratio	Lower	Upper
75	100		
53	119.5	95.6	143.4
89	46.3	37.0	55.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#82

4-Chlorotoluene

Concen: 50.273 ug/l




RT: 13.218 min Scan# 1915


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	91	Resp:	379524
Ion	Ratio	Lower	Upper
91	100		
126	31.8	15.9	47.7





#83

tert-Butylbenzene

Concen: 52.090 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:119 Resp: 357563

Ion Ratio Lower Upper

119 100

91 70.2 35.1 105.3





134 24.1 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#84

1,2,4-Trimethylbenzene

Concen: 54.275 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion:105 Resp: 442175

Ion Ratio Lower Upper

105 100

120 43.3 21.6 65.0





#85

sec-Butylbenzene

Concen: 52.719 ug/l

RT: 13.612 min Scan# 1982

Delta R.T. -0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:105 Resp: 501561

Ion Ratio Lower Upper

105 100





134 19.3 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#86

p-Isopropyltoluene

Concen: 49.240 ug/l

RT: 13.723 min Scan# 2001

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion:119 Resp: 421052


Ion Ratio Lower Upper

119 100

134 25.3 12.7 38.0

91 25.4 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 48.204 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	146	Resp:	229518
Ion	Ratio	Lower	Upper
146	100		
111	42.9	21.4	64.3
148	64.6	32.3	96.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#88

1,4-Dichlorobenzene

Concen: 46.416 ug/l




RT: 13.812 min Scan# 2016


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	146	Resp:	229084
Ion	Ratio	Lower	Upper
146	100		
111	42.5	21.3	63.7
148	64.7	32.4	97.0





#89

n-Butylbenzene

Concen: 53.348 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	91	Resp:	364111
Ion	Ratio	Lower	Upper
91	100		
92	51.5	25.8	77.3
134	23.4	11.7	35.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90

Hexachloroethane

Concen: 46.712 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	117	Resp:	84602
Ion	Ratio	Lower	Upper
117	100		
201	67.3	33.7	101.0





#91

1,2-Dichlorobenzene

Concen: 48.012 ug/l

RT: 14.100 min Scan# 2065

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	146	Resp:	228024
Ion	Ratio	Lower	Upper
146	100		
111	43.4	21.7	65.1
148	62.8	31.4	94.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 50.971 ug/l




RT: 14.717 min Scan# 2170


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	75	Resp:	32536
Ion	Ratio	Lower	Upper
75	100		
155	72.8	36.4	109.2
157	90.7	45.4	136.1





#93

1,2,4-Trichlorobenzene

Concen: 51.857 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	180	Resp:	114468
Ion Ratio	Lower	Upper	
180	100		
182	93.5	46.8	140.3
145	32.0	16.0	48.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene

Concen: 45.947 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	225	Resp:	53846
Ion Ratio	Lower	Upper	
225	100		
223	61.4	30.7	92.1
227	61.7	30.9	92.5





#95

Naphthalene

Concen: 55.243 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19





Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:	128	Resp:	363882
Ion	Ratio	Lower	Upper
128	100		
127	13.3	10.6	16.0
129	11.0	8.8	13.2

Manual Integrations
APPROVEDReviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

#96

1,2,3-Trichlorobenzene

Concen: 51.601 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Tgt Ion:	180	Resp:	115251
Ion	Ratio	Lower	Upper
180	100		
182	94.8	47.4	142.2
145	33.8	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085440.D
 Acq On : 14 Jan 2025 15:43
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:43:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	200796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	340593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143526	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.571	65	60593	18.694	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	37.380%#
35) Dibromofluoromethane	8.165	113	45655	19.322	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	38.640%#
50) Toluene-d8	10.559	98	164391	19.581	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	39.160%#
62) 4-Bromofluorobenzene	12.847	95	55798	19.430	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	38.860%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	54666	20.107	ug/l	98
3) Chloromethane	2.359	50	58426	19.848	ug/l	100
4) Vinyl Chloride	2.512	62	58403	19.739	ug/l	99
5) Bromomethane	2.954	94	36440	20.389	ug/l	94
6) Chloroethane	3.118	64	37594	20.041	ug/l	93
7) Trichlorofluoromethane	3.501	101	88121	20.526	ug/l	99
8) Diethyl Ether	3.965	74	29138	19.647	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	48933	20.237	ug/l	99
10) Methyl Iodide	4.589	142	56567	20.428	ug/l	93
11) Tert butyl alcohol	5.506	59	37984	102.342	ug/l	100
12) 1,1-Dichloroethene	4.342	96	44670	20.728	ug/l	95
13) Acrolein	4.171	56	49649	97.995	ug/l	99
14) Allyl chloride	5.024	41	67705	19.362	ug/l	96
15) Acrylonitrile	5.712	53	121626	103.171	ug/l	99
16) Acetone	4.424	43	101065	96.502	ug/l	96
17) Carbon Disulfide	4.712	76	132300	19.939	ug/l	98
18) Methyl Acetate	5.018	43	60873	19.115	ug/l	99
19) Methyl tert-butyl Ether	5.795	73	148847	21.272	ug/l	100
20) Methylene Chloride	5.271	84	52813	20.370	ug/l	97
21) trans-1,2-Dichloroethene	5.789	96	46115	20.023	ug/l	98
22) Diisopropyl ether	6.671	45	166719	21.480	ug/l	97
23) Vinyl Acetate	6.594	43	588567m	107.984	ug/l	
24) 1,1-Dichloroethane	6.565	63	96875	20.469	ug/l #	96
25) 2-Butanone	7.477	43	159952	103.786	ug/l	98
26) 2,2-Dichloropropane	7.483	77	79145	20.684	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	57454	21.180	ug/l	98
28) Bromochloromethane	7.812	49	41181	18.703	ug/l	99
29) Tetrahydrofuran	7.836	42	104792	107.247	ug/l	99
30) Chloroform	7.959	83	99655	20.374	ug/l	99
31) Cyclohexane	8.253	56	82980	20.173	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	87659	20.431	ug/l	98
36) 1,1-Dichloropropene	8.371	75	67949	20.490	ug/l	98
37) Ethyl Acetate	7.553	43	66443	19.850	ug/l	96
38) Carbon Tetrachloride	8.359	117	78842	20.767	ug/l	96
39) Methylcyclohexane	9.594	83	65012	20.862	ug/l	96
40) Benzene	8.600	78	207985	20.873	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085440.D
 Acq On : 14 Jan 2025 15:43
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:43:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	36578	21.004	ug/l	99
42) 1,2-Dichloroethane	8.665	62	78402	20.891	ug/l	99
43) Isopropyl Acetate	8.683	43	109131	20.347	ug/l	98
44) Trichloroethene	9.347	130	48009	20.699	ug/l	96
45) 1,2-Dichloropropane	9.618	63	52833	20.758	ug/l	93
46) Dibromomethane	9.700	93	37721	20.540	ug/l	98
47) Bromodichloromethane	9.882	83	78918	21.093	ug/l	96
48) Methyl methacrylate	9.677	41	50224	20.808	ug/l	96
49) 1,4-Dioxane	9.688	88	17140	421.686	ug/l	95
51) 4-Methyl-2-Pentanone	10.441	43	336911	108.249	ug/l	100
52) Toluene	10.630	92	125172	21.680	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	74129	20.965	ug/l	96
54) cis-1,3-Dichloropropene	10.306	75	81897	21.684	ug/l	99
55) 1,1,2-Trichloroethane	11.012	97	48039	21.025	ug/l	97
56) Ethyl methacrylate	10.871	69	70934	19.649	ug/l	99
57) 1,3-Dichloropropane	11.159	76	84210	21.196	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	153681	105.991	ug/l	100
59) 2-Hexanone	11.194	43	240595	109.863	ug/l	98
60) Dibromochloromethane	11.353	129	56069	20.326	ug/l	96
61) 1,2-Dibromoethane	11.465	107	48465	21.309	ug/l	96
64) Tetrachloroethene	11.100	164	43243	21.408	ug/l	96
65) Chlorobenzene	11.888	112	136809	21.077	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	48370	20.304	ug/l	98
67) Ethyl Benzene	11.959	91	229978	21.753	ug/l	98
68) m/p-Xylenes	12.071	106	177755	45.494	ug/l	100
69) o-Xylene	12.394	106	84548	22.643	ug/l	98
70) Styrene	12.412	104	140581	22.749	ug/l	99
71) Bromoform	12.576	173	36961	21.683	ug/l #	96
73) Isopropylbenzene	12.694	105	211354	21.819	ug/l	100
74) N-amyl acetate	12.494	43	94292	21.656	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	68151	19.917	ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	64180m	22.036	ug/l	
77) Bromobenzene	12.976	156	52022	20.556	ug/l	99
78) n-propylbenzene	13.035	91	252610	22.028	ug/l	100
79) 2-Chlorotoluene	13.123	91	158511	21.358	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	180294	22.533	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	22178m	20.569	ug/l	
82) 4-Chlorotoluene	13.218	91	161880	21.906	ug/l	100
83) tert-Butylbenzene	13.435	119	147541	21.958	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	183144	22.965	ug/l	99
85) sec-Butylbenzene	13.612	105	209133	22.457	ug/l	100
86) p-Isopropyltoluene	13.729	119	168888	21.661	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	97652	20.952	ug/l	100
88) 1,4-Dichlorobenzene	13.806	146	98321	20.351	ug/l	99
89) n-Butylbenzene	14.053	91	142673	21.355	ug/l	99
90) Hexachloroethane	14.329	117	35205	19.858	ug/l	98
91) 1,2-Dichlorobenzene	14.100	146	94954	20.425	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12859	20.580	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	45878	21.233	ug/l	99
94) Hexachlorobutadiene	15.500	225	22719	19.805	ug/l	95
95) Naphthalene	15.635	128	134793	20.906	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	45468	20.797	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085440.D
Acq On : 14 Jan 2025 15:43
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed



```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\  
Data File : VN085440.D  
Acq On    : 14 Jan 2025   15:43  
Operator  : JC\MD  
Sample    : VSTDICC020  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 6   Sample Multiplier: 1
```


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\M\$VOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:168 Resp: 200796
Ion Ratio Lower Upper
168 100
99 62.2 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#2
Dichlorodifluoromethane
Concen: 20.107 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 85 Resp: 54666
Ion Ratio Lower Upper
85 100
87 32.5 16.7 50.1





#3

Chloromethane

Concen: 19.848 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 50 Resp: 58426

Ion Ratio Lower Upper

50 100





52 34.0 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#4

Vinyl Chloride

Concen: 19.739 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 62 Resp: 58403

Ion Ratio Lower Upper

62 100

64 30.7 24.8 37.2





#5

Bromomethane

Concen: 20.389 ug/l

RT: 2.954 min Scan# 170

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 94 Resp: 36440

Ion Ratio Lower Upper

94 100





96 94.9 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#6

Chloroethane

Concen: 20.041 ug/l

RT: 3.118 min Scan# 198

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 64 Resp: 37594

Ion Ratio Lower Upper

64 100

66 34.3 24.6 36.8





#7

Trichlorofluoromethane

Concen: 20.526 ug/l

RT: 3.501 min Scan# 26

Delta R.T. 0.000 min

Lab File: VN085440.D


Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC020



Tgt Ion: 101 Resp: 88121
Ion Ratio Lower Upper
101 100
103 65.3 51.4 77.2

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#8

Diethyl Ether


Concen: 19.647 ug/l

RT: 3.965 min Scan# 342



Delta R.T. 0.012 min


Lab File: VN085440.D

Acq: 14 Jan 2025 15:43



Tgt Ion: 74 Resp: 29138
Ion Ratio Lower Upper
74 100
45 101.8 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.237 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 48933
Ion Ratio Lower Upper
101 100
85 46.6 37.8 56.8
151 72.1 58.8 88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#10
Methyl Iodide
Concen: 20.428 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:142 Resp: 56567
Ion Ratio Lower Upper
142 100
127 44.6 40.3 60.5
141 14.8 12.8 19.2





#11

Tert butyl alcohol

Concen: 102.342 ug/l

RT: 5.506 min Scan# 605

Delta R.T. -0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 59 Resp: 37984

Ion Ratio Lower Upper

59 100





57 10.7 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#12

1,1-Dichloroethene

Concen: 20.728 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 96 Resp: 44670


Ion Ratio Lower Upper

96 100

61 170.3 141.0 211.4

98 56.8 49.0 73.6





#13

Acrolein

Concen: 97.995 ug/l

RT: 4.171 min Scan# 377

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 49649

Ion Ratio Lower Upper

56 100





55 71.4 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#14

Allyl chloride

Concen: 19.362 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 41 Resp: 67705


Ion Ratio Lower Upper

41 100

39 85.0 64.4 96.6

76 38.6 30.5 45.7





#15

Acrylonitrile

Concen: 103.171 ug/l

RT: 5.712 min Scan# 639

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 53 Resp: 121626

Ion Ratio Lower Upper

53 100

52 81.1 65.5 98.3





51 36.1 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#16

Acetone

Concen: 96.502 ug/l

RT: 4.424 min Scan# 420

Delta R.T. 0.006 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 43 Resp: 101065

Ion Ratio Lower Upper

43 100

58 27.7 23.8 35.6





#17

Carbon Disulfide

Concen: 19.939 ug/l

RT: 4.712 min Scan# 468

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 132300

Ion Ratio Lower Upper

76 100





78 9.3 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#18

Methyl Acetate

Concen: 19.115 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 43 Resp: 60873

Ion Ratio Lower Upper

43 100

74 24.0 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 21.272 ug/l

RT: 5.795 min Scan# 65

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 73 Resp: 148847

Ion Ratio Lower Upper

73 100





57 22.1 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#20

Methylene Chloride

Concen: 20.370 ug/l

RT: 5.271 min Scan# 564

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 84 Resp: 52813




Ion Ratio Lower Upper


84 100

49 129.2 104.1 156.1

51 42.2 31.0 46.4

86 61.6 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 20.023 ug/l

RT: 5.789 min Scan# 651

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 96 Resp: 46115

Ion Ratio Lower Upper

96 100

61 146.1 115.7 173.5





98 64.5 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether

Concen: 21.480 ug/l

RT: 6.671 min Scan# 802

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 45 Resp: 166719




Ion Ratio Lower Upper


45 100

43 54.3 45.0 67.6

87 25.3 21.8 32.8

59 10.8 10.2 15.2





#23

Vinyl Acetate

Concen: 107.984 ug/l m

RT: 6.594 min Scan# 789

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 588567

Ion Ratio Lower Upper

43 100

86 10.1 8.4 12.6

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 20.469 ug/l

RT: 6.565 min Scan# 784

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 63 Resp: 96875


Ion Ratio Lower Upper

63 100

98 7.4 3.3 9.8

100 6.0 2.0 6.0#





#25

2-Butanone

Concen: 103.786 ug/l

RT: 7.477 min Scan# 93

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 159952

Ion Ratio Lower Upper

43 100

72 24.1 20.2 30.4

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#26

2,2-Dichloropropane

Concen: 20.684 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 77 Resp: 79145

Ion Ratio Lower Upper

77 100

97 22.0 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 21.180 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Tgt Ion: 96 Resp: 57454
 Ion Ratio Lower Upper
 96 100
 61 153.0 0.0 311.8
 98 63.7 0.0 126.0




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#28
 Bromochloromethane
 Concen: 18.703 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 49 Resp: 41181
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.4
 130 69.3 55.0 82.4





#29

Tetrahydrofuran

Concen: 107.247 ug/l

RT: 7.836 min Scan# 10

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 42 Resp: 104792

Ion Ratio Lower Upper

42 100

72 42.6 34.2 51.2





71 41.4 32.5 48.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#30

Chloroform

Concen: 20.374 ug/l

RT: 7.959 min Scan# 1021

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 83 Resp: 99655

Ion Ratio Lower Upper

83 100

85 65.8 51.8 77.6





#31

Cyclohexane

Concen: 20.173 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 82980

Ion Ratio Lower Upper

56 100

69 31.2 22.2 33.4





84 80.8 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#32

1,1,1-Trichloroethane

Concen: 20.431 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 97 Resp: 87659


Ion Ratio Lower Upper

97 100

99 64.2 49.8 74.6

61 50.3 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 18.694 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 60593

Ion Ratio Lower Upper

65 100





67 51.2 0.0 101.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion:114 Resp: 340593


Ion Ratio Lower Upper

114 100

63 21.6 0.0 47.6

88 16.0 0.0 32.6





#35

Dibromofluoromethane

Concen: 19.322 ug/l

RT: 8.165 min Scan# 1055

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 113 Resp: 45655

Ion Ratio Lower Upper

113 100

111 100.6 82.7 124.1





192 17.9 14.3 21.5

Manual Integrations

APPROVED

Reviewed By : John Carlone 01/15/2025

Supervised By : Mahesh Dadoda 01/15/2025



#36

1,1-Dichloropropene

Concen: 20.490 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 75 Resp: 67949


Ion Ratio Lower Upper

75 100

110 31.8 16.5 49.5

77 31.1 24.4 36.6





#37

Ethyl Acetate

Concen: 19.850 ug/l

RT: 7.553 min Scan# 952

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC020

Tgt Ion:	43	Resp:	66443
Ion	Ratio	Lower	Upper
43	100		
61	11.9	10.9	16.3
70	10.6	7.5	11.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#38

Carbon Tetrachloride

Concen: 20.767 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	117	Resp:	78842
Ion	Ratio	Lower	Upper
117	100		
119	98.9	75.4	113.2
121	28.9	24.6	37.0





#39

Methylcyclohexane

Concen: 20.862 ug/l

RT: 9.594 min Scan# 1299

Delta R.T. -0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC020

Tgt Ion:	83	Resp:	65012
Ion Ratio	Lower	Upper	
83	100		
55	83.4	62.6	94.0
98	46.4	37.7	56.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#40

Benzene

Concen: 20.873 ug/l




RT: 8.600 min Scan# 1130


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	78	Resp:	207985
Ion Ratio	Lower	Upper	
78	100		
77	24.4	19.0	28.6





#41

Methacrylonitrile

Concen: 21.004 ug/l

RT: 7.771 min Scan# 989

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 41 Resp: 36578

Ion Ratio Lower Upper

41 100

39 56.6 46.0 69.0

67 70.3 57.4 86.2





52 32.4 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#42

1,2-Dichloroethane

Concen: 20.891 ug/l

RT: 8.665 min Scan# 1141

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 62 Resp: 78402

Ion Ratio Lower Upper

62 100

98 8.3 0.0 17.0





#43

Isopropyl Acetate

Concen: 20.347 ug/l

RT: 8.683 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 109131

Ion Ratio Lower Upper

43 100

61 27.2 20.7 31.1





87 12.3 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#44

Trichloroethene

Concen: 20.699 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion:130 Resp: 48009

Ion Ratio Lower Upper

130 100

95 101.5 0.0 195.8





#45

1,2-Dichloropropane

Concen: 20.758 ug/l

RT: 9.618 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 52833

Ion Ratio Lower Upper

63 100





65 28.3 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#46

Dibromomethane

Concen: 20.540 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 93 Resp: 37721


Ion Ratio Lower Upper

93 100

95 82.0 64.7 97.1

174 88.4 69.0 103.4





#47

Bromodichloromethane

Concen: 21.093 ug/l

RT: 9.882 min Scan# 1348

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 83 Resp: 78918

Ion Ratio Lower Upper

83 100

85 67.8 51.2 76.8





127 8.6 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#48

Methyl methacrylate

Concen: 20.808 ug/l

RT: 9.677 min Scan# 1313

Delta R.T. -0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 41 Resp: 50224


Ion Ratio Lower Upper

41 100

69 84.3 64.7 97.1

39 64.2 49.0 73.6





#49

1,4-Dioxane

Concen: 421.686 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 88 Resp: 17140

Ion Ratio Lower Upper

88 100

43 34.9 26.6 39.8





58 69.0 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#50

Toluene-d8

Concen: 19.581 ug/l

RT: 10.559 min Scan# 1463

Delta R.T. -0.006 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 98 Resp: 164391

Ion Ratio Lower Upper

98 100

100 64.4 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 108.249 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 336911
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 21.680 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 92 Resp: 125172
Ion Ratio Lower Upper
92 100
91 175.1 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 20.965 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 74129

Ion Ratio Lower Upper

75 100

77 31.4 23.5 35.3

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#54

cis-1,3-Dichloropropene

Concen: 21.684 ug/l

RT: 10.306 min Scan# 1420

Delta R.T. -0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 75 Resp: 81897


Ion Ratio Lower Upper

75 100

77 32.4 25.0 37.4

39 54.4 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 21.025 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 97 Resp: 48039

Ion Ratio Lower Upper

97 100

83 84.1 69.0 103.6

85 54.6 44.6 66.8





99 64.8 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#56

Ethyl methacrylate

Concen: 19.649 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 69 Resp: 70934


Ion Ratio Lower Upper

69 100

41 67.6 54.6 82.0

39 41.7 32.4 48.6





#57

1,3-Dichloropropane

Concen: 21.196 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 84210

Ion Ratio Lower Upper

76 100





78 32.7 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#58

2-Chloroethyl Vinyl ether

Concen: 105.991 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 63 Resp: 153681

Ion Ratio Lower Upper

63 100

106 27.2 21.6 32.4





#59

2-Hexanone

Concen: 109.863 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 240595

Ion Ratio Lower Upper

43 100





58 51.1 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#60

Dibromochloromethane

Concen: 20.326 ug/l

RT: 11.353 min Scan# 1598

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion:129 Resp: 56069

Ion Ratio Lower Upper

129 100

127 80.4 38.6 115.8









#61
1,2-Dibromoethane
Concen: 21.309 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 107 Resp: 48465
Ion Ratio Lower Upper
107 100
109 90.6 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#62
4-Bromofluorobenzene
Concen: 19.430 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 95 Resp: 55798
Ion Ratio Lower Upper
95 100
174 75.2 0.0 145.0
176 73.2 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:117 Resp: 296298

Ion Ratio Lower Upper

117 100

82 59.3 48.6 72.8





119 32.2 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#64

Tetrachloroethene

Concen: 21.408 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:164 Resp: 43243




Ion Ratio Lower Upper


164 100

166 130.6 103.4 155.2

129 107.4 79.2 118.8

131 94.7 77.1 115.7





#65

Chlorobenzene

Concen: 21.077 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:112 Resp: 136809

Ion Ratio Lower Upper

112 100





114 32.9 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 20.304 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion:131 Resp: 48370


Ion Ratio Lower Upper

131 100

133 96.9 47.4 142.3

119 67.2 33.1 99.5





#67

Ethyl Benzene

Concen: 21.753 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 91 Resp: 229978

Ion Ratio Lower Upper

91 100





106 31.1 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#68

m/p-Xylenes

Concen: 45.494 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.006 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion:106 Resp: 177755

Ion Ratio Lower Upper

106 100

91 209.3 167.7 251.5





#69

o-Xylene

Concen: 22.643 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion:106 Resp: 84548

Ion Ratio Lower Upper

106 100





91 217.0 110.4 331.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#70

Styrene

Concen: 22.749 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion:104 Resp: 140581


Ion Ratio Lower Upper

104 100

78 54.4 42.5 63.7

103 54.8 43.8 65.8





#71

Bromoform

Concen: 21.683 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC020

Tgt Ion:	173	Resp:	36961
Ion Ratio	Lower	Upper	
173	100		
175	51.7	24.4	73.2
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	152	Resp:	143526
Ion Ratio	Lower	Upper	
152	100		
115	61.4	31.1	93.3
150	163.9	0.0	343.6





#73

Isopropylbenzene

Concen: 21.819 ug/l

RT: 12.694 min Scan# 18

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 105 Resp: 211354

Ion Ratio Lower Upper

105 100

120 25.5 12.8 38.3

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#74

N-ethyl acetate

Concen: 21.656 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 94292




Ion Ratio Lower Upper


43 100

70 41.4 34.0 51.0

55 27.9 21.4 32.2

61 23.1 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 19.917 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC020

Tgt Ion:	83	Resp:	68151
Ion	Ratio	Lower	Upper
83	100		
131	9.9	4.8	14.4
85	65.5	32.2	96.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#76

1,2,3-Trichloropropane

Concen: 22.036 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	75	Resp:	64180
Ion	Ratio	Lower	Upper
75	100		
77	186.8	109.7	329.2





#77

Bromobenzene

Concen: 20.556 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. 0.000 min

Lab File: VN085440.D


Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC020



Tgt Ion:	156	Resp:	52022
Ion	Ratio	Lower	Upper
156	100		
77	230.4	114.1	342.4
158	97.0	48.9	146.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#78

n-propylbenzene


Concen: 22.028 ug/l

RT: 13.035 min Scan# 1884



Delta R.T. 0.006 min


Lab File: VN085440.D

Acq: 14 Jan 2025 15:43



Tgt Ion:	91	Resp:	252610
Ion	Ratio	Lower	Upper
91	100		
120	21.8	10.9	32.6





#79

2-Chlorotoluene

Concen: 21.358 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 91 Resp: 158511

Ion Ratio Lower Upper

91 100

126 31.5 15.7 47.1

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#80

1,3,5-Trimethylbenzene

Concen: 22.533 ug/l

RT: 13.171 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion:105 Resp: 180294

Ion Ratio Lower Upper

105 100

120 47.9 23.9 71.7





#81

trans-1,4-Dichloro-2-butene

Concen: 20.569 ug/l m

RT: 12.735 min Scan# 1833

Delta R.T. 0.000 min

Lab File: VN085440.D


Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC020



Tgt Ion:	75	Resp:	22178
Ion	Ratio	Lower	Upper
75	100		
53	107.2	95.6	143.4
89	46.9	37.0	55.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#82

4-Chlorotoluene


Concen: 21.906 ug/l

RT: 13.218 min Scan# 1915



Delta R.T. 0.000 min


Lab File: VN085440.D

Acq: 14 Jan 2025 15:43



Tgt Ion:	91	Resp:	161880
Ion	Ratio	Lower	Upper
91	100		
126	31.6	15.9	47.7





#83

tert-Butylbenzene

Concen: 21.958 ug/l

RT: 13.435 min Scan# 19

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:119 Resp: 147541

Ion Ratio Lower Upper

119 100

91 69.5 35.1 105.3





134 25.5 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#84

1,2,4-Trimethylbenzene

Concen: 22.965 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion:105 Resp: 183144

Ion Ratio Lower Upper

105 100

120 42.8 21.6 65.0





#85

sec-Butylbenzene

Concen: 22.457 ug/l

RT: 13.612 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:105 Resp: 209133

Ion Ratio Lower Upper

105 100





134 19.2 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#86

p-Isopropyltoluene

Concen: 21.661 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion:119 Resp: 168888


Ion Ratio Lower Upper

119 100

134 26.4 12.7 38.0

91 25.5 12.7 38.1









1,3-Dichlorobenzene
Concen: 20.952 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:	146	Resp:	97652
Ion	Ratio	Lower	Upper
146	100		
111	42.9	21.4	64.3
148	64.3	32.3	96.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



1,4-Dichlorobenzene
Concen: 20.351 ug/l
RT: 13.806 min Scan# 2015
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:	146	Resp:	98321
Ion	Ratio	Lower	Upper
146	100		
111	42.2	21.3	63.7
148	65.1	32.4	97.0





#89

n-Butylbenzene

Concen: 21.355 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 91 Resp: 142673

Ion Ratio Lower Upper

91 100

92 51.0 25.8 77.3





134 23.0 11.7 35.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#90

Hexachloroethane

Concen: 19.858 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion:117 Resp: 35205

Ion Ratio Lower Upper

117 100

201 68.8 33.7 101.0





#91

1,2-Dichlorobenzene

Concen: 20.425 ug/l

RT: 14.100 min Scan# 2065

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC020

Tgt Ion:	146	Resp:	94954
Ion	Ratio	Lower	Upper
146	100		
111	43.5	21.7	65.1
148	63.5	31.4	94.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 20.580 ug/l




RT: 14.717 min Scan# 2170


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	75	Resp:	12859
Ion	Ratio	Lower	Upper
75	100		
155	72.7	36.4	109.2
157	90.9	45.4	136.1





#93

1,2,4-Trichlorobenzene

Concen: 21.233 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC020

Tgt Ion:	180	Resp:	45878
Ion Ratio	Lower	Upper	
180	100		
182	94.2	46.8	140.3
145	31.4	16.0	48.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene

Concen: 19.805 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	225	Resp:	22719
Ion Ratio	Lower	Upper	
225	100		
223	64.3	30.7	92.1
227	66.3	30.9	92.5





#95

Naphthalene

Concen: 20.906 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion:	128	Resp:	134793
Ion	Ratio	Lower	Upper
128	100		
127	13.3	10.6	16.0
129	10.6	8.8	13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96

1,2,3-Trichlorobenzene

Concen: 20.797 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	180	Resp:	45468
Ion	Ratio	Lower	Upper
180	100		
182	97.9	47.4	142.2
145	35.7	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085441.D
 Acq On : 14 Jan 2025 16:07
 Operator : JC\MD
 Sample : VSTDICC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC010

Quant Time: Jan 15 01:44:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	195889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	340403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132934	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.571	65	29856	9.442	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	18.880%#
35) Dibromofluoromethane	8.165	113	21110	8.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	17.880%#
50) Toluene-d8	10.565	98	73223	8.727	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	17.460%#
62) 4-Bromofluorobenzene	12.847	95	24327	8.476	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	16.960%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	26124	9.850	ug/l	96
3) Chloromethane	2.359	50	27143	9.452	ug/l	99
4) Vinyl Chloride	2.512	62	27862	9.653	ug/l	100
5) Bromomethane	2.965	94	17139	9.830	ug/l	87
6) Chloroethane	3.130	64	16791	9.175	ug/l	96
7) Trichlorofluoromethane	3.501	101	40755	9.731	ug/l	98
8) Diethyl Ether	3.953	74	12415	8.581	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.371	101	22982	9.742	ug/l	98
10) Methyl Iodide	4.589	142	25085	9.286	ug/l	99
11) Tert butyl alcohol	5.512	59	17962	49.608	ug/l	97
12) 1,1-Dichloroethene	4.342	96	20615	9.806	ug/l	99
13) Acrolein	4.177	56	25568	51.729	ug/l	99
14) Allyl chloride	5.018	41	32240	9.451	ug/l	97
15) Acrylonitrile	5.712	53	53959	46.918	ug/l	98
16) Acetone	4.424	43	48475	47.446	ug/l	98
17) Carbon Disulfide	4.712	76	60201	9.300	ug/l #	94
18) Methyl Acetate	5.030	43	30500	9.817	ug/l	98
19) Methyl tert-butyl Ether	5.789	73	65207	9.552	ug/l	91
20) Methylene Chloride	5.277	84	23723	9.379	ug/l	96
21) trans-1,2-Dichloroethene	5.783	96	21127	9.403	ug/l	97
22) Diisopropyl ether	6.671	45	71857	9.490	ug/l	97
23) Vinyl Acetate	6.600	43	250575	47.125	ug/l	98
24) 1,1-Dichloroethane	6.565	63	43099	9.335	ug/l	95
25) 2-Butanone	7.477	43	71191	47.350	ug/l	99
26) 2,2-Dichloropropane	7.483	77	36105	9.672	ug/l	99
27) cis-1,2-Dichloroethene	7.489	96	25049	9.466	ug/l	97
28) Bromochloromethane	7.806	49	19037	8.863	ug/l	97
29) Tetrahydrofuran	7.836	42	45641	47.881	ug/l	98
30) Chloroform	7.965	83	45792	9.596	ug/l	100
31) Cyclohexane	8.253	56	40211	10.020	ug/l #	93
32) 1,1,1-Trichloroethane	8.165	97	39170	9.358	ug/l	99
36) 1,1-Dichloropropene	8.371	75	30015	9.056	ug/l	97
37) Ethyl Acetate	7.559	43	30760	9.195	ug/l	97
38) Carbon Tetrachloride	8.359	117	36015	9.492	ug/l	94
39) Methylcyclohexane	9.594	83	27721	8.900	ug/l	95
40) Benzene	8.600	78	93680	9.407	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085441.D
 Acq On : 14 Jan 2025 16:07
 Operator : JC\MD
 Sample : VSTDICC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC010

Quant Time: Jan 15 01:44:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	16859	9.686	ug/l	96
42) 1,2-Dichloroethane	8.665	62	35569	9.483	ug/l	99
43) Isopropyl Acetate	8.683	43	48926	9.127	ug/l	99
44) Trichloroethene	9.347	130	21113	9.108	ug/l	95
45) 1,2-Dichloropropane	9.618	63	22747	8.942	ug/l	97
46) Dibromomethane	9.706	93	17260	9.404	ug/l	99
47) Bromodichloromethane	9.883	83	35020	9.365	ug/l	98
48) Methyl methacrylate	9.677	41	21873	9.067	ug/l	97
49) 1,4-Dioxane	9.694	88	7734	190.382	ug/l	98
51) 4-Methyl-2-Pentanone	10.441	43	146990	47.254	ug/l	98
52) Toluene	10.630	92	54976	9.527	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	32780	9.276	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	35904	9.512	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	21060	9.222	ug/l	98
56) Ethyl methacrylate	10.871	69	28454	9.042	ug/l	95
57) 1,3-Dichloropropane	11.159	76	37768	9.512	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	63110	43.550	ug/l	98
59) 2-Hexanone	11.194	43	101318	46.291	ug/l	98
60) Dibromochloromethane	11.353	129	25029	9.079	ug/l	98
61) 1,2-Dibromoethane	11.465	107	21055	9.263	ug/l	100
64) Tetrachloroethene	11.100	164	19934	9.928	ug/l	97
65) Chlorobenzene	11.888	112	61680	9.560	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.959	131	21853	9.229	ug/l	98
67) Ethyl Benzene	11.959	91	99248	9.445	ug/l	97
68) m/p-Xylenes	12.065	106	72431	18.650	ug/l	99
69) o-Xylene	12.394	106	34382	9.264	ug/l	99
70) Styrene	12.406	104	56285	9.163	ug/l	98
71) Bromoform	12.576	173	16071	9.485	ug/l #	98
73) Isopropylbenzene	12.694	105	87003	9.697	ug/l	99
74) N-amyl acetate	12.494	43	37343	9.260	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	30770	9.709	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	25443m	9.432	ug/l	
77) Bromobenzene	12.976	156	22784	9.720	ug/l	97
78) n-propylbenzene	13.035	91	102262	9.628	ug/l	98
79) 2-Chlorotoluene	13.123	91	68346	9.943	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	71656	9.669	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	9236	9.248	ug/l	100
82) 4-Chlorotoluene	13.218	91	66773	9.756	ug/l	97
83) tert-Butylbenzene	13.435	119	59455	9.553	ug/l	98
84) 1,2,4-Trimethylbenzene	13.476	105	72046	9.754	ug/l	99
85) sec-Butylbenzene	13.612	105	83662	9.699	ug/l	99
86) p-Isopropyltoluene	13.729	119	67556	9.674	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	41843	9.693	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	42728	9.549	ug/l	99
89) n-Butylbenzene	14.053	91	57632	9.314	ug/l	97
90) Hexachloroethane	14.329	117	15261	9.294	ug/l	97
91) 1,2-Dichlorobenzene	14.106	146	40726	9.458	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5639	9.744	ug/l	95
93) 1,2,4-Trichlorobenzene	15.388	180	18720	9.354	ug/l	99
94) Hexachlorobutadiene	15.500	225	10491	9.874	ug/l	98
95) Naphthalene	15.635	128	52054	8.717	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	19450	9.605	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085441.D
Acq On : 14 Jan 2025 16:07
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:44:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085441.D
Acq On : 14 Jan 2025 16:07
Operator : JC\MD
Sample : VSTDIC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:44:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion:168 Resp: 195889
Ion Ratio Lower Upper
168 100
99 61.7 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#2
Dichlorodifluoromethane
Concen: 9.850 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07


Tgt Ion: 85 Resp: 26124
Ion Ratio Lower Upper
85 100
87 35.8 16.7 50.1





#3
Chloromethane
Concen: 9.452 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion: 50 Resp: 27143
Ion Ratio Lower Upper
50 100
52 34.3 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#4
Vinyl Chloride
Concen: 9.653 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 62 Resp: 27862
Ion Ratio Lower Upper
62 100
64 31.2 24.8 37.2





#5

Bromomethane

Concen: 9.830 ug/l

RT: 2.965 min Scan# 171

Delta R.T. 0.012 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 94 Resp: 17139

Ion Ratio Lower Upper

94 100





96 102.1 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#6

Chloroethane

Concen: 9.175 ug/l

RT: 3.130 min Scan# 200

Delta R.T. 0.012 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 64 Resp: 16791

Ion Ratio Lower Upper

64 100

66 33.1 24.6 36.8





#7

Trichlorofluoromethane

Concen: 9.731 ug/l

RT: 3.501 min Scan# 263

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:101 Resp: 40755

Ion Ratio Lower Upper

101 100





103 65.9 51.4 77.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#8

Diethyl Ether

Concen: 8.581 ug/l

RT: 3.953 min Scan# 340

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 74 Resp: 12415

Ion Ratio Lower Upper

74 100

45 105.3 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 9.742 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion:101 Resp: 22982
Ion Ratio Lower Upper
101 100
85 45.8 37.8 56.8
151 75.2 58.8 88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#10
Methyl Iodide
Concen: 9.286 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:142 Resp: 25085
Ion Ratio Lower Upper
142 100
127 50.9 40.3 60.5
141 15.0 12.8 19.2





#11

Tert butyl alcohol

Concen: 49.608 ug/l

RT: 5.512 min Scan# 605

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 59 Resp: 17962

Ion Ratio Lower Upper

59 100





57 11.9 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#12

1,1-Dichloroethene

Concen: 9.806 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 96 Resp: 20615


Ion Ratio Lower Upper

96 100

61 176.8 141.0 211.4

98 60.5 49.0 73.6





#13

Acrolein

Concen: 51.729 ug/l

RT: 4.177 min Scan# 377

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 56 Resp: 25568

Ion Ratio Lower Upper

56 100





55 70.9 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#14

Allyl chloride

Concen: 9.451 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 41 Resp: 32240


Ion Ratio Lower Upper

41 100

39 78.3 64.4 96.6

76 35.1 30.5 45.7





#15

Acrylonitrile

Concen: 46.918 ug/l

RT: 5.712 min Scan# 63

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC010

Tgt Ion:	53	Resp:	53959
Ion	Ratio	Lower	Upper
53	100		
52	83.3	65.5	98.3
51	38.8	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#16

Acetone

Concen: 47.446 ug/l




RT: 4.424 min Scan# 420


Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	43	Resp:	48475
Ion	Ratio	Lower	Upper
43	100		
58	30.8	23.8	35.6





#17

Carbon Disulfide

Concen: 9.300 ug/l

RT: 4.712 min Scan# 46

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 76 Resp: 60201

Ion Ratio Lower Upper

76 100





78 10.7 6.9 10.3#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#18

Methyl Acetate

Concen: 9.817 ug/l

RT: 5.030 min Scan# 523

Delta R.T. 0.012 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 43 Resp: 30500

Ion Ratio Lower Upper

43 100

74 24.3 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 9.552 ug/l

RT: 5.789 min Scan# 65

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 73 Resp: 65207

Ion Ratio Lower Upper

73 100





57 18.1 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#20

Methylene Chloride

Concen: 9.379 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion: 84 Resp: 23723




Ion Ratio Lower Upper


84 100

49 132.4 104.1 156.1

51 34.4 31.0 46.4

86 70.0 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 9.403 ug/l

RT: 5.783 min Scan# 651

Delta R.T. 0.000 min

Lab File: VN085441.D


Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC010



Tgt Ion:	96	Resp:	21127
Ion	Ratio	Lower	Upper
96	100		
61	140.8	115.7	173.5
98	60.5	49.8	74.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether


Concen: 9.490 ug/l

RT: 6.671 min Scan# 802



Delta R.T. 0.006 min


Lab File: VN085441.D

Acq: 14 Jan 2025 16:07



Tgt Ion:	45	Resp:	71857
Ion	Ratio	Lower	Upper
45	100		
43	53.2	45.0	67.6
87	28.8	21.8	32.8
59	13.2	10.2	15.2





#23

Vinyl Acetate

Concen: 47.125 ug/l

RT: 6.600 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 43 Resp: 250575

Ion Ratio Lower Upper

43 100





86 9.8 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 9.335 ug/l

RT: 6.565 min Scan# 784

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 63 Resp: 43099


Ion Ratio Lower Upper

63 100

98 8.8 3.3 9.8

100 5.0 2.0 6.0





#25

2-Butanone

Concen: 47.350 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 43 Resp: 71191

Ion Ratio Lower Upper

43 100





72 24.8 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#26

2,2-Dichloropropane

Concen: 9.672 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 77 Resp: 36105

Ion Ratio Lower Upper

77 100

97 20.8 10.7 32.1





#27

cis-1,2-Dichloroethene

Concen: 9.466 ug/l

RT: 7.489 min Scan# 940

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 96 Resp: 25049

Ion Ratio Lower Upper

96 100

61 151.8 0.0 311.8





98 64.7 0.0 126.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#28

Bromochloromethane

Concen: 8.863 ug/l

RT: 7.806 min Scan# 995

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 49 Resp: 19037


Ion Ratio Lower Upper

49 100

129 0.3 0.0 4.4

130 70.8 55.0 82.4









#29
Tetrahydrofuran
Concen: 47.881 ug/l
RT: 7.836 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 42 Resp: 45641
Ion Ratio Lower Upper
42 100
72 41.6 34.2 51.2
71 39.2 32.5 48.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#30
Chloroform
Concen: 9.596 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 83 Resp: 45792
Ion Ratio Lower Upper
83 100
85 64.9 51.8 77.6





#31

Cyclohexane

Concen: 10.020 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 56 Resp: 40211

Ion Ratio Lower Upper

56 100

69 34.2 22.2 33.4





84 87.2 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#32

1,1,1-Trichloroethane

Concen: 9.358 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 97 Resp: 39170


Ion Ratio Lower Upper

97 100

99 62.6 49.8 74.6

61 50.2 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 9.442 ug/l

RT: 8.571 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 65 Resp: 29856

Ion Ratio Lower Upper

65 100





67 49.5 0.0 101.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:114 Resp: 340403


Ion Ratio Lower Upper

114 100

63 21.1 0.0 47.6

88 15.5 0.0 32.6





#35

Dibromofluoromethane

Concen: 8.939 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 113 Resp: 21110

Ion Ratio Lower Upper

113 100

111 103.4 82.7 124.1





192 16.4 14.3 21.5

Manual Integrations

APPROVED

Reviewed By : John Carlone 01/15/2025

Supervised By : Mahesh Dadoda 01/15/2025



#36

1,1-Dichloropropene

Concen: 9.056 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 75 Resp: 30015


Ion Ratio Lower Upper

75 100

110 34.3 16.5 49.5

77 32.9 24.4 36.6





#37

Ethyl Acetate

Concen: 9.195 ug/l

RT: 7.559 min Scan# 952

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC010

Tgt Ion:	43	Resp:	30760
Ion	Ratio	Lower	Upper
43	100		
61	12.2	10.9	16.3
70	8.4	7.5	11.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#38

Carbon Tetrachloride

Concen: 9.492 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	117	Resp:	36015
Ion	Ratio	Lower	Upper
117	100		
119	101.2	75.4	113.2
121	29.1	24.6	37.0





#39

Methylcyclohexane

Concen: 8.900 ug/l

RT: 9.594 min Scan# 12

Delta R.T. -0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 83 Resp: 27721

Ion Ratio Lower Upper

83 100

55 83.8 62.6 94.0





98 45.8 37.7 56.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#40

Benzene

Concen: 9.407 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 78 Resp: 93680

Ion Ratio Lower Upper

78 100

77 22.0 19.0 28.6





#41

Methacrylonitrile

Concen: 9.686 ug/l

RT: 7.777 min Scan# 990

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC010

Tgt Ion:	41	Resp:	16859
Ion	Ratio	Lower	Upper
41	100		
39	57.3	46.0	69.0
67	66.2	57.4	86.2
52	30.4	25.5	38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#42

1,2-Dichloroethane

Concen: 9.483 ug/l




RT: 8.665 min Scan# 1141


Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	62	Resp:	35569
Ion	Ratio	Lower	Upper
62	100		
98	8.2	0.0	17.0





#43

Isopropyl Acetate

Concen: 9.127 ug/l

RT: 8.683 min Scan# 1144

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 43 Resp: 48926

Ion Ratio Lower Upper

43 100

61 26.2 20.7 31.1





87 12.8 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#44

Trichloroethene

Concen: 9.108 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion:130 Resp: 21113

Ion Ratio Lower Upper

130 100

95 92.9 0.0 195.8





#45

1,2-Dichloropropane

Concen: 8.942 ug/l

RT: 9.618 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 63 Resp: 22747

Ion Ratio Lower Upper

63 100





65 30.4 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#46

Dibromomethane

Concen: 9.404 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 93 Resp: 17260


Ion Ratio Lower Upper

93 100

95 81.5 64.7 97.1

174 87.8 69.0 103.4





#47

Bromodichloromethane

Concen: 9.365 ug/l

RT: 9.883 min Scan# 1348

Delta R.T. 0.001 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC010

Tgt Ion:	83	Resp:	35020
Ion	Ratio	Lower	Upper
83	100		
85	62.6	51.2	76.8
127	6.9	6.5	9.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#48

Methyl methacrylate

Concen: 9.067 ug/l




RT: 9.677 min Scan# 1313


Delta R.T. -0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	41	Resp:	21873
Ion	Ratio	Lower	Upper
41	100		
69	80.0	64.7	97.1
39	65.2	49.0	73.6





#49

1,4-Dioxane

Concen: 190.382 ug/l

RT: 9.694 min Scan# 1315

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 88 Resp: 7734

Ion Ratio Lower Upper

88 100

43 35.2 26.6 39.8





58 73.3 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#50

Toluene-d8

Concen: 8.727 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 98 Resp: 73223

Ion Ratio Lower Upper

98 100

100 64.5 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 47.254 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 43 Resp: 146996
Ion Ratio Lower Upper
43 100
58 37.1 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 9.527 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 92 Resp: 54976
Ion Ratio Lower Upper
92 100
91 173.7 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 9.276 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 75 Resp: 32780

Ion Ratio Lower Upper

75 100





77 27.2 23.5 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#54

cis-1,3-Dichloropropene

Concen: 9.512 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 75 Resp: 35904


Ion Ratio Lower Upper

75 100

77 30.0 25.0 37.4

39 53.1 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 9.222 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 97 Resp: 21060

Ion Ratio Lower Upper

97 100

83 88.3 69.0 103.6

85 57.3 44.6 66.8





99 60.2 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#56

Ethyl methacrylate

Concen: 9.042 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 69 Resp: 28454


Ion Ratio Lower Upper

69 100

41 72.1 54.6 82.0

39 43.4 32.4 48.6





#57

1,3-Dichloropropane

Concen: 9.512 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 76 Resp: 37768

Ion Ratio Lower Upper

76 100





78 31.1 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#58

2-Chloroethyl Vinyl ether

Concen: 43.550 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 63 Resp: 63110

Ion Ratio Lower Upper

63 100

106 26.1 21.6 32.4





#59

2-Hexanone

Concen: 46.291 ug/l

RT: 11.194 min Scan# 15

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 43 Resp: 101318

Ion Ratio Lower Upper

43 100





58 50.9 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#60

Dibromochloromethane

Concen: 9.079 ug/l

RT: 11.353 min Scan# 1598

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion:129 Resp: 25029

Ion Ratio Lower Upper

129 100

127 79.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 9.263 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 107 Resp: 21055
Ion Ratio Lower Upper
107 100
109 94.6 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#62
4-Bromofluorobenzene
Concen: 8.476 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 95 Resp: 24327
Ion Ratio Lower Upper
95 100
174 73.9 0.0 145.0
176 72.1 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion:117 Resp: 294508

Ion Ratio Lower Upper

117 100

82 58.7 48.6 72.8





119 32.2 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#64

Tetrachloroethene

Concen: 9.928 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:164 Resp: 19934




Ion Ratio Lower Upper


164 100

166 132.2 103.4 155.2

129 98.3 79.2 118.8

131 90.5 77.1 115.7





#65

Chlorobenzene

Concen: 9.560 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:112 Resp: 61680

Ion Ratio Lower Upper

112 100





114 27.6 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 9.229 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:131 Resp: 21853


Ion Ratio Lower Upper

131 100

133 98.5 47.4 142.3

119 66.0 33.1 99.5





#67

Ethyl Benzene

Concen: 9.445 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 91 Resp: 99248

Ion Ratio Lower Upper

91 100





106 28.2 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#68

m/p-Xylenes

Concen: 18.650 ug/l

RT: 12.065 min Scan# 1719

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion:106 Resp: 72431

Ion Ratio Lower Upper

106 100

91 211.7 167.7 251.5





#69

o-Xylene

Concen: 9.264 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion:106 Resp: 34382

Ion Ratio Lower Upper

106 100





91 219.5 110.4 331.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#70

Styrene

Concen: 9.163 ug/l

RT: 12.406 min Scan# 1777

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:104 Resp: 56285


Ion Ratio Lower Upper

104 100

78 53.8 42.5 63.7

103 56.8 43.8 65.8





#71

Bromoform

Concen: 9.485 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC010

Tgt Ion:	173	Resp:	16071
Ion Ratio	Lower	Upper	
173	100		
175	50.0	24.4	73.2
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	152	Resp:	132934
Ion Ratio	Lower	Upper	
152	100		
115	62.3	31.1	93.3
150	159.8	0.0	343.6





#73

Isopropylbenzene

Concen: 9.697 ug/l

RT: 12.694 min Scan# 18

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion:105 Resp: 87003

Ion Ratio Lower Upper

105 100





120 26.1 12.8 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#74

N-amyI acetate

Concen: 9.260 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 37343




Ion Ratio Lower Upper


43 100

70 41.4 34.0 51.0

55 28.5 21.4 32.2

61 22.0 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 9.709 ug/l

RT: 12.935 min Scan# 18

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion:	83	Resp:	30770
Ion	Ratio	Lower	Upper
83	100		
131	9.6	4.8	14.4
85	63.6	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#76

1,2,3-Trichloropropane

Concen: 9.432 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	75	Resp:	25443
Ion	Ratio	Lower	Upper
75	100		
77	198.5	109.7	329.2





#77

Bromobenzene

Concen: 9.720 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. 0.000 min

Lab File: VN085441.D


Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC010



Tgt Ion:	156	Resp:	22784
Ion	Ratio	Lower	Upper
156	100		
77	221.6	114.1	342.4
158	98.7	48.9	146.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#78

n-propylbenzene


Concen: 9.628 ug/l

RT: 13.035 min Scan# 1884



Delta R.T. 0.006 min


Lab File: VN085441.D

Acq: 14 Jan 2025 16:07



Tgt Ion:	91	Resp:	102262
Ion	Ratio	Lower	Upper
91	100		
120	20.9	10.9	32.6





#79

2-Chlorotoluene

Concen: 9.943 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 91 Resp: 68346

Ion Ratio Lower Upper

91 100





126 31.5 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#80

1,3,5-Trimethylbenzene

Concen: 9.669 ug/l

RT: 13.171 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion:105 Resp: 71656

Ion Ratio Lower Upper

105 100

120 46.8 23.9 71.7





#81

trans-1,4-Dichloro-2-butene

Concen: 9.248 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. 0.000 min

Lab File: VN085441.D


Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC010



Tgt Ion:	75	Resp:	9236
Ion	Ratio	Lower	Upper
75	100		
53	119.6	95.6	143.4
89	46.6	37.0	55.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#82

4-Chlorotoluene


Concen: 9.756 ug/l

RT: 13.218 min Scan# 1915



Delta R.T. 0.000 min


Lab File: VN085441.D

Acq: 14 Jan 2025 16:07



Tgt Ion:	91	Resp:	66773
Ion	Ratio	Lower	Upper
91	100		
126	33.6	15.9	47.7





#83

tert-Butylbenzene

Concen: 9.553 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:119 Resp: 59455

Ion Ratio Lower Upper

119 100

91 70.0 35.1 105.3





134 27.8 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#84

1,2,4-Trimethylbenzene

Concen: 9.754 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion:105 Resp: 72046

Ion Ratio Lower Upper

105 100

120 43.8 21.6 65.0





#85

sec-Butylbenzene

Concen: 9.699 ug/l

RT: 13.612 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:105 Resp: 83662

Ion Ratio Lower Upper

105 100





134 18.9 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#86

p-Isopropyltoluene

Concen: 9.674 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:119 Resp: 67556


Ion Ratio Lower Upper

119 100

134 25.2 12.7 38.0

91 25.0 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 9.693 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC010

Tgt Ion:	146	Resp:	41843
Ion Ratio	Lower	Upper	
146	100		
111	43.1	21.4	64.3
148	63.0	32.3	96.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#88

1,4-Dichlorobenzene

Concen: 9.549 ug/l




RT: 13.812 min Scan# 2016


Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	146	Resp:	42728
Ion Ratio	Lower	Upper	
146	100		
111	44.0	21.3	63.7
148	64.6	32.4	97.0





#89

n-Butylbenzene

Concen: 9.314 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC010

Tgt Ion:	91	Resp:	57632
Ion	Ratio	Lower	Upper
91	100		
92	49.2	25.8	77.3
134	22.8	11.7	35.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90

Hexachloroethane

Concen: 9.294 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07


Tgt Ion:	117	Resp:	15261
Ion	Ratio	Lower	Upper
117	100		
201	65.0	33.7	101.0





#91
1,2-Dichlorobenzene
Concen: 9.458 ug/l
RT: 14.106 min Scan# 2065
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion: 146 Resp: 40726
Ion Ratio Lower Upper
146 100
111 42.9 21.7 65.1
148 62.9 31.4 94.2




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 9.744 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 5639
Ion Ratio Lower Upper
75 100
155 67.9 36.4 109.2
157 94.1 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 9.354 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:180 Resp: 18720

Ion Ratio Lower Upper

180 100

182 92.4 46.8 140.3





145 31.7 16.0 48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene

Concen: 9.874 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:225 Resp: 10491


Ion Ratio Lower Upper

225 100

223 63.3 30.7 92.1

227 62.1 30.9 92.5





#95

Naphthalene

Concen: 8.717 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC010

Tgt Ion:	128	Resp:	52054
Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.6	16.0
129	11.2	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96

1,2,3-Trichlorobenzene

Concen: 9.605 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	180	Resp:	19450
Ion	Ratio	Lower	Upper
180	100		
182	90.3	47.4	142.2
145	34.7	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085442.D
 Acq On : 14 Jan 2025 16:31
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:45:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	186556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324413	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126086	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.577	65	17046	5.660	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	11.320%#
35) Dibromofluoromethane	8.165	113	12102	5.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	10.760%#
50) Toluene-d8	10.565	98	41328	5.168	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	10.340%#
62) 4-Bromofluorobenzene	12.847	95	13520	4.943	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	9.880%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	13199	5.225	ug/l	97
3) Chloromethane	2.359	50	14526	5.311	ug/l	96
4) Vinyl Chloride	2.512	62	14564	5.298	ug/l	98
5) Bromomethane	2.953	94	9788	5.895	ug/l	98
6) Chloroethane	3.124	64	9430	5.411	ug/l	99
7) Trichlorofluoromethane	3.501	101	20098	5.039	ug/l	92
8) Diethyl Ether	3.959	74	6652	4.828	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	11917	5.305	ug/l	97
10) Methyl Iodide	4.583	142	12331	4.793	ug/l	99
11) Tert butyl alcohol	5.518	59	9075	26.318	ug/l #	94
12) 1,1-Dichloroethene	4.342	96	10429	5.209	ug/l	96
13) Acrolein	4.171	56	9430	20.033	ug/l	97
14) Allyl chloride	5.030	41	16741	5.153	ug/l	97
15) Acrylonitrile	5.718	53	27869	25.445	ug/l	97
16) Acetone	4.430	43	25122	25.819	ug/l	97
17) Carbon Disulfide	4.706	76	32068	5.202	ug/l	96
18) Methyl Acetate	5.024	43	15106	5.106	ug/l	97
19) Methyl tert-butyl Ether	5.800	73	31439	4.836	ug/l	94
20) Methylene Chloride	5.271	84	12981	5.389	ug/l #	83
21) trans-1,2-Dichloroethene	5.783	96	10614	4.960	ug/l #	78
22) Diisopropyl ether	6.665	45	35785	4.962	ug/l #	96
23) Vinyl Acetate	6.600	43	118439m	23.389	ug/l	
24) 1,1-Dichloroethane	6.565	63	22865	5.200	ug/l	99
25) 2-Butanone	7.483	43	36127	25.231	ug/l	99
26) 2,2-Dichloropropane	7.488	77	18398	5.175	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	12477	4.951	ug/l	99
28) Bromochloromethane	7.806	49	11096	5.424	ug/l	99
29) Tetrahydrofuran	7.841	42	21983	24.215	ug/l	99
30) Chloroform	7.959	83	23384	5.146	ug/l	98
31) Cyclohexane	8.247	56	22343	5.846	ug/l #	92
32) 1,1,1-Trichloroethane	8.159	97	21416	5.372	ug/l	91
36) 1,1-Dichloropropene	8.371	75	15886	5.029	ug/l	96
37) Ethyl Acetate	7.553	43	16879m	5.294	ug/l	
38) Carbon Tetrachloride	8.359	117	18317	5.065	ug/l	96
39) Methylcyclohexane	9.600	83	14175	4.776	ug/l	98
40) Benzene	8.600	78	47833	5.040	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085442.D
 Acq On : 14 Jan 2025 16:31
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:45:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	7380	4.449	ug/l	96
42) 1,2-Dichloroethane	8.671	62	18628	5.211	ug/l	98
43) Isopropyl Acetate	8.688	43	24792	4.853	ug/l	98
44) Trichloroethene	9.347	130	11133	5.039	ug/l	86
45) 1,2-Dichloropropane	9.618	63	12591	5.194	ug/l	92
46) Dibromomethane	9.700	93	8230	4.705	ug/l	91
47) Bromodichloromethane	9.888	83	18456	5.179	ug/l #	96
48) Methyl methacrylate	9.677	41	10574	4.599	ug/l	95
49) 1,4-Dioxane	9.694	88	3619	93.477	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	71889	24.250	ug/l	98
52) Toluene	10.629	92	27098	4.928	ug/l	98
53) t-1,3-Dichloropropene	10.829	75	17099	5.077	ug/l	95
54) cis-1,3-Dichloropropene	10.306	75	17440	4.848	ug/l	97
55) 1,1,2-Trichloroethane	11.012	97	11311	5.197	ug/l	96
56) Ethyl methacrylate	10.871	69	14094	5.627	ug/l	94
57) 1,3-Dichloropropane	11.159	76	18945	5.006	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	33210	24.047	ug/l	97
59) 2-Hexanone	11.194	43	49021	23.501	ug/l	94
60) Dibromochloromethane	11.353	129	13635	5.190	ug/l	96
61) 1,2-Dibromoethane	11.471	107	10419	4.810	ug/l	98
64) Tetrachloroethene	11.100	164	9756	5.071	ug/l	91
65) Chlorobenzene	11.888	112	31312	5.065	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	11620	5.121	ug/l	98
67) Ethyl Benzene	11.965	91	48216	4.789	ug/l	98
68) m/p-Xylenes	12.065	106	34771	9.344	ug/l	99
69) o-Xylene	12.394	106	16419	4.617	ug/l	98
70) Styrene	12.412	104	26218	4.454	ug/l	99
71) Bromoform	12.576	173	8017	4.938	ug/l #	99
73) Isopropylbenzene	12.694	105	39811	4.678	ug/l	99
74) N-amyl acetate	12.488	43	17577	4.595	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.935	83	15483	5.151	ug/l	97
76) 1,2,3-Trichloropropane	12.994	75	12173m	4.758	ug/l	
77) Bromobenzene	12.982	156	11091	4.989	ug/l	99
78) n-propylbenzene	13.035	91	45450	4.512	ug/l	99
79) 2-Chlorotoluene	13.123	91	31932	4.898	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	31891	4.537	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.729	75	4229	4.465	ug/l	90
82) 4-Chlorotoluene	13.218	91	31976	4.926	ug/l	100
83) tert-Butylbenzene	13.435	119	27597	4.675	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	32054	4.575	ug/l	98
85) sec-Butylbenzene	13.612	105	38148	4.663	ug/l	98
86) p-Isopropyltoluene	13.723	119	29856	4.584	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	20876	5.099	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	21982	5.179	ug/l	97
89) n-Butylbenzene	14.053	91	27437	4.675	ug/l	96
90) Hexachloroethane	14.329	117	7767	4.987	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	20177	4.940	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.729	75	2872	5.232	ug/l	90
93) 1,2,4-Trichlorobenzene	15.388	180	9037	4.761	ug/l	94
94) Hexachlorobutadiene	15.500	225	5563	5.520	ug/l	98
95) Naphthalene	15.641	128	25047	4.422	ug/l	96
96) 1,2,3-Trichlorobenzene	15.841	180	8737	4.549	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085442.D
Acq On : 14 Jan 2025 16:31
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:45:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085442.D
Acq On : 14 Jan 2025 16:31
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025


Quant Time: Jan 15 01:45:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 168 Resp: 186556
Ion Ratio Lower Upper
168 100
99 65.9 53.6 80.4




Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 5.225 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 85 Resp: 13199
Ion Ratio Lower Upper
85 100
87 31.5 16.7 50.1



Abundance Scan 69 (2.359 min): VN085439.D\data.ms (-62)



#3

Chloromethane

Concen: 5.311 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 50 Resp: 14526

Ion Ratio Lower Upper

50 100

52 31.6 27.0 40.6


Manual Integrations

APPROVED


Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025


Abundance Scan 69 (2.359 min): VN085442.D\data.ms




Abundance Scan 69 (2.359 min): VN085442.D\data.ms (-18)



Abundance



Abundance Scan 95 (2.512 min): VN085439.D\data.ms (-87)



#4

Vinyl Chloride

Concen: 5.298 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31


Tgt Ion: 62 Resp: 14564

Ion Ratio Lower Upper


62 100

64 29.9 24.8 37.2


Abundance Scan 95 (2.512 min): VN085442.D\data.ms




Abundance Scan 95 (2.512 min): VN085442.D\data.ms (-44)



Abundance





#5

Bromomethane

Concen: 5.895 ug/l

RT: 2.953 min Scan# 170

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 94 Resp: 9788

Ion Ratio Lower Upper

94 100





96 87.7 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#6

Chloroethane

Concen: 5.411 ug/l

RT: 3.124 min Scan# 199

Delta R.T. 0.006 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 64 Resp: 9430

Ion Ratio Lower Upper

64 100

66 31.4 24.6 36.8





#7

Trichlorofluoromethane

Concen: 5.039 ug/l

RT: 3.501 min Scan# 26

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion:101 Resp: 20098

Ion Ratio Lower Upper

101 100





103 70.8 51.4 77.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#8

Diethyl Ether

Concen: 4.828 ug/l

RT: 3.959 min Scan# 341

Delta R.T. 0.006 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 74 Resp: 6652

Ion Ratio Lower Upper

74 100

45 99.5 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.305 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:101 Resp: 11917
Ion Ratio Lower Upper
101 100
85 46.5 37.8 56.8
151 76.7 58.8 88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#10
Methyl Iodide
Concen: 4.793 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:142 Resp: 12331
Ion Ratio Lower Upper
142 100
127 50.9 40.3 60.5
141 15.0 12.8 19.2





#11

Tert butyl alcohol

Concen: 26.318 ug/l

RT: 5.518 min Scan# 605

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 59 Resp: 9075

Ion Ratio Lower Upper

59 100





57 8.5 8.6 13.0#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#12

1,1-Dichloroethene

Concen: 5.209 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 96 Resp: 10429


Ion Ratio Lower Upper

96 100

61 172.7 141.0 211.4

98 69.0 49.0 73.6





#13

Acrolein

Concen: 20.033 ug/l

RT: 4.171 min Scan# 377

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 56 Resp: 9430

Ion Ratio Lower Upper

56 100





55 73.2 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#14

Allyl chloride

Concen: 5.153 ug/l

RT: 5.030 min Scan# 523

Delta R.T. 0.012 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 41 Resp: 16741


Ion Ratio Lower Upper

41 100

39 77.4 64.4 96.6

76 36.6 30.5 45.7





#15

Acrylonitrile

Concen: 25.445 ug/l

RT: 5.718 min Scan# 64

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC005

Tgt Ion:	53	Resp:	27869
Ion	Ratio	Lower	Upper
53	100		
52	85.5	65.5	98.3
51	38.5	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#16

Acetone

Concen: 25.819 ug/l




RT: 4.430 min Scan# 421


Delta R.T. 0.012 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	43	Resp:	25122
Ion	Ratio	Lower	Upper
43	100		
58	27.9	23.8	35.6





#17

Carbon Disulfide

Concen: 5.202 ug/l

RT: 4.706 min Scan# 468

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 32068

Ion Ratio Lower Upper

76 100





78 9.9 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#18

Methyl Acetate

Concen: 5.106 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 43 Resp: 15106

Ion Ratio Lower Upper

43 100

74 22.0 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 4.836 ug/l

RT: 5.800 min Scan# 65

Delta R.T. 0.012 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 73 Resp: 31439

Ion Ratio Lower Upper

73 100





57 25.2 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#20

Methylene Chloride

Concen: 5.389 ug/l

RT: 5.271 min Scan# 564

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion: 84 Resp: 12981




Ion Ratio Lower Upper


84 100

49 114.5 104.1 156.1

51 32.1 31.0 46.4

86 42.9 51.9 77.9#





#21

trans-1,2-Dichloroethene

Concen: 4.960 ug/l

RT: 5.783 min Scan# 651

Delta R.T. 0.000 min

Lab File: VN085442.D


Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005



Tgt Ion: 96 Resp: 10614

Ion Ratio Lower Upper

96 100

61 118.1 115.7 173.5




98 44.7 49.8 74.8#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether


Concen: 4.962 ug/l

RT: 6.665 min Scan# 801

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31



Tgt Ion: 45 Resp: 35785



Ion Ratio Lower Upper


45 100

43 58.3 45.0 67.6

87 24.2 21.8 32.8

59 9.7 10.2 15.2#





#23

Vinyl Acetate

Concen: 23.389 ug/l m

RT: 6.600 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 43 Resp: 118439

Ion Ratio Lower Upper

43 100





86 9.0 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 5.200 ug/l

RT: 6.565 min Scan# 784

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 63 Resp: 22865


Ion Ratio Lower Upper

63 100

98 6.8 3.3 9.8

100 4.0 2.0 6.0





#25

2-Butanone

Concen: 25.231 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 36127

Ion Ratio Lower Upper

43 100

72 24.9 20.2 30.4

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#26

2,2-Dichloropropane

Concen: 5.175 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 77 Resp: 18398

Ion Ratio Lower Upper

77 100

97 21.2 10.7 32.1









#27
cis-1,2-Dichloroethene
Concen: 4.951 ug/l
RT: 7.488 min Scan# 940
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion	Ratio	Lower	Upper
96	100		
61	154.5	0.0	311.8
98	62.8	0.0	126.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#28
Bromochloromethane
Concen: 5.424 ug/l
RT: 7.806 min Scan# 995
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion	Ratio	Lower	Upper
49	100		
129	0.7	0.0	4.4
130	69.5	55.0	82.4





#29

Tetrahydrofuran

Concen: 24.215 ug/l

RT: 7.841 min Scan# 10

Delta R.T. 0.012 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 42 Resp: 21983

Ion Ratio Lower Upper

42 100

72 41.5 34.2 51.2





71 41.0 32.5 48.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#30

Chloroform

Concen: 5.146 ug/l

RT: 7.959 min Scan# 1021

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 83 Resp: 23384

Ion Ratio Lower Upper

83 100

85 63.4 51.8 77.6





#31

Cyclohexane

Concen: 5.846 ug/l

RT: 8.247 min Scan# 1071

Delta R.T. -0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 22343

Ion Ratio Lower Upper

56 100

69 42.3 22.2 33.4





84 84.3 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#32

1,1,1-Trichloroethane

Concen: 5.372 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. -0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 97 Resp: 21416


Ion Ratio Lower Upper

97 100

99 50.5 49.8 74.6

61 50.1 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 5.660 ug/l

RT: 8.577 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 17046

Ion Ratio Lower Upper

65 100





67 49.2 0.0 101.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion:114 Resp: 324413


Ion Ratio Lower Upper

114 100

63 23.3 0.0 47.6

88 16.0 0.0 32.6





#35

Dibromofluoromethane

Concen: 5.377 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:113 Resp: 12102

Ion Ratio Lower Upper

113 100

111 107.4 82.7 124.1





192 19.7 14.3 21.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#36

1,1-Dichloropropene

Concen: 5.029 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 75 Resp: 15886


Ion Ratio Lower Upper

75 100

110 30.3 16.5 49.5

77 29.2 24.4 36.6





#37

Ethyl Acetate

Concen: 5.294 ug/l m

RT: 7.553 min Scan# 952

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	43	Resp:	16879
Ion	Ratio	Lower	Upper
43	100		
61	12.3	10.9	16.3
70	8.4	7.5	11.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#38

Carbon Tetrachloride

Concen: 5.065 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	117	Resp:	18317
Ion	Ratio	Lower	Upper
117	100		
119	91.1	75.4	113.2
121	27.3	24.6	37.0





#39

Methylcyclohexane

Concen: 4.776 ug/l

RT: 9.600 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC005

Tgt Ion:	83	Resp:	14175
Ion Ratio	Lower	Upper	
83	100		
55	76.5	62.6	94.0
98	47.7	37.7	56.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#40

Benzene

Concen: 5.040 ug/l




RT: 8.600 min Scan# 1130


Delta R.T. -0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	78	Resp:	47833
Ion Ratio	Lower	Upper	
78	100		
77	23.2	19.0	28.6





#41

Methacrylonitrile

Concen: 4.449 ug/l

RT: 7.777 min Scan# 99

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 41 Resp: 7380

Ion Ratio Lower Upper

41 100

39 60.5 46.0 69.0

67 74.6 57.4 86.2





52 35.3 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#42

1,2-Dichloroethane

Concen: 5.211 ug/l

RT: 8.671 min Scan# 1142

Delta R.T. 0.006 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 62 Resp: 18628

Ion Ratio Lower Upper

62 100

98 7.8 0.0 17.0





#43

Isopropyl Acetate

Concen: 4.853 ug/l

RT: 8.688 min Scan# 1144

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	43	Resp:	24792
Ion	Ratio	Lower	Upper
43	100		
61	27.0	20.7	31.1
87	12.3	9.8	14.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#44

Trichloroethene

Concen: 5.039 ug/l




RT: 9.347 min Scan# 1257


Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	130	Resp:	11133
Ion	Ratio	Lower	Upper
130	100		
95	112.2	0.0	195.8





#45

1,2-Dichloropropane

Concen: 5.194 ug/l

RT: 9.618 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 63 Resp: 12591

Ion Ratio Lower Upper

63 100





65 27.8 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#46

Dibromomethane

Concen: 4.705 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 93 Resp: 8230


Ion Ratio Lower Upper

93 100

95 88.7 64.7 97.1

174 95.0 69.0 103.4





#47

Bromodichloromethane

Concen: 5.179 ug/l

RT: 9.888 min Scan# 1348

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	83	Resp:	18456
Ion	Ratio	Lower	Upper
83	100		
85	67.0	51.2	76.8
127	6.1	6.5	9.7#

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#48

Methyl methacrylate

Concen: 4.599 ug/l




RT: 9.677 min Scan# 1313


Delta R.T. -0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	41	Resp:	10574
Ion	Ratio	Lower	Upper
41	100		
69	84.9	64.7	97.1
39	65.3	49.0	73.6





#49

1,4-Dioxane

Concen: 93.477 ug/l

RT: 9.694 min Scan# 1315

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 88 Resp: 3619

Ion Ratio Lower Upper

88 100

43 32.3 26.6 39.8





58 77.9 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#50

Toluene-d8

Concen: 5.168 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 98 Resp: 41328

Ion Ratio Lower Upper

98 100

100 60.9 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 24.250 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 71889
Ion Ratio Lower Upper
43 100
58 36.9 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 4.928 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 92 Resp: 27098
Ion Ratio Lower Upper
92 100
91 176.4 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 5.077 ug/l

RT: 10.829 min Scan# 1510

Delta R.T. -0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 75 Resp: 17099

Ion Ratio Lower Upper

75 100





77 32.2 23.5 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#54

cis-1,3-Dichloropropene

Concen: 4.848 ug/l

RT: 10.306 min Scan# 1420

Delta R.T. -0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 75 Resp: 17440


Ion Ratio Lower Upper

75 100

77 27.5 25.0 37.4

39 53.2 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 5.197 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 11311

Ion Ratio Lower Upper

97 100

83 82.7 69.0 103.6

85 51.3 44.6 66.8





99 62.4 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#56

Ethyl methacrylate

Concen: 5.627 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 69 Resp: 14094


Ion Ratio Lower Upper

69 100

41 73.8 54.6 82.0

39 42.6 32.4 48.6





#57

1,3-Dichloropropane

Concen: 5.006 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 18945

Ion Ratio Lower Upper

76 100





78 32.2 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#58

2-Chloroethyl Vinyl ether

Concen: 24.047 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 63 Resp: 33210

Ion Ratio Lower Upper

63 100

106 25.2 21.6 32.4





#59

2-Hexanone

Concen: 23.501 ug/l

RT: 11.194 min Scan# 1570

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 49021

Ion Ratio Lower Upper

43 100





58 48.3 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#60

Dibromochloromethane

Concen: 5.190 ug/l

RT: 11.353 min Scan# 1598

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion:129 Resp: 13635

Ion Ratio Lower Upper

129 100

127 73.9 38.6 115.8









#61
1,2-Dibromoethane
Concen: 4.810 ug/l
RT: 11.471 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 107 Resp: 10419
Ion Ratio Lower Upper
107 100
109 96.8 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#62
4-Bromofluorobenzene
Concen: 4.943 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 95 Resp: 13520
Ion Ratio Lower Upper
95 100
174 73.2 0.0 145.0
176 68.0 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:117 Resp: 282203

Ion Ratio Lower Upper

117 100

82 58.3 48.6 72.8





119 32.1 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#64

Tetrachloroethene

Concen: 5.071 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:164 Resp: 9756




Ion Ratio Lower Upper


164 100

166 130.6 103.4 155.2

129 111.5 79.2 118.8

131 110.5 77.1 115.7





#65

Chlorobenzene

Concen: 5.065 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:112 Resp: 31312

Ion Ratio Lower Upper

112 100





114 30.7 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 5.121 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion:131 Resp: 11620


Ion Ratio Lower Upper

131 100

133 92.7 47.4 142.3

119 65.0 33.1 99.5





#67

Ethyl Benzene

Concen: 4.789 ug/l

RT: 11.965 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 48216

Ion Ratio Lower Upper

91 100





106 28.9 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#68

m/p-Xylenes

Concen: 9.344 ug/l

RT: 12.065 min Scan# 1719

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion:106 Resp: 34771

Ion Ratio Lower Upper

106 100

91 207.9 167.7 251.5





#69

o-Xylene

Concen: 4.617 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:106 Resp: 16419

Ion Ratio Lower Upper

106 100





91 224.0 110.4 331.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#70

Styrene

Concen: 4.454 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion:104 Resp: 26218


Ion Ratio Lower Upper

104 100

78 52.6 42.5 63.7

103 54.2 43.8 65.8





#71

Bromoform

Concen: 4.938 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC005

Tgt Ion:	173	Resp:	8017
Ion Ratio	Lower	Upper	
173	100		
175	48.3	24.4	73.2
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	152	Resp:	126086
Ion Ratio	Lower	Upper	
152	100		
115	62.0	31.1	93.3
150	160.7	0.0	343.6





#73

Isopropylbenzene

Concen: 4.678 ug/l

RT: 12.694 min Scan# 1825

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:105 Resp: 39811

Ion Ratio Lower Upper

105 100





120 26.1 12.8 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#74

N-ethyl acetate

Concen: 4.595 ug/l

RT: 12.488 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 17577




Ion Ratio Lower Upper


43 100

70 39.5 34.0 51.0

55 30.8 21.4 32.2

61 22.8 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 5.151 ug/l

RT: 12.935 min Scan# 18

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion:	83	Resp:	15483
Ion Ratio	Lower	Upper	
83	100		
131	9.5	4.8	14.4
85	67.5	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#76

1,2,3-Trichloropropane

Concen: 4.758 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	75	Resp:	12173
Ion Ratio	Lower	Upper	
75	100		
77	209.1	109.7	329.2





#77

Bromobenzene

Concen: 4.989 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.006 min

Lab File: VN085442.D


Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC005



Tgt Ion:	156	Resp:	11091
Ion	Ratio	Lower	Upper
156	100		
77	229.5	114.1	342.4
158	100.8	48.9	146.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#78

n-propylbenzene

Concen: 4.512 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	91	Resp:	45450
Ion	Ratio	Lower	Upper
91	100		
120	21.1	10.9	32.6





#79

2-Chlorotoluene

Concen: 4.898 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion: 91 Resp: 31932

Ion Ratio Lower Upper

91 100

126 32.6 15.7 47.1

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#80

1,3,5-Trimethylbenzene

Concen: 4.537 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion:105 Resp: 31891

Ion Ratio Lower Upper

105 100

120 48.1 23.9 71.7





#81

trans-1,4-Dichloro-2-butene

Concen: 4.465 ug/l

RT: 12.729 min Scan# 1833

Delta R.T. -0.006 min

Lab File: VN085442.D


Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC005



Tgt Ion:	75	Resp:	4229
Ion	Ratio	Lower	Upper
75	100		
53	132.6	95.6	143.4
89	49.7	37.0	55.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#82

4-Chlorotoluene


Concen: 4.926 ug/l

RT: 13.218 min Scan# 1915



Delta R.T. 0.000 min


Lab File: VN085442.D

Acq: 14 Jan 2025 16:31



Tgt Ion:	91	Resp:	31976
Ion	Ratio	Lower	Upper
91	100		
126	31.9	15.9	47.7





#83

tert-Butylbenzene

Concen: 4.675 ug/l

RT: 13.435 min Scan# 19

Delta R.T. 0.000 min

Lab File: VN085442.D


Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC005



Tgt Ion:	119	Resp:	27597
Ion	Ratio	Lower	Upper
119	100		
91	71.1	35.1	105.3
134	24.8	12.0	36.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#84

1,2,4-Trimethylbenzene


Concen: 4.575 ug/l

RT: 13.482 min Scan# 1960



Delta R.T. 0.006 min


Lab File: VN085442.D

Acq: 14 Jan 2025 16:31



Tgt Ion:	105	Resp:	32054
Ion	Ratio	Lower	Upper
105	100		
120	42.0	21.6	65.0





#85

sec-Butylbenzene

Concen: 4.663 ug/l

RT: 13.612 min Scan# 1982

Delta R.T. -0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:105 Resp: 38148

Ion Ratio Lower Upper

105 100





134 20.1 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#86

p-Isopropyltoluene

Concen: 4.584 ug/l

RT: 13.723 min Scan# 2001

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion:119 Resp: 29856


Ion Ratio Lower Upper

119 100

134 26.5 12.7 38.0

91 26.5 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 5.099 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:146 Resp: 20876

Ion Ratio Lower Upper

146 100

111 42.3 21.4 64.3





148 62.7 32.3 96.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#88

1,4-Dichlorobenzene

Concen: 5.179 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion:146 Resp: 21982


Ion Ratio Lower Upper

146 100

111 47.1 21.3 63.7

148 65.5 32.4 97.0





#89

n-Butylbenzene

Concen: 4.675 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC005

Tgt Ion:	91	Resp:	27437
Ion Ratio	Lower	Upper	
91	100		
92	48.7	25.8	77.3
134	21.4	11.7	35.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90

Hexachloroethane

Concen: 4.987 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	117	Resp:	7767
Ion Ratio	Lower	Upper	
117	100		
201	66.9	33.7	101.0





#91

1,2-Dichlorobenzene

Concen: 4.940 ug/l

RT: 14.106 min Scan# 2065

Delta R.T. 0.006 min

Lab File: VN085442.D


Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC005



Tgt Ion:	146	Resp:	20177
Ion	Ratio	Lower	Upper
146	100		
111	44.0	21.7	65.1
148	62.8	31.4	94.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#92

1,2-Dibromo-3-Chloropropane


Concen: 5.232 ug/l

RT: 14.729 min Scan# 2172



Delta R.T. 0.012 min


Lab File: VN085442.D

Acq: 14 Jan 2025 16:31



Tgt Ion:	75	Resp:	2872
Ion	Ratio	Lower	Upper
75	100		
155	64.2	36.4	109.2
157	99.8	45.4	136.1





#93

1,2,4-Trichlorobenzene

Concen: 4.761 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt	Ion:180	Resp:	9037
Ion	Ratio	Lower	Upper
180	100		
182	100.5	46.8	140.3
145	33.0	16.0	48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene

Concen: 5.520 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt	Ion:225	Resp:	5563
Ion	Ratio	Lower	Upper
225	100		
223	59.2	30.7	92.1
227	60.1	30.9	92.5





#95

Naphthalene

Concen: 4.422 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	128	Resp:	25047
Ion Ratio	Lower	Upper	
128	100		
127	14.7	10.6	16.0
129	9.1	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96

1,2,3-Trichlorobenzene

Concen: 4.549 ug/l




RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Tgt Ion:	180	Resp:	8737
Ion Ratio	Lower	Upper	
180	100		
182	95.6	47.4	142.2
145	35.5	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085443.D
 Acq On : 14 Jan 2025 17:19
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Jan 15 01:46:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302938	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	99539	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	2382	1.054	ug/l	91
3) Chloromethane	2.359	50	2801	1.145	ug/l	93
4) Vinyl Chloride	2.512	62	2732	1.111	ug/l #	81
6) Chloroethane	3.124	64	1809	1.160	ug/l	94
7) Trichlorofluoromethane	3.512	101	3860	1.082	ug/l	88
8) Diethyl Ether	3.953	74	1526	1.238	ug/l	81
9) 1,1,2-Trichlorotrifluo...	4.359	101	2156m	1.073	ug/l	
12) 1,1-Dichloroethene	4.330	96	1660	0.927	ug/l #	67
14) Allyl chloride	5.024	41	3179	1.094	ug/l #	85
15) Acrylonitrile	5.718	53	4967	5.070	ug/l #	70
16) Acetone	4.436	43	5102	5.862	ug/l	92
17) Carbon Disulfide	4.706	76	6601	1.197	ug/l #	87
18) Methyl Acetate	5.030	43	2908	1.099	ug/l #	80
19) Methyl tert-butyl Ether	5.789	73	5156	0.887	ug/l #	87
20) Methylene Chloride	5.265	84	2190	1.016	ug/l #	87
21) trans-1,2-Dichloroethene	5.794	96	2110	1.102	ug/l #	66
22) Diisopropyl ether	6.677	45	5691	0.882	ug/l #	88
23) Vinyl Acetate	6.606	43	18526m	4.090	ug/l	
24) 1,1-Dichloroethane	6.583	63	4020m	1.022	ug/l	
25) 2-Butanone	7.483	43	6448m	5.034	ug/l	
26) 2,2-Dichloropropane	7.488	77	3104	0.976	ug/l	92
27) cis-1,2-Dichloroethene	7.477	96	2185	0.969	ug/l	92
28) Bromochloromethane	7.812	49	2082	1.138	ug/l #	89
29) Tetrahydrofuran	7.835	42	3693	4.548	ug/l	98
30) Chloroform	7.965	83	4250	1.045	ug/l #	80
32) 1,1,1-Trichloroethane	8.165	97	3678	1.031	ug/l #	56
36) 1,1-Dichloropropene	8.371	75	3084	1.046	ug/l #	84
37) Ethyl Acetate	7.577	43	3056m	1.026	ug/l	
38) Carbon Tetrachloride	8.359	117	3436	1.018	ug/l	90
39) Methylcyclohexane	9.600	83	2403	0.867	ug/l #	87
40) Benzene	8.606	78	8480	0.957	ug/l	91
41) Methacrylonitrile	7.771	41	1458	0.941	ug/l #	76
42) 1,2-Dichloroethane	8.665	62	3131	0.938	ug/l	91
43) Isopropyl Acetate	8.682	43	4803	1.007	ug/l #	87
44) Trichloroethene	9.347	130	2133	1.034	ug/l	71
45) 1,2-Dichloropropane	9.618	63	2246	0.992	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085443.D
 Acq On : 14 Jan 2025 17:19
 Operator : JC\MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:46:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.706	93	1769	1.083	ug/l	92
47) Bromodichloromethane	9.882	83	2935	0.882	ug/l	90
48) Methyl methacrylate	9.677	41	1886	0.878	ug/l	96
49) 1,4-Dioxane	9.688	88	673	18.616	ug/l #	47
51) 4-Methyl-2-Pentanone	10.441	43	11523	4.163	ug/l	93
52) Toluene	10.618	92	4180	0.814	ug/l	93
53) t-1,3-Dichloropropene	10.835	75	2522	0.802	ug/l	92
54) cis-1,3-Dichloropropene	10.312	75	2725	0.811	ug/l #	83
55) 1,1,2-Trichloroethane	11.018	97	1901	0.935	ug/l	91
56) Ethyl methacrylate	10.871	69	1923	2.471	ug/l #	88
57) 1,3-Dichloropropane	11.165	76	3113	0.881	ug/l	95
58) 2-Chloroethyl Vinyl ether	10.153	63	4929	3.822	ug/l	91
59) 2-Hexanone	11.194	43	7905	4.058	ug/l	87
60) Dibromochloromethane	11.359	129	2339	0.953	ug/l	98
61) 1,2-Dibromoethane	11.465	107	2022	1.000	ug/l	90
64) Tetrachloroethene	11.106	164	1658	0.947	ug/l #	83
65) Chlorobenzene	11.888	112	5401	0.960	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.959	131	2185	1.058	ug/l #	64
67) Ethyl Benzene	11.959	91	7347	0.802	ug/l	94
68) m/p-Xylenes	12.070	106	5059	1.494	ug/l	100
69) o-Xylene	12.394	106	2478	0.766	ug/l	97
70) Styrene	12.406	104	3812	0.712	ug/l	99
71) Bromoform	12.582	173	1208	0.818	ug/l #	88
73) Isopropylbenzene	12.688	105	5506	0.820	ug/l	98
74) N-amyl acetate	12.494	43	2716	0.899	ug/l #	87
75) 1,1,1,2-Tetrachloroethane	12.935	83	2616	1.102	ug/l #	96
76) 1,2,3-Trichloropropane	12.982	75	2198m	1.088	ug/l	
77) Bromobenzene	12.976	156	1733	0.987	ug/l	97
78) n-propylbenzene	13.029	91	6424	0.808	ug/l	93
79) 2-Chlorotoluene	13.117	91	4483	0.871	ug/l	97
80) 1,3,5-Trimethylbenzene	13.165	105	4301	0.775	ug/l	90
82) 4-Chlorotoluene	13.223	91	4268	0.833	ug/l	99
83) tert-Butylbenzene	13.435	119	3753	0.805	ug/l	96
84) 1,2,4-Trimethylbenzene	13.482	105	3758	0.679	ug/l	93
85) sec-Butylbenzene	13.612	105	4614	0.714	ug/l	99
86) p-Isopropyltoluene	13.723	119	3538	0.716	ug/l	98
87) 1,3-Dichlorobenzene	13.723	146	3038	0.940	ug/l	95
88) 1,4-Dichlorobenzene	13.806	146	3517m	1.050	ug/l	
89) n-Butylbenzene	14.053	91	3527	0.761	ug/l	93
90) Hexachloroethane	14.341	117	1355	1.102	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	3515	1.090	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.711	75	402	0.928	ug/l	74
93) 1,2,4-Trichlorobenzene	15.394	180	1310	0.874	ug/l #	80
94) Hexachlorobutadiene	15.500	225	822	1.033	ug/l	89
95) Naphthalene	15.641	128	4210	0.941	ug/l #	90
96) 1,2,3-Trichlorobenzene	15.835	180	1494	0.985	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085443.D
Acq On : 14 Jan 2025 17:19
Operator : JC\MD
Sample : VSTDIC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:46:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion:168 Resp: 166880
Ion Ratio Lower Upper
168 100
99 68.6 53.6 80.4




Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 1.054 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 85 Resp: 2382
Ion Ratio Lower Upper
85 100
87 28.5 16.7 50.1



Abundance Scan 69 (2.359 min): VN085439.D\data.ms (-62)



#3

Chloromethane

Concen: 1.145 ug/l

RT: 2.359 min Scan# 69

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19


Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Abundance Scan 69 (2.359 min): VN085443.D\data.ms



Tgt Ion: 50 Resp: 2801

Ion Ratio Lower Upper

50 100

52 29.7 27.0 40.6


Manual Integrations

APPROVED


Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025


Abundance Scan 69 (2.359 min): VN085443.D\data.ms (-18)



Abundance



Abundance Scan 95 (2.512 min): VN085439.D\data.ms (-87)



#4

Vinyl Chloride

Concen: 1.111 ug/l

RT: 2.512 min Scan# 95

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19


Tgt Ion: 62 Resp: 2732

Ion Ratio Lower Upper


62 100

64 41.4 24.8 37.2


Abundance Scan 95 (2.512 min): VN085443.D\data.ms




Abundance




Abundance Scan 95 (2.512 min): VN085443.D\data.ms (-44)





#6
Chloroethane
Concen: 1.160 ug/l
RT: 3.124 min Scan# 198
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 64 Resp: 1809
Ion Ratio Lower Upper
64 100
66 34.1 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#7
Trichlorofluoromethane
Concen: 1.082 ug/l
RT: 3.512 min Scan# 265
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:101 Resp: 3860
Ion Ratio Lower Upper
101 100
103 55.0 51.4 77.2









#8
Diethyl Ether
Concen: 1.238 ug/l
RT: 3.953 min Scan# 340
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 74 Resp: 1526
Ion Ratio Lower Upper
74 100
45 80.9 49.7 149.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.073 ug/l m
RT: 4.359 min Scan# 409
Delta R.T. -0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:101 Resp: 2156
Ion Ratio Lower Upper
101 100
85 35.3 37.8 56.8#
151 66.4 58.8 88.2





#12

1,1-Dichloroethene

Concen: 0.927 ug/l

RT: 4.330 min Scan# 406

Delta R.T. -0.012 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC001



Tgt Ion:	96	Resp:	1660
Ion	Ratio	Lower	Upper
96	100		
61	122.1	141.0	211.4#
98	48.1	49.0	73.6#

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#14

Allyl chloride


Concen: 1.094 ug/l

RT: 5.024 min Scan# 522



Delta R.T. 0.006 min


Lab File: VN085443.D

Acq: 14 Jan 2025 17:19



Tgt Ion:	41	Resp:	3179
Ion	Ratio	Lower	Upper
41	100		
39	73.1	64.4	96.6
76	21.4	30.5	45.7#





#15

Acrylonitrile

Concen: 5.070 ug/l

RT: 5.718 min Scan# 64

Delta R.T. 0.006 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC001



Tgt Ion:	53	Resp:	4967
Ion	Ratio	Lower	Upper
53	100		
52	45.1	65.5	98.3
51	33.3	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#16

Acetone


Concen: 5.862 ug/l

RT: 4.436 min Scan# 422



Delta R.T. 0.018 min


Lab File: VN085443.D

Acq: 14 Jan 2025 17:19



Tgt Ion:	43	Resp:	5102
Ion	Ratio	Lower	Upper
43	100		
58	33.8	23.8	35.6





#17

Carbon Disulfide

Concen: 1.197 ug/l

RT: 4.706 min Scan# 468

Delta R.T. -0.000 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC001



Tgt Ion	Ratio	Lower	Upper
76	100		
78	13.2	6.9	10.3

**Manual Integrations
APPROVED**

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#18

Methyl Acetate


Concen: 1.099 ug/l

RT: 5.030 min Scan# 523



Delta R.T. 0.012 min


Lab File: VN085443.D

Acq: 14 Jan 2025 17:19



Tgt Ion	Ratio	Lower	Upper
43	100		
74	13.8	18.6	28.0





#19

Methyl tert-butyl Ether

Concen: 0.887 ug/l

RT: 5.789 min Scan# 65

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 73 Resp: 5156

Ion Ratio Lower Upper

73 100





57 28.4 17.8 26.8#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#20

Methylene Chloride

Concen: 1.016 ug/l

RT: 5.265 min Scan# 563

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 84 Resp: 2190




Ion Ratio Lower Upper


84 100

49 115.6 104.1 156.1

51 36.9 31.0 46.4

86 48.9 51.9 77.9#





#21

trans-1,2-Dichloroethene

Concen: 1.102 ug/l

RT: 5.794 min Scan# 651

Delta R.T. 0.012 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 2110

Ion Ratio Lower Upper

96 100

61 98.7 115.7 173.5#





98 41.8 49.8 74.8#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether

Concen: 0.882 ug/l

RT: 6.677 min Scan# 803

Delta R.T. 0.012 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 45 Resp: 5691




Ion Ratio Lower Upper


45 100

43 65.9 45.0 67.6

87 35.0 21.8 32.8#

59 11.3 10.2 15.2





#23

Vinyl Acetate

Concen: 4.090 ug/l m

RT: 6.606 min Scan# 791

Delta R.T. 0.012 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 43 Resp: 18526

Ion Ratio Lower Upper

43 100





86 9.8 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 1.022 ug/l m

RT: 6.583 min Scan# 787

Delta R.T. 0.018 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 63 Resp: 4020


Ion Ratio Lower Upper

63 100

98 0.0 3.3 9.8#

100 0.0 2.0 6.0#





#25

2-Butanone

Concen: 5.034 ug/l m

RT: 7.483 min Scan# 940

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 6448

Ion Ratio Lower Upper

43 100

72 18.0 20.2 30.4#

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#26

2,2-Dichloropropane

Concen: 0.976 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 77 Resp: 3104

Ion Ratio Lower Upper

77 100

97 17.7 10.7 32.1





#27

cis-1,2-Dichloroethene

Concen: 0.969 ug/l

RT: 7.477 min Scan# 93

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 2185

Ion Ratio Lower Upper

96 100

61 166.1 0.0 311.8





98 68.7 0.0 126.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#28

Bromochloromethane

Concen: 1.138 ug/l

RT: 7.812 min Scan# 996

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 49 Resp: 2082


Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.4

130 78.0 55.0 82.4









#29
Tetrahydrofuran
Concen: 4.548 ug/l
RT: 7.835 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 42 Resp: 3693
Ion Ratio Lower Upper
42 100
72 43.1 34.2 51.2
71 37.9 32.5 48.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#30
Chloroform
Concen: 1.045 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion: 83 Resp: 4250
Ion Ratio Lower Upper
83 100
85 80.2 51.8 77.6#





#32
1,1,1-Trichloroethane
Concen: 1.031 ug/l
RT: 8.165 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 97 Resp: 3678
Ion Ratio Lower Upper
97 100
99 0.0 49.8 74.6#
61 52.2 41.4 62.2

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 8.57 min

Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 65
Sig Exp Ratio
65 100
67 50.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 12

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:114 Resp: 302938

Ion Ratio Lower Upper

114 100

63 22.3 0.0 47.6





88 16.8 0.0 32.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 8.16 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19


Tgt Ion: 113


Sig Exp Ratio

113 100

111 103.4

192 17.9





#36

1,1-Dichloropropene

Concen: 1.046 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 75 Resp: 3084

Ion Ratio Lower Upper

75 100

110 30.1 16.5 49.5





77 15.9 24.4 36.6#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#37

Ethyl Acetate

Concen: 1.026 ug/l m

RT: 7.577 min Scan# 956

Delta R.T. 0.023 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 43 Resp: 3056


Ion Ratio Lower Upper

43 100

61 4.3 10.9 16.3#

70 0.0 7.5 11.3#





#38

Carbon Tetrachloride

Concen: 1.018 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 117 Resp: 3436

Ion Ratio Lower Upper

117 100

119 106.6 75.4 113.2





121 31.0 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#39

Methylcyclohexane

Concen: 0.867 ug/l

RT: 9.600 min Scan# 1300

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 83 Resp: 2403


Ion Ratio Lower Upper

83 100

55 72.7 62.6 94.0

98 31.8 37.7 56.5#





#40

Benzene

Concen: 0.957 ug/l

RT: 8.606 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 78 Resp: 8480

Ion Ratio Lower Upper

78 100





77 28.3 19.0 28.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#41

Methacrylonitrile

Concen: 0.941 ug/l

RT: 7.771 min Scan# 989

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 41 Resp: 1458




Ion Ratio Lower Upper


41 100

39 37.6 46.0 69.0#

67 53.8 57.4 86.2#

52 18.5 25.5 38.3#





#42

1,2-Dichloroethane

Concen: 0.938 ug/l

RT: 8.665 min Scan# 1141

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 62 Resp: 3131

Ion Ratio Lower Upper

62 100





98 5.4 0.0 17.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#43

Isopropyl Acetate

Concen: 1.007 ug/l

RT: 8.682 min Scan# 1144

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 43 Resp: 4803


Ion Ratio Lower Upper

43 100

61 18.9 20.7 31.1#

87 7.3 9.8 14.8#





#44

Trichloroethene

Concen: 1.034 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001



Tgt Ion:130 Resp: 2133

Ion Ratio Lower Upper

130 100




95 126.1 0.0 195.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#45

1,2-Dichloropropane


Concen: 0.992 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19






Tgt Ion: 63 Resp: 2246

Ion Ratio Lower Upper

63 100

65 32.6 25.6 38.4





#46

Dibromomethane

Concen: 1.083 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC001



Tgt Ion:	93	Resp:	1769
Ion	Ratio	Lower	Upper
93	100		
95	86.5	64.7	97.1
174	77.4	69.0	103.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#47

Bromodichloromethane


Concen: 0.882 ug/l

RT: 9.882 min Scan# 1348



Delta R.T. 0.000 min


Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion:	83	Resp:	2935
Ion	Ratio	Lower	Upper
83	100		
85	72.9	51.2	76.8
127	9.0	6.5	9.7





#48
Methyl methacrylate
Concen: 0.878 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 41 Resp: 1886
Ion Ratio Lower Upper
41 100
69 76.2 64.7 97.1
39 62.8 49.0 73.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#49
1,4-Dioxane
Concen: 18.616 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 88 Resp: 673
Ion Ratio Lower Upper
88 100
43 0.0 26.6 39.8#
58 31.1 59.5 89.3#





#50

Toluene-d8

Concen: 0.000 ug/l

Expected RT: 10.56 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 98

Sig Exp Ratio

98	100
100	65.3

Instrument :

MSVOA_N

ClientSampleId :



VSTDICC001

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#51

4-Methyl-2-Pentanone

Concen: 4.163 ug/l

RT: 10.441 min Scan# 1443

Delta R.T. 0.000 min




Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 11523

Ion Ratio Lower Upper

43	100		
58	33.6	30.5	45.7





#52

Toluene

Concen: 0.814 ug/l

RT: 10.618 min Scan# 1473

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 92 Resp: 4180

Ion Ratio Lower Upper

92 100





91 183.9 139.2 208.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#53

t-1,3-Dichloropropene

Concen: 0.802 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. -0.000 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 75 Resp: 2522

Ion Ratio Lower Upper

75 100

77 33.5 23.5 35.3





#54

cis-1,3-Dichloropropene

Concen: 0.811 ug/l

RT: 10.312 min Scan# 14

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 75 Resp: 2725

Ion Ratio Lower Upper

75 100

77 39.0 25.0 37.4#





39 67.6 43.1 64.7#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#55

1,1,2-Trichloroethane

Concen: 0.935 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 97 Resp: 1901




Ion Ratio Lower Upper


97 100

83 101.5 69.0 103.6

85 60.3 44.6 66.8

99 59.9 48.6 72.8





#56

Ethyl methacrylate

Concen: 2.471 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion:	69	Resp:	1923
Ion	Ratio	Lower	Upper
69	100		
41	74.4	54.6	82.0
39	52.4	32.4	48.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#57

1,3-Dichloropropane

Concen: 0.881 ug/l




RT: 11.165 min Scan# 1566


Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion:	76	Resp:	3113
Ion	Ratio	Lower	Upper
76	100		
78	34.9	25.6	38.4





#58

2-Chloroethyl Vinyl ether

Concen: 3.822 ug/l

RT: 10.153 min Scan# 1395

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 63 Resp: 4929

Ion Ratio Lower Upper

63 100





106 31.7 21.6 32.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#59

2-Hexanone

Concen: 4.058 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 43 Resp: 7905

Ion Ratio Lower Upper

43 100

58 42.9 26.2 78.6





#60

Dibromochloromethane

Concen: 0.953 ug/l

RT: 11.359 min Scan# 1598

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion:129 Resp: 2339

Ion Ratio Lower Upper

129 100

127 78.9 38.6 115.8

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#61

1,2-Dibromoethane

Concen: 1.000 ug/l

RT: 11.465 min Scan# 1617

Delta R.T. -0.000 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion:107 Resp: 2022

Ion Ratio Lower Upper

107 100

109 84.8 75.9 113.9





#62

4-Bromofluorobenzene

Concen: 0.000 ug/l

Expected RT: 12.85 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 95

Sig Exp Ratio

95 100

174 72.5

176 71.2

Instrument :

MSVOA_N

ClientSampleId :



VSTDICC001

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion:117 Resp: 256851


Ion Ratio Lower Upper

117 100

82 59.0 48.6 72.8

119 32.2 26.6 39.8





#64

Tetrachloroethene

Concen: 0.947 ug/l

RT: 11.106 min Scan# 1555

Delta R.T. 0.006 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001



Tgt Ion:164 Resp: 1658

Ion Ratio Lower Upper

164 100

166 152.1 103.4 155.2

129 128.3 79.2 118.8#




131 96.9 77.1 115.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#65

Chlorobenzene

Concen: 0.960 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. -0.000 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion:112 Resp: 5401

Ion Ratio Lower Upper

112 100

114 35.3 25.3 37.9





#66

1,1,1,2-Tetrachloroethane

Concen: 1.058 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001






Tgt Ion:	131	Resp:	2185
Ion Ratio	Lower	Upper	
131	100		
133	97.7	47.4	142.3
119	0.0	33.1	99.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#67

Ethyl Benzene


Concen: 0.802 ug/l

RT: 11.959 min Scan# 1701



Delta R.T. -0.000 min


Lab File: VN085443.D

Acq: 14 Jan 2025 17:19



Tgt Ion:	91	Resp:	7347
Ion Ratio	Lower	Upper	
91	100		
106	32.8	23.8	35.8





#68

m/p-Xylenes

Concen: 1.494 ug/l

RT: 12.070 min Scan# 1719

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:106 Resp: 5059

Ion Ratio Lower Upper

106 100





91 209.3 167.7 251.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#69

o-Xylene

Concen: 0.766 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. -0.000 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion:106 Resp: 2478

Ion Ratio Lower Upper

106 100

91 225.4 110.4 331.2





#70

Styrene

Concen: 0.712 ug/l

RT: 12.406 min Scan# 17

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion:	104	Resp:	3812
Ion Ratio	Lower	Upper	
104	100		
78	52.9	42.5	63.7
103	53.7	43.8	65.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#71

Bromoform

Concen: 0.818 ug/l




RT: 12.582 min Scan# 1807


Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion:	173	Resp:	1208
Ion Ratio	Lower	Upper	
173	100		
175	40.7	24.4	73.2
254	0.0	0.0	0.0






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC001




Tgt Ion:152 Resp: 99539
Ion Ratio Lower Upper
152 100
115 63.8 31.1 93.3
150 158.5 0.0 343.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#73
Isopropylbenzene
Concen: 0.820 ug/l
RT: 12.688 min Scan# 1825
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19



Tgt Ion:105 Resp: 5506
Ion Ratio Lower Upper
105 100
120 26.7 12.8 38.3





#74

N-amy1 acetate

Concen: 0.899 ug/l

RT: 12.494 min Scan# 1791

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 2716

Ion Ratio Lower Upper

43 100

70 30.4 34.0 51.0#

55 24.2 21.4 32.2





61 17.6 19.0 28.4#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#75

1,1,2,2-Tetrachloroethane

Concen: 1.102 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 83 Resp: 2616


Ion Ratio Lower Upper

83 100

131 2.4 4.8 14.4#

85 65.7 32.2 96.6





#76

1,2,3-Trichloropropane

Concen: 1.088 ug/l m

RT: 12.982 min Scan# 18

Delta R.T. -0.006 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001



Tgt Ion: 75 Resp: 2198

Ion Ratio Lower Upper

75 100




77 184.2 109.7 329.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#77

Bromobenzene


Concen: 0.987 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19





Tgt Ion:156 Resp: 1733


Ion Ratio Lower Upper

156 100

77 233.6 114.1 342.4

158 94.9 48.9 146.8





#78

n-propylbenzene

Concen: 0.808 ug/l

RT: 13.029 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 91 Resp: 6424

Ion Ratio Lower Upper

91 100





120 18.2 10.9 32.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#79

2-Chlorotoluene

Concen: 0.871 ug/l

RT: 13.117 min Scan# 1898

Delta R.T. -0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 91 Resp: 4483

Ion Ratio Lower Upper

91 100

126 32.9 15.7 47.1





#80

1,3,5-Trimethylbenzene

Concen: 0.775 ug/l

RT: 13.165 min Scan# 19

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 105 Resp: 4301

Ion Ratio Lower Upper

105 100





120 41.2 23.9 71.7

Manual Integrations

APPROVED

Reviewed By : John Carlone 01/15/2025

Supervised By : Mahesh Dadoda 01/15/2025



#82

4-Chlorotoluene

Concen: 0.833 ug/l

RT: 13.223 min Scan# 1916

Delta R.T. 0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 91 Resp: 4268

Ion Ratio Lower Upper

91 100

126 31.0 15.9 47.7





#83

tert-Butylbenzene

Concen: 0.805 ug/l

RT: 13.435 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC001



Tgt Ion:	119	Resp:	3753
Ion	Ratio	Lower	Upper
119	100		
91	65.9	35.1	105.3
134	24.9	12.0	36.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#84

1,2,4-Trimethylbenzene


Concen: 0.679 ug/l

RT: 13.482 min Scan# 1960



Delta R.T. 0.006 min


Lab File: VN085443.D

Acq: 14 Jan 2025 17:19



Tgt Ion:	105	Resp:	3758
Ion	Ratio	Lower	Upper
105	100		
120	47.7	21.6	65.0





#85

sec-Butylbenzene

Concen: 0.714 ug/l

RT: 13.612 min Scan# 1982

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion:105 Resp: 4614

Ion Ratio Lower Upper

105 100





134 19.9 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#86

p-Isopropyltoluene

Concen: 0.716 ug/l

RT: 13.723 min Scan# 2001

Delta R.T. 0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion:119 Resp: 3538


Ion Ratio Lower Upper

119 100

134 25.6 12.7 38.0

91 23.2 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 0.940 ug/l

RT: 13.723 min Scan# 20

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC001

Tgt	Ion	146	Resp:	3038
Ion	Ratio	Lower	Upper	
146	100			
111	43.0	21.4	64.3	
148	70.9	32.3	96.9	

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#88

1,4-Dichlorobenzene

Concen: 1.050 ug/l m




RT: 13.806 min Scan# 2015


Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt	Ion	146	Resp:	3517
Ion	Ratio	Lower	Upper	
146	100			
111	37.2	21.3	63.7	
148	61.2	32.4	97.0	





#89

n-Butylbenzene

Concen: 0.761 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt	Ion	91	Resp:	3527
Ion	Ratio	Lower	Upper	
91	100			
92	57.2	25.8	77.3	
134	25.5	11.7	35.1	

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90

Hexachloroethane

Concen: 1.102 ug/l




RT: 14.341 min Scan# 2106


Delta R.T. 0.012 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt	Ion	117	Resp:	1355
Ion	Ratio	Lower	Upper	
117	100			
201	66.5	33.7	101.0	





#91

1,2-Dichlorobenzene

Concen: 1.090 ug/l

RT: 14.106 min Scan# 2065

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt	Ion	146	Resp:	3515
Ion	Ratio	Lower	Upper	
146	100			
111	43.4	21.7	65.1	
148	57.8	31.4	94.2	

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.928 ug/l




RT: 14.711 min Scan# 2169


Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt	Ion	75	Resp:	402
Ion	Ratio	Lower	Upper	
75	100			
155	48.0	36.4	109.2	
157	112.7	45.4	136.1	





#93

1,2,4-Trichlorobenzene

Concen: 0.874 ug/l

RT: 15.394 min Scan# 2284

Delta R.T. 0.006 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC001



Tgt Ion:	180	Resp:	1310
Ion Ratio	Lower	Upper	
180	100		
182	106.3	46.8	140.3
145	9.7	16.0	48.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene


Concen: 1.033 ug/l

RT: 15.500 min Scan# 2303



Delta R.T. -0.000 min


Lab File: VN085443.D

Acq: 14 Jan 2025 17:19



Tgt Ion:	225	Resp:	822
Ion Ratio	Lower	Upper	
225	100		
223	50.7	30.7	92.1
227	55.7	30.9	92.5





#95

Naphthalene

Concen: 0.941 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt Ion:	128	Resp:	4210
Ion	Ratio	Lower	Upper
128	100		
127	11.1	10.6	16.0
129	5.3	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96

1,2,3-Trichlorobenzene

Concen: 0.985 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion:	180	Resp:	1494
Ion	Ratio	Lower	Upper
180	100		
182	90.2	47.4	142.2
145	11.8	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	219757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	370487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	320948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	158525	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.571	65	176594	49.782	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.560%
35) Dibromofluoromethane	8.165	113	130487	50.768	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	10.565	98	477630	52.302	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.600%
62) 4-Bromofluorobenzene	12.847	95	165249	52.899	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.800%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	130242	43.772	ug/l	100
3) Chloromethane	2.360	50	138152	42.883	ug/l	99
4) Vinyl Chloride	2.513	62	138316	42.715	ug/l	99
5) Bromomethane	2.954	94	80019	40.910	ug/l	91
6) Chloroethane	3.124	64	86837	42.299	ug/l	99
7) Trichlorofluoromethane	3.501	101	206526	43.956	ug/l	99
8) Diethyl Ether	3.959	74	69545	42.846	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	118310	44.706	ug/l	98
10) Methyl Iodide	4.595	142	140190	46.259	ug/l	98
11) Tert butyl alcohol	5.518	59	83933	206.633	ug/l	99
12) 1,1-Dichloroethene	4.342	96	106880	45.317	ug/l	98
13) Acrolein	4.177	56	135720	244.765	ug/l	99
14) Allyl chloride	5.024	41	170308	44.501	ug/l	99
15) Acrylonitrile	5.712	53	294062	227.920	ug/l	99
16) Acetone	4.418	43	245248	213.971	ug/l	98
17) Carbon Disulfide	4.712	76	300215	41.342	ug/l	100
18) Methyl Acetate	5.018	43	156938	45.028	ug/l	100
19) Methyl tert-butyl Ether	5.795	73	365819	47.768	ug/l	100
20) Methylene Chloride	5.271	84	122059	43.017	ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	112597	44.672	ug/l	93
22) Diisopropyl ether	6.665	45	399206	46.996	ug/l	99
23) Vinyl Acetate	6.601	43	1471068	247.459	ug/l	100
24) 1,1-Dichloroethane	6.571	63	228489	44.114	ug/l	98
25) 2-Butanone	7.477	43	386125	228.924	ug/l	100
26) 2,2-Dichloropropane	7.489	77	199303	47.593	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	134920	45.446	ug/l	98
28) Bromochloromethane	7.812	49	106323	44.122	ug/l	100
29) Tetrahydrofuran	7.836	42	255650	239.065	ug/l	99
30) Chloroform	7.965	83	235923	44.071	ug/l	98
31) Cyclohexane	8.253	56	193034	42.878	ug/l	93
32) 1,1,1-Trichloroethane	8.165	97	208172	44.332	ug/l	99
36) 1,1-Dichloropropene	8.365	75	161834	44.863	ug/l	98
37) Ethyl Acetate	7.559	43	160656	44.123	ug/l	97
38) Carbon Tetrachloride	8.359	117	183592	44.456	ug/l	99
39) Methylcyclohexane	9.600	83	168431	49.687	ug/l	98
40) Benzene	8.600	78	494871	45.657	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	89326	47.155	ug/l	99
42) 1,2-Dichloroethane	8.665	62	185663	45.481	ug/l	100
43) Isopropyl Acetate	8.683	43	265107	45.440	ug/l	99
44) Trichloroethene	9.347	130	110058	43.622	ug/l	96
45) 1,2-Dichloropropane	9.618	63	123876	44.744	ug/l	97
46) Dibromomethane	9.700	93	88325	44.214	ug/l	98
47) Bromodichloromethane	9.889	83	185808	45.655	ug/l	99
48) Methyl methacrylate	9.677	41	131142	49.948	ug/l	97
49) 1,4-Dioxane	9.689	88	35051	792.760	ug/l	98
51) 4-Methyl-2-Pentanone	10.441	43	832426	245.877	ug/l	99
52) Toluene	10.630	92	301911	48.073	ug/l	99
53) t-1,3-Dichloropropene	10.830	75	183680	47.757	ug/l	94
54) cis-1,3-Dichloropropene	10.312	75	196307	47.783	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	114757	46.173	ug/l	97
56) Ethyl methacrylate	10.871	69	182170	43.762	ug/l	98
57) 1,3-Dichloropropane	11.159	76	202583	46.877	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	403949	256.116	ug/l	99
59) 2-Hexanone	11.194	43	598433	251.213	ug/l	100
60) Dibromochloromethane	11.353	129	137831	45.935	ug/l	99
61) 1,2-Dibromoethane	11.465	107	110926	44.837	ug/l	100
64) Tetrachloroethene	11.100	164	98211	44.886	ug/l	96
65) Chlorobenzene	11.888	112	322723	45.901	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	115379	44.712	ug/l	99
67) Ethyl Benzene	11.959	91	577059	50.391	ug/l	97
68) m/p-Xylenes	12.071	106	437367	103.341	ug/l	100
69) o-Xylene	12.394	106	210698	52.093	ug/l	99
70) Styrene	12.406	104	355459	53.103	ug/l	100
71) Bromoform	12.577	173	90693	49.119	ug/l #	100
73) Isopropylbenzene	12.694	105	534428	49.951	ug/l	99
74) N-amyl acetate	12.488	43	234342	48.730	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	164882	43.627	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	134976m	41.891	ug/l	
77) Bromobenzene	12.977	156	128968	46.138	ug/l	97
78) n-propylbenzene	13.029	91	645977	51.001	ug/l	100
79) 2-Chlorotoluene	13.124	91	392404	47.870	ug/l	99
80) 1,3,5-Trimethylbenzene	13.171	105	456048	51.604	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	58623m	49.226	ug/l	
82) 4-Chlorotoluene	13.218	91	398110	48.776	ug/l	100
83) tert-Butylbenzene	13.435	119	378478	50.997	ug/l	98
84) 1,2,4-Trimethylbenzene	13.477	105	457340	51.922	ug/l	100
85) sec-Butylbenzene	13.612	105	532578	51.777	ug/l	100
86) p-Isopropyltoluene	13.729	119	440284	47.795	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	236387	45.920	ug/l	99
88) 1,4-Dichlorobenzene	13.806	146	235319	44.102	ug/l	99
89) n-Butylbenzene	14.053	91	373227	50.578	ug/l	99
90) Hexachloroethane	14.329	117	88538	45.215	ug/l	100
91) 1,2-Dichlorobenzene	14.100	146	231646	45.113	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.718	75	31899	46.221	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	109157	45.739	ug/l	98
94) Hexachlorobutadiene	15.500	225	53358	42.112	ug/l	99
95) Naphthalene	15.641	128	335535	47.116	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	107583	44.551	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\  
Data File : VN085445.D  
Acq On    : 14 Jan 2025   18:06  
Operator  : JC\MD  
Sample    : VSTDICV050  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 12   Sample Multiplier: 1
```


Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025


Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425



Tgt Ion:168 Resp: 219757
Ion Ratio Lower Upper
168 100
99 66.9 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#2
Dichlorodifluoromethane
Concen: 43.772 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 85 Resp: 130242
Ion Ratio Lower Upper
85 100
87 33.2 16.7 50.1





#3

Chloromethane

Concen: 42.883 ug/l

RT: 2.360 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 50 Resp: 138152

Ion Ratio Lower Upper

50 100





52 34.4 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#4

Vinyl Chloride

Concen: 42.715 ug/l

RT: 2.513 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 62 Resp: 138316

Ion Ratio Lower Upper

62 100

64 31.4 24.8 37.2





#5

Bromomethane

Concen: 40.910 ug/l

RT: 2.954 min Scan# 170

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 94 Resp: 80019

Ion Ratio Lower Upper

94 100





96 98.2 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#6

Chloroethane

Concen: 42.299 ug/l

RT: 3.124 min Scan# 199

Delta R.T. 0.006 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 64 Resp: 86837

Ion Ratio Lower Upper

64 100

66 31.1 24.6 36.8





#7

Trichlorofluoromethane

Concen: 43.956 ug/l

RT: 3.501 min Scan# 26

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion:101 Resp: 206526

Ion Ratio Lower Upper

101 100





103 65.1 51.4 77.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#8

Diethyl Ether

Concen: 42.846 ug/l

RT: 3.959 min Scan# 341

Delta R.T. 0.006 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 74 Resp: 69545

Ion Ratio Lower Upper

74 100

45 99.0 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.706 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Tgt Ion:101 Resp: 118310
Ion Ratio Lower Upper
101 100
85 45.9 37.8 56.8
151 71.5 58.8 88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#10
Methyl Iodide
Concen: 46.259 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:142 Resp: 140190
Ion Ratio Lower Upper
142 100
127 49.2 40.3 60.5
141 13.9 12.8 19.2





#11

Tert butyl alcohol

Concen: 206.633 ug/l

RT: 5.518 min Scan# 605

Delta R.T. 0.006 min

Lab File: VN085445.D


Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425



Tgt Ion: 59 Resp: 83933

Ion Ratio Lower Upper

59 100




57 10.4 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#12

1,1-Dichloroethene

Concen: 45.317 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 96 Resp: 106880


Ion Ratio Lower Upper

96 100

61 171.9 141.0 211.4

98 61.0 49.0 73.6





#13

Acrolein

Concen: 244.765 ug/l

RT: 4.177 min Scan# 37

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 56 Resp: 135720

Ion Ratio Lower Upper

56 100





55 70.9 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#14

Allyl chloride

Concen: 44.501 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 41 Resp: 170308


Ion Ratio Lower Upper

41 100

39 78.9 64.4 96.6

76 38.2 30.5 45.7





#15

Acrylonitrile

Concen: 227.920 ug/l

RT: 5.712 min Scan# 639

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 53 Resp: 294062

Ion Ratio Lower Upper

53 100

52 82.7 65.5 98.3





51 36.8 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#16

Acetone

Concen: 213.971 ug/l

RT: 4.418 min Scan# 419

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 43 Resp: 245248

Ion Ratio Lower Upper

43 100

58 28.4 23.8 35.6





#17

Carbon Disulfide

Concen: 41.342 ug/l

RT: 4.712 min Scan# 468

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion: 76 Resp: 300215

Ion Ratio Lower Upper

76 100

78 8.7 6.9 10.3

Instrument :

MSVOA_N

ClientSampleId :





ICVVN011425

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#18

Methyl Acetate

Concen: 45.028 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 43 Resp: 156938

Ion Ratio Lower Upper

43 100

74 23.4 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 47.768 ug/l

RT: 5.795 min Scan# 65

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 73 Resp: 365819

Ion Ratio Lower Upper

73 100





57 22.3 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#20

Methylene Chloride

Concen: 43.017 ug/l

RT: 5.271 min Scan# 564

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion: 84 Resp: 122059




Ion Ratio Lower Upper


84 100

49 131.3 104.1 156.1

51 41.0 31.0 46.4

86 63.3 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 44.672 ug/l

RT: 5.783 min Scan# 651

Delta R.T. 0.000 min

Lab File: VN085445.D


Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :




ICVVN011425



Tgt Ion:	96	Resp:	112597
Ion	Ratio	Lower	Upper
96	100		
61	134.3	115.7	173.5
98	59.1	49.8	74.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether


Concen: 46.996 ug/l

RT: 6.665 min Scan# 801



Delta R.T. 0.000 min


Lab File: VN085445.D

Acq: 14 Jan 2025 18:06



Tgt Ion:	45	Resp:	399206
Ion	Ratio	Lower	Upper
45	100		
43	56.0	45.0	67.6
87	26.8	21.8	32.8
59	12.2	10.2	15.2





#23

Vinyl Acetate

Concen: 247.459 ug/l

RT: 6.601 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 43 Resp: 1471068

Ion Ratio Lower Upper

43 100





86 10.7 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 44.114 ug/l

RT: 6.571 min Scan# 785

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 63 Resp: 228489


Ion Ratio Lower Upper

63 100

98 7.5 3.3 9.8

100 4.2 2.0 6.0





#25

2-Butanone

Concen: 228.924 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 43 Resp: 386125

Ion Ratio Lower Upper

43 100





72 25.4 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#26

2,2-Dichloropropane

Concen: 47.593 ug/l

RT: 7.489 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 77 Resp: 199303

Ion Ratio Lower Upper

77 100


97 20.9 10.7 32.1





#27
cis-1,2-Dichloroethene
Concen: 45.446 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425



Tgt Ion: 96 Resp: 134920
Ion Ratio Lower Upper
96 100
61 158.0 0.0 311.8
98 64.1 0.0 126.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#28
Bromochloromethane
Concen: 44.122 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 49 Resp: 106323
Ion Ratio Lower Upper
49 100
129 1.8 0.0 4.4
130 68.6 55.0 82.4





#29

Tetrahydrofuran

Concen: 239.065 ug/l

RT: 7.836 min Scan# 10

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 42 Resp: 255650

Ion Ratio Lower Upper

42 100

72 42.5 34.2 51.2





71 39.9 32.5 48.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#30

Chloroform

Concen: 44.071 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 83 Resp: 235923

Ion Ratio Lower Upper

83 100

85 63.5 51.8 77.6





#31

Cyclohexane

Concen: 42.878 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 56 Resp: 193034

Ion Ratio Lower Upper

56 100

69 30.5 22.2 33.4





84 89.4 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#32

1,1,1-Trichloroethane

Concen: 44.332 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 97 Resp: 208172


Ion Ratio Lower Upper

97 100

99 61.8 49.8 74.6

61 52.8 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 49.782 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 65 Resp: 176594

Ion Ratio Lower Upper

65 100





67 50.9 0.0 101.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:114 Resp: 370487


Ion Ratio Lower Upper

114 100

63 21.6 0.0 47.6

88 15.8 0.0 32.6





#35

Dibromofluoromethane

Concen: 50.768 ug/l

RT: 8.165 min Scan# 1055

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion:113 Resp: 130487

Ion Ratio Lower Upper

113 100

111 102.5 82.7 124.1





192 17.7 14.3 21.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#36

1,1-Dichloropropene

Concen: 44.863 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 75 Resp: 161834


Ion Ratio Lower Upper

75 100

110 32.4 16.5 49.5

77 31.8 24.4 36.6





#37

Ethyl Acetate

Concen: 44.123 ug/l

RT: 7.559 min Scan# 952

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 43 Resp: 160656

Ion Ratio Lower Upper

43 100

61 11.9 10.9 16.3





70 10.2 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#38

Carbon Tetrachloride

Concen: 44.456 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:117 Resp: 183592


Ion Ratio Lower Upper

117 100

119 94.4 75.4 113.2

121 29.0 24.6 37.0





#39

Methylcyclohexane

Concen: 49.687 ug/l

RT: 9.600 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 83 Resp: 168431

Ion Ratio Lower Upper

83 100

55 80.9 62.6 94.0





98 46.9 37.7 56.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#40

Benzene

Concen: 45.657 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 78 Resp: 494871

Ion Ratio Lower Upper

78 100

77 23.5 19.0 28.6





#41

Methacrylonitrile

Concen: 47.155 ug/l

RT: 7.777 min Scan# 99

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 41 Resp: 89326

Ion Ratio Lower Upper

41 100

39 56.6 46.0 69.0

67 70.9 57.4 86.2





52 32.6 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#42

1,2-Dichloroethane

Concen: 45.481 ug/l

RT: 8.665 min Scan# 1141

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 62 Resp: 185663

Ion Ratio Lower Upper

62 100

98 8.3 0.0 17.0





#43

Isopropyl Acetate

Concen: 45.440 ug/l

RT: 8.683 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :





ICVVN011425

Tgt Ion:	43	Resp:	265107
Ion	Ratio	Lower	Upper
43	100		
61	26.7	20.7	31.1
87	12.9	9.8	14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#44

Trichloroethene

Concen: 43.622 ug/l




RT: 9.347 min Scan# 1257


Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion:	130	Resp:	110058
Ion	Ratio	Lower	Upper
130	100		
95	101.9	0.0	195.8





#45

1,2-Dichloropropane

Concen: 44.744 ug/l

RT: 9.618 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 63 Resp: 123876

Ion Ratio Lower Upper

63 100





65 33.5 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#46

Dibromomethane

Concen: 44.214 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 93 Resp: 88325


Ion Ratio Lower Upper

93 100

95 82.9 64.7 97.1

174 88.0 69.0 103.4





#47

Bromodichloromethane

Concen: 45.655 ug/l

RT: 9.889 min Scan# 1348

Delta R.T. 0.007 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 83 Resp: 185808

Ion Ratio Lower Upper

83 100

85 63.5 51.2 76.8





127 7.8 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#48

Methyl methacrylate

Concen: 49.948 ug/l

RT: 9.677 min Scan# 1313

Delta R.T. -0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 41 Resp: 131142


Ion Ratio Lower Upper

41 100

69 78.6 64.7 97.1

39 63.3 49.0 73.6





#49

1,4-Dioxane

Concen: 792.760 ug/l

RT: 9.689 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 88 Resp: 35051

Ion Ratio Lower Upper

88 100

43 33.8 26.6 39.8





58 76.1 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#50

Toluene-d8

Concen: 52.302 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 98 Resp: 477630

Ion Ratio Lower Upper

98 100

100 64.7 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 245.877 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Tgt Ion: 43 Resp: 832426
Ion Ratio Lower Upper
43 100
58 37.8 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 48.073 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 92 Resp: 301911
Ion Ratio Lower Upper
92 100
91 175.7 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 47.757 ug/l

RT: 10.830 min Scan# 1510

Delta R.T. -0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 75 Resp: 183680

Ion Ratio Lower Upper

75 100





77 32.6 23.5 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#54

cis-1,3-Dichloropropene

Concen: 47.783 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 75 Resp: 196307


Ion Ratio Lower Upper

75 100

77 30.2 25.0 37.4

39 51.9 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 46.173 ug/l

RT: 11.012 min Scan# 15

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 97 Resp: 114757

Ion Ratio Lower Upper

97 100

83 83.7 69.0 103.6

85 53.2 44.6 66.8





99 62.8 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#56

Ethyl methacrylate

Concen: 43.762 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 69 Resp: 182170


Ion Ratio Lower Upper

69 100

41 70.4 54.6 82.0

39 40.2 32.4 48.6









#57
1,3-Dichloropropane
Concen: 46.877 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Tgt Ion: 76 Resp: 202583
Ion Ratio Lower Upper
76 100
78 32.1 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#58
2-Chloroethyl Vinyl ether
Concen: 256.116 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 63 Resp: 403949
Ion Ratio Lower Upper
63 100
106 26.4 21.6 32.4





#59

2-Hexanone

Concen: 251.213 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 43 Resp: 598433

Ion Ratio Lower Upper

43 100





58 52.1 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#60

Dibromochloromethane

Concen: 45.935 ug/l

RT: 11.353 min Scan# 1598

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion:129 Resp: 137831

Ion Ratio Lower Upper

129 100

127 78.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 44.837 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Tgt Ion: 107 Resp: 110926
Ion Ratio Lower Upper
107 100
109 94.7 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#62
4-Bromofluorobenzene
Concen: 52.899 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 95 Resp: 165249
Ion Ratio Lower Upper
95 100
174 73.1 0.0 145.0
176 70.8 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion:117 Resp: 320948

Ion Ratio Lower Upper

117 100

82 60.1 48.6 72.8





119 32.4 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#64

Tetrachloroethene

Concen: 44.886 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion:164 Resp: 98211




Ion Ratio Lower Upper


164 100

166 124.4 103.4 155.2

129 104.8 79.2 118.8

131 93.1 77.1 115.7





#65

Chlorobenzene

Concen: 45.901 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion:112 Resp: 322723

Ion Ratio Lower Upper

112 100





114 32.4 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 44.712 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:131 Resp: 115379


Ion Ratio Lower Upper

131 100

133 95.8 47.4 142.3

119 65.2 33.1 99.5





#67

Ethyl Benzene

Concen: 50.391 ug/l

RT: 11.959 min Scan# 17

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 91 Resp: 577059

Ion Ratio Lower Upper

91 100





106 31.2 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#68

m/p-Xylenes

Concen: 103.341 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.006 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion:106 Resp: 437367

Ion Ratio Lower Upper

106 100

91 210.1 167.7 251.5





#69

o-Xylene

Concen: 52.093 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion:106 Resp: 210698

Ion Ratio Lower Upper

106 100





91 219.9 110.4 331.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#70

Styrene

Concen: 53.103 ug/l

RT: 12.406 min Scan# 1777

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:104 Resp: 355459


Ion Ratio Lower Upper

104 100

78 52.8 42.5 63.7

103 54.5 43.8 65.8





#71

Bromoform

Concen: 49.119 ug/l

RT: 12.577 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N





ClientSampleId :

ICVVN011425

Tgt Ion:	173	Resp:	90693
Ion Ratio	Lower	Upper	
173	100		
175	48.7	24.4	73.2
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion:	152	Resp:	158525
Ion Ratio	Lower	Upper	
152	100		
115	62.2	31.1	93.3
150	171.0	0.0	343.6





#73

Isopropylbenzene

Concen: 49.951 ug/l

RT: 12.694 min Scan# 1825

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion: 105 Resp: 534428

Ion Ratio Lower Upper

105 100

120 25.2 12.8 38.3

Instrument :

MSVOA_N

ClientSampleId :





ICVVN011425

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#74

N-ethyl acetate

Concen: 48.730 ug/l

RT: 12.488 min Scan# 1791

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 234342




Ion Ratio Lower Upper


43 100

70 42.2 34.0 51.0

55 26.5 21.4 32.2

61 23.8 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 43.627 ug/l

RT: 12.935 min Scan# 18

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :





ICVVN011425

Tgt Ion:	83	Resp:	164882
Ion	Ratio	Lower	Upper
83	100		
131	9.9	4.8	14.4
85	64.6	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#76

1,2,3-Trichloropropane

Concen: 41.891 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion:	75	Resp:	134976
Ion	Ratio	Lower	Upper
75	100		
77	213.9	109.7	329.2





#77

Bromobenzene

Concen: 46.138 ug/l

RT: 12.977 min Scan# 1874

Delta R.T. 0.000 min

Lab File: VN085445.D


Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :




ICVVN011425



Tgt Ion:	156	Resp:	128968
Ion Ratio	Lower	Upper	
156	100		
77	223.8	114.1	342.4
158	95.7	48.9	146.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#78

n-propylbenzene

Concen: 51.001 ug/l




RT: 13.029 min Scan# 1883


Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion:	91	Resp:	645977
Ion Ratio	Lower	Upper	
91	100		
120	21.7	10.9	32.6





#79

2-Chlorotoluene

Concen: 47.870 ug/l

RT: 13.124 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion: 91 Resp: 392404

Ion Ratio Lower Upper

91 100





126 31.8 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#80

1,3,5-Trimethylbenzene

Concen: 51.604 ug/l

RT: 13.171 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion:105 Resp: 456048

Ion Ratio Lower Upper

105 100

120 48.1 23.9 71.7





#81

trans-1,4-Dichloro-2-butene

Concen: 49.226 ug/l m

RT: 12.735 min Scan# 1833

Delta R.T. 0.000 min

Lab File: VN085445.D


Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :




ICVVN011425



Tgt Ion:	75	Resp:	58623
Ion	Ratio	Lower	Upper
75	100		
53	118.0	95.6	143.4
89	43.3	37.0	55.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#82

4-Chlorotoluene

Concen: 48.776 ug/l




RT: 13.218 min Scan# 1915


Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06


Tgt Ion:	91	Resp:	398110
Ion	Ratio	Lower	Upper
91	100		
126	31.6	15.9	47.7





#83
tert-Butylbenzene
Concen: 50.997 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425



Tgt Ion:119 Resp: 378478
Ion Ratio Lower Upper
119 100
91 70.6 35.1 105.3
134 26.7 12.0 36.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#84
1,2,4-Trimethylbenzene
Concen: 51.922 ug/l
RT: 13.477 min Scan# 1959
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:105 Resp: 457340
Ion Ratio Lower Upper
105 100
120 43.6 21.6 65.0





#85

sec-Butylbenzene

Concen: 51.777 ug/l

RT: 13.612 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425

Tgt Ion:105 Resp: 532578

Ion Ratio Lower Upper

105 100





134 19.2 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#86

p-Isopropyltoluene

Concen: 47.795 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:119 Resp: 440284


Ion Ratio Lower Upper

119 100

134 25.3 12.7 38.0

91 25.0 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 45.920 ug/l

RT: 13.729 min Scan# 20

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N





ClientSampleId :

ICVVN011425

Tgt Ion:	146	Resp:	236387
Ion	Ratio	Lower	Upper
146	100		
111	42.6	21.4	64.3
148	63.8	32.3	96.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#88

1,4-Dichlorobenzene

Concen: 44.102 ug/l




RT: 13.806 min Scan# 2015


Delta R.T. -0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion:	146	Resp:	235319
Ion	Ratio	Lower	Upper
146	100		
111	41.9	21.3	63.7
148	63.5	32.4	97.0





#89

n-Butylbenzene

Concen: 50.578 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N





ClientSampleId :

ICVVN011425

Tgt Ion:	91	Resp:	373227
Ion	Ratio	Lower	Upper
91	100		
92	51.9	25.8	77.3
134	23.8	11.7	35.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90

Hexachloroethane

Concen: 45.215 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06


Tgt Ion:	117	Resp:	88538
Ion	Ratio	Lower	Upper
117	100		
201	67.2	33.7	101.0





#91
1,2-Dichlorobenzene
Concen: 45.113 ug/l
RT: 14.100 min Scan# 2065
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425



Tgt Ion: 146 Resp: 231646
Ion Ratio Lower Upper
146 100
111 43.9 21.7 65.1
148 63.1 31.4 94.2




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 46.221 ug/l
RT: 14.718 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 31899
Ion Ratio Lower Upper
75 100
155 69.0 36.4 109.2
157 92.3 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 45.739 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN085445.D


Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :




ICVVN011425



Tgt Ion:	180	Resp:	109157
Ion	Ratio	Lower	Upper
180	100		
182	96.3	46.8	140.3
145	32.7	16.0	48.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene


Concen: 42.112 ug/l

RT: 15.500 min Scan# 2303



Delta R.T. 0.000 min


Lab File: VN085445.D

Acq: 14 Jan 2025 18:06



Tgt Ion:	225	Resp:	53358
Ion	Ratio	Lower	Upper
225	100		
223	59.8	30.7	92.1
227	61.2	30.9	92.5





#95

Naphthalene

Concen: 47.116 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N





ClientSampleId :

ICVVN011425

Tgt Ion:	128	Resp:	335535
Ion Ratio	Lower	Upper	
128	100		
127	13.2	10.6	16.0
129	11.1	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96

1,2,3-Trichlorobenzene

Concen: 44.551 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Tgt Ion:	180	Resp:	107583
Ion Ratio	Lower	Upper	
180	100		
182	95.6	47.4	142.2
145	34.3	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.677	0.593	12.4	104	0.00
3 P	Chloromethane	0.733	0.629	14.2	102	0.00
4 C	Vinyl Chloride	0.737	0.629	14.7#	101	0.00
5 T	Bromomethane	0.445	0.364	18.2	96	0.00
6 T	Chloroethane	0.467	0.395	15.4	103	0.00
7 T	Trichlorofluoromethane	1.069	0.940	12.1	104	0.00
8 T	Diethyl Ether	0.369	0.316	14.4	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.538	10.6	109	0.00
10 T	Methyl Iodide	0.690	0.638	7.5	98	0.00
11 T	Tert butyl alcohol	0.092	0.076	17.4	90	0.00
12 CM	1,1-Dichloroethene	0.537	0.486	9.5#	101	0.00
13 T	Acrolein	0.126	0.124	1.6	97	0.00
14 T	Allyl chloride	0.871	0.775	11.0	101	0.00
15 T	Acrylonitrile	0.294	0.268	8.8	99	0.00
16 T	Acetone	0.261	0.223	14.6	98	0.00
17 T	Carbon Disulfide	1.652	1.366	17.3	102	0.00
18 T	Methyl Acetate	0.793	0.714	10.0	100	0.00
19 T	Methyl tert-butyl Ether	1.742	1.665	4.4	98	0.00
20 T	Methylene Chloride	0.646	0.555	14.1	97	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.512	10.6	102	0.00
22 T	Diisopropyl ether	1.933	1.817	6.0	98	0.00
23 T	Vinyl Acetate	1.353	1.339	1.0	99	0.00
24 P	1,1-Dichloroethane	1.178	1.040	11.7	98	0.00
25 T	2-Butanone	0.384	0.351	8.6	99	0.00
26 T	2,2-Dichloropropane	0.953	0.907	4.8	107	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.614	9.0	99	0.00
28 T	Bromochloromethane	0.548	0.484	11.7	98	0.00
29 T	Tetrahydrofuran	0.243	0.233	4.1	98	0.00
30 C	Chloroform	1.218	1.074	11.8#	101	0.00
31 T	Cyclohexane	1.024	0.878	14.3	110	0.00
32 T	1,1,1-Trichloroethane	1.068	0.947	11.3	103	0.00
33 S	1,2-Dichloroethane-d4	0.807	0.804	0.4	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.347	0.352	-1.4	107	0.00
36 T	1,1-Dichloropropene	0.487	0.437	10.3	103	0.00
37 T	Ethyl Acetate	0.491	0.434	11.6	97	0.00
38 T	Carbon Tetrachloride	0.557	0.496	11.0	102	0.00
39 T	Methylcyclohexane	0.457	0.455	0.4	107	0.00
40 TM	Benzene	1.463	1.336	8.7	100	0.00
41 T	Methacrylonitrile	0.256	0.241	5.9	97	0.00
42 TM	1,2-Dichloroethane	0.551	0.501	9.1	100	0.00
43 T	Isopropyl Acetate	0.787	0.716	9.0	96	0.00
44 TM	Trichloroethene	0.341	0.297	12.9	100	0.00
45 C	1,2-Dichloropropane	0.374	0.334	10.7#	98	0.00
46 T	Dibromomethane	0.270	0.238	11.9	97	0.00
47 T	Bromodichloromethane	0.549	0.502	8.6	98	0.00
48 T	Methyl methacrylate	0.354	0.354	0.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.005	16.7	81	0.00
50 S	Toluene-d8	1.232	1.289	-4.6	111	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.449	1.8	99	0.00
52 CM	Toluene	0.848	0.815	3.9#	102	0.00
53 T	t-1,3-Dichloropropene	0.519	0.496	4.4	98	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.530	4.3	98	0.00
55 T	1,1,2-Trichloroethane	0.335	0.310	7.5	99	0.00
56 T	Ethyl methacrylate	0.469	0.492	-4.9	99	0.00
57 T	1,3-Dichloropropane	0.583	0.547	6.2	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.218	-2.3	99	0.00
59 T	2-Hexanone	0.321	0.323	-0.6	99	0.00
60 T	Dibromochloromethane	0.405	0.372	8.1	98	0.00
61 T	1,2-Dibromoethane	0.334	0.299	10.5	98	0.00
62 S	4-Bromofluorobenzene	0.422	0.446	-5.7	108	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64 T	Tetrachloroethene	0.341	0.306	10.3	102	0.00
65 PM	Chlorobenzene	1.095	1.006	8.1	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.359	10.7	99	0.00
67 C	Ethyl Benzene	1.784	1.798	-0.8#	104	0.00
68 T	m/p-Xylenes	0.659	0.681	-3.3	104	0.00
69 T	o-Xylene	0.630	0.656	-4.1	104	0.00
70 T	Styrene	1.043	1.108	-6.2	102	0.00
71 P	Bromoform	0.288	0.283	1.7	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.375	3.371	0.1	106	0.00
74 T	N-amyl acetate	1.517	1.478	2.6	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.192	1.040	12.8	98	0.00
76 T	1,2,3-Trichloropropane	1.016	0.851	16.2	103	0.00
77 T	Bromobenzene	0.882	0.814	7.7	103	0.00
78 T	n-propylbenzene	3.995	4.075	-2.0	106	0.00
79 T	2-Chlorotoluene	2.586	2.475	4.3	104	0.00
80 T	1,3,5-Trimethylbenzene	2.787	2.877	-3.2	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.370	1.6	103	0.00
82 T	4-Chlorotoluene	2.574	2.511	2.4	105	0.00
83 T	tert-Butylbenzene	2.341	2.387	-2.0	106	0.00
84 T	1,2,4-Trimethylbenzene	2.778	2.885	-3.9	103	0.00
85 T	sec-Butylbenzene	3.244	3.360	-3.6	106	0.00
86 T	p-Isopropyltoluene	2.631	2.777	-5.5	105	0.00
87 T	1,3-Dichlorobenzene	1.624	1.491	8.2	103	0.00
88 T	1,4-Dichlorobenzene	1.683	1.484	11.8	103	0.00
89 T	n-Butylbenzene	2.327	2.354	-1.2	103	0.00
90 T	Hexachloroethane	0.618	0.559	9.5	105	0.00
91 T	1,2-Dichlorobenzene	1.620	1.461	9.8	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.201	7.8	98	0.00
93 T	1,2,4-Trichlorobenzene	0.753	0.689	8.5	95	0.00
94 T	Hexachlorobutadiene	0.400	0.337	15.8	99	0.00
95 T	Naphthalene	2.246	2.117	5.7	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.679	10.9	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	50.000	43.772	12.5	104	0.00
3 P	Chloromethane	50.000	42.883	14.2	102	0.00
4 C	Vinyl Chloride	50.000	42.715	14.6#	101	0.00
5 T	Bromomethane	50.000	40.910	18.2	96	0.00
6 T	Chloroethane	50.000	42.299	15.4	103	0.00
7 T	Trichlorofluoromethane	50.000	43.956	12.1	104	0.00
8 T	Diethyl Ether	50.000	42.846	14.3	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	44.706	10.6	109	0.00
10 T	Methyl Iodide	50.000	46.259	7.5	98	0.00
11 T	Tert butyl alcohol	250.000	206.633	17.3	90	0.00
12 CM	1,1-Dichloroethene	50.000	45.317	9.4#	101	0.00
13 T	Acrolein	250.000	244.765	2.1	97	0.00
14 T	Allyl chloride	50.000	44.501	11.0	101	0.00
15 T	Acrylonitrile	250.000	227.920	8.8	99	0.00
16 T	Acetone	250.000	213.971	14.4	98	0.00
17 T	Carbon Disulfide	50.000	41.342	17.3	102	0.00
18 T	Methyl Acetate	50.000	45.028	9.9	100	0.00
19 T	Methyl tert-butyl Ether	50.000	47.768	4.5	98	0.00
20 T	Methylene Chloride	50.000	43.017	14.0	97	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.672	10.7	102	0.00
22 T	Diisopropyl ether	50.000	46.996	6.0	98	0.00
23 T	Vinyl Acetate	250.000	247.459	1.0	99	0.00
24 P	1,1-Dichloroethane	50.000	44.114	11.8	98	0.00
25 T	2-Butanone	250.000	228.924	8.4	99	0.00
26 T	2,2-Dichloropropane	50.000	47.593	4.8	107	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.446	9.1	99	0.00
28 T	Bromochloromethane	50.000	44.122	11.8	98	0.00
29 T	Tetrahydrofuran	250.000	239.065	4.4	98	0.00
30 C	Chloroform	50.000	44.071	11.9#	101	0.00
31 T	Cyclohexane	50.000	42.878	14.2	110	0.00
32 T	1,1,1-Trichloroethane	50.000	44.332	11.3	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.782	0.4	107	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	109	0.00
35 S	Dibromofluoromethane	50.000	50.768	-1.5	107	0.00
36 T	1,1-Dichloropropene	50.000	44.863	10.3	103	0.00
37 T	Ethyl Acetate	50.000	44.123	11.8	97	0.00
38 T	Carbon Tetrachloride	50.000	44.456	11.1	102	0.00
39 T	Methylcyclohexane	50.000	49.687	0.6	107	0.00
40 TM	Benzene	50.000	45.657	8.7	100	0.00
41 T	Methacrylonitrile	50.000	47.155	5.7	97	0.00
42 TM	1,2-Dichloroethane	50.000	45.481	9.0	100	0.00
43 T	Isopropyl Acetate	50.000	45.440	9.1	96	0.00
44 TM	Trichloroethene	50.000	43.622	12.8	100	0.00
45 C	1,2-Dichloropropane	50.000	44.744	10.5#	98	0.00
46 T	Dibromomethane	50.000	44.214	11.6	97	0.00
47 T	Bromodichloromethane	50.000	45.655	8.7	98	0.00
48 T	Methyl methacrylate	50.000	49.948	0.1	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	792.760	20.7	81	0.00
50 S	Toluene-d8	50.000	52.302	-4.6	111	0.00
51 T	4-Methyl-2-Pentanone	250.000	245.877	1.6	99	0.00
52 CM	Toluene	50.000	48.073	3.9#	102	0.00
53 T	t-1,3-Dichloropropene	50.000	47.757	4.5	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.783	4.4	98	0.00
55 T	1,1,2-Trichloroethane	50.000	46.173	7.7	99	0.00
56 T	Ethyl methacrylate	50.000	43.762	12.5	99	0.00
57 T	1,3-Dichloropropane	50.000	46.877	6.2	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	256.116	-2.4	99	0.00
59 T	2-Hexanone	250.000	251.213	-0.5	99	0.00
60 T	Dibromochloromethane	50.000	45.935	8.1	98	0.00
61 T	1,2-Dibromoethane	50.000	44.837	10.3	98	0.00
62 S	4-Bromofluorobenzene	50.000	52.899	-5.8	108	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	108	0.00
64 T	Tetrachloroethene	50.000	44.886	10.2	102	0.00
65 PM	Chlorobenzene	50.000	45.901	8.2	101	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	44.712	10.6	99	0.00
67 C	Ethyl Benzene	50.000	50.391	-0.8#	104	0.00
68 T	m/p-Xylenes	100.000	103.341	-3.3	104	0.00
69 T	o-Xylene	50.000	52.093	-4.2	104	0.00
70 T	Styrene	50.000	53.103	-6.2	102	0.00
71 P	Bromoform	50.000	49.119	1.8	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	108	0.00
73 T	Isopropylbenzene	50.000	49.951	0.1	106	0.00
74 T	N-amyl acetate	50.000	48.730	2.5	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	43.627	12.7	98	0.00
76 T	1,2,3-Trichloropropane	50.000	41.891	16.2	103	0.00
77 T	Bromobenzene	50.000	46.138	7.7	103	0.00
78 T	n-propylbenzene	50.000	51.001	-2.0	106	0.00
79 T	2-Chlorotoluene	50.000	47.870	4.3	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.604	-3.2	106	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.226	1.5	103	0.00
82 T	4-Chlorotoluene	50.000	48.776	2.4	105	0.00
83 T	tert-Butylbenzene	50.000	50.997	-2.0	106	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.922	-3.8	103	0.00
85 T	sec-Butylbenzene	50.000	51.777	-3.6	106	0.00
86 T	p-Isopropyltoluene	50.000	47.795	4.4	105	0.00
87 T	1,3-Dichlorobenzene	50.000	45.920	8.2	103	0.00
88 T	1,4-Dichlorobenzene	50.000	44.102	11.8	103	0.00
89 T	n-Butylbenzene	50.000	50.578	-1.2	103	0.00
90 T	Hexachloroethane	50.000	45.215	9.6	105	0.00
91 T	1,2-Dichlorobenzene	50.000	45.113	9.8	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.221	7.6	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	45.739	8.5	95	0.00
94 T	Hexachlorobutadiene	50.000	42.112	15.8	99	0.00
95 T	Naphthalene	50.000	47.116	5.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	44.551	10.9	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199

Instrument ID: MSVOA_N Calibration Date/Time: 01/28/2025 09:53

Lab File ID: VN085526.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.733	0.642	0.1	-12.41	20
Vinyl Chloride	0.737	0.665		-9.77	20
Bromomethane	0.445	0.389		-12.58	20
Chloroethane	0.467	0.439		-6	20
Trichlorofluoromethane	1.069	1.071		0.19	20
1,1,2-Trichlorotrifluoroethane	0.602	0.618		2.66	20
1,1-Dichloroethene	0.537	0.526		-2.05	20
Acetone	0.261	0.263		0.77	20
Carbon Disulfide	1.652	1.391		-15.8	20
Methyl tert-butyl Ether	1.742	1.940		11.37	20
Methylene Chloride	0.646	0.645		-0.16	20
trans-1,2-Dichloroethene	0.573	0.570		-0.52	20
1,1-Dichloroethane	1.178	1.202	0.1	2.04	20
2-Butanone	0.384	0.399		3.91	20
Carbon Tetrachloride	0.557	0.602		8.08	20
cis-1,2-Dichloroethene	0.675	0.700		3.7	20
Chloroform	1.218	1.248		2.46	20
1,1,1-Trichloroethane	1.068	1.105		3.46	20
Methylcyclohexane	0.457	0.513		12.25	20
Benzene	1.463	1.575		7.66	20
1,2-Dichloroethane	0.551	0.585		6.17	20
Trichloroethene	0.341	0.359		5.28	20
1,2-Dichloropropane	0.374	0.407		8.82	20
Bromodichloromethane	0.549	0.613		11.66	20
4-Methyl-2-Pentanone	0.457	0.535		17.07	20
Toluene	0.848	0.975		14.98	20
t-1,3-Dichloropropene	0.519	0.611		17.73	20
cis-1,3-Dichloropropene	0.554	0.639		15.34	20
1,1,2-Trichloroethane	0.335	0.375		11.94	20
2-Hexanone	0.321	0.386		20.25	20
Dibromochloromethane	0.405	0.460		13.58	20
Tetrachloroethene	0.341	0.361		5.86	20
Chlorobenzene	1.095	1.171	0.3	6.94	20
Ethyl Benzene	1.784	2.053		15.08	20
m/p-Xylenes	0.659	0.785		19.12	20
o-Xylene	0.630	0.739		17.3	20
Styrene	1.043	1.279		22.63	20
Bromoform	0.288	0.334	0.1	15.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199

Instrument ID: MSVOA_N Calibration Date/Time: 01/28/2025 09:53

Lab File ID: VN085526.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.375	3.736		10.7	20
1,1,2,2-Tetrachloroethane	1.192	1.170	0.3	-1.85	20
1,3-Dichlorobenzene	1.624	1.699		4.62	20
1,4-Dichlorobenzene	1.683	1.700		1.01	20
1,2-Dichlorobenzene	1.620	1.644		1.48	20
1,2-Dichloroethane-d4	0.807	0.777		-3.72	20
Dibromofluoromethane	0.347	0.355		2.31	20
Toluene-d8	1.232	1.296		5.2	20
4-Bromofluorobenzene	0.422	0.467		10.66	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085526.D
 Acq On : 28 Jan 2025 09:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/29/2025
 Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	203638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	333806	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	302551	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	155713	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.577	65	158323	48.164	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.320%
35) Dibromofluoromethane	8.159	113	118662	51.240	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.480%
50) Toluene-d8	10.565	98	432537	52.569	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.847	95	155870	55.379	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.760%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	126280	45.800	ug/l	98
3) Chloromethane	2.359	50	130825	43.823	ug/l	99
4) Vinyl Chloride	2.512	62	135494	45.155	ug/l	98
5) Bromomethane	2.948	94	79311	43.757	ug/l	98
6) Chloroethane	3.118	64	89399	46.993	ug/l	100
7) Trichlorofluoromethane	3.495	101	218167	50.109	ug/l	99
8) Diethyl Ether	3.959	74	73428	48.819	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.371	101	125888	51.335	ug/l	99
10) Methyl Iodide	4.583	142	144081	51.306	ug/l	97
11) Tert butyl alcohol	5.518	59	110319	293.091	ug/l	100
12) 1,1-Dichloroethene	4.336	96	107139	49.022	ug/l	98
13) Acrolein	4.171	56	134509	261.783	ug/l	100
14) Allyl chloride	5.018	41	172649	48.683	ug/l	98
15) Acrylonitrile	5.712	53	312035	260.994	ug/l	99
16) Acetone	4.418	43	267575	251.929	ug/l	97
17) Carbon Disulfide	4.712	76	283263	42.095	ug/l	98
18) Methyl Acetate	5.018	43	174889	54.151	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	395014	55.663	ug/l	98
20) Methylene Chloride	5.277	84	131284	49.931	ug/l	99
21) trans-1,2-Dichloroethene	5.783	96	116065	49.692	ug/l	97
22) Diisopropyl ether	6.665	45	426926	54.237	ug/l	98
23) Vinyl Acetate	6.600	43	1520538	276.027	ug/l	99
24) 1,1-Dichloroethane	6.565	63	244849	51.014	ug/l	98
25) 2-Butanone	7.477	43	406723	260.223	ug/l	100
26) 2,2-Dichloropropane	7.483	77	223673	57.640	ug/l	100
27) cis-1,2-Dichloroethene	7.483	96	142501	51.799	ug/l	99
28) Bromochloromethane	7.812	49	101613	45.505	ug/l	98
29) Tetrahydrofuran	7.830	42	272222	274.712	ug/l	100
30) Chloroform	7.959	83	254209	51.246	ug/l	100
31) Cyclohexane	8.253	56	189607	45.451	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	225078	51.727	ug/l	98
36) 1,1-Dichloropropene	8.365	75	170755	52.537	ug/l	99
37) Ethyl Acetate	7.553	43	173618	52.923	ug/l	97
38) Carbon Tetrachloride	8.359	117	200977	54.013	ug/l	99
39) Methylcyclohexane	9.600	83	171154	56.039	ug/l	99
40) Benzene	8.600	78	525669	53.828	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085526.D
 Acq On : 28 Jan 2025 09:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/29/2025
 Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	95095	55.717	ug/l	99
42) 1,2-Dichloroethane	8.665	62	195201	53.072	ug/l	99
43) Isopropyl Acetate	8.683	43	287826	54.755	ug/l	100
44) Trichloroethene	9.347	130	119734	52.672	ug/l	95
45) 1,2-Dichloropropane	9.618	63	135814	54.447	ug/l	97
46) Dibromomethane	9.706	93	95715	53.179	ug/l	99
47) Bromodichloromethane	9.883	83	204517	55.774	ug/l	99
48) Methyl methacrylate	9.677	41	136479	57.693	ug/l	97
49) 1,4-Dioxane	9.688	88	40607	1019.345	ug/l	100
51) 4-Methyl-2-Pentanone	10.441	43	892528	292.599	ug/l	100
52) Toluene	10.630	92	325356	57.499	ug/l	100
53) t-1,3-Dichloropropene	10.830	75	204105	58.899	ug/l	96
54) cis-1,3-Dichloropropene	10.306	75	213245	57.610	ug/l	99
55) 1,1,2-Trichloroethane	11.012	97	125068	55.851	ug/l	98
56) Ethyl methacrylate	10.871	69	191924	50.844	ug/l	99
57) 1,3-Dichloropropane	11.159	76	219323	56.327	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	426978	300.465	ug/l	99
59) 2-Hexanone	11.194	43	643969	300.034	ug/l	99
60) Dibromochloromethane	11.359	129	153599	56.815	ug/l	99
61) 1,2-Dibromoethane	11.465	107	122098	54.776	ug/l	97
64) Tetrachloroethene	11.100	164	109264	52.974	ug/l	98
65) Chlorobenzene	11.888	112	354308	53.457	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	130012	53.447	ug/l	99
67) Ethyl Benzene	11.959	91	621051	57.531	ug/l	99
68) m/p-Xylenes	12.065	106	475266	119.124	ug/l	99
69) o-Xylene	12.394	106	223578	58.639	ug/l	100
70) Styrene	12.406	104	387081	61.343	ug/l	99
71) Bromoform	12.576	173	101032	58.046	ug/l #	99
73) Isopropylbenzene	12.694	105	581759	55.357	ug/l	100
74) N-amyl acetate	12.494	43	249973	52.919	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	182235	49.089	ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	147100m	46.478	ug/l	
77) Bromobenzene	12.976	156	138788	50.547	ug/l	99
78) n-propylbenzene	13.035	91	701218	56.362	ug/l	100
79) 2-Chlorotoluene	13.123	91	424851	52.764	ug/l	99
80) 1,3,5-Trimethylbenzene	13.171	105	494666	56.984	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	63601	54.370	ug/l	92
82) 4-Chlorotoluene	13.218	91	432608	53.960	ug/l	98
83) tert-Butylbenzene	13.435	119	405613	55.641	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	501298	57.941	ug/l	100
85) sec-Butylbenzene	13.612	105	580942	57.499	ug/l	100
86) p-Isopropyltoluene	13.729	119	491281	53.532	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	264631	52.335	ug/l	99
88) 1,4-Dichlorobenzene	13.806	146	264719	50.507	ug/l	98
89) n-Butylbenzene	14.053	91	420498	58.013	ug/l	99
90) Hexachloroethane	14.329	117	97383	50.630	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	255916	50.739	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	34993	51.620	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	127164	54.246	ug/l	100
94) Hexachlorobutadiene	15.500	225	63906	51.348	ug/l	99
95) Naphthalene	15.641	128	367063	52.473	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	121516	51.230	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085526.D
Acq On : 28 Jan 2025 09:53
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:33:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085526.D
Acq On : 28 Jan 2025 09:53
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025


Quant Time: Jan 29 00:33:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 168 Resp: 203638
Ion Ratio Lower Upper
168 100
99 63.5 53.6 80.4




Manual Integrations
APPROVED

Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025




#2
Dichlorodifluoromethane
Concen: 45.800 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 85 Resp: 126280
Ion Ratio Lower Upper
85 100
87 32.1 16.7 50.1



Abundance Scan 69 (2.359 min): VN085439.D\data.ms (-62)



#3

Chloromethane

Concen: 43.823 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 50 Resp: 130825

Ion Ratio Lower Upper

50 100


52 33.0 27.0 40.6

Manual Integrations



APPROVED

Reviewed By :John Carlone 01/29/2025


Supervised By :Semsettin Yesilyurt 01/29/2025



Abundance Scan 69 (2.359 min): VN085526.D\data.ms (-18)



Abundance Scan 95 (2.512 min): VN085439.D\data.ms (-87)



#4

Vinyl Chloride

Concen: 45.155 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN085526.D


Acq: 28 Jan 2025 09:53

Tgt Ion: 62 Resp: 135494



Ion Ratio Lower Upper


62 100

64 31.9 24.8 37.2




Abundance Scan 95 (2.512 min): VN085526.D\data.ms (-44)





#5
Bromomethane
Concen: 43.757 ug/l
RT: 2.948 min Scan# 169
Delta R.T. -0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 79311
Ion Ratio Lower Upper
94 100
96 91.4 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#6
Chloroethane
Concen: 46.993 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53


Tgt Ion: 64 Resp: 89399
Ion Ratio Lower Upper
64 100
66 30.6 24.6 36.8





#7
Trichlorofluoromethane
Concen: 50.109 ug/l
RT: 3.495 min Scan# 263
Delta R.T. -0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 218167
Ion Ratio Lower Upper
101 100
103 63.5 51.4 77.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#8
Diethyl Ether
Concen: 48.819 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 74 Resp: 73428
Ion Ratio Lower Upper
74 100
45 96.9 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.335 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 125888
Ion Ratio Lower Upper
101 100
85 46.4 37.8 56.8
151 73.8 58.8 88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#10
Methyl Iodide
Concen: 51.306 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:142 Resp: 144081
Ion Ratio Lower Upper
142 100
127 52.8 40.3 60.5
141 14.5 12.8 19.2





#11

Tert butyl alcohol

Concen: 293.091 ug/l

RT: 5.518 min Scan# 605

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 110319

Ion Ratio Lower Upper

59 100





57 10.8 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#12

1,1-Dichloroethene

Concen: 49.022 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 96 Resp: 107139


Ion Ratio Lower Upper

96 100

61 180.3 141.0 211.4

98 61.9 49.0 73.6





#13

Acrolein

Concen: 261.783 ug/l

RT: 4.171 min Scan# 377

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 134509

Ion Ratio Lower Upper

56 100





55 70.5 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#14

Allyl chloride

Concen: 48.683 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 41 Resp: 172649


Ion Ratio Lower Upper

41 100

39 82.7 64.4 96.6

76 38.6 30.5 45.7





#15

Acrylonitrile

Concen: 260.994 ug/l

RT: 5.712 min Scan# 639

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 53 Resp: 312035

Ion Ratio Lower Upper

53 100

52 82.4 65.5 98.3





51 36.9 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#16

Acetone

Concen: 251.929 ug/l

RT: 4.418 min Scan# 419

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 43 Resp: 267575

Ion Ratio Lower Upper

43 100

58 28.2 23.8 35.6





#17

Carbon Disulfide

Concen: 42.095 ug/l

RT: 4.712 min Scan# 468

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 283263

Ion Ratio Lower Upper

76 100





78 9.2 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#18

Methyl Acetate

Concen: 54.151 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 43 Resp: 174889

Ion Ratio Lower Upper

43 100

74 23.1 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 55.663 ug/l

RT: 5.789 min Scan# 65

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 73 Resp: 395014

Ion Ratio Lower Upper

73 100





57 21.3 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#20

Methylene Chloride

Concen: 49.931 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Tgt Ion: 84 Resp: 131284




Ion Ratio Lower Upper


84 100

49 128.3 104.1 156.1

51 40.2 31.0 46.4

86 64.3 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 49.692 ug/l

RT: 5.783 min Scan# 651

Delta R.T. 0.000 min

Lab File: VN085526.D


Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050



Tgt Ion: 96 Resp: 116065

Ion Ratio Lower Upper

96 100

61 148.4 115.7 173.5




98 59.7 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#22

Diisopropyl ether


Concen: 54.237 ug/l

RT: 6.665 min Scan# 801

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53



Tgt Ion: 45 Resp: 426926



Ion Ratio Lower Upper


45 100

43 55.8 45.0 67.6

87 25.6 21.8 32.8

59 11.3 10.2 15.2





#23

Vinyl Acetate

Concen: 276.027 ug/l

RT: 6.600 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 1520538





Ion	Ratio	Lower	Upper
43	100		
86	10.1	8.4	12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#24

1,1-Dichloroethane

Concen: 51.014 ug/l

RT: 6.565 min Scan# 784




Delta R.T. 0.000 min


Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Tgt Ion: 63 Resp: 244849

Ion	Ratio	Lower	Upper
63	100		
98	7.1	3.3	9.8
100	4.4	2.0	6.0





#25

2-Butanone

Concen: 260.223 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 406723

Ion Ratio Lower Upper

43 100





72 25.0 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#26

2,2-Dichloropropane

Concen: 57.640 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 77 Resp: 223673

Ion Ratio Lower Upper

77 100

97 21.3 10.7 32.1









#27
cis-1,2-Dichloroethene
Concen: 51.799 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 96 Resp: 142501
Ion Ratio Lower Upper
96 100
61 157.2 0.0 311.8
98 64.9 0.0 126.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#28
Bromochloromethane
Concen: 45.505 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53


Tgt Ion: 49 Resp: 101613
Ion Ratio Lower Upper
49 100
129 1.9 0.0 4.4
130 70.7 55.0 82.4





#29
Tetrahydrofuran
Concen: 274.712 ug/l
RT: 7.830 min Scan# 999
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 42 Resp: 272222
Ion Ratio Lower Upper
42 100
72 43.0 34.2 51.2
71 40.6 32.5 48.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025




#30
Chloroform
Concen: 51.246 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53



Tgt Ion: 83 Resp: 254209
Ion Ratio Lower Upper
83 100
85 64.5 51.8 77.6





#31

Cyclohexane

Concen: 45.451 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 189607

Ion Ratio Lower Upper

56 100

69 30.2 22.2 33.4





84 85.8 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#32

1,1,1-Trichloroethane

Concen: 51.727 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 97 Resp: 225078


Ion Ratio Lower Upper

97 100

99 64.9 49.8 74.6


61 50.9 41.4 62.2





#33
1,2-Dichloroethane-d4
Concen: 48.164 ug/l
RT: 8.577 min Scan# 1125
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 65 Resp: 158323
Ion Ratio Lower Upper
65 100
67 51.1 0.0 101.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.094 min Scan# 1214
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53



Tgt Ion: 114 Resp: 333806
Ion Ratio Lower Upper
114 100
63 22.7 0.0 47.6
88 16.4 0.0 32.6





#35

Dibromofluoromethane

Concen: 51.240 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 118662

Ion Ratio Lower Upper

113 100

111 103.2 82.7 124.1





192 18.0 14.3 21.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#36

1,1-Dichloropropene

Concen: 52.537 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 75 Resp: 170755


Ion Ratio Lower Upper

75 100

110 32.6 16.5 49.5

77 31.6 24.4 36.6





#37

Ethyl Acetate

Concen: 52.923 ug/l

RT: 7.553 min Scan# 952

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 173618

Ion Ratio Lower Upper

43 100

61 11.8 10.9 16.3





70 10.0 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#38

Carbon Tetrachloride

Concen: 54.013 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion:117 Resp: 200977


Ion Ratio Lower Upper

117 100

119 96.0 75.4 113.2

121 31.1 24.6 37.0





#39

Methylcyclohexane

Concen: 56.039 ug/l

RT: 9.600 min Scan# 1300

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 171154

Ion Ratio Lower Upper

83 100

55 79.2 62.6 94.0





98 47.9 37.7 56.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#40

Benzene

Concen: 53.828 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 78 Resp: 525669

Ion Ratio Lower Upper

78 100

77 23.9 19.0 28.6





#41

Methacrylonitrile

Concen: 55.717 ug/l

RT: 7.771 min Scan# 989

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 41 Resp: 95095

Ion Ratio Lower Upper

41 100

39 56.3 46.0 69.0

67 71.1 57.4 86.2





52 32.2 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#42

1,2-Dichloroethane

Concen: 53.072 ug/l

RT: 8.665 min Scan# 1141

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 62 Resp: 195201

Ion Ratio Lower Upper

62 100

98 8.8 0.0 17.0





#43

Isopropyl Acetate

Concen: 54.755 ug/l

RT: 8.683 min Scan# 1144

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 287826

Ion Ratio Lower Upper

43 100

61 25.9 20.7 31.1





87 12.6 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#44

Trichloroethene

Concen: 52.672 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion:130 Resp: 119734

Ion Ratio Lower Upper

130 100

95 103.0 0.0 195.8





#45

1,2-Dichloropropane

Concen: 54.447 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 135814

Ion Ratio Lower Upper

63 100





65 33.9 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#46

Dibromomethane

Concen: 53.179 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 93 Resp: 95715


Ion Ratio Lower Upper

93 100

95 81.2 64.7 97.1

174 87.0 69.0 103.4





#47

Bromodichloromethane

Concen: 55.774 ug/l

RT: 9.883 min Scan# 13

Delta R.T. 0.001 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 204517

Ion Ratio Lower Upper

83 100

85 65.1 51.2 76.8





127 8.4 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#48

Methyl methacrylate

Concen: 57.693 ug/l

RT: 9.677 min Scan# 1313

Delta R.T. -0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 41 Resp: 136479


Ion Ratio Lower Upper

41 100

69 83.2 64.7 97.1

39 64.1 49.0 73.6





#49

1,4-Dioxane

Concen: 1019.345 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 40607

Ion Ratio Lower Upper

88 100

43 33.0 26.6 39.8





58 74.6 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#50

Toluene-d8

Concen: 52.569 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 98 Resp: 432537

Ion Ratio Lower Upper

98 100

100 65.0 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 292.599 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 892528
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#52
Toluene
Concen: 57.499 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 92 Resp: 325356
Ion Ratio Lower Upper
92 100
91 173.6 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 58.899 ug/l

RT: 10.830 min Scan# 1510

Delta R.T. -0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 75 Resp: 204105

Ion Ratio Lower Upper

75 100





77 31.6 23.5 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#54

cis-1,3-Dichloropropene

Concen: 57.610 ug/l

RT: 10.306 min Scan# 1420

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 75 Resp: 213245


Ion Ratio Lower Upper

75 100

77 30.8 25.0 37.4


39 52.6 43.1 64.7





#55
1,1,2-Trichloroethane
Concen: 55.851 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 125068
Ion Ratio Lower Upper
97 100
83 84.4 69.0 103.6
85 57.9 44.6 66.8
99 62.7 48.6 72.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025




#56
Ethyl methacrylate
Concen: 50.844 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53



Tgt Ion: 69 Resp: 191924
Ion Ratio Lower Upper
69 100
41 68.4 54.6 82.0
39 42.2 32.4 48.6





#57

1,3-Dichloropropane

Concen: 56.327 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 219323

Ion Ratio Lower Upper

76 100





78 31.5 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#58

2-Chloroethyl Vinyl ether

Concen: 300.465 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 63 Resp: 426978

Ion Ratio Lower Upper

63 100

106 27.8 21.6 32.4









#59
2-Hexanone
Concen: 300.034 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 643969
Ion Ratio Lower Upper
43 100
58 51.8 26.2 78.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#60
Dibromochloromethane
Concen: 56.815 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:129 Resp: 153599
Ion Ratio Lower Upper
129 100
127 76.7 38.6 115.8









#61
1,2-Dibromoethane
Concen: 54.776 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 107 Resp: 122098
Ion Ratio Lower Upper
107 100
109 92.4 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025



#62
4-Bromofluorobenzene
Concen: 55.379 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 95 Resp: 155870
Ion Ratio Lower Upper
95 100
174 73.4 0.0 145.0
176 71.9 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:	117	Resp:	302551
Ion Ratio	Lower	Upper	
117	100		
82	58.5	48.6	72.8
119	31.2	26.6	39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#64
Tetrachloroethene
Concen: 52.974 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:	164	Resp:	109264
Ion Ratio	Lower	Upper	
164	100		
166	125.7	103.4	155.2
129	101.8	79.2	118.8
131	96.7	77.1	115.7





#65

Chlorobenzene

Concen: 53.457 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:112 Resp: 354308

Ion Ratio Lower Upper

112 100





114 31.2 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 53.447 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion:131 Resp: 130012


Ion Ratio Lower Upper

131 100

133 95.0 47.4 142.3

119 64.2 33.1 99.5





#67

Ethyl Benzene

Concen: 57.531 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 621051

Ion Ratio Lower Upper

91 100





106 30.2 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#68

m/p-Xylenes

Concen: 119.124 ug/l

RT: 12.065 min Scan# 1719

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion:106 Resp: 475266

Ion Ratio Lower Upper

106 100

91 207.6 167.7 251.5









#69
o-Xylene
Concen: 58.639 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 223578
Ion Ratio Lower Upper
106 100
91 220.9 110.4 331.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#70
Styrene
Concen: 61.343 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:104 Resp: 387081
Ion Ratio Lower Upper
104 100
78 52.7 42.5 63.7
103 55.4 43.8 65.8





#71

Bromoform

Concen: 58.046 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:173 Resp: 101032

Ion Ratio Lower Upper

173 100

175 49.5 24.4 73.2





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion:152 Resp: 155713


Ion Ratio Lower Upper

152 100

115 62.3 31.1 93.3

150 172.0 0.0 343.6





#73

Isopropylbenzene

Concen: 55.357 ug/l

RT: 12.694 min Scan# 18

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 581759

Ion Ratio Lower Upper

105 100





120 25.5 12.8 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#74

N-ethyl acetate

Concen: 52.919 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Tgt Ion: 43 Resp: 249973




Ion Ratio Lower Upper


43 100

70 42.1 34.0 51.0

55 26.4 21.4 32.2

61 22.9 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 49.089 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion:	83	Resp:	182235
Ion	Ratio	Lower	Upper
83	100		
131	9.9	4.8	14.4
85	63.9	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#76

1,2,3-Trichloropropane

Concen: 46.478 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Tgt Ion:	75	Resp:	147100
Ion	Ratio	Lower	Upper
75	100		
77	216.6	109.7	329.2





#77

Bromobenzene

Concen: 50.547 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. 0.000 min

Lab File: VN085526.D


Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050






Tgt Ion:	156	Resp:	138788
Ion	Ratio	Lower	Upper
156	100		
77	229.6	114.1	342.4
158	98.7	48.9	146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#78

n-propylbenzene


Concen: 56.362 ug/l

RT: 13.035 min Scan# 1884



Delta R.T. 0.006 min


Lab File: VN085526.D

Acq: 28 Jan 2025 09:53



Tgt Ion:	91	Resp:	701218
Ion	Ratio	Lower	Upper
91	100		
120	21.9	10.9	32.6





#79

2-Chlorotoluene

Concen: 52.764 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 424851

Ion Ratio Lower Upper

91 100





126 32.2 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#80

1,3,5-Trimethylbenzene

Concen: 56.984 ug/l

RT: 13.171 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion:105 Resp: 494666

Ion Ratio Lower Upper

105 100

120 47.7 23.9 71.7





#81

trans-1,4-Dichloro-2-butene

Concen: 54.370 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 75 Resp: 63601

Ion Ratio Lower Upper

75 100

53 107.1 95.6 143.4





89 46.8 37.0 55.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#82

4-Chlorotoluene

Concen: 53.960 ug/l

RT: 13.218 min Scan# 1915

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 91 Resp: 432608

Ion Ratio Lower Upper

91 100

126 32.8 15.9 47.7





#83

tert-Butylbenzene

Concen: 55.641 ug/l

RT: 13.435 min Scan# 19

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 405613

Ion Ratio Lower Upper

119 100

91 69.8 35.1 105.3





134 27.3 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#84

1,2,4-Trimethylbenzene

Concen: 57.941 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion:105 Resp: 501298

Ion Ratio Lower Upper

105 100

120 43.4 21.6 65.0





#85

sec-Butylbenzene

Concen: 57.499 ug/l

RT: 13.612 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 580942

Ion Ratio Lower Upper

105 100





134 19.4 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#86

p-Isopropyltoluene

Concen: 53.532 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion:119 Resp: 491281


Ion Ratio Lower Upper

119 100

134 25.5 12.7 38.0

91 24.7 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 52.335 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 264631

Ion Ratio Lower Upper

146 100

111 43.3 21.4 64.3





148 64.1 32.3 96.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#88

1,4-Dichlorobenzene

Concen: 50.507 ug/l

RT: 13.806 min Scan# 2015

Delta R.T. -0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion:146 Resp: 264719


Ion Ratio Lower Upper

146 100

111 40.8 21.3 63.7

148 63.7 32.4 97.0









#89
n-Butylbenzene
Concen: 58.013 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:	91	Resp:	420498
Ion	Ratio	Lower	Upper
91	100		
92	52.1	25.8	77.3
134	24.2	11.7	35.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#90
Hexachloroethane
Concen: 50.630 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53


Tgt Ion:	117	Resp:	97383
Ion	Ratio	Lower	Upper
117	100		
201	67.0	33.7	101.0





#91
1,2-Dichlorobenzene
Concen: 50.739 ug/l
RT: 14.106 min Scan# 2065
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 255916
Ion Ratio Lower Upper
146 100
111 43.8 21.7 65.1
148 62.7 31.4 94.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.620 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 75 Resp: 34993
Ion Ratio Lower Upper
75 100
155 69.5 36.4 109.2
157 90.0 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 54.246 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN085526.D


Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050



Tgt Ion:180 Resp: 127164

Ion Ratio Lower Upper

180 100

182 93.6 46.8 140.3




145 32.5 16.0 48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#94

Hexachlorobutadiene


Concen: 51.348 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53





Tgt Ion:225 Resp: 63906


Ion Ratio Lower Upper

225 100

223 62.8 30.7 92.1

227 62.3 30.9 92.5





#95

Naphthalene

Concen: 52.473 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:128 Resp: 367063

Ion Ratio Lower Upper

128 100

127 12.9 10.6 16.0





129 10.5 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#96

1,2,3-Trichlorobenzene

Concen: 51.230 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion:180 Resp: 121516

Ion Ratio Lower Upper

180 100

182 96.0 47.4 142.2

145 34.5 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085526.D
 Acq On : 28 Jan 2025 09:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.677	0.620	8.4	101	0.00
3 P	Chloromethane	0.733	0.642	12.4	96	0.00
4 C	Vinyl Chloride	0.737	0.665	9.8#	99	0.00
5 T	Bromomethane	0.445	0.389	12.6	95	0.00
6 T	Chloroethane	0.467	0.439	6.0	106	0.00
7 T	Trichlorofluoromethane	1.069	1.071	-0.2	110	0.00
8 T	Diethyl Ether	0.369	0.361	2.2	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.618	-2.7	116	0.00
10 T	Methyl Iodide	0.690	0.708	-2.6	100	0.00
11 T	Tert butyl alcohol	0.092	0.108	-17.4	119	0.00
12 CM	1,1-Dichloroethene	0.537	0.526	2.0#	101	0.00
13 T	Acrolein	0.126	0.132	-4.8	97	0.00
14 T	Allyl chloride	0.871	0.848	2.6	102	0.00
15 T	Acrylonitrile	0.294	0.306	-4.1	105	0.00
16 T	Acetone	0.261	0.263	-0.8	106	0.00
17 T	Carbon Disulfide	1.652	1.391	15.8	96	0.00
18 T	Methyl Acetate	0.793	0.859	-8.3	111	0.00
19 T	Methyl tert-butyl Ether	1.742	1.940	-11.4	106	0.00
20 T	Methylene Chloride	0.646	0.645	0.2	105	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.570	0.5	105	0.00
22 T	Diisopropyl ether	1.933	2.096	-8.4	105	0.00
23 T	Vinyl Acetate	1.353	1.493	-10.3	102	0.00
24 P	1,1-Dichloroethane	1.178	1.202	-2.0	105	0.00
25 T	2-Butanone	0.384	0.399	-3.9	105	0.00
26 T	2,2-Dichloropropane	0.953	1.098	-15.2	120	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.700	-3.7	105	0.00
28 T	Bromochloromethane	0.548	0.499	8.9	94	0.00
29 T	Tetrahydrofuran	0.243	0.267	-9.9	105	0.00
30 C	Chloroform	1.218	1.248	-2.5#	109	0.00
31 T	Cyclohexane	1.024	0.931	9.1	108	0.00
32 T	1,1,1-Trichloroethane	1.068	1.105	-3.5	111	0.00
33 S	1,2-Dichloroethane-d4	0.807	0.777	3.7	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.347	0.355	-2.3	98	0.00
36 T	1,1-Dichloropropene	0.487	0.512	-5.1	109	0.00
37 T	Ethyl Acetate	0.491	0.520	-5.9	105	0.00
38 T	Carbon Tetrachloride	0.557	0.602	-8.1	112	0.00
39 T	Methylcyclohexane	0.457	0.513	-12.3	109	0.00
40 TM	Benzene	1.463	1.575	-7.7	107	0.00
41 T	Methacrylonitrile	0.256	0.285	-11.3	104	0.00
42 TM	1,2-Dichloroethane	0.551	0.585	-6.2	105	0.00
43 T	Isopropyl Acetate	0.787	0.862	-9.5	104	0.00
44 TM	Trichloroethene	0.341	0.359	-5.3	109	0.00
45 C	1,2-Dichloropropane	0.374	0.407	-8.8#	108	0.00
46 T	Dibromomethane	0.270	0.287	-6.3	106	0.00
47 T	Bromodichloromethane	0.549	0.613	-11.7	108	0.00
48 T	Methyl methacrylate	0.354	0.409	-15.5	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085526.D
 Acq On : 28 Jan 2025 09:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	94	0.00
50 S	Toluene-d8	1.232	1.296	-5.2	100	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.535	-17.1	107	0.00
52 CM	Toluene	0.848	0.975	-15.0#	110	0.00
53 T	t-1,3-Dichloropropene	0.519	0.611	-17.7	109	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.639	-15.3	107	0.00
55 T	1,1,2-Trichloroethane	0.335	0.375	-11.9	108	0.00
56 T	Ethyl methacrylate	0.469	0.575	-22.6	104	0.00
57 T	1,3-Dichloropropane	0.583	0.657	-12.7	107	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.256	-20.2	104	0.00
59 T	2-Hexanone	0.321	0.386	-20.2	106	0.00
60 T	Dibromochloromethane	0.405	0.460	-13.6	109	0.00
61 T	1,2-Dibromoethane	0.334	0.366	-9.6	108	0.00
62 S	4-Bromofluorobenzene	0.422	0.467	-10.7	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.341	0.361	-5.9	114	0.00
65 PM	Chlorobenzene	1.095	1.171	-6.9	111	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.430	-7.0	112	0.00
67 C	Ethyl Benzene	1.784	2.053	-15.1#	112	0.00
68 T	m/p-Xylenes	0.659	0.785	-19.1	113	0.00
69 T	o-Xylene	0.630	0.739	-17.3	110	0.00
70 T	Styrene	1.043	1.279	-22.6	111	0.00
71 P	Bromoform	0.288	0.334	-16.0	109	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
73 T	Isopropylbenzene	3.375	3.736	-10.7	115	0.00
74 T	N-amyl acetate	1.517	1.605	-5.8	105	0.00
75 P	1,1,2,2-Tetrachloroethane	1.192	1.170	1.8	109	0.00
76 T	1,2,3-Trichloropropane	1.016	0.945	7.0	112	0.00
77 T	Bromobenzene	0.882	0.891	-1.0	110	0.00
78 T	n-propylbenzene	3.995	4.503	-12.7	115	0.00
79 T	2-Chlorotoluene	2.586	2.728	-5.5	112	0.00
80 T	1,3,5-Trimethylbenzene	2.787	3.177	-14.0	115	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.408	-8.5	111	0.00
82 T	4-Chlorotoluene	2.574	2.778	-7.9	114	0.00
83 T	tert-Butylbenzene	2.341	2.605	-11.3	113	0.00
84 T	1,2,4-Trimethylbenzene	2.778	3.219	-15.9	113	0.00
85 T	sec-Butylbenzene	3.244	3.731	-15.0	116	0.00
86 T	p-Isopropyltoluene	2.631	3.155	-19.9	117	0.00
87 T	1,3-Dichlorobenzene	1.624	1.699	-4.6	115	0.00
88 T	1,4-Dichlorobenzene	1.683	1.700	-1.0	116	0.00
89 T	n-Butylbenzene	2.327	2.700	-16.0	115	0.00
90 T	Hexachloroethane	0.618	0.625	-1.1	115	0.00
91 T	1,2-Dichlorobenzene	1.620	1.644	-1.5	112	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.225	-3.2	108	0.00
93 T	1,2,4-Trichlorobenzene	0.753	0.817	-8.5	111	0.00
94 T	Hexachlorobutadiene	0.400	0.410	-2.5	119	0.00
95 T	Naphthalene	2.246	2.357	-4.9	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085526.D
Acq On : 28 Jan 2025 09:53
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.780	-2.4	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085526.D
 Acq On : 28 Jan 2025 09:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	50.000	45.800	8.4	101	0.00
3 P	Chloromethane	50.000	43.823	12.4	96	0.00
4 C	Vinyl Chloride	50.000	45.155	9.7#	99	0.00
5 T	Bromomethane	50.000	43.757	12.5	95	0.00
6 T	Chloroethane	50.000	46.993	6.0	106	0.00
7 T	Trichlorofluoromethane	50.000	50.109	-0.2	110	0.00
8 T	Diethyl Ether	50.000	48.819	2.4	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.335	-2.7	116	0.00
10 T	Methyl Iodide	50.000	51.306	-2.6	100	0.00
11 T	Tert butyl alcohol	250.000	293.091	-17.2	119	0.00
12 CM	1,1-Dichloroethene	50.000	49.022	2.0#	101	0.00
13 T	Acrolein	250.000	261.783	-4.7	97	0.00
14 T	Allyl chloride	50.000	48.683	2.6	102	0.00
15 T	Acrylonitrile	250.000	260.994	-4.4	105	0.00
16 T	Acetone	250.000	251.929	-0.8	106	0.00
17 T	Carbon Disulfide	50.000	42.095	15.8	96	0.00
18 T	Methyl Acetate	50.000	54.151	-8.3	111	0.00
19 T	Methyl tert-butyl Ether	50.000	55.663	-11.3	106	0.00
20 T	Methylene Chloride	50.000	49.931	0.1	105	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.692	0.6	105	0.00
22 T	Diisopropyl ether	50.000	54.237	-8.5	105	0.00
23 T	Vinyl Acetate	250.000	276.027	-10.4	102	0.00
24 P	1,1-Dichloroethane	50.000	51.014	-2.0	105	0.00
25 T	2-Butanone	250.000	260.223	-4.1	105	0.00
26 T	2,2-Dichloropropane	50.000	57.640	-15.3	120	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.799	-3.6	105	0.00
28 T	Bromochloromethane	50.000	45.505	9.0	94	0.00
29 T	Tetrahydrofuran	250.000	274.712	-9.9	105	0.00
30 C	Chloroform	50.000	51.246	-2.5#	109	0.00
31 T	Cyclohexane	50.000	45.451	9.1	108	0.00
32 T	1,1,1-Trichloroethane	50.000	51.727	-3.5	111	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.164	3.7	96	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	51.240	-2.5	98	0.00
36 T	1,1-Dichloropropene	50.000	52.537	-5.1	109	0.00
37 T	Ethyl Acetate	50.000	52.923	-5.8	105	0.00
38 T	Carbon Tetrachloride	50.000	54.013	-8.0	112	0.00
39 T	Methylcyclohexane	50.000	56.039	-12.1	109	0.00
40 TM	Benzene	50.000	53.828	-7.7	107	0.00
41 T	Methacrylonitrile	50.000	55.717	-11.4	104	0.00
42 TM	1,2-Dichloroethane	50.000	53.072	-6.1	105	0.00
43 T	Isopropyl Acetate	50.000	54.755	-9.5	104	0.00
44 TM	Trichloroethene	50.000	52.672	-5.3	109	0.00
45 C	1,2-Dichloropropane	50.000	54.447	-8.9#	108	0.00
46 T	Dibromomethane	50.000	53.179	-6.4	106	0.00
47 T	Bromodichloromethane	50.000	55.774	-11.5	108	0.00
48 T	Methyl methacrylate	50.000	57.693	-15.4	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085526.D
 Acq On : 28 Jan 2025 09:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1019.345	-1.9	94	0.00
50 S	Toluene-d8	50.000	52.569	-5.1	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	292.599	-17.0	107	0.00
52 CM	Toluene	50.000	57.499	-15.0#	110	0.00
53 T	t-1,3-Dichloropropene	50.000	58.899	-17.8	109	0.00
54 T	cis-1,3-Dichloropropene	50.000	57.610	-15.2	107	0.00
55 T	1,1,2-Trichloroethane	50.000	55.851	-11.7	108	0.00
56 T	Ethyl methacrylate	50.000	50.844	-1.7	104	0.00
57 T	1,3-Dichloropropane	50.000	56.327	-12.7	107	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	300.465	-20.2	104	0.00
59 T	2-Hexanone	250.000	300.034	-20.0	106	0.00
60 T	Dibromochloromethane	50.000	56.815	-13.6	109	0.00
61 T	1,2-Dibromoethane	50.000	54.776	-9.6	108	0.00
62 S	4-Bromofluorobenzene	50.000	55.379	-10.8	102	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	102	0.00
64 T	Tetrachloroethene	50.000	52.974	-5.9	114	0.00
65 PM	Chlorobenzene	50.000	53.457	-6.9	111	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.447	-6.9	112	0.00
67 C	Ethyl Benzene	50.000	57.531	-15.1#	112	0.00
68 T	m/p-Xylenes	100.000	119.124	-19.1	113	0.00
69 T	o-Xylene	50.000	58.639	-17.3	110	0.00
70 T	Styrene	50.000	61.343	-22.7	111	0.00
71 P	Bromoform	50.000	58.046	-16.1	109	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	106	0.00
73 T	Isopropylbenzene	50.000	55.357	-10.7	115	0.00
74 T	N-amyl acetate	50.000	52.919	-5.8	105	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.089	1.8	109	0.00
76 T	1,2,3-Trichloropropane	50.000	46.478	7.0	112	0.00
77 T	Bromobenzene	50.000	50.547	-1.1	110	0.00
78 T	n-propylbenzene	50.000	56.362	-12.7	115	0.00
79 T	2-Chlorotoluene	50.000	52.764	-5.5	112	0.00
80 T	1,3,5-Trimethylbenzene	50.000	56.984	-14.0	115	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	54.370	-8.7	111	0.00
82 T	4-Chlorotoluene	50.000	53.960	-7.9	114	0.00
83 T	tert-Butylbenzene	50.000	55.641	-11.3	113	0.00
84 T	1,2,4-Trimethylbenzene	50.000	57.941	-15.9	113	0.00
85 T	sec-Butylbenzene	50.000	57.499	-15.0	116	0.00
86 T	p-Isopropyltoluene	50.000	53.532	-7.1	117	0.00
87 T	1,3-Dichlorobenzene	50.000	52.335	-4.7	115	0.00
88 T	1,4-Dichlorobenzene	50.000	50.507	-1.0	116	0.00
89 T	n-Butylbenzene	50.000	58.013	-16.0	115	0.00
90 T	Hexachloroethane	50.000	50.630	-1.3	115	0.00
91 T	1,2-Dichlorobenzene	50.000	50.739	-1.5	112	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.620	-3.2	108	0.00
93 T	1,2,4-Trichlorobenzene	50.000	54.246	-8.5	111	0.00
94 T	Hexachlorobutadiene	50.000	51.348	-2.7	119	0.00
95 T	Naphthalene	50.000	52.473	-4.9	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085526.D
Acq On : 28 Jan 2025 09:53
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.230	-2.5	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199

Instrument ID: MSVOA_N Calibration Date/Time: 01/28/2025 18:14

Lab File ID: VN085546.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.733	0.606	0.1	-17.33	50
Vinyl Chloride	0.737	0.641		-13.03	50
Bromomethane	0.445	0.362		-18.65	50
Chloroethane	0.467	0.426		-8.78	50
Trichlorofluoromethane	1.069	1.034		-3.27	50
1,1,2-Trichlorotrifluoroethane	0.602	0.590		-1.99	50
1,1-Dichloroethene	0.537	0.523		-2.61	50
Acetone	0.261	0.285		9.19	50
Carbon Disulfide	1.652	1.307		-20.88	50
Methyl tert-butyl Ether	1.742	1.909		9.59	50
Methylene Chloride	0.646	0.637		-1.39	50
trans-1,2-Dichloroethene	0.573	0.549		-4.19	50
1,1-Dichloroethane	1.178	1.186	0.1	0.68	50
2-Butanone	0.384	0.451		17.45	50
Carbon Tetrachloride	0.557	0.573		2.87	50
cis-1,2-Dichloroethene	0.675	0.693		2.67	50
Chloroform	1.218	1.238		1.64	50
1,1,1-Trichloroethane	1.068	1.078		0.94	50
Methylcyclohexane	0.457	0.458		0.22	50
Benzene	1.463	1.498		2.39	50
1,2-Dichloroethane	0.551	0.571		3.63	50
Trichloroethene	0.341	0.341		0	50
1,2-Dichloropropane	0.374	0.391		4.55	50
Bromodichloromethane	0.549	0.602		9.65	50
4-Methyl-2-Pentanone	0.457	0.577		26.26	50
Toluene	0.848	0.923		8.84	50
t-1,3-Dichloropropene	0.519	0.580		11.75	50
cis-1,3-Dichloropropene	0.554	0.602		8.66	50
1,1,2-Trichloroethane	0.335	0.367		9.55	50
2-Hexanone	0.321	0.420		30.84	50
Dibromochloromethane	0.405	0.444		9.63	50
Tetrachloroethene	0.341	0.309		-9.38	50
Chlorobenzene	1.095	1.102	0.3	0.64	50
Ethyl Benzene	1.784	1.936		8.52	50
m/p-Xylenes	0.659	0.738		11.99	50
o-Xylene	0.630	0.697		10.64	50
Styrene	1.043	1.210		16.01	50
Bromoform	0.288	0.322	0.1	11.81	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG No.: Q1199

Instrument ID: MSVOA_N Calibration Date/Time: 01/28/2025 18:14

Lab File ID: VN085546.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.375	3.613		7.05	50
1,1,2,2-Tetrachloroethane	1.192	1.228	0.3	3.02	50
1,3-Dichlorobenzene	1.624	1.626		0.12	50
1,4-Dichlorobenzene	1.683	1.600		-4.93	50
1,2-Dichlorobenzene	1.620	1.592		-1.73	50
1,2-Dichloroethane-d4	0.807	0.882		9.29	50
Dibromofluoromethane	0.347	0.384		10.66	50
Toluene-d8	1.232	1.397		13.39	50
4-Bromofluorobenzene	0.422	0.503		19.19	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085546.D
 Acq On : 28 Jan 2025 18:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/29/2025
 Supervised By : Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 03:39:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	187799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	314055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143565	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.577	65	165633	54.638	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.280%
35) Dibromofluoromethane	8.165	113	120592	55.349	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.700%
50) Toluene-d8	10.565	98	438659	56.666	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	113.340%#
62) 4-Bromofluorobenzene	12.847	95	157860	59.614	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	119.220%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	109169	42.934	ug/l	94
3) Chloromethane	2.359	50	113732	41.310	ug/l	100
4) Vinyl Chloride	2.512	62	120445	43.525	ug/l	98
5) Bromomethane	2.936	94	67919	40.633	ug/l	97
6) Chloroethane	3.106	64	80051	45.629	ug/l	97
7) Trichlorofluoromethane	3.495	101	194108	48.343	ug/l	98
8) Diethyl Ether	3.959	74	65944	47.541	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	110829	49.006	ug/l	99
10) Methyl Iodide	4.589	142	128917	49.778	ug/l	98
11) Tert butyl alcohol	5.518	59	122061	351.637	ug/l	99
12) 1,1-Dichloroethene	4.336	96	98130	48.687	ug/l	97
13) Acrolein	4.177	56	129555	273.407	ug/l	100
14) Allyl chloride	5.024	41	155852	47.653	ug/l	98
15) Acrylonitrile	5.712	53	310806	281.891	ug/l	98
16) Acetone	4.424	43	267295	272.891	ug/l	99
17) Carbon Disulfide	4.712	76	245541	39.567	ug/l	99
18) Methyl Acetate	5.018	43	178998	60.097	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	358554	54.787	ug/l	99
20) Methylene Chloride	5.271	84	119628	49.335	ug/l	97
21) trans-1,2-Dichloroethene	5.783	96	103142	47.884	ug/l	97
22) Diisopropyl ether	6.671	45	392245	54.034	ug/l	99
23) Vinyl Acetate	6.600	43	1454703	286.348	ug/l	98
24) 1,1-Dichloroethane	6.565	63	222797	50.334	ug/l	98
25) 2-Butanone	7.482	43	423099	293.531	ug/l	99
26) 2,2-Dichloropropane	7.488	77	189076	52.834	ug/l	100
27) cis-1,2-Dichloroethene	7.482	96	130214	51.325	ug/l	99
28) Bromochloromethane	7.812	49	89552	43.487	ug/l	98
29) Tetrahydrofuran	7.835	42	275635	301.616	ug/l	100
30) Chloroform	7.965	83	232517	50.826	ug/l	99
31) Cyclohexane	8.253	56	166703	43.331	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	202362	50.429	ug/l	98
36) 1,1-Dichloropropene	8.371	75	152437	49.851	ug/l	99
37) Ethyl Acetate	7.559	43	171657	55.616	ug/l	98
38) Carbon Tetrachloride	8.359	117	179826	51.368	ug/l	98
39) Methylcyclohexane	9.600	83	143759	50.030	ug/l	98
40) Benzene	8.606	78	470586	51.218	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085546.D
 Acq On : 28 Jan 2025 18:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/29/2025
 Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 03:39:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	97709	60.848	ug/l	97
42) 1,2-Dichloroethane	8.671	62	179469	51.863	ug/l	100
43) Isopropyl Acetate	8.688	43	345380	69.836	ug/l #	91
44) Trichloroethene	9.347	130	106986	50.023	ug/l	97
45) 1,2-Dichloropropane	9.618	63	122928	52.380	ug/l	98
46) Dibromomethane	9.706	93	87054	51.408	ug/l	98
47) Bromodichloromethane	9.888	83	188980	54.778	ug/l	97
48) Methyl methacrylate	9.676	41	134702	60.523	ug/l	98
49) 1,4-Dioxane	9.694	88	46510	1240.952	ug/l #	92
51) 4-Methyl-2-Pentanone	10.441	43	906426	315.843	ug/l	100
52) Toluene	10.629	92	289726	54.423	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	182074	55.846	ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	188960	54.260	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	115272	54.714	ug/l	98
56) Ethyl methacrylate	10.871	69	183142	51.541	ug/l	98
57) 1,3-Dichloropropane	11.159	76	203386	55.519	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	389176	291.087	ug/l	100
59) 2-Hexanone	11.194	43	659035	326.364	ug/l	99
60) Dibromochloromethane	11.359	129	139526	54.855	ug/l	98
61) 1,2-Dibromoethane	11.470	107	112753	53.765	ug/l	99
64) Tetrachloroethene	11.100	164	88612	45.330	ug/l	99
65) Chlorobenzene	11.888	112	315889	50.289	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	117993	51.180	ug/l	98
67) Ethyl Benzene	11.965	91	555215	54.268	ug/l	99
68) m/p-Xylenes	12.070	106	423444	111.987	ug/l	99
69) o-Xylene	12.394	106	199877	55.313	ug/l	99
70) Styrene	12.412	104	347052	58.032	ug/l	100
71) Bromoform	12.576	173	92343	55.980	ug/l #	99
73) Isopropylbenzene	12.694	105	518634	53.526	ug/l	99
74) N-amyl acetate	12.494	43	244112	56.051	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	176292	51.507	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	145886m	49.995	ug/l	
77) Bromobenzene	12.976	156	127032	50.181	ug/l	100
78) n-propylbenzene	13.035	91	616293	53.728	ug/l	99
79) 2-Chlorotoluene	13.123	91	384370	51.776	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	432662	54.059	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	61346	56.880	ug/l	90
82) 4-Chlorotoluene	13.217	91	383756	51.917	ug/l	99
83) tert-Butylbenzene	13.435	119	354690	52.772	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	443501	55.598	ug/l	99
85) sec-Butylbenzene	13.617	105	504931	54.205	ug/l	100
86) p-Isopropyltoluene	13.729	119	422476	50.325	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	233419	50.068	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	229722	47.539	ug/l	100
89) n-Butylbenzene	14.053	91	351230	52.557	ug/l	99
90) Hexachloroethane	14.329	117	86748	48.917	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	228524	49.142	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	34650	55.439	ug/l	99
93) 1,2,4-Trichlorobenzene	15.388	180	108928	50.399	ug/l	99
94) Hexachlorobutadiene	15.500	225	49834	43.430	ug/l	95
95) Naphthalene	15.641	128	342729	53.141	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	107618	49.210	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085546.D
Acq On : 28 Jan 2025 18:14
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 03:39:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\  
Data File : VN085546.D  
Acq On    : 28 Jan 2025   18:14  
Operator  : JC\MD  
Sample    : VSTDCCC050  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 22   Sample Multiplier: 1
```


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025


Quant Time: Jan 29 03:39:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 168 Resp: 187799
Ion Ratio Lower Upper
168 100
99 62.0 53.6 80.4




Manual Integrations
APPROVED

Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025




#2
Dichlorodifluoromethane
Concen: 42.934 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 85 Resp: 109169
Ion Ratio Lower Upper
85 100
87 29.7 16.7 50.1



Abundance Scan 69 (2.359 min): VN085439.D\data.ms (-62)



#3

Chloromethane

Concen: 41.310 ug/l

RT: 2.359 min Scan# 69

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 50 Resp: 113732

Ion Ratio Lower Upper

50 100

52 33.6 27.0 40.6


Manual Integrations

APPROVED



Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025


Abundance Scan 69 (2.359 min): VN085546.D\data.ms



Abundance Scan 69 (2.359 min): VN085546.D\data.ms (-18)



Abundance Scan 95 (2.512 min): VN085439.D\data.ms (-87)



#4

Vinyl Chloride

Concen: 43.525 ug/l

RT: 2.512 min Scan# 95

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14


Tgt Ion: 62 Resp: 120445

Ion Ratio Lower Upper



62 100


64 32.0 24.8 37.2

Abundance Scan 95 (2.512 min): VN085546.D\data.ms




Abundance Scan 95 (2.512 min): VN085546.D\data.ms (-44)





#5
Bromomethane
Concen: 40.633 ug/l
RT: 2.936 min Scan# 16
Delta R.T. -0.018 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 67919
Ion Ratio Lower Upper
94 100
96 92.4 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#6
Chloroethane
Concen: 45.629 ug/l
RT: 3.106 min Scan# 196
Delta R.T. -0.012 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14


Tgt Ion: 64 Resp: 80051
Ion Ratio Lower Upper
64 100
66 32.1 24.6 36.8





#7
Trichlorofluoromethane
Concen: 48.343 ug/l
RT: 3.495 min Scan# 263
Delta R.T. -0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 101 Resp: 194108
Ion Ratio Lower Upper
101 100
103 66.0 51.4 77.2

Manual Integrations
APPROVED



Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025




#8
Diethyl Ether
Concen: 47.541 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14



Tgt Ion: 74 Resp: 65944
Ion Ratio Lower Upper
74 100
45 98.3 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.006 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion	101	Resp	110829
Ion	Ratio	Lower	Upper
101	100		
85	46.6	37.8	56.8
151	72.6	58.8	88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#10
Methyl Iodide
Concen: 49.778 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion	142	Resp	128917
Ion	Ratio	Lower	Upper
142	100		
127	48.3	40.3	60.5
141	16.0	12.8	19.2





#11

Tert butyl alcohol

Concen: 351.637 ug/l

RT: 5.518 min Scan# 605

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 122061

Ion Ratio Lower Upper

59 100





57 11.2 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#12

1,1-Dichloroethene

Concen: 48.687 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 96 Resp: 98130


Ion Ratio Lower Upper

96 100

61 180.0 141.0 211.4

98 63.4 49.0 73.6





#13

Acrolein

Concen: 273.407 ug/l

RT: 4.177 min Scan# 37

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 129555

Ion Ratio Lower Upper

56 100





55 70.6 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#14

Allyl chloride

Concen: 47.653 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 41 Resp: 155852


Ion Ratio Lower Upper

41 100

39 82.2 64.4 96.6

76 37.0 30.5 45.7





#15

Acrylonitrile

Concen: 281.891 ug/l

RT: 5.712 min Scan# 639

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 53 Resp: 310806

Ion Ratio Lower Upper

53 100

52 83.9 65.5 98.3





51 37.3 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#16

Acetone

Concen: 272.891 ug/l

RT: 4.424 min Scan# 420

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 43 Resp: 267295

Ion Ratio Lower Upper

43 100

58 30.0 23.8 35.6





#17

Carbon Disulfide

Concen: 39.567 ug/l

RT: 4.712 min Scan# 468

Delta R.T. 0.006 min

Lab File: VN085546.D


Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 76 Resp: 245541

Ion Ratio Lower Upper

76 100




78 8.9 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#18

Methyl Acetate

Concen: 60.097 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.000 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 43 Resp: 178998

Ion Ratio Lower Upper

43 100

74 22.9 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 54.787 ug/l

RT: 5.794 min Scan# 65

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 73 Resp: 358554

Ion Ratio Lower Upper

73 100





57 21.7 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#20

Methylene Chloride

Concen: 49.335 ug/l

RT: 5.271 min Scan# 564

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Tgt Ion: 84 Resp: 119628




Ion Ratio Lower Upper


84 100

49 133.6 104.1 156.1

51 36.4 31.0 46.4

86 62.6 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 47.884 ug/l

RT: 5.783 min Scan# 651

Delta R.T. -0.000 min

Lab File: VN085546.D


Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 96 Resp: 103142

Ion Ratio Lower Upper

96 100

61 146.0 115.7 173.5




98 66.8 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#22

Diisopropyl ether


Concen: 54.034 ug/l

RT: 6.671 min Scan# 802

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14



Tgt Ion: 45 Resp: 392245



Ion Ratio Lower Upper


45 100

43 54.9 45.0 67.6

87 26.8 21.8 32.8

59 12.6 10.2 15.2





#23

Vinyl Acetate

Concen: 286.348 ug/l

RT: 6.600 min Scan# 79

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 1454703

Ion Ratio Lower Upper

43 100





86 9.6 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#24

1,1-Dichloroethane

Concen: 50.334 ug/l

RT: 6.565 min Scan# 784

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 63 Resp: 222797


Ion Ratio Lower Upper

63 100

98 5.9 3.3 9.8

100 4.4 2.0 6.0





#25

2-Butanone

Concen: 293.531 ug/l

RT: 7.482 min Scan# 940

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Tgt Ion: 43 Resp: 423099

Ion Ratio Lower Upper

43 100

72 24.7 20.2 30.4

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#26

2,2-Dichloropropane

Concen: 52.834 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 77 Resp: 189076

Ion Ratio Lower Upper

77 100


97 21.4 10.7 32.1





#27
 cis-1,2-Dichloroethene
 Concen: 51.325 ug/l
 RT: 7.482 min Scan# 940
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 96 Resp: 130214
 Ion Ratio Lower Upper
 96 100
 61 155.9 0.0 311.8
 98 64.7 0.0 126.0




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/29/2025
 Supervised By :Semsettin Yesilyurt 01/29/2025



#28
 Bromochloromethane
 Concen: 43.487 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion: 49 Resp: 89552
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.4
 130 70.3 55.0 82.4









#29
Tetrahydrofuran
Concen: 301.616 ug/l
RT: 7.835 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 42 Resp: 275635
Ion Ratio Lower Upper
42 100
72 42.8 34.2 51.2
71 40.1 32.5 48.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#30
Chloroform
Concen: 50.826 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 83 Resp: 232517
Ion Ratio Lower Upper
83 100
85 65.3 51.8 77.6





#31

Cyclohexane

Concen: 43.331 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 166703

Ion Ratio Lower Upper

56 100

69 29.8 22.2 33.4





84 80.6 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#32

1,1,1-Trichloroethane

Concen: 50.429 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 97 Resp: 202362


Ion Ratio Lower Upper

97 100

99 65.0 49.8 74.6


61 51.3 41.4 62.2





#33
1,2-Dichloroethane-d4
Concen: 54.638 ug/l
RT: 8.577 min Scan# 1125
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 65 Resp: 165633
Ion Ratio Lower Upper
65 100
67 50.7 0.0 101.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:114 Resp: 314055
Ion Ratio Lower Upper
114 100
63 22.0 0.0 47.6
88 15.5 0.0 32.6





#35

Dibromofluoromethane

Concen: 55.349 ug/l

RT: 8.165 min Scan# 1055

Delta R.T. 0.006 min

Lab File: VN085546.D


Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 113 Resp: 120592

Ion Ratio Lower Upper

113 100

111 102.9 82.7 124.1




192 17.8 14.3 21.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#36

1,1-Dichloropropene


Concen: 49.851 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14





Tgt Ion: 75 Resp: 152437


Ion Ratio Lower Upper

75 100

110 32.6 16.5 49.5

77 29.9 24.4 36.6





#37

Ethyl Acetate

Concen: 55.616 ug/l

RT: 7.559 min Scan# 952

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 171657

Ion Ratio Lower Upper

43 100

61 12.5 10.9 16.3





70 9.5 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#38

Carbon Tetrachloride

Concen: 51.368 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion:117 Resp: 179826


Ion Ratio Lower Upper

117 100

119 95.0 75.4 113.2

121 28.1 24.6 37.0





#39

Methylcyclohexane

Concen: 50.030 ug/l

RT: 9.600 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 143759

Ion Ratio Lower Upper

83 100

55 77.8 62.6 94.0





98 50.5 37.7 56.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#40

Benzene

Concen: 51.218 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 78 Resp: 470586

Ion Ratio Lower Upper

78 100

77 24.3 19.0 28.6





#41

Methacrylonitrile

Concen: 60.848 ug/l

RT: 7.777 min Scan# 99

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 41 Resp: 97709

Ion Ratio Lower Upper

41 100

39 56.4 46.0 69.0

67 68.0 57.4 86.2





52 31.0 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#42

1,2-Dichloroethane

Concen: 51.863 ug/l

RT: 8.671 min Scan# 1142

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 62 Resp: 179469

Ion Ratio Lower Upper

62 100

98 8.6 0.0 17.0





#43

Isopropyl Acetate

Concen: 69.836 ug/l

RT: 8.688 min Scan# 1144

Delta R.T. 0.006 min

Lab File: VN085546.D


Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 43 Resp: 345386

Ion Ratio Lower Upper

43 100

61 20.6 20.7 31.1




87 10.2 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#44

Trichloroethene

Concen: 50.023 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion:130 Resp: 106986

Ion Ratio Lower Upper

130 100

95 101.0 0.0 195.8





#45

1,2-Dichloropropane

Concen: 52.380 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 122928

Ion Ratio Lower Upper

63 100





65 31.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#46

Dibromomethane

Concen: 51.408 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 93 Resp: 87054


Ion Ratio Lower Upper

93 100

95 83.2 64.7 97.1

174 88.2 69.0 103.4





#47

Bromodichloromethane

Concen: 54.778 ug/l

RT: 9.888 min Scan# 1348

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 188980

Ion Ratio Lower Upper

83 100

85 61.4 51.2 76.8





127 8.1 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#48

Methyl methacrylate

Concen: 60.523 ug/l

RT: 9.676 min Scan# 1313

Delta R.T. -0.001 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 41 Resp: 134702


Ion Ratio Lower Upper

41 100

69 80.5 64.7 97.1

39 64.5 49.0 73.6





#49

1,4-Dioxane

Concen: 1240.952 ug/l

RT: 9.694 min Scan# 1315

Delta R.T. 0.006 min

Lab File: VN085546.D


Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 88 Resp: 46510

Ion Ratio Lower Upper

88 100

43 23.4 26.6 39.8




58 71.6 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#50

Toluene-d8

Concen: 56.666 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 98 Resp: 438659

Ion Ratio Lower Upper

98 100

100 65.0 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 315.843 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 906426
Ion Ratio Lower Upper
43 100
58 38.0 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#52
Toluene
Concen: 54.423 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14


Tgt Ion: 92 Resp: 289726
Ion Ratio Lower Upper
92 100
91 174.4 139.2 208.8





#53
t-1,3-Dichloropropene
Concen: 55.846 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 182074
Ion Ratio Lower Upper
75 100
77 31.2 23.5 35.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#54
cis-1,3-Dichloropropene
Concen: 54.260 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 75 Resp: 188960
Ion Ratio Lower Upper
75 100
77 33.1 25.0 37.4
39 55.1 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 54.714 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 115272

Ion Ratio Lower Upper

97 100

83 85.0 69.0 103.6

85 53.3 44.6 66.8





99 60.7 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#56

Ethyl methacrylate

Concen: 51.541 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 69 Resp: 183142


Ion Ratio Lower Upper

69 100

41 69.8 54.6 82.0


39 41.8 32.4 48.6





#57
1,3-Dichloropropane
Concen: 55.519 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 203386
Ion Ratio Lower Upper
76 100
78 31.9 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#58
2-Chloroethyl Vinyl ether
Concen: 291.087 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 63 Resp: 389176
Ion Ratio Lower Upper
63 100
106 27.0 21.6 32.4





#59

2-Hexanone

Concen: 326.364 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 659035

Ion Ratio Lower Upper

43 100





58 52.0 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#60

Dibromochloromethane

Concen: 54.855 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion:129 Resp: 139526

Ion Ratio Lower Upper

129 100

127 78.8 38.6 115.8









#61
1,2-Dibromoethane
Concen: 53.765 ug/l
RT: 11.470 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 107 Resp: 112753
Ion Ratio Lower Upper
107 100
109 94.2 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025



#62
4-Bromofluorobenzene
Concen: 59.614 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 95 Resp: 157860
Ion Ratio Lower Upper
95 100
174 73.2 0.0 145.0
176 72.3 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:117 Resp: 286739

Ion Ratio Lower Upper

117 100

82 59.0 48.6 72.8





119 32.6 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#64

Tetrachloroethene

Concen: 45.330 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Tgt Ion:164 Resp: 88612




Ion Ratio Lower Upper


164 100

166 127.2 103.4 155.2

129 99.0 79.2 118.8

131 95.8 77.1 115.7









#65
Chlorobenzene
Concen: 50.289 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:112 Resp: 315889
Ion Ratio Lower Upper
112 100
114 32.3 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 51.180 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:131 Resp: 117993
Ion Ratio Lower Upper
131 100
133 96.4 47.4 142.3
119 65.0 33.1 99.5





#67

Ethyl Benzene

Concen: 54.268 ug/l

RT: 11.965 min Scan# 17

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 555215

Ion Ratio Lower Upper

91 100





106 30.5 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#68

m/p-Xylenes

Concen: 111.987 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion:106 Resp: 423444

Ion Ratio Lower Upper

106 100

91 207.7 167.7 251.5









#69
o-Xylene
Concen: 55.313 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 199877
Ion Ratio Lower Upper
106 100
91 222.8 110.4 331.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#70
Styrene
Concen: 58.032 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:104 Resp: 347052
Ion Ratio Lower Upper
104 100
78 53.0 42.5 63.7
103 55.1 43.8 65.8





#71

Bromoform

Concen: 55.980 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:173 Resp: 92343

Ion Ratio Lower Upper

173 100

175 49.3 24.4 73.2





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion:152 Resp: 143565


Ion Ratio Lower Upper

152 100

115 62.2 31.1 93.3

150 173.3 0.0 343.6





#73

Isopropylbenzene

Concen: 53.526 ug/l

RT: 12.694 min Scan# 18

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Tgt Ion: 105 Resp: 518634

Ion Ratio Lower Upper

105 100

120 25.8 12.8 38.3

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#74

N-ethyl acetate

Concen: 56.051 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Tgt Ion: 43 Resp: 244112




Ion Ratio Lower Upper


43 100

70 42.5 34.0 51.0

55 26.1 21.4 32.2

61 23.3 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 51.507 ug/l

RT: 12.935 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:	83	Resp:	176292
Ion	Ratio	Lower	Upper
83	100		
131	9.4	4.8	14.4
85	64.2	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#76

1,2,3-Trichloropropane

Concen: 49.995 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Tgt Ion:	75	Resp:	145886
Ion	Ratio	Lower	Upper
75	100		
77	199.3	109.7	329.2





#77

Bromobenzene

Concen: 50.181 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.000 min

Lab File: VN085546.D


Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 156 Resp: 127032

Ion Ratio Lower Upper

156 100

77 228.8 114.1 342.4




158 97.3 48.9 146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#78

n-propylbenzene

Concen: 53.728 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 91 Resp: 616293

Ion Ratio Lower Upper

91 100

120 22.1 10.9 32.6





#79

2-Chlorotoluene

Concen: 51.776 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 384376

Ion Ratio Lower Upper

91 100





126 32.3 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#80

1,3,5-Trimethylbenzene

Concen: 54.059 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion:105 Resp: 432662

Ion Ratio Lower Upper

105 100

120 48.3 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 56.880 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 75 Resp: 61346
Ion Ratio Lower Upper
75 100
53 105.0 95.6 143.4
89 45.8 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#82
4-Chlorotoluene
Concen: 51.917 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 91 Resp: 383756
Ion Ratio Lower Upper
91 100
126 32.3 15.9 47.7





#83

tert-Butylbenzene

Concen: 52.772 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:119 Resp: 354690

Ion Ratio Lower Upper

119 100

91 71.2 35.1 105.3





134 27.6 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#84

1,2,4-Trimethylbenzene

Concen: 55.598 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion:105 Resp: 443501

Ion Ratio Lower Upper

105 100

120 44.0 21.6 65.0





#85

sec-Butylbenzene

Concen: 54.205 ug/l

RT: 13.617 min Scan# 19

Delta R.T. 0.005 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 504931

Ion Ratio Lower Upper

105 100





134 19.1 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#86

p-Isopropyltoluene

Concen: 50.325 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion:119 Resp: 422476


Ion Ratio Lower Upper

119 100

134 25.2 12.7 38.0

91 25.0 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 50.068 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:146 Resp: 233419

Ion Ratio Lower Upper

146 100

111 42.4 21.4 64.3





148 64.2 32.3 96.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#88

1,4-Dichlorobenzene

Concen: 47.539 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion:146 Resp: 229722


Ion Ratio Lower Upper

146 100

111 42.1 21.3 63.7

148 64.8 32.4 97.0





#89

n-Butylbenzene

Concen: 52.557 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:	91	Resp:	351230
Ion	Ratio	Lower	Upper
91	100		
92	50.9	25.8	77.3
134	24.0	11.7	35.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#90

Hexachloroethane

Concen: 48.917 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14


Tgt Ion:	117	Resp:	86748
Ion	Ratio	Lower	Upper
117	100		
201	68.1	33.7	101.0





#91
1,2-Dichlorobenzene
Concen: 49.142 ug/l
RT: 14.106 min Scan# 2065
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 228524
Ion Ratio Lower Upper
146 100
111 43.9 21.7 65.1
148 62.5 31.4 94.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 55.439 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 75 Resp: 34650
Ion Ratio Lower Upper
75 100
155 71.7 36.4 109.2
157 91.5 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 50.399 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:180 Resp: 108928

Ion Ratio Lower Upper

180 100

182 92.9 46.8 140.3





145 31.5 16.0 48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#94

Hexachlorobutadiene

Concen: 43.430 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion:225 Resp: 49834


Ion Ratio Lower Upper

225 100

223 65.2 30.7 92.1

227 65.8 30.9 92.5





#95

Naphthalene

Concen: 53.141 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14





Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:	128	Resp:	342729
Ion Ratio	Lower	Upper	
128	100		
127	13.3	10.6	16.0
129	11.0	8.8	13.2

Manual Integrations
APPROVEDReviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025

#96

1,2,3-Trichlorobenzene

Concen: 49.210 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Tgt Ion:	180	Resp:	107618
Ion Ratio	Lower	Upper	
180	100		
182	95.6	47.4	142.2
145	34.4	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085546.D
 Acq On : 28 Jan 2025 18:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 03:39:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.677	0.581	14.2	87	0.00
3 P	Chloromethane	0.733	0.606	17.3	84	0.00
4 C	Vinyl Chloride	0.737	0.641	13.0#	88	0.00
5 T	Bromomethane	0.445	0.362	18.7	82	-0.02
6 T	Chloroethane	0.467	0.426	8.8	95	-0.01
7 T	Trichlorofluoromethane	1.069	1.034	3.3	98	0.00
8 T	Diethyl Ether	0.369	0.351	4.9	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.590	2.0	103	0.00
10 T	Methyl Iodide	0.690	0.686	0.6	90	0.00
11 T	Tert butyl alcohol	0.092	0.130	-41.3#	131	0.00
12 CM	1,1-Dichloroethene	0.537	0.523	2.6#	92	0.00
13 T	Acrolein	0.126	0.138	-9.5	93	0.00
14 T	Allyl chloride	0.871	0.830	4.7	92	0.00
15 T	Acrylonitrile	0.294	0.331	-12.6	105	0.00
16 T	Acetone	0.261	0.285	-9.2	106	0.00
17 T	Carbon Disulfide	1.652	1.307	20.9	83	0.00
18 T	Methyl Acetate	0.793	0.953	-20.2	114	0.00
19 T	Methyl tert-butyl Ether	1.742	1.909	-9.6	96	0.00
20 T	Methylene Chloride	0.646	0.637	1.4	95	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.549	4.2	93	0.00
22 T	Diisopropyl ether	1.933	2.089	-8.1	97	0.00
23 T	Vinyl Acetate	1.353	1.549	-14.5	98	0.00
24 P	1,1-Dichloroethane	1.178	1.186	-0.7	95	0.00
25 T	2-Butanone	0.384	0.451	-17.4	109	0.00
26 T	2,2-Dichloropropane	0.953	1.007	-5.7	101	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.693	-2.7	96	0.00
28 T	Bromochloromethane	0.548	0.477	13.0	83	0.00
29 T	Tetrahydrofuran	0.243	0.294	-21.0	106	0.00
30 C	Chloroform	1.218	1.238	-1.6#	99	0.00
31 T	Cyclohexane	1.024	0.888	13.3	95	0.00
32 T	1,1,1-Trichloroethane	1.068	1.078	-0.9	100	0.00
33 S	1,2-Dichloroethane-d4	0.807	0.882	-9.3	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.347	0.384	-10.7	99	0.00
36 T	1,1-Dichloropropene	0.487	0.485	0.4	97	0.00
37 T	Ethyl Acetate	0.491	0.547	-11.4	103	0.00
38 T	Carbon Tetrachloride	0.557	0.573	-2.9	100	0.00
39 T	Methylcyclohexane	0.457	0.458	-0.2	91	0.00
40 TM	Benzene	1.463	1.498	-2.4	96	0.00
41 T	Methacrylonitrile	0.256	0.311	-21.5	107	0.00
42 TM	1,2-Dichloroethane	0.551	0.571	-3.6	96	0.00
43 T	Isopropyl Acetate	0.787	1.100	-39.8#	125	0.00
44 TM	Trichloroethene	0.341	0.341	0.0	97	0.00
45 C	1,2-Dichloropropane	0.374	0.391	-4.5#	97	0.00
46 T	Dibromomethane	0.270	0.277	-2.6	96	0.00
47 T	Bromodichloromethane	0.549	0.602	-9.7	100	0.00
48 T	Methyl methacrylate	0.354	0.429	-21.2	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085546.D
 Acq On : 28 Jan 2025 18:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 03:39:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	108	0.00
50 S	Toluene-d8	1.232	1.397	-13.4	102	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.577	-26.3#	108	0.00
52 CM	Toluene	0.848	0.923	-8.8#	98	0.00
53 T	t-1,3-Dichloropropene	0.519	0.580	-11.8	97	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.602	-8.7	95	0.00
55 T	1,1,2-Trichloroethane	0.335	0.367	-9.6	100	0.00
56 T	Ethyl methacrylate	0.469	0.583	-24.3	99	0.00
57 T	1,3-Dichloropropane	0.583	0.648	-11.1	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.248	-16.4	95	0.00
59 T	2-Hexanone	0.321	0.420	-30.8#	109	0.00
60 T	Dibromochloromethane	0.405	0.444	-9.6	99	0.00
61 T	1,2-Dibromoethane	0.334	0.359	-7.5	99	0.00
62 S	4-Bromofluorobenzene	0.422	0.503	-19.2	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.341	0.309	9.4	92	0.00
65 PM	Chlorobenzene	1.095	1.102	-0.6	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.411	-2.2	101	0.00
67 C	Ethyl Benzene	1.784	1.936	-8.5#	100	0.00
68 T	m/p-Xylenes	0.659	0.738	-12.0	101	0.00
69 T	o-Xylene	0.630	0.697	-10.6	99	0.00
70 T	Styrene	1.043	1.210	-16.0	100	0.00
71 P	Bromoform	0.288	0.322	-11.8	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.375	3.613	-7.1	103	0.00
74 T	N-amyl acetate	1.517	1.700	-12.1	103	0.00
75 P	1,1,2,2-Tetrachloroethane	1.192	1.228	-3.0	105	0.00
76 T	1,2,3-Trichloropropane	1.016	1.016	0.0	111	0.00
77 T	Bromobenzene	0.882	0.885	-0.3	101	0.00
78 T	n-propylbenzene	3.995	4.293	-7.5	101	0.00
79 T	2-Chlorotoluene	2.586	2.677	-3.5	101	0.00
80 T	1,3,5-Trimethylbenzene	2.787	3.014	-8.1	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.427	-13.6	107	0.00
82 T	4-Chlorotoluene	2.574	2.673	-3.8	101	0.00
83 T	tert-Butylbenzene	2.341	2.471	-5.6	99	0.00
84 T	1,2,4-Trimethylbenzene	2.778	3.089	-11.2	100	0.00
85 T	sec-Butylbenzene	3.244	3.517	-8.4	101	0.00
86 T	p-Isopropyltoluene	2.631	2.943	-11.9	100	0.00
87 T	1,3-Dichlorobenzene	1.624	1.626	-0.1	102	0.00
88 T	1,4-Dichlorobenzene	1.683	1.600	4.9	100	0.00
89 T	n-Butylbenzene	2.327	2.446	-5.1	96	0.00
90 T	Hexachloroethane	0.618	0.604	2.3	103	0.00
91 T	1,2-Dichlorobenzene	1.620	1.592	1.7	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.241	-10.6	106	0.00
93 T	1,2,4-Trichlorobenzene	0.753	0.759	-0.8	95	0.00
94 T	Hexachlorobutadiene	0.400	0.347	13.3	93	0.00
95 T	Naphthalene	2.246	2.387	-6.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085546.D
Acq On : 28 Jan 2025 18:14
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 29 03:39:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.750	1.6	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085546.D
 Acq On : 28 Jan 2025 18:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 03:39:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	42.934	14.1	87	0.00
3 P	Chloromethane	50.000	41.310	17.4	84	0.00
4 C	Vinyl Chloride	50.000	43.525	13.0#	88	0.00
5 T	Bromomethane	50.000	40.633	18.7	82	-0.02
6 T	Chloroethane	50.000	45.629	8.7	95	-0.01
7 T	Trichlorofluoromethane	50.000	48.343	3.3	98	0.00
8 T	Diethyl Ether	50.000	47.541	4.9	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.006	2.0	103	0.00
10 T	Methyl Iodide	50.000	49.778	0.4	90	0.00
11 T	Tert butyl alcohol	250.000	351.637	-40.7#	131	0.00
12 CM	1,1-Dichloroethene	50.000	48.687	2.6#	92	0.00
13 T	Acrolein	250.000	273.407	-9.4	93	0.00
14 T	Allyl chloride	50.000	47.653	4.7	92	0.00
15 T	Acrylonitrile	250.000	281.891	-12.8	105	0.00
16 T	Acetone	250.000	272.891	-9.2	106	0.00
17 T	Carbon Disulfide	50.000	39.567	20.9	83	0.00
18 T	Methyl Acetate	50.000	60.097	-20.2	114	0.00
19 T	Methyl tert-butyl Ether	50.000	54.787	-9.6	96	0.00
20 T	Methylene Chloride	50.000	49.335	1.3	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.884	4.2	93	0.00
22 T	Diisopropyl ether	50.000	54.034	-8.1	97	0.00
23 T	Vinyl Acetate	250.000	286.348	-14.5	98	0.00
24 P	1,1-Dichloroethane	50.000	50.334	-0.7	95	0.00
25 T	2-Butanone	250.000	293.531	-17.4	109	0.00
26 T	2,2-Dichloropropane	50.000	52.834	-5.7	101	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.325	-2.7	96	0.00
28 T	Bromochloromethane	50.000	43.487	13.0	83	0.00
29 T	Tetrahydrofuran	250.000	301.616	-20.6	106	0.00
30 C	Chloroform	50.000	50.826	-1.7#	99	0.00
31 T	Cyclohexane	50.000	43.331	13.3	95	0.00
32 T	1,1,1-Trichloroethane	50.000	50.429	-0.9	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	54.638	-9.3	100	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35 S	Dibromofluoromethane	50.000	55.349	-10.7	99	0.00
36 T	1,1-Dichloropropene	50.000	49.851	0.3	97	0.00
37 T	Ethyl Acetate	50.000	55.616	-11.2	103	0.00
38 T	Carbon Tetrachloride	50.000	51.368	-2.7	100	0.00
39 T	Methylcyclohexane	50.000	50.030	-0.1	91	0.00
40 TM	Benzene	50.000	51.218	-2.4	96	0.00
41 T	Methacrylonitrile	50.000	60.848	-21.7	107	0.00
42 TM	1,2-Dichloroethane	50.000	51.863	-3.7	96	0.00
43 T	Isopropyl Acetate	50.000	69.836	-39.7#	125	0.00
44 TM	Trichloroethene	50.000	50.023	-0.0	97	0.00
45 C	1,2-Dichloropropane	50.000	52.380	-4.8#	97	0.00
46 T	Dibromomethane	50.000	51.408	-2.8	96	0.00
47 T	Bromodichloromethane	50.000	54.778	-9.6	100	0.00
48 T	Methyl methacrylate	50.000	60.523	-21.0	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085546.D
 Acq On : 28 Jan 2025 18:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 03:39:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1240.952	-24.1	108	0.00
50 S	Toluene-d8	50.000	56.666	-13.3	102	0.00
51 T	4-Methyl-2-Pentanone	250.000	315.843	-26.3#	108	0.00
52 CM	Toluene	50.000	54.423	-8.8#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	55.846	-11.7	97	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.260	-8.5	95	0.00
55 T	1,1,2-Trichloroethane	50.000	54.714	-9.4	100	0.00
56 T	Ethyl methacrylate	50.000	51.541	-3.1	99	0.00
57 T	1,3-Dichloropropane	50.000	55.519	-11.0	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	291.087	-16.4	95	0.00
59 T	2-Hexanone	250.000	326.364	-30.5#	109	0.00
60 T	Dibromochloromethane	50.000	54.855	-9.7	99	0.00
61 T	1,2-Dibromoethane	50.000	53.765	-7.5	99	0.00
62 S	4-Bromofluorobenzene	50.000	59.614	-19.2	103	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	96	0.00
64 T	Tetrachloroethene	50.000	45.330	9.3	92	0.00
65 PM	Chlorobenzene	50.000	50.289	-0.6	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.180	-2.4	101	0.00
67 C	Ethyl Benzene	50.000	54.268	-8.5#	100	0.00
68 T	m/p-Xylenes	100.000	111.987	-12.0	101	0.00
69 T	o-Xylene	50.000	55.313	-10.6	99	0.00
70 T	Styrene	50.000	58.032	-16.1	100	0.00
71 P	Bromoform	50.000	55.980	-12.0	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	53.526	-7.1	103	0.00
74 T	N-amyl acetate	50.000	56.051	-12.1	103	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.507	-3.0	105	0.00
76 T	1,2,3-Trichloropropane	50.000	49.995	0.0	111	0.00
77 T	Bromobenzene	50.000	50.181	-0.4	101	0.00
78 T	n-propylbenzene	50.000	53.728	-7.5	101	0.00
79 T	2-Chlorotoluene	50.000	51.776	-3.6	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	54.059	-8.1	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	56.880	-13.8	107	0.00
82 T	4-Chlorotoluene	50.000	51.917	-3.8	101	0.00
83 T	tert-Butylbenzene	50.000	52.772	-5.5	99	0.00
84 T	1,2,4-Trimethylbenzene	50.000	55.598	-11.2	100	0.00
85 T	sec-Butylbenzene	50.000	54.205	-8.4	101	0.00
86 T	p-Isopropyltoluene	50.000	50.325	-0.7	100	0.00
87 T	1,3-Dichlorobenzene	50.000	50.068	-0.1	102	0.00
88 T	1,4-Dichlorobenzene	50.000	47.539	4.9	100	0.00
89 T	n-Butylbenzene	50.000	52.557	-5.1	96	0.00
90 T	Hexachloroethane	50.000	48.917	2.2	103	0.00
91 T	1,2-Dichlorobenzene	50.000	49.142	1.7	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	55.439	-10.9	106	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.399	-0.8	95	0.00
94 T	Hexachlorobutadiene	50.000	43.430	13.1	93	0.00
95 T	Naphthalene	50.000	53.141	-6.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085546.D
Acq On : 28 Jan 2025 18:14
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 29 03:39:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.210	1.6	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085437.D
Acq On : 14 Jan 2025 14:22
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Title : SW846 8260
Last Update : Wed Jan 15 02:16:08 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843



Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	4864	PASS
75	95	30	60	58.0	12679	PASS
95	95	100	100	100.0	21856	PASS
96	95	5	9	7.8	1703	PASS
173	174	0.00	2	1.8	296	PASS
174	95	50	100	76.0	16613	PASS
175	174	5	9	7.1	1179	PASS
176	174	95	101	97.4	16185	PASS
177	176	5	9	6.7	1079	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085525.D
Acq On : 28 Jan 2025 09:17
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Title : SW846 8260
Last Update : Wed Jan 15 02:16:08 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	6102	PASS
75	95	30	60	55.9	16153	PASS
95	95	100	100	100.0	28877	PASS
96	95	5	9	7.4	2133	PASS
173	174	0.00	2	1.9	424	PASS
174	95	50	100	75.9	21904	PASS
175	174	5	9	7.2	1583	PASS
176	174	95	101	97.2	21288	PASS
177	176	5	9	6.8	1447	PASS

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:			
Project:	CTO WE13			Date Received:			
Client Sample ID:	VN0128WBL01			SDG No.:	Q1199		
Lab Sample ID:	VN0128WBL01			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085528.D	1		01/28/25 10:51	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN0128WBL01			SDG No.:	Q1199
Lab Sample ID:	VN0128WBL01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085528.D	1		01/28/25 10:51	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.7		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	52.2		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	50.1		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.2		85 - 114		88%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	184000	8.218				
540-36-3	1,4-Difluorobenzene	333000	9.1				
3114-55-4	Chlorobenzene-d5	291000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	116000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085528.D
 Acq On : 28 Jan 2025 10:51
 Operator : JC\MD
 Sample : VN0128WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBL01

Quant Time: Jan 29 00:35:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	183717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	332601	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	116250	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159230	53.693	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	107.380%	
35) Dibromofluoromethane	8.165	113	120495	52.220	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.440%	
50) Toluene-d8	10.565	98	410338	50.052	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.100%	
62) 4-Bromofluorobenzene	12.847	95	124045	44.232	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.460%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085528.D
Acq On : 28 Jan 2025 10:51
Operator : JC\MD
Sample : VN0128WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBL01


Quant Time: Jan 29 00:35:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085528.D
Acq: 28 Jan 2025 10:51

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBL01







Tgt Ion:168 Resp: 183717
Ion Ratio Lower Upper
168 100
99 63.9 53.6 80.4



#33
1,2-Dichloroethane-d4
Concen: 53.693 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN085528.D
Acq: 28 Jan 2025 10:51

Tgt Ion: 65 Resp: 159230
Ion Ratio Lower Upper
65 100
67 51.5 0.0 101.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1214

Delta R.T. 0.006 min

Lab File: VN085528.D

Acq: 28 Jan 2025 10:51

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBL01





Tgt Ion:114 Resp: 332601

Ion Ratio Lower Upper

114 100

63 22.6 0.0 47.6

88 17.2 0.0 32.6



#35

Dibromofluoromethane

Concen: 52.220 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085528.D

Acq: 28 Jan 2025 10:51




Tgt Ion:113 Resp: 120495


Ion Ratio Lower Upper

113 100

111 103.9 82.7 124.1

192 17.8 14.3 21.5





#50

Toluene-d8

Concen: 50.052 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085528.D

Acq: 28 Jan 2025 10:51

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBL01

Tgt Ion: 98 Resp: 410338

Ion Ratio Lower Upper

98 100

100 63.0 52.2 78.4



#62

4-Bromofluorobenzene

Concen: 44.232 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. 0.000 min

Lab File: VN085528.D

Acq: 28 Jan 2025 10:51




Tgt Ion: 95 Resp: 124045


Ion Ratio Lower Upper

95 100

174 74.0 0.0 145.0

176 70.5 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085528.D

Acq: 28 Jan 2025 10:51

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBL01


Tgt Ion:117 Resp: 291066

Ion Ratio Lower Upper

117 100

82 58.6 48.6 72.8

119 31.5 26.6 39.8



Abundance

150000



11.865

100000

50000

0

Time-->



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085528.D

Acq: 28 Jan 2025 10:51


Tgt Ion:152 Resp: 116250

Ion Ratio Lower Upper

152 100

115 60.3 31.1 93.3

150 155.3 0.0 343.6



Abundance

100000

80000


60000

40000

20000

0

Time-->



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085528.D
Acq On : 28 Jan 2025 10:51
Operator : JC\MD
Sample : VN0128WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBL01

Integration Parameters: RTEINT.P
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Title : SW846 8260

Signal : TIC: VN085528.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	8.165	1046	1056	1060	rBV	167943	401423	36.19%	7.042%
2	8.224	1060	1066	1082	rVB	267962	606129	54.65%	10.633%
3	8.577	1116	1126	1138	rBV	191499	427577	38.55%	7.501%
4	9.100	1205	1215	1229	rBV	407646	827937	74.65%	14.524%
5	10.565	1455	1464	1480	rBV	607166	1109139	100.00%	19.457%
6	11.865	1676	1685	1697	rBV	529971	948825	85.55%	16.645%
7	12.847	1845	1852	1860	rBV	378051	632238	57.00%	11.091%
8	13.788	2005	2012	2021	rBV	449400	747250	67.37%	13.108%




Sum of corrected areas: 5700518

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085528.D
Acq On : 28 Jan 2025 10:51
Operator : JC\MD
Sample : VN0128WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085528.D
Acq On : 28 Jan 2025 10:51
Operator : JC\MD
Sample : VN0128WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085528.D
Acq On : 28 Jan 2025 10:51
Operator : JC\MD
Sample : VN0128WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--
				#	RT Resp Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN0128WBS01			SDG No.:	Q1199
Lab Sample ID:	VN0128WBS01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085529.D	1		01/28/25 11:15	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	17.3		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.2		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	19.5		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.8		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	20.7		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.5		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.1		0.26	0.75	1.00	ug/L
67-64-1	Acetone	100		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.4		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.9		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	19.8		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.9		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.1		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	110		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	20.5		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.9		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	20.6		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.4		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.9		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.7		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.0		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	20.4		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.7		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	21.7		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	120		0.75	2.50	5.00	ug/L
108-88-3	Toluene	21.9		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	21.4		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	21.5		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.9		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	120		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN0128WBS01			SDG No.:	Q1199
Lab Sample ID:	VN0128WBS01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085529.D	1		01/28/25 11:15	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	21.2		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	21.7		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.6		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.8		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	43.1		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	20.9		0.14	0.50	1.00	ug/L
100-42-5	Styrene	21.3		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	22.5		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.9		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.6		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.8		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.8		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.4		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.1		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	53.2		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	54.1		89 - 112		108%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.0		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	186000	8.224				
540-36-3	1,4-Difluorobenzene	308000	9.1				
3114-55-4	Chlorobenzene-d5	271000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	127000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085529.D
 Acq On : 28 Jan 2025 11:15
 Operator : JC\MD
 Sample : VN0128WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/29/2025
 Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:35:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	307569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	271043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126721	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	153168	51.135	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.280%
35) Dibromofluoromethane	8.165	113	113571	53.225	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.460%
50) Toluene-d8	10.565	98	410056	54.088	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.180%
62) 4-Bromofluorobenzene	12.847	95	134929	52.029	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.060%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	45881	18.261	ug/l	99
3) Chloromethane	2.359	50	46936	17.254	ug/l	98
4) Vinyl Chloride	2.512	62	49715	18.182	ug/l	98
5) Bromomethane	2.965	94	32217	19.506	ug/l	93
6) Chloroethane	3.118	64	34288	19.780	ug/l	95
7) Trichlorofluoromethane	3.501	101	82151	20.707	ug/l	94
8) Diethyl Ether	3.959	74	26345	19.222	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.377	101	45774	20.484	ug/l	98
10) Methyl Iodide	4.589	142	50026	19.549	ug/l	95
11) Tert butyl alcohol	5.512	59	37334	108.849	ug/l	99
12) 1,1-Dichloroethene	4.342	96	37952	19.057	ug/l	93
13) Acrolein	4.177	56	39646	84.676	ug/l	98
14) Allyl chloride	5.024	41	59056	18.275	ug/l	98
15) Acrylonitrile	5.712	53	116207	106.667	ug/l	98
16) Acetone	4.424	43	101360	104.730	ug/l	94
17) Carbon Disulfide	4.706	76	100344	16.364	ug/l	98
18) Methyl Acetate	5.018	43	66988	22.762	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	135056	20.885	ug/l	97
20) Methylene Chloride	5.277	84	47487	19.820	ug/l	96
21) trans-1,2-Dichloroethene	5.789	96	40233	18.903	ug/l	94
22) Diisopropyl ether	6.671	45	147403	20.550	ug/l	97
23) Vinyl Acetate	6.600	43	516002	102.796	ug/l	99
24) 1,1-Dichloroethane	6.565	63	87952	20.110	ug/l	99
25) 2-Butanone	7.477	43	153642	107.876	ug/l	98
26) 2,2-Dichloropropane	7.488	77	78263	22.133	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	49822	19.875	ug/l	99
28) Bromochloromethane	7.812	49	32851	16.145	ug/l	96
29) Tetrahydrofuran	7.836	42	102010	112.971	ug/l	99
30) Chloroform	7.965	83	92905	20.553	ug/l	93
31) Cyclohexane	8.253	56	67069	17.643	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	80919	20.408	ug/l	97
36) 1,1-Dichloropropene	8.365	75	58916	19.673	ug/l	99
37) Ethyl Acetate	7.553	43	62638	20.722	ug/l	98
38) Carbon Tetrachloride	8.359	117	70429	20.543	ug/l	97
39) Methylcyclohexane	9.600	83	53089	18.865	ug/l	96
40) Benzene	8.606	78	186111	20.683	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085529.D
 Acq On : 28 Jan 2025 11:15
 Operator : JC\MD
 Sample : VN0128WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/29/2025
 Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:35:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	33670	21.410	ug/l	98
42) 1,2-Dichloroethane	8.665	62	71068	20.970	ug/l	100
43) Isopropyl Acetate	8.688	43	105036	21.686	ug/l	99
44) Trichloroethene	9.347	130	42674	20.374	ug/l	97
45) 1,2-Dichloropropane	9.618	63	47674	20.743	ug/l	98
46) Dibromomethane	9.706	93	33823	20.395	ug/l	97
47) Bromodichloromethane	9.882	83	73309	21.697	ug/l	97
48) Methyl methacrylate	9.677	41	47567	21.823	ug/l	98
49) 1,4-Dioxane	9.694	88	15432	420.430	ug/l	98
51) 4-Methyl-2-Pentanone	10.441	43	324778	115.555	ug/l	100
52) Toluene	10.629	92	113994	21.864	ug/l	99
53) t-1,3-Dichloropropene	10.829	75	68322	21.398	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	73429	21.530	ug/l	95
55) 1,1,2-Trichloroethane	11.012	97	45212	21.912	ug/l	94
56) Ethyl methacrylate	10.871	69	62727	19.281	ug/l	98
57) 1,3-Dichloropropane	11.159	76	78623	21.915	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	114984	87.817	ug/l	100
59) 2-Hexanone	11.194	43	234494	118.574	ug/l	98
60) Dibromochloromethane	11.359	129	52835	21.210	ug/l	99
61) 1,2-Dibromoethane	11.471	107	42361	20.625	ug/l	99
64) Tetrachloroethene	11.100	164	40019	21.658	ug/l	99
65) Chlorobenzene	11.888	112	122235	20.586	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	46034	21.124	ug/l	99
67) Ethyl Benzene	11.959	91	201123	20.797	ug/l	98
68) m/p-Xylenes	12.071	106	154069	43.106	ug/l	99
69) o-Xylene	12.394	106	71535	20.943	ug/l	100
70) Styrene	12.412	104	120125	21.250	ug/l	99
71) Bromoform	12.576	173	35108	22.516	ug/l #	99
73) Isopropylbenzene	12.694	105	178577	20.880	ug/l	98
74) N-ethyl acetate	12.488	43	80633	20.975	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	65405	21.649	ug/l	98
76) 1,2,3-Trichloropropane	12.988	75	60066m	23.321	ug/l	
77) Bromobenzene	12.976	156	47665	21.332	ug/l	99
78) n-propylbenzene	13.035	91	217855	21.517	ug/l	99
79) 2-Chlorotoluene	13.123	91	138138	21.081	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	154258	21.836	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	21965	23.073	ug/l	87
82) 4-Chlorotoluene	13.218	91	139420	21.369	ug/l	100
83) tert-Butylbenzene	13.435	119	124147	20.926	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	155548	22.092	ug/l	98
85) sec-Butylbenzene	13.612	105	178351	21.691	ug/l	100
86) p-Isopropyltoluene	13.729	119	143953	20.952	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	85729	20.833	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	84406	19.789	ug/l	99
89) n-Butylbenzene	14.053	91	115181	19.526	ug/l	100
90) Hexachloroethane	14.329	117	30558	19.522	ug/l	97
91) 1,2-Dichlorobenzene	14.100	146	83529	20.350	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12961	23.494	ug/l	91
93) 1,2,4-Trichlorobenzene	15.394	180	36081	18.913	ug/l	95
94) Hexachlorobutadiene	15.494	225	20163	19.907	ug/l	98
95) Naphthalene	15.641	128	104713	18.394	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	36634	18.978	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085529.D
Acq On : 28 Jan 2025 11:15
Operator : JC\MD
Sample : VN0128WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:35:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085529.D
Acq On : 28 Jan 2025 11:15
Operator : JC\MD
Sample : VN0128WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025


Quant Time: Jan 29 00:35:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01



Tgt Ion: 168 Resp: 185562
Ion Ratio Lower Upper
168 100
99 65.3 53.6 80.4




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025



#2
Dichlorodifluoromethane
Concen: 18.261 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15


Tgt Ion: 85 Resp: 45881
Ion Ratio Lower Upper
85 100
87 33.0 16.7 50.1





#3
Chloromethane
Concen: 17.254 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01



Tgt Ion: 50 Resp: 46936
Ion Ratio Lower Upper
50 100
52 35.2 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#4
Vinyl Chloride
Concen: 18.182 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15


Tgt Ion: 62 Resp: 49715
Ion Ratio Lower Upper
62 100
64 32.1 24.8 37.2





#5
Bromomethane
Concen: 19.506 ug/l
RT: 2.965 min Scan# 170
Delta R.T. 0.012 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01



Tgt Ion: 94 Resp: 32217
Ion Ratio Lower Upper
94 100
96 83.2 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#6
Chloroethane
Concen: 19.780 ug/l
RT: 3.118 min Scan# 198
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 64 Resp: 34288
Ion Ratio Lower Upper
64 100
66 33.3 24.6 36.8





#7

Trichlorofluoromethane

Concen: 20.707 ug/l

RT: 3.501 min Scan# 26

Delta R.T. 0.000 min

Lab File: VN085529.D


Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01



Tgt Ion:101 Resp: 82151

Ion Ratio Lower Upper

101 100




103 69.3 51.4 77.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#8

Diethyl Ether

Concen: 19.222 ug/l

RT: 3.959 min Scan# 341

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 74 Resp: 26345

Ion Ratio Lower Upper

74 100


45 97.4 49.7 149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.484 ug/l
RT: 4.377 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01




Tgt Ion:101 Resp: 45774
Ion Ratio Lower Upper
101 100
85 44.9 37.8 56.8
151 72.0 58.8 88.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025




#10
Methyl Iodide
Concen: 19.549 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15



Tgt Ion:142 Resp: 50026
Ion Ratio Lower Upper
142 100
127 54.4 40.3 60.5
141 16.2 12.8 19.2





#11

Tert butyl alcohol

Concen: 108.849 ug/l

RT: 5.512 min Scan# 605

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 59 Resp: 37334

Ion Ratio Lower Upper

59 100





57 10.4 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#12

1,1-Dichloroethene

Concen: 19.057 ug/l

RT: 4.342 min Scan# 406

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 96 Resp: 37952


Ion Ratio Lower Upper

96 100

61 185.5 141.0 211.4

98 66.4 49.0 73.6





#13

Acrolein

Concen: 84.676 ug/l

RT: 4.177 min Scan# 377

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 56 Resp: 39646

Ion Ratio Lower Upper

56 100





55 71.9 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#14

Allyl chloride

Concen: 18.275 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 41 Resp: 59056


Ion Ratio Lower Upper

41 100

39 83.5 64.4 96.6

76 38.4 30.5 45.7





#15

Acrylonitrile

Concen: 106.667 ug/l

RT: 5.712 min Scan# 639

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 53 Resp: 116207

Ion Ratio Lower Upper

53 100

52 83.3 65.5 98.3





51 36.3 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#16

Acetone

Concen: 104.730 ug/l

RT: 4.424 min Scan# 420

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 43 Resp: 101360

Ion Ratio Lower Upper

43 100

58 26.7 23.8 35.6





#17

Carbon Disulfide

Concen: 16.364 ug/l

RT: 4.706 min Scan# 468

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 76 Resp: 100344

Ion Ratio Lower Upper

76 100





78 9.2 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#18

Methyl Acetate

Concen: 22.762 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 43 Resp: 66988

Ion Ratio Lower Upper

43 100

74 23.6 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 20.885 ug/l

RT: 5.789 min Scan# 65

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 73 Resp: 135056

Ion Ratio Lower Upper

73 100





57 20.8 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#20

Methylene Chloride

Concen: 19.820 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Tgt Ion: 84 Resp: 47487




Ion Ratio Lower Upper


84 100

49 126.6 104.1 156.1

51 34.2 31.0 46.4

86 62.2 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 18.903 ug/l

RT: 5.789 min Scan# 651

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 96 Resp: 40233

Ion Ratio Lower Upper

96 100

61 152.9 115.7 173.5





98 66.4 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#22

Diisopropyl ether

Concen: 20.550 ug/l

RT: 6.671 min Scan# 802

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Tgt Ion: 45 Resp: 147403




Ion Ratio Lower Upper


45 100

43 55.8 45.0 67.6

87 31.0 21.8 32.8

59 11.8 10.2 15.2





#23

Vinyl Acetate

Concen: 102.796 ug/l

RT: 6.600 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 43 Resp: 516002

Ion Ratio Lower Upper

43 100





86 10.2 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#24

1,1-Dichloroethane

Concen: 20.110 ug/l

RT: 6.565 min Scan# 784

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 63 Resp: 87952


Ion Ratio Lower Upper

63 100

98 6.5 3.3 9.8

100 4.7 2.0 6.0





#25

2-Butanone

Concen: 107.876 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 43 Resp: 153642

Ion Ratio Lower Upper

43 100





72 26.4 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#26

2,2-Dichloropropane

Concen: 22.133 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 77 Resp: 78263

Ion Ratio Lower Upper

77 100


97 21.4 10.7 32.1





#27
cis-1,2-Dichloroethene
Concen: 19.875 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01



Tgt Ion: 96 Resp: 49822
Ion Ratio Lower Upper
96 100
61 157.4 0.0 311.8
98 62.7 0.0 126.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#28
Bromochloromethane
Concen: 16.145 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 49 Resp: 32851
Ion Ratio Lower Upper
49 100
129 1.5 0.0 4.4
130 72.0 55.0 82.4









#29
Tetrahydrofuran
Concen: 112.971 ug/l
RT: 7.836 min Scan# 1020
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Tgt Ion: 42 Resp: 102010
Ion Ratio Lower Upper
42 100
72 41.9 34.2 51.2
71 40.5 32.5 48.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#30
Chloroform
Concen: 20.553 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 83 Resp: 92905
Ion Ratio Lower Upper
83 100
85 59.4 51.8 77.6





#31

Cyclohexane

Concen: 17.643 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 56 Resp: 67069

Ion Ratio Lower Upper

56 100

69 29.0 22.2 33.4





84 84.8 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#32

1,1,1-Trichloroethane

Concen: 20.408 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 97 Resp: 80919


Ion Ratio Lower Upper

97 100

99 63.9 49.8 74.6


61 48.7 41.4 62.2





#33
1,2-Dichloroethane-d4
Concen: 51.135 ug/l
RT: 8.577 min Scan# 1125
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01




Tgt Ion: 65 Resp: 153168
Ion Ratio Lower Upper
65 100
67 50.0 0.0 101.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15



Tgt Ion: 114 Resp: 307569
Ion Ratio Lower Upper
114 100
63 22.9 0.0 47.6
88 16.7 0.0 32.6





#35

Dibromofluoromethane

Concen: 53.225 ug/l

RT: 8.165 min Scan# 1055

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 113 Resp: 113571

Ion Ratio Lower Upper

113 100

111 103.6 82.7 124.1





192 17.7 14.3 21.5

Manual Integrations

APPROVED

Reviewed By : John Carlone 01/29/2025

Supervised By : Semsettin Yesilyurt 01/29/2025



#36

1,1-Dichloropropene

Concen: 19.673 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 75 Resp: 58916


Ion Ratio Lower Upper

75 100

110 33.1 16.5 49.5

77 31.4 24.4 36.6





#37

Ethyl Acetate

Concen: 20.722 ug/l

RT: 7.553 min Scan# 952

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 43 Resp: 62638

Ion Ratio Lower Upper

43 100

61 13.1 10.9 16.3





70 10.3 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#38

Carbon Tetrachloride

Concen: 20.543 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion:117 Resp: 70429


Ion Ratio Lower Upper

117 100

119 97.9 75.4 113.2

121 30.7 24.6 37.0





#39

Methylcyclohexane

Concen: 18.865 ug/l

RT: 9.600 min Scan# 131

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N





ClientSampleId :

VN0128WBS01

Tgt Ion:	83	Resp:	53089
Ion	Ratio	Lower	Upper
83	100		
55	83.8	62.6	94.0
98	47.6	37.7	56.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#40

Benzene

Concen: 20.683 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Tgt Ion:	78	Resp:	186111
Ion	Ratio	Lower	Upper
78	100		
77	23.7	19.0	28.6





#41

Methacrylonitrile

Concen: 21.410 ug/l

RT: 7.777 min Scan# 990

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 41 Resp: 33670

Ion Ratio Lower Upper

41 100

39 60.5 46.0 69.0

67 71.9 57.4 86.2





52 32.1 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#42

1,2-Dichloroethane

Concen: 20.970 ug/l

RT: 8.665 min Scan# 1141

Delta R.T. -0.000 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 62 Resp: 71068

Ion Ratio Lower Upper

62 100

98 8.4 0.0 17.0





#43

Isopropyl Acetate

Concen: 21.686 ug/l

RT: 8.688 min Scan# 1144

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 43 Resp: 105036

Ion Ratio Lower Upper

43 100

61 25.2 20.7 31.1





87 12.1 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#44

Trichloroethene

Concen: 20.374 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. 0.000 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion:130 Resp: 42674

Ion Ratio Lower Upper

130 100

95 100.5 0.0 195.8





#45

1,2-Dichloropropane

Concen: 20.743 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 63 Resp: 47674

Ion Ratio Lower Upper

63 100





65 31.1 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#46

Dibromomethane

Concen: 20.395 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 93 Resp: 33823


Ion Ratio Lower Upper

93 100

95 81.5 64.7 97.1

174 90.4 69.0 103.4





#47

Bromodichloromethane

Concen: 21.697 ug/l

RT: 9.882 min Scan# 1348

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 83 Resp: 73309

Ion Ratio Lower Upper

83 100

85 61.0 51.2 76.8





127 8.4 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#48

Methyl methacrylate

Concen: 21.823 ug/l

RT: 9.677 min Scan# 1313

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 41 Resp: 47567


Ion Ratio Lower Upper

41 100

69 81.0 64.7 97.1

39 65.5 49.0 73.6





#49

1,4-Dioxane

Concen: 420.430 ug/l

RT: 9.694 min Scan# 1315

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 88 Resp: 15432

Ion Ratio Lower Upper

88 100

43 37.0 26.6 39.8





58 74.0 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#50

Toluene-d8

Concen: 54.088 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 98 Resp: 410056

Ion Ratio Lower Upper

98 100

100 64.1 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 115.555 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Tgt Ion: 43 Resp: 324778
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#52
Toluene
Concen: 21.864 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 92 Resp: 113994
Ion Ratio Lower Upper
92 100
91 172.4 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 21.398 ug/l

RT: 10.829 min Scan# 1510

Delta R.T. -0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 75 Resp: 68322

Ion Ratio Lower Upper

75 100





77 29.3 23.5 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#54

cis-1,3-Dichloropropene

Concen: 21.530 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 75 Resp: 73429


Ion Ratio Lower Upper

75 100

77 33.1 25.0 37.4

39 49.3 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 21.912 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 97 Resp: 45212

Ion Ratio Lower Upper

97 100

83 92.7 69.0 103.6

85 59.4 44.6 66.8





99 63.7 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#56

Ethyl methacrylate

Concen: 19.281 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 69 Resp: 62727


Ion Ratio Lower Upper

69 100

41 69.3 54.6 82.0


39 42.9 32.4 48.6





#57
1,3-Dichloropropane
Concen: 21.915 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01



Tgt Ion: 76 Resp: 78623
Ion Ratio Lower Upper
76 100
78 32.0 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#58
2-Chloroethyl Vinyl ether
Concen: 87.817 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 63 Resp: 114984
Ion Ratio Lower Upper
63 100
106 27.1 21.6 32.4





#59

2-Hexanone

Concen: 118.574 ug/l

RT: 11.194 min Scan# 15

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 43 Resp: 234494

Ion Ratio Lower Upper

43 100





58 51.0 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#60

Dibromochloromethane

Concen: 21.210 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion:129 Resp: 52835

Ion Ratio Lower Upper

129 100

127 75.9 38.6 115.8









#61
1,2-Dibromoethane
Concen: 20.625 ug/l
RT: 11.471 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Tgt Ion: 107 Resp: 42361
Ion Ratio Lower Upper
107 100
109 95.9 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025



#62
4-Bromofluorobenzene
Concen: 52.029 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 95 Resp: 134929
Ion Ratio Lower Upper
95 100
174 75.1 0.0 145.0
176 70.9 0.0 142.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion:117 Resp: 271043

Ion Ratio Lower Upper

117 100

82 58.5 48.6 72.8





119 31.9 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#64

Tetrachloroethene

Concen: 21.658 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Tgt Ion:164 Resp: 40019




Ion Ratio Lower Upper


164 100

166 128.2 103.4 155.2

129 95.9 79.2 118.8

131 96.1 77.1 115.7





#65

Chlorobenzene

Concen: 20.586 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion:112 Resp: 122235

Ion Ratio Lower Upper

112 100





114 31.6 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 21.124 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion:131 Resp: 46034


Ion Ratio Lower Upper

131 100

133 95.1 47.4 142.3

119 65.4 33.1 99.5





#67

Ethyl Benzene

Concen: 20.797 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 91 Resp: 201123

Ion Ratio Lower Upper

91 100





106 28.8 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#68

m/p-Xylenes

Concen: 43.106 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion:106 Resp: 154069

Ion Ratio Lower Upper

106 100

91 208.4 167.7 251.5





#69

o-Xylene

Concen: 20.943 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion:106 Resp: 71535

Ion Ratio Lower Upper

106 100





91 220.7 110.4 331.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#70

Styrene

Concen: 21.250 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion:104 Resp: 120125


Ion Ratio Lower Upper

104 100

78 53.0 42.5 63.7

103 55.9 43.8 65.8





#71

Bromoform

Concen: 22.516 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion:173 Resp: 35108

Ion Ratio Lower Upper

173 100

175 48.0 24.4 73.2





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion:152 Resp: 126721


Ion Ratio Lower Upper

152 100

115 63.7 31.1 93.3

150 163.7 0.0 343.6





#73

Isopropylbenzene

Concen: 20.880 ug/l

RT: 12.694 min Scan# 1825

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion:105 Resp: 178577

Ion Ratio Lower Upper

105 100





120 26.7 12.8 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#74

N-ethyl acetate

Concen: 20.975 ug/l

RT: 12.488 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Tgt Ion: 43 Resp: 80633




Ion Ratio Lower Upper


43 100

70 41.7 34.0 51.0

55 28.0 21.4 32.2

61 23.5 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 21.649 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBS01

Tgt Ion:	83	Resp:	65405
Ion	Ratio	Lower	Upper
83	100		
131	8.8	4.8	14.4
85	62.7	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#76

1,2,3-Trichloropropane

Concen: 23.321 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Tgt Ion:	75	Resp:	60066
Ion	Ratio	Lower	Upper
75	100		
77	181.2	109.7	329.2





#77

Bromobenzene

Concen: 21.332 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion:156 Resp: 47665

Ion Ratio Lower Upper

156 100

77 228.4 114.1 342.4





158 94.0 48.9 146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#78

n-propylbenzene

Concen: 21.517 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 91 Resp: 217855

Ion Ratio Lower Upper

91 100

120 22.1 10.9 32.6





#79

2-Chlorotoluene

Concen: 21.081 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion: 91 Resp: 138138

Ion Ratio Lower Upper

91 100





126 31.9 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#80

1,3,5-Trimethylbenzene

Concen: 21.836 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion:105 Resp: 154258

Ion Ratio Lower Upper

105 100

120 47.7 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 23.073 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Tgt Ion	Ratio	Lower	Upper
75	100		
53	101.0	95.6	143.4
89	44.0	37.0	55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#82
4-Chlorotoluene
Concen: 21.369 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion	Ratio	Lower	Upper
91	100		
126	31.6	15.9	47.7





#83

tert-Butylbenzene

Concen: 20.926 ug/l

RT: 13.435 min Scan# 19

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion:119 Resp: 124147

Ion Ratio Lower Upper

119 100

91 71.0 35.1 105.3





134 27.7 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#84

1,2,4-Trimethylbenzene

Concen: 22.092 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion:105 Resp: 155548

Ion Ratio Lower Upper

105 100

120 44.5 21.6 65.0









#85
sec-Butylbenzene
Concen: 21.691 ug/l
RT: 13.612 min Scan# 1982
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Tgt Ion:105 Resp: 178351
Ion Ratio Lower Upper
105 100
134 19.5 9.7 28.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#86
p-Isopropyltoluene
Concen: 20.952 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:119 Resp: 143953
Ion Ratio Lower Upper
119 100
134 25.1 12.7 38.0
91 25.5 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 20.833 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01





Tgt Ion:	146	Resp:	85729
Ion Ratio	Lower	Upper	
146	100		
111	41.8	21.4	64.3
148	62.6	32.3	96.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#88

1,4-Dichlorobenzene

Concen: 19.789 ug/l




RT: 13.812 min Scan# 2016


Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Tgt Ion:	146	Resp:	84406
Ion Ratio	Lower	Upper	
146	100		
111	43.9	21.3	63.7
148	65.4	32.4	97.0









#89
n-Butylbenzene
Concen: 19.526 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Tgt Ion:	91	Resp:	115181
Ion Ratio	Lower	Upper	
91	100		
92	51.5	25.8	77.3
134	23.8	11.7	35.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#90
Hexachloroethane
Concen: 19.522 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15


Tgt Ion:	117	Resp:	30558
Ion Ratio	Lower	Upper	
117	100		
201	70.0	33.7	101.0





#91
1,2-Dichlorobenzene
Concen: 20.350 ug/l
RT: 14.100 min Scan# 2065
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01



Tgt Ion: 146 Resp: 83529
Ion Ratio Lower Upper
146 100
111 42.3 21.7 65.1
148 63.4 31.4 94.2




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 23.494 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 75 Resp: 12961
Ion Ratio Lower Upper
75 100
155 68.3 36.4 109.2
157 79.8 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 18.913 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01

Tgt Ion:180 Resp: 36081

Ion Ratio Lower Upper

180 100

182 99.0 46.8 140.3





145 32.9 16.0 48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#94

Hexachlorobutadiene

Concen: 19.907 ug/l

RT: 15.494 min Scan# 2302

Delta R.T. -0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion:225 Resp: 20163


Ion Ratio Lower Upper

225 100

223 63.1 30.7 92.1

227 62.5 30.9 92.5





#95

Naphthalene

Concen: 18.394 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01





Tgt Ion:	128	Resp:	104713
Ion Ratio	Lower	Upper	
128	100		
127	13.6	10.6	16.0
129	11.0	8.8	13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#96

1,2,3-Trichlorobenzene

Concen: 18.978 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Tgt Ion:	180	Resp:	36634
Ion Ratio	Lower	Upper	
180	100		
182	97.1	47.4	142.2
145	34.3	16.9	50.7



Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN0128WBSD01			SDG No.:	Q1199
Lab Sample ID:	VN0128WBSD01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085530.D	1		01/28/25 11:48	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	17.7		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.8		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	18.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.6		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.5		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.5		0.26	0.75	1.00	ug/L
67-64-1	Acetone	110		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.2		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	21.9		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	19.7		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.9		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.4		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	110		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	20.0		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.2		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	20.6		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.1		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.7		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.2		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.5		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.4		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.7		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	21.7		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	120		0.75	2.50	5.00	ug/L
108-88-3	Toluene	21.5		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	21.5		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	21.8		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	120		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN0128WBSD01			SDG No.:	Q1199
Lab Sample ID:	VN0128WBSD01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085530.D	1		01/28/25 11:48	VN012825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	21.9		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	21.0		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.6		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.6		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	43.4		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	20.8		0.14	0.50	1.00	ug/L
100-42-5	Styrene	21.3		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	22.5		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	21.2		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.8		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	21.0		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.4		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.4		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.3		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	50.2		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	51.1		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.0		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	182000	8.224				
540-36-3	1,4-Difluorobenzene	310000	9.1				
3114-55-4	Chlorobenzene-d5	273000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	128000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085530.D
 Acq On : 28 Jan 2025 11:48
 Operator : JC\MD
 Sample : VN0128WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/29/2025
 Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:36:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	182154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	309854	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127977	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	147838	50.279	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.560%
35) Dibromofluoromethane	8.165	113	107958	50.222	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.440%
50) Toluene-d8	10.565	98	390099	51.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.160%
62) 4-Bromofluorobenzene	12.847	95	133187	50.978	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.960%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	43993	17.838	ug/l	98
3) Chloromethane	2.359	50	47315	17.719	ug/l	99
4) Vinyl Chloride	2.512	62	47895	17.844	ug/l	98
5) Bromomethane	2.959	94	29197	18.008	ug/l	93
6) Chloroethane	3.118	64	33369	19.610	ug/l	98
7) Trichlorofluoromethane	3.501	101	75960	19.504	ug/l	99
8) Diethyl Ether	3.965	74	26094	19.395	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.371	101	44549	20.309	ug/l	98
10) Methyl Iodide	4.583	142	50174	19.974	ug/l	99
11) Tert butyl alcohol	5.518	59	39612	117.652	ug/l	98
12) 1,1-Dichloroethene	4.342	96	38078	19.478	ug/l	96
13) Acrolein	4.183	56	40971	89.143	ug/l	96
14) Allyl chloride	5.024	41	61288	19.320	ug/l	98
15) Acrylonitrile	5.718	53	117305	109.689	ug/l	97
16) Acetone	4.424	43	102700	108.100	ug/l	95
17) Carbon Disulfide	4.712	76	97539	16.205	ug/l	99
18) Methyl Acetate	5.024	43	67889	23.500	ug/l	100
19) Methyl tert-butyl Ether	5.795	73	139282	21.942	ug/l	100
20) Methylene Chloride	5.277	84	46285	19.680	ug/l	93
21) trans-1,2-Dichloroethene	5.783	96	39520	18.916	ug/l	94
22) Diisopropyl ether	6.665	45	150475	21.371	ug/l	98
23) Vinyl Acetate	6.600	43	520716	105.676	ug/l	100
24) 1,1-Dichloroethane	6.565	63	87788	20.448	ug/l	99
25) 2-Butanone	7.477	43	154236	110.320	ug/l	96
26) 2,2-Dichloropropane	7.489	77	75260	21.682	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	49722	20.206	ug/l	98
28) Bromochloromethane	7.812	49	33640	16.842	ug/l	97
29) Tetrahydrofuran	7.841	42	101879	114.937	ug/l	100
30) Chloroform	7.965	83	91378	20.593	ug/l	99
31) Cyclohexane	8.253	56	65567	17.571	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	78366	20.134	ug/l	98
36) 1,1-Dichloropropene	8.371	75	57513	19.063	ug/l	99
37) Ethyl Acetate	7.553	43	64941	21.326	ug/l	99
38) Carbon Tetrachloride	8.359	117	69210	20.038	ug/l	96
39) Methylcyclohexane	9.600	83	52898	18.659	ug/l	99
40) Benzene	8.606	78	183341	20.225	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085530.D
 Acq On : 28 Jan 2025 11:48
 Operator : JC\MD
 Sample : VN0128WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/29/2025
 Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:36:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	33920	21.410	ug/l	98
42) 1,2-Dichloroethane	8.665	62	73450	21.514	ug/l	100
43) Isopropyl Acetate	8.683	43	105226	21.565	ug/l	100
44) Trichloroethene	9.353	130	40907	19.386	ug/l	99
45) 1,2-Dichloropropane	9.618	63	47937	20.703	ug/l	98
46) Dibromomethane	9.706	93	35054	20.981	ug/l	98
47) Bromodichloromethane	9.882	83	73992	21.738	ug/l	95
48) Methyl methacrylate	9.677	41	48128	21.918	ug/l	97
49) 1,4-Dioxane	9.688	88	17249	466.467	ug/l #	80
51) 4-Methyl-2-Pentanone	10.441	43	334158	118.016	ug/l	100
52) Toluene	10.624	92	113049	21.523	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	69056	21.468	ug/l	95
54) cis-1,3-Dichloropropene	10.312	75	74974	21.821	ug/l	95
55) 1,1,2-Trichloroethane	11.012	97	44789	21.547	ug/l	92
56) Ethyl methacrylate	10.871	69	65671	19.962	ug/l	98
57) 1,3-Dichloropropane	11.159	76	78972	21.850	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	129996	98.550	ug/l	100
59) 2-Hexanone	11.194	43	235784	118.347	ug/l	99
60) Dibromochloromethane	11.359	129	54955	21.899	ug/l	99
61) 1,2-Dibromoethane	11.465	107	43841	21.189	ug/l	98
64) Tetrachloroethene	11.100	164	39045	20.962	ug/l	97
65) Chlorobenzene	11.888	112	123448	20.625	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	47119	21.449	ug/l	97
67) Ethyl Benzene	11.959	91	200961	20.614	ug/l	99
68) m/p-Xylenes	12.071	106	156444	43.421	ug/l	98
69) o-Xylene	12.394	106	71566	20.785	ug/l	98
70) Styrene	12.412	104	121646	21.347	ug/l	98
71) Bromoform	12.576	173	35363	22.498	ug/l #	99
73) Isopropylbenzene	12.694	105	183110	21.200	ug/l	100
74) N-amyl acetate	12.494	43	84761	21.833	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	66402	21.763	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	61286m	23.561	ug/l	
77) Bromobenzene	12.976	156	47394	21.002	ug/l	100
78) n-propylbenzene	13.035	91	222149	21.726	ug/l	99
79) 2-Chlorotoluene	13.123	91	143547	21.691	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	155679	21.820	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	21365	22.223	ug/l	97
82) 4-Chlorotoluene	13.218	91	141708	21.506	ug/l	99
83) tert-Butylbenzene	13.435	119	126399	21.097	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	158115	22.236	ug/l	99
85) sec-Butylbenzene	13.612	105	176846	21.297	ug/l	100
86) p-Isopropyltoluene	13.729	119	144432	20.822	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	87193	20.981	ug/l	97
88) 1,4-Dichlorobenzene	13.812	146	87963	20.420	ug/l	98
89) n-Butylbenzene	14.053	91	121076	20.324	ug/l	99
90) Hexachloroethane	14.329	117	31810	20.123	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	84554	20.397	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12008	21.553	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	37673	19.554	ug/l	96
94) Hexachlorobutadiene	15.500	225	20093	19.644	ug/l	98
95) Naphthalene	15.641	128	110672	19.250	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	39684	20.356	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085530.D
Acq On : 28 Jan 2025 11:48
Operator : JC\MD
Sample : VN0128WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025

Quant Time: Jan 29 00:36:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085530.D
Acq On : 28 Jan 2025 11:48
Operator : JC\MD
Sample : VN0128WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Manual Integrations

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025


Quant Time: Jan 29 00:36:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01



Tgt Ion: 168 Resp: 182154
Ion Ratio Lower Upper
168 100
99 63.5 53.6 80.4




Manual Integrations
APPROVED

Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025




#2
Dichlorodifluoromethane
Concen: 17.838 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 85 Resp: 43993
Ion Ratio Lower Upper
85 100
87 32.3 16.7 50.1



Abundance Scan 69 (2.359 min): VN085439.D\data.ms (-62)



#3

Chloromethane

Concen: 17.719 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 50 Resp: 47315

Ion Ratio Lower Upper

50 100

52 33.4 27.0 40.6


Manual Integrations

APPROVED



Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025


Abundance Scan 69 (2.359 min): VN085530.D\data.ms



Abundance Scan 69 (2.359 min): VN085530.D\data.ms (-18)



Abundance Scan 95 (2.512 min): VN085439.D\data.ms (-87)



#4

Vinyl Chloride

Concen: 17.844 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48


Tgt Ion: 62 Resp: 47895

Ion Ratio Lower Upper



62 100


64 32.3 24.8 37.2

Abundance Scan 95 (2.512 min): VN085530.D\data.ms




Abundance Scan 95 (2.512 min): VN085530.D\data.ms (-44)





#5
Bromomethane
Concen: 18.008 ug/l
RT: 2.959 min Scan# 171
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01



Tgt Ion: 94 Resp: 29197
Ion Ratio Lower Upper
94 100
96 95.9 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#6
Chloroethane
Concen: 19.610 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 64 Resp: 33369
Ion Ratio Lower Upper
64 100
66 29.4 24.6 36.8





#7

Trichlorofluoromethane

Concen: 19.504 ug/l

RT: 3.501 min Scan# 263

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion:101 Resp: 75960

Ion Ratio Lower Upper

101 100





103 65.4 51.4 77.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#8

Diethyl Ether

Concen: 19.395 ug/l

RT: 3.965 min Scan# 342

Delta R.T. 0.012 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 74 Resp: 26094

Ion Ratio Lower Upper

74 100

45 100.4 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.309 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Tgt Ion:101 Resp: 44549
Ion Ratio Lower Upper
101 100
85 45.2 37.8 56.8
151 72.2 58.8 88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#10
Methyl Iodide
Concen: 19.974 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:142 Resp: 50174
Ion Ratio Lower Upper
142 100
127 50.2 40.3 60.5
141 13.9 12.8 19.2





#11

Tert butyl alcohol

Concen: 117.652 ug/l

RT: 5.518 min Scan# 605

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 59 Resp: 39612

Ion Ratio Lower Upper

59 100





57 10.2 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#12

1,1-Dichloroethene

Concen: 19.478 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 96 Resp: 38078


Ion Ratio Lower Upper

96 100

61 180.1 141.0 211.4

98 65.7 49.0 73.6





#13

Acrolein

Concen: 89.143 ug/l

RT: 4.183 min Scan# 377

Delta R.T. 0.012 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 56 Resp: 40971

Ion Ratio Lower Upper

56 100





55 73.7 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#14

Allyl chloride

Concen: 19.320 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 41 Resp: 61288


Ion Ratio Lower Upper

41 100

39 78.5 64.4 96.6

76 36.5 30.5 45.7





#15

Acrylonitrile

Concen: 109.689 ug/l

RT: 5.718 min Scan# 64

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 53 Resp: 117305

Ion Ratio Lower Upper

53 100

52 84.6 65.5 98.3





51 38.2 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#16

Acetone

Concen: 108.100 ug/l

RT: 4.424 min Scan# 420

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 43 Resp: 102700

Ion Ratio Lower Upper

43 100

58 26.8 23.8 35.6





#17

Carbon Disulfide

Concen: 16.205 ug/l

RT: 4.712 min Scan# 46

Delta R.T. 0.006 min

Lab File: VN085530.D


Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01



Tgt Ion: 76 Resp: 97539

Ion Ratio Lower Upper

76 100




78 8.1 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#18

Methyl Acetate

Concen: 23.500 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 43 Resp: 67889

Ion Ratio Lower Upper

43 100

74 23.5 18.6 28.0





#19

Methyl tert-butyl Ether

Concen: 21.942 ug/l

RT: 5.795 min Scan# 65

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 73 Resp: 139282

Ion Ratio Lower Upper

73 100





57 22.1 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#20

Methylene Chloride

Concen: 19.680 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Tgt Ion: 84 Resp: 46285




Ion Ratio Lower Upper


84 100

49 120.0 104.1 156.1

51 43.7 31.0 46.4

86 67.5 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 18.916 ug/l

RT: 5.783 min Scan# 651

Delta R.T. 0.000 min

Lab File: VN085530.D


Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01



Tgt Ion: 96 Resp: 39520

Ion Ratio Lower Upper

96 100

61 153.6 115.7 173.5




98 58.7 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#22

Diisopropyl ether


Concen: 21.371 ug/l

RT: 6.665 min Scan# 801

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48



Tgt Ion: 45 Resp: 150475



Ion Ratio Lower Upper


45 100

43 57.5 45.0 67.6

87 25.7 21.8 32.8

59 11.4 10.2 15.2





#23

Vinyl Acetate

Concen: 105.676 ug/l

RT: 6.600 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 43 Resp: 520716

Ion Ratio Lower Upper

43 100





86 10.6 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#24

1,1-Dichloroethane

Concen: 20.448 ug/l

RT: 6.565 min Scan# 784

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 63 Resp: 87788


Ion Ratio Lower Upper

63 100

98 7.0 3.3 9.8

100 4.1 2.0 6.0





#25

2-Butanone

Concen: 110.320 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 43 Resp: 154236

Ion Ratio Lower Upper

43 100





72 27.2 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#26

2,2-Dichloropropane

Concen: 21.682 ug/l

RT: 7.489 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 77 Resp: 75260

Ion Ratio Lower Upper

77 100


97 21.8 10.7 32.1





#27
cis-1,2-Dichloroethene
Concen: 20.206 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01



Tgt Ion: 96 Resp: 49722
Ion Ratio Lower Upper
96 100
61 158.8 0.0 311.8
98 64.8 0.0 126.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#28
Bromochloromethane
Concen: 16.842 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 49 Resp: 33640
Ion Ratio Lower Upper
49 100
129 1.3 0.0 4.4
130 71.1 55.0 82.4





#29

Tetrahydrofuran

Concen: 114.937 ug/l

RT: 7.841 min Scan# 101

Delta R.T. 0.012 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 42 Resp: 101879

Ion Ratio Lower Upper

42 100

72 43.2 34.2 51.2





71 40.7 32.5 48.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#30

Chloroform

Concen: 20.593 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 83 Resp: 91378

Ion Ratio Lower Upper

83 100

85 65.4 51.8 77.6





#31

Cyclohexane

Concen: 17.571 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 56 Resp: 65567

Ion Ratio Lower Upper

56 100

69 29.3 22.2 33.4





84 84.2 66.4 99.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#32

1,1,1-Trichloroethane

Concen: 20.134 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 97 Resp: 78366


Ion Ratio Lower Upper

97 100

99 63.6 49.8 74.6


61 50.8 41.4 62.2





#33
1,2-Dichloroethane-d4
Concen: 50.279 ug/l
RT: 8.577 min Scan# 1125
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01




Tgt Ion: 65 Resp: 147838
Ion Ratio Lower Upper
65 100
67 51.5 0.0 101.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Tgt Ion: 114 Resp: 309854
Ion Ratio Lower Upper
114 100
63 22.7 0.0 47.6
88 16.2 0.0 32.6





#35
Dibromofluoromethane
Concen: 50.222 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01



Tgt Ion: 113 Resp: 107958
Ion Ratio Lower Upper
113 100
111 101.5 82.7 124.1
192 17.7 14.3 21.5




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025



#36
1,1-Dichloropropene
Concen: 19.063 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 75 Resp: 57513
Ion Ratio Lower Upper
75 100
110 33.1 16.5 49.5
77 29.9 24.4 36.6





#37

Ethyl Acetate

Concen: 21.326 ug/l

RT: 7.553 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 43 Resp: 64941

Ion Ratio Lower Upper

43 100

61 13.2 10.9 16.3





70 10.2 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#38

Carbon Tetrachloride

Concen: 20.038 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion:117 Resp: 69210


Ion Ratio Lower Upper

117 100

119 98.5 75.4 113.2

121 29.3 24.6 37.0





#39

Methylcyclohexane

Concen: 18.659 ug/l

RT: 9.600 min Scan# 1311

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N





ClientSampleId :

VN0128WBSD01

Tgt Ion:	83	Resp:	52898
Ion	Ratio	Lower	Upper
83	100		
55	79.0	62.6	94.0
98	48.2	37.7	56.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#40

Benzene

Concen: 20.225 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Tgt Ion:	78	Resp:	183341
Ion	Ratio	Lower	Upper
78	100		
77	23.3	19.0	28.6





#41

Methacrylonitrile

Concen: 21.410 ug/l

RT: 7.777 min Scan# 99

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 41 Resp: 33920

Ion Ratio Lower Upper

41 100

39 55.6 46.0 69.0

67 73.8 57.4 86.2





52 32.7 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#42

1,2-Dichloroethane

Concen: 21.514 ug/l

RT: 8.665 min Scan# 1141

Delta R.T. 0.000 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 62 Resp: 73450

Ion Ratio Lower Upper

62 100

98 8.4 0.0 17.0





#43

Isopropyl Acetate

Concen: 21.565 ug/l

RT: 8.683 min Scan# 1144

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 43 Resp: 105226

Ion Ratio Lower Upper

43 100

61 26.0 20.7 31.1





87 12.7 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#44

Trichloroethene

Concen: 19.386 ug/l

RT: 9.353 min Scan# 1258

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion:130 Resp: 40907

Ion Ratio Lower Upper

130 100

95 97.3 0.0 195.8





#45

1,2-Dichloropropane

Concen: 20.703 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 63 Resp: 47937

Ion Ratio Lower Upper

63 100





65 30.8 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#46

Dibromomethane

Concen: 20.981 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 93 Resp: 35054


Ion Ratio Lower Upper

93 100

95 82.8 64.7 97.1

174 88.1 69.0 103.4





#47

Bromodichloromethane

Concen: 21.738 ug/l

RT: 9.882 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 83 Resp: 73992

Ion Ratio Lower Upper

83 100

85 60.1 51.2 76.8





127 8.5 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#48

Methyl methacrylate

Concen: 21.918 ug/l

RT: 9.677 min Scan# 1313

Delta R.T. -0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 41 Resp: 48128


Ion Ratio Lower Upper

41 100

69 81.9 64.7 97.1

39 65.4 49.0 73.6





#49

1,4-Dioxane

Concen: 466.467 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 88 Resp: 17249

Ion Ratio Lower Upper

88 100

43 0.0 26.6 39.8





58 72.4 59.5 89.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#50

Toluene-d8

Concen: 51.076 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 98 Resp: 390099

Ion Ratio Lower Upper

98 100

100 65.0 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 118.016 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Tgt Ion: 43 Resp: 334158
Ion Ratio Lower Upper
43 100
58 37.9 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#52
Toluene
Concen: 21.523 ug/l
RT: 10.624 min Scan# 1474
Delta R.T. -0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 92 Resp: 113049
Ion Ratio Lower Upper
92 100
91 170.5 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 21.468 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 75 Resp: 69056

Ion Ratio Lower Upper

75 100





77 32.1 23.5 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#54

cis-1,3-Dichloropropene

Concen: 21.821 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 75 Resp: 74974


Ion Ratio Lower Upper

75 100

77 31.2 25.0 37.4

39 48.6 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 21.547 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 97 Resp: 44789

Ion Ratio Lower Upper

97 100

83 92.0 69.0 103.6

85 61.6 44.6 66.8





99 68.5 48.6 72.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#56

Ethyl methacrylate

Concen: 19.962 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 69 Resp: 65671


Ion Ratio Lower Upper

69 100

41 70.1 54.6 82.0

39 41.9 32.4 48.6





#57

1,3-Dichloropropane

Concen: 21.850 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 76 Resp: 78972

Ion Ratio Lower Upper

76 100





78 31.8 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#58

2-Chloroethyl Vinyl ether

Concen: 98.550 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 63 Resp: 129996

Ion Ratio Lower Upper

63 100

106 27.2 21.6 32.4





#59

2-Hexanone

Concen: 118.347 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 43 Resp: 235784

Ion Ratio Lower Upper

43 100





58 52.0 26.2 78.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#60

Dibromochloromethane

Concen: 21.899 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion:129 Resp: 54955

Ion Ratio Lower Upper

129 100

127 76.6 38.6 115.8









#61
1,2-Dibromoethane
Concen: 21.189 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Tgt Ion: 107 Resp: 43841
Ion Ratio Lower Upper
107 100
109 93.3 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/29/2025
Supervised By : Semsettin Yesilyurt 01/29/2025



#62
4-Bromofluorobenzene
Concen: 50.978 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 95 Resp: 133187
Ion Ratio Lower Upper
95 100
174 75.7 0.0 145.0
176 70.9 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Tgt Ion:117 Resp: 273221
Ion Ratio Lower Upper
117 100
82 58.5 48.6 72.8
119 32.2 26.6 39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#64
Tetrachloroethene
Concen: 20.962 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:164 Resp: 39045
Ion Ratio Lower Upper
164 100
166 131.5 103.4 155.2
129 101.5 79.2 118.8
131 101.9 77.1 115.7





#65

Chlorobenzene

Concen: 20.625 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion:112 Resp: 123448

Ion Ratio Lower Upper

112 100





114 31.9 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 21.449 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion:131 Resp: 47119


Ion Ratio Lower Upper

131 100

133 92.4 47.4 142.3

119 63.4 33.1 99.5





#67

Ethyl Benzene

Concen: 20.614 ug/l

RT: 11.959 min Scan# 17

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 91 Resp: 200961

Ion Ratio Lower Upper

91 100





106 30.6 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#68

m/p-Xylenes

Concen: 43.421 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion:106 Resp: 156444

Ion Ratio Lower Upper

106 100

91 206.6 167.7 251.5





#69

o-Xylene

Concen: 20.785 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion:106 Resp: 71566

Ion Ratio Lower Upper

106 100





91 224.0 110.4 331.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#70

Styrene

Concen: 21.347 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion:104 Resp: 121646


Ion Ratio Lower Upper

104 100

78 54.5 42.5 63.7

103 56.3 43.8 65.8





#71

Bromoform

Concen: 22.498 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion:173 Resp: 35363

Ion Ratio Lower Upper

173 100

175 48.3 24.4 73.2





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion:152 Resp: 127977


Ion Ratio Lower Upper

152 100

115 62.9 31.1 93.3

150 165.8 0.0 343.6





#73

Isopropylbenzene

Concen: 21.200 ug/l

RT: 12.694 min Scan# 18

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 105 Resp: 183110

Ion Ratio Lower Upper

105 100





120 25.6 12.8 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#74

N-ethyl acetate

Concen: 21.833 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Tgt Ion: 43 Resp: 84761




Ion Ratio Lower Upper


43 100

70 42.4 34.0 51.0

55 27.8 21.4 32.2

61 23.0 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 21.763 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBSD01

Tgt Ion:	83	Resp:	66402
Ion	Ratio	Lower	Upper
83	100		
131	9.9	4.8	14.4
85	64.2	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#76

1,2,3-Trichloropropane

Concen: 23.561 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48


Tgt Ion:	75	Resp:	61286
Ion	Ratio	Lower	Upper
75	100		
77	176.6	109.7	329.2





#77
Bromobenzene
Concen: 21.002 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01



Tgt Ion: 156 Resp: 47394
Ion Ratio Lower Upper
156 100
77 228.3 114.1 342.4
158 96.4 48.9 146.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#78
n-propylbenzene
Concen: 21.726 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 91 Resp: 222149
Ion Ratio Lower Upper
91 100
120 21.0 10.9 32.6





#79

2-Chlorotoluene

Concen: 21.691 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 91 Resp: 143547

Ion Ratio Lower Upper

91 100





126 30.9 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#80

1,3,5-Trimethylbenzene

Concen: 21.820 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion:105 Resp: 155679

Ion Ratio Lower Upper

105 100

120 47.5 23.9 71.7





#81

trans-1,4-Dichloro-2-butene

Concen: 22.223 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. 0.000 min

Lab File: VN085530.D


Acq: 28 Jan 2025 11:48

Instrument :




MSVOA_N

ClientSampleId :

VN0128WBSD01



Tgt Ion:	75	Resp:	21365
Ion Ratio	Lower	Upper	
75	100		
53	115.4	95.6	143.4
89	44.8	37.0	55.6

Manual Integrations
APPROVEDReviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025

#82

4-Chlorotoluene


Concen: 21.506 ug/l

RT: 13.218 min Scan# 1915



Delta R.T. 0.000 min


Lab File: VN085530.D

Acq: 28 Jan 2025 11:48



Tgt Ion:	91	Resp:	141708
Ion Ratio	Lower	Upper	
91	100		
126	32.2	15.9	47.7





#83

tert-Butylbenzene

Concen: 21.097 ug/l

RT: 13.435 min Scan# 19

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion:119 Resp: 126399

Ion Ratio Lower Upper

119 100

91 70.5 35.1 105.3





134 27.3 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#84

1,2,4-Trimethylbenzene

Concen: 22.236 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion:105 Resp: 158115

Ion Ratio Lower Upper

105 100

120 43.9 21.6 65.0





#85

sec-Butylbenzene

Concen: 21.297 ug/l

RT: 13.612 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion:105 Resp: 176846

Ion Ratio Lower Upper

105 100





134 19.3 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#86

p-Isopropyltoluene

Concen: 20.822 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion:119 Resp: 144432


Ion Ratio Lower Upper

119 100

134 26.1 12.7 38.0

91 25.6 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 20.981 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01





Tgt Ion:	146	Resp:	87193
Ion	Ratio	Lower	Upper
146	100		
111	41.4	21.4	64.3
148	62.2	32.3	96.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#88

1,4-Dichlorobenzene

Concen: 20.420 ug/l




RT: 13.812 min Scan# 2016


Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Tgt Ion:	146	Resp:	87963
Ion	Ratio	Lower	Upper
146	100		
111	41.6	21.3	63.7
148	63.2	32.4	97.0





#89

n-Butylbenzene

Concen: 20.324 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion: 91 Resp: 121076

Ion Ratio Lower Upper

91 100

92 50.3 25.8 77.3





134 23.8 11.7 35.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#90

Hexachloroethane

Concen: 20.123 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion:117 Resp: 31810

Ion Ratio Lower Upper

117 100

201 66.3 33.7 101.0









#91
1,2-Dichlorobenzene
Concen: 20.397 ug/l
RT: 14.106 min Scan# 2065
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Tgt Ion	Ratio	Lower	Upper
146	100		
111	44.6	21.7	65.1
148	63.0	31.4	94.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/29/2025
Supervised By :Semsettin Yesilyurt 01/29/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.553 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion	Ratio	Lower	Upper
75	100		
155	73.1	36.4	109.2
157	95.1	45.4	136.1





#93

1,2,4-Trichlorobenzene

Concen: 19.554 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion:180 Resp: 37673

Ion Ratio Lower Upper

180 100

182 96.9 46.8 140.3





145 35.9 16.0 48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#94

Hexachlorobutadiene

Concen: 19.644 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion:225 Resp: 20093


Ion Ratio Lower Upper

225 100

223 63.7 30.7 92.1

227 60.8 30.9 92.5





#95

Naphthalene

Concen: 19.250 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01

Tgt Ion:128 Resp: 110672

Ion Ratio Lower Upper

128 100

127 13.4 10.6 16.0





129 11.0 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/29/2025

Supervised By :Semsettin Yesilyurt 01/29/2025



#96

1,2,3-Trichlorobenzene

Concen: 20.356 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion:180 Resp: 39684

Ion Ratio Lower Upper

180 100

182 93.4 47.4 142.2

145 33.7 16.9 50.7



Manual Integration Report

Sequence:	VN011425	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC100	VN085438.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:09 AM	MMDadoda	1/15/2025 12:55:19 PM	Peak Integrated by Software
VSTDICCC050	VN085439.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:14 AM	MMDadoda	1/15/2025 12:55:22 PM	Peak Integrated by Software
VSTDICC020	VN085440.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:19 AM	MMDadoda	1/15/2025 12:55:25 PM	Peak Integrated by Software
VSTDICC020	VN085440.D	trans-1,4-Dichloro-2-but ene	JOHN	1/15/2025 9:31:19 AM	MMDadoda	1/15/2025 12:55:25 PM	Peak Integrated by Software
VSTDICC020	VN085440.D	Vinyl Acetate	JOHN	1/15/2025 9:31:19 AM	MMDadoda	1/15/2025 12:55:25 PM	Peak Integrated by Software
VSTDICC010	VN085441.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:23 AM	MMDadoda	1/15/2025 12:55:29 PM	Peak Integrated by Software
VSTDICC005	VN085442.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:28 AM	MMDadoda	1/15/2025 12:55:34 PM	Peak Integrated by Software
VSTDICC005	VN085442.D	Ethyl Acetate	JOHN	1/15/2025 9:31:28 AM	MMDadoda	1/15/2025 12:55:34 PM	Peak Integrated by Software
VSTDICC005	VN085442.D	Vinyl Acetate	JOHN	1/15/2025 9:31:28 AM	MMDadoda	1/15/2025 12:55:34 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	1,1,2-Trichlorotrifluoroet hane	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	1,1-Dichloroethane	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	1,4-Dichlorobenzene	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software

Manual Integration Report

Sequence:

VN011425

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VN085443.D	2-Butanone	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	Ethyl Acetate	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	Vinyl Acetate	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICV050	VN085445.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:37 AM	MMDadoda	1/15/2025 12:55:43 PM	Peak Integrated by Software
VSTDICV050	VN085445.D	trans-1,4-Dichloro-2-butene	JOHN	1/15/2025 9:31:37 AM	MMDadoda	1/15/2025 12:55:43 PM	Peak Integrated by Software

Manual Integration Report

Sequence:

VN012825

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN085526.D	1,2,3-Trichloropropane	JOHN	1/29/2025 9:37:19 AM	SAM	1/29/2025 12:55:25 PM	Peak Integrated by Software
VN0128WBS01	VN085529.D	1,2,3-Trichloropropane	JOHN	1/29/2025 9:37:23 AM	SAM	1/29/2025 12:55:23 PM	Peak Integrated by Software
VN0128WBSD01	VN085530.D	1,2,3-Trichloropropane	JOHN	1/29/2025 9:37:28 AM	SAM	1/29/2025 12:55:21 PM	Peak Integrated by Software
VSTDCCC050	VN085546.D	1,2,3-Trichloropropane	JOHN	1/29/2025 9:37:40 AM	SAM	1/29/2025 12:55:20 PM	Peak Integrated by Software

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN011425

Review By	John Carlone	Review On	1/15/2025 9:31:51 AM		
Supervise By	Maresh Dadoda	Supervise On	1/15/2025 12:55:51 PM		
SubDirectory	VN011425	HP Acquire Method	MSVOA_N	HP Processing Method	82N011425W.M
STD. NAME		STD REF.#			
Tune/Reschk		VP132529			
Initial Calibration Stds		VP132530,VP132531,VP132532,VP132533,VP132534,VP132535			
CCC					
Internal Standard/PEM					
ICV/I.BLK		VP132544			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN085437.D	14 Jan 2025 14:22	JC\MD	Ok
2	VSTDICC100	VN085438.D	14 Jan 2025 14:56	JC\MD	Ok,M
3	VSTDICCC050	VN085439.D	14 Jan 2025 15:19	JC\MD	Ok,M
4	VSTDICC020	VN085440.D	14 Jan 2025 15:43	JC\MD	Ok,M
5	VSTDICC010	VN085441.D	14 Jan 2025 16:07	JC\MD	Ok,M
6	VSTDICC005	VN085442.D	14 Jan 2025 16:31	JC\MD	Ok,M
7	VSTDICC001	VN085443.D	14 Jan 2025 17:19	JC\MD	Ok,M
8	IBLK	VN085444.D	14 Jan 2025 17:42	JC\MD	Ok
9	VSTDICV050	VN085445.D	14 Jan 2025 18:06	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN012825

Review By	John Carlone	Review On	1/29/2025 9:40:00 AM
Supervise By	Semsettin Yesilyurt	Supervise On	1/29/2025 12:55:29 PM
SubDirectory	VN012825	HP Acquire Method	HP Processing Method 82N011425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132723		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132724,VP132725		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN085525.D	28 Jan 2025 09:17	JC\MD	Ok
2	VSTDCCC050	VN085526.D	28 Jan 2025 09:53	JC\MD	Ok,M
3	VN0128MBL01	VN085527.D	28 Jan 2025 10:27	JC\MD	Ok
4	VN0128WBL01	VN085528.D	28 Jan 2025 10:51	JC\MD	Ok
5	VN0128WBS01	VN085529.D	28 Jan 2025 11:15	JC\MD	Ok,M
6	VN0128WBSD01	VN085530.D	28 Jan 2025 11:48	JC\MD	Ok,M
7	IBLK	VN085531.D	28 Jan 2025 12:12	JC\MD	Ok
8	Q1180-04	VN085532.D	28 Jan 2025 12:36	JC\MD	Ok
9	Q1194-09	VN085533.D	28 Jan 2025 13:00	JC\MD	Ok
10	Q1199-01	VN085534.D	28 Jan 2025 13:24	JC\MD	Ok
11	Q1199-02	VN085535.D	28 Jan 2025 13:49	JC\MD	Ok
12	Q1199-03	VN085536.D	28 Jan 2025 14:13	JC\MD	Ok
13	Q1199-04	VN085537.D	28 Jan 2025 14:37	JC\MD	Ok
14	Q1199-05	VN085538.D	28 Jan 2025 15:01	JC\MD	Ok
15	Q1199-06	VN085539.D	28 Jan 2025 15:25	JC\MD	ReRun
16	Q1199-07	VN085540.D	28 Jan 2025 15:49	JC\MD	Ok
17	PB166233ZHE#11	VN085541.D	28 Jan 2025 16:13	JC\MD	Ok
18	PB166233ZHE#12	VN085542.D	28 Jan 2025 16:37	JC\MD	Ok
19	PB166233ZHE#13	VN085543.D	28 Jan 2025 17:01	JC\MD	Ok
20	IBLK	VN085544.D	28 Jan 2025 17:25	JC\MD	Ok
21	Q1199-06	VN085545.D	28 Jan 2025 17:50	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN012825

Review By	John Carlone	Review On	1/29/2025 9:40:00 AM
Supervise By	Semsettin Yesilyurt	Supervise On	1/29/2025 12:55:29 PM
SubDirectory	VN012825	HP Acquire Method	HP Processing Method 82N011425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132723		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132724,VP132725		

22	VSTDCCC050	VN085546.D	28 Jan 2025 18:14	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN011425

Review By	John Carlone	Review On	1/15/2025 9:31:51 AM
Supervise By	Maresh Dadoda	Supervise On	1/15/2025 12:55:51 PM
SubDirectory	VN011425	HP Acquire Method	MSVOA_N HP Processing Method 82N011425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132529 VP132530,VP132531,VP132532,VP132533,VP132534,VP132535 VP132544

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085437.D	14 Jan 2025 14:22		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN085438.D	14 Jan 2025 14:56		JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN085439.D	14 Jan 2025 15:19	Comp.#56 is on Linear Regression	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN085440.D	14 Jan 2025 15:43	Comp.#86 is on Quadratic Regression	JC\MD	Ok,M
5	VSTDICC010	VSTDICC010	VN085441.D	14 Jan 2025 16:07		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN085442.D	14 Jan 2025 16:31		JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN085443.D	14 Jan 2025 17:19		JC\MD	Ok,M
8	IBLK	IBLK	VN085444.D	14 Jan 2025 17:42		JC\MD	Ok
9	VSTDICV050	ICVVN011425	VN085445.D	14 Jan 2025 18:06		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN012825

Review By	John Carlone	Review On	1/29/2025 9:40:00 AM
Supervise By	Semsettin Yesilyurt	Supervise On	1/29/2025 12:55:29 PM
SubDirectory	VN012825	HP Acquire Method	HP Processing Method 82N011425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132723
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132724,VP132725

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085525.D	28 Jan 2025 09:17		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN085526.D	28 Jan 2025 09:53	V13516	JC\MD	Ok,M
3	VN0128MBL01	VN0128MBL01	VN085527.D	28 Jan 2025 10:27		JC\MD	Ok
4	VN0128WBL01	VN0128WBL01	VN085528.D	28 Jan 2025 10:51		JC\MD	Ok
5	VN0128WBS01	VN0128WBS01	VN085529.D	28 Jan 2025 11:15		JC\MD	Ok,M
6	VN0128WBSD01	VN0128WBSD01	VN085530.D	28 Jan 2025 11:48		JC\MD	Ok,M
7	IBLK	IBLK	VN085531.D	28 Jan 2025 12:12		JC\MD	Ok
8	Q1180-04	Storage-Blank-SAMPLE	VN085532.D	28 Jan 2025 12:36	vial B pH<2	JC\MD	Ok
9	Q1194-09	FB	VN085533.D	28 Jan 2025 13:00	vial B pH<2 FB	JC\MD	Ok
10	Q1199-01	BP-VPB-192-TB-20250	VN085534.D	28 Jan 2025 13:24	vial A pH<2 TB	JC\MD	Ok
11	Q1199-02	BP-VPB-192-EB-20250	VN085535.D	28 Jan 2025 13:49	vial A pH<2 EB	JC\MD	Ok
12	Q1199-03	BP-VPB-192-GW-280-2	VN085536.D	28 Jan 2025 14:13	vial A pH<2	JC\MD	Ok
13	Q1199-04	BP-VPB-192-GW-260-2	VN085537.D	28 Jan 2025 14:37	vial A pH<2	JC\MD	Ok
14	Q1199-05	BP-VPB-192-GW-240-2	VN085538.D	28 Jan 2025 15:01	vial A pH<2	JC\MD	Ok
15	Q1199-06	BP-VPB-192-GW-220-2	VN085539.D	28 Jan 2025 15:25	vial A pH<2 Surrogate Fail	JC\MD	ReRun
16	Q1199-07	BP-VPB-192-DUP-2025	VN085540.D	28 Jan 2025 15:49	vial A pH<2	JC\MD	Ok
17	PB166233ZHE#11	PB166233ZHE#11	VN085541.D	28 Jan 2025 16:13		JC\MD	Ok
18	PB166233ZHE#12	PB166233ZHE#12	VN085542.D	28 Jan 2025 16:37		JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN012825

Review By	John Carlone	Review On	1/29/2025 9:40:00 AM
Supervise By	Semsettin Yesilyurt	Supervise On	1/29/2025 12:55:29 PM
SubDirectory	VN012825	HP Acquire Method	HP Processing Method 82N011425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132723		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132724,VP132725		

19	PB166233ZHE#13	PB166233ZHE#13	VN085543.D	28 Jan 2025 17:01		JC\MD	Ok
20	IBLK	IBLK	VN085544.D	28 Jan 2025 17:25		JC\MD	Ok
21	Q1199-06	BP-VPB-192-GW-220-2	VN085545.D	28 Jan 2025 17:50	vial B pH<2	JC\MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VN085546.D	28 Jan 2025 18:14		JC\MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q1199

Test : VOCMS Group1

Prepbatch ID :

Sequence ID/Qc Batch ID: VN012825,

Standard ID :

VP130430,VP131746,VP131767,VP132035,VP132096,VP132468,VP132613,VP132723,VP132724,VP132725,

Chemical ID :

V13391,V13446,V13457,V13460,V13465,V13466,V13707,V14145,V14154,V14175,V14176,V14289,V14433,V14439,V14521,V14522,V14614,V14624,V14627,V14630,V14631,V14632,V14633,V14722,V14723,V14724,V14754,V14756,V14801,V14814,V14830,V14831,V14832,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP130430	09/20/2024	02/28/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								09/26/2024

FROM 0.40000ml of V13707 + 24.60000ml of V14145 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP131746	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								11/23/2024

FROM 0.50000ml of V14289 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP131767	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								11/27/2024

FROM 0.50000ml of V13391 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP132035	12/10/2024	06/10/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								12/12/2024

FROM 1.00000ml of V14630 + 1.00000ml of V14631 + 1.00000ml of V14632 + 1.00000ml of V14633 + 46.00000ml of V14614 = Final Quantity: 50.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP132096	12/12/2024	06/10/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 12/19/2024

FROM 1.00000ml of V13465 + 1.00000ml of V13466 + 1.50000ml of V13457 + 1.50000ml of V13460 + 20.00000ml of V14614 = Final
Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP132468	01/08/2025	02/07/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 01/17/2025

FROM 1.00000ml of V14832 + 1.50000ml of V14830 + 1.50000ml of V14831 + 21.00000ml of V14627 = Final Quantity: 25.000 ml



VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP132613	01/20/2025	02/28/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 01/29/2025
<u>FROM</u>	0.40000ml of V13446 + 1.00000ml of V14175 + 1.00000ml of V14176 + 1.00000ml of V14433 + 1.00000ml of V14439 + 1.00000ml of V14521 + 1.00000ml of V14522 + 1.00000ml of V14722 + 1.00000ml of V14754 + 1.00000ml of V14756 + 1.00000ml of V14801 + 1.00000ml of V14814 + 1.50000ml of V14723 + 1.50000ml of V14724 + 10.60000ml of V14624 = Final Quantity: 25.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP132723	01/28/2025	01/29/2025	John Carlone	None	None	Mahesh Dadoda 01/29/2025
<u>FROM</u>	39.98400ml of W3112 + 0.01600ml of VP131767 = Final Quantity: 40.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP132724	01/28/2025	01/29/2025	John Carlone	None	None	Mahesh Dadoda 01/29/2025
<u>FROM</u>	39.94450ml of W3112 + 0.00500ml of VP130430 + 0.00500ml of VP132096 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP132468 + 0.01250ml of VP132613 = Final Quantity: 40.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP132725	01/28/2025	01/29/2025	John Carlone	None	None	Mahesh Dadoda 01/29/2025
<u>FROM</u>	39.94450ml of W3112 + 0.00500ml of VP130430 + 0.00500ml of VP132096 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP132468 + 0.01250ml of VP132613 = Final Quantity: 40.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0181905	02/28/2025	01/10/2025 / SAM	01/23/2023 / SAM	V13446

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13457

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13460

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13465

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13466

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	06/10/2025	06/10/2024 / SAM	04/12/2023 / SAM	V13707

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	02/28/2025	08/29/2024 / SAM	02/06/2024 / SAM	V14145

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	05/18/2025	11/18/2024 / pedro	02/06/2024 / SAM	V14154

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	07/10/2025	01/10/2025 / SAM	02/20/2024 / SAM	V14175

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	07/10/2025	01/10/2025 / SAM	02/20/2024 / SAM	V14176

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	11/22/2025	11/22/2024 / SAM	04/15/2024 / SAM	V14289

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	07/10/2025	01/10/2025 / SAM	08/15/2024 / SAM	V14433

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	07/10/2025	01/10/2025 / SAM	08/15/2024 / SAM	V14439

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	07/10/2025	01/10/2025 / SAM	09/18/2024 / SAM	V14521

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	07/10/2025	01/10/2025 / SAM	09/18/2024 / SAM	V14522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	06/10/2025	12/10/2024 / SAM	11/26/2024 / SAM	V14614

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	07/13/2025	01/13/2025 / SAM	11/26/2024 / SAM	V14624

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	07/06/2025	01/06/2025 / SAM	11/26/2024 / SAM	V14627

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14630

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14631

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14632

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14633

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14722

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14723

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14724

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	05/31/2031	01/10/2025 / SAM	12/17/2024 / SAM	V14754

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14756

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220563	06/30/2026	01/10/2025 / SAM	01/08/2025 / SAM	V14801

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	07/10/2025	01/10/2025 / SAM	01/08/2025 / SAM	V14814

LOTS

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	010725	02/07/2025	01/08/2025 / SAM	01/08/2025 / SAM	V14830

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	010725	02/07/2025	01/08/2025 / SAM	01/08/2025 / SAM	V14831

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	010725	02/07/2025	01/08/2025 / SAM	01/08/2025 / SAM	V14832

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis




Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **010725**
Description: **Acrolein**

Expiration Date: **020725**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **5000**
NIST Test ID#: **6UTB**

Solvent(s): **Water** Lot# **072324Q**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): **10.0**

<i>Anthony Mahoney</i>		010725
Formulated By:	Anthony Mahoney	DATE
<i>Pedro L. Rentas</i>		010725
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossetto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#: 7732-18-5	% (optional)
Water		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible		NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products -	No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC. DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **010725**
Description: **Acrolein**

Expiration Date: **020725**Recommended Storage: **Refrigerate (4 °C)**Nominal Concentration (µg/mL): **5000**NIST Test ID#: **6UTB**

Solvent(s): **Water** Lot# **072324Q**

5E-05 Balance Uncertainty

0.001 Flask Uncertainty



Weight(s) shown below were combined and diluted to (mL): **10.0**

<i>Anthony Mahoney</i>		010725
Formulated By:	Anthony Mahoney	DATE
<i>Pedro L. Rentas</i>		010725
Reviewed By:	Pedro L. Rentas	DATE

5 vial
V 14823 to
V 14829

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossetto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))
Water

CAS#: 7732-18-5

% (optional)
> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible		NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products -	No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **010725**
Description: **Acrolein**

Expiration Date: **020725**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **5000**
NIST Test ID#: **6UTB**

Solvent(s): **Water** Lot# **072324Q**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): **10.0**

<i>Anthony Mahoney</i>		010725
Formulated By:	Anthony Mahoney	DATE
<i>Pedro L. Rentas</i>		010725
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossetto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#: 7732-18-5	% (optional)
Water		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible		NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products -	No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: **95317**
Lot Number: **021624**
Description: **Universal VOA Megamix**
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 8UTB

Solvent(s): **Methanol**
Lot#: **EG359-USQ12**

Weight(s) shown below were combined and diluted to (mL): **100.0** **0.021** Balance Uncertainty
0.021 Flask Uncertainty


Formulated By:	<i>Prashant Chauhan</i>	021624
	Prashant Chauhan	DATE
Reviewed By:	<i>Pedro L. Rentes</i>	021624
	Pedro L. Rentes	DATE

Compound	(RM#)	Lot	Di.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information			
														(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA)	LD50
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc.(µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	(+/-) (µg/mL)				
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg	
2. Allyl chloride (3-Chloropropene)	(0325)	102395	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg	
3. Carbon disulphide	(0060)	MKCR8581	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg	
4. cis-1,4-Dichloro-2-butene	(1198)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1478-11-5	N/A	N/A	
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A	
6. Diethyl ether	(0153)	IK18CA5000K	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A	
7. Ethyl methacrylate	(0381)	06126PK	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	N/A	
8. Iodomethane	(0489)	5HBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 14800mg/kg	
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 70mg/kg	
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 1200mg/kg	
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg	
12. Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg	
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg	
14. 2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg	
15. Perfluorooctane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	N/A	N/A	
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7800mg/m3/8H)	or-rat 43g/kg	
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 916mg/kg	
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 840mg/kg	
19. cis-1,2-Dichloroethane	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A	
20. trans-1,2-Dichloroethane	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg	
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg	
22. 1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.6	22.9	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg	
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg	
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg	
25. Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-66-3	50 ppm (240mg/m3) (CL)	or-rat 900mg/kg	
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-85-3	N/A	or-rat 105mg/kg	
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg	
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A	
29. Trichloroethane	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg	
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg	
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-6	0.001 ppm	or-rat 170mg/kg	
32. 1,2-Dibromomethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg	
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-08-2	50 ppm (8H)	or-rat 870mg/kg	
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-57-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg	
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	un-rat 3600mg/kg	
36. 1,1-Dichloropropane	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	22.9	563-56-6	N/A	N/A	
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A	
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.8	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A	
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	22.9	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg	
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	N/A	or-rat 670mg/kg	
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg	
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (46mg/m3/8H) (skin)	or-rat 836mg/kg	
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-rat 2400mg/kg	
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.6mg/kg	
45. Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4694mg/kg	
46. Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A	
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg	
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	99-87-8	N/A	or-rat 4750mg/kg	
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg	
50. Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg	
51. Styrene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-98-3	200 ppm	or-rat 5000mg/kg	
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	87-61-6	N/A	or-rat 1390mg/kg	
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 750mg/kg	
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	N/A	or-rat 5g/kg	
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-67-8	N/A	or-rat 5000mg/kg	
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg	
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	N/A	N/A	
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	N/A	or-rat 2240mg/kg	
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg	
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	60 ppm (250mg/m3/8H)	or-rat 3900mg/kg	
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg	
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	200											

Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.24
11	cis-1,2-Dichloroethane	18.00
12	Methacrylonitrile/methyl acrylate/Chloroform	18.46
13	Isobutane/1,1,1-Trichloroethane	19.01
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.49
17	Trichloroethane	23.58
18	1,2-Dichloropropane	24.38
19	Methyl methacrylate	24.52
20	Bromochloromethane	25.13
21	Dibromomethane/2-Hitropropane	25.46
22	cis-1,3-Dichloropropane	27.02
23	Toluene	28.05
24	Ethyl methacrylate/trans-1,3-Dichloropropane	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.24
27	Dibromochloromethane	31.16
28	1,2-Dibromomethane	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.66
32	o-Xylene	35.22
33	Styrene	35.39
34	Isopropylbenzene/Bromoforn	36.18
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.93
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.80
42	2-Chlorotoluene	38.82
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.74
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.53
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.46
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection. Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC. DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: **95317**
Lot Number: **021624**
Description: **Universal VOA Megamix**
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 8UTB

Solvent(s): **Methanol**
Lot#: **EG359-USQ12**

Weight(s) shown below were combined and diluted to (mL): **100.0** **0.021** Balance Uncertainty: **5E-05** Flask Uncertainty: **0.021**


		021624
Formulated By:	Prashant Chauhan	DATE
		021624
Reviewed By:	Pedro L. Rentes	DATE

Compound	(RM#)	Lot	Di.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information		
														(Solvent Safety Info. On Attached pg.)		
														CAS#	OSHA PEL (TWA)	LD50
Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	(+/-) (µg/mL)				
Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
Carbon disulfide	(0060)	MKCR8581	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	N/A	N/A
trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A
Diethyl ether	(0153)	IK18CA5000K	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
Ethyl methacrylate	(0381)	06126PK	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg
Iodomethane	(0489)	5HBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 70mg/kg
2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3480mg/kg
Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 1200mg/kg
Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 700mg/kg
2-Nitropropane	(0461)	14002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	N/A	N/A
1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7800mg/m3/8H)	or-rat 43g/kg
Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 910mg/kg
Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 840mg/kg
cis-1,2-Dichloroethane	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
trans-1,2-Dichloroethane	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg
Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-66-3	50 ppm (240mg/m3) (CL)	or-rat 900mg/kg
Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 105mg/kg
1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg
2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
Tetrachloroethane	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-6	0.001 ppm	or-rat 170mg/kg
1,2-Dibromomethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-08-2	50 ppm (8H)	or-rat 870mg/kg
1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2000.2	22.9	78-57-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg
1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	un-rat 3600mg/kg
1,1-Dichloropropane	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	29.7	563-56-6	N/A	N/A
cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.8	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	N/A	or-rat 670mg/kg
1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (46mg/m3/8H) (skin)	or-rat 630mg/kg
2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-rat 2400mg/kg
Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (50mg/m3/8H)	or-rat 149.6mg/kg
Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	104-51-8	1 ppm	or-rat 4694mg/kg
n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-98-1	N/A	or-rat 2699mg/kg
Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg
Tyrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-88-3	200 ppm	or-rat 5000mg/kg
2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	N/A	or-rat 1390mg/kg
2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 750mg/kg
3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	95-63-6	N/A	or-rat 5g/kg
Xylene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-67-8	N/A	or-rat 5000mg/kg
tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	98-06-6	100 ppm (435mg/m3/8H)	or-rat 5g/kg
sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	N/A	N/A
Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-49-8	75 ppm (350mg/m3/8H)	or-rat 2240mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	108-90-7	50 ppm (250mg/m3/8H)	or-rat 2290mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 3900mg/kg
2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	N/A
4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
propylbenzene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	106-46-7	75 ppm (450mg/m3/8H)	

Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.24
11	cis-1,2-Dichloroethane	18.00
12	Methacrylonitrile/methyl acrylate/Chloroform	18.46
13	Isobutane/1,1,1-Trichloroethane	19.01
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.49
17	Trichloroethane	23.58
18	1,2-Dichloropropane	24.38
19	Methyl methacrylate	24.52
20	Bromochloromethane	25.13
21	Dibromomethane/2-Hitroprene	25.46
22	cis-1,3-Dichloropropane	27.02
23	Toluene	28.05
24	Ethyl methacrylate/trans-1,3-Dichloropropane	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.24
27	Dibromochloromethane	31.16
28	1,2-Dibromomethane	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.66
32	o-Xylene	35.22
33	Styrene	35.39
34	Isopropylbenzene/Bromoforn	36.18
35	cis-1,4-Dichloro-3-butene	36.48
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.93
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.80
42	2-Chlorotoluene	38.82
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.74
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.53
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.46
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection. Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC. DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324QExpiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTBWeight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance

Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324QExpiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTBWeight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance

Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



ANAB ISO 17034 Accredited
AR-1539 Certificate Number
<https://absolutestandards.com>

CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

GHS/OSHA Compliant

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT. 06514	Emergency Telephone International Date Prepared/Revised	1-352-323-3500 January 1, 2024

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1
		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Methanol	67-56-1 TWA 200 ppm		
Skin notation	TWA 200 ppm		
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



ANAB ISO 17034 Accredited
AR-1539 Certificate Number
<https://absolutestandards.com>

CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

GHS/OSHA Compliant

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT. 06514	Emergency Telephone International Date Prepared/Revised	1-352-323-3500 January 1, 2024

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Components (Specific Chemical Identity; Common Name(s))	% (optional)
Methanol METHYL ALCOHOL CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Methanol	67-56-1 TWA 200 ppm		
Skin notation	TWA 200 ppm		
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 Lot No.: A0181905
Description : tert-Butanol Standard
tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : February 28, 2025 Storage: 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)												
1	tert-Butanol (TBA) CAS # 75-65-0 Purity 99% (Lot SHBM7694)	50,126.0 µg/mL	<table><tr><td>+/-</td><td>293.4988</td><td>µg/mL</td><td>Gravimetric</td></tr><tr><td>+/-</td><td>1,073.7654</td><td>µg/mL</td><td>Unstressed</td></tr><tr><td>+/-</td><td>1,104.9494</td><td>µg/mL</td><td>Stressed</td></tr></table>	+/-	293.4988	µg/mL	Gravimetric	+/-	1,073.7654	µg/mL	Unstressed	+/-	1,104.9494	µg/mL	Stressed
+/-	293.4988	µg/mL	Gravimetric												
+/-	1,073.7654	µg/mL	Unstressed												
+/-	1,104.9494	µg/mL	Stressed												

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:


200°C

Det. Temp:

250°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 16-Feb-2022

Balance: B442140311

Marlene Cowan - Operations Tech I

Date Passed: 21-Feb-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : December 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : December 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix
Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023

Balance: 1127510105

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek Review of SDS per April 18, 2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
CERTIFIED REFERENCE MATERIAL

30 ml
Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271



Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
CERTIFIED REFERENCE MATERIAL

30 ml
Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vol Rec 01/08/25
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

✓ 14793 to 14802



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-FL **Lot No.:** A0220563
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,060.0 µg/mL	+/- 278.5905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckal - Mix Technician

Date Mixed: 30-Dec-2024

Balance Serial # B345965662


Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pothos at 7:11 am, Jan 03, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (µeq/g)	≤ 0.3	0.2
Titration Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms


For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality



SHIPPING DOCUMENTS





284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1199 TETR06

Order Date : 1/27/2025 3:35:00 PM

Project Mgr :

Client Name : Tetra Tech NUS, Inc.

Project Name : CTO WE13

Report Type : Level 4

Client Contact : Ernie Wu

Receive DateTime : 1/27/2025 12:00:00 AM

EDD Type : ADAPT

Invoice Name : Tetra Tech NUS, Inc.

Purchase Order :

Hard Copy Date :

Invoice Contact : Ernie Wu

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1199-01	BP-VPB-192-TB-20250123	Water	01/23/2025	09:00					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1199-02	BP-VPB-192-EB-20250124	Water	01/24/2025	12:15					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1199-03	BP-VPB-192-GW-280-282	Water	01/24/2025	11:45					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1199-04	BP-VPB-192-GW-260-262	Water	01/23/2025	14:45					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1199-05	BP-VPB-192-GW-240-242	Water	01/23/2025	12:50					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1199-06	~~BP-VPB-192-220-222~~ BP-VPB-192-GW-220-222	Water	01/23/2025	10:55					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1199-07	BP-VPB-192-GW-DUP-20250123	Water	01/23/2025	12:00					
					VOCMS Group1		8260-Low	10 Bus. Days	

YG
02/03/25

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1199 TETR06

Order Date : 1/27/2025 3:35:00 PM

Project Mgr :

Client Name : Tetra Tech NUS, Inc.

Project Name : CTO WE13

Report Type : Level 4

Client Contact : Ernie Wu

Receive DateTime : 1/27/2025 12:00:00 AM

EDD Type : ADAPT

Invoice Name : Tetra Tech NUS, Inc.

Purchase Order :

Hard Copy Date :

Invoice Contact : Ernie Wu

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

Relinquished By :

Date / Time :

[Signature]
1-28-25 0845

Received By :

Date / Time :

[Signature]
01/28/25 0845 Ng # 4

Storage Area : VOA Refridgerator Room

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085438.D
 Acq On : 14 Jan 2025 14:56
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Jan 15 01:41:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	202613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	329363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146196	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.571	65	313645	95.899	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 191.800%#	
35) Dibromofluoromethane	8.159	113	236159	103.353	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 206.700%#	
50) Toluene-d8	10.565	98	881956	108.636	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 217.280%#	
62) 4-Bromofluorobenzene	12.847	95	313161	112.765	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 225.520%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	269023	98.065	ug/l	99
3) Chloromethane	2.359	50	275415	92.724	ug/l	98
4) Vinyl Chloride	2.512	62	282361	94.576	ug/l	99
5) Bromomethane	2.942	94	158888	88.105	ug/l	98
6) Chloroethane	3.112	64	176111	93.043	ug/l	99
7) Trichlorofluoromethane	3.494	101	423967	97.870	ug/l	100
8) Diethyl Ether	3.953	74	146889	98.154	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	239104	97.997	ug/l	99
10) Methyl Iodide	4.589	142	292983	104.856	ug/l	95
11) Tert butyl alcohol	5.512	59	173044	462.061	ug/l	99
12) 1,1-Dichloroethene	4.336	96	222117	102.146	ug/l	99
13) Acrolein	4.171	56	275350	538.600	ug/l	99
14) Allyl chloride	5.018	41	349717	99.111	ug/l	98
15) Acrylonitrile	5.712	53	585370	492.095	ug/l	99
16) Acetone	4.418	43	483221	457.268	ug/l	96
17) Carbon Disulfide	4.712	76	630294	94.140	ug/l	100
18) Methyl Acetate	5.012	43	304226	94.674	ug/l	99
19) Methyl tert-butyl Ether	5.788	73	743008	105.230	ug/l	100
20) Methylene Chloride	5.271	84	254912	97.440	ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	231503	99.618	ug/l	97
22) Diisopropyl ether	6.665	45	820871	104.812	ug/l	99
23) Vinyl Acetate	6.594	43	3029568	550.850	ug/l	99
24) 1,1-Dichloroethane	6.565	63	471831	98.803	ug/l	99
25) 2-Butanone	7.477	43	764979	491.912	ug/l	98
26) 2,2-Dichloropropane	7.482	77	388140	100.530	ug/l	100
27) cis-1,2-Dichloroethene	7.482	96	280197	102.367	ug/l	99
28) Bromochloromethane	7.806	49	214796	96.679	ug/l	100
29) Tetrahydrofuran	7.830	42	503459	510.635	ug/l	99
30) Chloroform	7.959	83	484941	98.253	ug/l	100
31) Cyclohexane	8.253	56	398641	96.041	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	426782	98.578	ug/l	98
36) 1,1-Dichloropropene	8.365	75	342961	106.945	ug/l	100
37) Ethyl Acetate	7.553	43	326802	100.961	ug/l	97
38) Carbon Tetrachloride	8.359	117	378390	103.065	ug/l	97
39) Methylcyclohexane	9.594	83	371734	123.354	ug/l	99
40) Benzene	8.600	78	1021393	106.001	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085438.D
 Acq On : 14 Jan 2025 14:56
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Jan 15 01:41:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	184326	109.454	ug/l	96
42) 1,2-Dichloroethane	8.665	62	374938	103.314	ug/l	100
43) Isopropyl Acetate	8.682	43	549969	106.035	ug/l	100
44) Trichloroethene	9.347	130	238296	106.242	ug/l	96
45) 1,2-Dichloropropane	9.618	63	256964	104.405	ug/l	97
46) Dibromomethane	9.706	93	180990	101.913	ug/l	99
47) Bromodichloromethane	9.882	83	388499	107.377	ug/l	98
48) Methyl methacrylate	9.676	41	267689	114.685	ug/l	99
49) 1,4-Dioxane	9.688	88	83879	2133.992	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	1643800	546.159	ug/l	100
52) Toluene	10.623	92	634743	113.689	ug/l	99
53) t-1,3-Dichloropropene	10.829	75	391238	114.424	ug/l	97
54) cis-1,3-Dichloropropene	10.306	75	410403	112.369	ug/l	99
55) 1,1,2-Trichloroethane	11.012	97	229522	103.879	ug/l	98
56) Ethyl methacrylate	10.870	69	383663	101.029	ug/l	99
57) 1,3-Dichloropropane	11.159	76	412886	107.469	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.153	63	848636	605.243	ug/l	100
59) 2-Hexanone	11.194	43	1178831	556.643	ug/l	99
60) Dibromochloromethane	11.353	129	283361	106.227	ug/l	99
61) 1,2-Dibromoethane	11.465	107	230104	104.623	ug/l	98
64) Tetrachloroethene	11.100	164	209870	102.980	ug/l	98
65) Chlorobenzene	11.888	112	677679	103.482	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	241593	100.517	ug/l	99
67) Ethyl Benzene	11.959	91	1239080	116.169	ug/l	99
68) m/p-Xylenes	12.070	106	927244	235.219	ug/l	99
69) o-Xylene	12.394	106	441279	117.135	ug/l	100
70) Styrene	12.406	104	760136	121.919	ug/l	99
71) Bromoform	12.576	173	186124	108.227	ug/l #	99
73) Isopropylbenzene	12.694	105	1146752	116.221	ug/l	100
74) N-amyl acetate	12.488	43	489217	110.308	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	327767	94.039	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	446237	150.413	ug/l	47
77) Bromobenzene	12.976	156	268616	104.200	ug/l	97
78) n-propylbenzene	13.029	91	1385353	118.600	ug/l	100
79) 2-Chlorotoluene	13.123	91	822258	108.767	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	953066	116.938	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	122508	111.544	ug/l	87
82) 4-Chlorotoluene	13.217	91	832317	110.574	ug/l	99
83) tert-Butylbenzene	13.435	119	797069	116.457	ug/l	98
84) 1,2,4-Trimethylbenzene	13.476	105	971729	119.625	ug/l	100
85) sec-Butylbenzene	13.611	105	1143885	120.587	ug/l	100
86) p-Isopropyltoluene	13.723	119	960747	100.111	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	502960	105.943	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	498709	101.341	ug/l	97
89) n-Butylbenzene	14.053	91	842368	123.781	ug/l	98
90) Hexachloroethane	14.329	117	188541	104.405	ug/l	100
91) 1,2-Dichlorobenzene	14.100	146	470919	99.445	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	63846	100.313	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	250787	113.946	ug/l	99
94) Hexachlorobutadiene	15.500	225	112899	96.619	ug/l	99
95) Naphthalene	15.635	128	756816	115.234	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	238860	107.256	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085438.D
Acq On : 14 Jan 2025 14:56
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.218 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion:168 Resp: 202613

Ion Ratio Lower Upper

168 100

99 64.9 53.6 80.4



#2

Dichlorodifluoromethane

Concen: 98.065 ug/l

RT: 2.124 min Scan# 29

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 85 Resp: 269023

Ion Ratio Lower Upper

85 100

87 32.7 16.7 50.1









#3
Chloromethane
Concen: 92.724 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 50 Resp: 275415
Ion Ratio Lower Upper
50 100
52 32.4 27.0 40.6



#4
Vinyl Chloride
Concen: 94.576 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 62 Resp: 282361
Ion Ratio Lower Upper
62 100
64 31.6 24.8 37.2









#5
Bromomethane
Concen: 88.105 ug/l
RT: 2.942 min Scan# 1
Delta R.T. -0.012 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 94 Resp: 158888
Ion Ratio Lower Upper
94 100
96 91.9 71.8 107.6



#6
Chloroethane
Concen: 93.043 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 64 Resp: 176111
Ion Ratio Lower Upper
64 100
66 31.1 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 97.870 ug/l
 RT: 3.494 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:101 Resp: 423967
 Ion Ratio Lower Upper
 101 100
 103 64.4 51.4 77.2



#8
 Diethyl Ether
 Concen: 98.154 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Tgt Ion: 74 Resp: 146889
 Ion Ratio Lower Upper
 74 100
 45 97.4 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 97.997 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 239104
Ion Ratio Lower Upper
101 100
85 47.1 37.8 56.8
151 71.8 58.8 88.2



#10
Methyl Iodide
Concen: 104.856 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:142 Resp: 292983
Ion Ratio Lower Upper
142 100
127 47.1 40.3 60.5
141 13.1 12.8 19.2









#11
Tert butyl alcohol
Concen: 462.061 ug/l
RT: 5.512 min Scan# 605
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 59 Resp: 173044
Ion Ratio Lower Upper
59 100
57 11.1 8.6 13.0



#12
1,1-Dichloroethene
Concen: 102.146 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 96 Resp: 222117
Ion Ratio Lower Upper
96 100
61 175.3 141.0 211.4
98 62.7 49.0 73.6





#13

Acrolein

Concen: 538.600 ug/l

RT: 4.171 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion: 56 Resp: 275350

Ion Ratio Lower Upper

56 100

55 71.1 56.3 84.5



#14

Allyl chloride

Concen: 99.111 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 41 Resp: 349717


Ion Ratio Lower Upper

41 100

39 83.3 64.4 96.6

76 37.9 30.5 45.7









#15
Acrylonitrile
Concen: 492.095 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 53 Resp: 585370
Ion Ratio Lower Upper
53 100
52 83.0 65.5 98.3
51 37.1 29.8 44.8



#16
Acetone
Concen: 457.268 ug/l
RT: 4.418 min Scan# 419
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 483221
Ion Ratio Lower Upper
43 100
58 27.3 23.8 35.6









#17
Carbon Disulfide
Concen: 94.140 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 76 Resp: 630294
Ion Ratio Lower Upper
76 100
78 8.7 6.9 10.3



#18
Methyl Acetate
Concen: 94.674 ug/l
RT: 5.012 min Scan# 520
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 304226
Ion Ratio Lower Upper
43 100
74 23.6 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 105.230 ug/l
RT: 5.788 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 73 Resp: 743008
Ion Ratio Lower Upper
73 100
57 22.3 17.8 26.8



#20
Methylene Chloride
Concen: 97.440 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 84 Resp: 254912
Ion Ratio Lower Upper
84 100
49 128.5 104.1 156.1
51 37.8 31.0 46.4
86 61.1 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 99.618 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 96 Resp: 231503
Ion Ratio Lower Upper
96 100
61 148.3 115.7 173.5
98 65.0 49.8 74.8



#22
Diisopropyl ether
Concen: 104.812 ug/l
RT: 6.665 min Scan# 801
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 45 Resp: 820871
Ion Ratio Lower Upper
45 100
43 55.7 45.0 67.6
87 27.2 21.8 32.8
59 11.6 10.2 15.2









#23
 Vinyl Acetate
 Concen: 550.850 ug/l
 RT: 6.594 min Scan# 789
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion: 43 Resp: 3029568
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 98.803 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Tgt Ion: 63 Resp: 471831
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.3 9.8
 100 4.6 2.0 6.0









#25
2-Butanone
Concen: 491.912 ug/l
RT: 7.477 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 764979
Ion Ratio Lower Upper
43 100
72 26.3 20.2 30.4



#26
2,2-Dichloropropane
Concen: 100.530 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 77 Resp: 388140
Ion Ratio Lower Upper
77 100
97 21.4 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 102.367 ug/l
 RT: 7.482 min Scan# 940
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion: 96 Resp: 280197
 Ion Ratio Lower Upper
 96 100
 61 156.8 0.0 311.8
 98 65.0 0.0 126.0



#28
 Bromochloromethane
 Concen: 96.679 ug/l
 RT: 7.806 min Scan# 995
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Tgt Ion: 49 Resp: 214796
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.4
 130 68.4 55.0 82.4









#29
Tetrahydrofuran
Concen: 510.635 ug/l
RT: 7.830 min Scan# 999
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 42 Resp: 503459
Ion Ratio Lower Upper
42 100
72 43.1 34.2 51.2
71 40.9 32.5 48.7



#30
Chloroform
Concen: 98.253 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 83 Resp: 484941
Ion Ratio Lower Upper
83 100
85 65.0 51.8 77.6









#31
Cyclohexane
Concen: 96.041 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 398641
Ion Ratio Lower Upper
56 100
69 29.2 22.2 33.4
84 86.4 66.4 99.6



#32
1,1,1-Trichloroethane
Concen: 98.578 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 97 Resp: 426782
Ion Ratio Lower Upper
97 100
99 64.2 49.8 74.6
61 50.6 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 95.899 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 65 Resp: 313645

Ion Ratio Lower Upper

65 100

67 51.1 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 114 Resp: 329363


Ion Ratio Lower Upper

114 100

63 21.2 0.0 47.6

88 16.2 0.0 32.6









#35
Dibromofluoromethane
Concen: 103.353 ug/l
RT: 8.159 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:113 Resp: 236159
Ion Ratio Lower Upper
113 100
111 102.3 82.7 124.1
192 17.4 14.3 21.5



#36
1,1-Dichloropropene
Concen: 106.945 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 342961
Ion Ratio Lower Upper
75 100
110 32.6 16.5 49.5
77 30.6 24.4 36.6





#37

Ethyl Acetate

Concen: 100.961 ug/l

RT: 7.553 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion: 43 Resp: 326802

Ion Ratio Lower Upper

43 100

61 12.0 10.9 16.3

70 9.6 7.5 11.3



#38

Carbon Tetrachloride

Concen: 103.065 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 117 Resp: 378390


Ion Ratio Lower Upper

117 100

119 98.1 75.4 113.2

121 30.2 24.6 37.0





#39

Methylcyclohexane

Concen: 123.354 ug/l

RT: 9.594 min Scan# 11299

Delta R.T. -0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion: 83 Resp: 371734

Ion Ratio Lower Upper

83 100

55 79.3 62.6 94.0

98 46.7 37.7 56.5



#40

Benzene

Concen: 106.001 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 78 Resp: 1021393

Ion Ratio Lower Upper

78 100

77 23.9 19.0 28.6









#41
Methacrylonitrile
Concen: 109.454 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 41 Resp: 184326




Ion	Ratio	Lower	Upper
41	100		
39	52.8	46.0	69.0
67	68.8	57.4	86.2
52	32.3	25.5	38.3




#42
1,2-Dichloroethane
Concen: 103.314 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 62 Resp: 374938

Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	17.0









#43
Isopropyl Acetate
Concen: 106.035 ug/l
RT: 8.682 min Scan# 1144
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 549969
Ion Ratio Lower Upper
43 100
61 25.7 20.7 31.1
87 12.2 9.8 14.8



#44
Trichloroethene
Concen: 106.242 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 130 Resp: 238296
Ion Ratio Lower Upper
130 100
95 101.5 0.0 195.8









#45
1,2-Dichloropropane
Concen: 104.405 ug/l
RT: 9.618 min Scan# 1303
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 63 Resp: 256964
Ion Ratio Lower Upper
63 100
65 30.3 25.6 38.4



#46
Dibromomethane
Concen: 101.913 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 93 Resp: 180990
Ion Ratio Lower Upper
93 100
95 81.6 64.7 97.1
174 87.5 69.0 103.4









#47
Bromodichloromethane
Concen: 107.377 ug/l
RT: 9.882 min Scan# 1348
Delta R.T. 0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 388499
Ion Ratio Lower Upper
83 100
85 65.9 51.2 76.8
127 8.4 6.5 9.7



#48
Methyl methacrylate
Concen: 114.685 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 41 Resp: 267689
Ion Ratio Lower Upper
41 100
69 80.9 64.7 97.1
39 63.1 49.0 73.6









#49
1,4-Dioxane
Concen: 2133.992 ug/l
RT: 9.688 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 88 Resp: 83879
Ion Ratio Lower Upper
88 100
43 31.3 26.6 39.8
58 72.1 59.5 89.3



#50
Toluene-d8
Concen: 108.636 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 98 Resp: 881956
Ion Ratio Lower Upper
98 100
100 64.5 52.2 78.4





#51

4-Methyl-2-Pentanone

Concen: 546.159 ug/l

RT: 10.441 min Scan# 1443

Delta R.T. 0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 1643800

Ion Ratio Lower Upper

43 100

58 38.4 30.5 45.7



#52

Toluene

Concen: 113.689 ug/l

RT: 10.623 min Scan# 1474

Delta R.T. -0.001 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 92 Resp: 634743

Ion Ratio Lower Upper

92 100

91 173.1 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 114.424 ug/l
 RT: 10.829 min Scan# 1509
 Delta R.T. -0.006 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100


Tgt Ion: 75 Resp: 391238
 Ion Ratio Lower Upper
 75 100
 77 31.0 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 112.369 ug/l
 RT: 10.306 min Scan# 1420
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 410403
 Ion Ratio Lower Upper
 75 100
 77 31.9 25.0 37.4
 39 54.3 43.1 64.7









#55
1,1,2-Trichloroethane
Concen: 103.879 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 97 Resp: 229522




Ion	Ratio	Lower	Upper
97	100		
83	88.4	69.0	103.6
85	56.3	44.6	66.8
99	62.9	48.6	72.8




#56
Ethyl methacrylate
Concen: 101.029 ug/l
RT: 10.870 min Scan# 1516
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 69 Resp: 383663

Ion	Ratio	Lower	Upper
69	100		
41	69.1	54.6	82.0
39	40.1	32.4	48.6





#57

1,3-Dichloropropane

Concen: 107.469 ug/l

RT: 11.159 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 76 Resp: 412886

Ion Ratio Lower Upper

76 100

78 32.2 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 605.243 ug/l

RT: 10.153 min Scan# 1394

Delta R.T. -0.006 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 63 Resp: 848636

Ion Ratio Lower Upper

63 100

106 26.8 21.6 32.4









#59
2-Hexanone
Concen: 556.643 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 1178831
Ion Ratio Lower Upper
43 100
58 53.2 26.2 78.6



#60
Dibromochloromethane
Concen: 106.227 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 129 Resp: 283361
Ion Ratio Lower Upper
129 100
127 78.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 104.623 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:107 Resp: 230104
Ion Ratio Lower Upper
107 100
109 93.1 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 112.765 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 95 Resp: 313161
Ion Ratio Lower Upper
95 100
174 73.4 0.0 145.0
176 70.5 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion:117 Resp: 298938
Ion Ratio Lower Upper
117 100
82 59.3 48.6 72.8
119 31.6 26.6 39.8



#64
Tetrachloroethene
Concen: 102.980 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:164 Resp: 209870
Ion Ratio Lower Upper
164 100
166 126.8 103.4 155.2
129 98.3 79.2 118.8
131 94.7 77.1 115.7









#65
Chlorobenzene
Concen: 103.482 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:112 Resp: 677679
Ion Ratio Lower Upper
112 100
114 32.5 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 100.517 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:131 Resp: 241593
Ion Ratio Lower Upper
131 100
133 94.3 47.4 142.3
119 64.9 33.1 99.5









#67
Ethyl Benzene
Concen: 116.169 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 91 Resp: 1239080
Ion Ratio Lower Upper
91 100
106 30.3 23.8 35.8



#68
m/p-Xylenes
Concen: 235.219 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:106 Resp: 927244
Ion Ratio Lower Upper
106 100
91 208.0 167.7 251.5









#69
o-Xylene
Concen: 117.135 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:106 Resp: 441279
Ion Ratio Lower Upper
106 100
91 221.3 110.4 331.2



#70
Styrene
Concen: 121.919 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:104 Resp: 760136
Ion Ratio Lower Upper
104 100
78 52.8 42.5 63.7
103 55.3 43.8 65.8









#71
Bromoform
Concen: 108.227 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion:173 Resp: 186124
Ion Ratio Lower Upper
173 100
175 48.4 24.4 73.2
254 0.1 0.0 0.0#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:152 Resp: 146196
Ion Ratio Lower Upper
152 100
115 62.7 31.1 93.3
150 193.1 0.0 343.6









#73
Isopropylbenzene
Concen: 116.221 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:105 Resp: 1146752
Ion Ratio Lower Upper
105 100
120 25.7 12.8 38.3



#74
N-ethyl acetate
Concen: 110.308 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 489217
Ion Ratio Lower Upper
43 100
70 41.9 34.0 51.0
55 25.7 21.4 32.2
61 23.5 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 94.039 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion: 83 Resp: 327767

Ion Ratio Lower Upper

83 100

131 9.8 4.8 14.4

85 64.1 32.2 96.6



#76

1,2,3-Trichloropropane

Concen: 150.413 ug/l

RT: 12.988 min Scan# 1876

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 75 Resp: 446237

Ion Ratio Lower Upper

75 100

77 133.9 109.7 329.2









#77
Bromobenzene
Concen: 104.200 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:156 Resp: 268616
Ion Ratio Lower Upper
156 100
77 222.4 114.1 342.4
158 96.3 48.9 146.8



#78
n-propylbenzene
Concen: 118.600 ug/l
RT: 13.029 min Scan# 1883
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 91 Resp: 1385353
Ion Ratio Lower Upper
91 100
120 21.8 10.9 32.6









#79
2-Chlorotoluene
Concen: 108.767 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 822258
Ion Ratio Lower Upper
91 100
126 31.8 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 116.938 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 105 Resp: 953066
Ion Ratio Lower Upper
105 100
120 47.8 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 111.544 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 75 Resp: 122508
Ion Ratio Lower Upper
75 100
53 100.9 95.6 143.4
89 44.4 37.0 55.6



#82
4-Chlorotoluene
Concen: 110.574 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 91 Resp: 832317
Ion Ratio Lower Upper
91 100
126 32.5 15.9 47.7





#83

tert-Butylbenzene

Concen: 116.457 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:119 Resp: 797069

Ion Ratio Lower Upper

119 100

91 70.7 35.1 105.3

134 26.6 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 119.625 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion:105 Resp: 971729

Ion Ratio Lower Upper

105 100

120 43.4 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 120.587 ug/l
 RT: 13.611 min Scan# 1982
 Delta R.T. -0.001 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:105 Resp: 1143885
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 100.111 ug/l
 RT: 13.723 min Scan# 2001
 Delta R.T. 0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Tgt Ion:119 Resp: 960747
 Ion Ratio Lower Upper
 119 100
 134 25.3 12.7 38.0
 91 25.0 12.7 38.1









#87
 1,3-Dichlorobenzene
 Concen: 105.943 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:146 Resp: 502960
 Ion Ratio Lower Upper
 146 100
 111 42.8 21.4 64.3
 148 63.6 32.3 96.9



#88
 1,4-Dichlorobenzene
 Concen: 101.341 ug/l
 RT: 13.811 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Tgt Ion:146 Resp: 498709
 Ion Ratio Lower Upper
 146 100
 111 40.1 21.3 63.7
 148 63.1 32.4 97.0





#89

n-Butylbenzene

Concen: 123.781 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion: 91 Resp: 842368

Ion Ratio Lower Upper

91 100

92 52.8 25.8 77.3

134 24.3 11.7 35.1



#90

Hexachloroethane

Concen: 104.405 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.000 min

Lab File: VN085438.D




Acq: 14 Jan 2025 14:56


Tgt Ion: 117 Resp: 188541

Ion Ratio Lower Upper

117 100

201 67.2 33.7 101.0









#91
1,2-Dichlorobenzene
Concen: 99.445 ug/l
RT: 14.100 min Scan# 2065
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:146 Resp: 470919
Ion Ratio Lower Upper
146 100
111 43.9 21.7 65.1
148 64.0 31.4 94.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 100.313 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 63846
Ion Ratio Lower Upper
75 100
155 73.5 36.4 109.2
157 93.0 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 113.946 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:180 Resp: 250787

Ion Ratio Lower Upper

180 100

182 94.9 46.8 140.3

145 32.6 16.0 48.0



#94

Hexachlorobutadiene

Concen: 96.619 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:225 Resp: 112899


Ion Ratio Lower Upper

225 100

223 61.8 30.7 92.1

227 63.6 30.9 92.5





#95

Naphthalene

Concen: 115.234 ug/l

RT: 15.635 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion:128 Resp: 756816

Ion Ratio Lower Upper

128 100

127 13.1 10.6 16.0

129 10.8 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 107.256 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:180 Resp: 238860

Ion Ratio Lower Upper

180 100

182 95.0 47.4 142.2

145 34.3 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085439.D
 Acq On : 14 Jan 2025 15:19
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Jan 15 01:42:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	199403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	339872	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146624	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	165768	51.500	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.000%	
35) Dibromofluoromethane	8.159	113	121554	51.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.100%	
50) Toluene-d8	10.565	98	430698	51.411	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.820%	
62) 4-Bromofluorobenzene	12.847	95	152568	53.239	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.480%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	125504	46.486	ug/l	100
3) Chloromethane	2.359	50	135631	46.398	ug/l	100
4) Vinyl Chloride	2.512	62	136831	46.569	ug/l	100
5) Bromomethane	2.953	94	83199	46.877	ug/l	100
6) Chloroethane	3.118	64	84517	45.371	ug/l	100
7) Trichlorofluoromethane	3.501	101	198726	46.613	ug/l	100
8) Diethyl Ether	3.953	74	71764	48.726	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	108081	45.010	ug/l	100
10) Methyl Iodide	4.589	142	143383	52.142	ug/l	100
11) Tert butyl alcohol	5.512	59	92841	251.894	ug/l	100
12) 1,1-Dichloroethene	4.342	96	106247	49.647	ug/l	100
13) Acrolein	4.171	56	139234	276.734	ug/l	100
14) Allyl chloride	5.018	41	168654	48.567	ug/l	100
15) Acrylonitrile	5.712	53	296781	253.507	ug/l	100
16) Acetone	4.418	43	251271	241.603	ug/l	100
17) Carbon Disulfide	4.706	76	294604	44.710	ug/l	100
18) Methyl Acetate	5.018	43	157485	49.798	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	373561	53.758	ug/l	100
20) Methylene Chloride	5.271	84	125508	48.748	ug/l	100
21) trans-1,2-Dichloroethene	5.783	96	110684	48.395	ug/l	100
22) Diisopropyl ether	6.665	45	406253	52.707	ug/l	100
23) Vinyl Acetate	6.594	43	1491031	275.471	ug/l	100
24) 1,1-Dichloroethane	6.565	63	233336	49.648	ug/l	100
25) 2-Butanone	7.477	43	388457	253.815	ug/l	100
26) 2,2-Dichloropropane	7.483	77	186591	49.106	ug/l	100
27) cis-1,2-Dichloroethene	7.483	96	136229	50.571	ug/l	100
28) Bromochloromethane	7.806	49	108148	49.461	ug/l	100
29) Tetrahydrofuran	7.830	42	259683	267.624	ug/l	100
30) Chloroform	7.959	83	234271	48.229	ug/l	100
31) Cyclohexane	8.253	56	175588	42.984	ug/l	100
32) 1,1,1-Trichloroethane	8.165	97	202590	47.547	ug/l	100
36) 1,1-Dichloropropene	8.365	75	157031	47.453	ug/l	100
37) Ethyl Acetate	7.553	43	165866	49.658	ug/l	100
38) Carbon Tetrachloride	8.359	117	180199	47.564	ug/l	100
39) Methylcyclohexane	9.600	83	157234	50.562	ug/l	100
40) Benzene	8.600	78	492606	49.542	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085439.D
 Acq On : 14 Jan 2025 15:19
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Jan 15 01:42:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	91711	52.775	ug/l	100
42) 1,2-Dichloroethane	8.665	62	186065	49.685	ug/l	100
43) Isopropyl Acetate	8.683	43	276228	51.611	ug/l	100
44) Trichloroethene	9.347	130	109953	47.506	ug/l	100
45) 1,2-Dichloropropane	9.618	63	126084	49.644	ug/l	100
46) Dibromomethane	9.706	93	90670	49.476	ug/l	100
47) Bromodichloromethane	9.882	83	189915	50.867	ug/l	100
48) Methyl methacrylate	9.677	41	133399	55.385	ug/l	100
49) 1,4-Dioxane	9.688	88	43048	1061.334	ug/l	100
51) 4-Methyl-2-Pentanone	10.441	43	836707	269.403	ug/l	100
52) Toluene	10.624	92	295792	51.341	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	187439	53.125	ug/l	100
54) cis-1,3-Dichloropropene	10.306	75	199771	53.006	ug/l	100
55) 1,1,2-Trichloroethane	11.012	97	115462	50.641	ug/l	100
56) Ethyl methacrylate	10.871	69	184775	48.182	ug/l	100
57) 1,3-Dichloropropane	11.159	76	204590	51.606	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	409594	283.088	ug/l	100
59) 2-Hexanone	11.188	43	606740	277.643	ug/l	100
60) Dibromochloromethane	11.353	129	140704	51.116	ug/l	100
61) 1,2-Dibromoethane	11.465	107	113541	50.028	ug/l	100
64) Tetrachloroethene	11.100	164	95879	47.295	ug/l	100
65) Chlorobenzene	11.888	112	319992	49.121	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	116488	48.722	ug/l	100
67) Ethyl Benzene	11.959	91	555322	52.339	ug/l	100
68) m/p-Xylenes	12.065	106	420702	107.286	ug/l	100
69) o-Xylene	12.394	106	202597	54.062	ug/l	100
70) Styrene	12.406	104	348720	56.227	ug/l	100
71) Bromoform	12.576	173	92366	53.993	ug/l #	100
73) Isopropylbenzene	12.688	105	505619	51.094	ug/l	100
74) N-amyl acetate	12.488	43	237870	53.478	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	167856	48.019	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	219507	73.773	ug/l	45
77) Bromobenzene	12.976	156	125800	48.657	ug/l	100
78) n-propylbenzene	13.029	91	609022	51.986	ug/l	100
79) 2-Chlorotoluene	13.123	91	378981	49.985	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	431021	52.730	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	57184	51.914	ug/l	100
82) 4-Chlorotoluene	13.218	91	379524	50.273	ug/l	100
83) tert-Butylbenzene	13.435	119	357563	52.090	ug/l	100
84) 1,2,4-Trimethylbenzene	13.476	105	442175	54.275	ug/l	100
85) sec-Butylbenzene	13.612	105	501561	52.719	ug/l	100
86) p-Isopropyltoluene	13.723	119	421052	49.240	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	229518	48.204	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	229084	46.416	ug/l	100
89) n-Butylbenzene	14.053	91	364111	53.348	ug/l	100
90) Hexachloroethane	14.329	117	84602	46.712	ug/l	100
91) 1,2-Dichlorobenzene	14.100	146	228024	48.012	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	32536	50.971	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	114468	51.857	ug/l	100
94) Hexachlorobutadiene	15.500	225	53846	45.947	ug/l	100
95) Naphthalene	15.635	128	363882	55.243	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	115251	51.601	ug/l	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085439.D
Acq On : 14 Jan 2025 15:19
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Quant Time: Jan 15 01:42:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:168 Resp: 199403
Ion Ratio Lower Upper
168 100
99 67.0 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 46.486 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 85 Resp: 125504
Ion Ratio Lower Upper
85 100
87 33.4 16.7 50.1









#3
Chloromethane
Concen: 46.398 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 50 Resp: 135631
Ion Ratio Lower Upper
50 100
52 33.8 27.0 40.6



#4
Vinyl Chloride
Concen: 46.569 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 62 Resp: 136831
Ion Ratio Lower Upper
62 100
64 31.0 24.8 37.2









#5
Bromomethane
Concen: 46.877 ug/l
RT: 2.953 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 83199
Ion Ratio Lower Upper
94 100
96 89.7 71.8 107.6



#6
Chloroethane
Concen: 45.371 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 64 Resp: 84517
Ion Ratio Lower Upper
64 100
66 30.7 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 46.613 ug/l
 RT: 3.501 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:101 Resp: 198726
 Ion Ratio Lower Upper
 101 100
 103 64.3 51.4 77.2



#8
 Diethyl Ether
 Concen: 48.726 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion: 74 Resp: 71764
 Ion Ratio Lower Upper
 74 100
 45 99.4 49.7 149.1









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 45.010 ug/l
 RT: 4.371 min Scan# 411
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:101 Resp: 108081
 Ion Ratio Lower Upper
 101 100
 85 47.3 37.8 56.8
 151 73.5 58.8 88.2



#10
 Methyl Iodide
 Concen: 52.142 ug/l
 RT: 4.589 min Scan# 448
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion:142 Resp: 143383
 Ion Ratio Lower Upper
 142 100
 127 50.4 40.3 60.5
 141 16.0 12.8 19.2









#11
Tert butyl alcohol
Concen: 251.894 ug/l
RT: 5.512 min Scan# 605
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 59 Resp: 92841
Ion Ratio Lower Upper
59 100
57 10.8 8.6 13.0



#12
1,1-Dichloroethene
Concen: 49.647 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 96 Resp: 106247
Ion Ratio Lower Upper
96 100
61 176.2 141.0 211.4
98 61.3 49.0 73.6









#13
Acrolein
Concen: 276.734 ug/l
RT: 4.171 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 139234
Ion Ratio Lower Upper
56 100
55 70.4 56.3 84.5



#14
Allyl chloride
Concen: 48.567 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 41 Resp: 168654
Ion Ratio Lower Upper
41 100
39 80.5 64.4 96.6
76 38.1 30.5 45.7









#15
Acrylonitrile
Concen: 253.507 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 53 Resp: 296781
Ion Ratio Lower Upper
53 100
52 81.9 65.5 98.3
51 37.3 29.8 44.8



#16
Acetone
Concen: 241.603 ug/l
RT: 4.418 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 251271
Ion Ratio Lower Upper
43 100
58 29.7 23.8 35.6









#17
Carbon Disulfide
Concen: 44.710 ug/l
RT: 4.706 min Scan# 468
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 76 Resp: 294604
Ion Ratio Lower Upper
76 100
78 8.6 6.9 10.3



#18
Methyl Acetate
Concen: 49.798 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 157485
Ion Ratio Lower Upper
43 100
74 23.3 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 53.758 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 373561
Ion Ratio Lower Upper
73 100
57 22.3 17.8 26.8



#20
Methylene Chloride
Concen: 48.748 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 84 Resp: 125508
Ion Ratio Lower Upper
84 100
49 130.1 104.1 156.1
51 38.7 31.0 46.4
86 64.9 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 48.395 ug/l
RT: 5.783 min Scan# 651
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 110684
Ion Ratio Lower Upper
96 100
61 144.6 115.7 173.5
98 62.3 49.8 74.8



#22
Diisopropyl ether
Concen: 52.707 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 45 Resp: 406253
Ion Ratio Lower Upper
45 100
43 56.3 45.0 67.6
87 27.3 21.8 32.8
59 12.7 10.2 15.2









#23
 Vinyl Acetate
 Concen: 275.471 ug/l
 RT: 6.594 min Scan# 789
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 1491031
 Ion Ratio Lower Upper
 43 100
 86 10.5 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 49.648 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion: 63 Resp: 233336
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.3 9.8
 100 4.0 2.0 6.0









#25
2-Butanone
Concen: 253.815 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 388457
Ion Ratio Lower Upper
43 100
72 25.3 20.2 30.4



#26
2,2-Dichloropropane
Concen: 49.106 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 77 Resp: 186591
Ion Ratio Lower Upper
77 100
97 21.4 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 50.571 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 96 Resp: 136229
 Ion Ratio Lower Upper
 96 100
 61 155.9 0.0 311.8
 98 63.0 0.0 126.0



#28
 Bromochloromethane
 Concen: 49.461 ug/l
 RT: 7.806 min Scan# 995
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion: 49 Resp: 108148
 Ion Ratio Lower Upper
 49 100
 129 2.2 0.0 4.4
 130 68.7 55.0 82.4





#29

Tetrahydrofuran

Concen: 267.624 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 259683

Ion Ratio Lower Upper

42 100

72 42.7 34.2 51.2

71 40.6 32.5 48.7



#30

Chloroform

Concen: 48.229 ug/l

RT: 7.959 min Scan# 1021

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 83 Resp: 234271

Ion Ratio Lower Upper

83 100

85 64.7 51.8 77.6









#31
Cyclohexane
Concen: 42.984 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 175588
Ion Ratio Lower Upper
56 100
69 27.8 22.2 33.4
84 83.0 66.4 99.6



#32
1,1,1-Trichloroethane
Concen: 47.547 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 97 Resp: 202590
Ion Ratio Lower Upper
97 100
99 62.2 49.8 74.6
61 51.8 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 51.500 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 65 Resp: 165768

Ion Ratio Lower Upper

65 100

67 50.8 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 114 Resp: 339872


Ion Ratio Lower Upper

114 100

63 23.8 0.0 47.6

88 16.3 0.0 32.6





#35

Dibromofluoromethane

Concen: 51.552 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:113 Resp: 121554

Ion Ratio Lower Upper

113 100

111 103.4 82.7 124.1

192 17.9 14.3 21.5



#36

1,1-Dichloropropene

Concen: 47.453 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 75 Resp: 157031


Ion Ratio Lower Upper

75 100

110 33.0 16.5 49.5

77 30.5 24.4 36.6









#37
Ethyl Acetate
Concen: 49.658 ug/l
RT: 7.553 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 165866
Ion Ratio Lower Upper
43 100
61 13.6 10.9 16.3
70 9.4 7.5 11.3



#38
Carbon Tetrachloride
Concen: 47.564 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 117 Resp: 180199
Ion Ratio Lower Upper
117 100
119 94.3 75.4 113.2
121 30.8 24.6 37.0









#39
Methylcyclohexane
Concen: 50.562 ug/l
RT: 9.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 157234
Ion Ratio Lower Upper
83 100
55 78.3 62.6 94.0
98 47.1 37.7 56.5



#40
Benzene
Concen: 49.542 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 78 Resp: 492606
Ion Ratio Lower Upper
78 100
77 23.8 19.0 28.6









#41
Methacrylonitrile
Concen: 52.775 ug/l
RT: 7.771 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:	41	Resp:	91711
Ion	Ratio	Lower	Upper
41	100		
39	57.5	46.0	69.0
67	71.8	57.4	86.2
52	31.9	25.5	38.3



#42
1,2-Dichloroethane
Concen: 49.685 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:	62	Resp:	186065
Ion	Ratio	Lower	Upper
62	100		
98	8.5	0.0	17.0









#43
Isopropyl Acetate
Concen: 51.611 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 276228
Ion Ratio Lower Upper
43 100
61 25.9 20.7 31.1
87 12.3 9.8 14.8



#44
Trichloroethene
Concen: 47.506 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:130 Resp: 109953
Ion Ratio Lower Upper
130 100
95 97.9 0.0 195.8





#45

1,2-Dichloropropane

Concen: 49.644 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 63 Resp: 126084

Ion Ratio Lower Upper

63 100

65 32.0 25.6 38.4



#46

Dibromomethane

Concen: 49.476 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 93 Resp: 90670


Ion Ratio Lower Upper

93 100

95 80.9 64.7 97.1

174 86.2 69.0 103.4









#47
Bromodichloromethane
Concen: 50.867 ug/l
RT: 9.882 min Scan# 1348
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 189915
Ion Ratio Lower Upper
83 100
85 64.0 51.2 76.8
127 8.1 6.5 9.7



#48
Methyl methacrylate
Concen: 55.385 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 41 Resp: 133399
Ion Ratio Lower Upper
41 100
69 80.9 64.7 97.1
39 61.3 49.0 73.6





#49

1,4-Dioxane

Concen: 1061.334 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 88 Resp: 43048

Ion Ratio Lower Upper

88 100

43 33.2 26.6 39.8

58 74.4 59.5 89.3



#50

Toluene-d8

Concen: 51.411 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 98 Resp: 430698

Ion Ratio Lower Upper

98 100

100 65.3 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 269.403 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 836707
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7



#52
Toluene
Concen: 51.341 ug/l
RT: 10.624 min Scan# 1474
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 92 Resp: 295792
Ion Ratio Lower Upper
92 100
91 174.0 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 53.125 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 75 Resp: 187439
 Ion Ratio Lower Upper
 75 100
 77 29.4 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 53.006 ug/l
 RT: 10.306 min Scan# 1420
 Delta R.T. -0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 199771
 Ion Ratio Lower Upper
 75 100
 77 31.2 25.0 37.4
 39 53.9 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 50.641 ug/l

RT: 11.012 min Scan# 115462

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 115462





Ion Ratio Lower Upper

97 100

83 86.3 69.0 103.6

85 55.7 44.6 66.8

99 60.7 48.6 72.8



#56

Ethyl methacrylate

Concen: 48.182 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion: 69 Resp: 184775


Ion Ratio Lower Upper

69 100

41 68.3 54.6 82.0

39 40.5 32.4 48.6





#57

1,3-Dichloropropane

Concen: 51.606 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 204590

Ion Ratio Lower Upper

76 100

78 32.0 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 283.088 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion: 63 Resp: 409594

Ion Ratio Lower Upper

63 100

106 27.0 21.6 32.4









#59
2-Hexanone
Concen: 277.643 ug/l
RT: 11.188 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 606740
Ion Ratio Lower Upper
43 100
58 52.4 26.2 78.6



#60
Dibromochloromethane
Concen: 51.116 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:129 Resp: 140704
Ion Ratio Lower Upper
129 100
127 77.2 38.6 115.8









#61
1,2-Dibromoethane
Concen: 50.028 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 113541
Ion Ratio Lower Upper
107 100
109 94.9 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 53.239 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 95 Resp: 152568
Ion Ratio Lower Upper
95 100
174 72.5 0.0 145.0
176 71.2 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:117 Resp: 297366
Ion Ratio Lower Upper
117 100
82 60.7 48.6 72.8
119 33.2 26.6 39.8



#64
Tetrachloroethene
Concen: 47.295 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:164 Resp: 95879
Ion Ratio Lower Upper
164 100
166 129.3 103.4 155.2
129 99.0 79.2 118.8
131 96.4 77.1 115.7





#65

Chlorobenzene

Concen: 49.121 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:112 Resp: 319992

Ion Ratio Lower Upper

112 100

114 31.6 25.3 37.9



#66

1,1,1,2-Tetrachloroethane

Concen: 48.722 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion:131 Resp: 116488


Ion Ratio Lower Upper

131 100

133 94.9 47.4 142.3

119 66.3 33.1 99.5









#67
Ethyl Benzene
Concen: 52.339 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 555322
Ion Ratio Lower Upper
91 100
106 29.8 23.8 35.8



#68
m/p-Xylenes
Concen: 107.286 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:106 Resp: 420702
Ion Ratio Lower Upper
106 100
91 209.6 167.7 251.5









#69
o-Xylene
Concen: 54.062 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:106 Resp: 202597
Ion Ratio Lower Upper
106 100
91 220.8 110.4 331.2



#70
Styrene
Concen: 56.227 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:104 Resp: 348720
Ion Ratio Lower Upper
104 100
78 53.1 42.5 63.7
103 54.8 43.8 65.8





#71

Bromoform

Concen: 53.993 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:173 Resp: 92366

Ion Ratio Lower Upper

173 100

175 48.8 24.4 73.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19




Tgt Ion:152 Resp: 146624


Ion Ratio Lower Upper

152 100

115 62.2 31.1 93.3

150 171.8 0.0 343.6









#73
Isopropylbenzene
Concen: 51.094 ug/l
RT: 12.688 min Scan# 1825
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:105 Resp: 505619
Ion Ratio Lower Upper
105 100
120 25.5 12.8 38.3



#74
N-ethyl acetate
Concen: 53.478 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 237870
Ion Ratio Lower Upper
43 100
70 42.5 34.0 51.0
55 26.8 21.4 32.2
61 23.7 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 48.019 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 83 Resp: 167856
 Ion Ratio Lower Upper
 83 100
 131 9.6 4.8 14.4
 85 64.4 32.2 96.6



#76
 1,2,3-Trichloropropane
 Concen: 73.773 ug/l
 RT: 12.988 min Scan# 1876
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 219507
 Ion Ratio Lower Upper
 75 100
 77 130.9 109.7 329.2









#77
Bromobenzene
Concen: 48.657 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:156 Resp: 125800
Ion Ratio Lower Upper
156 100
77 228.3 114.1 342.4
158 97.9 48.9 146.8



#78
n-propylbenzene
Concen: 51.986 ug/l
RT: 13.029 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 91 Resp: 609022
Ion Ratio Lower Upper
91 100
120 21.7 10.9 32.6









#79
2-Chlorotoluene
Concen: 49.985 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 378981
Ion Ratio Lower Upper
91 100
126 31.4 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 52.730 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 105 Resp: 431021
Ion Ratio Lower Upper
105 100
120 47.8 23.9 71.7









#81
 trans-1,4-Dichloro-2-butene
 Concen: 51.914 ug/l
 RT: 12.735 min Scan# 1833
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:	75	Resp:	57184
Ion	Ratio	Lower	Upper
75	100		
53	119.5	95.6	143.4
89	46.3	37.0	55.6



#82
 4-Chlorotoluene
 Concen: 50.273 ug/l
 RT: 13.218 min Scan# 1915
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion:	91	Resp:	379524
Ion	Ratio	Lower	Upper
91	100		
126	31.8	15.9	47.7





#83

tert-Butylbenzene

Concen: 52.090 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085439.D

Acq: 14 Jan 2025 15:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:119 Resp: 357563

Ion Ratio Lower Upper

119 100

91 70.2 35.1 105.3

134 24.1 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 54.275 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. 0.000 min

Lab File: VN085439.D




Acq: 14 Jan 2025 15:19


Tgt Ion:105 Resp: 442175

Ion Ratio Lower Upper

105 100

120 43.3 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 52.719 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:105 Resp: 501561
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 49.240 ug/l
 RT: 13.723 min Scan# 2001
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion:119 Resp: 421052
 Ion Ratio Lower Upper
 119 100
 134 25.3 12.7 38.0
 91 25.4 12.7 38.1









#87
 1,3-Dichlorobenzene
 Concen: 48.204 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:146 Resp: 229518
 Ion Ratio Lower Upper
 146 100
 111 42.9 21.4 64.3
 148 64.6 32.3 96.9



#88
 1,4-Dichlorobenzene
 Concen: 46.416 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion:146 Resp: 229084
 Ion Ratio Lower Upper
 146 100
 111 42.5 21.3 63.7
 148 64.7 32.4 97.0









#89
n-Butylbenzene
Concen: 53.348 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 364111
Ion Ratio Lower Upper
91 100
92 51.5 25.8 77.3
134 23.4 11.7 35.1



#90
Hexachloroethane
Concen: 46.712 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 117 Resp: 84602
Ion Ratio Lower Upper
117 100
201 67.3 33.7 101.0









#91
1,2-Dichlorobenzene
Concen: 48.012 ug/l
RT: 14.100 min Scan# 2065
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 228024
Ion Ratio Lower Upper
146 100
111 43.4 21.7 65.1
148 62.8 31.4 94.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 50.971 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 32536
Ion Ratio Lower Upper
75 100
155 72.8 36.4 109.2
157 90.7 45.4 136.1









#93
 1,2,4-Trichlorobenzene
 Concen: 51.857 ug/l
 RT: 15.388 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:180 Resp: 114468
 Ion Ratio Lower Upper
 180 100
 182 93.5 46.8 140.3
 145 32.0 16.0 48.0



#94
 Hexachlorobutadiene
 Concen: 45.947 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. 0.000 min
 Lab File: VN085439.D
 Acq: 14 Jan 2025 15:19

Tgt Ion:225 Resp: 53846
 Ion Ratio Lower Upper
 225 100
 223 61.4 30.7 92.1
 227 61.7 30.9 92.5









#95
Naphthalene
Concen: 55.243 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:128 Resp: 363882
Ion Ratio Lower Upper
128 100
127 13.3 10.6 16.0
129 11.0 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 51.601 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:180 Resp: 115251
Ion Ratio Lower Upper
180 100
182 94.8 47.4 142.2
145 33.8 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085440.D
 Acq On : 14 Jan 2025 15:43
 Operator : JC\MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Quant Time: Jan 15 01:43:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	200796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	340593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	60593	18.694	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	37.380%#		
35) Dibromofluoromethane	8.165	113	45655	19.322	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	38.640%#		
50) Toluene-d8	10.559	98	164391	19.581	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	39.160%#		
62) 4-Bromofluorobenzene	12.847	95	55798	19.430	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	38.860%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	54666	20.107	ug/l	98
3) Chloromethane	2.359	50	58426	19.848	ug/l	100
4) Vinyl Chloride	2.512	62	58403	19.739	ug/l	99
5) Bromomethane	2.954	94	36440	20.389	ug/l	94
6) Chloroethane	3.118	64	37594	20.041	ug/l	93
7) Trichlorofluoromethane	3.501	101	88121	20.526	ug/l	99
8) Diethyl Ether	3.965	74	29138	19.647	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	48933	20.237	ug/l	99
10) Methyl Iodide	4.589	142	56567	20.428	ug/l	93
11) Tert butyl alcohol	5.506	59	37984	102.342	ug/l	100
12) 1,1-Dichloroethene	4.342	96	44670	20.728	ug/l	95
13) Acrolein	4.171	56	49649	97.995	ug/l	99
14) Allyl chloride	5.024	41	67705	19.362	ug/l	96
15) Acrylonitrile	5.712	53	121626	103.171	ug/l	99
16) Acetone	4.424	43	101065	96.502	ug/l	96
17) Carbon Disulfide	4.712	76	132300	19.939	ug/l	98
18) Methyl Acetate	5.018	43	60873	19.115	ug/l	99
19) Methyl tert-butyl Ether	5.795	73	148847	21.272	ug/l	100
20) Methylene Chloride	5.271	84	52813	20.370	ug/l	97
21) trans-1,2-Dichloroethene	5.789	96	46115	20.023	ug/l	98
22) Diisopropyl ether	6.671	45	166719	21.480	ug/l	97
23) Vinyl Acetate	6.594	43	666942	122.364	ug/l	99
24) 1,1-Dichloroethane	6.565	63	96875	20.469	ug/l #	96
25) 2-Butanone	7.477	43	159952	103.786	ug/l	98
26) 2,2-Dichloropropane	7.483	77	79145	20.684	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	57454	21.180	ug/l	98
28) Bromochloromethane	7.812	49	41181	18.703	ug/l	99
29) Tetrahydrofuran	7.836	42	104792	107.247	ug/l	99
30) Chloroform	7.959	83	99655	20.374	ug/l	99
31) Cyclohexane	8.253	56	82980	20.173	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	87659	20.431	ug/l	98
36) 1,1-Dichloropropene	8.371	75	67949	20.490	ug/l	98
37) Ethyl Acetate	7.553	43	66443	19.850	ug/l	96
38) Carbon Tetrachloride	8.359	117	78842	20.767	ug/l	96
39) Methylcyclohexane	9.594	83	65012	20.862	ug/l	96
40) Benzene	8.600	78	207985	20.873	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085440.D
 Acq On : 14 Jan 2025 15:43
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Jan 15 01:43:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	36578	21.004	ug/l	99
42) 1,2-Dichloroethane	8.665	62	78402	20.891	ug/l	99
43) Isopropyl Acetate	8.683	43	109131	20.347	ug/l	98
44) Trichloroethene	9.347	130	48009	20.699	ug/l	96
45) 1,2-Dichloropropane	9.618	63	52833	20.758	ug/l	93
46) Dibromomethane	9.700	93	37721	20.540	ug/l	98
47) Bromodichloromethane	9.882	83	78918	21.093	ug/l	96
48) Methyl methacrylate	9.677	41	50224	20.808	ug/l	96
49) 1,4-Dioxane	9.688	88	17140	421.686	ug/l	95
51) 4-Methyl-2-Pentanone	10.441	43	336911	108.249	ug/l	100
52) Toluene	10.630	92	125172	21.680	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	74129	20.965	ug/l	96
54) cis-1,3-Dichloropropene	10.306	75	81897	21.684	ug/l	99
55) 1,1,2-Trichloroethane	11.012	97	48039	21.025	ug/l	97
56) Ethyl methacrylate	10.871	69	70934	19.649	ug/l	99
57) 1,3-Dichloropropane	11.159	76	84210	21.196	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	153681	105.991	ug/l	100
59) 2-Hexanone	11.194	43	240595	109.863	ug/l	98
60) Dibromochloromethane	11.353	129	56069	20.326	ug/l	96
61) 1,2-Dibromoethane	11.465	107	48465	21.309	ug/l	96
64) Tetrachloroethene	11.100	164	43243	21.408	ug/l	96
65) Chlorobenzene	11.888	112	136809	21.077	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	48370	20.304	ug/l	98
67) Ethyl Benzene	11.959	91	229978	21.753	ug/l	98
68) m/p-Xylenes	12.071	106	177755	45.494	ug/l	100
69) o-Xylene	12.394	106	84548	22.643	ug/l	98
70) Styrene	12.412	104	140581	22.749	ug/l	99
71) Bromoform	12.576	173	36961	21.683	ug/l #	96
73) Isopropylbenzene	12.694	105	211354	21.819	ug/l	100
74) N-amyl acetate	12.494	43	94292	21.656	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	68151	19.917	ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	90004	30.902	ug/l	46
77) Bromobenzene	12.976	156	52022	20.556	ug/l	99
78) n-propylbenzene	13.035	91	252610	22.028	ug/l	100
79) 2-Chlorotoluene	13.123	91	158511	21.358	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	180294	22.533	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	25146	23.322	ug/l #	82
82) 4-Chlorotoluene	13.218	91	161880	21.906	ug/l	100
83) tert-Butylbenzene	13.435	119	147541	21.958	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	183144	22.965	ug/l	99
85) sec-Butylbenzene	13.612	105	209133	22.457	ug/l	100
86) p-Isopropyltoluene	13.729	119	168888	21.661	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	97652	20.952	ug/l	100
88) 1,4-Dichlorobenzene	13.806	146	98321	20.351	ug/l	99
89) n-Butylbenzene	14.053	91	142673	21.355	ug/l	99
90) Hexachloroethane	14.329	117	35205	19.858	ug/l	98
91) 1,2-Dichlorobenzene	14.100	146	94954	20.425	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12859	20.580	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	45878	21.233	ug/l	99
94) Hexachlorobutadiene	15.500	225	22719	19.805	ug/l	95
95) Naphthalene	15.635	128	134793	20.906	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	45468	20.797	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085440.D
Acq On : 14 Jan 2025 15:43
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:168 Resp: 200796
Ion Ratio Lower Upper
168 100
99 62.2 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 20.107 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 85 Resp: 54666
Ion Ratio Lower Upper
85 100
87 32.5 16.7 50.1









#3
 Chloromethane
 Concen: 19.848 ug/l
 RT: 2.359 min Scan# 69
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion: 50 Resp: 58426
 Ion Ratio Lower Upper
 50 100
 52 34.0 27.0 40.6



#4
 Vinyl Chloride
 Concen: 19.739 ug/l
 RT: 2.512 min Scan# 95
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 62 Resp: 58403
 Ion Ratio Lower Upper
 62 100
 64 30.7 24.8 37.2









#5
Bromomethane
Concen: 20.389 ug/l
RT: 2.954 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 94 Resp: 36440
Ion Ratio Lower Upper
94 100
96 94.9 71.8 107.6



#6
Chloroethane
Concen: 20.041 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 64 Resp: 37594
Ion Ratio Lower Upper
64 100
66 34.3 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 20.526 ug/l
 RT: 3.501 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion:101 Resp: 88121
 Ion Ratio Lower Upper
 101 100
 103 65.3 51.4 77.2



#8
 Diethyl Ether
 Concen: 19.647 ug/l
 RT: 3.965 min Scan# 342
 Delta R.T. 0.012 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 74 Resp: 29138
 Ion Ratio Lower Upper
 74 100
 45 101.8 49.7 149.1









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 20.237 ug/l
 RT: 4.371 min Scan# 411
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion:101 Resp: 48933
 Ion Ratio Lower Upper
 101 100
 85 46.6 37.8 56.8
 151 72.1 58.8 88.2



#10
 Methyl Iodide
 Concen: 20.428 ug/l
 RT: 4.589 min Scan# 448
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion:142 Resp: 56567
 Ion Ratio Lower Upper
 142 100
 127 44.6 40.3 60.5
 141 14.8 12.8 19.2









#11
Tert butyl alcohol
Concen: 102.342 ug/l
RT: 5.506 min Scan# 605
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 59 Resp: 37984
Ion Ratio Lower Upper
59 100
57 10.7 8.6 13.0



#12
1,1-Dichloroethene
Concen: 20.728 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 96 Resp: 44670
Ion Ratio Lower Upper
96 100
61 170.3 141.0 211.4
98 56.8 49.0 73.6





#13

Acrolein

Concen: 97.995 ug/l

RT: 4.171 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Tgt Ion: 56 Resp: 49649

Ion Ratio Lower Upper

56 100

55 71.4 56.3 84.5



#14

Allyl chloride

Concen: 19.362 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 41 Resp: 67705


Ion Ratio Lower Upper

41 100

39 85.0 64.4 96.6

76 38.6 30.5 45.7









#15
Acrylonitrile
Concen: 103.171 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 53 Resp: 121626
Ion Ratio Lower Upper
53 100
52 81.1 65.5 98.3
51 36.1 29.8 44.8



#16
Acetone
Concen: 96.502 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 101065
Ion Ratio Lower Upper
43 100
58 27.7 23.8 35.6









#17
Carbon Disulfide
Concen: 19.939 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 76 Resp: 132300
Ion Ratio Lower Upper
76 100
78 9.3 6.9 10.3



#18
Methyl Acetate
Concen: 19.115 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 60873
Ion Ratio Lower Upper
43 100
74 24.0 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 21.272 ug/l
RT: 5.795 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 73 Resp: 148847
Ion Ratio Lower Upper
73 100
57 22.1 17.8 26.8



#20
Methylene Chloride
Concen: 20.370 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 84 Resp: 52813
Ion Ratio Lower Upper
84 100
49 129.2 104.1 156.1
51 42.2 31.0 46.4
86 61.6 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 20.023 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 96 Resp: 46115
Ion Ratio Lower Upper
96 100
61 146.1 115.7 173.5
98 64.5 49.8 74.8



#22
Diisopropyl ether
Concen: 21.480 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 45 Resp: 166719
Ion Ratio Lower Upper
45 100
43 54.3 45.0 67.6
87 25.3 21.8 32.8
59 10.8 10.2 15.2









#23
 Vinyl Acetate
 Concen: 122.364 ug/l
 RT: 6.594 min Scan# 789
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion: 43 Resp: 666942
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 20.469 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 63 Resp: 96875
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.3 9.8
 100 6.0 2.0 6.0#









#25
2-Butanone
Concen: 103.786 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 159952
Ion Ratio Lower Upper
43 100
72 24.1 20.2 30.4



#26
2,2-Dichloropropane
Concen: 20.684 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 77 Resp: 79145
Ion Ratio Lower Upper
77 100
97 22.0 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 21.180 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Tgt Ion: 96 Resp: 57454




Ion	Ratio	Lower	Upper
96	100		
61	153.0	0.0	311.8
98	63.7	0.0	126.0




#28
 Bromochloromethane
 Concen: 18.703 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 49 Resp: 41181

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.4
130	69.3	55.0	82.4





#29

Tetrahydrofuran

Concen: 107.247 ug/l

RT: 7.836 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion: 42 Resp: 104792

Ion Ratio Lower Upper

42 100

72 42.6 34.2 51.2

71 41.4 32.5 48.7



#30

Chloroform

Concen: 20.374 ug/l

RT: 7.959 min Scan# 1021

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 83 Resp: 99655

Ion Ratio Lower Upper

83 100

85 65.8 51.8 77.6









#31
Cyclohexane
Concen: 20.173 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:	56	Resp:	82980
Ion Ratio	Lower	Upper	
56	100		
69	31.2	22.2	33.4
84	80.8	66.4	99.6



#32
1,1,1-Trichloroethane
Concen: 20.431 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:	97	Resp:	87659
Ion Ratio	Lower	Upper	
97	100		
99	64.2	49.8	74.6
61	50.3	41.4	62.2





#33

1,2-Dichloroethane-d4

Concen: 18.694 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Tgt Ion: 65 Resp: 60593

Ion Ratio Lower Upper

65 100

67 51.2 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 114 Resp: 340593


Ion Ratio Lower Upper

114 100

63 21.6 0.0 47.6

88 16.0 0.0 32.6









#35
Dibromofluoromethane
Concen: 19.322 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:113 Resp: 45655
Ion Ratio Lower Upper
113 100
111 100.6 82.7 124.1
192 17.9 14.3 21.5



#36
1,1-Dichloropropene
Concen: 20.490 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 67949
Ion Ratio Lower Upper
75 100
110 31.8 16.5 49.5
77 31.1 24.4 36.6









#37
Ethyl Acetate
Concen: 19.850 ug/l
RT: 7.553 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 66443
Ion Ratio Lower Upper
43 100
61 11.9 10.9 16.3
70 10.6 7.5 11.3



#38
Carbon Tetrachloride
Concen: 20.767 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 117 Resp: 78842
Ion Ratio Lower Upper
117 100
119 98.9 75.4 113.2
121 28.9 24.6 37.0





#39

Methylcyclohexane

Concen: 20.862 ug/l

RT: 9.594 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion: 83 Resp: 65012

Ion Ratio Lower Upper

83 100

55 83.4 62.6 94.0

98 46.4 37.7 56.5



#40

Benzene

Concen: 20.873 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 78 Resp: 207985

Ion Ratio Lower Upper

78 100

77 24.4 19.0 28.6









#41
Methacrylonitrile
Concen: 21.004 ug/l
RT: 7.771 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:	41	Resp:	36578
Ion	Ratio	Lower	Upper
41	100		
39	56.6	46.0	69.0
67	70.3	57.4	86.2
52	32.4	25.5	38.3



#42
1,2-Dichloroethane
Concen: 20.891 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:	62	Resp:	78402
Ion	Ratio	Lower	Upper
62	100		
98	8.3	0.0	17.0









#43
Isopropyl Acetate
Concen: 20.347 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 109131
Ion Ratio Lower Upper
43 100
61 27.2 20.7 31.1
87 12.3 9.8 14.8



#44
Trichloroethene
Concen: 20.699 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:130 Resp: 48009
Ion Ratio Lower Upper
130 100
95 101.5 0.0 195.8





#45

1,2-Dichloropropane

Concen: 20.758 ug/l

RT: 9.618 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion: 63 Resp: 52833

Ion Ratio Lower Upper

63 100

65 28.3 25.6 38.4



#46

Dibromomethane

Concen: 20.540 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 93 Resp: 37721


Ion Ratio Lower Upper

93 100

95 82.0 64.7 97.1

174 88.4 69.0 103.4









#47
Bromodichloromethane
Concen: 21.093 ug/l
RT: 9.882 min Scan# 1348
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 83 Resp: 78918
Ion Ratio Lower Upper
83 100
85 67.8 51.2 76.8
127 8.6 6.5 9.7



#48
Methyl methacrylate
Concen: 20.808 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 41 Resp: 50224
Ion Ratio Lower Upper
41 100
69 84.3 64.7 97.1
39 64.2 49.0 73.6





#49

1,4-Dioxane

Concen: 421.686 ug/l

RT: 9.688 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion: 88 Resp: 17140

Ion Ratio Lower Upper

88 100

43 34.9 26.6 39.8

58 69.0 59.5 89.3



#50

Toluene-d8

Concen: 19.581 ug/l

RT: 10.559 min Scan# 1463

Delta R.T. -0.006 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion: 98 Resp: 164391

Ion Ratio Lower Upper

98 100

100 64.4 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 108.249 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 336911
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7



#52
Toluene
Concen: 21.680 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 92 Resp: 125172
Ion Ratio Lower Upper
92 100
91 175.1 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 20.965 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion: 75 Resp: 74129
 Ion Ratio Lower Upper
 75 100
 77 31.4 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 21.684 ug/l
 RT: 10.306 min Scan# 1420
 Delta R.T. -0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 81897
 Ion Ratio Lower Upper
 75 100
 77 32.4 25.0 37.4
 39 54.4 43.1 64.7









#55
1,1,2-Trichloroethane
Concen: 21.025 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:	97	Resp:	48039
Ion	Ratio	Lower	Upper
97	100		
83	84.1	69.0	103.6
85	54.6	44.6	66.8
99	64.8	48.6	72.8



#56
Ethyl methacrylate
Concen: 19.649 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:	69	Resp:	70934
Ion	Ratio	Lower	Upper
69	100		
41	67.6	54.6	82.0
39	41.7	32.4	48.6









#57
1,3-Dichloropropane
Concen: 21.196 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 76 Resp: 84210
Ion Ratio Lower Upper
76 100
78 32.7 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 105.991 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 63 Resp: 153681
Ion Ratio Lower Upper
63 100
106 27.2 21.6 32.4









#59
2-Hexanone
Concen: 109.863 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 240595
Ion Ratio Lower Upper
43 100
58 51.1 26.2 78.6



#60
Dibromochloromethane
Concen: 20.326 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 129 Resp: 56069
Ion Ratio Lower Upper
129 100
127 80.4 38.6 115.8









#61
1,2-Dibromoethane
Concen: 21.309 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:107 Resp: 48465
Ion Ratio Lower Upper
107 100
109 90.6 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 19.430 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 95 Resp: 55798
Ion Ratio Lower Upper
95 100
174 75.2 0.0 145.0
176 73.2 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:117 Resp: 296298
Ion Ratio Lower Upper
117 100
82 59.3 48.6 72.8
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 21.408 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:164 Resp: 43243
Ion Ratio Lower Upper
164 100
166 130.6 103.4 155.2
129 107.4 79.2 118.8
131 94.7 77.1 115.7





#65

Chlorobenzene

Concen: 21.077 ug/l

RT: 11.888 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion:112 Resp: 136809

Ion Ratio Lower Upper

112 100

114 32.9 25.3 37.9



#66

1,1,1,2-Tetrachloroethane

Concen: 20.304 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion:131 Resp: 48370


Ion Ratio Lower Upper

131 100

133 96.9 47.4 142.3

119 67.2 33.1 99.5









#67
Ethyl Benzene
Concen: 21.753 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 91 Resp: 229978
Ion Ratio Lower Upper
91 100
106 31.1 23.8 35.8



#68
m/p-Xylenes
Concen: 45.494 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:106 Resp: 177755
Ion Ratio Lower Upper
106 100
91 209.3 167.7 251.5









#69
o-Xylene
Concen: 22.643 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:106 Resp: 84548
Ion Ratio Lower Upper
106 100
91 217.0 110.4 331.2



#70
Styrene
Concen: 22.749 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:104 Resp: 140581
Ion Ratio Lower Upper
104 100
78 54.4 42.5 63.7
103 54.8 43.8 65.8





#71

Bromoform

Concen: 21.683 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:173 Resp: 36961

Ion Ratio Lower Upper

173 100

175 51.7 24.4 73.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion:152 Resp: 143526


Ion Ratio Lower Upper

152 100

115 61.4 31.1 93.3

150 163.9 0.0 343.6









#73
Isopropylbenzene
Concen: 21.819 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 211354
Ion Ratio Lower Upper
105 100
120 25.5 12.8 38.3



#74
N-ethyl acetate
Concen: 21.656 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 94292
Ion Ratio Lower Upper
43 100
70 41.4 34.0 51.0
55 27.9 21.4 32.2
61 23.1 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 19.917 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion: 83 Resp: 68151
 Ion Ratio Lower Upper
 83 100
 131 9.9 4.8 14.4
 85 65.5 32.2 96.6



#76
 1,2,3-Trichloropropane
 Concen: 30.902 ug/l
 RT: 12.988 min Scan# 1876
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 90004
 Ion Ratio Lower Upper
 75 100
 77 133.2 109.7 329.2









#77
Bromobenzene
Concen: 20.556 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:156 Resp: 52022
Ion Ratio Lower Upper
156 100
77 230.4 114.1 342.4
158 97.0 48.9 146.8



#78
n-propylbenzene
Concen: 22.028 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 91 Resp: 252610
Ion Ratio Lower Upper
91 100
120 21.8 10.9 32.6









#79
2-Chlorotoluene
Concen: 21.358 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 91 Resp: 158511
Ion Ratio Lower Upper
91 100
126 31.5 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 22.533 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:105 Resp: 180294
Ion Ratio Lower Upper
105 100
120 47.9 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 23.322 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 75 Resp: 25146
Ion Ratio Lower Upper
75 100
53 94.6 95.6 143.4#
89 41.3 37.0 55.6



#82
4-Chlorotoluene
Concen: 21.906 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 91 Resp: 161880
Ion Ratio Lower Upper
91 100
126 31.6 15.9 47.7





#83

tert-Butylbenzene

Concen: 21.958 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:119 Resp: 147541

Ion Ratio Lower Upper

119 100

91 69.5 35.1 105.3

134 25.5 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 22.965 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085440.D




Acq: 14 Jan 2025 15:43


Tgt Ion:105 Resp: 183144

Ion Ratio Lower Upper

105 100

120 42.8 21.6 65.0









#85
sec-Butylbenzene
Concen: 22.457 ug/l
RT: 13.612 min Scan# 1982
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 209133
Ion Ratio Lower Upper
105 100
134 19.2 9.7 28.9



#86
p-Isopropyltoluene
Concen: 21.661 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:119 Resp: 168888
Ion Ratio Lower Upper
119 100
134 26.4 12.7 38.0
91 25.5 12.7 38.1









#87
 1,3-Dichlorobenzene
 Concen: 20.952 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion:146 Resp: 97652
 Ion Ratio Lower Upper
 146 100
 111 42.9 21.4 64.3
 148 64.3 32.3 96.9



#88
 1,4-Dichlorobenzene
 Concen: 20.351 ug/l
 RT: 13.806 min Scan# 2015
 Delta R.T. -0.006 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion:146 Resp: 98321
 Ion Ratio Lower Upper
 146 100
 111 42.2 21.3 63.7
 148 65.1 32.4 97.0









#89
n-Butylbenzene
Concen: 21.355 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 142673
Ion Ratio Lower Upper
91 100
92 51.0 25.8 77.3
134 23.0 11.7 35.1



#90
Hexachloroethane
Concen: 19.858 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 117 Resp: 35205
Ion Ratio Lower Upper
117 100
201 68.8 33.7 101.0









#91
 1,2-Dichlorobenzene
 Concen: 20.425 ug/l
 RT: 14.100 min Scan# 2065
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion:146	Resp:	94954
Ion Ratio	Lower	Upper
146	100	
111	43.5	21.7 65.1
148	63.5	31.4 94.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 20.580 ug/l
 RT: 14.717 min Scan# 2170
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 75	Resp:	12859
Ion Ratio	Lower	Upper
75	100	
155	72.7	36.4 109.2
157	90.9	45.4 136.1









#93
1,2,4-Trichlorobenzene
Concen: 21.233 ug/l
RT: 15.388 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:180 Resp: 45878
Ion Ratio Lower Upper
180 100
182 94.2 46.8 140.3
145 31.4 16.0 48.0



#94
Hexachlorobutadiene
Concen: 19.805 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:225 Resp: 22719
Ion Ratio Lower Upper
225 100
223 64.3 30.7 92.1
227 66.3 30.9 92.5





#95

Naphthalene

Concen: 20.906 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:128 Resp: 134793

Ion Ratio Lower Upper

128 100

127 13.3 10.6 16.0

129 10.6 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 20.797 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion:180 Resp: 45468

Ion Ratio Lower Upper

180 100

182 97.9 47.4 142.2

145 35.7 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085441.D
 Acq On : 14 Jan 2025 16:07
 Operator : JC\MD
 Sample : VSTDIC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010

Quant Time: Jan 15 01:44:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	195889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	340403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	29856	9.442	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	18.880%#
35) Dibromofluoromethane	8.165	113	21110	8.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	17.880%#
50) Toluene-d8	10.565	98	73223	8.727	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	17.460%#
62) 4-Bromofluorobenzene	12.847	95	24327	8.476	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	16.960%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	26124	9.850	ug/l	96
3) Chloromethane	2.359	50	27143	9.452	ug/l	99
4) Vinyl Chloride	2.512	62	27862	9.653	ug/l	100
5) Bromomethane	2.965	94	17139	9.830	ug/l	87
6) Chloroethane	3.130	64	16791	9.175	ug/l	96
7) Trichlorofluoromethane	3.501	101	40755	9.731	ug/l	98
8) Diethyl Ether	3.953	74	12415	8.581	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.371	101	22982	9.742	ug/l	98
10) Methyl Iodide	4.589	142	25085	9.286	ug/l	99
11) Tert butyl alcohol	5.512	59	17962	49.608	ug/l	97
12) 1,1-Dichloroethene	4.342	96	20615	9.806	ug/l	99
13) Acrolein	4.177	56	25568	51.729	ug/l	99
14) Allyl chloride	5.018	41	32240	9.451	ug/l	97
15) Acrylonitrile	5.712	53	53959	46.918	ug/l	98
16) Acetone	4.424	43	48475	47.446	ug/l	98
17) Carbon Disulfide	4.712	76	60201	9.300	ug/l #	94
18) Methyl Acetate	5.030	43	30500	9.817	ug/l	98
19) Methyl tert-butyl Ether	5.789	73	65207	9.552	ug/l	91
20) Methylene Chloride	5.277	84	23723	9.379	ug/l	96
21) trans-1,2-Dichloroethene	5.783	96	21127	9.403	ug/l	97
22) Diisopropyl ether	6.671	45	71857	9.490	ug/l	97
23) Vinyl Acetate	6.600	43	250575	47.125	ug/l	98
24) 1,1-Dichloroethane	6.565	63	43099	9.335	ug/l	95
25) 2-Butanone	7.477	43	71191	47.350	ug/l	99
26) 2,2-Dichloropropane	7.483	77	36105	9.672	ug/l	99
27) cis-1,2-Dichloroethene	7.489	96	25049	9.466	ug/l	97
28) Bromochloromethane	7.806	49	19037	8.863	ug/l	97
29) Tetrahydrofuran	7.836	42	45641	47.881	ug/l	98
30) Chloroform	7.965	83	45792	9.596	ug/l	100
31) Cyclohexane	8.253	56	40211	10.020	ug/l #	93
32) 1,1,1-Trichloroethane	8.165	97	39170	9.358	ug/l	99
36) 1,1-Dichloropropene	8.371	75	30015	9.056	ug/l	97
37) Ethyl Acetate	7.559	43	30760	9.195	ug/l	97
38) Carbon Tetrachloride	8.359	117	36015	9.492	ug/l	94
39) Methylcyclohexane	9.594	83	27721	8.900	ug/l	95
40) Benzene	8.600	78	93680	9.407	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085441.D
 Acq On : 14 Jan 2025 16:07
 Operator : JC\MD
 Sample : VSTDICC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC010

Quant Time: Jan 15 01:44:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	16859	9.686	ug/l	96
42) 1,2-Dichloroethane	8.665	62	35569	9.483	ug/l	99
43) Isopropyl Acetate	8.683	43	48926	9.127	ug/l	99
44) Trichloroethene	9.347	130	21113	9.108	ug/l	95
45) 1,2-Dichloropropane	9.618	63	22747	8.942	ug/l	97
46) Dibromomethane	9.706	93	17260	9.404	ug/l	99
47) Bromodichloromethane	9.883	83	35020	9.365	ug/l	98
48) Methyl methacrylate	9.677	41	21873	9.067	ug/l	97
49) 1,4-Dioxane	9.694	88	7734	190.382	ug/l	98
51) 4-Methyl-2-Pentanone	10.441	43	146990	47.254	ug/l	98
52) Toluene	10.630	92	54976	9.527	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	32780	9.276	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	35904	9.512	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	21060	9.222	ug/l	98
56) Ethyl methacrylate	10.871	69	28454	9.042	ug/l	95
57) 1,3-Dichloropropane	11.159	76	37768	9.512	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	63110	43.550	ug/l	98
59) 2-Hexanone	11.194	43	101318	46.291	ug/l	98
60) Dibromochloromethane	11.353	129	25029	9.079	ug/l	98
61) 1,2-Dibromoethane	11.465	107	21055	9.263	ug/l	100
64) Tetrachloroethene	11.100	164	19934	9.928	ug/l	97
65) Chlorobenzene	11.888	112	61680	9.560	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.959	131	21853	9.229	ug/l	98
67) Ethyl Benzene	11.959	91	99248	9.445	ug/l	97
68) m/p-Xylenes	12.065	106	72431	18.650	ug/l	99
69) o-Xylene	12.394	106	34382	9.264	ug/l	99
70) Styrene	12.406	104	56285	9.163	ug/l	98
71) Bromoform	12.576	173	16071	9.485	ug/l #	98
73) Isopropylbenzene	12.694	105	87003	9.697	ug/l	99
74) N-amyl acetate	12.494	43	37343	9.260	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	30770	9.709	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	39589	14.675	ug/l	43
77) Bromobenzene	12.976	156	22784	9.720	ug/l	97
78) n-propylbenzene	13.035	91	102262	9.628	ug/l	98
79) 2-Chlorotoluene	13.123	91	68346	9.943	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	71656	9.669	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	9236	9.248	ug/l	100
82) 4-Chlorotoluene	13.218	91	66773	9.756	ug/l	97
83) tert-Butylbenzene	13.435	119	59455	9.553	ug/l	98
84) 1,2,4-Trimethylbenzene	13.476	105	72046	9.754	ug/l	99
85) sec-Butylbenzene	13.612	105	83662	9.699	ug/l	99
86) p-Isopropyltoluene	13.729	119	67556	9.674	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	41843	9.693	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	42728	9.549	ug/l	99
89) n-Butylbenzene	14.053	91	57632	9.314	ug/l	97
90) Hexachloroethane	14.329	117	15261	9.294	ug/l	97
91) 1,2-Dichlorobenzene	14.106	146	40726	9.458	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5639	9.744	ug/l	95
93) 1,2,4-Trichlorobenzene	15.388	180	18720	9.354	ug/l	99
94) Hexachlorobutadiene	15.500	225	10491	9.874	ug/l	98
95) Naphthalene	15.635	128	52054	8.717	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	19450	9.605	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085441.D
Acq On : 14 Jan 2025 16:07
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Quant Time: Jan 15 01:44:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:168 Resp: 195889
Ion Ratio Lower Upper
168 100
99 61.7 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 9.850 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 85 Resp: 26124
Ion Ratio Lower Upper
85 100
87 35.8 16.7 50.1









#3
Chloromethane
Concen: 9.452 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 50 Resp: 27143
Ion Ratio Lower Upper
50 100
52 34.3 27.0 40.6



#4
Vinyl Chloride
Concen: 9.653 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 62 Resp: 27862
Ion Ratio Lower Upper
62 100
64 31.2 24.8 37.2









#5
Bromomethane
Concen: 9.830 ug/l
RT: 2.965 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 94 Resp: 17139
Ion Ratio Lower Upper
94 100
96 102.1 71.8 107.6



#6
Chloroethane
Concen: 9.175 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 64 Resp: 16791
Ion Ratio Lower Upper
64 100
66 33.1 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 9.731 ug/l
 RT: 3.501 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010


Tgt Ion:101 Resp: 40755
 Ion Ratio Lower Upper
 101 100
 103 65.9 51.4 77.2



#8
 Diethyl Ether
 Concen: 8.581 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07

Tgt Ion: 74 Resp: 12415
 Ion Ratio Lower Upper
 74 100
 45 105.3 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 9.742 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:101 Resp: 22982
Ion Ratio Lower Upper
101 100
85 45.8 37.8 56.8
151 75.2 58.8 88.2



#10
Methyl Iodide
Concen: 9.286 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:142 Resp: 25085
Ion Ratio Lower Upper
142 100
127 50.9 40.3 60.5
141 15.0 12.8 19.2









#11
Tert butyl alcohol
Concen: 49.608 ug/l
RT: 5.512 min Scan# 605
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 59 Resp: 17962
Ion Ratio Lower Upper
59 100
57 11.9 8.6 13.0



#12
1,1-Dichloroethene
Concen: 9.806 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 96 Resp: 20615
Ion Ratio Lower Upper
96 100
61 176.8 141.0 211.4
98 60.5 49.0 73.6









#13
Acrolein
Concen: 51.729 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010


Tgt Ion: 56 Resp: 25568
Ion Ratio Lower Upper
56 100
55 70.9 56.3 84.5



#14
Allyl chloride
Concen: 9.451 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 41 Resp: 32240
Ion Ratio Lower Upper
41 100
39 78.3 64.4 96.6
76 35.1 30.5 45.7









#15
Acrylonitrile
Concen: 46.918 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010


Tgt Ion:	53	Resp:	53959
Ion	Ratio	Lower	Upper
53	100		
52	83.3	65.5	98.3
51	38.8	29.8	44.8



#16
Acetone
Concen: 47.446 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:	43	Resp:	48475
Ion	Ratio	Lower	Upper
43	100		
58	30.8	23.8	35.6









#17
Carbon Disulfide
Concen: 9.300 ug/l
RT: 4.712 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 76 Resp: 60201
Ion Ratio Lower Upper
76 100
78 10.7 6.9 10.3#



#18
Methyl Acetate
Concen: 9.817 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 30500
Ion Ratio Lower Upper
43 100
74 24.3 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 9.552 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 73 Resp: 65207
Ion Ratio Lower Upper
73 100
57 18.1 17.8 26.8



#20
Methylene Chloride
Concen: 9.379 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 84 Resp: 23723
Ion Ratio Lower Upper
84 100
49 132.4 104.1 156.1
51 34.4 31.0 46.4
86 70.0 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 9.403 ug/l
RT: 5.783 min Scan# 651
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 96 Resp: 21127
Ion Ratio Lower Upper
96 100
61 140.8 115.7 173.5
98 60.5 49.8 74.8



#22
Diisopropyl ether
Concen: 9.490 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 45 Resp: 71857
Ion Ratio Lower Upper
45 100
43 53.2 45.0 67.6
87 28.8 21.8 32.8
59 13.2 10.2 15.2









#23
 Vinyl Acetate
 Concen: 47.125 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010


Tgt Ion: 43 Resp: 250575
 Ion Ratio Lower Upper
 43 100
 86 9.8 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 9.335 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07

Tgt Ion: 63 Resp: 43099
 Ion Ratio Lower Upper
 63 100
 98 8.8 3.3 9.8
 100 5.0 2.0 6.0









#25
2-Butanone
Concen: 47.350 ug/l
RT: 7.477 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 71191
Ion Ratio Lower Upper
43 100
72 24.8 20.2 30.4



#26
2,2-Dichloropropane
Concen: 9.672 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 77 Resp: 36105
Ion Ratio Lower Upper
77 100
97 20.8 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 9.466 ug/l
 RT: 7.489 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010


Tgt Ion: 96 Resp: 25049
 Ion Ratio Lower Upper
 96 100
 61 151.8 0.0 311.8
 98 64.7 0.0 126.0



#28
 Bromochloromethane
 Concen: 8.863 ug/l
 RT: 7.806 min Scan# 995
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07

Tgt Ion: 49 Resp: 19037
 Ion Ratio Lower Upper
 49 100
 129 0.3 0.0 4.4
 130 70.8 55.0 82.4









#29
Tetrahydrofuran
Concen: 47.881 ug/l
RT: 7.836 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 42 Resp: 45641
Ion Ratio Lower Upper
42 100
72 41.6 34.2 51.2
71 39.2 32.5 48.7



#30
Chloroform
Concen: 9.596 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 83 Resp: 45792
Ion Ratio Lower Upper
83 100
85 64.9 51.8 77.6









#31
Cyclohexane
Concen: 10.020 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 56 Resp: 40211
Ion Ratio Lower Upper
56 100
69 34.2 22.2 33.4#
84 87.2 66.4 99.6



#32
1,1,1-Trichloroethane
Concen: 9.358 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 97 Resp: 39170
Ion Ratio Lower Upper
97 100
99 62.6 49.8 74.6
61 50.2 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 9.442 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion: 65 Resp: 29856

Ion Ratio Lower Upper

65 100

67 49.5 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 114 Resp: 340403


Ion Ratio Lower Upper

114 100

63 21.1 0.0 47.6

88 15.5 0.0 32.6





#35

Dibromofluoromethane

Concen: 8.939 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:113 Resp: 21110

Ion Ratio Lower Upper

113 100

111 103.4 82.7 124.1

192 16.4 14.3 21.5



#36

1,1-Dichloropropene

Concen: 9.056 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 75 Resp: 30015


Ion Ratio Lower Upper

75 100

110 34.3 16.5 49.5

77 32.9 24.4 36.6









#37
Ethyl Acetate
Concen: 9.195 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 30760
Ion Ratio Lower Upper
43 100
61 12.2 10.9 16.3
70 8.4 7.5 11.3



#38
Carbon Tetrachloride
Concen: 9.492 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 117 Resp: 36015
Ion Ratio Lower Upper
117 100
119 101.2 75.4 113.2
121 29.1 24.6 37.0





#39

Methylcyclohexane

Concen: 8.900 ug/l

RT: 9.594 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010





Tgt Ion: 83 Resp: 27721

Ion Ratio Lower Upper

83 100

55 83.8 62.6 94.0

98 45.8 37.7 56.5



#40

Benzene

Concen: 9.407 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 78 Resp: 93680

Ion Ratio Lower Upper

78 100

77 22.0 19.0 28.6









#41
Methacrylonitrile
Concen: 9.686 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:	41	Resp:	16859
Ion	Ratio	Lower	Upper
41	100		
39	57.3	46.0	69.0
67	66.2	57.4	86.2
52	30.4	25.5	38.3



#42
1,2-Dichloroethane
Concen: 9.483 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:	62	Resp:	35569
Ion	Ratio	Lower	Upper
62	100		
98	8.2	0.0	17.0









#43
Isopropyl Acetate
Concen: 9.127 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010


Tgt Ion: 43 Resp: 48926
Ion Ratio Lower Upper
43 100
61 26.2 20.7 31.1
87 12.8 9.8 14.8



#44
Trichloroethene
Concen: 9.108 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:130 Resp: 21113
Ion Ratio Lower Upper
130 100
95 92.9 0.0 195.8





#45

1,2-Dichloropropane

Concen: 8.942 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion: 63 Resp: 22747

Ion Ratio Lower Upper

63 100

65 30.4 25.6 38.4



#46

Dibromomethane

Concen: 9.404 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 93 Resp: 17260


Ion Ratio Lower Upper

93 100

95 81.5 64.7 97.1

174 87.8 69.0 103.4









#47
Bromodichloromethane
Concen: 9.365 ug/l
RT: 9.883 min Scan# 1348
Delta R.T. 0.001 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 83 Resp: 35020
Ion Ratio Lower Upper
83 100
85 62.6 51.2 76.8
127 6.9 6.5 9.7



#48
Methyl methacrylate
Concen: 9.067 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 41 Resp: 21873
Ion Ratio Lower Upper
41 100
69 80.0 64.7 97.1
39 65.2 49.0 73.6





#49

1,4-Dioxane

Concen: 190.382 ug/l

RT: 9.694 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion: 88 Resp: 7734

Ion Ratio Lower Upper

88 100

43 35.2 26.6 39.8

58 73.3 59.5 89.3



#50

Toluene-d8

Concen: 8.727 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 98 Resp: 73223

Ion Ratio Lower Upper

98 100

100 64.5 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 47.254 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 146990
Ion Ratio Lower Upper
43 100
58 37.1 30.5 45.7



#52
Toluene
Concen: 9.527 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 92 Resp: 54976
Ion Ratio Lower Upper
92 100
91 173.7 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 9.276 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010


Tgt Ion: 75 Resp: 32780
 Ion Ratio Lower Upper
 75 100
 77 27.2 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 9.512 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.006 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 35904
 Ion Ratio Lower Upper
 75 100
 77 30.0 25.0 37.4
 39 53.1 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 9.222 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 97 Resp: 21060





Ion Ratio Lower Upper

97 100

83 88.3 69.0 103.6

85 57.3 44.6 66.8

99 60.2 48.6 72.8



#56

Ethyl methacrylate

Concen: 9.042 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 69 Resp: 28454


Ion Ratio Lower Upper

69 100

41 72.1 54.6 82.0

39 43.4 32.4 48.6









#57
1,3-Dichloropropane
Concen: 9.512 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 76 Resp: 37768
Ion Ratio Lower Upper
76 100
78 31.1 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 43.550 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 63 Resp: 63110
Ion Ratio Lower Upper
63 100
106 26.1 21.6 32.4









#59
2-Hexanone
Concen: 46.291 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 101318
Ion Ratio Lower Upper
43 100
58 50.9 26.2 78.6



#60
Dibromochloromethane
Concen: 9.079 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 129 Resp: 25029
Ion Ratio Lower Upper
129 100
127 79.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 9.263 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:107 Resp: 21055
Ion Ratio Lower Upper
107 100
109 94.6 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 8.476 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 95 Resp: 24327
Ion Ratio Lower Upper
95 100
174 73.9 0.0 145.0
176 72.1 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1177
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:117 Resp: 294508
Ion Ratio Lower Upper
117 100
82 58.7 48.6 72.8
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 9.928 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:164 Resp: 19934
Ion Ratio Lower Upper
164 100
166 132.2 103.4 155.2
129 98.3 79.2 118.8
131 90.5 77.1 115.7





#65

Chlorobenzene

Concen: 9.560 ug/l

RT: 11.888 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion:112 Resp: 61680

Ion Ratio Lower Upper

112 100

114 27.6 25.3 37.9



#66

1,1,1,2-Tetrachloroethane

Concen: 9.229 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:131 Resp: 21853


Ion Ratio Lower Upper

131 100

133 98.5 47.4 142.3

119 66.0 33.1 99.5









#67
Ethyl Benzene
Concen: 9.445 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 91 Resp: 99248
Ion Ratio Lower Upper
91 100
106 28.2 23.8 35.8



#68
m/p-Xylenes
Concen: 18.650 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:106 Resp: 72431
Ion Ratio Lower Upper
106 100
91 211.7 167.7 251.5









#69
o-Xylene
Concen: 9.264 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:106 Resp: 34382
Ion Ratio Lower Upper
106 100
91 219.5 110.4 331.2



#70
Styrene
Concen: 9.163 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:104 Resp: 56285
Ion Ratio Lower Upper
104 100
78 53.8 42.5 63.7
103 56.8 43.8 65.8





#71

Bromoform

Concen: 9.485 ug/l

RT: 12.576 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:173 Resp: 16071

Ion Ratio Lower Upper

173 100

175 50.0 24.4 73.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:152 Resp: 132934


Ion Ratio Lower Upper

152 100

115 62.3 31.1 93.3

150 159.8 0.0 343.6









#73
Isopropylbenzene
Concen: 9.697 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:105 Resp: 87003
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.3



#74
N-ethyl acetate
Concen: 9.260 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 37343
Ion Ratio Lower Upper
43 100
70 41.4 34.0 51.0
55 28.5 21.4 32.2
61 22.0 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 9.709 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 83 Resp: 30770
Ion Ratio Lower Upper
83 100
131 9.6 4.8 14.4
85 63.6 32.2 96.6



#76
1,2,3-Trichloropropane
Concen: 14.675 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 39589
Ion Ratio Lower Upper
75 100
77 127.6 109.7 329.2









#77
Bromobenzene
Concen: 9.720 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:156 Resp: 22784
Ion Ratio Lower Upper
156 100
77 221.6 114.1 342.4
158 98.7 48.9 146.8



#78
n-propylbenzene
Concen: 9.628 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 102262
Ion Ratio Lower Upper
91 100
120 20.9 10.9 32.6









#79
2-Chlorotoluene
Concen: 9.943 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 91 Resp: 68346
Ion Ratio Lower Upper
91 100
126 31.5 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 9.669 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 105 Resp: 71656
Ion Ratio Lower Upper
105 100
120 46.8 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 9.248 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 75 Resp: 9236
Ion Ratio Lower Upper
75 100
53 119.6 95.6 143.4
89 46.6 37.0 55.6



#82
4-Chlorotoluene
Concen: 9.756 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 66773
Ion Ratio Lower Upper
91 100
126 33.6 15.9 47.7





#83

tert-Butylbenzene

Concen: 9.553 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:119 Resp: 59455

Ion Ratio Lower Upper

119 100

91 70.0 35.1 105.3

134 27.8 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 9.754 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion:105 Resp: 72046

Ion Ratio Lower Upper

105 100

120 43.8 21.6 65.0





#85

sec-Butylbenzene

Concen: 9.699 ug/l

RT: 13.612 min Scan# 1982

Delta R.T. -0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion:105 Resp: 83662

Ion Ratio Lower Upper

105 100

134 18.9 9.7 28.9



#86

p-Isopropyltoluene

Concen: 9.674 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:119 Resp: 67556


Ion Ratio Lower Upper

119 100

134 25.2 12.7 38.0

91 25.0 12.7 38.1









#87
1,3-Dichlorobenzene
Concen: 9.693 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:146 Resp: 41843
Ion Ratio Lower Upper
146 100
111 43.1 21.4 64.3
148 63.0 32.3 96.9



#88
1,4-Dichlorobenzene
Concen: 9.549 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:146 Resp: 42728
Ion Ratio Lower Upper
146 100
111 44.0 21.3 63.7
148 64.6 32.4 97.0





#89

n-Butylbenzene

Concen: 9.314 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion: 91 Resp: 57632

Ion Ratio Lower Upper

91 100

92 49.2 25.8 77.3

134 22.8 11.7 35.1



#90

Hexachloroethane

Concen: 9.294 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 117 Resp: 15261

Ion Ratio Lower Upper

117 100

201 65.0 33.7 101.0





#91

1,2-Dichlorobenzene

Concen: 9.458 ug/l

RT: 14.106 min Scan# 2065

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:146 Resp: 40726

Ion Ratio Lower Upper

146 100

111 42.9 21.7 65.1

148 62.9 31.4 94.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 9.744 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 75 Resp: 5639


Ion Ratio Lower Upper

75 100

155 67.9 36.4 109.2

157 94.1 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 9.354 ug/l

RT: 15.388 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:180 Resp: 18720

Ion Ratio Lower Upper

180 100

182 92.4 46.8 140.3

145 31.7 16.0 48.0



#94

Hexachlorobutadiene

Concen: 9.874 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:225 Resp: 10491


Ion Ratio Lower Upper

225 100

223 63.3 30.7 92.1

227 62.1 30.9 92.5









#95
Naphthalene
Concen: 8.717 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010

Tgt Ion:128 Resp: 52054
Ion Ratio Lower Upper
128 100
127 12.6 10.6 16.0
129 11.2 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 9.605 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:180 Resp: 19450
Ion Ratio Lower Upper
180 100
182 90.3 47.4 142.2
145 34.7 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085442.D
 Acq On : 14 Jan 2025 16:31
 Operator : JC\MD
 Sample : VSTDIC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005

Quant Time: Jan 15 01:45:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	186556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324413	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126086	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	17046	5.660	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	11.320%#
35) Dibromofluoromethane	8.165	113	12102	5.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	10.760%#
50) Toluene-d8	10.565	98	41328	5.168	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	10.340%#
62) 4-Bromofluorobenzene	12.847	95	13520	4.943	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	9.880%#

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	13199	5.225	ug/l	97
3) Chloromethane	2.359	50	14526	5.311	ug/l	96
4) Vinyl Chloride	2.512	62	14564	5.298	ug/l	98
5) Bromomethane	2.953	94	9788	5.895	ug/l	98
6) Chloroethane	3.124	64	9430	5.411	ug/l	99
7) Trichlorofluoromethane	3.501	101	20098	5.039	ug/l	92
8) Diethyl Ether	3.959	74	6652	4.828	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	11917	5.305	ug/l	97
10) Methyl Iodide	4.583	142	12331	4.793	ug/l	99
11) Tert butyl alcohol	5.518	59	9075	26.318	ug/l #	94
12) 1,1-Dichloroethene	4.342	96	10429	5.209	ug/l	96
13) Acrolein	4.171	56	9430	20.033	ug/l	97
14) Allyl chloride	5.030	41	16741	5.153	ug/l	97
15) Acrylonitrile	5.718	53	27869	25.445	ug/l	97
16) Acetone	4.430	43	25122	25.819	ug/l	97
17) Carbon Disulfide	4.706	76	32068	5.202	ug/l	96
18) Methyl Acetate	5.024	43	15106	5.106	ug/l	97
19) Methyl tert-butyl Ether	5.800	73	31439	4.836	ug/l	94
20) Methylene Chloride	5.271	84	12981	5.389	ug/l #	83
21) trans-1,2-Dichloroethene	5.783	96	10614	4.960	ug/l #	78
22) Diisopropyl ether	6.665	45	35785	4.962	ug/l #	96
23) Vinyl Acetate	6.600	43	139425	27.533	ug/l	96
24) 1,1-Dichloroethane	6.565	63	22865	5.200	ug/l	99
25) 2-Butanone	7.483	43	36127	25.231	ug/l	99
26) 2,2-Dichloropropane	7.488	77	18398	5.175	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	12477	4.951	ug/l	99
28) Bromochloromethane	7.806	49	11096	5.424	ug/l	99
29) Tetrahydrofuran	7.841	42	21983	24.215	ug/l	99
30) Chloroform	7.959	83	23384	5.146	ug/l	98
31) Cyclohexane	8.247	56	22343	5.846	ug/l #	92
32) 1,1,1-Trichloroethane	8.159	97	21416	5.372	ug/l	91
36) 1,1-Dichloropropene	8.371	75	15886	5.029	ug/l	96
37) Ethyl Acetate	7.553	43	15336	4.810	ug/l	100
38) Carbon Tetrachloride	8.359	117	18317	5.065	ug/l	96
39) Methylcyclohexane	9.600	83	14175	4.776	ug/l	98
40) Benzene	8.600	78	47833	5.040	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085442.D
 Acq On : 14 Jan 2025 16:31
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Quant Time: Jan 15 01:45:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	7380	4.449	ug/l	96
42) 1,2-Dichloroethane	8.671	62	18628	5.211	ug/l	98
43) Isopropyl Acetate	8.688	43	24792	4.853	ug/l	98
44) Trichloroethene	9.347	130	11133	5.039	ug/l	86
45) 1,2-Dichloropropane	9.618	63	12591	5.194	ug/l	92
46) Dibromomethane	9.700	93	8230	4.705	ug/l	91
47) Bromodichloromethane	9.888	83	18456	5.179	ug/l #	96
48) Methyl methacrylate	9.677	41	10574	4.599	ug/l	95
49) 1,4-Dioxane	9.694	88	3619	93.477	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	71889	24.250	ug/l	98
52) Toluene	10.629	92	27098	4.928	ug/l	98
53) t-1,3-Dichloropropene	10.829	75	17099	5.077	ug/l	95
54) cis-1,3-Dichloropropene	10.306	75	17440	4.848	ug/l	97
55) 1,1,2-Trichloroethane	11.012	97	11311	5.197	ug/l	96
56) Ethyl methacrylate	10.871	69	14094	5.627	ug/l	94
57) 1,3-Dichloropropane	11.159	76	18945	5.006	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	33210	24.047	ug/l	97
59) 2-Hexanone	11.194	43	49021	23.501	ug/l	94
60) Dibromochloromethane	11.353	129	13635	5.190	ug/l	96
61) 1,2-Dibromoethane	11.471	107	10419	4.810	ug/l	98
64) Tetrachloroethene	11.100	164	9756	5.071	ug/l	91
65) Chlorobenzene	11.888	112	31312	5.065	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	11620	5.121	ug/l	98
67) Ethyl Benzene	11.965	91	48216	4.789	ug/l	98
68) m/p-Xylenes	12.065	106	34771	9.344	ug/l	99
69) o-Xylene	12.394	106	16419	4.617	ug/l	98
70) Styrene	12.412	104	26218	4.454	ug/l	99
71) Bromoform	12.576	173	8017	4.938	ug/l #	99
73) Isopropylbenzene	12.694	105	39811	4.678	ug/l	99
74) N-amyl acetate	12.488	43	17577	4.595	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.935	83	15483	5.151	ug/l	97
76) 1,2,3-Trichloropropane	12.994	75	18624	7.279	ug/l	48
77) Bromobenzene	12.982	156	11091	4.989	ug/l	99
78) n-propylbenzene	13.035	91	45450	4.512	ug/l	99
79) 2-Chlorotoluene	13.123	91	31932	4.898	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	31891	4.537	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.729	75	4229	4.465	ug/l	90
82) 4-Chlorotoluene	13.218	91	31976	4.926	ug/l	100
83) tert-Butylbenzene	13.435	119	27597	4.675	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	32054	4.575	ug/l	98
85) sec-Butylbenzene	13.612	105	38148	4.663	ug/l	98
86) p-Isopropyltoluene	13.723	119	29856	4.584	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	20876	5.099	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	21982	5.179	ug/l	97
89) n-Butylbenzene	14.053	91	27437	4.675	ug/l	96
90) Hexachloroethane	14.329	117	7767	4.987	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	20177	4.940	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.729	75	2872	5.232	ug/l	90
93) 1,2,4-Trichlorobenzene	15.388	180	9037	4.761	ug/l	94
94) Hexachlorobutadiene	15.500	225	5563	5.520	ug/l	98
95) Naphthalene	15.641	128	25047	4.422	ug/l	96
96) 1,2,3-Trichlorobenzene	15.841	180	8737	4.549	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085442.D
Acq On : 14 Jan 2025 16:31
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Quant Time: Jan 15 01:45:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:168 Resp: 186556
Ion Ratio Lower Upper
168 100
99 65.9 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 5.225 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 85 Resp: 13199
Ion Ratio Lower Upper
85 100
87 31.5 16.7 50.1









#3
Chloromethane
Concen: 5.311 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 50 Resp: 14526
Ion Ratio Lower Upper
50 100
52 31.6 27.0 40.6



#4
Vinyl Chloride
Concen: 5.298 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 62 Resp: 14564
Ion Ratio Lower Upper
62 100
64 29.9 24.8 37.2









#5
Bromomethane
Concen: 5.895 ug/l
RT: 2.953 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 94 Resp: 9788
Ion Ratio Lower Upper
94 100
96 87.7 71.8 107.6



#6
Chloroethane
Concen: 5.411 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 64 Resp: 9430
Ion Ratio Lower Upper
64 100
66 31.4 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 5.039 ug/l
 RT: 3.501 min Scan# 209
 Delta R.T. 0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:101 Resp: 20098
 Ion Ratio Lower Upper
 101 100
 103 70.8 51.4 77.2



#8
 Diethyl Ether
 Concen: 4.828 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Tgt Ion: 74 Resp: 6652
 Ion Ratio Lower Upper
 74 100
 45 99.5 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.305 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:101 Resp: 11917
Ion Ratio Lower Upper
101 100
85 46.5 37.8 56.8
151 76.7 58.8 88.2



#10
Methyl Iodide
Concen: 4.793 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:142 Resp: 12331
Ion Ratio Lower Upper
142 100
127 50.9 40.3 60.5
141 15.0 12.8 19.2









#11
Tert butyl alcohol
Concen: 26.318 ug/l
RT: 5.518 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 59 Resp: 9075
Ion Ratio Lower Upper
59 100
57 8.5 8.6 13.0#



#12
1,1-Dichloroethene
Concen: 5.209 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 96 Resp: 10429
Ion Ratio Lower Upper
96 100
61 172.7 141.0 211.4
98 69.0 49.0 73.6





#13

Acrolein

Concen: 20.033 ug/l

RT: 4.171 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Tgt Ion: 56 Resp: 9430

Ion Ratio Lower Upper

56 100

55 73.2 56.3 84.5



#14

Allyl chloride

Concen: 5.153 ug/l

RT: 5.030 min Scan# 523

Delta R.T. 0.012 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 41 Resp: 16741


Ion Ratio Lower Upper

41 100

39 77.4 64.4 96.6

76 36.6 30.5 45.7









#15
Acrylonitrile
Concen: 25.445 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 53	Resp:	27869
Ion	Ratio	Lower Upper
53	100	
52	85.5	65.5 98.3
51	38.5	29.8 44.8



#16
Acetone
Concen: 25.819 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.012 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43	Resp:	25122
Ion	Ratio	Lower Upper
43	100	
58	27.9	23.8 35.6









#17
Carbon Disulfide
Concen: 5.202 ug/l
RT: 4.706 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 76 Resp: 32068
Ion Ratio Lower Upper
76 100
78 9.9 6.9 10.3



#18
Methyl Acetate
Concen: 5.106 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 15106
Ion Ratio Lower Upper
43 100
74 22.0 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 4.836 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.012 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 31439
Ion Ratio Lower Upper
73 100
57 25.2 17.8 26.8



#20
Methylene Chloride
Concen: 5.389 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 84 Resp: 12981
Ion Ratio Lower Upper
84 100
49 114.5 104.1 156.1
51 32.1 31.0 46.4
86 42.9 51.9 77.9#









#21
trans-1,2-Dichloroethene
Concen: 4.960 ug/l
RT: 5.783 min Scan# 651
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 96 Resp: 10614
Ion Ratio Lower Upper
96 100
61 118.1 115.7 173.5
98 44.7 49.8 74.8#



#22
Diisopropyl ether
Concen: 4.962 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 45 Resp: 35785
Ion Ratio Lower Upper
45 100
43 58.3 45.0 67.6
87 24.2 21.8 32.8
59 9.7 10.2 15.2#









#23
 Vinyl Acetate
 Concen: 27.533 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 43 Resp: 139425
 Ion Ratio Lower Upper
 43 100
 86 9.0 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 5.200 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Tgt Ion: 63 Resp: 22865
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.3 9.8
 100 4.0 2.0 6.0









#25
2-Butanone
Concen: 25.231 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 36127
Ion Ratio Lower Upper
43 100
72 24.9 20.2 30.4



#26
2,2-Dichloropropane
Concen: 5.175 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 77 Resp: 18398
Ion Ratio Lower Upper
77 100
97 21.2 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 4.951 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005

Tgt Ion: 96 Resp: 12477




Ion	Ratio	Lower	Upper
96	100		
61	154.5	0.0	311.8
98	62.8	0.0	126.0




#28
 Bromochloromethane
 Concen: 5.424 ug/l
 RT: 7.806 min Scan# 995
 Delta R.T. -0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Tgt Ion: 49 Resp: 11096

Ion	Ratio	Lower	Upper
49	100		
129	0.7	0.0	4.4
130	69.5	55.0	82.4









#29
Tetrahydrofuran
Concen: 24.215 ug/l
RT: 7.841 min Scan# 1001
Delta R.T. 0.012 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 42 Resp: 21983
Ion Ratio Lower Upper
42 100
72 41.5 34.2 51.2
71 41.0 32.5 48.7



#30
Chloroform
Concen: 5.146 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 83 Resp: 23384
Ion Ratio Lower Upper
83 100
85 63.4 51.8 77.6









#31
Cyclohexane
Concen: 5.846 ug/l
RT: 8.247 min Scan# 1070
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 56 Resp: 22343
Ion Ratio Lower Upper
56 100
69 42.3 22.2 33.4#
84 84.3 66.4 99.6



#32
1,1,1-Trichloroethane
Concen: 5.372 ug/l
RT: 8.159 min Scan# 1055
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 97 Resp: 21416
Ion Ratio Lower Upper
97 100
99 50.5 49.8 74.6
61 50.1 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 5.660 ug/l

RT: 8.577 min Scan# 1125

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 65 Resp: 17046

Ion Ratio Lower Upper

65 100

67 49.2 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 114 Resp: 324413


Ion Ratio Lower Upper

114 100

63 23.3 0.0 47.6

88 16.0 0.0 32.6





#35

Dibromofluoromethane

Concen: 5.377 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:113 Resp: 12102

Ion Ratio Lower Upper

113 100

111 107.4 82.7 124.1

192 19.7 14.3 21.5



#36

1,1-Dichloropropene

Concen: 5.029 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 75 Resp: 15886


Ion Ratio Lower Upper

75 100

110 30.3 16.5 49.5

77 29.2 24.4 36.6









#37
Ethyl Acetate
Concen: 4.810 ug/l
RT: 7.553 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 15336
Ion Ratio Lower Upper
43 100
61 13.5 10.9 16.3
70 9.2 7.5 11.3



#38
Carbon Tetrachloride
Concen: 5.065 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 117 Resp: 18317
Ion Ratio Lower Upper
117 100
119 91.1 75.4 113.2
121 27.3 24.6 37.0





#39

Methylcyclohexane

Concen: 4.776 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion: 83 Resp: 14175

Ion Ratio Lower Upper

83 100

55 76.5 62.6 94.0

98 47.7 37.7 56.5



#40

Benzene

Concen: 5.040 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 78 Resp: 47833

Ion Ratio Lower Upper

78 100

77 23.2 19.0 28.6









#41
Methacrylonitrile
Concen: 4.449 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 41	Resp:	7380	
Ion	Ratio	Lower	Upper
41	100		
39	60.5	46.0	69.0
67	74.6	57.4	86.2
52	35.3	25.5	38.3



#42
1,2-Dichloroethane
Concen: 5.211 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 62	Resp:	18628	
Ion	Ratio	Lower	Upper
62	100		
98	7.8	0.0	17.0









#43
Isopropyl Acetate
Concen: 4.853 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 24792
Ion Ratio Lower Upper
43 100
61 27.0 20.7 31.1
87 12.3 9.8 14.8



#44
Trichloroethene
Concen: 5.039 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 130 Resp: 11133
Ion Ratio Lower Upper
130 100
95 112.2 0.0 195.8





#45

1,2-Dichloropropane

Concen: 5.194 ug/l

RT: 9.618 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 63 Resp: 12591

Ion Ratio Lower Upper

63 100

65 27.8 25.6 38.4



#46

Dibromomethane

Concen: 4.705 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 93 Resp: 8230


Ion Ratio Lower Upper

93 100

95 88.7 64.7 97.1

174 95.0 69.0 103.4









#47
Bromodichloromethane
Concen: 5.179 ug/l
RT: 9.888 min Scan# 1348
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 83 Resp: 18456
Ion Ratio Lower Upper
83 100
85 67.0 51.2 76.8
127 6.1 6.5 9.7#



#48
Methyl methacrylate
Concen: 4.599 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 41 Resp: 10574
Ion Ratio Lower Upper
41 100
69 84.9 64.7 97.1
39 65.3 49.0 73.6





#49

1,4-Dioxane

Concen: 93.477 ug/l

RT: 9.694 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 88 Resp: 3619

Ion Ratio Lower Upper

88 100

43 32.3 26.6 39.8

58 77.9 59.5 89.3



#50

Toluene-d8

Concen: 5.168 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 98 Resp: 41328

Ion Ratio Lower Upper

98 100

100 60.9 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 24.250 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 71889
Ion Ratio Lower Upper
43 100
58 36.9 30.5 45.7



#52
Toluene
Concen: 4.928 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 92 Resp: 27098
Ion Ratio Lower Upper
92 100
91 176.4 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 5.077 ug/l
 RT: 10.829 min Scan# 1510
 Delta R.T. -0.006 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 75 Resp: 17099
 Ion Ratio Lower Upper
 75 100
 77 32.2 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 4.848 ug/l
 RT: 10.306 min Scan# 1420
 Delta R.T. -0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 17440
 Ion Ratio Lower Upper
 75 100
 77 27.5 25.0 37.4
 39 53.2 43.1 64.7





#55

1,1,2-Trichloroethane

Concen: 5.197 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 11311





Ion Ratio Lower Upper

97 100

83 82.7 69.0 103.6

85 51.3 44.6 66.8

99 62.4 48.6 72.8



#56

Ethyl methacrylate

Concen: 5.627 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion: 69 Resp: 14094


Ion Ratio Lower Upper

69 100

41 73.8 54.6 82.0

39 42.6 32.4 48.6





#57

1,3-Dichloropropane

Concen: 5.006 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 76 Resp: 18945

Ion Ratio Lower Upper

76 100

78 32.2 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 24.047 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 63 Resp: 33210

Ion Ratio Lower Upper

63 100

106 25.2 21.6 32.4









#59
2-Hexanone
Concen: 23.501 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 49021
Ion Ratio Lower Upper
43 100
58 48.3 26.2 78.6



#60
Dibromochloromethane
Concen: 5.190 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 129 Resp: 13635
Ion Ratio Lower Upper
129 100
127 73.9 38.6 115.8









#61
1,2-Dibromoethane
Concen: 4.810 ug/l
RT: 11.471 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:107 Resp: 10419
Ion Ratio Lower Upper
107 100
109 96.8 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 4.943 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 95 Resp: 13520
Ion Ratio Lower Upper
95 100
174 73.2 0.0 145.0
176 68.0 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:117 Resp: 282203
Ion Ratio Lower Upper
117 100
82 58.3 48.6 72.8
119 32.1 26.6 39.8



#64
Tetrachloroethene
Concen: 5.071 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:164 Resp: 9756
Ion Ratio Lower Upper
164 100
166 130.6 103.4 155.2
129 111.5 79.2 118.8
131 110.5 77.1 115.7









#65
Chlorobenzene
Concen: 5.065 ug/l
RT: 11.888 min Scan# 111
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:112 Resp: 31312
Ion Ratio Lower Upper
112 100
114 30.7 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 5.121 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:131 Resp: 11620
Ion Ratio Lower Upper
131 100
133 92.7 47.4 142.3
119 65.0 33.1 99.5









#67
Ethyl Benzene
Concen: 4.789 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 48216
Ion Ratio Lower Upper
91 100
106 28.9 23.8 35.8



#68
m/p-Xylenes
Concen: 9.344 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 106 Resp: 34771
Ion Ratio Lower Upper
106 100
91 207.9 167.7 251.5









#69
o-Xylene
Concen: 4.617 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:106 Resp: 16419
Ion Ratio Lower Upper
106 100
91 224.0 110.4 331.2



#70
Styrene
Concen: 4.454 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:104 Resp: 26218
Ion Ratio Lower Upper
104 100
78 52.6 42.5 63.7
103 54.2 43.8 65.8





#71

Bromoform

Concen: 4.938 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:173 Resp: 8017

Ion Ratio Lower Upper

173 100

175 48.3 24.4 73.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion:152 Resp: 126086


Ion Ratio Lower Upper

152 100

115 62.0 31.1 93.3

150 160.7 0.0 343.6









#73
Isopropylbenzene
Concen: 4.678 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:105 Resp: 39811
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.3



#74
N-ethyl acetate
Concen: 4.595 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 17577
Ion Ratio Lower Upper
43 100
70 39.5 34.0 51.0
55 30.8 21.4 32.2
61 22.8 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 5.151 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 83 Resp: 15483
 Ion Ratio Lower Upper
 83 100
 131 9.5 4.8 14.4
 85 67.5 32.2 96.6



#76
 1,2,3-Trichloropropane
 Concen: 7.279 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 18624
 Ion Ratio Lower Upper
 75 100
 77 136.7 109.7 329.2









#77
Bromobenzene
Concen: 4.989 ug/l
RT: 12.982 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:156 Resp: 11091
Ion Ratio Lower Upper
156 100
77 229.5 114.1 342.4
158 100.8 48.9 146.8



#78
n-propylbenzene
Concen: 4.512 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 91 Resp: 45450
Ion Ratio Lower Upper
91 100
120 21.1 10.9 32.6









#79
2-Chlorotoluene
Concen: 4.898 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 31932
Ion Ratio Lower Upper
91 100
126 32.6 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 4.537 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 105 Resp: 31891
Ion Ratio Lower Upper
105 100
120 48.1 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 4.465 ug/l
RT: 12.729 min Scan# 1832
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 75 Resp: 4229
Ion Ratio Lower Upper
75 100
53 132.6 95.6 143.4
89 49.7 37.0 55.6



#82
4-Chlorotoluene
Concen: 4.926 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 91 Resp: 31976
Ion Ratio Lower Upper
91 100
126 31.9 15.9 47.7





#83

tert-Butylbenzene

Concen: 4.675 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:119 Resp: 27597

Ion Ratio Lower Upper

119 100

91 71.1 35.1 105.3

134 24.8 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 4.575 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion:105 Resp: 32054

Ion Ratio Lower Upper

105 100

120 42.0 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 4.663 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 38148
 Ion Ratio Lower Upper
 105 100
 134 20.1 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 4.584 ug/l
 RT: 13.723 min Scan# 2001
 Delta R.T. 0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Tgt Ion:119 Resp: 29856
 Ion Ratio Lower Upper
 119 100
 134 26.5 12.7 38.0
 91 26.5 12.7 38.1









#87
 1,3-Dichlorobenzene
 Concen: 5.099 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:146 Resp: 20876
 Ion Ratio Lower Upper
 146 100
 111 42.3 21.4 64.3
 148 62.7 32.3 96.9



#88
 1,4-Dichlorobenzene
 Concen: 5.179 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. 0.000 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Tgt Ion:146 Resp: 21982
 Ion Ratio Lower Upper
 146 100
 111 47.1 21.3 63.7
 148 65.5 32.4 97.0





#89

n-Butylbenzene

Concen: 4.675 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 91 Resp: 27437

Ion Ratio Lower Upper

91 100

92 48.7 25.8 77.3

134 21.4 11.7 35.1



#90

Hexachloroethane

Concen: 4.987 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN085442.D




Acq: 14 Jan 2025 16:31


Tgt Ion: 117 Resp: 7767

Ion Ratio Lower Upper

117 100

201 66.9 33.7 101.0









#91
 1,2-Dichlorobenzene
 Concen: 4.940 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. 0.006 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:146 Resp: 20177
 Ion Ratio Lower Upper
 146 100
 111 44.0 21.7 65.1
 148 62.8 31.4 94.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 5.232 ug/l
 RT: 14.729 min Scan# 2172
 Delta R.T. 0.012 min
 Lab File: VN085442.D
 Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 2872
 Ion Ratio Lower Upper
 75 100
 155 64.2 36.4 109.2
 157 99.8 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 4.761 ug/l

RT: 15.388 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:180 Resp: 9037

Ion Ratio Lower Upper

180 100

182 100.5 46.8 140.3

145 33.0 16.0 48.0



#94

Hexachlorobutadiene

Concen: 5.520 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion:225 Resp: 5563


Ion Ratio Lower Upper

225 100

223 59.2 30.7 92.1

227 60.1 30.9 92.5





#95

Naphthalene

Concen: 4.422 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:128 Resp: 25047

Ion Ratio Lower Upper

128 100

127 14.7 10.6 16.0

129 9.1 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 4.549 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN085442.D

Acq: 14 Jan 2025 16:31




Tgt Ion:180 Resp: 8737

Ion Ratio Lower Upper

180 100

182 95.6 47.4 142.2

145 35.5 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085443.D
 Acq On : 14 Jan 2025 17:19
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Jan 15 01:46:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302938	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	99539	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	3273	1.215	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	2.440%#
35) Dibromofluoromethane	8.159	113	2258	1.074	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	2.140%#
50) Toluene-d8	10.565	98	7164	0.959	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	1.920%#
62) 4-Bromofluorobenzene	12.841	95	2428	0.951	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	1.900%#

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.124	85	2382	1.054	ug/l	91
3) Chloromethane	2.359	50	2801	1.145	ug/l	93
4) Vinyl Chloride	2.512	62	2732	1.111	ug/l #	81
5) Bromomethane	2.971	94	2279	1.534	ug/l	97
6) Chloroethane	3.124	64	1809	1.160	ug/l	94
7) Trichlorofluoromethane	3.512	101	3860	1.082	ug/l	88
8) Diethyl Ether	3.953	74	1526	1.238	ug/l	81
9) 1,1,2-Trichlorotrifluo...	4.359	101	1308	0.651	ug/l #	68
10) Methyl Iodide	4.600	142	2478	1.077	ug/l #	87
11) Tert butyl alcohol	5.512	59	1314	4.260	ug/l #	71
12) 1,1-Dichloroethene	4.330	96	1660	0.927	ug/l #	67
13) Acrolein	4.177	56	1510	3.586	ug/l	86
14) Allyl chloride	5.024	41	3179	1.094	ug/l #	85
15) Acrylonitrile	5.718	53	4967	5.070	ug/l #	70
16) Acetone	4.436	43	5102	5.862	ug/l	92
17) Carbon Disulfide	4.706	76	6601	1.197	ug/l #	87
18) Methyl Acetate	5.030	43	2908	1.099	ug/l #	80
19) Methyl tert-butyl Ether	5.789	73	5156	0.887	ug/l #	87
20) Methylene Chloride	5.265	84	2190	1.016	ug/l #	87
21) trans-1,2-Dichloroethene	5.794	96	2110	1.102	ug/l #	66
22) Diisopropyl ether	6.677	45	5691	0.882	ug/l #	88
23) Vinyl Acetate	6.606	43	22593	4.988	ug/l	98
24) 1,1-Dichloroethane	6.553	63	2873	0.730	ug/l #	84
25) 2-Butanone	7.483	43	5203	4.062	ug/l #	90
26) 2,2-Dichloropropane	7.488	77	3104	0.976	ug/l	92
27) cis-1,2-Dichloroethene	7.477	96	2185	0.969	ug/l	92
28) Bromochloromethane	7.812	49	2082	1.138	ug/l #	89
29) Tetrahydrofuran	7.835	42	3693	4.548	ug/l	98
30) Chloroform	7.965	83	4250	1.045	ug/l #	80
31) Cyclohexane	8.218	56	7152	2.092	ug/l #	1
32) 1,1,1-Trichloroethane	8.165	97	3678	1.031	ug/l #	56
36) 1,1-Dichloropropene	8.371	75	3084	1.046	ug/l #	84
37) Ethyl Acetate	7.553	43	911	0.306	ug/l #	88
38) Carbon Tetrachloride	8.359	117	3436	1.018	ug/l	90
39) Methylcyclohexane	9.600	83	2403	0.867	ug/l #	87
40) Benzene	8.606	78	8480	0.957	ug/l	91

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085443.D
 Acq On : 14 Jan 2025 17:19
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Jan 15 01:46:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	1458	0.941	ug/l #	76
42) 1,2-Dichloroethane	8.665	62	3131	0.938	ug/l	91
43) Isopropyl Acetate	8.682	43	4803	1.007	ug/l #	87
44) Trichloroethene	9.347	130	2133	1.034	ug/l	71
45) 1,2-Dichloropropane	9.618	63	2246	0.992	ug/l	99
46) Dibromomethane	9.706	93	1769	1.083	ug/l	92
47) Bromodichloromethane	9.882	83	2935	0.882	ug/l	90
48) Methyl methacrylate	9.677	41	1886	0.878	ug/l	96
49) 1,4-Dioxane	9.688	88	673	18.616	ug/l #	47
51) 4-Methyl-2-Pentanone	10.441	43	11523	4.163	ug/l	93
52) Toluene	10.618	92	4180	0.814	ug/l	93
53) t-1,3-Dichloropropene	10.835	75	2522	0.802	ug/l	92
54) cis-1,3-Dichloropropene	10.312	75	2725	0.811	ug/l #	83
55) 1,1,2-Trichloroethane	11.018	97	1901	0.935	ug/l	91
56) Ethyl methacrylate	10.871	69	1923	2.471	ug/l #	88
57) 1,3-Dichloropropane	11.165	76	3113	0.881	ug/l	95
58) 2-Chloroethyl Vinyl ether	10.153	63	4929	3.822	ug/l	91
59) 2-Hexanone	11.194	43	7905	4.058	ug/l	87
60) Dibromochloromethane	11.359	129	2339	0.953	ug/l	98
61) 1,2-Dibromoethane	11.465	107	2022	1.000	ug/l	90
64) Tetrachloroethene	11.106	164	1658	0.947	ug/l #	83
65) Chlorobenzene	11.888	112	5401	0.960	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.959	131	2185	1.058	ug/l #	64
67) Ethyl Benzene	11.959	91	7347	0.802	ug/l	94
68) m/p-Xylenes	12.070	106	5059	1.494	ug/l	100
69) o-Xylene	12.394	106	2478	0.766	ug/l	97
70) Styrene	12.406	104	3812	0.712	ug/l	99
71) Bromoform	12.582	173	1208	0.818	ug/l #	88
73) Isopropylbenzene	12.688	105	5506	0.820	ug/l	98
74) N-amyl acetate	12.494	43	2716	0.899	ug/l #	87
75) 1,1,2,2-Tetrachloroethane	12.935	83	2616	1.102	ug/l #	96
76) 1,2,3-Trichloropropane	12.982	75	3212	1.590	ug/l	42
77) Bromobenzene	12.976	156	1733	0.987	ug/l	97
78) n-propylbenzene	13.029	91	6424	0.808	ug/l	93
79) 2-Chlorotoluene	13.117	91	4483	0.871	ug/l	97
80) 1,3,5-Trimethylbenzene	13.165	105	4301	0.775	ug/l	90
81) trans-1,4-Dichloro-2-b...	12.735	75	695	0.929	ug/l #	76
82) 4-Chlorotoluene	13.223	91	4268	0.833	ug/l	99
83) tert-Butylbenzene	13.435	119	3753	0.805	ug/l	96
84) 1,2,4-Trimethylbenzene	13.482	105	3758	0.679	ug/l	93
85) sec-Butylbenzene	13.612	105	4614	0.714	ug/l	99
86) p-Isopropyltoluene	13.723	119	3538	0.716	ug/l	98
87) 1,3-Dichlorobenzene	13.723	146	3038	0.940	ug/l	95
88) 1,4-Dichlorobenzene	13.723	146	3038	0.907	ug/l	95
89) n-Butylbenzene	14.053	91	3527	0.761	ug/l	93
90) Hexachloroethane	14.341	117	1355	1.102	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	3515	1.090	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.711	75	402	0.928	ug/l	74
93) 1,2,4-Trichlorobenzene	15.394	180	1310	0.874	ug/l #	80
94) Hexachlorobutadiene	15.500	225	822	1.033	ug/l	89
95) Naphthalene	15.641	128	4210	0.941	ug/l #	90
96) 1,2,3-Trichlorobenzene	15.835	180	1494	0.985	ug/l #	86

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085443.D
Acq On : 14 Jan 2025 17:19
Operator : JC\MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Quant Time: Jan 15 01:46:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 15 01:46:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:168 Resp: 166880
Ion Ratio Lower Upper
168 100
99 68.6 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 1.054 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 85 Resp: 2382
Ion Ratio Lower Upper
85 100
87 28.5 16.7 50.1









#3
Chloromethane
Concen: 1.145 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 50 Resp: 2801
Ion Ratio Lower Upper
50 100
52 29.7 27.0 40.6



#4
Vinyl Chloride
Concen: 1.111 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 62 Resp: 2732
Ion Ratio Lower Upper
62 100
64 41.4 24.8 37.2#









#5
Bromomethane
Concen: 1.534 ug/l
RT: 2.971 min Scan# 11
Delta R.T. 0.018 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 94 Resp: 2279
Ion Ratio Lower Upper
94 100
96 92.9 71.8 107.6



#6
Chloroethane
Concen: 1.160 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 64 Resp: 1809
Ion Ratio Lower Upper
64 100
66 34.1 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 1.082 ug/l
 RT: 3.512 min Scan# 21
 Delta R.T. 0.012 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion:101 Resp: 3860
 Ion Ratio Lower Upper
 101 100
 103 55.0 51.4 77.2



#8
 Diethyl Ether
 Concen: 1.238 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. -0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion: 74 Resp: 1526
 Ion Ratio Lower Upper
 74 100
 45 80.9 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.651 ug/l
RT: 4.359 min Scan# 409
Delta R.T. -0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:101 Resp: 1308
Ion Ratio Lower Upper
101 100
85 58.3 37.8 56.8#
151 109.4 58.8 88.2#



#10
Methyl Iodide
Concen: 1.077 ug/l
RT: 4.600 min Scan# 450
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:142 Resp: 2478
Ion Ratio Lower Upper
142 100
127 47.0 40.3 60.5
141 0.0 12.8 19.2#









#11
Tert butyl alcohol
Concen: 4.260 ug/l
RT: 5.512 min Scan# 605
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 59 Resp: 1314
Ion Ratio Lower Upper
59 100
57 0.0 8.6 13.0#



#12
1,1-Dichloroethene
Concen: 0.927 ug/l
RT: 4.330 min Scan# 404
Delta R.T. -0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 96 Resp: 1660
Ion Ratio Lower Upper
96 100
61 122.1 141.0 211.4#
98 48.1 49.0 73.6#





#13

Acrolein

Concen: 3.586 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion: 56 Resp: 1510

Ion Ratio Lower Upper

56 100

55 81.8 56.3 84.5



#14

Allyl chloride

Concen: 1.094 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 41 Resp: 3179


Ion Ratio Lower Upper

41 100

39 73.1 64.4 96.6

76 21.4 30.5 45.7#









#15
Acrylonitrile
Concen: 5.070 ug/l
RT: 5.718 min Scan# 64
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 53 Resp: 4967
Ion Ratio Lower Upper
53 100
52 45.1 65.5 98.3#
51 33.3 29.8 44.8



#16
Acetone
Concen: 5.862 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.018 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 5102
Ion Ratio Lower Upper
43 100
58 33.8 23.8 35.6









#17
Carbon Disulfide
Concen: 1.197 ug/l
RT: 4.706 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 76 Resp: 6601
Ion Ratio Lower Upper
76 100
78 13.2 6.9 10.3#



#18
Methyl Acetate
Concen: 1.099 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 2908
Ion Ratio Lower Upper
43 100
74 13.8 18.6 28.0#









#19
Methyl tert-butyl Ether
Concen: 0.887 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 73 Resp: 5156
Ion Ratio Lower Upper
73 100
57 28.4 17.8 26.8#



#20
Methylene Chloride
Concen: 1.016 ug/l
RT: 5.265 min Scan# 563
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 84 Resp: 2190
Ion Ratio Lower Upper
84 100
49 115.6 104.1 156.1
51 36.9 31.0 46.4
86 48.9 51.9 77.9#









#21
trans-1,2-Dichloroethene
Concen: 1.102 ug/l
RT: 5.794 min Scan# 610
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 96 Resp: 2110
Ion Ratio Lower Upper
96 100
61 98.7 115.7 173.5#
98 41.8 49.8 74.8#



#22
Diisopropyl ether
Concen: 0.882 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 45 Resp: 5691
Ion Ratio Lower Upper
45 100
43 65.9 45.0 67.6
87 35.0 21.8 32.8#
59 11.3 10.2 15.2









#23
 Vinyl Acetate
 Concen: 4.988 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.012 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 43 Resp: 22593
 Ion Ratio Lower Upper
 43 100
 86 9.8 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 0.730 ug/l
 RT: 6.553 min Scan# 782
 Delta R.T. -0.012 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion: 63 Resp: 2873
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.3 9.8#
 100 0.0 2.0 6.0#









#25
2-Butanone
Concen: 4.062 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 5203
Ion Ratio Lower Upper
43 100
72 20.1 20.2 30.4#



#26
2,2-Dichloropropane
Concen: 0.976 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 77 Resp: 3104
Ion Ratio Lower Upper
77 100
97 17.7 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 0.969 ug/l
 RT: 7.477 min Scan# 91
 Delta R.T. -0.006 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 96 Resp: 2185
 Ion Ratio Lower Upper
 96 100
 61 166.1 0.0 311.8
 98 68.7 0.0 126.0



#28
 Bromochloromethane
 Concen: 1.138 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion: 49 Resp: 2082
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 4.4
 130 78.0 55.0 82.4





#29

Tetrahydrofuran

Concen: 4.548 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 42 Resp: 3693

Ion Ratio Lower Upper

42 100

72 43.1 34.2 51.2

71 37.9 32.5 48.7



#30

Chloroform

Concen: 1.045 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 83 Resp: 4250

Ion Ratio Lower Upper

83 100

85 80.2 51.8 77.6#









#31
Cyclohexane
Concen: 2.092 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.035 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 56 Resp: 7152
Ion Ratio Lower Upper
56 100
69 177.3 22.2 33.4#
84 137.6 66.4 99.6#



#32
1,1,1-Trichloroethane
Concen: 1.031 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 97 Resp: 3678
Ion Ratio Lower Upper
97 100
99 0.0 49.8 74.6#
61 52.2 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 1.215 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 65 Resp: 3273

Ion Ratio Lower Upper

65 100

67 54.5 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 114 Resp: 302938


Ion Ratio Lower Upper

114 100

63 22.3 0.0 47.6

88 16.8 0.0 32.6









#35
Dibromofluoromethane
Concen: 1.074 ug/l
RT: 8.159 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:113 Resp: 2258
Ion Ratio Lower Upper
113 100
111 116.7 82.7 124.1
192 3.1 14.3 21.5#



#36
1,1-Dichloropropene
Concen: 1.046 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 75 Resp: 3084
Ion Ratio Lower Upper
75 100
110 30.1 16.5 49.5
77 15.9 24.4 36.6#









#37
Ethyl Acetate
Concen: 0.306 ug/l
RT: 7.553 min Scan# 911
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43	Resp:	911
Ion Ratio	Lower	Upper
43	100	
61	14.3	10.9 16.3
70	0.0	7.5 11.3#



#38
Carbon Tetrachloride
Concen: 1.018 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 117	Resp:	3436
Ion Ratio	Lower	Upper
117	100	
119	106.6	75.4 113.2
121	31.0	24.6 37.0





#39

Methylcyclohexane

Concen: 0.867 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 2403

Ion Ratio Lower Upper

83 100

55 72.7 62.6 94.0

98 31.8 37.7 56.5#



#40

Benzene

Concen: 0.957 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 78 Resp: 8480

Ion Ratio Lower Upper

78 100

77 28.3 19.0 28.6









#41
Methacrylonitrile
Concen: 0.941 ug/l
RT: 7.771 min Scan# 989
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 41	Resp: 1458
Ion Ratio	Lower Upper
41 100	
39 37.6	46.0 69.0#
67 53.8	57.4 86.2#
52 18.5	25.5 38.3#



#42
1,2-Dichloroethane
Concen: 0.938 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 62	Resp: 3131
Ion Ratio	Lower Upper
62 100	
98 5.4	0.0 17.0





#43

Isopropyl Acetate

Concen: 1.007 ug/l

RT: 8.682 min Scan# 1144

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 43 Resp: 4803

Ion Ratio Lower Upper

43 100

61 18.9 20.7 31.1#

87 7.3 9.8 14.8#



#44

Trichloroethene

Concen: 1.034 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion:130 Resp: 2133

Ion Ratio Lower Upper

130 100

95 126.1 0.0 195.8





#45

1,2-Dichloropropane

Concen: 0.992 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 63 Resp: 2246

Ion Ratio Lower Upper

63 100

65 32.6 25.6 38.4



#46

Dibromomethane

Concen: 1.083 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 93 Resp: 1769


Ion Ratio Lower Upper

93 100

95 86.5 64.7 97.1

174 77.4 69.0 103.4









#47
Bromodichloromethane
Concen: 0.882 ug/l
RT: 9.882 min Scan# 1348
Delta R.T. 0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	83	Resp:	2935
Ion	Ratio	Lower	Upper
83	100		
85	72.9	51.2	76.8
127	9.0	6.5	9.7



#48
Methyl methacrylate
Concen: 0.878 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:	41	Resp:	1886
Ion	Ratio	Lower	Upper
41	100		
69	76.2	64.7	97.1
39	62.8	49.0	73.6





#49

1,4-Dioxane

Concen: 18.616 ug/l

RT: 9.688 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 88 Resp: 673

Ion Ratio Lower Upper

88 100

43 0.0 26.6 39.8#

58 31.1 59.5 89.3#



#50

Toluene-d8

Concen: 0.959 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 98 Resp: 7164

Ion Ratio Lower Upper

98 100

100 61.9 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 4.163 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 11523
Ion Ratio Lower Upper
43 100
58 33.6 30.5 45.7



#52
Toluene
Concen: 0.814 ug/l
RT: 10.618 min Scan# 1473
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 92 Resp: 4180
Ion Ratio Lower Upper
92 100
91 183.9 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 0.802 ug/l
 RT: 10.835 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 75 Resp: 2522
 Ion Ratio Lower Upper
 75 100
 77 33.5 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 0.811 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.006 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion: 75 Resp: 2725
 Ion Ratio Lower Upper
 75 100
 77 39.0 25.0 37.4#
 39 67.6 43.1 64.7#









#55
1,1,2-Trichloroethane
Concen: 0.935 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	97	Resp:	1901
Ion	Ratio	Lower	Upper
97	100		
83	101.5	69.0	103.6
85	60.3	44.6	66.8
99	59.9	48.6	72.8



#56
Ethyl methacrylate
Concen: 2.471 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:	69	Resp:	1923
Ion	Ratio	Lower	Upper
69	100		
41	74.4	54.6	82.0
39	52.4	32.4	48.6





#57

1,3-Dichloropropane

Concen: 0.881 ug/l

RT: 11.165 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 76 Resp: 3113

Ion Ratio Lower Upper

76 100

78 34.9 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 3.822 ug/l

RT: 10.153 min Scan# 1394

Delta R.T. -0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 63 Resp: 4929

Ion Ratio Lower Upper

63 100

106 31.7 21.6 32.4





#59

2-Hexanone

Concen: 4.058 ug/l

RT: 11.194 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 43 Resp: 7905

Ion Ratio Lower Upper

43 100

58 42.9 26.2 78.6



#60

Dibromochloromethane

Concen: 0.953 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 129 Resp: 2339

Ion Ratio Lower Upper

129 100

127 78.9 38.6 115.8









#61
1,2-Dibromoethane
Concen: 1.000 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:107 Resp: 2022
Ion Ratio Lower Upper
107 100
109 84.8 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 0.951 ug/l
RT: 12.841 min Scan# 1851
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 95 Resp: 2428
Ion Ratio Lower Upper
95 100
174 67.5 0.0 145.0
176 67.3 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:117 Resp: 256851
Ion Ratio Lower Upper
117 100
82 59.0 48.6 72.8
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 0.947 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:164 Resp: 1658
Ion Ratio Lower Upper
164 100
166 152.1 103.4 155.2
129 128.3 79.2 118.8#
131 96.9 77.1 115.7









#65
Chlorobenzene
Concen: 0.960 ug/l
RT: 11.888 min Scan# 111
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:112 Resp: 5401
Ion Ratio Lower Upper
112 100
114 35.3 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 1.058 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:131 Resp: 2185
Ion Ratio Lower Upper
131 100
133 97.7 47.4 142.3
119 0.0 33.1 99.5#









#67
Ethyl Benzene
Concen: 0.802 ug/l
RT: 11.959 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 7347
Ion Ratio Lower Upper
91 100
106 32.8 23.8 35.8



#68
m/p-Xylenes
Concen: 1.494 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:106 Resp: 5059
Ion Ratio Lower Upper
106 100
91 209.3 167.7 251.5









#69
o-Xylene
Concen: 0.766 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:106 Resp: 2478
Ion Ratio Lower Upper
106 100
91 225.4 110.4 331.2



#70
Styrene
Concen: 0.712 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:104 Resp: 3812
Ion Ratio Lower Upper
104 100
78 52.9 42.5 63.7
103 53.7 43.8 65.8





#71

Bromoform

Concen: 0.818 ug/l

RT: 12.582 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:173 Resp: 1208

Ion Ratio Lower Upper

173 100

175 40.7 24.4 73.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion:152 Resp: 99539


Ion Ratio Lower Upper

152 100

115 63.8 31.1 93.3

150 158.5 0.0 343.6









#73
Isopropylbenzene
Concen: 0.820 ug/l
RT: 12.688 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 5506
Ion Ratio Lower Upper
105 100
120 26.7 12.8 38.3



#74
N-ethyl acetate
Concen: 0.899 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 2716
Ion Ratio Lower Upper
43 100
70 30.4 34.0 51.0#
55 24.2 21.4 32.2
61 17.6 19.0 28.4#





#75

1,1,2,2-Tetrachloroethane

Concen: 1.102 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 2616

Ion Ratio Lower Upper

83 100

131 2.4 4.8 14.4#

85 65.7 32.2 96.6



#76

1,2,3-Trichloropropane

Concen: 1.590 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 75 Resp: 3212

Ion Ratio Lower Upper

75 100

77 126.0 109.7 329.2









#77
Bromobenzene
Concen: 0.987 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:156 Resp: 1733
Ion Ratio Lower Upper
156 100
77 233.6 114.1 342.4
158 94.9 48.9 146.8



#78
n-propylbenzene
Concen: 0.808 ug/l
RT: 13.029 min Scan# 1883
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 91 Resp: 6424
Ion Ratio Lower Upper
91 100
120 18.2 10.9 32.6









#79
2-Chlorotoluene
Concen: 0.871 ug/l
RT: 13.117 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 4483
Ion Ratio Lower Upper
91 100
126 32.9 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 0.775 ug/l
RT: 13.165 min Scan# 1906
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 105 Resp: 4301
Ion Ratio Lower Upper
105 100
120 41.2 23.9 71.7









#81
 trans-1,4-Dichloro-2-butene
 Concen: 0.929 ug/l
 RT: 12.735 min Scan# 1833
 Delta R.T. -0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 75 Resp: 695
 Ion Ratio Lower Upper
 75 100
 53 101.2 95.6 143.4
 89 18.6 37.0 55.6#



#82
 4-Chlorotoluene
 Concen: 0.833 ug/l
 RT: 13.223 min Scan# 1916
 Delta R.T. 0.006 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion: 91 Resp: 4268
 Ion Ratio Lower Upper
 91 100
 126 31.0 15.9 47.7





#83

tert-Butylbenzene

Concen: 0.805 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:119 Resp: 3753

Ion Ratio Lower Upper

119 100

91 65.9 35.1 105.3

134 24.9 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 0.679 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion:105 Resp: 3758

Ion Ratio Lower Upper

105 100

120 47.7 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 0.714 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion:105 Resp: 4614
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 0.716 ug/l
 RT: 13.723 min Scan# 2001
 Delta R.T. 0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion:119 Resp: 3538
 Ion Ratio Lower Upper
 119 100
 134 25.6 12.7 38.0
 91 23.2 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 0.940 ug/l

RT: 13.723 min Scan# 2001

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:146 Resp: 3038

Ion Ratio Lower Upper

146 100

111 43.0 21.4 64.3

148 70.9 32.3 96.9



#88

1,4-Dichlorobenzene

Concen: 0.907 ug/l

RT: 13.723 min Scan# 2001

Delta R.T. -0.088 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion:146 Resp: 3038


Ion Ratio Lower Upper

146 100

111 43.0 21.3 63.7

148 70.9 32.4 97.0





#89

n-Butylbenzene

Concen: 0.761 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 91 Resp: 3527

Ion Ratio Lower Upper

91 100

92 57.2 25.8 77.3

134 25.5 11.7 35.1



#90

Hexachloroethane

Concen: 1.102 ug/l

RT: 14.341 min Scan# 2106

Delta R.T. 0.012 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 117 Resp: 1355

Ion Ratio Lower Upper

117 100

201 66.5 33.7 101.0









#91
1,2-Dichlorobenzene
Concen: 1.090 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:146 Resp: 3515
Ion Ratio Lower Upper
146 100
111 43.4 21.7 65.1
148 57.8 31.4 94.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 0.928 ug/l
RT: 14.711 min Scan# 2169
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 75 Resp: 402
Ion Ratio Lower Upper
75 100
155 48.0 36.4 109.2
157 112.7 45.4 136.1









#93
 1,2,4-Trichlorobenzene
 Concen: 0.874 ug/l
 RT: 15.394 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion:180 Resp: 1310
 Ion Ratio Lower Upper
 180 100
 182 106.3 46.8 140.3
 145 9.7 16.0 48.0#



#94
 Hexachlorobutadiene
 Concen: 1.033 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. -0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion:225 Resp: 822
 Ion Ratio Lower Upper
 225 100
 223 50.7 30.7 92.1
 227 55.7 30.9 92.5









#95
Naphthalene
Concen: 0.941 ug/l
RT: 15.641 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion:128 Resp: 4210
Ion Ratio Lower Upper
128 100
127 11.1 10.6 16.0
129 5.3 8.8 13.2#



#96
1,2,3-Trichlorobenzene
Concen: 0.985 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:180 Resp: 1494
Ion Ratio Lower Upper
180 100
182 90.2 47.4 142.2
145 11.8 16.9 50.7#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	219757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	370487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	320948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	158525	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	176594	49.782	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.560%
35) Dibromofluoromethane	8.165	113	130487	50.768	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	10.565	98	477630	52.302	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.600%
62) 4-Bromofluorobenzene	12.847	95	165249	52.899	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.800%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.124	85	130242	43.772	ug/l	100
3) Chloromethane	2.360	50	138152	42.883	ug/l	99
4) Vinyl Chloride	2.513	62	138316	42.715	ug/l	99
5) Bromomethane	2.954	94	80019	40.910	ug/l	91
6) Chloroethane	3.124	64	86837	42.299	ug/l	99
7) Trichlorofluoromethane	3.501	101	206526	43.956	ug/l	99
8) Diethyl Ether	3.959	74	69545	42.846	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	118310	44.706	ug/l	98
10) Methyl Iodide	4.595	142	140190	46.259	ug/l	98
11) Tert butyl alcohol	5.518	59	83933	206.633	ug/l	99
12) 1,1-Dichloroethene	4.342	96	106880	45.317	ug/l	98
13) Acrolein	4.177	56	135720	244.765	ug/l	99
14) Allyl chloride	5.024	41	170308	44.501	ug/l	99
15) Acrylonitrile	5.712	53	294062	227.920	ug/l	99
16) Acetone	4.418	43	245248	213.971	ug/l	98
17) Carbon Disulfide	4.712	76	300215	41.342	ug/l	100
18) Methyl Acetate	5.018	43	156938	45.028	ug/l	100
19) Methyl tert-butyl Ether	5.795	73	365819	47.768	ug/l	100
20) Methylene Chloride	5.271	84	122059	43.017	ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	112597	44.672	ug/l	93
22) Diisopropyl ether	6.665	45	399206	46.996	ug/l	99
23) Vinyl Acetate	6.601	43	1471068	247.459	ug/l	100
24) 1,1-Dichloroethane	6.571	63	228489	44.114	ug/l	98
25) 2-Butanone	7.477	43	386125	228.924	ug/l	100
26) 2,2-Dichloropropane	7.489	77	199303	47.593	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	134920	45.446	ug/l	98
28) Bromochloromethane	7.812	49	106323	44.122	ug/l	100
29) Tetrahydrofuran	7.836	42	255650	239.065	ug/l	99
30) Chloroform	7.965	83	235923	44.071	ug/l	98
31) Cyclohexane	8.253	56	193034	42.878	ug/l	93
32) 1,1,1-Trichloroethane	8.165	97	208172	44.332	ug/l	99
36) 1,1-Dichloropropene	8.365	75	161834	44.863	ug/l	98
37) Ethyl Acetate	7.559	43	160656	44.123	ug/l	97
38) Carbon Tetrachloride	8.359	117	183592	44.456	ug/l	99
39) Methylcyclohexane	9.600	83	168431	49.687	ug/l	98
40) Benzene	8.600	78	494871	45.657	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	89326	47.155	ug/l	99
42) 1,2-Dichloroethane	8.665	62	185663	45.481	ug/l	100
43) Isopropyl Acetate	8.683	43	265107	45.440	ug/l	99
44) Trichloroethene	9.347	130	110058	43.622	ug/l	96
45) 1,2-Dichloropropane	9.618	63	123876	44.744	ug/l	97
46) Dibromomethane	9.700	93	88325	44.214	ug/l	98
47) Bromodichloromethane	9.889	83	185808	45.655	ug/l	99
48) Methyl methacrylate	9.677	41	131142	49.948	ug/l	97
49) 1,4-Dioxane	9.689	88	35051	792.760	ug/l	98
51) 4-Methyl-2-Pentanone	10.441	43	832426	245.877	ug/l	99
52) Toluene	10.630	92	301911	48.073	ug/l	99
53) t-1,3-Dichloropropene	10.830	75	183680	47.757	ug/l	94
54) cis-1,3-Dichloropropene	10.312	75	196307	47.783	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	114757	46.173	ug/l	97
56) Ethyl methacrylate	10.871	69	182170	43.762	ug/l	98
57) 1,3-Dichloropropane	11.159	76	202583	46.877	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	403949	256.116	ug/l	99
59) 2-Hexanone	11.194	43	598433	251.213	ug/l	100
60) Dibromochloromethane	11.353	129	137831	45.935	ug/l	99
61) 1,2-Dibromoethane	11.465	107	110926	44.837	ug/l	100
64) Tetrachloroethene	11.100	164	98211	44.886	ug/l	96
65) Chlorobenzene	11.888	112	322723	45.901	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	115379	44.712	ug/l	99
67) Ethyl Benzene	11.959	91	577059	50.391	ug/l	97
68) m/p-Xylenes	12.071	106	437367	103.341	ug/l	100
69) o-Xylene	12.394	106	210698	52.093	ug/l	99
70) Styrene	12.406	104	355459	53.103	ug/l	100
71) Bromoform	12.577	173	90693	49.119	ug/l #	100
73) Isopropylbenzene	12.694	105	534428	49.951	ug/l	99
74) N-amyl acetate	12.488	43	234342	48.730	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	164882	43.627	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	221480	68.739	ug/l	44
77) Bromobenzene	12.977	156	128968	46.138	ug/l	97
78) n-propylbenzene	13.029	91	645977	51.001	ug/l	100
79) 2-Chlorotoluene	13.124	91	392404	47.870	ug/l	99
80) 1,3,5-Trimethylbenzene	13.171	105	456048	51.604	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	66842	56.127	ug/l	86
82) 4-Chlorotoluene	13.218	91	398110	48.776	ug/l	100
83) tert-Butylbenzene	13.435	119	378478	50.997	ug/l	98
84) 1,2,4-Trimethylbenzene	13.477	105	457340	51.922	ug/l	100
85) sec-Butylbenzene	13.612	105	532578	51.777	ug/l	100
86) p-Isopropyltoluene	13.729	119	440284	47.795	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	236387	45.920	ug/l	99
88) 1,4-Dichlorobenzene	13.806	146	235319	44.102	ug/l	99
89) n-Butylbenzene	14.053	91	373227	50.578	ug/l	99
90) Hexachloroethane	14.329	117	88538	45.215	ug/l	100
91) 1,2-Dichlorobenzene	14.100	146	231646	45.113	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.718	75	31899	46.221	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	109157	45.739	ug/l	98
94) Hexachlorobutadiene	15.500	225	53358	42.112	ug/l	99
95) Naphthalene	15.641	128	335535	47.116	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	107583	44.551	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICV011425


Tgt Ion:168 Resp: 219757
Ion Ratio Lower Upper
168 100
99 66.9 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 43.772 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 85 Resp: 130242
Ion Ratio Lower Upper
85 100
87 33.2 16.7 50.1









#3
Chloromethane
Concen: 42.883 ug/l
RT: 2.360 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 50 Resp: 138152
Ion Ratio Lower Upper
50 100
52 34.4 27.0 40.6



#4
Vinyl Chloride
Concen: 42.715 ug/l
RT: 2.513 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 62 Resp: 138316
Ion Ratio Lower Upper
62 100
64 31.4 24.8 37.2









#5
Bromomethane
Concen: 40.910 ug/l
RT: 2.954 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 94 Resp: 80019
Ion Ratio Lower Upper
94 100
96 98.2 71.8 107.6



#6
Chloroethane
Concen: 42.299 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 64 Resp: 86837
Ion Ratio Lower Upper
64 100
66 31.1 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 43.956 ug/l
 RT: 3.501 min Scan# 206
 Delta R.T. 0.000 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425


Tgt Ion:101 Resp: 206526
 Ion Ratio Lower Upper
 101 100
 103 65.1 51.4 77.2



#8
 Diethyl Ether
 Concen: 42.846 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06

Tgt Ion: 74 Resp: 69545
 Ion Ratio Lower Upper
 74 100
 45 99.0 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.706 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion:101 Resp: 118310
Ion Ratio Lower Upper
101 100
85 45.9 37.8 56.8
151 71.5 58.8 88.2



#10
Methyl Iodide
Concen: 46.259 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:142 Resp: 140190
Ion Ratio Lower Upper
142 100
127 49.2 40.3 60.5
141 13.9 12.8 19.2









#11
Tert butyl alcohol
Concen: 206.633 ug/l
RT: 5.518 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICV011425


Tgt Ion: 59 Resp: 83933
Ion Ratio Lower Upper
59 100
57 10.4 8.6 13.0



#12
1,1-Dichloroethene
Concen: 45.317 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 96 Resp: 106880
Ion Ratio Lower Upper
96 100
61 171.9 141.0 211.4
98 61.0 49.0 73.6





#13

Acrolein

Concen: 244.765 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :





ICVVN011425

Tgt Ion: 56 Resp: 135720

Ion Ratio Lower Upper

56 100

55 70.9 56.3 84.5



#14

Allyl chloride

Concen: 44.501 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 41 Resp: 170308


Ion Ratio Lower Upper

41 100

39 78.9 64.4 96.6

76 38.2 30.5 45.7









#15
Acrylonitrile
Concen: 227.920 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 53 Resp: 294062
Ion Ratio Lower Upper
53 100
52 82.7 65.5 98.3
51 36.8 29.8 44.8



#16
Acetone
Concen: 213.971 ug/l
RT: 4.418 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 245248
Ion Ratio Lower Upper
43 100
58 28.4 23.8 35.6









#17
Carbon Disulfide
Concen: 41.342 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 76 Resp: 300215
Ion Ratio Lower Upper
76 100
78 8.7 6.9 10.3



#18
Methyl Acetate
Concen: 45.028 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 156938
Ion Ratio Lower Upper
43 100
74 23.4 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 47.768 ug/l
RT: 5.795 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVNVN011425


Tgt Ion: 73 Resp: 365819
Ion Ratio Lower Upper
73 100
57 22.3 17.8 26.8



#20
Methylene Chloride
Concen: 43.017 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 84 Resp: 122059
Ion Ratio Lower Upper
84 100
49 131.3 104.1 156.1
51 41.0 31.0 46.4
86 63.3 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 44.672 ug/l
RT: 5.783 min Scan# 651
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVNV011425


Tgt Ion: 96 Resp: 112597
Ion Ratio Lower Upper
96 100
61 134.3 115.7 173.5
98 59.1 49.8 74.8



#22
Diisopropyl ether
Concen: 46.996 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 45 Resp: 399206
Ion Ratio Lower Upper
45 100
43 56.0 45.0 67.6
87 26.8 21.8 32.8
59 12.2 10.2 15.2









#23
 Vinyl Acetate
 Concen: 247.459 ug/l
 RT: 6.601 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425


Tgt Ion: 43 Resp: 1471068
 Ion Ratio Lower Upper
 43 100
 86 10.7 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 44.114 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.006 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06

Tgt Ion: 63 Resp: 228489
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.3 9.8
 100 4.2 2.0 6.0









#25
2-Butanone
Concen: 228.924 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 43 Resp: 386125
Ion Ratio Lower Upper
43 100
72 25.4 20.2 30.4



#26
2,2-Dichloropropane
Concen: 47.593 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 77 Resp: 199303
Ion Ratio Lower Upper
77 100
97 20.9 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 45.446 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425


Tgt Ion: 96 Resp: 134920
 Ion Ratio Lower Upper
 96 100
 61 158.0 0.0 311.8
 98 64.1 0.0 126.0



#28
 Bromochloromethane
 Concen: 44.122 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06

Tgt Ion: 49 Resp: 106323
 Ion Ratio Lower Upper
 49 100
 129 1.8 0.0 4.4
 130 68.6 55.0 82.4









#29
Tetrahydrofuran
Concen: 239.065 ug/l
RT: 7.836 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 42 Resp: 255650
Ion Ratio Lower Upper
42 100
72 42.5 34.2 51.2
71 39.9 32.5 48.7



#30
Chloroform
Concen: 44.071 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 83 Resp: 235923
Ion Ratio Lower Upper
83 100
85 63.5 51.8 77.6









#31
Cyclohexane
Concen: 42.878 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 56 Resp: 193034
Ion Ratio Lower Upper
56 100
69 30.5 22.2 33.4
84 89.4 66.4 99.6



#32
1,1,1-Trichloroethane
Concen: 44.332 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 97 Resp: 208172
Ion Ratio Lower Upper
97 100
99 61.8 49.8 74.6
61 52.8 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 49.782 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :





ICVVN011425

Tgt Ion: 65 Resp: 176594

Ion Ratio Lower Upper

65 100

67 50.9 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 114 Resp: 370487


Ion Ratio Lower Upper

114 100

63 21.6 0.0 47.6

88 15.8 0.0 32.6









#35
Dibromofluoromethane
Concen: 50.768 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion:113 Resp: 130487
Ion Ratio Lower Upper
113 100
111 102.5 82.7 124.1
192 17.7 14.3 21.5



#36
1,1-Dichloropropene
Concen: 44.863 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 161834
Ion Ratio Lower Upper
75 100
110 32.4 16.5 49.5
77 31.8 24.4 36.6









#37
Ethyl Acetate
Concen: 44.123 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 43 Resp: 160656
Ion Ratio Lower Upper
43 100
61 11.9 10.9 16.3
70 10.2 7.5 11.3



#38
Carbon Tetrachloride
Concen: 44.456 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 117 Resp: 183592
Ion Ratio Lower Upper
117 100
119 94.4 75.4 113.2
121 29.0 24.6 37.0





#39

Methylcyclohexane

Concen: 49.687 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425





Tgt Ion: 83 Resp: 168431

Ion Ratio Lower Upper

83 100

55 80.9 62.6 94.0

98 46.9 37.7 56.5



#40

Benzene

Concen: 45.657 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 78 Resp: 494871

Ion Ratio Lower Upper

78 100

77 23.5 19.0 28.6









#41
Methacrylonitrile
Concen: 47.155 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion:	41	Resp:	89326
Ion	Ratio	Lower	Upper
41	100		
39	56.6	46.0	69.0
67	70.9	57.4	86.2
52	32.6	25.5	38.3



#42
1,2-Dichloroethane
Concen: 45.481 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:	62	Resp:	185663
Ion	Ratio	Lower	Upper
62	100		
98	8.3	0.0	17.0









#43
Isopropyl Acetate
Concen: 45.440 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 43 Resp: 265107
Ion Ratio Lower Upper
43 100
61 26.7 20.7 31.1
87 12.9 9.8 14.8



#44
Trichloroethene
Concen: 43.622 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 130 Resp: 110058
Ion Ratio Lower Upper
130 100
95 101.9 0.0 195.8





#45

1,2-Dichloropropane

Concen: 44.744 ug/l

RT: 9.618 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :





ICVVN011425

Tgt Ion: 63 Resp: 123876

Ion Ratio Lower Upper

63 100

65 33.5 25.6 38.4



#46

Dibromomethane

Concen: 44.214 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion: 93 Resp: 88325


Ion Ratio Lower Upper

93 100

95 82.9 64.7 97.1

174 88.0 69.0 103.4









#47
Bromodichloromethane
Concen: 45.655 ug/l
RT: 9.889 min Scan# 1349
Delta R.T. 0.007 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 83 Resp: 185808
Ion Ratio Lower Upper
83 100
85 63.5 51.2 76.8
127 7.8 6.5 9.7



#48
Methyl methacrylate
Concen: 49.948 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 41 Resp: 131142
Ion Ratio Lower Upper
41 100
69 78.6 64.7 97.1
39 63.3 49.0 73.6





#49

1,4-Dioxane

Concen: 792.760 ug/l

RT: 9.689 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425





Tgt Ion: 88 Resp: 35051

Ion Ratio Lower Upper

88 100

43 33.8 26.6 39.8

58 76.1 59.5 89.3



#50

Toluene-d8

Concen: 52.302 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085445.D




Acq: 14 Jan 2025 18:06


Tgt Ion: 98 Resp: 477630

Ion Ratio Lower Upper

98 100

100 64.7 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 245.877 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 43 Resp: 832426
Ion Ratio Lower Upper
43 100
58 37.8 30.5 45.7



#52
Toluene
Concen: 48.073 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 92 Resp: 301911
Ion Ratio Lower Upper
92 100
91 175.7 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 47.757 ug/l
 RT: 10.830 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06




Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN011425


Tgt Ion: 75 Resp: 183680
 Ion Ratio Lower Upper
 75 100
 77 32.6 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 47.783 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.006 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 196307
 Ion Ratio Lower Upper
 75 100
 77 30.2 25.0 37.4
 39 51.9 43.1 64.7









#55
1,1,2-Trichloroethane
Concen: 46.173 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Tgt Ion: 97 Resp: 114757




Ion	Ratio	Lower	Upper
97	100		
83	83.7	69.0	103.6
85	53.2	44.6	66.8
99	62.8	48.6	72.8




#56
Ethyl methacrylate
Concen: 43.762 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 69 Resp: 182170

Ion	Ratio	Lower	Upper
69	100		
41	70.4	54.6	82.0
39	40.2	32.4	48.6









#57
1,3-Dichloropropane
Concen: 46.877 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 76 Resp: 202583
Ion Ratio Lower Upper
76 100
78 32.1 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 256.116 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 63 Resp: 403949
Ion Ratio Lower Upper
63 100
106 26.4 21.6 32.4









#59
2-Hexanone
Concen: 251.213 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 43 Resp: 598433
Ion Ratio Lower Upper
43 100
58 52.1 26.2 78.6



#60
Dibromochloromethane
Concen: 45.935 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 129 Resp: 137831
Ion Ratio Lower Upper
129 100
127 78.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 44.837 ug/l
RT: 11.465 min Scan# 1109
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion:107 Resp: 110926
Ion Ratio Lower Upper
107 100
109 94.7 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 52.899 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 95 Resp: 165249
Ion Ratio Lower Upper
95 100
174 73.1 0.0 145.0
176 70.8 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion:117 Resp: 320948
Ion Ratio Lower Upper
117 100
82 60.1 48.6 72.8
119 32.4 26.6 39.8



#64
Tetrachloroethene
Concen: 44.886 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:164 Resp: 98211
Ion Ratio Lower Upper
164 100
166 124.4 103.4 155.2
129 104.8 79.2 118.8
131 93.1 77.1 115.7









#65
Chlorobenzene
Concen: 45.901 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion:112 Resp: 322723
Ion Ratio Lower Upper
112 100
114 32.4 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 44.712 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:131 Resp: 115379
Ion Ratio Lower Upper
131 100
133 95.8 47.4 142.3
119 65.2 33.1 99.5









#67
Ethyl Benzene
Concen: 50.391 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 91 Resp: 577059
Ion Ratio Lower Upper
91 100
106 31.2 23.8 35.8



#68
m/p-Xylenes
Concen: 103.341 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:106 Resp: 437367
Ion Ratio Lower Upper
106 100
91 210.1 167.7 251.5





#69

o-Xylene

Concen: 52.093 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :





ICVVN011425

Tgt Ion:106 Resp: 210698

Ion Ratio Lower Upper

106 100

91 219.9 110.4 331.2



#70

Styrene

Concen: 53.103 ug/l

RT: 12.406 min Scan# 1777

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:104 Resp: 355459


Ion Ratio Lower Upper

104 100

78 52.8 42.5 63.7

103 54.5 43.8 65.8





#71

Bromoform

Concen: 49.119 ug/l

RT: 12.577 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425





Tgt Ion:173 Resp: 90693

Ion Ratio Lower Upper

173 100

175 48.7 24.4 73.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:152 Resp: 158525


Ion Ratio Lower Upper

152 100

115 62.2 31.1 93.3

150 171.0 0.0 343.6









#73
Isopropylbenzene
Concen: 49.951 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion:105 Resp: 534428
Ion Ratio Lower Upper
105 100
120 25.2 12.8 38.3



#74
N-ethyl acetate
Concen: 48.730 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 234342
Ion Ratio Lower Upper
43 100
70 42.2 34.0 51.0
55 26.5 21.4 32.2
61 23.8 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 43.627 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 83 Resp: 164882
Ion Ratio Lower Upper
83 100
131 9.9 4.8 14.4
85 64.6 32.2 96.6



#76
1,2,3-Trichloropropane
Concen: 68.739 ug/l
RT: 12.988 min Scan# 1876
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 221480
Ion Ratio Lower Upper
75 100
77 130.3 109.7 329.2









#77
Bromobenzene
Concen: 46.138 ug/l
RT: 12.977 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion:156 Resp: 128968
Ion Ratio Lower Upper
156 100
77 223.8 114.1 342.4
158 95.7 48.9 146.8



#78
n-propylbenzene
Concen: 51.001 ug/l
RT: 13.029 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 91 Resp: 645977
Ion Ratio Lower Upper
91 100
120 21.7 10.9 32.6









#79
2-Chlorotoluene
Concen: 47.870 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 91 Resp: 392404
Ion Ratio Lower Upper
91 100
126 31.8 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 51.604 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 105 Resp: 456048
Ion Ratio Lower Upper
105 100
120 48.1 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 56.127 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 75 Resp: 66842
Ion Ratio Lower Upper
75 100
53 103.5 95.6 143.4
89 38.0 37.0 55.6



#82
4-Chlorotoluene
Concen: 48.776 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 91 Resp: 398110
Ion Ratio Lower Upper
91 100
126 31.6 15.9 47.7









#83
 tert-Butylbenzene
 Concen: 50.997 ug/l
 RT: 13.435 min Scan# 1952
 Delta R.T. 0.000 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425


Tgt Ion:119 Resp: 378478
 Ion Ratio Lower Upper
 119 100
 91 70.6 35.1 105.3
 134 26.7 12.0 36.1



#84
 1,2,4-Trimethylbenzene
 Concen: 51.922 ug/l
 RT: 13.477 min Scan# 1959
 Delta R.T. 0.000 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06

Tgt Ion:105 Resp: 457340
 Ion Ratio Lower Upper
 105 100
 120 43.6 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 51.777 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425


Tgt Ion:105 Resp: 532578
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 47.795 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.006 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06

Tgt Ion:119 Resp: 440284
 Ion Ratio Lower Upper
 119 100
 134 25.3 12.7 38.0
 91 25.0 12.7 38.1









#87
 1,3-Dichlorobenzene
 Concen: 45.920 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425


Tgt Ion:146 Resp: 236387
 Ion Ratio Lower Upper
 146 100
 111 42.6 21.4 64.3
 148 63.8 32.3 96.9



#88
 1,4-Dichlorobenzene
 Concen: 44.102 ug/l
 RT: 13.806 min Scan# 2015
 Delta R.T. -0.006 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06

Tgt Ion:146 Resp: 235319
 Ion Ratio Lower Upper
 146 100
 111 41.9 21.3 63.7
 148 63.5 32.4 97.0









#89
n-Butylbenzene
Concen: 50.578 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06




Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425


Tgt Ion: 91 Resp: 373227
Ion Ratio Lower Upper
91 100
92 51.9 25.8 77.3
134 23.8 11.7 35.1



#90
Hexachloroethane
Concen: 45.215 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 117 Resp: 88538
Ion Ratio Lower Upper
117 100
201 67.2 33.7 101.0









#91
 1,2-Dichlorobenzene
 Concen: 45.113 ug/l
 RT: 14.100 min Scan# 2065
 Delta R.T. 0.000 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN011425


Tgt Ion:146 Resp: 231646
 Ion Ratio Lower Upper
 146 100
 111 43.9 21.7 65.1
 148 63.1 31.4 94.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 46.221 ug/l
 RT: 14.718 min Scan# 2170
 Delta R.T. 0.000 min
 Lab File: VN085445.D
 Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 31899
 Ion Ratio Lower Upper
 75 100
 155 69.0 36.4 109.2
 157 92.3 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 45.739 ug/l

RT: 15.388 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425





Tgt Ion:180 Resp: 109157

Ion Ratio Lower Upper

180 100

182 96.3 46.8 140.3

145 32.7 16.0 48.0



#94

Hexachlorobutadiene

Concen: 42.112 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:225 Resp: 53358


Ion Ratio Lower Upper

225 100

223 59.8 30.7 92.1

227 61.2 30.9 92.5





#95

Naphthalene

Concen: 47.116 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.006 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06

Instrument :

MSVOA_N

ClientSampleId :

ICVVN011425





Tgt Ion:128 Resp: 335535

Ion Ratio Lower Upper

128 100

127 13.2 10.6 16.0

129 11.1 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 44.551 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085445.D

Acq: 14 Jan 2025 18:06




Tgt Ion:180 Resp: 107583

Ion Ratio Lower Upper

180 100

182 95.6 47.4 142.2

145 34.3 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085526.D
 Acq On : 28 Jan 2025 09:53
 Operator : JC\MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDC0050

Quant Time: Jan 29 00:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	203638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	333806	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	302551	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	155713	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	158323	48.164	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.320%
35) Dibromofluoromethane	8.159	113	118662	51.240	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.480%
50) Toluene-d8	10.565	98	432537	52.569	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.847	95	155870	55.379	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.760%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	126280	45.800	ug/l	98
3) Chloromethane	2.359	50	130825	43.823	ug/l	99
4) Vinyl Chloride	2.512	62	135494	45.155	ug/l	98
5) Bromomethane	2.948	94	79311	43.757	ug/l	98
6) Chloroethane	3.118	64	89399	46.993	ug/l	100
7) Trichlorofluoromethane	3.495	101	218167	50.109	ug/l	99
8) Diethyl Ether	3.959	74	73428	48.819	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.371	101	125888	51.335	ug/l	99
10) Methyl Iodide	4.583	142	144081	51.306	ug/l	97
11) Tert butyl alcohol	5.518	59	110319	293.091	ug/l	100
12) 1,1-Dichloroethene	4.336	96	107139	49.022	ug/l	98
13) Acrolein	4.171	56	134509	261.783	ug/l	100
14) Allyl chloride	5.018	41	172649	48.683	ug/l	98
15) Acrylonitrile	5.712	53	312035	260.994	ug/l	99
16) Acetone	4.418	43	267575	251.929	ug/l	97
17) Carbon Disulfide	4.712	76	283263	42.095	ug/l	98
18) Methyl Acetate	5.018	43	174889	54.151	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	395014	55.663	ug/l	98
20) Methylene Chloride	5.277	84	131284	49.931	ug/l	99
21) trans-1,2-Dichloroethene	5.783	96	116065	49.692	ug/l	97
22) Diisopropyl ether	6.665	45	426926	54.237	ug/l	98
23) Vinyl Acetate	6.600	43	1520538	276.027	ug/l	99
24) 1,1-Dichloroethane	6.565	63	244849	51.014	ug/l	98
25) 2-Butanone	7.477	43	406723	260.223	ug/l	100
26) 2,2-Dichloropropane	7.483	77	223673	57.640	ug/l	100
27) cis-1,2-Dichloroethene	7.483	96	142501	51.799	ug/l	99
28) Bromochloromethane	7.812	49	101613	45.505	ug/l	98
29) Tetrahydrofuran	7.830	42	272222	274.712	ug/l	100
30) Chloroform	7.959	83	254209	51.246	ug/l	100
31) Cyclohexane	8.253	56	189607	45.451	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	225078	51.727	ug/l	98
36) 1,1-Dichloropropene	8.365	75	170755	52.537	ug/l	99
37) Ethyl Acetate	7.553	43	173618	52.923	ug/l	97
38) Carbon Tetrachloride	8.359	117	200977	54.013	ug/l	99
39) Methylcyclohexane	9.600	83	171154	56.039	ug/l	99
40) Benzene	8.600	78	525669	53.828	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085526.D
 Acq On : 28 Jan 2025 09:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	95095	55.717	ug/l	99
42) 1,2-Dichloroethane	8.665	62	195201	53.072	ug/l	99
43) Isopropyl Acetate	8.683	43	287826	54.755	ug/l	100
44) Trichloroethene	9.347	130	119734	52.672	ug/l	95
45) 1,2-Dichloropropane	9.618	63	135814	54.447	ug/l	97
46) Dibromomethane	9.706	93	95715	53.179	ug/l	99
47) Bromodichloromethane	9.883	83	204517	55.774	ug/l	99
48) Methyl methacrylate	9.677	41	136479	57.693	ug/l	97
49) 1,4-Dioxane	9.688	88	40607	1019.345	ug/l	100
51) 4-Methyl-2-Pentanone	10.441	43	892528	292.599	ug/l	100
52) Toluene	10.630	92	325356	57.499	ug/l	100
53) t-1,3-Dichloropropene	10.830	75	204105	58.899	ug/l	96
54) cis-1,3-Dichloropropene	10.306	75	213245	57.610	ug/l	99
55) 1,1,2-Trichloroethane	11.012	97	125068	55.851	ug/l	98
56) Ethyl methacrylate	10.871	69	191924	50.844	ug/l	99
57) 1,3-Dichloropropane	11.159	76	219323	56.327	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	426978	300.465	ug/l	99
59) 2-Hexanone	11.194	43	643969	300.034	ug/l	99
60) Dibromochloromethane	11.359	129	153599	56.815	ug/l	99
61) 1,2-Dibromoethane	11.465	107	122098	54.776	ug/l	97
64) Tetrachloroethene	11.100	164	109264	52.974	ug/l	98
65) Chlorobenzene	11.888	112	354308	53.457	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	130012	53.447	ug/l	99
67) Ethyl Benzene	11.959	91	621051	57.531	ug/l	99
68) m/p-Xylenes	12.065	106	475266	119.124	ug/l	99
69) o-Xylene	12.394	106	223578	58.639	ug/l	100
70) Styrene	12.406	104	387081	61.343	ug/l	99
71) Bromoform	12.576	173	101032	58.046	ug/l #	99
73) Isopropylbenzene	12.694	105	581759	55.357	ug/l	100
74) N-amyl acetate	12.494	43	249973	52.919	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	182235	49.089	ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	238136	75.243	ug/l	47
77) Bromobenzene	12.976	156	138788	50.547	ug/l	99
78) n-propylbenzene	13.035	91	701218	56.362	ug/l	100
79) 2-Chlorotoluene	13.123	91	424851	52.764	ug/l	99
80) 1,3,5-Trimethylbenzene	13.171	105	494666	56.984	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	63601	54.370	ug/l	92
82) 4-Chlorotoluene	13.218	91	432608	53.960	ug/l	98
83) tert-Butylbenzene	13.435	119	405613	55.641	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	501298	57.941	ug/l	100
85) sec-Butylbenzene	13.612	105	580942	57.499	ug/l	100
86) p-Isopropyltoluene	13.729	119	491281	53.532	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	264631	52.335	ug/l	99
88) 1,4-Dichlorobenzene	13.806	146	264719	50.507	ug/l	98
89) n-Butylbenzene	14.053	91	420498	58.013	ug/l	99
90) Hexachloroethane	14.329	117	97383	50.630	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	255916	50.739	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	34993	51.620	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	127164	54.246	ug/l	100
94) Hexachlorobutadiene	15.500	225	63906	51.348	ug/l	99
95) Naphthalene	15.641	128	367063	52.473	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	121516	51.230	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085526.D
Acq On : 28 Jan 2025 09:53
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085526.D
Acq On : 28 Jan 2025 09:53
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Quant Time: Jan 29 00:33:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 203638
Ion Ratio Lower Upper
168 100
99 63.5 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 45.800 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 85 Resp: 126280
Ion Ratio Lower Upper
85 100
87 32.1 16.7 50.1









#3
 Chloromethane
 Concen: 43.823 ug/l
 RT: 2.359 min Scan# 69
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 50 Resp: 130825
 Ion Ratio Lower Upper
 50 100
 52 33.0 27.0 40.6



#4
 Vinyl Chloride
 Concen: 45.155 ug/l
 RT: 2.512 min Scan# 95
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Tgt Ion: 62 Resp: 135494
 Ion Ratio Lower Upper
 62 100
 64 31.9 24.8 37.2









#5
Bromomethane
Concen: 43.757 ug/l
RT: 2.948 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 79311
Ion Ratio Lower Upper
94 100
96 91.4 71.8 107.6



#6
Chloroethane
Concen: 46.993 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 64 Resp: 89399
Ion Ratio Lower Upper
64 100
66 30.6 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 50.109 ug/l
 RT: 3.495 min Scan# 210
 Delta R.T. -0.006 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 218167
 Ion Ratio Lower Upper
 101 100
 103 63.5 51.4 77.2



#8
 Diethyl Ether
 Concen: 48.819 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Tgt Ion: 74 Resp: 73428
 Ion Ratio Lower Upper
 74 100
 45 96.9 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.335 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 125888
Ion Ratio Lower Upper
101 100
85 46.4 37.8 56.8
151 73.8 58.8 88.2



#10
Methyl Iodide
Concen: 51.306 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:142 Resp: 144081
Ion Ratio Lower Upper
142 100
127 52.8 40.3 60.5
141 14.5 12.8 19.2









#11
Tert butyl alcohol
Concen: 293.091 ug/l
RT: 5.518 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 110319
Ion Ratio Lower Upper
59 100
57 10.8 8.6 13.0



#12
1,1-Dichloroethene
Concen: 49.022 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 96 Resp: 107139
Ion Ratio Lower Upper
96 100
61 180.3 141.0 211.4
98 61.9 49.0 73.6





#13

Acrolein

Concen: 261.783 ug/l

RT: 4.171 min Scan# 377

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 56 Resp: 134509

Ion Ratio Lower Upper

56 100

55 70.5 56.3 84.5



#14

Allyl chloride

Concen: 48.683 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 41 Resp: 172649


Ion Ratio Lower Upper

41 100

39 82.7 64.4 96.6

76 38.6 30.5 45.7









#15
Acrylonitrile
Concen: 260.994 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 312035
Ion Ratio Lower Upper
53 100
52 82.4 65.5 98.3
51 36.9 29.8 44.8



#16
Acetone
Concen: 251.929 ug/l
RT: 4.418 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 43 Resp: 267575
Ion Ratio Lower Upper
43 100
58 28.2 23.8 35.6









#17
Carbon Disulfide
Concen: 42.095 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 283263
Ion Ratio Lower Upper
76 100
78 9.2 6.9 10.3



#18
Methyl Acetate
Concen: 54.151 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 43 Resp: 174889
Ion Ratio Lower Upper
43 100
74 23.1 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 55.663 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 395014
Ion Ratio Lower Upper
73 100
57 21.3 17.8 26.8



#20
Methylene Chloride
Concen: 49.931 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 84 Resp: 131284
Ion Ratio Lower Upper
84 100
49 128.3 104.1 156.1
51 40.2 31.0 46.4
86 64.3 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 49.692 ug/l
RT: 5.783 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 116065
Ion Ratio Lower Upper
96 100
61 148.4 115.7 173.5
98 59.7 49.8 74.8



#22
Diisopropyl ether
Concen: 54.237 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 45 Resp: 426926
Ion Ratio Lower Upper
45 100
43 55.8 45.0 67.6
87 25.6 21.8 32.8
59 11.3 10.2 15.2









#23
 Vinyl Acetate
 Concen: 276.027 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 1520538
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 51.014 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Tgt Ion: 63 Resp: 244849
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.3 9.8
 100 4.4 2.0 6.0









#25
2-Butanone
Concen: 260.223 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 406723
Ion Ratio Lower Upper
43 100
72 25.0 20.2 30.4



#26
2,2-Dichloropropane
Concen: 57.640 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 77 Resp: 223673
Ion Ratio Lower Upper
77 100
97 21.3 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 51.799 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 96 Resp: 142501
 Ion Ratio Lower Upper
 96 100
 61 157.2 0.0 311.8
 98 64.9 0.0 126.0



#28
 Bromochloromethane
 Concen: 45.505 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Tgt Ion: 49 Resp: 101613
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.4
 130 70.7 55.0 82.4





#29

Tetrahydrofuran

Concen: 274.712 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 272222

Ion Ratio Lower Upper

42 100

72 43.0 34.2 51.2

71 40.6 32.5 48.7



#30

Chloroform

Concen: 51.246 ug/l

RT: 7.959 min Scan# 1021

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 83 Resp: 254209

Ion Ratio Lower Upper

83 100

85 64.5 51.8 77.6









#31
Cyclohexane
Concen: 45.451 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 189607
Ion Ratio Lower Upper
56 100
69 30.2 22.2 33.4
84 85.8 66.4 99.6



#32
1,1,1-Trichloroethane
Concen: 51.727 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 97 Resp: 225078
Ion Ratio Lower Upper
97 100
99 64.9 49.8 74.6
61 50.9 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 48.164 ug/l

RT: 8.577 min Scan# 1125

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 65 Resp: 158323

Ion Ratio Lower Upper

65 100

67 51.1 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 114 Resp: 333806


Ion Ratio Lower Upper

114 100

63 22.7 0.0 47.6

88 16.4 0.0 32.6









#35
Dibromofluoromethane
Concen: 51.240 ug/l
RT: 8.159 min Scan# 1055
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 118662
Ion Ratio Lower Upper
113 100
111 103.2 82.7 124.1
192 18.0 14.3 21.5



#36
1,1-Dichloropropene
Concen: 52.537 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 75 Resp: 170755
Ion Ratio Lower Upper
75 100
110 32.6 16.5 49.5
77 31.6 24.4 36.6





#37

Ethyl Acetate

Concen: 52.923 ug/l

RT: 7.553 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 173618

Ion Ratio Lower Upper

43 100

61 11.8 10.9 16.3

70 10.0 7.5 11.3



#38

Carbon Tetrachloride

Concen: 54.013 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 117 Resp: 200977


Ion Ratio Lower Upper

117 100

119 96.0 75.4 113.2

121 31.1 24.6 37.0





#39

Methylcyclohexane

Concen: 56.039 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 171154

Ion Ratio Lower Upper

83 100

55 79.2 62.6 94.0

98 47.9 37.7 56.5



#40

Benzene

Concen: 53.828 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 78 Resp: 525669

Ion Ratio Lower Upper

78 100

77 23.9 19.0 28.6









#41
Methacrylonitrile
Concen: 55.717 ug/l
RT: 7.771 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	95095
Ion	Ratio	Lower	Upper
41	100		
39	56.3	46.0	69.0
67	71.1	57.4	86.2
52	32.2	25.5	38.3



#42
1,2-Dichloroethane
Concen: 53.072 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:	62	Resp:	195201
Ion	Ratio	Lower	Upper
62	100		
98	8.8	0.0	17.0









#43
Isopropyl Acetate
Concen: 54.755 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 287826
Ion Ratio Lower Upper
43 100
61 25.9 20.7 31.1
87 12.6 9.8 14.8



#44
Trichloroethene
Concen: 52.672 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:130 Resp: 119734
Ion Ratio Lower Upper
130 100
95 103.0 0.0 195.8





#45

1,2-Dichloropropane

Concen: 54.447 ug/l

RT: 9.618 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 135814

Ion Ratio Lower Upper

63 100

65 33.9 25.6 38.4



#46

Dibromomethane

Concen: 53.179 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 93 Resp: 95715


Ion Ratio Lower Upper

93 100

95 81.2 64.7 97.1

174 87.0 69.0 103.4





#47

Bromodichloromethane

Concen: 55.774 ug/l

RT: 9.883 min Scan# 1

Delta R.T. 0.001 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 204517

Ion Ratio Lower Upper

83 100

85 65.1 51.2 76.8

127 8.4 6.5 9.7



#48

Methyl methacrylate

Concen: 57.693 ug/l

RT: 9.677 min Scan# 1313

Delta R.T. -0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion: 41 Resp: 136479


Ion Ratio Lower Upper

41 100

69 83.2 64.7 97.1

39 64.1 49.0 73.6





#49

1,4-Dioxane

Concen: 1019.345 ug/l

RT: 9.688 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 40607

Ion Ratio Lower Upper

88 100

43 33.0 26.6 39.8

58 74.6 59.5 89.3



#50

Toluene-d8

Concen: 52.569 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion: 98 Resp: 432537

Ion Ratio Lower Upper

98 100

100 65.0 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 292.599 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 892528
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7



#52
Toluene
Concen: 57.499 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 92 Resp: 325356
Ion Ratio Lower Upper
92 100
91 173.6 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 58.899 ug/l
 RT: 10.830 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 75 Resp: 204105
 Ion Ratio Lower Upper
 75 100
 77 31.6 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 57.610 ug/l
 RT: 10.306 min Scan# 1420
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Tgt Ion: 75 Resp: 213245
 Ion Ratio Lower Upper
 75 100
 77 30.8 25.0 37.4
 39 52.6 43.1 64.7









#55
1,1,2-Trichloroethane
Concen: 55.851 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 97 Resp: 125068




Ion	Ratio	Lower	Upper
97	100		
83	84.4	69.0	103.6
85	57.9	44.6	66.8
99	62.7	48.6	72.8




#56
Ethyl methacrylate
Concen: 50.844 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 69 Resp: 191924

Ion	Ratio	Lower	Upper
69	100		
41	68.4	54.6	82.0
39	42.2	32.4	48.6









#57
1,3-Dichloropropane
Concen: 56.327 ug/l
RT: 11.159 min Scan# 1115
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 219323
Ion Ratio Lower Upper
76 100
78 31.5 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 300.465 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 63 Resp: 426978
Ion Ratio Lower Upper
63 100
106 27.8 21.6 32.4









#59
2-Hexanone
Concen: 300.034 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 643969
Ion Ratio Lower Upper
43 100
58 51.8 26.2 78.6



#60
Dibromochloromethane
Concen: 56.815 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 129 Resp: 153599
Ion Ratio Lower Upper
129 100
127 76.7 38.6 115.8









#61
1,2-Dibromoethane
Concen: 54.776 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 122098
Ion Ratio Lower Upper
107 100
109 92.4 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 55.379 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 95 Resp: 155870
Ion Ratio Lower Upper
95 100
174 73.4 0.0 145.0
176 71.9 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 302551
Ion Ratio Lower Upper
117 100
82 58.5 48.6 72.8
119 31.2 26.6 39.8



#64
Tetrachloroethene
Concen: 52.974 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:164 Resp: 109264
Ion Ratio Lower Upper
164 100
166 125.7 103.4 155.2
129 101.8 79.2 118.8
131 96.7 77.1 115.7









#65
Chlorobenzene
Concen: 53.457 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 354308
Ion Ratio Lower Upper
112 100
114 31.2 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 53.447 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:131 Resp: 130012
Ion Ratio Lower Upper
131 100
133 95.0 47.4 142.3
119 64.2 33.1 99.5





#67

Ethyl Benzene

Concen: 57.531 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 621051

Ion Ratio Lower Upper

91 100

106 30.2 23.8 35.8



#68

m/p-Xylenes

Concen: 119.124 ug/l

RT: 12.065 min Scan# 1719

Delta R.T. 0.000 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion:106 Resp: 475266

Ion Ratio Lower Upper

106 100

91 207.6 167.7 251.5









#69
o-Xylene
Concen: 58.639 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 223578
Ion Ratio Lower Upper
106 100
91 220.9 110.4 331.2



#70
Styrene
Concen: 61.343 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:104 Resp: 387081
Ion Ratio Lower Upper
104 100
78 52.7 42.5 63.7
103 55.4 43.8 65.8









#71
Bromoform
Concen: 58.046 ug/l
RT: 12.576 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 101032
Ion Ratio Lower Upper
173 100
175 49.5 24.4 73.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion:152 Resp: 155713
Ion Ratio Lower Upper
152 100
115 62.3 31.1 93.3
150 172.0 0.0 343.6









#73
Isopropylbenzene
Concen: 55.357 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 581759
Ion Ratio Lower Upper
105 100
120 25.5 12.8 38.3



#74
N-ethyl acetate
Concen: 52.919 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 43 Resp: 249973
Ion Ratio Lower Upper
43 100
70 42.1 34.0 51.0
55 26.4 21.4 32.2
61 22.9 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 49.089 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53




Instrument : MSVOA_N
 ClientSampleId : VSTDCCC050


Tgt Ion:	83	Resp:	182235
Ion Ratio	Lower	Upper	
83	100		
131	9.9	4.8	14.4
85	63.9	32.2	96.6



#76
 1,2,3-Trichloropropane
 Concen: 75.243 ug/l
 RT: 12.988 min Scan# 1876
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Tgt Ion:	75	Resp:	238136
Ion Ratio	Lower	Upper	
75	100		
77	133.8	109.7	329.2









#77
Bromobenzene
Concen: 50.547 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 138788
Ion Ratio Lower Upper
156 100
77 229.6 114.1 342.4
158 98.7 48.9 146.8



#78
n-propylbenzene
Concen: 56.362 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 91 Resp: 701218
Ion Ratio Lower Upper
91 100
120 21.9 10.9 32.6









#79
2-Chlorotoluene
Concen: 52.764 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 424851
Ion Ratio Lower Upper
91 100
126 32.2 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 56.984 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 105 Resp: 494666
Ion Ratio Lower Upper
105 100
120 47.7 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 54.370 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 63601
Ion Ratio Lower Upper
75 100
53 107.1 95.6 143.4
89 46.8 37.0 55.6



#82
4-Chlorotoluene
Concen: 53.960 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 91 Resp: 432608
Ion Ratio Lower Upper
91 100
126 32.8 15.9 47.7





#83

tert-Butylbenzene

Concen: 55.641 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:119 Resp: 405613

Ion Ratio Lower Upper

119 100

91 69.8 35.1 105.3

134 27.3 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 57.941 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085526.D




Acq: 28 Jan 2025 09:53


Tgt Ion:105 Resp: 501298

Ion Ratio Lower Upper

105 100

120 43.4 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 57.499 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 580942
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 53.532 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.006 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Tgt Ion:119 Resp: 491281
 Ion Ratio Lower Upper
 119 100
 134 25.5 12.7 38.0
 91 24.7 12.7 38.1









#87
 1,3-Dichlorobenzene
 Concen: 52.335 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:146 Resp: 264631
 Ion Ratio Lower Upper
 146 100
 111 43.3 21.4 64.3
 148 64.1 32.3 96.9



#88
 1,4-Dichlorobenzene
 Concen: 50.507 ug/l
 RT: 13.806 min Scan# 2015
 Delta R.T. -0.006 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Tgt Ion:146 Resp: 264719
 Ion Ratio Lower Upper
 146 100
 111 40.8 21.3 63.7
 148 63.7 32.4 97.0









#89
n-Butylbenzene
Concen: 58.013 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 420498
Ion Ratio Lower Upper
91 100
92 52.1 25.8 77.3
134 24.2 11.7 35.1



#90
Hexachloroethane
Concen: 50.630 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53

Tgt Ion: 117 Resp: 97383
Ion Ratio Lower Upper
117 100
201 67.0 33.7 101.0









#91
1,2-Dichlorobenzene
Concen: 50.739 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.006 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 255916
Ion Ratio Lower Upper
146 100
111 43.8 21.7 65.1
148 62.7 31.4 94.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.620 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085526.D
Acq: 28 Jan 2025 09:53


Tgt Ion: 75 Resp: 34993
Ion Ratio Lower Upper
75 100
155 69.5 36.4 109.2
157 90.0 45.4 136.1








#93
 1,2,4-Trichlorobenzene
 Concen: 54.246 ug/l
 RT: 15.388 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050




Tgt Ion:180 Resp: 127164

Ion	Ratio	Lower	Upper
180	100		
182	93.6	46.8	140.3
145	32.5	16.0	48.0






#94
 Hexachlorobutadiene
 Concen: 51.348 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. 0.000 min
 Lab File: VN085526.D
 Acq: 28 Jan 2025 09:53



Tgt Ion:225 Resp: 63906

Ion	Ratio	Lower	Upper
225	100		
223	62.8	30.7	92.1
227	62.3	30.9	92.5





#95

Naphthalene

Concen: 52.473 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:128 Resp: 367063

Ion Ratio Lower Upper

128 100

127 12.9 10.6 16.0

129 10.5 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 51.230 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085526.D

Acq: 28 Jan 2025 09:53




Tgt Ion:180 Resp: 121516

Ion Ratio Lower Upper

180 100

182 96.0 47.4 142.2

145 34.5 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085529.D
 Acq On : 28 Jan 2025 11:15
 Operator : JC\MD
 Sample : VN0128WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01

Quant Time: Jan 29 00:35:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	307569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	271043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126721	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	153168	51.135	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.280%	
35) Dibromofluoromethane	8.165	113	113571	53.225	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.460%	
50) Toluene-d8	10.565	98	410056	54.088	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.180%	
62) 4-Bromofluorobenzene	12.847	95	134929	52.029	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.060%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	45881	18.261	ug/l	99
3) Chloromethane	2.359	50	46936	17.254	ug/l	98
4) Vinyl Chloride	2.512	62	49715	18.182	ug/l	98
5) Bromomethane	2.965	94	32217	19.506	ug/l	93
6) Chloroethane	3.118	64	34288	19.780	ug/l	95
7) Trichlorofluoromethane	3.501	101	82151	20.707	ug/l	94
8) Diethyl Ether	3.959	74	26345	19.222	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.377	101	45774	20.484	ug/l	98
10) Methyl Iodide	4.589	142	50026	19.549	ug/l	95
11) Tert butyl alcohol	5.512	59	37334	108.849	ug/l	99
12) 1,1-Dichloroethene	4.342	96	37952	19.057	ug/l	93
13) Acrolein	4.177	56	39646	84.676	ug/l	98
14) Allyl chloride	5.024	41	59056	18.275	ug/l	98
15) Acrylonitrile	5.712	53	116207	106.667	ug/l	98
16) Acetone	4.424	43	101360	104.730	ug/l	94
17) Carbon Disulfide	4.706	76	100344	16.364	ug/l	98
18) Methyl Acetate	5.018	43	66988	22.762	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	135056	20.885	ug/l	97
20) Methylene Chloride	5.277	84	47487	19.820	ug/l	96
21) trans-1,2-Dichloroethene	5.789	96	40233	18.903	ug/l	94
22) Diisopropyl ether	6.671	45	147403	20.550	ug/l	97
23) Vinyl Acetate	6.600	43	516002	102.796	ug/l	99
24) 1,1-Dichloroethane	6.565	63	87952	20.110	ug/l	99
25) 2-Butanone	7.477	43	153642	107.876	ug/l	98
26) 2,2-Dichloropropane	7.488	77	78263	22.133	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	49822	19.875	ug/l	99
28) Bromochloromethane	7.812	49	32851	16.145	ug/l	96
29) Tetrahydrofuran	7.836	42	102010	112.971	ug/l	99
30) Chloroform	7.965	83	92905	20.553	ug/l	93
31) Cyclohexane	8.253	56	67069	17.643	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	80919	20.408	ug/l	97
36) 1,1-Dichloropropene	8.365	75	58916	19.673	ug/l	99
37) Ethyl Acetate	7.553	43	62638	20.722	ug/l	98
38) Carbon Tetrachloride	8.359	117	70429	20.543	ug/l	97
39) Methylcyclohexane	9.600	83	53089	18.865	ug/l	96
40) Benzene	8.606	78	186111	20.683	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085529.D
 Acq On : 28 Jan 2025 11:15
 Operator : JC\MD
 Sample : VN0128WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01

Quant Time: Jan 29 00:35:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	33670	21.410	ug/l	98
42) 1,2-Dichloroethane	8.665	62	71068	20.970	ug/l	100
43) Isopropyl Acetate	8.688	43	105036	21.686	ug/l	99
44) Trichloroethene	9.347	130	42674	20.374	ug/l	97
45) 1,2-Dichloropropane	9.618	63	47674	20.743	ug/l	98
46) Dibromomethane	9.706	93	33823	20.395	ug/l	97
47) Bromodichloromethane	9.882	83	73309	21.697	ug/l	97
48) Methyl methacrylate	9.677	41	47567	21.823	ug/l	98
49) 1,4-Dioxane	9.694	88	15432	420.430	ug/l	98
51) 4-Methyl-2-Pentanone	10.441	43	324778	115.555	ug/l	100
52) Toluene	10.629	92	113994	21.864	ug/l	99
53) t-1,3-Dichloropropene	10.829	75	68322	21.398	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	73429	21.530	ug/l	95
55) 1,1,2-Trichloroethane	11.012	97	45212	21.912	ug/l	94
56) Ethyl methacrylate	10.871	69	62727	19.281	ug/l	98
57) 1,3-Dichloropropane	11.159	76	78623	21.915	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	114984	87.817	ug/l	100
59) 2-Hexanone	11.194	43	234494	118.574	ug/l	98
60) Dibromochloromethane	11.359	129	52835	21.210	ug/l	99
61) 1,2-Dibromoethane	11.471	107	42361	20.625	ug/l	99
64) Tetrachloroethene	11.100	164	40019	21.658	ug/l	99
65) Chlorobenzene	11.888	112	122235	20.586	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	46034	21.124	ug/l	99
67) Ethyl Benzene	11.959	91	201123	20.797	ug/l	98
68) m/p-Xylenes	12.071	106	154069	43.106	ug/l	99
69) o-Xylene	12.394	106	71535	20.943	ug/l	100
70) Styrene	12.412	104	120125	21.250	ug/l	99
71) Bromoform	12.576	173	35108	22.516	ug/l #	99
73) Isopropylbenzene	12.694	105	178577	20.880	ug/l	98
74) N-amyl acetate	12.488	43	80633	20.975	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	65405	21.649	ug/l	98
76) 1,2,3-Trichloropropane	12.988	75	82646	32.088	ug/l	45
77) Bromobenzene	12.976	156	47665	21.332	ug/l	99
78) n-propylbenzene	13.035	91	217855	21.517	ug/l	99
79) 2-Chlorotoluene	13.123	91	138138	21.081	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	154258	21.836	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	21965	23.073	ug/l	87
82) 4-Chlorotoluene	13.218	91	139420	21.369	ug/l	100
83) tert-Butylbenzene	13.435	119	124147	20.926	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	155548	22.092	ug/l	98
85) sec-Butylbenzene	13.612	105	178351	21.691	ug/l	100
86) p-Isopropyltoluene	13.729	119	143953	20.952	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	85729	20.833	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	84406	19.789	ug/l	99
89) n-Butylbenzene	14.053	91	115181	19.526	ug/l	100
90) Hexachloroethane	14.329	117	30558	19.522	ug/l	97
91) 1,2-Dichlorobenzene	14.100	146	83529	20.350	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12961	23.494	ug/l	91
93) 1,2,4-Trichlorobenzene	15.394	180	36081	18.913	ug/l	95
94) Hexachlorobutadiene	15.494	225	20163	19.907	ug/l	98
95) Naphthalene	15.641	128	104713	18.394	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	36634	18.978	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085529.D
Acq On : 28 Jan 2025 11:15
Operator : JC\MD
Sample : VN0128WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01

Quant Time: Jan 29 00:35:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085529.D
 Acq On : 28 Jan 2025 11:15
 Operator : JC\MD
 Sample : VN0128WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01

Quant Time: Jan 29 00:35:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:168 Resp: 185562
Ion Ratio Lower Upper
168 100
99 65.3 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 18.261 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 85 Resp: 45881
Ion Ratio Lower Upper
85 100
87 33.0 16.7 50.1









#3
Chloromethane
Concen: 17.254 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 50 Resp: 46936
Ion Ratio Lower Upper
50 100
52 35.2 27.0 40.6



#4
Vinyl Chloride
Concen: 18.182 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 62 Resp: 49715
Ion Ratio Lower Upper
62 100
64 32.1 24.8 37.2









#5
Bromomethane
Concen: 19.506 ug/l
RT: 2.965 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 94 Resp: 32217
Ion Ratio Lower Upper
94 100
96 83.2 71.8 107.6



#6
Chloroethane
Concen: 19.780 ug/l
RT: 3.118 min Scan# 198
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 64 Resp: 34288
Ion Ratio Lower Upper
64 100
66 33.3 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 20.707 ug/l
 RT: 3.501 min Scan# 210
 Delta R.T. 0.000 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01


Tgt Ion:101 Resp: 82151
 Ion Ratio Lower Upper
 101 100
 103 69.3 51.4 77.2



#8
 Diethyl Ether
 Concen: 19.222 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15

Tgt Ion: 74 Resp: 26345
 Ion Ratio Lower Upper
 74 100
 45 97.4 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.484 ug/l
RT: 4.377 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:101 Resp: 45774
Ion Ratio Lower Upper
101 100
85 44.9 37.8 56.8
151 72.0 58.8 88.2



#10
Methyl Iodide
Concen: 19.549 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:142 Resp: 50026
Ion Ratio Lower Upper
142 100
127 54.4 40.3 60.5
141 16.2 12.8 19.2









#11
Tert butyl alcohol
Concen: 108.849 ug/l
RT: 5.512 min Scan# 605
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 59 Resp: 37334
Ion Ratio Lower Upper
59 100
57 10.4 8.6 13.0



#12
1,1-Dichloroethene
Concen: 19.057 ug/l
RT: 4.342 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 96 Resp: 37952
Ion Ratio Lower Upper
96 100
61 185.5 141.0 211.4
98 66.4 49.0 73.6





#13

Acrolein

Concen: 84.676 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBS01

Tgt Ion: 56 Resp: 39646

Ion Ratio Lower Upper

56 100

55 71.9 56.3 84.5



#14

Allyl chloride

Concen: 18.275 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 41 Resp: 59056


Ion Ratio Lower Upper

41 100

39 83.5 64.4 96.6

76 38.4 30.5 45.7









#15
Acrylonitrile
Concen: 106.667 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 53 Resp: 116207
Ion Ratio Lower Upper
53 100
52 83.3 65.5 98.3
51 36.3 29.8 44.8



#16
Acetone
Concen: 104.730 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 43 Resp: 101360
Ion Ratio Lower Upper
43 100
58 26.7 23.8 35.6









#17
Carbon Disulfide
Concen: 16.364 ug/l
RT: 4.706 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 76 Resp: 100344
Ion Ratio Lower Upper
76 100
78 9.2 6.9 10.3



#18
Methyl Acetate
Concen: 22.762 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 43 Resp: 66988
Ion Ratio Lower Upper
43 100
74 23.6 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 20.885 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 73 Resp: 135056
Ion Ratio Lower Upper
73 100
57 20.8 17.8 26.8



#20
Methylene Chloride
Concen: 19.820 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 84 Resp: 47487
Ion Ratio Lower Upper
84 100
49 126.6 104.1 156.1
51 34.2 31.0 46.4
86 62.2 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 18.903 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 96 Resp: 40233
Ion Ratio Lower Upper
96 100
61 152.9 115.7 173.5
98 66.4 49.8 74.8



#22
Diisopropyl ether
Concen: 20.550 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 45 Resp: 147403
Ion Ratio Lower Upper
45 100
43 55.8 45.0 67.6
87 31.0 21.8 32.8
59 11.8 10.2 15.2









#23
 Vinyl Acetate
 Concen: 102.796 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01


Tgt Ion: 43 Resp: 516002
 Ion Ratio Lower Upper
 43 100
 86 10.2 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 20.110 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15

Tgt Ion: 63 Resp: 87952
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.3 9.8
 100 4.7 2.0 6.0









#25
2-Butanone
Concen: 107.876 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 43 Resp: 153642
Ion Ratio Lower Upper
43 100
72 26.4 20.2 30.4



#26
2,2-Dichloropropane
Concen: 22.133 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 77 Resp: 78263
Ion Ratio Lower Upper
77 100
97 21.4 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 19.875 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01


Tgt Ion: 96 Resp: 49822
 Ion Ratio Lower Upper
 96 100
 61 157.4 0.0 311.8
 98 62.7 0.0 126.0



#28
 Bromochloromethane
 Concen: 16.145 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15

Tgt Ion: 49 Resp: 32851
 Ion Ratio Lower Upper
 49 100
 129 1.5 0.0 4.4
 130 72.0 55.0 82.4









#29
Tetrahydrofuran
Concen: 112.971 ug/l
RT: 7.836 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 42 Resp: 102010
Ion Ratio Lower Upper
42 100
72 41.9 34.2 51.2
71 40.5 32.5 48.7



#30
Chloroform
Concen: 20.553 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 83 Resp: 92905
Ion Ratio Lower Upper
83 100
85 59.4 51.8 77.6









#31
Cyclohexane
Concen: 17.643 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:	56	Resp:	67069
Ion Ratio	Lower	Upper	
56	100		
69	29.0	22.2	33.4
84	84.8	66.4	99.6



#32
1,1,1-Trichloroethane
Concen: 20.408 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:	97	Resp:	80919
Ion Ratio	Lower	Upper	
97	100		
99	63.9	49.8	74.6
61	48.7	41.4	62.2





#33

1,2-Dichloroethane-d4

Concen: 51.135 ug/l

RT: 8.577 min Scan# 1125

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBS01

Tgt Ion: 65 Resp: 153168

Ion Ratio Lower Upper

65 100

67 50.0 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 114 Resp: 307569


Ion Ratio Lower Upper

114 100

63 22.9 0.0 47.6

88 16.7 0.0 32.6





#35

Dibromofluoromethane

Concen: 53.225 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01





Tgt Ion:113 Resp: 113571

Ion Ratio Lower Upper

113 100

111 103.6 82.7 124.1

192 17.7 14.3 21.5



#36

1,1-Dichloropropene

Concen: 19.673 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 75 Resp: 58916


Ion Ratio Lower Upper

75 100

110 33.1 16.5 49.5

77 31.4 24.4 36.6









#37
Ethyl Acetate
Concen: 20.722 ug/l
RT: 7.553 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 43 Resp: 62638
Ion Ratio Lower Upper
43 100
61 13.1 10.9 16.3
70 10.3 7.5 11.3



#38
Carbon Tetrachloride
Concen: 20.543 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 117 Resp: 70429
Ion Ratio Lower Upper
117 100
119 97.9 75.4 113.2
121 30.7 24.6 37.0





#39

Methylcyclohexane

Concen: 18.865 ug/l

RT: 9.600 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01





Tgt Ion: 83 Resp: 53089

Ion Ratio Lower Upper

83 100

55 83.8 62.6 94.0

98 47.6 37.7 56.5



#40

Benzene

Concen: 20.683 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 78 Resp: 186111

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.6









#41
Methacrylonitrile
Concen: 21.410 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:	41	Resp:	33670
Ion	Ratio	Lower	Upper
41	100		
39	60.5	46.0	69.0
67	71.9	57.4	86.2
52	32.1	25.5	38.3



#42
1,2-Dichloroethane
Concen: 20.970 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:	62	Resp:	71068
Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	17.0









#43
Isopropyl Acetate
Concen: 21.686 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 43 Resp: 105036
Ion Ratio Lower Upper
43 100
61 25.2 20.7 31.1
87 12.1 9.8 14.8



#44
Trichloroethene
Concen: 20.374 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:130 Resp: 42674
Ion Ratio Lower Upper
130 100
95 100.5 0.0 195.8





#45

1,2-Dichloropropane

Concen: 20.743 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBS01

Tgt Ion: 63 Resp: 47674

Ion Ratio Lower Upper

63 100

65 31.1 25.6 38.4



#46

Dibromomethane

Concen: 20.395 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion: 93 Resp: 33823


Ion Ratio Lower Upper

93 100

95 81.5 64.7 97.1

174 90.4 69.0 103.4









#47
Bromodichloromethane
Concen: 21.697 ug/l
RT: 9.882 min Scan# 1348
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 83 Resp: 73309
Ion Ratio Lower Upper
83 100
85 61.0 51.2 76.8
127 8.4 6.5 9.7



#48
Methyl methacrylate
Concen: 21.823 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 41 Resp: 47567
Ion Ratio Lower Upper
41 100
69 81.0 64.7 97.1
39 65.5 49.0 73.6









#49
1,4-Dioxane
Concen: 420.430 ug/l
RT: 9.694 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 88 Resp: 15432
Ion Ratio Lower Upper
88 100
43 37.0 26.6 39.8
58 74.0 59.5 89.3



#50
Toluene-d8
Concen: 54.088 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 98 Resp: 410056
Ion Ratio Lower Upper
98 100
100 64.1 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 115.555 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 43 Resp: 324778
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7



#52
Toluene
Concen: 21.864 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 92 Resp: 113994
Ion Ratio Lower Upper
92 100
91 172.4 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 21.398 ug/l
 RT: 10.829 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01


Tgt Ion: 75 Resp: 68322
 Ion Ratio Lower Upper
 75 100
 77 29.3 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 21.530 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15

Tgt Ion: 75 Resp: 73429
 Ion Ratio Lower Upper
 75 100
 77 33.1 25.0 37.4
 39 49.3 43.1 64.7









#55
1,1,2-Trichloroethane
Concen: 21.912 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 97	Resp: 45212
Ion Ratio	Lower Upper
97 100	
83 92.7	69.0 103.6
85 59.4	44.6 66.8
99 63.7	48.6 72.8



#56
Ethyl methacrylate
Concen: 19.281 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 69	Resp: 62727		
Ion Ratio	Lower	Upper	
69 100			
41 69.3	54.6	82.0	
39 42.9	32.4	48.6	





#57

1,3-Dichloropropane

Concen: 21.915 ug/l

RT: 11.159 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBS01

Tgt Ion: 76 Resp: 78623

Ion Ratio Lower Upper

76 100

78 32.0 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 87.817 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion: 63 Resp: 114984

Ion Ratio Lower Upper

63 100

106 27.1 21.6 32.4









#59
2-Hexanone
Concen: 118.574 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 43 Resp: 234494
Ion Ratio Lower Upper
43 100
58 51.0 26.2 78.6



#60
Dibromochloromethane
Concen: 21.210 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 129 Resp: 52835
Ion Ratio Lower Upper
129 100
127 75.9 38.6 115.8









#61
1,2-Dibromoethane
Concen: 20.625 ug/l
RT: 11.471 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:107 Resp: 42361
Ion Ratio Lower Upper
107 100
109 95.9 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 52.029 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 95 Resp: 134929
Ion Ratio Lower Upper
95 100
174 75.1 0.0 145.0
176 70.9 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1170
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:117 Resp: 271043
Ion Ratio Lower Upper
117 100
82 58.5 48.6 72.8
119 31.9 26.6 39.8



#64
Tetrachloroethene
Concen: 21.658 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:164 Resp: 40019
Ion Ratio Lower Upper
164 100
166 128.2 103.4 155.2
129 95.9 79.2 118.8
131 96.1 77.1 115.7









#65
Chlorobenzene
Concen: 20.586 ug/l
RT: 11.888 min Scan# 111
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:112 Resp: 122235
Ion Ratio Lower Upper
112 100
114 31.6 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 21.124 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:131 Resp: 46034
Ion Ratio Lower Upper
131 100
133 95.1 47.4 142.3
119 65.4 33.1 99.5









#67
Ethyl Benzene
Concen: 20.797 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 91 Resp: 201123
Ion Ratio Lower Upper
91 100
106 28.8 23.8 35.8



#68
m/p-Xylenes
Concen: 43.106 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:106 Resp: 154069
Ion Ratio Lower Upper
106 100
91 208.4 167.7 251.5









#69
o-Xylene
Concen: 20.943 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:106 Resp: 71535
Ion Ratio Lower Upper
106 100
91 220.7 110.4 331.2



#70
Styrene
Concen: 21.250 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:104 Resp: 120125
Ion Ratio Lower Upper
104 100
78 53.0 42.5 63.7
103 55.9 43.8 65.8









#71
Bromoform
Concen: 22.516 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:173 Resp: 35108
Ion Ratio Lower Upper
173 100
175 48.0 24.4 73.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion:152 Resp: 126721
Ion Ratio Lower Upper
152 100
115 63.7 31.1 93.3
150 163.7 0.0 343.6









#73
Isopropylbenzene
Concen: 20.880 ug/l
RT: 12.694 min Scan# 1825
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:105 Resp: 178577
Ion Ratio Lower Upper
105 100
120 26.7 12.8 38.3



#74
N-ethyl acetate
Concen: 20.975 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 43 Resp: 80633
Ion Ratio Lower Upper
43 100
70 41.7 34.0 51.0
55 28.0 21.4 32.2
61 23.5 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 21.649 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01


Tgt Ion: 83 Resp: 65405
 Ion Ratio Lower Upper
 83 100
 131 8.8 4.8 14.4
 85 62.7 32.2 96.6



#76
 1,2,3-Trichloropropane
 Concen: 32.088 ug/l
 RT: 12.988 min Scan# 1876
 Delta R.T. -0.000 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15

Tgt Ion: 75 Resp: 82646
 Ion Ratio Lower Upper
 75 100
 77 131.7 109.7 329.2









#77
Bromobenzene
Concen: 21.332 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:156 Resp: 47665
Ion Ratio Lower Upper
156 100
77 228.4 114.1 342.4
158 94.0 48.9 146.8



#78
n-propylbenzene
Concen: 21.517 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 91 Resp: 217855
Ion Ratio Lower Upper
91 100
120 22.1 10.9 32.6





#79

2-Chlorotoluene

Concen: 21.081 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBS01

Tgt Ion: 91 Resp: 138138

Ion Ratio Lower Upper

91 100

126 31.9 15.7 47.1



#80

1,3,5-Trimethylbenzene

Concen: 21.836 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion:105 Resp: 154258

Ion Ratio Lower Upper

105 100

120 47.7 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 23.073 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 75 Resp: 21965
Ion Ratio Lower Upper
75 100
53 101.0 95.6 143.4
89 44.0 37.0 55.6



#82
4-Chlorotoluene
Concen: 21.369 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 91 Resp: 139420
Ion Ratio Lower Upper
91 100
126 31.6 15.9 47.7





#83

tert-Butylbenzene

Concen: 20.926 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01





Tgt Ion:119 Resp: 124147

Ion Ratio Lower Upper

119 100

91 71.0 35.1 105.3

134 27.7 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 22.092 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085529.D




Acq: 28 Jan 2025 11:15


Tgt Ion:105 Resp: 155548

Ion Ratio Lower Upper

105 100

120 44.5 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 21.691 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01


Tgt Ion:105 Resp: 178351
 Ion Ratio Lower Upper
 105 100
 134 19.5 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 20.952 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15

Tgt Ion:119 Resp: 143953
 Ion Ratio Lower Upper
 119 100
 134 25.1 12.7 38.0
 91 25.5 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 20.833 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01





Tgt Ion:146 Resp: 85729

Ion Ratio Lower Upper

146 100

111 41.8 21.4 64.3

148 62.6 32.3 96.9



#88

1,4-Dichlorobenzene

Concen: 19.789 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion:146 Resp: 84406


Ion Ratio Lower Upper

146 100

111 43.9 21.3 63.7

148 65.4 32.4 97.0









#89
n-Butylbenzene
Concen: 19.526 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion: 91 Resp: 115181
Ion Ratio Lower Upper
91 100
92 51.5 25.8 77.3
134 23.8 11.7 35.1



#90
Hexachloroethane
Concen: 19.522 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 117 Resp: 30558
Ion Ratio Lower Upper
117 100
201 70.0 33.7 101.0









#91
1,2-Dichlorobenzene
Concen: 20.350 ug/l
RT: 14.100 min Scan# 2065
Delta R.T. 0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBS01


Tgt Ion:146 Resp: 83529
Ion Ratio Lower Upper
146 100
111 42.3 21.7 65.1
148 63.4 31.4 94.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 23.494 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085529.D
Acq: 28 Jan 2025 11:15

Tgt Ion: 75 Resp: 12961
Ion Ratio Lower Upper
75 100
155 68.3 36.4 109.2
157 79.8 45.4 136.1









#93
 1,2,4-Trichlorobenzene
 Concen: 18.913 ug/l
 RT: 15.394 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBS01


Tgt Ion:180 Resp: 36081
 Ion Ratio Lower Upper
 180 100
 182 99.0 46.8 140.3
 145 32.9 16.0 48.0



#94
 Hexachlorobutadiene
 Concen: 19.907 ug/l
 RT: 15.494 min Scan# 2302
 Delta R.T. -0.006 min
 Lab File: VN085529.D
 Acq: 28 Jan 2025 11:15

Tgt Ion:225 Resp: 20163
 Ion Ratio Lower Upper
 225 100
 223 63.1 30.7 92.1
 227 62.5 30.9 92.5





#95

Naphthalene

Concen: 18.394 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBS01





Tgt Ion:128 Resp: 104713

Ion Ratio Lower Upper

128 100

127 13.6 10.6 16.0

129 11.0 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 18.978 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN085529.D

Acq: 28 Jan 2025 11:15




Tgt Ion:180 Resp: 36634

Ion Ratio Lower Upper

180 100

182 97.1 47.4 142.2

145 34.3 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085530.D
 Acq On : 28 Jan 2025 11:48
 Operator : JC\MD
 Sample : VN0128WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01

Quant Time: Jan 29 00:36:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	182154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	309854	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127977	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	147838	50.279	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.560%	
35) Dibromofluoromethane	8.165	113	107958	50.222	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.440%	
50) Toluene-d8	10.565	98	390099	51.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.160%	
62) 4-Bromofluorobenzene	12.847	95	133187	50.978	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.960%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	43993	17.838	ug/l	98
3) Chloromethane	2.359	50	47315	17.719	ug/l	99
4) Vinyl Chloride	2.512	62	47895	17.844	ug/l	98
5) Bromomethane	2.959	94	29197	18.008	ug/l	93
6) Chloroethane	3.118	64	33369	19.610	ug/l	98
7) Trichlorofluoromethane	3.501	101	75960	19.504	ug/l	99
8) Diethyl Ether	3.965	74	26094	19.395	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.371	101	44549	20.309	ug/l	98
10) Methyl Iodide	4.583	142	50174	19.974	ug/l	99
11) Tert butyl alcohol	5.518	59	39612	117.652	ug/l	98
12) 1,1-Dichloroethene	4.342	96	38078	19.478	ug/l	96
13) Acrolein	4.183	56	40971	89.143	ug/l	96
14) Allyl chloride	5.024	41	61288	19.320	ug/l	98
15) Acrylonitrile	5.718	53	117305	109.689	ug/l	97
16) Acetone	4.424	43	102700	108.100	ug/l	95
17) Carbon Disulfide	4.712	76	97539	16.205	ug/l	99
18) Methyl Acetate	5.024	43	67889	23.500	ug/l	100
19) Methyl tert-butyl Ether	5.795	73	139282	21.942	ug/l	100
20) Methylene Chloride	5.277	84	46285	19.680	ug/l	93
21) trans-1,2-Dichloroethene	5.783	96	39520	18.916	ug/l	94
22) Diisopropyl ether	6.665	45	150475	21.371	ug/l	98
23) Vinyl Acetate	6.600	43	520716	105.676	ug/l	100
24) 1,1-Dichloroethane	6.565	63	87788	20.448	ug/l	99
25) 2-Butanone	7.477	43	154236	110.320	ug/l	96
26) 2,2-Dichloropropane	7.489	77	75260	21.682	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	49722	20.206	ug/l	98
28) Bromochloromethane	7.812	49	33640	16.842	ug/l	97
29) Tetrahydrofuran	7.841	42	101879	114.937	ug/l	100
30) Chloroform	7.965	83	91378	20.593	ug/l	99
31) Cyclohexane	8.253	56	65567	17.571	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	78366	20.134	ug/l	98
36) 1,1-Dichloropropene	8.371	75	57513	19.063	ug/l	99
37) Ethyl Acetate	7.553	43	64941	21.326	ug/l	99
38) Carbon Tetrachloride	8.359	117	69210	20.038	ug/l	96
39) Methylcyclohexane	9.600	83	52898	18.659	ug/l	99
40) Benzene	8.606	78	183341	20.225	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085530.D
 Acq On : 28 Jan 2025 11:48
 Operator : JC\MD
 Sample : VN0128WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01

Quant Time: Jan 29 00:36:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	33920	21.410	ug/l	98
42) 1,2-Dichloroethane	8.665	62	73450	21.514	ug/l	100
43) Isopropyl Acetate	8.683	43	105226	21.565	ug/l	100
44) Trichloroethene	9.353	130	40907	19.386	ug/l	99
45) 1,2-Dichloropropane	9.618	63	47937	20.703	ug/l	98
46) Dibromomethane	9.706	93	35054	20.981	ug/l	98
47) Bromodichloromethane	9.882	83	73992	21.738	ug/l	95
48) Methyl methacrylate	9.677	41	48128	21.918	ug/l	97
49) 1,4-Dioxane	9.688	88	17249	466.467	ug/l #	80
51) 4-Methyl-2-Pentanone	10.441	43	334158	118.016	ug/l	100
52) Toluene	10.624	92	113049	21.523	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	69056	21.468	ug/l	95
54) cis-1,3-Dichloropropene	10.312	75	74974	21.821	ug/l	95
55) 1,1,2-Trichloroethane	11.012	97	44789	21.547	ug/l	92
56) Ethyl methacrylate	10.871	69	65671	19.962	ug/l	98
57) 1,3-Dichloropropane	11.159	76	78972	21.850	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	129996	98.550	ug/l	100
59) 2-Hexanone	11.194	43	235784	118.347	ug/l	99
60) Dibromochloromethane	11.359	129	54955	21.899	ug/l	99
61) 1,2-Dibromoethane	11.465	107	43841	21.189	ug/l	98
64) Tetrachloroethene	11.100	164	39045	20.962	ug/l	97
65) Chlorobenzene	11.888	112	123448	20.625	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	47119	21.449	ug/l	97
67) Ethyl Benzene	11.959	91	200961	20.614	ug/l	99
68) m/p-Xylenes	12.071	106	156444	43.421	ug/l	98
69) o-Xylene	12.394	106	71566	20.785	ug/l	98
70) Styrene	12.412	104	121646	21.347	ug/l	98
71) Bromoform	12.576	173	35363	22.498	ug/l #	99
73) Isopropylbenzene	12.694	105	183110	21.200	ug/l	100
74) N-amyl acetate	12.494	43	84761	21.833	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	66402	21.763	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	84360	32.432	ug/l	43
77) Bromobenzene	12.976	156	47394	21.002	ug/l	100
78) n-propylbenzene	13.035	91	222149	21.726	ug/l	99
79) 2-Chlorotoluene	13.123	91	143547	21.691	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	155679	21.820	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	21365	22.223	ug/l	97
82) 4-Chlorotoluene	13.218	91	141708	21.506	ug/l	99
83) tert-Butylbenzene	13.435	119	126399	21.097	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	158115	22.236	ug/l	99
85) sec-Butylbenzene	13.612	105	176846	21.297	ug/l	100
86) p-Isopropyltoluene	13.729	119	144432	20.822	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	87193	20.981	ug/l	97
88) 1,4-Dichlorobenzene	13.812	146	87963	20.420	ug/l	98
89) n-Butylbenzene	14.053	91	121076	20.324	ug/l	99
90) Hexachloroethane	14.329	117	31810	20.123	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	84554	20.397	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12008	21.553	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	37673	19.554	ug/l	96
94) Hexachlorobutadiene	15.500	225	20093	19.644	ug/l	98
95) Naphthalene	15.641	128	110672	19.250	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	39684	20.356	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085530.D
Acq On : 28 Jan 2025 11:48
Operator : JC\MD
Sample : VN0128WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Quant Time: Jan 29 00:36:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:168 Resp: 182154
Ion Ratio Lower Upper
168 100
99 63.5 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 17.838 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 85 Resp: 43993
Ion Ratio Lower Upper
85 100
87 32.3 16.7 50.1









#3
 Chloromethane
 Concen: 17.719 ug/l
 RT: 2.359 min Scan# 69
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion: 50 Resp: 47315
 Ion Ratio Lower Upper
 50 100
 52 33.4 27.0 40.6



#4
 Vinyl Chloride
 Concen: 17.844 ug/l
 RT: 2.512 min Scan# 95
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion: 62 Resp: 47895
 Ion Ratio Lower Upper
 62 100
 64 32.3 24.8 37.2









#5
Bromomethane
Concen: 18.008 ug/l
RT: 2.959 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 94 Resp: 29197
Ion Ratio Lower Upper
94 100
96 95.9 71.8 107.6



#6
Chloroethane
Concen: 19.610 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 64 Resp: 33369
Ion Ratio Lower Upper
64 100
66 29.4 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 19.504 ug/l
 RT: 3.501 min Scan# 210
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion:101 Resp: 75960
 Ion Ratio Lower Upper
 101 100
 103 65.4 51.4 77.2



#8
 Diethyl Ether
 Concen: 19.395 ug/l
 RT: 3.965 min Scan# 342
 Delta R.T. 0.012 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion: 74 Resp: 26094
 Ion Ratio Lower Upper
 74 100
 45 100.4 49.7 149.1









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 20.309 ug/l
 RT: 4.371 min Scan# 411
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion:101 Resp: 44549
 Ion Ratio Lower Upper
 101 100
 85 45.2 37.8 56.8
 151 72.2 58.8 88.2



#10
 Methyl Iodide
 Concen: 19.974 ug/l
 RT: 4.583 min Scan# 447
 Delta R.T. -0.006 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion:142 Resp: 50174
 Ion Ratio Lower Upper
 142 100
 127 50.2 40.3 60.5
 141 13.9 12.8 19.2









#11
Tert butyl alcohol
Concen: 117.652 ug/l
RT: 5.518 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 59 Resp: 39612
Ion Ratio Lower Upper
59 100
57 10.2 8.6 13.0



#12
1,1-Dichloroethene
Concen: 19.478 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 96 Resp: 38078
Ion Ratio Lower Upper
96 100
61 180.1 141.0 211.4
98 65.7 49.0 73.6









#13
Acrolein
Concen: 89.143 ug/l
RT: 4.183 min Scan# 31
Delta R.T. 0.012 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 56 Resp: 40971
Ion Ratio Lower Upper
56 100
55 73.7 56.3 84.5



#14
Allyl chloride
Concen: 19.320 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 41 Resp: 61288
Ion Ratio Lower Upper
41 100
39 78.5 64.4 96.6
76 36.5 30.5 45.7









#15
Acrylonitrile
Concen: 109.689 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 53 Resp: 117305
Ion Ratio Lower Upper
53 100
52 84.6 65.5 98.3
51 38.2 29.8 44.8



#16
Acetone
Concen: 108.100 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 43 Resp: 102700
Ion Ratio Lower Upper
43 100
58 26.8 23.8 35.6









#17
Carbon Disulfide
Concen: 16.205 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 76 Resp: 97539
Ion Ratio Lower Upper
76 100
78 8.1 6.9 10.3



#18
Methyl Acetate
Concen: 23.500 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 43 Resp: 67889
Ion Ratio Lower Upper
43 100
74 23.5 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 21.942 ug/l
RT: 5.795 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 73 Resp: 139282
Ion Ratio Lower Upper
73 100
57 22.1 17.8 26.8



#20
Methylene Chloride
Concen: 19.680 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 84 Resp: 46285
Ion Ratio Lower Upper
84 100
49 120.0 104.1 156.1
51 43.7 31.0 46.4
86 67.5 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 18.916 ug/l
RT: 5.783 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 96 Resp: 39520
Ion Ratio Lower Upper
96 100
61 153.6 115.7 173.5
98 58.7 49.8 74.8



#22
Diisopropyl ether
Concen: 21.371 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 45 Resp: 150475
Ion Ratio Lower Upper
45 100
43 57.5 45.0 67.6
87 25.7 21.8 32.8
59 11.4 10.2 15.2









#23
 Vinyl Acetate
 Concen: 105.676 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion: 43 Resp: 520716
 Ion Ratio Lower Upper
 43 100
 86 10.6 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 20.448 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion: 63 Resp: 87788
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.3 9.8
 100 4.1 2.0 6.0









#25
2-Butanone
Concen: 110.320 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 43 Resp: 154236
Ion Ratio Lower Upper
43 100
72 27.2 20.2 30.4



#26
2,2-Dichloropropane
Concen: 21.682 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 77 Resp: 75260
Ion Ratio Lower Upper
77 100
97 21.8 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 20.206 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion: 96 Resp: 49722
 Ion Ratio Lower Upper
 96 100
 61 158.8 0.0 311.8
 98 64.8 0.0 126.0



#28
 Bromochloromethane
 Concen: 16.842 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion: 49 Resp: 33640
 Ion Ratio Lower Upper
 49 100
 129 1.3 0.0 4.4
 130 71.1 55.0 82.4





#29

Tetrahydrofuran

Concen: 114.937 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. 0.012 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01





Tgt Ion: 42 Resp: 101879

Ion Ratio Lower Upper

42 100

72 43.2 34.2 51.2

71 40.7 32.5 48.7



#30

Chloroform

Concen: 20.593 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 83 Resp: 91378

Ion Ratio Lower Upper

83 100

85 65.4 51.8 77.6









#31
Cyclohexane
Concen: 17.571 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:	56	Resp:	65567
Ion Ratio	Lower	Upper	
56	100		
69	29.3	22.2	33.4
84	84.2	66.4	99.6



#32
1,1,1-Trichloroethane
Concen: 20.134 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:	97	Resp:	78366
Ion Ratio	Lower	Upper	
97	100		
99	63.6	49.8	74.6
61	50.8	41.4	62.2









#33
 1,2-Dichloroethane-d4
 Concen: 50.279 ug/l
 RT: 8.577 min Scan# 1125
 Delta R.T. 0.006 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion: 65 Resp: 147838
 Ion Ratio Lower Upper
 65 100
 67 51.5 0.0 101.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. 0.006 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion: 114 Resp: 309854
 Ion Ratio Lower Upper
 114 100
 63 22.7 0.0 47.6
 88 16.2 0.0 32.6









#35
Dibromofluoromethane
Concen: 50.222 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:113 Resp: 107958
Ion Ratio Lower Upper
113 100
111 101.5 82.7 124.1
192 17.7 14.3 21.5



#36
1,1-Dichloropropene
Concen: 19.063 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 75 Resp: 57513
Ion Ratio Lower Upper
75 100
110 33.1 16.5 49.5
77 29.9 24.4 36.6









#37
Ethyl Acetate
Concen: 21.326 ug/l
RT: 7.553 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 43 Resp: 64941
Ion Ratio Lower Upper
43 100
61 13.2 10.9 16.3
70 10.2 7.5 11.3



#38
Carbon Tetrachloride
Concen: 20.038 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 117 Resp: 69210
Ion Ratio Lower Upper
117 100
119 98.5 75.4 113.2
121 29.3 24.6 37.0





#39

Methylcyclohexane

Concen: 18.659 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01





Tgt Ion: 83 Resp: 52898

Ion Ratio Lower Upper

83 100

55 79.0 62.6 94.0

98 48.2 37.7 56.5



#40

Benzene

Concen: 20.225 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 78 Resp: 183341

Ion Ratio Lower Upper

78 100

77 23.3 19.0 28.6









#41
Methacrylonitrile
Concen: 21.410 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:	41	Resp:	33920
Ion	Ratio	Lower	Upper
41	100		
39	55.6	46.0	69.0
67	73.8	57.4	86.2
52	32.7	25.5	38.3



#42
1,2-Dichloroethane
Concen: 21.514 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:	62	Resp:	73450
Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	17.0









#43
Isopropyl Acetate
Concen: 21.565 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 43 Resp: 105226
Ion Ratio Lower Upper
43 100
61 26.0 20.7 31.1
87 12.7 9.8 14.8



#44
Trichloroethene
Concen: 19.386 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:130 Resp: 40907
Ion Ratio Lower Upper
130 100
95 97.3 0.0 195.8





#45

1,2-Dichloropropane

Concen: 20.703 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBSD01

Tgt Ion: 63 Resp: 47937

Ion Ratio Lower Upper

63 100

65 30.8 25.6 38.4



#46

Dibromomethane

Concen: 20.981 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 93 Resp: 35054


Ion Ratio Lower Upper

93 100

95 82.8 64.7 97.1

174 88.1 69.0 103.4









#47
Bromodichloromethane
Concen: 21.738 ug/l
RT: 9.882 min Scan# 1348
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 83 Resp: 73992
Ion Ratio Lower Upper
83 100
85 60.1 51.2 76.8
127 8.5 6.5 9.7



#48
Methyl methacrylate
Concen: 21.918 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 41 Resp: 48128
Ion Ratio Lower Upper
41 100
69 81.9 64.7 97.1
39 65.4 49.0 73.6





#49

1,4-Dioxane

Concen: 466.467 ug/l

RT: 9.688 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01





Tgt Ion: 88 Resp: 17249

Ion Ratio Lower Upper

88 100

43 0.0 26.6 39.8#

58 72.4 59.5 89.3



#50

Toluene-d8

Concen: 51.076 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 98 Resp: 390099

Ion Ratio Lower Upper

98 100

100 65.0 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 118.016 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 43 Resp: 334158
Ion Ratio Lower Upper
43 100
58 37.9 30.5 45.7



#52
Toluene
Concen: 21.523 ug/l
RT: 10.624 min Scan# 1474
Delta R.T. -0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 92 Resp: 113049
Ion Ratio Lower Upper
92 100
91 170.5 139.2 208.8









#53
t-1,3-Dichloropropene
Concen: 21.468 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 75 Resp: 69056
Ion Ratio Lower Upper
75 100
77 32.1 23.5 35.3



#54
cis-1,3-Dichloropropene
Concen: 21.821 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 75 Resp: 74974
Ion Ratio Lower Upper
75 100
77 31.2 25.0 37.4
39 48.6 43.1 64.7









#55
1,1,2-Trichloroethane
Concen: 21.547 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:	97	Resp:	44789
Ion	Ratio	Lower	Upper
97	100		
83	92.0	69.0	103.6
85	61.6	44.6	66.8
99	68.5	48.6	72.8



#56
Ethyl methacrylate
Concen: 19.962 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:	69	Resp:	65671
Ion	Ratio	Lower	Upper
69	100		
41	70.1	54.6	82.0
39	41.9	32.4	48.6





#57

1,3-Dichloropropane

Concen: 21.850 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :





VN0128WBSD01

Tgt Ion: 76 Resp: 78972

Ion Ratio Lower Upper

76 100

78 31.8 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 98.550 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN085530.D




Acq: 28 Jan 2025 11:48


Tgt Ion: 63 Resp: 129996

Ion Ratio Lower Upper

63 100

106 27.2 21.6 32.4









#59
2-Hexanone
Concen: 118.347 ug/l
RT: 11.194 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 43 Resp: 235784
Ion Ratio Lower Upper
43 100
58 52.0 26.2 78.6



#60
Dibromochloromethane
Concen: 21.899 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 129 Resp: 54955
Ion Ratio Lower Upper
129 100
127 76.6 38.6 115.8









#61
1,2-Dibromoethane
Concen: 21.189 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:107 Resp: 43841
Ion Ratio Lower Upper
107 100
109 93.3 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 50.978 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 95 Resp: 133187
Ion Ratio Lower Upper
95 100
174 75.7 0.0 145.0
176 70.9 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1170
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:117 Resp: 273221
Ion Ratio Lower Upper
117 100
82 58.5 48.6 72.8
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 20.962 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:164 Resp: 39045
Ion Ratio Lower Upper
164 100
166 131.5 103.4 155.2
129 101.5 79.2 118.8
131 101.9 77.1 115.7









#65
Chlorobenzene
Concen: 20.625 ug/l
RT: 11.888 min Scan# 111
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:112 Resp: 123448
Ion Ratio Lower Upper
112 100
114 31.9 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 21.449 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:131 Resp: 47119
Ion Ratio Lower Upper
131 100
133 92.4 47.4 142.3
119 63.4 33.1 99.5









#67
Ethyl Benzene
Concen: 20.614 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 91 Resp: 200961
Ion Ratio Lower Upper
91 100
106 30.6 23.8 35.8



#68
m/p-Xylenes
Concen: 43.421 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:106 Resp: 156444
Ion Ratio Lower Upper
106 100
91 206.6 167.7 251.5









#69
o-Xylene
Concen: 20.785 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:106 Resp: 71566
Ion Ratio Lower Upper
106 100
91 224.0 110.4 331.2



#70
Styrene
Concen: 21.347 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:104 Resp: 121646
Ion Ratio Lower Upper
104 100
78 54.5 42.5 63.7
103 56.3 43.8 65.8





#71

Bromoform

Concen: 22.498 ug/l

RT: 12.576 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01





Tgt Ion:173 Resp: 35363

Ion Ratio Lower Upper

173 100

175 48.3 24.4 73.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion:152 Resp: 127977


Ion Ratio Lower Upper

152 100

115 62.9 31.1 93.3

150 165.8 0.0 343.6









#73
Isopropylbenzene
Concen: 21.200 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:105 Resp: 183110
Ion Ratio Lower Upper
105 100
120 25.6 12.8 38.3



#74
N-ethyl acetate
Concen: 21.833 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 43 Resp: 84761
Ion Ratio Lower Upper
43 100
70 42.4 34.0 51.0
55 27.8 21.4 32.2
61 23.0 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 21.763 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion: 83 Resp: 66402
 Ion Ratio Lower Upper
 83 100
 131 9.9 4.8 14.4
 85 64.2 32.2 96.6



#76
 1,2,3-Trichloropropane
 Concen: 32.432 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion: 75 Resp: 84360
 Ion Ratio Lower Upper
 75 100
 77 128.3 109.7 329.2









#77
Bromobenzene
Concen: 21.002 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:156 Resp: 47394
Ion Ratio Lower Upper
156 100
77 228.3 114.1 342.4
158 96.4 48.9 146.8



#78
n-propylbenzene
Concen: 21.726 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 91 Resp: 222149
Ion Ratio Lower Upper
91 100
120 21.0 10.9 32.6









#79
2-Chlorotoluene
Concen: 21.691 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 91 Resp: 143547
Ion Ratio Lower Upper
91 100
126 30.9 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 21.820 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:105 Resp: 155679
Ion Ratio Lower Upper
105 100
120 47.5 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 22.223 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 75 Resp: 21365
Ion Ratio Lower Upper
75 100
53 115.4 95.6 143.4
89 44.8 37.0 55.6



#82
4-Chlorotoluene
Concen: 21.506 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 91 Resp: 141708
Ion Ratio Lower Upper
91 100
126 32.2 15.9 47.7









#83
tert-Butylbenzene
Concen: 21.097 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion:119 Resp: 126399
Ion Ratio Lower Upper
119 100
91 70.5 35.1 105.3
134 27.3 12.0 36.1



#84
1,2,4-Trimethylbenzene
Concen: 22.236 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:105 Resp: 158115
Ion Ratio Lower Upper
105 100
120 43.9 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 21.297 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion:105 Resp: 176846
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 20.822 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.006 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion:119 Resp: 144432
 Ion Ratio Lower Upper
 119 100
 134 26.1 12.7 38.0
 91 25.6 12.7 38.1









#87
 1,3-Dichlorobenzene
 Concen: 20.981 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion:146 Resp: 87193
 Ion Ratio Lower Upper
 146 100
 111 41.4 21.4 64.3
 148 62.2 32.3 96.9



#88
 1,4-Dichlorobenzene
 Concen: 20.420 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion:146 Resp: 87963
 Ion Ratio Lower Upper
 146 100
 111 41.6 21.3 63.7
 148 63.2 32.4 97.0









#89
n-Butylbenzene
Concen: 20.324 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01


Tgt Ion: 91 Resp: 121076
Ion Ratio Lower Upper
91 100
92 50.3 25.8 77.3
134 23.8 11.7 35.1



#90
Hexachloroethane
Concen: 20.123 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion: 117 Resp: 31810
Ion Ratio Lower Upper
117 100
201 66.3 33.7 101.0





#91

1,2-Dichlorobenzene

Concen: 20.397 ug/l

RT: 14.106 min Scan# 2065

Delta R.T. 0.006 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48

Instrument :

MSVOA_N

ClientSampleId :

VN0128WBSD01





Tgt Ion:146 Resp: 84554

Ion Ratio Lower Upper

146 100

111 44.6 21.7 65.1

148 63.0 31.4 94.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 21.553 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. 0.000 min

Lab File: VN085530.D

Acq: 28 Jan 2025 11:48




Tgt Ion: 75 Resp: 12008


Ion Ratio Lower Upper

75 100

155 73.1 36.4 109.2

157 95.1 45.4 136.1









#93
 1,2,4-Trichlorobenzene
 Concen: 19.554 ug/l
 RT: 15.388 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBSD01


Tgt Ion:180 Resp: 37673
 Ion Ratio Lower Upper
 180 100
 182 96.9 46.8 140.3
 145 35.9 16.0 48.0



#94
 Hexachlorobutadiene
 Concen: 19.644 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. 0.000 min
 Lab File: VN085530.D
 Acq: 28 Jan 2025 11:48

Tgt Ion:225 Resp: 20093
 Ion Ratio Lower Upper
 225 100
 223 63.7 30.7 92.1
 227 60.8 30.9 92.5









#95
Naphthalene
Concen: 19.250 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.006 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48




Instrument :
MSVOA_N
ClientSampleId :
VN0128WBSD01

Tgt Ion:128 Resp: 110672
Ion Ratio Lower Upper
128 100
127 13.4 10.6 16.0
129 11.0 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 20.356 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085530.D
Acq: 28 Jan 2025 11:48

Tgt Ion:180 Resp: 39684
Ion Ratio Lower Upper
180 100
182 93.4 47.4 142.2
145 33.7 16.9 50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085546.D
 Acq On : 28 Jan 2025 18:14
 Operator : JC\MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDC0050EC

Quant Time: Jan 29 00:43:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	187799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	314055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143565	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	165633	54.638	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.280%	
35) Dibromofluoromethane	8.165	113	120592	55.349	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.700%	
50) Toluene-d8	10.565	98	438659	56.666	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 113.340%#	
62) 4-Bromofluorobenzene	12.847	95	157860	59.614	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 119.220%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	109169	42.934	ug/l	94
3) Chloromethane	2.359	50	113732	41.310	ug/l	100
4) Vinyl Chloride	2.512	62	120445	43.525	ug/l	98
5) Bromomethane	2.936	94	67919	40.633	ug/l	97
6) Chloroethane	3.106	64	80051	45.629	ug/l	97
7) Trichlorofluoromethane	3.495	101	194108	48.343	ug/l	98
8) Diethyl Ether	3.959	74	65944	47.541	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	110829	49.006	ug/l	99
10) Methyl Iodide	4.589	142	128917	49.778	ug/l	98
11) Tert butyl alcohol	5.518	59	122061	351.637	ug/l	99
12) 1,1-Dichloroethene	4.336	96	98130	48.687	ug/l	97
13) Acrolein	4.177	56	129555	273.407	ug/l	100
14) Allyl chloride	5.024	41	155852	47.653	ug/l	98
15) Acrylonitrile	5.712	53	310806	281.891	ug/l	98
16) Acetone	4.424	43	267295	272.891	ug/l	99
17) Carbon Disulfide	4.712	76	245541	39.567	ug/l	99
18) Methyl Acetate	5.018	43	178998	60.097	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	358554	54.787	ug/l	99
20) Methylene Chloride	5.271	84	119628	49.335	ug/l	97
21) trans-1,2-Dichloroethene	5.783	96	103142	47.884	ug/l	97
22) Diisopropyl ether	6.671	45	392245	54.034	ug/l	99
23) Vinyl Acetate	6.600	43	1454703	286.348	ug/l	98
24) 1,1-Dichloroethane	6.565	63	222797	50.334	ug/l	98
25) 2-Butanone	7.482	43	423099	293.531	ug/l	99
26) 2,2-Dichloropropane	7.488	77	189076	52.834	ug/l	100
27) cis-1,2-Dichloroethene	7.482	96	130214	51.325	ug/l	99
28) Bromochloromethane	7.812	49	89552	43.487	ug/l	98
29) Tetrahydrofuran	7.835	42	275635	301.616	ug/l	100
30) Chloroform	7.965	83	232517	50.826	ug/l	99
31) Cyclohexane	8.253	56	166703	43.331	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	202362	50.429	ug/l	98
36) 1,1-Dichloropropene	8.371	75	152437	49.851	ug/l	99
37) Ethyl Acetate	7.559	43	171657	55.616	ug/l	98
38) Carbon Tetrachloride	8.359	117	179826	51.368	ug/l	98
39) Methylcyclohexane	9.600	83	143759	50.030	ug/l	98
40) Benzene	8.606	78	470586	51.218	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085546.D
 Acq On : 28 Jan 2025 18:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jan 29 00:43:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	97709	60.848	ug/l	97
42) 1,2-Dichloroethane	8.671	62	179469	51.863	ug/l	100
43) Isopropyl Acetate	8.688	43	345380	69.836	ug/l #	91
44) Trichloroethene	9.347	130	106986	50.023	ug/l	97
45) 1,2-Dichloropropane	9.618	63	122928	52.380	ug/l	98
46) Dibromomethane	9.706	93	87054	51.408	ug/l	98
47) Bromodichloromethane	9.888	83	188980	54.778	ug/l	97
48) Methyl methacrylate	9.676	41	134702	60.523	ug/l	98
49) 1,4-Dioxane	9.694	88	46510	1240.952	ug/l #	92
51) 4-Methyl-2-Pentanone	10.441	43	906426	315.843	ug/l	100
52) Toluene	10.629	92	289726	54.423	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	182074	55.846	ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	188960	54.260	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	115272	54.714	ug/l	98
56) Ethyl methacrylate	10.871	69	183142	51.541	ug/l	98
57) 1,3-Dichloropropane	11.159	76	203386	55.519	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	389176	291.087	ug/l	100
59) 2-Hexanone	11.194	43	659035	326.364	ug/l	99
60) Dibromochloromethane	11.359	129	139526	54.855	ug/l	98
61) 1,2-Dibromoethane	11.470	107	112753	53.765	ug/l	99
64) Tetrachloroethene	11.100	164	88612	45.330	ug/l	99
65) Chlorobenzene	11.888	112	315889	50.289	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	117993	51.180	ug/l	98
67) Ethyl Benzene	11.965	91	555215	54.268	ug/l	99
68) m/p-Xylenes	12.070	106	423444	111.987	ug/l	99
69) o-Xylene	12.394	106	199877	55.313	ug/l	99
70) Styrene	12.412	104	347052	58.032	ug/l	100
71) Bromoform	12.576	173	92343	55.980	ug/l #	99
73) Isopropylbenzene	12.694	105	518634	53.526	ug/l	99
74) N-amyl acetate	12.494	43	244112	56.051	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	176292	51.507	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	228606	78.343	ug/l	42
77) Bromobenzene	12.976	156	127032	50.181	ug/l	100
78) n-propylbenzene	13.035	91	616293	53.728	ug/l	99
79) 2-Chlorotoluene	13.123	91	384370	51.776	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	432662	54.059	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	61346	56.880	ug/l	90
82) 4-Chlorotoluene	13.217	91	383756	51.917	ug/l	99
83) tert-Butylbenzene	13.435	119	354690	52.772	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	443501	55.598	ug/l	99
85) sec-Butylbenzene	13.617	105	504931	54.205	ug/l	100
86) p-Isopropyltoluene	13.729	119	422476	50.325	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	233419	50.068	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	229722	47.539	ug/l	100
89) n-Butylbenzene	14.053	91	351230	52.557	ug/l	99
90) Hexachloroethane	14.329	117	86748	48.917	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	228524	49.142	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	34650	55.439	ug/l	99
93) 1,2,4-Trichlorobenzene	15.388	180	108928	50.399	ug/l	99
94) Hexachlorobutadiene	15.500	225	49834	43.430	ug/l	95
95) Naphthalene	15.641	128	342729	53.141	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	107618	49.210	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085546.D
Acq On : 28 Jan 2025 18:14
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Quant Time: Jan 29 00:43:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 29 00:43:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 187799
Ion Ratio Lower Upper
168 100
99 62.0 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 42.934 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 85 Resp: 109169
Ion Ratio Lower Upper
85 100
87 29.7 16.7 50.1









#3
 Chloromethane
 Concen: 41.310 ug/l
 RT: 2.359 min Scan# 69
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 50 Resp: 113732
 Ion Ratio Lower Upper
 50 100
 52 33.6 27.0 40.6



#4
 Vinyl Chloride
 Concen: 43.525 ug/l
 RT: 2.512 min Scan# 95
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion: 62 Resp: 120445
 Ion Ratio Lower Upper
 62 100
 64 32.0 24.8 37.2









#5
Bromomethane
Concen: 40.633 ug/l
RT: 2.936 min Scan# 11
Delta R.T. -0.018 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 67919
Ion Ratio Lower Upper
94 100
96 92.4 71.8 107.6



#6
Chloroethane
Concen: 45.629 ug/l
RT: 3.106 min Scan# 196
Delta R.T. -0.012 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 64 Resp: 80051
Ion Ratio Lower Upper
64 100
66 32.1 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 48.343 ug/l
 RT: 3.495 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 194108
 Ion Ratio Lower Upper
 101 100
 103 66.0 51.4 77.2



#8
 Diethyl Ether
 Concen: 47.541 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion: 74 Resp: 65944
 Ion Ratio Lower Upper
 74 100
 45 98.3 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.006 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 110829
Ion Ratio Lower Upper
101 100
85 46.6 37.8 56.8
151 72.6 58.8 88.2



#10
Methyl Iodide
Concen: 49.778 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:142 Resp: 128917
Ion Ratio Lower Upper
142 100
127 48.3 40.3 60.5
141 16.0 12.8 19.2









#11
Tert butyl alcohol
Concen: 351.637 ug/l
RT: 5.518 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 122061
Ion Ratio Lower Upper
59 100
57 11.2 8.6 13.0



#12
1,1-Dichloroethene
Concen: 48.687 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 96 Resp: 98130
Ion Ratio Lower Upper
96 100
61 180.0 141.0 211.4
98 63.4 49.0 73.6









#13
Acrolein
Concen: 273.407 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 129555
Ion Ratio Lower Upper
56 100
55 70.6 56.3 84.5



#14
Allyl chloride
Concen: 47.653 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 41 Resp: 155852
Ion Ratio Lower Upper
41 100
39 82.2 64.4 96.6
76 37.0 30.5 45.7









#15
Acrylonitrile
Concen: 281.891 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 310806
Ion Ratio Lower Upper
53 100
52 83.9 65.5 98.3
51 37.3 29.8 44.8



#16
Acetone
Concen: 272.891 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 43 Resp: 267295
Ion Ratio Lower Upper
43 100
58 30.0 23.8 35.6









#17
Carbon Disulfide
Concen: 39.567 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 245541
Ion Ratio Lower Upper
76 100
78 8.9 6.9 10.3



#18
Methyl Acetate
Concen: 60.097 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 43 Resp: 178998
Ion Ratio Lower Upper
43 100
74 22.9 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 54.787 ug/l
RT: 5.794 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 358554
Ion Ratio Lower Upper
73 100
57 21.7 17.8 26.8



#20
Methylene Chloride
Concen: 49.335 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 84 Resp: 119628
Ion Ratio Lower Upper
84 100
49 133.6 104.1 156.1
51 36.4 31.0 46.4
86 62.6 51.9 77.9









#21
 trans-1,2-Dichloroethene
 Concen: 47.884 ug/l
 RT: 5.783 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 96 Resp: 103142
 Ion Ratio Lower Upper
 96 100
 61 146.0 115.7 173.5
 98 66.8 49.8 74.8



#22
 Diisopropyl ether
 Concen: 54.034 ug/l
 RT: 6.671 min Scan# 802
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion: 45 Resp: 392245
 Ion Ratio Lower Upper
 45 100
 43 54.9 45.0 67.6
 87 26.8 21.8 32.8
 59 12.6 10.2 15.2









#23
 Vinyl Acetate
 Concen: 286.348 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1454703
 Ion Ratio Lower Upper
 43 100
 86 9.6 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 50.334 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion: 63 Resp: 222797
 Ion Ratio Lower Upper
 63 100
 98 5.9 3.3 9.8
 100 4.4 2.0 6.0









#25
2-Butanone
Concen: 293.531 ug/l
RT: 7.482 min Scan# 940
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 423099
Ion Ratio Lower Upper
43 100
72 24.7 20.2 30.4



#26
2,2-Dichloropropane
Concen: 52.834 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 77 Resp: 189076
Ion Ratio Lower Upper
77 100
97 21.4 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 51.325 ug/l
 RT: 7.482 min Scan# 940
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 96 Resp: 130214




Ion	Ratio	Lower	Upper
96	100		
61	155.9	0.0	311.8
98	64.7	0.0	126.0




#28
 Bromochloromethane
 Concen: 43.487 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion: 49 Resp: 89552

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.4
130	70.3	55.0	82.4





#29

Tetrahydrofuran

Concen: 301.616 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 275635

Ion Ratio Lower Upper

42 100

72 42.8 34.2 51.2

71 40.1 32.5 48.7



#30

Chloroform

Concen: 50.826 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 83 Resp: 232517

Ion Ratio Lower Upper

83 100

85 65.3 51.8 77.6









#31
Cyclohexane
Concen: 43.331 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 166703
Ion Ratio Lower Upper
56 100
69 29.8 22.2 33.4
84 80.6 66.4 99.6



#32
1,1,1-Trichloroethane
Concen: 50.429 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 97 Resp: 202362
Ion Ratio Lower Upper
97 100
99 65.0 49.8 74.6
61 51.3 41.4 62.2









#33
 1,2-Dichloroethane-d4
 Concen: 54.638 ug/l
 RT: 8.577 min Scan# 1125
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 65 Resp: 165633
 Ion Ratio Lower Upper
 65 100
 67 50.7 0.0 101.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion: 114 Resp: 314055
 Ion Ratio Lower Upper
 114 100
 63 22.0 0.0 47.6
 88 15.5 0.0 32.6









#35
Dibromofluoromethane
Concen: 55.349 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 120592
Ion Ratio Lower Upper
113 100
111 102.9 82.7 124.1
192 17.8 14.3 21.5



#36
1,1-Dichloropropene
Concen: 49.851 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 75 Resp: 152437
Ion Ratio Lower Upper
75 100
110 32.6 16.5 49.5
77 29.9 24.4 36.6









#37
Ethyl Acetate
Concen: 55.616 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 171657
Ion Ratio Lower Upper
43 100
61 12.5 10.9 16.3
70 9.5 7.5 11.3



#38
Carbon Tetrachloride
Concen: 51.368 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 117 Resp: 179826
Ion Ratio Lower Upper
117 100
119 95.0 75.4 113.2
121 28.1 24.6 37.0





#39

Methylcyclohexane

Concen: 50.030 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 143759

Ion Ratio Lower Upper

83 100

55 77.8 62.6 94.0

98 50.5 37.7 56.5



#40

Benzene

Concen: 51.218 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 78 Resp: 470586

Ion Ratio Lower Upper

78 100

77 24.3 19.0 28.6









#41
Methacrylonitrile
Concen: 60.848 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	97709
Ion	Ratio	Lower	Upper
41	100		
39	56.4	46.0	69.0
67	68.0	57.4	86.2
52	31.0	25.5	38.3



#42
1,2-Dichloroethane
Concen: 51.863 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:	62	Resp:	179469
Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	17.0





#43

Isopropyl Acetate

Concen: 69.836 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 345380

Ion Ratio Lower Upper

43 100

61 20.6 20.7 31.1#

87 10.2 9.8 14.8



#44

Trichloroethene

Concen: 50.023 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion:130 Resp: 106986

Ion Ratio Lower Upper

130 100

95 101.0 0.0 195.8





#45

1,2-Dichloropropane

Concen: 52.380 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 122928

Ion Ratio Lower Upper

63 100

65 31.0 25.6 38.4



#46

Dibromomethane

Concen: 51.408 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 93 Resp: 87054


Ion Ratio Lower Upper

93 100

95 83.2 64.7 97.1

174 88.2 69.0 103.4









#47
Bromodichloromethane
Concen: 54.778 ug/l
RT: 9.888 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 188980
Ion Ratio Lower Upper
83 100
85 61.4 51.2 76.8
127 8.1 6.5 9.7



#48
Methyl methacrylate
Concen: 60.523 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.001 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 41 Resp: 134702
Ion Ratio Lower Upper
41 100
69 80.5 64.7 97.1
39 64.5 49.0 73.6





#49

1,4-Dioxane

Concen: 1240.952 ug/l

RT: 9.694 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 46510

Ion Ratio Lower Upper

88 100

43 23.4 26.6 39.8#

58 71.6 59.5 89.3



#50

Toluene-d8

Concen: 56.666 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 98 Resp: 438659

Ion Ratio Lower Upper

98 100

100 65.0 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 315.843 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 906426
Ion Ratio Lower Upper
43 100
58 38.0 30.5 45.7



#52
Toluene
Concen: 54.423 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 92 Resp: 289726
Ion Ratio Lower Upper
92 100
91 174.4 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 55.846 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 75 Resp: 182074
 Ion Ratio Lower Upper
 75 100
 77 31.2 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 54.260 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion: 75 Resp: 188960
 Ion Ratio Lower Upper
 75 100
 77 33.1 25.0 37.4
 39 55.1 43.1 64.7









#55
1,1,2-Trichloroethane
Concen: 54.714 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 97 Resp: 115272




Ion	Ratio	Lower	Upper
97	100		
83	85.0	69.0	103.6
85	53.3	44.6	66.8
99	60.7	48.6	72.8




#56
Ethyl methacrylate
Concen: 51.541 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 69 Resp: 183142

Ion	Ratio	Lower	Upper
69	100		
41	69.8	54.6	82.0
39	41.8	32.4	48.6





#57

1,3-Dichloropropane

Concen: 55.519 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 203386

Ion Ratio Lower Upper

76 100

78 31.9 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 291.087 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 63 Resp: 389176

Ion Ratio Lower Upper

63 100

106 27.0 21.6 32.4









#59
2-Hexanone
Concen: 326.364 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 659035
Ion Ratio Lower Upper
43 100
58 52.0 26.2 78.6



#60
Dibromochloromethane
Concen: 54.855 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:129 Resp: 139526
Ion Ratio Lower Upper
129 100
127 78.8 38.6 115.8









#61
1,2-Dibromoethane
Concen: 53.765 ug/l
RT: 11.470 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 112753
Ion Ratio Lower Upper
107 100
109 94.2 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 59.614 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 95 Resp: 157860
Ion Ratio Lower Upper
95 100
174 73.2 0.0 145.0
176 72.3 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 286739
Ion Ratio Lower Upper
117 100
82 59.0 48.6 72.8
119 32.6 26.6 39.8



#64
Tetrachloroethene
Concen: 45.330 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:164 Resp: 88612
Ion Ratio Lower Upper
164 100
166 127.2 103.4 155.2
129 99.0 79.2 118.8
131 95.8 77.1 115.7









#65
Chlorobenzene
Concen: 50.289 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 315889
Ion Ratio Lower Upper
112 100
114 32.3 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 51.180 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:131 Resp: 117993
Ion Ratio Lower Upper
131 100
133 96.4 47.4 142.3
119 65.0 33.1 99.5





#67

Ethyl Benzene

Concen: 54.268 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 555215

Ion Ratio Lower Upper

91 100

106 30.5 23.8 35.8



#68

m/p-Xylenes

Concen: 111.987 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion:106 Resp: 423444

Ion Ratio Lower Upper

106 100

91 207.7 167.7 251.5









#69
o-Xylene
Concen: 55.313 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 199877
Ion Ratio Lower Upper
106 100
91 222.8 110.4 331.2



#70
Styrene
Concen: 58.032 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:104 Resp: 347052
Ion Ratio Lower Upper
104 100
78 53.0 42.5 63.7
103 55.1 43.8 65.8









#71
 Bromoform
 Concen: 55.980 ug/l
 RT: 12.576 min Scan# 1806
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:	173	Resp:	92343
Ion Ratio	Lower	Upper	
173	100		
175	49.3	24.4	73.2
254	0.0	0.0	0.0



#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.788 min Scan# 2012
 Delta R.T. -0.000 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion:	152	Resp:	143565
Ion Ratio	Lower	Upper	
152	100		
115	62.2	31.1	93.3
150	173.3	0.0	343.6





#73

Isopropylbenzene

Concen: 53.526 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 518634

Ion Ratio Lower Upper

105 100

120 25.8 12.8 38.3



#74

N-ethyl acetate

Concen: 56.051 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Tgt Ion: 43 Resp: 244112




Ion Ratio Lower Upper


43 100

70 42.5 34.0 51.0

55 26.1 21.4 32.2

61 23.3 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 51.507 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 176292

Ion Ratio Lower Upper

83 100

131 9.4 4.8 14.4

85 64.2 32.2 96.6



#76

1,2,3-Trichloropropane

Concen: 78.343 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. 0.006 min

Lab File: VN085546.D




Acq: 28 Jan 2025 18:14


Tgt Ion: 75 Resp: 228606

Ion Ratio Lower Upper

75 100

77 127.2 109.7 329.2









#77
Bromobenzene
Concen: 50.181 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:156 Resp: 127032
Ion Ratio Lower Upper
156 100
77 228.8 114.1 342.4
158 97.3 48.9 146.8



#78
n-propylbenzene
Concen: 53.728 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 91 Resp: 616293
Ion Ratio Lower Upper
91 100
120 22.1 10.9 32.6









#79
2-Chlorotoluene
Concen: 51.776 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 384370
Ion Ratio Lower Upper
91 100
126 32.3 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 54.059 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:105 Resp: 432662
Ion Ratio Lower Upper
105 100
120 48.3 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 56.880 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 61346
Ion Ratio Lower Upper
75 100
53 105.0 95.6 143.4
89 45.8 37.0 55.6



#82
4-Chlorotoluene
Concen: 51.917 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 91 Resp: 383756
Ion Ratio Lower Upper
91 100
126 32.3 15.9 47.7









#83
tert-Butylbenzene
Concen: 52.772 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:119 Resp: 354690
Ion Ratio Lower Upper
119 100
91 71.2 35.1 105.3
134 27.6 12.0 36.1



#84
1,2,4-Trimethylbenzene
Concen: 55.598 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:105 Resp: 443501
Ion Ratio Lower Upper
105 100
120 44.0 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 54.205 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. 0.005 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 504931
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 50.325 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.006 min
 Lab File: VN085546.D
 Acq: 28 Jan 2025 18:14

Tgt Ion:119 Resp: 422476
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.7 38.0
 91 25.0 12.7 38.1









#87
1,3-Dichlorobenzene
Concen: 50.068 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 233419
Ion Ratio Lower Upper
146 100
111 42.4 21.4 64.3
148 64.2 32.3 96.9



#88
1,4-Dichlorobenzene
Concen: 47.539 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:146 Resp: 229722
Ion Ratio Lower Upper
146 100
111 42.1 21.3 63.7
148 64.8 32.4 97.0









#89
n-Butylbenzene
Concen: 52.557 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 351230
Ion Ratio Lower Upper
91 100
92 50.9 25.8 77.3
134 24.0 11.7 35.1



#90
Hexachloroethane
Concen: 48.917 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion: 117 Resp: 86748
Ion Ratio Lower Upper
117 100
201 68.1 33.7 101.0





#91

1,2-Dichlorobenzene

Concen: 49.142 ug/l

RT: 14.106 min Scan# 2065

Delta R.T. 0.006 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 228524

Ion Ratio Lower Upper

146 100

111 43.9 21.7 65.1

148 62.5 31.4 94.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 55.439 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion: 75 Resp: 34650


Ion Ratio Lower Upper

75 100

155 71.7 36.4 109.2

157 91.5 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 50.399 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 108928

Ion Ratio Lower Upper

180 100

182 92.9 46.8 140.3

145 31.5 16.0 48.0



#94

Hexachlorobutadiene

Concen: 43.430 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN085546.D

Acq: 28 Jan 2025 18:14




Tgt Ion:225 Resp: 49834


Ion Ratio Lower Upper

225 100

223 65.2 30.7 92.1

227 65.8 30.9 92.5









#95
Naphthalene
Concen: 53.141 ug/l
RT: 15.641 min Scan# 2326
Delta R.T. 0.006 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:128 Resp: 342729
Ion Ratio Lower Upper
128 100
127 13.3 10.6 16.0
129 11.0 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 49.210 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN085546.D
Acq: 28 Jan 2025 18:14

Tgt Ion:180 Resp: 107618
Ion Ratio Lower Upper
180 100
182 95.6 47.4 142.2
145 34.4 16.9 50.7

