

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1199

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN085530.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0128WBSD01	Chloromethane	20	17.7	ug/L	89	3		50	139	20
	Vinyl chloride	20	17.8	ug/L	89	2		58	137	20
	Bromomethane	20	18.0	ug/L	90	9		53	141	20
	Chloroethane	20	19.6	ug/L	98	1		60	138	20
	Trichlorofluoromethane	20	19.5	ug/L	98	6		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	20.3	ug/L	102	1		70	136	20
	1,1-Dichloroethene	20	19.5	ug/L	98	2		71	131	20
	Acetone	100	110	ug/L	110	10		39	160	20
	Carbon disulfide	20	16.2	ug/L	81	1		64	133	20
	Methyl tert-butyl Ether	20	21.9	ug/L	110	6		71	124	20
	Methylene Chloride	20	19.7	ug/L	99	0		74	124	20
	trans-1,2-Dichloroethene	20	18.9	ug/L	95	0		75	124	20
	1,1-Dichloroethane	20	20.4	ug/L	102	1		77	125	20
	2-Butanone	100	110	ug/L	110	0		56	143	20
	Carbon Tetrachloride	20	20.0	ug/L	100	3		72	136	20
	cis-1,2-Dichloroethene	20	20.2	ug/L	101	1		78	123	20
	Chloroform	20	20.6	ug/L	103	0		79	124	20
	1,1,1-Trichloroethane	20	20.1	ug/L	101	1		74	131	20
	Methylcyclohexane	20	18.7	ug/L	94	1		72	132	20
	Benzene	20	20.2	ug/L	101	3		79	120	20
	1,2-Dichloroethane	20	21.5	ug/L	108	3		73	128	20
	Trichloroethene	20	19.4	ug/L	97	5		79	123	20
	1,2-Dichloropropane	20	20.7	ug/L	104	0		78	122	20
	Bromodichloromethane	20	21.7	ug/L	109	0		79	125	20
	4-Methyl-2-Pentanone	100	120	ug/L	120	0		67	130	20
	Toluene	20	21.5	ug/L	108	2		80	121	20
	t-1,3-Dichloropropene	20	21.5	ug/L	108	1		73	127	20
	cis-1,3-Dichloropropene	20	21.8	ug/L	109	1		75	124	20
	1,1,2-Trichloroethane	20	21.5	ug/L	108	2		80	119	20
	2-Hexanone	100	120	ug/L	120	0		57	139	20
	Dibromochloromethane	20	21.9	ug/L	110	4		74	126	20
	Tetrachloroethene	20	21.0	ug/L	105	4		74	129	20
	Chlorobenzene	20	20.6	ug/L	103	0		82	118	20
	Ethyl Benzene	20	20.6	ug/L	103	1		79	121	20
	m/p-Xylenes	40	43.4	ug/L	109	1		80	121	20
	o-Xylene	20	20.8	ug/L	104	0		78	122	20
	Styrene	20	21.3	ug/L	106	0		78	123	20
	Bromoform	20	22.5	ug/L	113	0		66	130	20
	Isopropylbenzene	20	21.2	ug/L	106	2		72	131	20
	1,1,2,2-Tetrachloroethane	20	21.8	ug/L	109	1		71	121	20
	1,3-Dichlorobenzene	20	21.0	ug/L	105	1		80	119	20
	1,4-Dichlorobenzene	20	20.4	ug/L	102	3		79	118	20
	1,2-Dichlorobenzene	20	20.4	ug/L	102	0		80	119	20