

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1199 SAS No.: Q1199 SDG NO.: Q1199
Lab File ID: VN085525.D BFB Injection Date: 01/28/2025
Instrument ID: MSVOA_N BFB Injection Time: 09:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.1
75	30.0 - 60.0% of mass 95	55.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.4
173	Less than 2.0% of mass 174	1.5 (1.9) 1
174	50.0 - 100.0% of mass 95	75.9
175	5.0 - 9.0% of mass 174	5.5 (7.2) 1
176	95.0 - 101.0% of mass 174	73.7 (97.2) 1
177	5.0 - 9.0% of mass 176	5 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN085526.D	01/28/2025	09:53
VN0128WBL01	VN0128WBL01	VN085528.D	01/28/2025	10:51
VN0128WBS01	VN0128WBS01	VN085529.D	01/28/2025	11:15
VN0128WBSD01	VN0128WBSD01	VN085530.D	01/28/2025	11:48
BP-VPB-192-TB-20250123	Q1199-01	VN085534.D	01/28/2025	13:24
BP-VPB-192-EB-20250124	Q1199-02	VN085535.D	01/28/2025	13:49
BP-VPB-192-GW-280-282	Q1199-03	VN085536.D	01/28/2025	14:13
BP-VPB-192-GW-260-262	Q1199-04	VN085537.D	01/28/2025	14:37
BP-VPB-192-GW-240-242	Q1199-05	VN085538.D	01/28/2025	15:01
BP-VPB-192-DUP-20250123	Q1199-07	VN085540.D	01/28/2025	15:49
BP-VPB-192-GW-220-222	Q1199-06	VN085545.D	01/28/2025	17:50
VSTDCCC050EC	VSTDCCC050	VN085546.D	01/28/2025	18:14