

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME2964
 Instrument ID: P8 Start Date: 02/06/2025
 Analytical Method: ICP-MS Run Batch: LB134612
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.0	0	20	20	0	1000	960	-4
Antimony	0	0.0	0	2	2.2	10	50	50	0
Arsenic	0	0.0	0	1	1.1	10	50	49	-2
Barium	0	0.0	0	10	11	10	250	250	0
Beryllium	0	0.0	0	1	1.1	10	50	51	2
Cadmium	0	0.0	0	1	1.2	20	50	50	0
Calcium	0	0.0	0	500	510	2	5000	4800	-4
Chromium	0	0.0	0	2	2.1	5	50	49	-2
Cobalt	0	0.0	0	1	1.2	20	50	50	0
Copper	0	0.0	0	2	2.2	10	500	520	4
Iron	0	0.0	0	50	52	4	2500	2500	0
Lead	0	0.0	0	1	1.0	0	250	250	0
Magnesium	0	0.0	0	500	530	6	5000	5000	0
Manganese	0	0.0	0	1	1.1	10	500	490	-2
Nickel	0	0.0	0	1	1.1	10	50	53	6
Potassium	0	0.0	0	500	510	2	2500	2400	-4
Selenium	0	0.0	0	5	5.3	6	50	51	2
Silver	0	0.0	0	1	1.1	10	50	51	2
Sodium	0	0.0	0	500	520	4	5000	5000	0
Thallium	0	0.0	0	1	1.0	0	50	50	0
Vanadium	0	0.0	0	5	5.3	6	50	48	-4
Zinc	0	0.0	0	5	5.4	8	500	520	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME2964
 Instrument ID: P8 Start Date: 02/06/2025
 Analytical Method: ICP-MS Run Batch: LB134612
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	2500	2400	-4	5000	4800	-4	10000	9700	-3
Antimony	125	130	4	250	270	8	500	500	0
Arsenic	125	120	-4	250	250	0	500	510	2
Barium	625	630	1	1250	1300	4	2500	2500	0
Beryllium	125	130	4	250	260	4	500	500	0
Cadmium	125	120	-4	250	250	0	500	500	0
Calcium	12500	12000	-4	25000	24000	-4	50000	47000	-6
Chromium	125	120	-4	250	240	-4	500	520	4
Cobalt	125	120	-4	250	260	4	500	520	4
Copper	1250	1300	4	2500	2600	4	5000	5200	4
Iron	6250	6100	-2	12500	12000	-4	25000	26000	4
Lead	625	630	1	1250	1300	4	2500	2500	0
Magnesium	12500	13000	4	25000	25000	0	50000	51000	2
Manganese	1250	1300	4	2500	2500	0	5000	5100	2
Nickel	125	130	4	250	270	8	500	490	-2
Potassium	6250	6000	-4	12500	12000	-4	25000	24000	-4
Selenium	125	130	4	250	260	4	500	500	0
Silver	125	130	4	250	260	4	500	500	0
Sodium	12500	12000	-4	25000	25000	0	50000	50000	0
Thallium	125	120	-4	250	250	0	500	500	0
Vanadium	125	120	-4	250	240	-4	500	510	2
Zinc	1250	1300	4	2500	2700	8	5000	5200	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME2964
 Instrument ID: P8 Start Date: 02/06/2025
 Analytical Method: ICP-MS Run Batch: LB134612
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	20000	19000	-5	100000	100000	0			
Antimony	1000	1000	0						
Arsenic	1000	1000	0						
Barium	5000	5000	0						
Beryllium	1000	1000	0						
Cadmium	1000	1000	0						
Calcium	100000	96000	-4	500000	500000	0			
Chromium	1000	990	-1						
Cobalt	1000	990	-1						
Copper	10000	9900	-1						
Iron	50000	50000	0	250000	250000	0			
Lead	5000	5000	0						
Magnesium	100000	97000	-3	500000	500000	0			
Manganese	10000	9900	-1						
Nickel	1000	1000	0						
Potassium	50000	48000	-4	250000	250000	0			
Selenium	1000	1000	0						
Silver	1000	1000	0						
Sodium	100000	97000	-3	500000	500000	0			
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	9800	-2						