

ME2964A

POST-DIGESTION/DISTILLATION SPIKE SAMPLE RECOVERY

Lab Name:	<u>Alliance Technical Group, LLC</u>	Contract:	<u>68HERH20D0011</u>
Lab Code:	<u>ACE</u>	Case No.:	<u>51900</u>
		MA No. :	<u>SDG No.: ME2964</u>
Matrix :	<u>Water</u>	Analytical Method:	<u>ICP-MS</u>
% Solids:	<u></u>		
Concentration Units (µg/L, mg/L or mg/kg dry weight):		ug/L	

Analyte	Control Limit %R	Spiked Sample Result (SSR)		Sample Result (SR)		Spike Added (SA)	%R	Q
Antimony	75 - 125	4.3		2.0	U	4.0	108	
Selenium	75 - 125	12		1.7	J	10.0	103	