

FORM 9-IN
METHOD DETECTION LIMIT

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51900 MA No.: SDG No.: ME2964
Analytical Method: ICP-MS Instrument ID: P8
Preparation Method: 200.8
Concentration Units ($\mu\text{g/L}$, μg or mg/kg): $\mu\text{g/L}$

Analyte	Wavelength/Mass	MDL	Date Analyzed
Aluminum	27	7.0	12/11/2023
Antimony	121	0.097	12/11/2023
Arsenic	75	0.12	12/11/2023
Barium	137	0.34	12/11/2023
Beryllium	9	0.071	12/11/2023
Cadmium	111	0.26	12/11/2023
Calcium	44	60.0	12/11/2023
Chromium	52	0.099	12/11/2023
Cobalt	59	0.050	12/11/2023
Copper	63	0.42	12/11/2023
Iron	57	14.0	12/11/2023
Lead	208	0.21	12/11/2023
Magnesium	24	29.0	12/11/2023
Manganese	55	0.35	12/11/2023
Nickel	60	0.19	12/11/2023
Potassium	39	110	12/11/2023
Selenium	82	0.90	12/11/2023
Silver	107	0.11	12/11/2023
Sodium	23	240	12/11/2023
Thallium	205	0.15	12/11/2023
Vanadium	51	0.077	12/11/2023
Zinc	66	0.73	12/11/2023

FORM 9-IN
METHOD DETECTION LIMIT

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51900 MA No.: SDG No.: ME2964
Analytical Method: ICP-MS Instrument ID: P8
Preparation Method: 200.8
Concentration Units ($\mu\text{g/L}$, μg or mg/kg): $\mu\text{g/L}$

Analyte	Wavelength/Mass	MDL	Date Analyzed
---------	-----------------	-----	---------------