



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions SDG No.: Q1201
Contract: WEST04 Lab Code: CHEM Case No.: Q1201 SAS No.: Q1201
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	99	40 - 160	CV	02/06/2025	12:01	LB134599
CRI	Aluminum	19.0	20.0	95	70 - 130	P	02/06/2025	17:05	LB134616
	Antimony	2.07	2.0	104	70 - 130	P	02/06/2025	17:05	LB134616
	Arsenic	1.06	1.0	106	70 - 130	P	02/06/2025	17:05	LB134616
	Barium	10.1	10.0	101	70 - 130	P	02/06/2025	17:05	LB134616
	Beryllium	1.03	1.0	103	70 - 130	P	02/06/2025	17:05	LB134616
	Cadmium	1.10	1.0	110	70 - 130	P	02/06/2025	17:05	LB134616
	Calcium	475	500	95	70 - 130	P	02/06/2025	17:05	LB134616
	Chromium	1.90	2.0	95	70 - 130	P	02/06/2025	17:05	LB134616
	Cobalt	1.09	1.0	109	50 - 150	P	02/06/2025	17:05	LB134616
	Copper	2.14	2.0	107	70 - 130	P	02/06/2025	17:05	LB134616
	Iron	49.2	50.0	98	70 - 130	P	02/06/2025	17:05	LB134616
	Lead	1.00	1.0	100	70 - 130	P	02/06/2025	17:05	LB134616
	Magnesium	495	500	99	70 - 130	P	02/06/2025	17:05	LB134616
	Manganese	0.99	1.0	99	50 - 150	P	02/06/2025	17:05	LB134616
	Nickel	1.27	1.0	127	70 - 130	P	02/06/2025	17:05	LB134616
	Potassium	494	500	99	70 - 130	P	02/06/2025	17:05	LB134616
	Selenium	5.03	5.0	101	70 - 130	P	02/06/2025	17:05	LB134616
	Silver	1.00	1.0	100	70 - 130	P	02/06/2025	17:05	LB134616
	Sodium	521	500	104	70 - 130	P	02/06/2025	17:05	LB134616
	Thallium	0.97	1.0	97	70 - 130	P	02/06/2025	17:05	LB134616
	Vanadium	4.96	5.0	99	70 - 130	P	02/06/2025	17:05	LB134616
	Zinc	4.66	5.0	93	50 - 150	P	02/06/2025	17:05	LB134616