

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1201
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1201 **SAS No.:** Q1201
Matrix: Water **Sample ID:** Q1201-01 **Client ID:** TAPHHA-MW12-012725-00-T2DUP
Percent Solids for Sample: NA **Duplicate ID** Q1201-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1201
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1201 **SAS No.:** Q1201
Matrix: Water **Sample ID:** Q1201-01MS **Client ID:** TAPHHA-MW12-012725-00-T2MSD
Percent Solids for Sample: NA **Duplicate ID** Q1201-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.28		3.13		5		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1201
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1201 **SAS No.:** Q1201
Matrix: Water **Sample ID:** Q1211-02 **Client ID:** TAPIAL2-MW03-012825-00-T3DUP
Percent Solids for Sample: NA **Duplicate ID** Q1211-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	93.0		270		98	*	P
Antimony	ug/L	20	1.00	U	1.00	U			P
Arsenic	ug/L	20	2.26		2.40		6		P
Barium	ug/L	20	20.4		20.6		1		P
Beryllium	ug/L	20	0.080	J	0.50	U	200.0		P
Cadmium	ug/L	20	0.50	U	0.50	U			P
Calcium	ug/L	20	6660		6590		1		P
Chromium	ug/L	20	0.60	J	0.79	J	28		P
Cobalt	ug/L	20	3.61		3.70		2		P
Copper	ug/L	20	0.42	J	0.50	J	19		P
Iron	ug/L	20	18500		18900		2		P
Lead	ug/L	20	0.17	J	0.19	J	11		P
Magnesium	ug/L	20	1440		1470		2		P
Manganese	ug/L	20	541		545		1		P
Nickel	ug/L	20	1.00		1.15		14		P
Potassium	ug/L	20	1250		1290		3		P
Selenium	ug/L	20	2.50	U	2.50	U			P
Silver	ug/L	20	0.50	U	0.50	U			P
Sodium	ug/L	20	864		890		3		P
Thallium	ug/L	20	0.50	U	0.50	U			P
Vanadium	ug/L	20	1.50	J	1.75	J	15		P
Zinc	ug/L	20	2.02	J	2.14	J	6		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1201
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1201 **SAS No.:** Q1201
Matrix: Water **Sample ID:** Q1211-02MS **Client ID:** TAPIAL2-MW03-012825-00-T3MSD
Percent Solids for Sample: NA **Duplicate ID** Q1211-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	4120		4220		2		P
Antimony	ug/L	20	246		250		2		P
Arsenic	ug/L	20	2.14		2.24		5		P
Barium	ug/L	20	1110		1120		1		P
Beryllium	ug/L	20	232		234		1		P
Cadmium	ug/L	20	242		248		2		P
Calcium	ug/L	20	27900		28000		0		P
Chromium	ug/L	20	231		240		4		P
Cobalt	ug/L	20	241		246		2		P
Copper	ug/L	20	2310		2370		3		P
Iron	ug/L	20	27000		27100		0		P
Lead	ug/L	20	1180		1200		2		P
Magnesium	ug/L	20	23600		24100		2		P
Manganese	ug/L	20	2780		2850		2		P
Nickel	ug/L	20	223		229		3		P
Potassium	ug/L	20	11300		11500		2		P
Selenium	ug/L	20	234		241		3		P
Silver	ug/L	20	43.0		43.6		1		P
Sodium	ug/L	20	23300		23800		2		P
Thallium	ug/L	20	232		239		3		P
Vanadium	ug/L	20	240		245		2		P
Zinc	ug/L	20	2310		2350		2		P