

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2975
 Initial Calibration Verification Source : WP111792
 Continuing Calibration Verification Source : WP111793
 Run Batch: LB134600 Analytical Method: Spectrophotometry
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Cyanide	99.0	96	97		250	240	95		240	94	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2975
Initial Calibration Verification Source : WP111792
Continuing Calibration Verification Source : WP111793
Run Batch: LB134600 Analytical Method: Spectrophotometry
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Cyanide					250	240	95				