EPA SAMPLE NO.

ME2975S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech Cor	nsulting Group	Contract:	EPW14030	
Lab Code:	ACE	Case No.: 51900	MA No.:		SDG No.: ME2975
Matrix :	Water		Analytical	Method:	Spectrophotometry
% Solids:					
Concentrat	ion Units (µg	/L, mg/L or mg/kg dry v	veight): ug/]	L	

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Cyanide	75 - 125	94		10.0	U	100	94	