

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51900 MA No.: \_\_\_\_\_ SDG No.: ME2975  
Instrument ID: CV1 Start Date: 02/18/2025  
Analytical Method: CVAA Run Batch: LB134741  
Concentration Units: ug/L

| Analyte | True | Found | %D | True | Found | %D  | True | Found | %D |
|---------|------|-------|----|------|-------|-----|------|-------|----|
| Mercury | 0    | 0.0   | 0  | 0.2  | 0.18  | -10 | 2.5  | 2.5   | -1 |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51900 MA No.:                      SDG No.: ME2975  
Instrument ID: CV1 Start Date: 02/18/2025  
Analytical Method: CVAA Run Batch: LB134741  
Concentration Units: ug/L

| Analyte | True | Found | %D | True | Found | %D | True | Found | %D |
|---------|------|-------|----|------|-------|----|------|-------|----|
| Mercury | 5    | 5.0   | 1  | 7.5  | 7.7   | 3  | 10   | 9.8   | -2 |