

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2975
 Initial Calibration Verification Source : MP84554
 Continuing Calibration Verification Source : MP84556
 Run Batch: LB134741 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV008				ID: CCV081				ID: CCV082		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.7	91		5.0	5.1	101		5.1	101	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2975
 Initial Calibration Verification Source : MP84554
 Continuing Calibration Verification Source : MP84556
 Run Batch: LB134741 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV083				ID: CCV084		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	4.9	98		4.9	99	