



- 2a -

Client:	<u>RU2 Engineering, LLC</u>	SDG No.:	<u>Q1206</u>	
Contract:	<u>RUTW01</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1206</u> SAS No.: <u>Q1206</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	<u>PLASMA-PURE</u>			

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV75	Mercury	4.12	4.0	103	90 - 110	CV	01/30/2025	09:48	LB134479

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: RU2 Engineering, LLC **SDG No.:** Q1206
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1206 **SAS No.:** Q1206
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV87	Mercury	5.27	5.0	105	90 - 110	CV	01/30/2025	09:52	LB134479
CCV88	Mercury	5.35	5.0	107	90 - 110	CV	01/30/2025	10:27	LB134479
CCV89	Mercury	5.04	5.0	101	90 - 110	CV	01/30/2025	10:53	LB134479
CCV90	Mercury	5.07	5.0	101	90 - 110	CV	01/30/2025	11:15	LB134479



- 2a -

Client:	<u>RU2 Engineering, LLC</u>	SDG No.:	<u>Q1206</u>	
Contract:	<u>RUTW01</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1206</u> SAS No.: <u>Q1206</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
ICV01	Arsenic	1040	1000	104	90 - 110	P	01/30/2025	11:49	LB134493
	Barium	470	520	90	90 - 110	P	01/30/2025	11:49	LB134493
	Cadmium	502	510	98	90 - 110	P	01/30/2025	11:49	LB134493
	Chromium	522	520	100	90 - 110	P	01/30/2025	11:49	LB134493
	Lead	1000	1000	100	90 - 110	P	01/30/2025	11:49	LB134493
	Selenium	1040	1000	104	90 - 110	P	01/30/2025	11:49	LB134493
	Silver	252	250	101	90 - 110	P	01/30/2025	11:49	LB134493

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: RU2 Engineering, LLC **SDG No.:** Q1206
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1206 **SAS No.:** Q1206
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	21.3	20.0	107	80 - 120	P	01/30/2025	11:56	LB134493
	Barium	83.9	100	84	80 - 120	P	01/30/2025	11:56	LB134493
	Cadmium	5.60	6.0	93	80 - 120	P	01/30/2025	11:56	LB134493
	Chromium	9.80	10.0	98	80 - 120	P	01/30/2025	11:56	LB134493
	Lead	12.2	12.0	102	80 - 120	P	01/30/2025	11:56	LB134493
	Selenium	19.1	20.0	96	80 - 120	P	01/30/2025	11:56	LB134493
	Silver	10.4	10.0	104	80 - 120	P	01/30/2025	11:56	LB134493

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: RU2 Engineering, LLC **SDG No.:** Q1206

Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1206 **SAS No.:** Q1206

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4950	5000	99	90 - 110	P	01/30/2025	12:41	LB134493
	Barium	9130	10000	91	90 - 110	P	01/30/2025	12:41	LB134493
	Cadmium	2420	2500	97	90 - 110	P	01/30/2025	12:41	LB134493
	Chromium	988	1000	99	90 - 110	P	01/30/2025	12:41	LB134493
	Lead	4840	5000	97	90 - 110	P	01/30/2025	12:41	LB134493
	Selenium	5010	5000	100	90 - 110	P	01/30/2025	12:41	LB134493
	Silver	1210	1250	97	90 - 110	P	01/30/2025	12:41	LB134493
CCV02	Arsenic	5000	5000	100	90 - 110	P	01/30/2025	13:33	LB134493
	Barium	9200	10000	92	90 - 110	P	01/30/2025	13:33	LB134493
	Cadmium	2420	2500	97	90 - 110	P	01/30/2025	13:33	LB134493
	Chromium	988	1000	99	90 - 110	P	01/30/2025	13:33	LB134493
	Lead	4830	5000	97	90 - 110	P	01/30/2025	13:33	LB134493
	Selenium	5050	5000	101	90 - 110	P	01/30/2025	13:33	LB134493
	Silver	1200	1250	96	90 - 110	P	01/30/2025	13:33	LB134493
CCV03	Arsenic	4880	5000	98	90 - 110	P	01/30/2025	14:34	LB134493
	Barium	9430	10000	94	90 - 110	P	01/30/2025	14:34	LB134493
	Cadmium	2490	2500	100	90 - 110	P	01/30/2025	14:34	LB134493
	Chromium	985	1000	98	90 - 110	P	01/30/2025	14:34	LB134493
	Lead	4910	5000	98	90 - 110	P	01/30/2025	14:34	LB134493
	Selenium	5010	5000	100	90 - 110	P	01/30/2025	14:34	LB134493
	Silver	1210	1250	97	90 - 110	P	01/30/2025	14:34	LB134493
CCV04	Arsenic	4890	5000	98	90 - 110	P	01/30/2025	15:25	LB134493
	Barium	9350	10000	94	90 - 110	P	01/30/2025	15:25	LB134493
	Cadmium	2370	2500	95	90 - 110	P	01/30/2025	15:25	LB134493
	Chromium	1000	1000	100	90 - 110	P	01/30/2025	15:25	LB134493
	Lead	4750	5000	95	90 - 110	P	01/30/2025	15:25	LB134493
	Selenium	4950	5000	99	90 - 110	P	01/30/2025	15:25	LB134493
	Silver	1220	1250	98	90 - 110	P	01/30/2025	15:25	LB134493
CCV05	Arsenic	4830	5000	97	90 - 110	P	01/30/2025	16:24	LB134493
	Barium	9330	10000	93	90 - 110	P	01/30/2025	16:24	LB134493
	Cadmium	2370	2500	95	90 - 110	P	01/30/2025	16:24	LB134493
	Chromium	991	1000	99	90 - 110	P	01/30/2025	16:24	LB134493
	Lead	4740	5000	95	90 - 110	P	01/30/2025	16:24	LB134493
	Selenium	4860	5000	97	90 - 110	P	01/30/2025	16:24	LB134493

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: RU2 Engineering, LLC **SDG No.:** Q1206
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1206 **SAS No.:** Q1206
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1210	1250	97	90 - 110	P	01/30/2025	16:24	LB134493
CCV06	Arsenic	4970	5000	99	90 - 110	P	01/30/2025	17:19	LB134493
	Barium	9260	10000	93	90 - 110	P	01/30/2025	17:19	LB134493
	Cadmium	2440	2500	98	90 - 110	P	01/30/2025	17:19	LB134493
	Chromium	1010	1000	101	90 - 110	P	01/30/2025	17:19	LB134493
	Lead	4880	5000	98	90 - 110	P	01/30/2025	17:19	LB134493
	Selenium	4970	5000	100	90 - 110	P	01/30/2025	17:19	LB134493
	Silver	1230	1250	99	90 - 110	P	01/30/2025	17:19	LB134493
CCV07	Arsenic	4970	5000	99	90 - 110	P	01/30/2025	18:11	LB134493
	Barium	9620	10000	96	90 - 110	P	01/30/2025	18:11	LB134493
	Cadmium	2440	2500	98	90 - 110	P	01/30/2025	18:11	LB134493
	Chromium	1020	1000	102	90 - 110	P	01/30/2025	18:11	LB134493
	Lead	4880	5000	98	90 - 110	P	01/30/2025	18:11	LB134493
	Selenium	5000	5000	100	90 - 110	P	01/30/2025	18:11	LB134493
	Silver	1270	1250	101	90 - 110	P	01/30/2025	18:11	LB134493
CCV08	Arsenic	5000	5000	100	90 - 110	P	01/30/2025	18:45	LB134493
	Barium	9470	10000	95	90 - 110	P	01/30/2025	18:45	LB134493
	Cadmium	2430	2500	97	90 - 110	P	01/30/2025	18:45	LB134493
	Chromium	1010	1000	101	90 - 110	P	01/30/2025	18:45	LB134493
	Lead	4860	5000	97	90 - 110	P	01/30/2025	18:45	LB134493
	Selenium	5030	5000	101	90 - 110	P	01/30/2025	18:45	LB134493
	Silver	1250	1250	100	90 - 110	P	01/30/2025	18:45	LB134493