

SOIL DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RI

Lab Name: Chemtech **Client:** RU2 Engineering, LLC
Lab Code: CHEM **Cas No:** Q1206 **SAS No :** Q1206 **SDG No:** Q1206
Matrix Spike - EPA Sample No : PB166348BS **Datafile:** FE052143.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
DRO	6662	0	6174	93	68-131