

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB166333BS

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Lab Sample ID: PB166333BS Date(s) Analyzed: 01/29/2025 01/29/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO109238.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.793	4.743	4.843	173	169	3.61
COLUMN 1	2	4.813	4.763	4.863	171		
	3	4.869	4.819	4.919	169		
	4	4.99	4.94	5.04	170		
	5	5.247	5.197	5.297	164		
	1	4.78	4.73	4.83	181		
COLUMN 2	2	4.799	4.749	4.849	158	163	
	3	4.976	4.926	5.026	163		
	4	5.017	4.967	5.067	158		
	5	5.23	5.18	5.28	156		
Aroclor-1260	1	6.29	6.24	6.34	173	159	1.27
COLUMN 1	2	6.478	6.428	6.528	170		
	3	6.848	6.798	6.898	148		
	4	7.109	7.059	7.159	152		
	5	7.35	7.3	7.4	153		
	1	6.265	6.215	6.315	167	157	
COLUMN 2	2	6.452	6.402	6.502	164		
	3	6.605	6.555	6.655	166		
	4	7.077	7.027	7.127	142		
	5	7.318	7.268	7.368	145		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

VNJ-236MS

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1206

SAS No.: Q1206

SDG NO.: Q1206

Lab Sample ID: Q1205-01MS

Date(s) Analyzed: 01/29/2025

01/29/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO109245.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.793	4.743	4.843	189	182	3.92
COLUMN 1	2	4.812	4.762	4.862	184		
	3	4.868	4.818	4.918	185		
	4	4.99	4.94	5.04	181		
	5	5.247	5.197	5.297	173		
	1	4.78	4.73	4.83	168		
COLUMN 2	2	4.8	4.75	4.85	181	175	
	3	4.976	4.926	5.026	179		
	4	5.017	4.967	5.067	180		
	5	5.231	5.181	5.281	169		
Aroclor-1260	1	6.29	6.24	6.34	190	175	1.14
COLUMN 1	2	6.48	6.43	6.53	198		
	3	6.848	6.798	6.898	151		
	4	7.11	7.06	7.16	180		
	5	7.35	7.3	7.4	156		
	1	6.265	6.215	6.315	187	177	
COLUMN 2	2	6.452	6.402	6.502	197		
	3	6.606	6.556	6.656	174		
	4	7.078	7.028	7.128	177		
	5	7.318	7.268	7.368	153		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

VNJ-236MSD

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Lab Sample ID: Q1205-01MSD Date(s) Analyzed: 01/29/2025 01/29/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO109246.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.793	4.743	4.843	185	178	0.56
COLUMN 1	2	4.813	4.763	4.863	182		
	3	4.869	4.819	4.919	182		
	4	4.99	4.94	5.04	176		
	5	5.247	5.197	5.297	168		
	1	4.781	4.731	4.831	190		
COLUMN 2	2	4.8	4.75	4.85	176	177	
	3	4.976	4.926	5.026	176		
	4	5.018	4.968	5.068	178		
	5	5.231	5.181	5.281	167		
Aroclor-1260	1	6.29	6.24	6.34	187	172	1.16
COLUMN 1	2	6.479	6.429	6.529	199		
	3	6.847	6.797	6.897	146		
	4	7.109	7.059	7.159	176		
	5	7.349	7.299	7.399	154		
	1	6.265	6.215	6.315	185	174	
COLUMN 2	2	6.452	6.402	6.502	193		
	3	6.606	6.556	6.656	172		
	4	7.078	7.028	7.128	169		
	5	7.318	7.268	7.368	151		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

JPP-16.3-012725

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Lab Sample ID: Q1206-07 Date(s) Analyzed: 01/29/2025 01/29/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO109248.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	5.608	5.558	5.658	54.6	41.2	6.12
COLUMN 1	2	5.754	5.704	5.804	39.4		
	3	6.159	6.109	6.209	28.6		
	4	6.389	6.339	6.439	24.6		
	5	6.81	6.76	6.86	58.5		
	1	5.587	5.537	5.637	59.1		
COLUMN 2	2	5.731	5.681	5.781	42.1	43.8	
	3	6.134	6.084	6.184	32.1		
	4	6.361	6.311	6.411	24.2		
	5	6.779	6.729	6.829	61.2		
Aroclor-1260	1	6.291	6.241	6.341	44.1	27.5	8.7
COLUMN 1	2	6.48	6.43	6.53	32.1		
	3	6.849	6.799	6.899	22.2		
	4	7.11	7.06	7.16	17.2		
	5	7.349	7.299	7.399	22.0		
	1	6.265	6.215	6.315	45.7		
COLUMN 2	2	6.453	6.403	6.503	33.7	30.0	
	3	6.605	6.555	6.655	29.8		
	4	7.078	7.028	7.128	17.9		
	5	7.316	7.266	7.366	22.9		