

### CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/29/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 11:53 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.80       | 4.80      | 4.70      | 4.90 | 0.00       |
| Aroclor-1016-2       | (2) | 4.82       | 4.81      | 4.71      | 4.91 | -0.01      |
| Aroclor-1016-3       | (3) | 4.87       | 4.87      | 4.77      | 4.97 | 0.00       |
| Aroclor-1016-4       | (4) | 4.99       | 4.99      | 4.89      | 5.09 | 0.00       |
| Aroclor-1016-5       | (5) | 5.25       | 5.25      | 5.15      | 5.35 | 0.00       |
| Aroclor-1260-1       | (1) | 6.29       | 6.29      | 6.19      | 6.39 | 0.00       |
| Aroclor-1260-2       | (2) | 6.48       | 6.48      | 6.38      | 6.58 | 0.00       |
| Aroclor-1260-3       | (3) | 6.85       | 6.85      | 6.75      | 6.95 | 0.00       |
| Aroclor-1260-4       | (4) | 7.11       | 7.11      | 7.01      | 7.21 | 0.00       |
| Aroclor-1260-5       | (5) | 7.35       | 7.35      | 7.25      | 7.45 | 0.00       |
| Tetrachloro-m-xylene |     | 3.70       | 3.70      | 3.60      | 3.80 | 0.00       |
| Decachlorobiphenyl   |     | 8.76       | 8.76      | 8.66      | 8.86 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/29/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 11:53 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.78       | 4.80      | 4.70      | 4.90 | 0.02       |
| Aroclor-1016-2       | (2) | 4.80       | 4.82      | 4.72      | 4.92 | 0.02       |
| Aroclor-1016-3       | (3) | 4.98       | 5.00      | 4.90      | 5.10 | 0.02       |
| Aroclor-1016-4       | (4) | 5.02       | 5.04      | 4.94      | 5.14 | 0.02       |
| Aroclor-1016-5       | (5) | 5.23       | 5.25      | 5.15      | 5.35 | 0.02       |
| Aroclor-1260-1       | (1) | 6.27       | 6.29      | 6.19      | 6.39 | 0.02       |
| Aroclor-1260-2       | (2) | 6.45       | 6.47      | 6.37      | 6.57 | 0.02       |
| Aroclor-1260-3       | (3) | 6.61       | 6.63      | 6.53      | 6.73 | 0.02       |
| Aroclor-1260-4       | (4) | 7.08       | 7.10      | 7.00      | 7.20 | 0.02       |
| Aroclor-1260-5       | (5) | 7.32       | 7.34      | 7.24      | 7.44 | 0.02       |
| Tetrachloro-m-xylene |     | 3.70       | 3.72      | 3.62      | 3.82 | 0.02       |
| Decachlorobiphenyl   |     | 8.72       | 8.73      | 8.63      | 8.83 | 0.01       |