

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO109254.D Time Analyzed: 00:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.795	4.695	4.895	505.990	500.000	1.2
Aroclor-1016-2	4.815	4.714	4.914	509.560	500.000	1.9
Aroclor-1016-3	4.871	4.770	4.970	498.380	500.000	-0.3
Aroclor-1016-4	4.992	4.891	5.091	501.440	500.000	0.3
Aroclor-1016-5	5.249	5.149	5.349	514.980	500.000	3.0
Aroclor-1260-1	6.292	6.190	6.390	453.860	500.000	-9.2
Aroclor-1260-2	6.481	6.379	6.579	431.160	500.000	-13.8
Aroclor-1260-3	6.849	6.747	6.947	415.340	500.000	-16.9
Aroclor-1260-4	7.109	7.008	7.208	437.080	500.000	-12.6
Aroclor-1260-5	7.350	7.249	7.449	451.440	500.000	-9.7
Decachlorobiphenyl	8.762	8.658	8.858	38.710	50.000	-22.6
Tetrachloro-m-xylene	3.700	3.600	3.800	50.770	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO109254.D Time Analyzed: 00:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.780	4.703	4.903	533.530	500.000	6.7
Aroclor-1016-2	4.800	4.722	4.922	527.150	500.000	5.4
Aroclor-1016-3	4.975	4.898	5.098	529.220	500.000	5.8
Aroclor-1016-4	5.018	4.940	5.140	507.710	500.000	1.5
Aroclor-1016-5	5.231	5.153	5.353	546.030	500.000	9.2
Aroclor-1260-1	6.265	6.186	6.386	517.230	500.000	3.4
Aroclor-1260-2	6.452	6.373	6.573	499.300	500.000	-0.1
Aroclor-1260-3	6.605	6.527	6.727	481.950	500.000	-3.6
Aroclor-1260-4	7.077	6.999	7.199	450.300	500.000	-9.9
Aroclor-1260-5	7.318	7.239	7.439	446.240	500.000	-10.8
Decachlorobiphenyl	8.712	8.633	8.833	42.230	50.000	-15.5
Tetrachloro-m-xylene	3.696	3.618	3.818	54.730	50.000	9.5