

LAB CHRONICLE

OrderID:	Q1206	OrderDate:	1/28/2025 11:18:51 AM
Client:	RU2 Engineering, LLC	Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR
Contact:	Rutu Manani	Location:	E11,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1206-01	JPP-20.1-012725	SOIL			01/27/25			01/28/25
			Diesel Range Organics	8015D		01/29/25	01/30/25	
			Gasoline Range Organics	8015D			01/29/25	
Q1206-03	JPP-20.1-012725	SOIL			01/27/25			01/28/25
			PCB	8082A		01/29/25	01/29/25	
			Pesticide-TCL	8081B		01/29/25	01/30/25	
Q1206-04	JPP-20.1-012725	TCLP			01/27/25			01/28/25
			TCLP Herbicide	8151A		01/29/25	01/30/25	
			TCLP Pesticide	8081B		01/29/25	02/03/25	
Q1206-05	JPP-16.3-012725	SOIL			01/27/25			01/28/25
			Diesel Range Organics	8015D		01/29/25	01/30/25	
			Gasoline Range Organics	8015D			01/29/25	
Q1206-07	JPP-16.3-012725	SOIL			01/27/25			01/28/25
			PCB	8082A		01/29/25	01/29/25	
			Pesticide-TCL	8081B		01/29/25	02/04/25	
Q1206-08	JPP-16.3-012725	TCLP			01/27/25			01/28/25
			TCLP Herbicide	8151A		01/29/25	01/30/25	
			TCLP Pesticide	8081B		01/29/25	01/30/25	

Hit Summary Sheet
SW-846

SDG No.: Q1206

Order ID: Q1206

Client: RU2 Engineering, LLC

Project ID: NYCDDC SANTWOBR Brooklyn Bri

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : JPP-20.1-012725								
Q1206-03	JPP-20.1-012725	SOIL	Aldrin	0.54	JP	0.16	2.00	ug/kg
Q1206-03	JPP-20.1-012725	SOIL	4,4-DDE	3.30		0.15	2.00	ug/kg
Q1206-03	JPP-20.1-012725	SOIL	4,4-DDD	0.88	JP	0.22	2.00	ug/kg
Q1206-03	JPP-20.1-012725	SOIL	4,4-DDT	7.90		0.20	2.00	ug/kg
Total Concentration:				12.620				
Client ID : JPP-16.3-012725								
Q1206-07	JPP-16.3-012725	SOIL	Heptachlor	0.25	J	0.20	2.00	ug/kg
Q1206-07	JPP-16.3-012725	SOIL	Heptachlor epoxide	0.70	JP	0.27	2.00	ug/kg
Q1206-07	JPP-16.3-012725	SOIL	4,4-DDE	3.10		0.15	2.00	ug/kg
Q1206-07	JPP-16.3-012725	SOIL	Endrin	0.84	JP	0.19	2.00	ug/kg
Q1206-07	JPP-16.3-012725	SOIL	4,4-DDT	1.90	JP	0.20	2.00	ug/kg
Q1206-07	JPP-16.3-012725	SOIL	alpha-Chlordane	2.40	P	0.20	2.00	ug/kg
Q1206-07	JPP-16.3-012725	SOIL	gamma-Chlordane	1.40	JP	0.22	2.00	ug/kg
Total Concentration:				10.590				



SAMPLE DATA

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/27/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	JPP-20.1-012725	SDG No.:	Q1206
Lab Sample ID:	Q1206-03	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	85.5
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093897.D	1	01/29/25 08:55	01/30/25 16:24	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	2.00	U	0.21	2.00	ug/kg
319-85-7	beta-BHC	2.00	U	0.57	2.00	ug/kg
319-86-8	delta-BHC	2.00	U	0.55	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	2.00	U	0.22	2.00	ug/kg
76-44-8	Heptachlor	2.00	U	0.20	2.00	ug/kg
309-00-2	Aldrin	0.54	JP	0.16	2.00	ug/kg
1024-57-3	Heptachlor epoxide	2.00	U	0.27	2.00	ug/kg
959-98-8	Endosulfan I	2.00	U	0.20	2.00	ug/kg
60-57-1	Dieldrin	2.00	U	0.18	2.00	ug/kg
72-55-9	4,4-DDE	3.30		0.15	2.00	ug/kg
72-20-8	Endrin	2.00	U	0.19	2.00	ug/kg
33213-65-9	Endosulfan II	2.00	U	0.35	2.00	ug/kg
72-54-8	4,4-DDD	0.88	JP	0.22	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	2.00	U	0.15	2.00	ug/kg
50-29-3	4,4-DDT	7.90		0.20	2.00	ug/kg
72-43-5	Methoxychlor	2.00	U	0.44	2.00	ug/kg
53494-70-5	Endrin ketone	2.00	U	0.26	2.00	ug/kg
7421-93-4	Endrin aldehyde	2.00	U	0.46	2.00	ug/kg
5103-71-9	alpha-Chlordane	2.00	U	0.20	2.00	ug/kg
5103-74-2	gamma-Chlordane	2.00	U	0.22	2.00	ug/kg
8001-35-2	Toxaphene	38.5	U	6.10	38.5	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	13.9		10 - 148	69%	SPK: 20
877-09-8	Tetrachloro-m-xylene	13.0		10 - 159	65%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/27/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	JPP-20.1-012725	SDG No.:	Q1206
Lab Sample ID:	Q1206-03	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	85.5
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093897.D	1	01/29/25 08:55	01/30/25 16:24	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/27/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	JPP-16.3-012725	SDG No.:	Q1206
Lab Sample ID:	Q1206-07	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	84.5
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Decanted:	
		Final Vol:	10000
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094040.D	1	01/29/25 08:55	02/04/25 09:47	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	2.00	U	0.21	2.00	ug/kg
319-85-7	beta-BHC	2.00	U	0.58	2.00	ug/kg
319-86-8	delta-BHC	2.00	U	0.56	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	2.00	U	0.22	2.00	ug/kg
76-44-8	Heptachlor	0.25	J	0.20	2.00	ug/kg
309-00-2	Aldrin	2.00	U	0.17	2.00	ug/kg
1024-57-3	Heptachlor epoxide	0.70	JP	0.27	2.00	ug/kg
959-98-8	Endosulfan I	2.00	U	0.20	2.00	ug/kg
60-57-1	Dieldrin	2.00	U	0.18	2.00	ug/kg
72-55-9	4,4-DDE	3.10		0.15	2.00	ug/kg
72-20-8	Endrin	0.84	JP	0.19	2.00	ug/kg
33213-65-9	Endosulfan II	2.00	U	0.35	2.00	ug/kg
72-54-8	4,4-DDD	2.00	U	0.22	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	2.00	U	0.15	2.00	ug/kg
50-29-3	4,4-DDT	1.90	JP	0.20	2.00	ug/kg
72-43-5	Methoxychlor	2.00	U	0.45	2.00	ug/kg
53494-70-5	Endrin ketone	2.00	U	0.26	2.00	ug/kg
7421-93-4	Endrin aldehyde	2.00	U	0.46	2.00	ug/kg
5103-71-9	alpha-Chlordane	2.40	P	0.20	2.00	ug/kg
5103-74-2	gamma-Chlordane	1.40	JP	0.22	2.00	ug/kg
8001-35-2	Toxaphene	39.0	U	6.20	39.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	28.9		10 - 148	145%	SPK: 20
877-09-8	Tetrachloro-m-xylene	13.8		10 - 159	69%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/27/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	JPP-16.3-012725	SDG No.:	Q1206
Lab Sample ID:	Q1206-07	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	84.5
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094040.D	1	01/29/25 08:55	02/04/25 09:47	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093874.D	PIBLK-PL093874.D	Decachlorobiphenyl	1	20	22.5	113		43	140
		Tetrachloro-m-xylene	1	20	19.9	100		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	18.1	91		77	126
Q1209-05MS	WC-5MS	Decachlorobiphenyl	1	20	22.2	111		10	148
		Tetrachloro-m-xylene	1	20	20.2	101		10	159
		Decachlorobiphenyl	2	20	21.5	107		10	148
		Tetrachloro-m-xylene	2	20	19.4	97		10	159
Q1209-05MSD	WC-5MSD	Decachlorobiphenyl	1	20	22.2	111		10	148
		Tetrachloro-m-xylene	1	20	20.4	102		10	159
		Decachlorobiphenyl	2	20	21.6	108		10	148
		Tetrachloro-m-xylene	2	20	19.8	99		10	159
PB166334BL	PB166334BL	Decachlorobiphenyl	1	20	24.8	124		10	148
		Tetrachloro-m-xylene	1	20	22.0	110		10	159
		Decachlorobiphenyl	2	20	24.4	122		10	148
		Tetrachloro-m-xylene	2	20	20.6	103		10	159
PB166334BS	PB166334BS	Decachlorobiphenyl	1	20	19.6	98		10	148
		Tetrachloro-m-xylene	1	20	16.9	84		10	159
		Decachlorobiphenyl	2	20	19.0	95		10	148
		Tetrachloro-m-xylene	2	20	15.9	80		10	159
I.BLK-PL093887.D	PIBLK-PL093887.D	Decachlorobiphenyl	1	20	21.4	107		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	21.2	106		43	140
		Tetrachloro-m-xylene	2	20	17.5	88		77	126
Q1206-03	JPP-20.1-012725	Decachlorobiphenyl	1	20	13.9	69		10	148
		Tetrachloro-m-xylene	1	20	13.0	65		10	159
		Decachlorobiphenyl	2	20	10.9	54		10	148
		Tetrachloro-m-xylene	2	20	12.7	64		10	159
I.BLK-PL093904.D	PIBLK-PL093904.D	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	19.1	96		77	126
		Decachlorobiphenyl	2	20	16.0	80		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
I.BLK-PL094036.D	PIBLK-PL094036.D	Decachlorobiphenyl	1	20	23.0	115		43	140
		Tetrachloro-m-xylene	1	20	22.4	112		77	126
		Decachlorobiphenyl	2	20	23.3	116		43	140
		Tetrachloro-m-xylene	2	20	21.9	110		77	126
Q1206-07	JPP-16.3-012725	Decachlorobiphenyl	1	20	28.9	145		10	148

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1206-07	JPP-16.3-012725	Tetrachloro-m-xylene	1	20	13.1	66		10	159
		Decachlorobiphenyl	2	20	7.45	37		10	148
		Tetrachloro-m-xylene	2	20	13.8	69		10	159
I.BLK-PL094051.D	PIBLK-PL094051.D	Decachlorobiphenyl	1	20	20.1	101		43	140
		Tetrachloro-m-xylene	1	20	21.9	109		77	126
		Decachlorobiphenyl	2	20	19.6	98		43	140
		Tetrachloro-m-xylene	2	20	21.4	107		77	126

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL093882.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q1209-05MS	WC-5MS											
	alpha-BHC	18.83	0	17.1	ug/kg	91				60	144	
	beta-BHC	18.83	0	17.9	ug/kg	95				54	143	
	delta-BHC	18.83	0	17.3	ug/kg	92				47	144	
	gamma-BHC (Lindane)	18.83	0	17.1	ug/kg	91				61	140	
	Heptachlor	18.83	0	18.1	ug/kg	96				63	135	
	Aldrin	18.83	0	17.1	ug/kg	91				49	139	
	Heptachlor epoxide	18.83	0	17.8	ug/kg	95				32	180	
	Endosulfan I	18.83	0	18.3	ug/kg	97				56	142	
	Dieldrin	18.83	0	18.1	ug/kg	96				47	161	
	4,4'-DDE	18.83	0	18.9	ug/kg	100				55	136	
	Endrin	18.83	0	18.8	ug/kg	100				57	139	
	Endosulfan II	18.83	0	18.7	ug/kg	99				40	163	
	4,4'-DDD	18.83	0	19.8	ug/kg	105				37	192	
	Endosulfan sulfate	18.83	0	18.7	ug/kg	99				62	139	
	4,4'-DDT	18.83	0	19.8	ug/kg	105				51	146	
	Methoxychlor	18.83	0	20.1	ug/kg	107				54	136	
	Endrin ketone	18.83	0	18.8	ug/kg	100				60	129	
	Endrin aldehyde	18.83	0	17.7	ug/kg	94				59	132	
	alpha-Chlordane	18.83	0	18.4	ug/kg	98				30	192	
	gamma-Chlordane	18.83	0	18.4	ug/kg	98				44	175	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL093883.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q1209-05MSD	WC-5MSD											
	alpha-BHC	18.85	0	17.4	ug/kg	92		1		60	144	20
	beta-BHC	18.85	0	18.0	ug/kg	95		0		54	143	20
	delta-BHC	18.85	0	17.5	ug/kg	93		1		47	144	20
	gamma-BHC (Lindane)	18.85	0	17.3	ug/kg	92		1		61	140	20
	Heptachlor	18.85	0	18.1	ug/kg	96		0		63	135	20
	Aldrin	18.85	0	17.4	ug/kg	92		1		49	139	20
	Heptachlor epoxide	18.85	0	18.0	ug/kg	95		0		32	180	20
	Endosulfan I	18.85	0	18.5	ug/kg	98		1		56	142	20
	Dieldrin	18.85	0	18.3	ug/kg	97		1		47	161	20
	4,4'-DDE	18.85	0	19.2	ug/kg	102		2		55	136	20
	Endrin	18.85	0	19.2	ug/kg	102		2		57	139	20
	Endosulfan II	18.85	0	18.9	ug/kg	100		1		40	163	20
	4,4'-DDD	18.85	0	20.1	ug/kg	107		2		37	192	20
	Endosulfan sulfate	18.85	0	18.9	ug/kg	100		1		62	139	20
	4,4'-DDT	18.85	0	19.9	ug/kg	106		1		51	146	20
	Methoxychlor	18.85	0	20.2	ug/kg	107		0		54	136	20
	Endrin ketone	18.85	0	19.0	ug/kg	101		1		60	129	20
	Endrin aldehyde	18.85	0	17.9	ug/kg	95		1		59	132	20
	alpha-Chlordane	18.85	0	18.6	ug/kg	99		1		30	192	20
	gamma-Chlordane	18.85	0	18.5	ug/kg	98		0		44	175	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: **8081B** Datafile : PL093885.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166334BS	alpha-BHC	16.65	14.2	ug/kg	85				84	123	
	beta-BHC	16.65	14.6	ug/kg	88				82	123	
	delta-BHC	16.65	14.3	ug/kg	86				83	126	
	gamma-BHC (Lindane)	16.65	14.1	ug/kg	85				83	125	
	Heptachlor	16.65	15.2	ug/kg	91				83	122	
	Aldrin	16.65	14.3	ug/kg	86				82	124	
	Heptachlor epoxide	16.65	14.7	ug/kg	88				83	120	
	Endosulfan I	16.65	15.1	ug/kg	91				81	124	
	Dieldrin	16.65	15.0	ug/kg	90				85	121	
	4,4'-DDE	16.65	15.7	ug/kg	94				81	123	
	Endrin	16.65	15.1	ug/kg	91				76	130	
	Endosulfan II	16.65	15.6	ug/kg	94				80	125	
	4,4'-DDD	16.65	16.5	ug/kg	99				80	131	
	Endosulfan sulfate	16.65	15.6	ug/kg	94				81	122	
	4,4'-DDT	16.65	16.3	ug/kg	98				70	129	
	Methoxychlor	16.65	16.2	ug/kg	97				60	119	
	Endrin ketone	16.65	15.8	ug/kg	95				77	132	
	Endrin aldehyde	16.65	15.1	ug/kg	91				79	124	
	alpha-Chlordane	16.65	15.1	ug/kg	91				84	120	
	gamma-Chlordane	16.65	15.2	ug/kg	91				83	122	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166334BL

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206

SAS No.: Q1206 SDG NO.: Q1206

Lab Sample ID: PB166334BL

Lab File ID: PL093884.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 01/29/2025

Date Analyzed (1): 01/30/2025

Date Analyzed (2): 01/30/2025

Time Analyzed (1): 11:51

Time Analyzed (2): 11:51

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
WC-5MS	Q1209-05MS	PL093882.D	01/30/2025	01/30/2025
WC-5MSD	Q1209-05MSD	PL093883.D	01/30/2025	01/30/2025
PB166334BS	PB166334BS	PL093885.D	01/30/2025	01/30/2025
JPP-20.1-012725	Q1206-03	PL093897.D	01/30/2025	01/30/2025
JPP-16.3-012725	Q1206-07	PL094040.D	02/04/2025	02/04/2025

COMMENTS: _____



QC SAMPLE DATA

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166334BL	SDG No.:	Q1206
Lab Sample ID:	PB166334BL	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	100
Sample Wt/Vol:	30.02 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
PH :			
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093884.D	1	01/29/25 08:55	01/30/25 11:51	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	1.70	U	0.18	1.70	ug/kg
319-85-7	beta-BHC	1.70	U	0.49	1.70	ug/kg
319-86-8	delta-BHC	1.70	U	0.47	1.70	ug/kg
58-89-9	gamma-BHC (Lindane)	1.70	U	0.19	1.70	ug/kg
76-44-8	Heptachlor	1.70	U	0.17	1.70	ug/kg
309-00-2	Aldrin	1.70	U	0.14	1.70	ug/kg
1024-57-3	Heptachlor epoxide	1.70	U	0.23	1.70	ug/kg
959-98-8	Endosulfan I	1.70	U	0.17	1.70	ug/kg
60-57-1	Dieldrin	1.70	U	0.15	1.70	ug/kg
72-55-9	4,4-DDE	1.70	U	0.13	1.70	ug/kg
72-20-8	Endrin	1.70	U	0.16	1.70	ug/kg
33213-65-9	Endosulfan II	1.70	U	0.30	1.70	ug/kg
72-54-8	4,4-DDD	1.70	U	0.19	1.70	ug/kg
1031-07-8	Endosulfan Sulfate	1.70	U	0.13	1.70	ug/kg
50-29-3	4,4-DDT	1.70	U	0.17	1.70	ug/kg
72-43-5	Methoxychlor	1.70	U	0.38	1.70	ug/kg
53494-70-5	Endrin ketone	1.70	U	0.22	1.70	ug/kg
7421-93-4	Endrin aldehyde	1.70	U	0.39	1.70	ug/kg
5103-71-9	alpha-Chlordane	1.70	U	0.17	1.70	ug/kg
5103-74-2	gamma-Chlordane	1.70	U	0.19	1.70	ug/kg
8001-35-2	Toxaphene	33.0	U	5.20	33.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	24.8		10 - 148	124%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.0		10 - 159	110%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166334BL	SDG No.:	Q1206
Lab Sample ID:	PB166334BL	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	100
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
PH :		Test:	Pesticide-TCL
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093884.D	1	01/29/25 08:55	01/30/25 11:51	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/21/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/21/25
Client Sample ID:	PIBLK-PL093725.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL093725.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093725.D	1		01/21/25	PL012125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.1		43 - 140	111%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.8		77 - 126	104%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/21/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/21/25
Client Sample ID:	PIBLK-PL093725.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL093725.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093725.D	1		01/21/25	PL012125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	PIBLK-PL093874.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL093874.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093874.D	1		01/30/25	pl013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.5		43 - 140	113%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.9		77 - 126	100%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	PIBLK-PL093874.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL093874.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093874.D	1		01/30/25	pl013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	PIBLK-PL093887.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL093887.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093887.D	1		01/30/25	pl013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.4		43 - 140	107%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.4		77 - 126	92%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	PIBLK-PL093887.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL093887.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093887.D	1		01/30/25	pl013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	PIBLK-PL093904.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL093904.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093904.D	1		01/30/25	pl013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.2		43 - 140	101%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.1		77 - 126	96%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	PIBLK-PL093904.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL093904.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093904.D	1		01/30/25	pl013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	02/04/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	02/04/25
Client Sample ID:	PIBLK-PL094036.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL094036.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094036.D	1		02/04/25	pl020425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	23.3		43 - 140	116%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.4		77 - 126	112%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	02/04/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	02/04/25
Client Sample ID:	PIBLK-PL094036.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL094036.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094036.D	1		02/04/25	pl020425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	02/04/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	02/04/25
Client Sample ID:	PIBLK-PL094051.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL094051.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094051.D	1		02/04/25	pl020425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.1		43 - 140	101%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.9		77 - 126	109%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	02/04/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	02/04/25
Client Sample ID:	PIBLK-PL094051.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PL094051.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094051.D	1		02/04/25	pl020425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166334BS	SDG No.:	Q1206
Lab Sample ID:	PB166334BS	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
PH :			
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093885.D	1	01/29/25 08:55	01/30/25 12:05	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	14.2		0.18	1.70	ug/kg
319-85-7	beta-BHC	14.6		0.49	1.70	ug/kg
319-86-8	delta-BHC	14.3		0.47	1.70	ug/kg
58-89-9	gamma-BHC (Lindane)	14.1		0.19	1.70	ug/kg
76-44-8	Heptachlor	15.2		0.17	1.70	ug/kg
309-00-2	Aldrin	14.3		0.14	1.70	ug/kg
1024-57-3	Heptachlor epoxide	14.7		0.23	1.70	ug/kg
959-98-8	Endosulfan I	15.1		0.17	1.70	ug/kg
60-57-1	Dieldrin	15.0		0.15	1.70	ug/kg
72-55-9	4,4-DDE	15.7		0.13	1.70	ug/kg
72-20-8	Endrin	15.1		0.16	1.70	ug/kg
33213-65-9	Endosulfan II	15.6		0.30	1.70	ug/kg
72-54-8	4,4-DDD	16.5		0.19	1.70	ug/kg
1031-07-8	Endosulfan Sulfate	15.6		0.13	1.70	ug/kg
50-29-3	4,4-DDT	16.3		0.17	1.70	ug/kg
72-43-5	Methoxychlor	16.2		0.38	1.70	ug/kg
53494-70-5	Endrin ketone	15.8		0.22	1.70	ug/kg
7421-93-4	Endrin aldehyde	15.1		0.39	1.70	ug/kg
5103-71-9	alpha-Chlordane	15.1		0.17	1.70	ug/kg
5103-74-2	gamma-Chlordane	15.2		0.19	1.70	ug/kg
8001-35-2	Toxaphene	33.0	U	5.20	33.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.6		10 - 148	98%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.9		10 - 159	84%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166334BS	SDG No.:	Q1206
Lab Sample ID:	PB166334BS	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	100 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:		Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093885.D	1	01/29/25 08:55	01/30/25 12:05	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/28/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	WC-5MS	SDG No.:	Q1206
Lab Sample ID:	Q1209-05MS	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	88.4
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093882.D	1	01/29/25 08:55	01/30/25 11:25	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	17.1		0.20	1.90	ug/kg
319-85-7	beta-BHC	17.9		0.55	1.90	ug/kg
319-86-8	delta-BHC	17.3		0.53	1.90	ug/kg
58-89-9	gamma-BHC (Lindane)	17.1		0.21	1.90	ug/kg
76-44-8	Heptachlor	18.1		0.19	1.90	ug/kg
309-00-2	Aldrin	17.1		0.16	1.90	ug/kg
1024-57-3	Heptachlor epoxide	17.8		0.26	1.90	ug/kg
959-98-8	Endosulfan I	18.3		0.19	1.90	ug/kg
60-57-1	Dieldrin	18.1		0.17	1.90	ug/kg
72-55-9	4,4-DDE	18.9		0.15	1.90	ug/kg
72-20-8	Endrin	18.8		0.18	1.90	ug/kg
33213-65-9	Endosulfan II	18.7		0.34	1.90	ug/kg
72-54-8	4,4-DDD	19.8		0.21	1.90	ug/kg
1031-07-8	Endosulfan Sulfate	18.7		0.15	1.90	ug/kg
50-29-3	4,4-DDT	19.8		0.19	1.90	ug/kg
72-43-5	Methoxychlor	20.1		0.43	1.90	ug/kg
53494-70-5	Endrin ketone	18.8		0.25	1.90	ug/kg
7421-93-4	Endrin aldehyde	17.7		0.44	1.90	ug/kg
5103-71-9	alpha-Chlordane	18.4		0.19	1.90	ug/kg
5103-74-2	gamma-Chlordane	18.4		0.21	1.90	ug/kg
8001-35-2	Toxaphene	37.3	U	5.90	37.3	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.2		10 - 148	111%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.2		10 - 159	101%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/28/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	WC-5MS	SDG No.:	Q1206
Lab Sample ID:	Q1209-05MS	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	88.4
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093882.D	1	01/29/25 08:55	01/30/25 11:25	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/28/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	WC-5MSD	SDG No.:	Q1206
Lab Sample ID:	Q1209-05MSD	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	88.4
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093883.D	1	01/29/25 08:55	01/30/25 11:38	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	17.4		0.20	1.90	ug/kg
319-85-7	beta-BHC	18.0		0.55	1.90	ug/kg
319-86-8	delta-BHC	17.5		0.53	1.90	ug/kg
58-89-9	gamma-BHC (Lindane)	17.3		0.21	1.90	ug/kg
76-44-8	Heptachlor	18.1		0.19	1.90	ug/kg
309-00-2	Aldrin	17.4		0.16	1.90	ug/kg
1024-57-3	Heptachlor epoxide	18.0		0.26	1.90	ug/kg
959-98-8	Endosulfan I	18.5		0.19	1.90	ug/kg
60-57-1	Dieldrin	18.3		0.17	1.90	ug/kg
72-55-9	4,4-DDE	19.2		0.15	1.90	ug/kg
72-20-8	Endrin	19.2		0.18	1.90	ug/kg
33213-65-9	Endosulfan II	18.9		0.34	1.90	ug/kg
72-54-8	4,4-DDD	20.1		0.21	1.90	ug/kg
1031-07-8	Endosulfan Sulfate	18.9		0.15	1.90	ug/kg
50-29-3	4,4-DDT	19.9		0.19	1.90	ug/kg
72-43-5	Methoxychlor	20.2		0.43	1.90	ug/kg
53494-70-5	Endrin ketone	19.0		0.25	1.90	ug/kg
7421-93-4	Endrin aldehyde	17.9		0.44	1.90	ug/kg
5103-71-9	alpha-Chlordane	18.6		0.19	1.90	ug/kg
5103-74-2	gamma-Chlordane	18.5		0.21	1.90	ug/kg
8001-35-2	Toxaphene	37.3	U	5.90	37.3	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.2		10 - 148	111%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.4		10 - 159	102%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/28/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	WC-5MSD	SDG No.:	Q1206
Lab Sample ID:	Q1209-05MSD	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	88.4
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093883.D	1	01/29/25 08:55	01/30/25 11:38	PB166334

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



CALIBRATION SUMMARY

A
B
C
D
E
F
G
H

Calibration Times: **10:57** **11:51**

LAB FILE ID: RT 100 = PL093728.D RT 075 = PL093729.D
RT 050 = PL093730.D RT 025 = PL093731.D RT 005 = PL093732.D

[illegible]

A
B
C
D
E
F
G
H

Calibration Times: **10:57** **11:51**

LAB FILE ID: RT 100 = PL093728.D RT 075 = PL093729.D
RT 050 = PL093730.D RT 025 = PL093731.D RT 005 = PL093732.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Instrument ID: ECD_L

Calibration Date(s): 01/21/2025 01/21/2025

Calibration Times: 10:57 11:51

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 =	PL093728.D	CF 075 =	PL093729.D		
CF 050 =		PL093730.D	CF 025 =	PL093731.D	CF 005 =	PL093732.D	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	1660930000	1679650000	1932670000	1802720000	2426830000	1900560000	17
4,4'-DDE	2179870000	2169930000	2489080000	2321590000	3012520000	2434600000	14
4,4'-DDT	1755570000	1766710000	2016720000	1907120000	2414170000	1972060000	14
Aldrin	2924220000	2896750000	3292630000	3099660000	4146570000	3271970000	16
alpha-BHC	3537700000	3490280000	3918110000	3562830000	4660310000	3833850000	13
alpha-Chlordane	2458070000	2458490000	2788200000	2666580000	3570690000	2788400000	16
beta-BHC	1393460000	1394440000	1618290000	1508890000	2121530000	1607320000	19
Decachlorobiphenyl	1768480000	1816480000	2098320000	2018470000	2757820000	2091910000	19
delta-BHC	3233860000	3194550000	3605880000	3303370000	4188780000	3505290000	12
Dieldrin	2456580000	2440810000	2788190000	2639340000	3554340000	2775850000	17
Endosulfan I	2304400000	2298550000	2637060000	2528610000	3445650000	2642860000	18
Endosulfan II	2084130000	2100600000	2413950000	2287820000	3160260000	2409350000	18
Endosulfan sulfate	1923100000	1945070000	2248580000	2190510000	3011450000	2263740000	20
Endrin	2079430000	2060990000	2363220000	2218560000	3001890000	2344820000	17
Endrin aldehyde	1673120000	1696040000	1958970000	1896570000	2495580000	1944060000	17
Endrin ketone	2196850000	2205550000	2539700000	2413910000	3257130000	2522630000	17
gamma-BHC (Lindane)	3375960000	3339350000	3767250000	3460830000	4470850000	3682850000	13
gamma-Chlordane	2455830000	2471830000	2815630000	2678390000	3515170000	2787370000	16
Heptachlor	2922500000	2901690000	3325290000	3144100000	4093120000	3277340000	15
Heptachlor epoxide	2568680000	2575960000	2953630000	2835830000	3935020000	2973820000	19
Methoxychlor	907284000	922109000	1080370000	1020090000	1287130000	1043400000	15
Tetrachloro-m-xylene	2397870000	2402980000	2740040000	2595500000	3327420000	2692760000	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Instrument ID: ECD_L **Calibration Date(s):** 01/21/2025 01/21/2025
Calibration Times: 10:57 11:51

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL093728.D</u>		CF 075 = <u>PL093729.D</u>			
CF 050 = <u>PL093730.D</u>		CF 025 = <u>PL093731.D</u>		CF 005 = <u>PL093732.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3134000000	3054730000	3379160000	2910470000	3304320000	3156540000	6
4,4'-DDE	3891920000	3807640000	4253650000	3749010000	4345130000	4009470000	7
4,4'-DDT	3270010000	3177800000	3542860000	3046890000	3232670000	3254050000	6
Aldrin	4482990000	4370810000	4856520000	4222470000	4876190000	4561800000	6
alpha-BHC	4914190000	4768640000	5271080000	4480730000	5010260000	4888980000	6
alpha-Chlordane	4056970000	3962110000	4424110000	3914810000	4574820000	4186560000	7
beta-BHC	1863440000	1842720000	2072180000	1889740000	2319100000	1997440000	10
Decachlorobiphenyl	3226690000	3193800000	3627020000	3320620000	4152210000	3504070000	11
delta-BHC	4741230000	4607910000	5098810000	4368820000	4939430000	4751240000	6
Dieldrin	4189300000	4076770000	4553570000	3958830000	4699760000	4295650000	7
Endosulfan I	3734100000	3661580000	4099030000	3635320000	4254550000	3876920000	7
Endosulfan II	3553260000	3487640000	3912960000	3484510000	4080760000	3703830000	7
Endosulfan sulfate	3408630000	3353240000	3757030000	3348270000	3963240000	3566080000	8
Endrin	3607760000	3481170000	3870730000	3406140000	4097610000	3692680000	8
Endrin aldehyde	2861460000	2820180000	3183430000	2892290000	3465840000	3044640000	9
Endrin ketone	3965120000	3881890000	4400080000	3907370000	4821740000	4195240000	10
gamma-BHC (Lindane)	4713370000	4597010000	5084610000	4384810000	4926270000	4741210000	6
gamma-Chlordane	4137240000	4016860000	4483010000	3935490000	4615500000	4237620000	7
Heptachlor	4505180000	4413750000	4924840000	4345980000	5084220000	4654790000	7
Heptachlor epoxide	4026840000	3946880000	4424170000	3927960000	4575440000	4180260000	7
Methoxychlor	1651870000	1634200000	1870410000	1643810000	2140390000	1788140000	12
Tetrachloro-m-xylene	3101220000	3058550000	3437230000	3066200000	3657590000	3264160000	8

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Instrument ID: ECD_L Date(s) Analyzed: 01/21/2025 01/21/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	23446000
		2	6.44	6.34	6.54	14767200
		3	7.06	6.96	7.16	75896000
		4	7.15	7.05	7.25	57345100
		5	7.93	7.83	8.03	43067100

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Instrument ID: ECD_L Date(s) Analyzed: 01/21/2025 01/21/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.00	4.90	5.10	27057100
		2	5.33	5.23	5.43	23947200
		3	5.68	5.58	5.78	24726400
		4	6.60	6.50	6.70	84987200
		5	7.04	6.94	7.14	80238300

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 09:58 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.01
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.91	4.81	5.01	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.64	7.54	7.74	0.00
Endrin aldehyde	6.92	6.92	6.82	7.02	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 09:58 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.79	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.03	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093876.D Time Analyzed: 09:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.710	6.608	6.808	50.170	50.000	0.3
4,4'-DDE	6.192	6.091	6.291	49.240	50.000	-1.5
4,4'-DDT	7.023	6.922	7.122	50.330	50.000	0.7
Aldrin	5.257	5.156	5.356	45.120	50.000	-9.8
alpha-BHC	3.995	3.895	4.095	45.760	50.000	-8.5
alpha-Chlordane	6.019	5.917	6.117	47.200	50.000	-5.6
beta-BHC	4.526	4.425	4.625	47.030	50.000	-5.9
Decachlorobiphenyl	9.055	8.953	9.153	48.590	50.000	-2.8
delta-BHC	4.773	4.672	4.872	45.450	50.000	-9.1
Dieldrin	6.344	6.243	6.443	46.380	50.000	-7.2
Endosulfan I	6.069	5.967	6.167	46.440	50.000	-7.1
Endosulfan II	6.794	6.692	6.892	46.940	50.000	-6.1
Endosulfan sulfate	7.158	7.057	7.257	46.630	50.000	-6.7
Endrin	6.574	6.472	6.672	48.110	50.000	-3.8
Endrin aldehyde	6.924	6.823	7.023	45.880	50.000	-8.2
Endrin ketone	7.643	7.542	7.742	47.110	50.000	-5.8
gamma-BHC (Lindane)	4.327	4.227	4.427	45.140	50.000	-9.7
gamma-Chlordane	5.939	5.838	6.038	47.300	50.000	-5.4
Heptachlor	4.915	4.814	5.014	47.560	50.000	-4.9
Heptachlor epoxide	5.683	5.582	5.782	45.130	50.000	-9.7
Methoxychlor	7.500	7.398	7.598	51.550	50.000	3.1
Tetrachloro-m-xylene	3.539	3.439	3.639	45.390	50.000	-9.2

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093876.D Time Analyzed: 09:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.785	5.684	5.884	51.350	50.000	2.7
4,4'-DDE	5.230	5.130	5.330	49.780	50.000	-0.4
4,4'-DDT	6.034	5.934	6.134	52.120	50.000	4.2
Aldrin	4.224	4.125	4.325	45.460	50.000	-9.1
alpha-BHC	3.277	3.177	3.377	46.900	50.000	-6.2
alpha-Chlordane	5.040	4.940	5.140	48.570	50.000	-2.9
beta-BHC	3.907	3.807	4.007	47.660	50.000	-4.7
Decachlorobiphenyl	7.910	7.810	8.010	49.800	50.000	-0.4
delta-BHC	4.135	4.036	4.236	45.620	50.000	-8.8
Dieldrin	5.361	5.261	5.461	47.730	50.000	-4.5
Endosulfan I	5.097	4.996	5.196	46.270	50.000	-7.5
Endosulfan II	5.932	5.831	6.031	49.200	50.000	-1.6
Endosulfan sulfate	6.334	6.233	6.433	48.600	50.000	-2.8
Endrin	5.636	5.536	5.736	48.870	50.000	-2.3
Endrin aldehyde	6.111	6.010	6.210	45.870	50.000	-8.3
Endrin ketone	6.838	6.739	6.939	48.600	50.000	-2.8
gamma-BHC (Lindane)	3.607	3.507	3.707	46.010	50.000	-8.0
gamma-Chlordane	4.977	4.877	5.077	48.760	50.000	-2.5
Heptachlor	3.945	3.845	4.045	47.440	50.000	-5.1
Heptachlor epoxide	4.727	4.627	4.827	46.460	50.000	-7.1
Methoxychlor	6.610	6.509	6.709	49.830	50.000	-0.3
Tetrachloro-m-xylene	2.774	2.674	2.874	45.470	50.000	-9.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 13:53 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.01
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.91	4.81	5.01	-0.01
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.35	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.65	7.64	7.54	7.74	0.00
Endrin aldehyde	6.93	6.92	6.82	7.02	-0.01
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 13:53 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.79	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.04	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093888.D Time Analyzed: 13:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.710	6.608	6.808	51.000	50.000	2.0
4,4'-DDE	6.193	6.091	6.291	48.290	50.000	-3.4
4,4'-DDT	7.024	6.922	7.122	47.220	50.000	-5.6
Aldrin	5.258	5.156	5.356	44.250	50.000	-11.5
alpha-BHC	3.995	3.895	4.095	45.350	50.000	-9.3
alpha-Chlordane	6.019	5.917	6.117	46.270	50.000	-7.5
beta-BHC	4.526	4.425	4.625	46.190	50.000	-7.6
Decachlorobiphenyl	9.056	8.953	9.153	48.490	50.000	-3.0
delta-BHC	4.773	4.672	4.872	44.920	50.000	-10.2
Dieldrin	6.345	6.243	6.443	45.520	50.000	-9.0
Endosulfan I	6.069	5.967	6.167	45.280	50.000	-9.4
Endosulfan II	6.794	6.692	6.892	46.220	50.000	-7.6
Endosulfan sulfate	7.159	7.057	7.257	45.860	50.000	-8.3
Endrin	6.573	6.472	6.672	43.820	50.000	-12.4
Endrin aldehyde	6.925	6.823	7.023	45.230	50.000	-9.5
Endrin ketone	7.645	7.542	7.742	46.160	50.000	-7.7
gamma-BHC (Lindane)	4.328	4.227	4.427	44.710	50.000	-10.6
gamma-Chlordane	5.940	5.838	6.038	46.270	50.000	-7.5
Heptachlor	4.916	4.814	5.014	46.250	50.000	-7.5
Heptachlor epoxide	5.684	5.582	5.782	44.430	50.000	-11.1
Methoxychlor	7.500	7.398	7.598	46.980	50.000	-6.0
Tetrachloro-m-xylene	3.539	3.439	3.639	45.400	50.000	-9.2

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093888.D Time Analyzed: 13:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.785	5.684	5.884	50.720	50.000	1.4
4,4'-DDE	5.230	5.130	5.330	47.940	50.000	-4.1
4,4'-DDT	6.035	5.934	6.134	46.590	50.000	-6.8
Aldrin	4.224	4.125	4.325	43.990	50.000	-12.0
alpha-BHC	3.277	3.177	3.377	45.710	50.000	-8.6
alpha-Chlordane	5.041	4.940	5.140	46.630	50.000	-6.7
beta-BHC	3.907	3.807	4.007	46.220	50.000	-7.6
Decachlorobiphenyl	7.910	7.810	8.010	48.940	50.000	-2.1
delta-BHC	4.135	4.036	4.236	44.430	50.000	-11.1
Dieldrin	5.361	5.261	5.461	45.590	50.000	-8.8
Endosulfan I	5.097	4.996	5.196	43.510	50.000	-13.0
Endosulfan II	5.932	5.831	6.031	47.250	50.000	-5.5
Endosulfan sulfate	6.334	6.233	6.433	46.850	50.000	-6.3
Endrin	5.637	5.536	5.736	45.280	50.000	-9.4
Endrin aldehyde	6.111	6.010	6.210	44.000	50.000	-12.0
Endrin ketone	6.839	6.739	6.939	47.490	50.000	-5.0
gamma-BHC (Lindane)	3.607	3.507	3.707	44.600	50.000	-10.8
gamma-Chlordane	4.977	4.877	5.077	46.810	50.000	-6.4
Heptachlor	3.945	3.845	4.045	45.400	50.000	-9.2
Heptachlor epoxide	4.727	4.627	4.827	44.970	50.000	-10.1
Methoxychlor	6.610	6.509	6.709	45.430	50.000	-9.1
Tetrachloro-m-xylene	2.774	2.674	2.874	44.780	50.000	-10.4

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 19:06 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.05	9.05	8.95	9.15	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	4.00	3.90	4.10	0.01
beta-BHC	4.52	4.53	4.43	4.63	0.01
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.91	4.91	4.81	5.01	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.64	7.54	7.74	0.00
Endrin aldehyde	6.92	6.92	6.82	7.02	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 19:06 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.13	4.14	4.04	4.24	0.01
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.94	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.01
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.78	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.03	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093906.D Time Analyzed: 19:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.708	6.608	6.808	51.280	50.000	2.6
4,4'-DDE	6.190	6.091	6.291	47.710	50.000	-4.6
4,4'-DDT	7.021	6.922	7.122	43.720	50.000	-12.6
Aldrin	5.255	5.156	5.356	44.150	50.000	-11.7
alpha-BHC	3.993	3.895	4.095	45.670	50.000	-8.7
alpha-Chlordane	6.017	5.917	6.117	45.590	50.000	-8.8
beta-BHC	4.524	4.425	4.625	46.550	50.000	-6.9
Decachlorobiphenyl	9.053	8.953	9.153	45.970	50.000	-8.1
delta-BHC	4.771	4.672	4.872	45.080	50.000	-9.8
Dieldrin	6.343	6.243	6.443	44.310	50.000	-11.4
Endosulfan I	6.067	5.967	6.167	44.670	50.000	-10.7
Endosulfan II	6.792	6.692	6.892	44.900	50.000	-10.2
Endosulfan sulfate	7.157	7.057	7.257	44.430	50.000	-11.1
Endrin	6.571	6.472	6.672	41.650	50.000	-16.7
Endrin aldehyde	6.922	6.823	7.023	44.090	50.000	-11.8
Endrin ketone	7.642	7.542	7.742	45.310	50.000	-9.4
gamma-BHC (Lindane)	4.326	4.227	4.427	45.020	50.000	-10.0
gamma-Chlordane	5.937	5.838	6.038	46.090	50.000	-7.8
Heptachlor	4.914	4.814	5.014	45.190	50.000	-9.6
Heptachlor epoxide	5.681	5.582	5.782	43.240	50.000	-13.5
Methoxychlor	7.498	7.398	7.598	44.720	50.000	-10.6
Tetrachloro-m-xylene	3.537	3.439	3.639	45.460	50.000	-9.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093906.D Time Analyzed: 19:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.784	5.684	5.884	51.980	50.000	4.0
4,4'-DDE	5.229	5.130	5.330	49.710	50.000	-0.6
4,4'-DDT	6.034	5.934	6.134	43.660	50.000	-12.7
Aldrin	4.223	4.125	4.325	45.790	50.000	-8.4
alpha-BHC	3.276	3.177	3.377	47.430	50.000	-5.1
alpha-Chlordane	5.039	4.940	5.140	48.080	50.000	-3.8
beta-BHC	3.906	3.807	4.007	48.150	50.000	-3.7
Decachlorobiphenyl	7.909	7.810	8.010	45.120	50.000	-9.8
delta-BHC	4.134	4.036	4.236	46.050	50.000	-7.9
Dieldrin	5.360	5.261	5.461	46.390	50.000	-7.2
Endosulfan I	5.095	4.996	5.196	44.980	50.000	-10.0
Endosulfan II	5.931	5.831	6.031	46.350	50.000	-7.3
Endosulfan sulfate	6.333	6.233	6.433	45.570	50.000	-8.9
Endrin	5.635	5.536	5.736	44.130	50.000	-11.7
Endrin aldehyde	6.110	6.010	6.210	43.640	50.000	-12.7
Endrin ketone	6.837	6.739	6.939	45.440	50.000	-9.1
gamma-BHC (Lindane)	3.606	3.507	3.707	46.060	50.000	-7.9
gamma-Chlordane	4.976	4.877	5.077	48.720	50.000	-2.6
Heptachlor	3.944	3.845	4.045	45.820	50.000	-8.4
Heptachlor epoxide	4.726	4.627	4.827	46.510	50.000	-7.0
Methoxychlor	6.609	6.509	6.709	42.130	50.000	-15.7
Tetrachloro-m-xylene	2.774	2.674	2.874	46.540	50.000	-6.9

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 02/04/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 09:14 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.01
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.91	4.81	5.01	-0.01
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.80	6.79	6.69	6.89	-0.01
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.64	7.54	7.74	0.00
Endrin aldehyde	6.93	6.92	6.82	7.02	-0.01
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 02/04/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 09:14 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.78	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.03	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL094038.D Time Analyzed: 09:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.710	6.608	6.808	56.810	50.000	13.6
4,4'-DDE	6.192	6.091	6.291	54.800	50.000	9.6
4,4'-DDT	7.024	6.922	7.122	48.210	50.000	-3.6
Aldrin	5.257	5.156	5.356	52.720	50.000	5.4
alpha-BHC	3.995	3.895	4.095	52.230	50.000	4.5
alpha-Chlordane	6.018	5.917	6.117	51.120	50.000	2.2
beta-BHC	4.526	4.425	4.625	51.890	50.000	3.8
Decachlorobiphenyl	9.056	8.953	9.153	47.740	50.000	-4.5
delta-BHC	4.773	4.672	4.872	53.360	50.000	6.7
Dieldrin	6.344	6.243	6.443	50.600	50.000	1.2
Endosulfan I	6.069	5.967	6.167	50.660	50.000	1.3
Endosulfan II	6.795	6.692	6.892	49.090	50.000	-1.8
Endosulfan sulfate	7.159	7.057	7.257	48.700	50.000	-2.6
Endrin	6.573	6.472	6.672	48.440	50.000	-3.1
Endrin aldehyde	6.925	6.823	7.023	48.990	50.000	-2.0
Endrin ketone	7.644	7.542	7.742	48.440	50.000	-3.1
gamma-BHC (Lindane)	4.328	4.227	4.427	51.900	50.000	3.8
gamma-Chlordane	5.939	5.838	6.038	51.540	50.000	3.1
Heptachlor	4.916	4.814	5.014	51.240	50.000	2.5
Heptachlor epoxide	5.684	5.582	5.782	50.560	50.000	1.1
Methoxychlor	7.500	7.398	7.598	48.720	50.000	-2.6
Tetrachloro-m-xylene	3.539	3.439	3.639	51.420	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL094038.D Time Analyzed: 09:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.784	5.684	5.884	59.220	50.000	18.4
4,4'-DDE	5.229	5.130	5.330	56.530	50.000	13.1
4,4'-DDT	6.034	5.934	6.134	50.630	50.000	1.3
Aldrin	4.224	4.125	4.325	53.070	50.000	6.1
alpha-BHC	3.277	3.177	3.377	53.160	50.000	6.3
alpha-Chlordane	5.040	4.940	5.140	53.190	50.000	6.4
beta-BHC	3.907	3.807	4.007	52.950	50.000	5.9
Decachlorobiphenyl	7.910	7.810	8.010	52.230	50.000	4.5
delta-BHC	4.135	4.036	4.236	53.410	50.000	6.8
Dieldrin	5.361	5.261	5.461	53.220	50.000	6.4
Endosulfan I	5.097	4.996	5.196	50.920	50.000	1.8
Endosulfan II	5.931	5.831	6.031	53.410	50.000	6.8
Endosulfan sulfate	6.333	6.233	6.433	52.360	50.000	4.7
Endrin	5.636	5.536	5.736	53.510	50.000	7.0
Endrin aldehyde	6.110	6.010	6.210	51.970	50.000	3.9
Endrin ketone	6.838	6.739	6.939	53.020	50.000	6.0
gamma-BHC (Lindane)	3.607	3.507	3.707	52.640	50.000	5.3
gamma-Chlordane	4.977	4.877	5.077	53.910	50.000	7.8
Heptachlor	3.945	3.845	4.045	50.310	50.000	0.6
Heptachlor epoxide	4.727	4.627	4.827	52.780	50.000	5.6
Methoxychlor	6.609	6.509	6.709	48.950	50.000	-2.1
Tetrachloro-m-xylene	2.774	2.674	2.874	52.080	50.000	4.2

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 02/04/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 14:50 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	4.00	3.90	4.10	0.01
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.91	4.81	5.01	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.64	7.54	7.74	0.00
Endrin aldehyde	6.92	6.92	6.82	7.02	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 02/04/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 14:50 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.94	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.78	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.03	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL05 Date Analyzed: 02/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL094052.D Time Analyzed: 14:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.709	6.608	6.808	53.640	50.000	7.3
4,4'-DDE	6.192	6.091	6.291	51.850	50.000	3.7
4,4'-DDT	7.023	6.922	7.122	42.020	50.000	-16.0
Aldrin	5.256	5.156	5.356	50.810	50.000	1.6
alpha-BHC	3.994	3.895	4.095	51.110	50.000	2.2
alpha-Chlordane	6.018	5.917	6.117	48.500	50.000	-3.0
beta-BHC	4.525	4.425	4.625	50.410	50.000	0.8
Decachlorobiphenyl	9.055	8.953	9.153	45.130	50.000	-9.7
delta-BHC	4.772	4.672	4.872	51.820	50.000	3.6
Dieldrin	6.343	6.243	6.443	47.720	50.000	-4.6
Endosulfan I	6.068	5.967	6.167	48.260	50.000	-3.5
Endosulfan II	6.793	6.692	6.892	46.180	50.000	-7.6
Endosulfan sulfate	7.158	7.057	7.257	45.500	50.000	-9.0
Endrin	6.572	6.472	6.672	44.960	50.000	-10.1
Endrin aldehyde	6.923	6.823	7.023	45.100	50.000	-9.8
Endrin ketone	7.643	7.542	7.742	44.860	50.000	-10.3
gamma-BHC (Lindane)	4.327	4.227	4.427	50.110	50.000	0.2
gamma-Chlordane	5.939	5.838	6.038	49.830	50.000	-0.3
Heptachlor	4.915	4.814	5.014	47.450	50.000	-5.1
Heptachlor epoxide	5.683	5.582	5.782	48.530	50.000	-2.9
Methoxychlor	7.499	7.398	7.598	43.350	50.000	-13.3
Tetrachloro-m-xylene	3.538	3.439	3.639	50.380	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL05 Date Analyzed: 02/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL094052.D Time Analyzed: 14:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.784	5.684	5.884	56.310	50.000	12.6
4,4'-DDE	5.227	5.130	5.330	54.840	50.000	9.7
4,4'-DDT	6.034	5.934	6.134	43.890	50.000	-12.2
Aldrin	4.224	4.125	4.325	52.930	50.000	5.9
alpha-BHC	3.276	3.177	3.377	53.360	50.000	6.7
alpha-Chlordane	5.039	4.940	5.140	51.920	50.000	3.8
beta-BHC	3.906	3.807	4.007	52.630	50.000	5.3
Decachlorobiphenyl	7.909	7.810	8.010	43.570	50.000	-12.9
delta-BHC	4.135	4.036	4.236	53.220	50.000	6.4
Dieldrin	5.360	5.261	5.461	51.590	50.000	3.2
Endosulfan I	5.096	4.996	5.196	48.810	50.000	-2.4
Endosulfan II	5.931	5.831	6.031	50.270	50.000	0.5
Endosulfan sulfate	6.333	6.233	6.433	49.010	50.000	-2.0
Endrin	5.636	5.536	5.736	50.920	50.000	1.8
Endrin aldehyde	6.110	6.010	6.210	47.340	50.000	-5.3
Endrin ketone	6.838	6.739	6.939	46.510	50.000	-7.0
gamma-BHC (Lindane)	3.606	3.507	3.707	52.180	50.000	4.4
gamma-Chlordane	4.976	4.877	5.077	52.700	50.000	5.4
Heptachlor	3.944	3.845	4.045	48.050	50.000	-3.9
Heptachlor epoxide	4.726	4.627	4.827	51.700	50.000	3.4
Methoxychlor	6.609	6.509	6.709	42.000	50.000	-16.0
Tetrachloro-m-xylene	2.774	2.674	2.874	52.500	50.000	5.0

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093726.D Date Analyzed: 01/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:30

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.052	8.950	9.150	18.070	20.000	-9.7
Tetrachloro-m-xylene	3.538	3.490	3.590	18.530	20.000	-7.4
alpha-BHC	3.994	3.940	4.040	9.490	10.000	-5.1
beta-BHC	4.525	4.470	4.580	9.790	10.000	-2.1
gamma-BHC (Lindane)	4.326	4.280	4.380	9.300	10.000	-7.0
Endrin	6.572	6.500	6.640	41.270	50.000	-17.5
4,4'-DDT	7.022	6.950	7.090	82.410	100.000	-17.6
Methoxychlor	7.498	7.430	7.570	190.380	250.000	-23.8

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093726.D Date Analyzed: 01/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:30

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.909	7.810	8.010	17.950	20.000	-10.3
Tetrachloro-m-xylene	2.775	2.720	2.830	17.900	20.000	-10.5
alpha-BHC	3.277	3.230	3.330	8.620	10.000	-13.8
beta-BHC	3.907	3.860	3.960	9.800	10.000	-2.0
gamma-BHC (Lindane)	3.607	3.560	3.660	8.300	10.000	-17.0
Endrin	5.636	5.570	5.710	42.700	50.000	-14.6
4,4'-DDT	6.034	5.960	6.100	96.510	100.000	-3.5
Methoxychlor	6.609	6.540	6.680	209.940	250.000	-16.0

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093875.D **Date Analyzed:** 01/30/2025

Lab Sample No.(PEM): PEM **Time Analyzed:** 09:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.056	8.960	9.160	23.080	20.000	15.4
Tetrachloro-m-xylene	3.539	3.490	3.590	21.630	20.000	8.2
alpha-BHC	3.995	3.940	4.050	11.160	10.000	11.6
beta-BHC	4.525	4.470	4.580	11.630	10.000	16.3
gamma-BHC (Lindane)	4.327	4.280	4.380	11.090	10.000	10.9
Endrin	6.575	6.500	6.650	50.070	50.000	0.1
4,4'-DDT	7.025	6.950	7.100	105.300	100.000	5.3
Methoxychlor	7.501	7.430	7.570	246.960	250.000	-1.2

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093875.D **Date Analyzed:** 01/30/2025

Lab Sample No.(PEM): PEM **Time Analyzed:** 09:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.910	7.810	8.010	22.210	20.000	11.1
Tetrachloro-m-xylene	2.774	2.720	2.820	20.410	20.000	2.1
alpha-BHC	3.276	3.230	3.330	9.790	10.000	-2.1
beta-BHC	3.907	3.860	3.960	11.070	10.000	10.7
gamma-BHC (Lindane)	3.606	3.560	3.660	9.520	10.000	-4.8
Endrin	5.637	5.570	5.710	50.920	50.000	1.8
4,4'-DDT	6.035	5.960	6.110	119.020	100.000	19.0
Methoxychlor	6.610	6.540	6.680	257.770	250.000	3.1

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL094037.D **Date Analyzed:** 02/04/2025

Lab Sample No.(PEM): PEM **Time Analyzed:** 09:00

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.057	8.960	9.160	20.890	20.000	4.5
Tetrachloro-m-xylene	3.539	3.490	3.590	22.330	20.000	11.7
alpha-BHC	3.995	3.940	4.050	11.650	10.000	16.5
beta-BHC	4.526	4.480	4.580	11.780	10.000	17.8
gamma-BHC (Lindane)	4.327	4.280	4.380	11.270	10.000	12.7
Endrin	6.574	6.500	6.640	46.750	50.000	-6.5
4,4'-DDT	7.025	6.950	7.100	94.850	100.000	-5.2
Methoxychlor	7.502	7.430	7.570	219.360	250.000	-12.3

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL094037.D **Date Analyzed:** 02/04/2025

Lab Sample No.(PEM): PEM **Time Analyzed:** 09:00

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.911	7.810	8.010	22.030	20.000	10.2
Tetrachloro-m-xylene	2.774	2.720	2.820	21.750	20.000	8.8
alpha-BHC	3.276	3.230	3.330	10.410	10.000	4.1
beta-BHC	3.907	3.860	3.960	11.650	10.000	16.5
gamma-BHC (Lindane)	3.607	3.560	3.660	9.960	10.000	-0.4
Endrin	5.637	5.570	5.710	52.100	50.000	4.2
4,4'-DDT	6.035	5.960	6.110	107.610	100.000	7.6
Methoxychlor	6.610	6.540	6.680	234.420	250.000	-6.2

Analytical Sequence

Client: RU2 Engineering, LLC	SDG No.: Q1206
Project: NYCDDC SANTWOBR Brooklyn Bridge BI	Instrument ID: ECD_L
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 01/21/2025 01/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	01/21/2025	10:16	PL093725.D	9.05	3.54
PEM	PEM	01/21/2025	10:30	PL093726.D	9.05	3.54
RESCHK	RESCHK	01/21/2025	10:43	PL093727.D	9.05	3.54
PSTDICC100	PSTDICC100	01/21/2025	10:57	PL093728.D	9.05	3.54
PSTDICC075	PSTDICC075	01/21/2025	11:10	PL093729.D	9.05	3.54
PSTDICC050	PSTDICC050	01/21/2025	11:24	PL093730.D	9.05	3.54
PSTDICC025	PSTDICC025	01/21/2025	11:38	PL093731.D	9.05	3.54
PSTDICC005	PSTDICC005	01/21/2025	11:51	PL093732.D	9.05	3.54
PCHLORICC500	PCHLORICC500	01/21/2025	12:32	PL093735.D	9.05	3.54
PTOXICC500	PTOXICC500	01/21/2025	13:39	PL093740.D	9.05	3.54
IBLK	IBLK	01/30/2025	09:31	PL093874.D	9.06	3.54
PEM	PEM	01/30/2025	09:45	PL093875.D	9.06	3.54
PSTDCCC050	PSTDCCC050	01/30/2025	09:58	PL093876.D	9.06	3.54
WC-5MS	Q1209-05MS	01/30/2025	11:25	PL093882.D	9.05	3.54
WC-5MSD	Q1209-05MSD	01/30/2025	11:38	PL093883.D	9.05	3.54
PB166334BL	PB166334BL	01/30/2025	11:51	PL093884.D	9.05	3.54
PB166334BS	PB166334BS	01/30/2025	12:05	PL093885.D	9.05	3.54
IBLK	IBLK	01/30/2025	13:40	PL093887.D	9.06	3.54
PSTDCCC050	PSTDCCC050	01/30/2025	13:53	PL093888.D	9.06	3.54
JPP-20.1-012725	Q1206-03	01/30/2025	16:24	PL093897.D	9.05	3.54
IBLK	IBLK	01/30/2025	17:59	PL093904.D	9.05	3.54
PSTDCCC050	PSTDCCC050	01/30/2025	19:06	PL093906.D	9.05	3.54
IBLK	IBLK	02/04/2025	08:47	PL094036.D	9.06	3.54
PEM	PEM	02/04/2025	09:00	PL094037.D	9.06	3.54
PSTDCCC050	PSTDCCC050	02/04/2025	09:14	PL094038.D	9.06	3.54
JPP-16.3-012725	Q1206-07	02/04/2025	09:47	PL094040.D	9.06	3.54
IBLK	IBLK	02/04/2025	14:37	PL094051.D	9.05	3.54
PSTDCCC050	PSTDCCC050	02/04/2025	14:50	PL094052.D	9.06	3.54

Analytical Sequence

Client: RU2 Engineering, LLC	SDG No.: Q1206
Project: NYCDDC SANTWOBR Brooklyn Bridge BI	Instrument ID: ECD_L
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 01/21/2025 01/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	01/21/2025	10:16	PL093725.D	7.91	2.78
PEM	PEM	01/21/2025	10:30	PL093726.D	7.91	2.78
RESCHK	RESCHK	01/21/2025	10:43	PL093727.D	7.91	2.77
PSTDICC100	PSTDICC100	01/21/2025	10:57	PL093728.D	7.91	2.78
PSTDICC075	PSTDICC075	01/21/2025	11:10	PL093729.D	7.91	2.77
PSTDICC050	PSTDICC050	01/21/2025	11:24	PL093730.D	7.91	2.77
PSTDICC025	PSTDICC025	01/21/2025	11:38	PL093731.D	7.91	2.77
PSTDICC005	PSTDICC005	01/21/2025	11:51	PL093732.D	7.91	2.77
PCHLORICC500	PCHLORICC500	01/21/2025	12:32	PL093735.D	7.91	2.77
PTOXICC500	PTOXICC500	01/21/2025	13:39	PL093740.D	7.91	2.77
IBLK	IBLK	01/30/2025	09:31	PL093874.D	7.91	2.77
PEM	PEM	01/30/2025	09:45	PL093875.D	7.91	2.77
PSTDCCC050	PSTDCCC050	01/30/2025	09:58	PL093876.D	7.91	2.77
WC-5MS	Q1209-05MS	01/30/2025	11:25	PL093882.D	7.91	2.77
WC-5MSD	Q1209-05MSD	01/30/2025	11:38	PL093883.D	7.91	2.77
PB166334BL	PB166334BL	01/30/2025	11:51	PL093884.D	7.91	2.77
PB166334BS	PB166334BS	01/30/2025	12:05	PL093885.D	7.91	2.77
IBLK	IBLK	01/30/2025	13:40	PL093887.D	7.91	2.77
PSTDCCC050	PSTDCCC050	01/30/2025	13:53	PL093888.D	7.91	2.77
JPP-20.1-012725	Q1206-03	01/30/2025	16:24	PL093897.D	7.91	2.77
IBLK	IBLK	01/30/2025	17:59	PL093904.D	7.91	2.77
PSTDCCC050	PSTDCCC050	01/30/2025	19:06	PL093906.D	7.91	2.77
IBLK	IBLK	02/04/2025	08:47	PL094036.D	7.91	2.77
PEM	PEM	02/04/2025	09:00	PL094037.D	7.91	2.77
PSTDCCC050	PSTDCCC050	02/04/2025	09:14	PL094038.D	7.91	2.77
JPP-16.3-012725	Q1206-07	02/04/2025	09:47	PL094040.D	7.91	2.77
IBLK	IBLK	02/04/2025	14:37	PL094051.D	7.91	2.77
PSTDCCC050	PSTDCCC050	02/04/2025	14:50	PL094052.D	7.91	2.77

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-16.3-012725

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1206

SAS No.: Q1206

SDG NO.: Q1206

Lab Sample ID: Q1206-07

Date(s) Analyzed: 02/04/2025 02/04/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.02	6.97	7.07	0.65	97.6
	2	6.03	5.98	6.08	1.90	
Heptachlor epoxide	1	5.68	5.63	5.73	0.70	36.5
	2	4.73	4.68	4.78	0.48	
gamma-Chlordane	1	5.94	5.89	5.99	1.40	67
	2	4.97	4.92	5.02	0.70	
alpha-Chlordane	1	6.02	5.97	6.07	2.40	82.4
	2	5.04	4.99	5.09	1.00	
4,4'-DDE	1	6.19	6.14	6.24	2.90	6.7
	2	5.23	5.18	5.28	3.10	
Endrin	1	6.57	6.52	6.62	0.84	56.6
	2	5.64	5.59	5.69	0.47	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-20.1-012725

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1206

SAS No.: Q1206

SDG NO.: Q1206

Lab Sample ID: Q1206-03

Date(s) Analyzed: 01/30/2025 01/30/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.71	6.66	6.76	0.88	54.2
	2	5.78	5.73	5.83	0.50	
4,4'-DDT	1	7.02	6.97	7.07	7.90	15
	2	6.03	5.98	6.08	6.80	
Aldrin	1	5.25	5.20	5.30	0.54	28.1
	2	4.22	4.17	4.27	0.40	
4,4'-DDE	1	6.19	6.14	6.24	3.00	9.5
	2	5.23	5.18	5.28	3.30	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB166334BS

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1206

SAS No.: Q1206

SDG NO.: Q1206

Lab Sample ID: PB166334BS

Date(s) Analyzed: 01/30/2025

01/30/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
delta-BHC	1	4.77	4.72	4.82	14.3	2.8
	2	4.13	4.08	4.18	13.9	
Endosulfan I	1	6.07	6.02	6.12	14.8	2
	2	5.10	5.05	5.15	15.1	
Dieldrin	1	6.34	6.29	6.39	14.9	0.7
	2	5.36	5.31	5.41	15.0	
Endrin aldehyde	1	6.92	6.87	6.97	15.1	1.3
	2	6.11	6.06	6.16	14.9	
Methoxychlor	1	7.50	7.45	7.55	16.2	4.4
	2	6.61	6.56	6.66	15.5	
Endrin ketone	1	7.64	7.59	7.69	15.6	1.3
	2	6.84	6.79	6.89	15.8	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	14.1	1.4
	2	3.61	3.56	3.66	13.9	
Heptachlor	1	4.91	4.86	4.96	15.2	2.7
	2	3.94	3.89	3.99	14.8	
Heptachlor epoxide	1	5.68	5.63	5.73	14.3	2.8
	2	4.73	4.68	4.78	14.7	
gamma-Chlordane	1	5.94	5.89	5.99	15.1	0.7
	2	4.97	4.92	5.02	15.2	
Endrin	1	6.57	6.52	6.62	14.6	3.4
	2	5.63	5.58	5.68	15.1	
4,4'-DDD	1	6.71	6.66	6.76	16.5	2.5
	2	5.78	5.73	5.83	16.1	
4,4'-DDE	1	6.19	6.14	6.24	15.7	1.9
	2	5.23	5.18	5.28	15.4	
4,4'-DDT	1	7.02	6.97	7.07	16.3	3.1
	2	6.03	5.98	6.08	15.8	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB166334BS

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1206

SAS No.: Q1206

SDG NO.: Q1206

Lab Sample ID: PB166334BS

Date(s) Analyzed: 01/30/2025

01/30/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
alpha-BHC	1	3.99	3.94	4.04	14.2	0
	2	3.28	3.23	3.33	14.2	
Aldrin	1	5.26	5.21	5.31	14.3	1.4
	2	4.22	4.17	4.27	14.1	
alpha-Chlordane	1	6.02	5.97	6.07	15.1	0
	2	5.04	4.99	5.09	15.1	
Endosulfan II	1	6.79	6.74	6.84	15.2	2.6
	2	5.93	5.88	5.98	15.6	
Endosulfan sulfate	1	7.16	7.11	7.21	15.3	1.9
	2	6.33	6.28	6.38	15.6	
beta-BHC	1	4.52	4.47	4.57	14.6	0.7
	2	3.91	3.86	3.96	14.5	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-5MS

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1206

SAS No.: Q1206

SDG NO.: Q1206

Lab Sample ID: Q1209-05MS

Date(s) Analyzed: 01/30/2025

01/30/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	18.1	3.3
	2	5.93	5.88	5.98	18.7	
4,4'-DDD	1	6.71	6.66	6.76	19.8	3.6
	2	5.78	5.73	5.83	19.1	
4,4'-DDT	1	7.02	6.97	7.07	19.7	0.5
	2	6.03	5.98	6.08	19.8	
Endrin aldehyde	1	6.92	6.87	6.97	17.7	0
	2	6.11	6.06	6.16	17.7	
Endosulfan sulfate	1	7.16	7.11	7.21	18.3	2.2
	2	6.33	6.28	6.38	18.7	
Methoxychlor	1	7.50	7.45	7.55	20.1	4.1
	2	6.61	6.56	6.66	19.3	
Endrin ketone	1	7.64	7.59	7.69	18.8	1.1
	2	6.84	6.79	6.89	18.6	
alpha-BHC	1	3.99	3.94	4.04	17.0	0.6
	2	3.28	3.23	3.33	17.1	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	17.1	1.2
	2	3.61	3.56	3.66	16.9	
Heptachlor	1	4.91	4.86	4.96	18.1	1.1
	2	3.94	3.89	3.99	17.9	
Aldrin	1	5.26	5.21	5.31	17.1	0
	2	4.22	4.17	4.27	17.1	
beta-BHC	1	4.53	4.48	4.58	17.9	1.1
	2	3.91	3.86	3.96	17.7	
delta-BHC	1	4.77	4.72	4.82	17.3	2.3
	2	4.14	4.09	4.19	16.9	
Heptachlor epoxide	1	5.68	5.63	5.73	17.2	3.4
	2	4.73	4.68	4.78	17.8	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-5MS

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206

SAS No.: Q1206 SDG NO.: Q1206

Lab Sample ID: Q1209-05MS

Date(s) Analyzed: 01/30/2025 01/30/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	17.8	2.8
	2	5.10	5.05	5.15	18.3	
gamma-Chlordane	1	5.94	5.89	5.99	18.0	2.2
	2	4.98	4.93	5.03	18.4	
alpha-Chlordane	1	6.02	5.97	6.07	18.0	2.2
	2	5.04	4.99	5.09	18.4	
4,4'-DDE	1	6.19	6.14	6.24	18.9	1.1
	2	5.23	5.18	5.28	18.7	
Dieldrin	1	6.34	6.29	6.39	17.8	1.7
	2	5.36	5.31	5.41	18.1	
Endrin	1	6.57	6.52	6.62	18.1	3.8
	2	5.63	5.58	5.68	18.8	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-5MSD

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206

SAS No.: Q1206 SDG NO.: Q1206

Lab Sample ID: Q1209-05MSD

Date(s) Analyzed: 01/30/2025 01/30/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	18.3	3.2
	2	5.93	5.88	5.98	18.9	
4,4'-DDD	1	6.71	6.66	6.76	20.1	3
	2	5.78	5.73	5.83	19.5	
4,4'-DDT	1	7.02	6.97	7.07	19.9	0
	2	6.03	5.98	6.08	19.9	
Endrin aldehyde	1	6.92	6.87	6.97	17.9	1.1
	2	6.11	6.06	6.16	17.7	
Endosulfan sulfate	1	7.16	7.11	7.21	18.7	1.1
	2	6.33	6.28	6.38	18.9	
Methoxychlor	1	7.50	7.45	7.55	20.2	5.1
	2	6.61	6.56	6.66	19.2	
Endrin ketone	1	7.64	7.59	7.69	19.0	0
	2	6.84	6.79	6.89	19.0	
alpha-BHC	1	3.99	3.94	4.04	17.3	0.6
	2	3.28	3.23	3.33	17.4	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	17.3	0.6
	2	3.61	3.56	3.66	17.2	
Heptachlor	1	4.91	4.86	4.96	18.1	0
	2	3.94	3.89	3.99	18.1	
Aldrin	1	5.26	5.21	5.31	17.4	1.2
	2	4.22	4.17	4.27	17.2	
beta-BHC	1	4.52	4.47	4.57	18.0	0
	2	3.91	3.86	3.96	18.0	
delta-BHC	1	4.77	4.72	4.82	17.5	2.3
	2	4.13	4.08	4.18	17.1	
Heptachlor epoxide	1	5.68	5.63	5.73	17.4	3.4
	2	4.73	4.68	4.78	18.0	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-5MSD

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206

SAS No.: Q1206 SDG NO.: Q1206

Lab Sample ID: Q1209-05MSD

Date(s) Analyzed: 01/30/2025 01/30/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	18.0	2.7
	2	5.10	5.05	5.15	18.5	
gamma-Chlordane	1	5.94	5.89	5.99	18.2	1.6
	2	4.98	4.93	5.03	18.5	
alpha-Chlordane	1	6.02	5.97	6.07	18.3	1.6
	2	5.04	4.99	5.09	18.6	
4,4'-DDE	1	6.19	6.14	6.24	19.2	1.6
	2	5.23	5.18	5.28	18.9	
Dieldrin	1	6.34	6.29	6.39	18.0	1.7
	2	5.36	5.31	5.41	18.3	
Endrin	1	6.57	6.52	6.62	18.2	5.3
	2	5.64	5.59	5.69	19.2	