

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093874.D	PIBLK-PL093874.D	Decachlorobiphenyl	1	20	22.5	113		43	140
		Tetrachloro-m-xylene	1	20	19.9	100		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	18.1	91		77	126
Q1209-05MS	WC-5MS	Decachlorobiphenyl	1	20	22.2	111		10	148
		Tetrachloro-m-xylene	1	20	20.2	101		10	159
		Decachlorobiphenyl	2	20	21.5	107		10	148
		Tetrachloro-m-xylene	2	20	19.4	97		10	159
Q1209-05MSD	WC-5MSD	Decachlorobiphenyl	1	20	22.2	111		10	148
		Tetrachloro-m-xylene	1	20	20.4	102		10	159
		Decachlorobiphenyl	2	20	21.6	108		10	148
		Tetrachloro-m-xylene	2	20	19.8	99		10	159
PB166334BL	PB166334BL	Decachlorobiphenyl	1	20	24.8	124		10	148
		Tetrachloro-m-xylene	1	20	22.0	110		10	159
		Decachlorobiphenyl	2	20	24.4	122		10	148
		Tetrachloro-m-xylene	2	20	20.6	103		10	159
PB166334BS	PB166334BS	Decachlorobiphenyl	1	20	19.6	98		10	148
		Tetrachloro-m-xylene	1	20	16.9	84		10	159
		Decachlorobiphenyl	2	20	19.0	95		10	148
		Tetrachloro-m-xylene	2	20	15.9	80		10	159
I.BLK-PL093887.D	PIBLK-PL093887.D	Decachlorobiphenyl	1	20	21.4	107		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	21.2	106		43	140
		Tetrachloro-m-xylene	2	20	17.5	88		77	126
Q1206-03	JPP-20.1-012725	Decachlorobiphenyl	1	20	13.9	69		10	148
		Tetrachloro-m-xylene	1	20	13.0	65		10	159
		Decachlorobiphenyl	2	20	10.9	54		10	148
		Tetrachloro-m-xylene	2	20	12.7	64		10	159
I.BLK-PL093904.D	PIBLK-PL093904.D	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	19.1	96		77	126
		Decachlorobiphenyl	2	20	16.0	80		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
I.BLK-PL094036.D	PIBLK-PL094036.D	Decachlorobiphenyl	1	20	23.0	115		43	140
		Tetrachloro-m-xylene	1	20	22.4	112		77	126
		Decachlorobiphenyl	2	20	23.3	116		43	140
		Tetrachloro-m-xylene	2	20	21.9	110		77	126
Q1206-07	JPP-16.3-012725	Decachlorobiphenyl	1	20	28.9	145		10	148

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1206-07	JPP-16.3-012725	Tetrachloro-m-xylene	1	20	13.1	66		10	159
		Decachlorobiphenyl	2	20	7.45	37		10	148
		Tetrachloro-m-xylene	2	20	13.8	69		10	159
I.BLK-PL094051.D	PIBLK-PL094051.D	Decachlorobiphenyl	1	20	20.1	101		43	140
		Tetrachloro-m-xylene	1	20	21.9	109		77	126
		Decachlorobiphenyl	2	20	19.6	98		43	140
		Tetrachloro-m-xylene	2	20	21.4	107		77	126