

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL093882.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	WC-5MS											
Q1209-05MS	alpha-BHC	18.83	0	17.1	ug/kg	91				60	144	
	beta-BHC	18.83	0	17.9	ug/kg	95				54	143	
	delta-BHC	18.83	0	17.3	ug/kg	92				47	144	
	gamma-BHC (Lindane)	18.83	0	17.1	ug/kg	91				61	140	
	Heptachlor	18.83	0	18.1	ug/kg	96				63	135	
	Aldrin	18.83	0	17.1	ug/kg	91				49	139	
	Heptachlor epoxide	18.83	0	17.8	ug/kg	95				32	180	
	Endosulfan I	18.83	0	18.3	ug/kg	97				56	142	
	Dieldrin	18.83	0	18.1	ug/kg	96				47	161	
	4,4'-DDE	18.83	0	18.9	ug/kg	100				55	136	
	Endrin	18.83	0	18.8	ug/kg	100				57	139	
	Endosulfan II	18.83	0	18.7	ug/kg	99				40	163	
	4,4'-DDD	18.83	0	19.8	ug/kg	105				37	192	
	Endosulfan sulfate	18.83	0	18.7	ug/kg	99				62	139	
	4,4'-DDT	18.83	0	19.8	ug/kg	105				51	146	
	Methoxychlor	18.83	0	20.1	ug/kg	107				54	136	
	Endrin ketone	18.83	0	18.8	ug/kg	100				60	129	
	Endrin aldehyde	18.83	0	17.7	ug/kg	94				59	132	
	alpha-Chlordane	18.83	0	18.4	ug/kg	98				30	192	
	gamma-Chlordane	18.83	0	18.4	ug/kg	98				44	175	