

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: **8081B** Datafile : PL093885.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166334BS	alpha-BHC	16.65	14.2	ug/kg	85				84	123	
	beta-BHC	16.65	14.6	ug/kg	88				82	123	
	delta-BHC	16.65	14.3	ug/kg	86				83	126	
	gamma-BHC (Lindane)	16.65	14.1	ug/kg	85				83	125	
	Heptachlor	16.65	15.2	ug/kg	91				83	122	
	Aldrin	16.65	14.3	ug/kg	86				82	124	
	Heptachlor epoxide	16.65	14.7	ug/kg	88				83	120	
	Endosulfan I	16.65	15.1	ug/kg	91				81	124	
	Dieldrin	16.65	15.0	ug/kg	90				85	121	
	4,4'-DDE	16.65	15.7	ug/kg	94				81	123	
	Endrin	16.65	15.1	ug/kg	91				76	130	
	Endosulfan II	16.65	15.6	ug/kg	94				80	125	
	4,4'-DDD	16.65	16.5	ug/kg	99				80	131	
	Endosulfan sulfate	16.65	15.6	ug/kg	94				81	122	
	4,4'-DDT	16.65	16.3	ug/kg	98				70	129	
	Methoxychlor	16.65	16.2	ug/kg	97				60	119	
	Endrin ketone	16.65	15.8	ug/kg	95				77	132	
	Endrin aldehyde	16.65	15.1	ug/kg	91				79	124	
	alpha-Chlordane	16.65	15.1	ug/kg	91				84	120	
	gamma-Chlordane	16.65	15.2	ug/kg	91				83	122	