

DATA PACKAGE

GC SEMI-VOLATILES

PROJECT NAME : NYCDDC SANTWOBR BROOKLYN BRIDGE BBMCR

RU2 ENGINEERING, LLC

2 Melinda Drive

Monroe Township, NJ - 08831

Phone No: 732-261-2236

ORDER ID : Q1206

ATTENTION : Rutu Manani



Laboratory Certification ID # 20012



1) TCLP HERBICIDE Data	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For TCLP Herbicide	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) MS/MSD Summary	15
9.3) LCS/LCSD Summary	17
9.4) Method Blank Summary	18
10) Sample Data	19
10.1) PB166318TB	20
10.2) JPP-20.1-012725	24
10.3) JPP-16.3-012725	28
11) Calibration Data Summary	32
11.1) Initial Calibration Data	33
11.1.1) PS011425	33
11.2) Continued Calibration Data	97
11.2.1) PS028989.D	97
11.2.2) PS028998.D	111
11.3) Analytical Seq	125
12) Compound Detection Summary	127
13) QC Sample Data	130
13.1) Method Blank Data	131
13.2) PIBLK Data	135
13.3) LCS Data	147
13.4) MS Data	158
13.5) MSD Data	169
14) Manual Integration	180
15) Analytical Runlogs	183
16) Extraction Logs	188
16.1) PB166318.pdf	188

Table Of Contents for Q1206

16.2) PB166318IC.pdf	192
16.3) PB166382.pdf	193
16.4) PB166382IC.pdf	195
17) Standard Prep Logs	196
18) Shipping Document	246
18.1) Chain Of Custody	247
18.2) Lab Certificate	248
18.3) Internal COC	3

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : Q1206

Project ID : NYCDDC SANTWOBR Brooklyn Bridge BBMCR

Client : RU2 Engineering, LLC

Lab Sample Number

Q1206-01
Q1206-02
Q1206-03
Q1206-04
Q1206-05
Q1206-06
Q1206-07
Q1206-08

Client Sample Number

JPP-20.1-012725
JPP-20.1-012725
JPP-20.1-012725
JPP-20.1-012725
JPP-16.3-012725
JPP-16.3-012725
JPP-16.3-012725
JPP-16.3-012725

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 2/1/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

RU2 Engineering, LLC

Project Name: NYCDDC SANTWOBR Brooklyn Bridge BBMCR

Project # N/A

Chemtech Project # Q1206

Test Name: TCLP Herbicide

A. Number of Samples and Date of Receipt:

8 Solid samples were received on 01/28/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Diesel Range Organics, Gasoline Range Organics, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, Paint Filter, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL and VOCMS Group1. This data package contains results for TCLP Herbicide.

C. Analytical Techniques:

The analysis was performed on instrument ECD_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0.5 µm df, Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 µm df, Catalog #: 11324. The analysis of TCLP Herbicides was based on method 8151A and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS {Q1206-04MS} with File ID: PS028994.D recoveries met the requirements for all compounds except for 2,4,5-TP(Silvex)[158%] Due to matrix interference.

The MSD {Q1206-04MSD} with File ID: PS028995.D recoveries met the acceptable requirements except for 2,4,5-TP(Silvex)[157%] Due to matrix interference.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1206

MATRIX: TCLP

METHOD: 8151A/3510/1311

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The MS {Q1206-04MS} with File ID: PS028994.D recoveries met the requirements for all compounds except for 2,4,5-TP(Silvex)[158%] Due to matrix interference. The MSD {Q1206-04MSD} with File ID: PS028995.D recoveries met the acceptable requirements except for 2,4,5-TP(Silvex)[157%] Due to matrix interference. The Blank Spike met requirements for all samples . The RPD met criteria .		✓	
7. Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8. Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

		NA	NO	YES
9.	Analysis Holding Time Met			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			

ADDITIONAL COMMENTS:

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1206

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 02/01/2025

LAB CHRONICLE

OrderID:	Q1206	OrderDate:	1/28/2025 11:18:51 AM
Client:	RU2 Engineering, LLC	Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR
Contact:	Rutu Manani	Location:	E11,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1206-01	JPP-20.1-012725	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/27/25	01/29/25	01/30/25 01/29/25	01/28/25
Q1206-03	JPP-20.1-012725	SOIL	PCB	8082A	01/27/25	01/29/25	01/29/25	01/28/25
Q1206-04	JPP-20.1-012725	TCLP	TCLP Herbicide	8151A	01/27/25	01/29/25	01/30/25	01/28/25
Q1206-05	JPP-16.3-012725	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/27/25	01/29/25	01/30/25 01/29/25	01/28/25
Q1206-07	JPP-16.3-012725	SOIL	PCB	8082A	01/27/25	01/29/25	01/29/25	01/28/25
Q1206-08	JPP-16.3-012725	TCLP	TCLP Herbicide	8151A	01/27/25	01/29/25	01/30/25	01/28/25

Hit Summary Sheet
SW-846

SDG No.: Q1206

Order ID: Q1206

Client: RU2 Engineering, LLC

Project ID: NYCDDC SANTWOBR Brooklyn Bri

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								

Total Concentration: 0.000



QC SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS028900.D	PIBLK-PS028900.D	2,4-DCAA	1	500	474	95		39	175
		2,4-DCAA	2	500	492	98		39	175
I.BLK-PS028988.D	PIBLK-PS028988.D	2,4-DCAA	1	500	512	102		39	175
		2,4-DCAA	2	500	489	98		39	175
PB166382BL	PB166382BL	2,4-DCAA	1	500	445	89		39	175
		2,4-DCAA	2	500	404	81		39	175
PB166382BS	PB166382BS	2,4-DCAA	1	500	510	102		39	175
		2,4-DCAA	2	500	465	93		39	175
PB166318TB	PB166318TB	2,4-DCAA	1	500	462	92		39	175
		2,4-DCAA	2	500	313	63		39	175
Q1206-04	JPP-20.1-012725	2,4-DCAA	1	500	385	77		39	175
		2,4-DCAA	2	500	241	48		39	175
Q1206-04MS	JPP-20.1-012725MS	2,4-DCAA	1	500	424	85		39	175
		2,4-DCAA	2	500	267	53		39	175
Q1206-04MSD	JPP-20.1-012725MSD	2,4-DCAA	1	500	424	85		39	175
		2,4-DCAA	2	500	265	53		39	175
Q1206-08	JPP-16.3-012725	2,4-DCAA	1	500	382	76		39	175
		2,4-DCAA	2	500	263	53		39	175
I.BLK-PS028997.D	PIBLK-PS028997.D	2,4-DCAA	1	500	510	102		39	175
		2,4-DCAA	2	500	499	100		39	175

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8151A

DataFile : PS028994.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JPP-20.1-012725MS											
Q1206-04MS	2,4-D	50	0	50.4	ug/L	101				65	135	
	2,4,5-TP(Silvex)	50	0	79.2	ug/L	158	*			62	139	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8151A

DataFile : PS028995.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JPP-20.1-012725MSD											
Q1206-04MSD	2,4-D	50	0	49.9	ug/L	100		1		65	135	20
	2,4,5-TP(Silvex)	50	0	78.7	ug/L	157	*	1		62	139	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8151A

Datafile : PS028991.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB166382BS	2,4-D	5	4.80	ug/L	96				83	130		
	2,4,5-TP(Silvex)	5	4.90	ug/L	98				78	127		

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166382BL

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206

SAS No.: Q1206 SDG NO.: Q1206

Lab Sample ID: PB166382BL

Lab File ID: PS028990.D

Matrix: (soil/water) water

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 01/29/2025

Date Analyzed (1): 01/30/2025

Date Analyzed (2): 01/30/2025

Time Analyzed (1): 13:44

Time Analyzed (2): 13:44

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column (1): RTX-CLP ID: 0.32 (mm)

GC Column (2): RTX-CLP2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB166382BS	PB166382BS	PS028991.D	01/30/2025	01/30/2025
PB166318TB	PB166318TB	PS028992.D	01/30/2025	01/30/2025
JPP-20.1-012725	Q1206-04	PS028993.D	01/30/2025	01/30/2025
JPP-20.1-012725MS	Q1206-04MS	PS028994.D	01/30/2025	01/30/2025
JPP-20.1-012725MSD	Q1206-04MSD	PS028995.D	01/30/2025	01/30/2025
JPP-16.3-012725	Q1206-08	PS028996.D	01/30/2025	01/30/2025

COMMENTS: _____



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Report of Analysis

Client:	RU2 Engineering, LLC		Date Collected:		
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR		Date Received:	01/29/25	
Client Sample ID:	PB166318TB		SDG No.:	Q1206	
Lab Sample ID:	PB166318TB		Matrix:	TCLP	
Analytical Method:	SW8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028992.D	1	01/29/25 12:09	01/30/25 14:32	PB166382

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	20.0	U	4.90	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	20.0	U	4.50	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	462		39 - 175	92%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 14:32
Operator : AR\AJ
Sample : PB166318TB
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB166318TB

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

4) S 2,4-DCAA 7.193 7.670 1287.2E6 348.8E6 462.337m 312.635 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 14:32
Operator : AR\AJ
Sample : PB166318TB
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
PB166318TB


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.193 min
Delta R.T.: -0.005 min
Response: 1287150834
Conc: 462.34 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166318TB

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025



#4 2,4-DCAA

R.T.: 7.670 min
Delta R.T.: -0.007 min
Response: 348841704
Conc: 312.63 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/27/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	JPP-20.1-012725	SDG No.:	Q1206
Lab Sample ID:	Q1206-04	Matrix:	TCLP
Analytical Method:	SW8151A	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	8151A	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028993.D	1	01/29/25 12:09	01/30/25 14:56	PB166382

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	20.0	U	4.90	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	20.0	U	4.50	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	385		39 - 175	77%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 14:56
Operator : AR\AJ
Sample : Q1206-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.194	7.671	1070.8E6	268.9E6	384.617	240.970 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 14:56
Operator : AR\AJ
Sample : Q1206-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.194 min
Delta R.T.: -0.004 min
Response: 1070779631
Conc: 384.62 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725



#4 2,4-DCAA

R.T.: 7.671 min
Delta R.T.: -0.006 min
Response: 268877060
Conc: 240.97 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/27/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	JPP-16.3-012725	SDG No.:	Q1206
Lab Sample ID:	Q1206-08	Matrix:	TCLP
Analytical Method:	SW8151A	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	8151A	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028996.D	1	01/29/25 12:09	01/30/25 16:07	PB166382

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	20.0	U	4.90	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	20.0	U	4.50	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	382		39 - 175	76%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
 Data File : PS028996.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2025 16:07
 Operator : AR\AJ
 Sample : Q1206-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 JPP-16.3-012725

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 05:21:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.193	7.669	1063.4E6	293.3E6	381.954	262.884 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 16:07
Operator : AR\AJ
Sample : Q1206-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
JPP-16.3-012725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.193 min
Delta R.T.: -0.005 min
Response: 1063365149
Conc: 381.95 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-16.3-012725



#4 2,4-DCAA

R.T.: 7.669 min
Delta R.T.: -0.008 min
Response: 293329512
Conc: 262.88 ng/ml



CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

RETENTION TIMES OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Instrument ID: ECD_S Calibration Date(s): 01/14/2025 01/14/2025

Calibration Times: 10:31 12:07

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS028901.D</u>	RT 500 = <u>PS028902.D</u>
	RT 750 = <u>PS028903.D</u>	RT 1000 = <u>PS028904.D</u>
		RT 1500 = <u>PS028905.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.19	9.19	9.19	9.19	9.19	9.19	9.09	9.29
2,4-D	8.32	8.32	8.32	8.32	8.32	8.32	8.22	8.42
2,4-DCAA	7.20	7.20	7.20	7.20	7.20	7.20	7.10	7.30

RETENTION TIMES OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Instrument ID: ECD_S **Calibration Date(s):** 01/14/2025 01/14/2025

Calibration Times: 10:31 12:07

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS028901.D</u>	RT 500 = <u>PS028902.D</u>
RT 750 = <u>PS028903.D</u>	RT 1000 = <u>PS028904.D</u>	RT 1500 = <u>PS028905.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW	
							FROM	TO
2,4,5-TP(Silvex)	9.81	9.81	9.81	9.81	9.81	9.81	9.71	9.91
2,4-D	8.92	8.91	8.91	8.91	8.91	8.91	8.81	9.01
2,4-DCAA	7.68	7.68	7.68	7.68	7.68	7.68	7.58	7.78

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Instrument ID: ECD_S **Calibration Date(s):** 01/14/2025 01/14/2025
Calibration Times: 10:31 12:07

GC Column: RTX-CLP **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS028901.D</u>		CF 500 = <u>PS028902.D</u>			
CF 750 = <u>PS028903.D</u>		CF 1000 = <u>PS028904.D</u>		CF 1500 = <u>PS028905.D</u>			
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	21246200000	19217800000	18444300000	17622300000	16707400000	18647600000	9
2,4-D	3794730000	3389210000	3238030000	3095840000	2967500000	3297060000	10
2,4-DCAA	3179220000	2766210000	2659700000	2530920000	2413760000	2709960000	11

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Instrument ID: ECD_S **Calibration Date(s):** 01/14/2025 01/14/2025
Calibration Times: 10:31 12:07

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS028901.D</u>	CF 500 = <u>PS028902.D</u>				
CF 750 = <u>PS028903.D</u>		CF 1000 = <u>PS028904.D</u>	CF 1500 = <u>PS028905.D</u>				
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	9615710000	9419870000	9409010000	9233020000	9015720000	9338670000	2
2,4-D	1602310000	1486700000	1468930000	1440130000	1429250000	1485460000	5
2,4-DCAA	1189550000	1103610000	1095350000	1074740000	1070080000	1106670000	4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
 Data File : PS028901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2025 10:31
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 11:39:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 11:39:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.198	7.679	635.8E6	237.9E6	217.795	208.246
Target Compounds							
1) T	Dalapon	2.617	2.667	556.0E6	386.3E6	184.941	186.638
2) T	3,5-DICHL...	6.375	6.644	837.5E6	324.0E6	201.140	192.113
3) T	4-Nitroph...	6.997	7.209	352.0E6	174.3E6	192.726	190.729
5) T	DICAMBA	7.383	7.875	2411.2E6	1035.2E6	197.792	185.939
6) T	MCPD	7.561	7.975	110.1E6	54294236	16.847	18.257
7) T	MCPA	7.709	8.216	180.7E6	79751056	18.378	18.719
8) T	DICHLORPROP	8.087	8.587	684.7E6	277.9E6	206.038	193.740
9) T	2,4-D	8.318	8.915	713.4E6	301.2E6	202.882	196.165
10) T	Pentachlo...	8.613	9.438	10288.3E6	4528.0E6	203.718	192.995
11) T	2,4,5-TP ...	9.189	9.814	4036.8E6	1827.0E6	203.413	192.064
12) T	2,4,5-T	9.482	10.231	4028.8E6	1754.4E6	202.825	192.500
13) T	2,4-DB	10.053	10.796	727.5E6	194.8E6	199.797	193.297
14) T	DINOSEB	11.255	11.173	3520.9E6	1259.6E6	203.928	192.833
15) T	Picloram	11.067	12.257	6433.5E6	2426.6E6	198.927	182.722
16) T	DCPA	11.550	12.211	6140.1E6	2196.1E6	206.127	192.334



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
Data File : PS028901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2025 10:31
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 11:39:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 11:39:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.617 min
Delta R.T.: 0.000 min
Response: 555991432
Conc: 184.94 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC200




#1 Dalapon

R.T.: 2.667 min
Delta R.T.: 0.000 min
Response: 386313856
Conc: 186.64 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.375 min
Delta R.T.: 0.000 min
Response: 837542116
Conc: 201.14 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.644 min
Delta R.T.: 0.000 min
Response: 323964765
Conc: 192.11 ng/ml



#3 4-Nitrophenol


R.T.: 6.997 min
Delta R.T.: 0.000 min
Response: 352019894
Conc: 192.73 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC200




#3 4-Nitrophenol

R.T.: 7.209 min
Delta R.T.: 0.000 min
Response: 174316954
Conc: 190.73 ng/ml




#4 2,4-DCAA

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 635843662
Conc: 217.79 ng/ml



#4 2,4-DCAA


R.T.: 7.679 min
Delta R.T.: 0.000 min
Response: 237909654
Conc: 208.25 ng/ml



#5 DICAMBA


R.T.: 7.383 min
Delta R.T.: 0.000 min
Response: 2411181004
Conc: 197.79 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC200




#5 DICAMBA

R.T.: 7.875 min
Delta R.T.: 0.000 min
Response: 1035181205
Conc: 185.94 ng/ml




#6 MCPP

R.T.: 7.561 min
Delta R.T.: 0.000 min
Response: 110118735
Conc: 16.85 ug/ml



#6 MCPP


R.T.: 7.975 min
Delta R.T.: 0.000 min
Response: 54294236
Conc: 18.26 ug/ml



#7 MCPA


R.T.: 7.709 min
Delta R.T.: 0.000 min
Response: 180716111
Conc: 18.38 ug/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC200




#7 MCPA

R.T.: 8.216 min
Delta R.T.: 0.000 min
Response: 79751056
Conc: 18.72 ug/ml




#8 DICHLORPROP

R.T.: 8.087 min
Delta R.T.: 0.000 min
Response: 684683212
Conc: 206.04 ng/ml



#8 DICHLORPROP


R.T.: 8.587 min
Delta R.T.: 0.000 min
Response: 277889937
Conc: 193.74 ng/ml



#9 2,4-D


R.T.: 8.318 min
Delta R.T.: 0.000 min
Response: 713408528
Conc: 202.88 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC200




#9 2,4-D

R.T.: 8.915 min
Delta R.T.: 0.000 min
Response: 301234806
Conc: 196.16 ng/ml




#10 Pentachlorophenol

R.T.: 8.613 min
Delta R.T.: 0.000 min
Response: 10288346428
Conc: 203.72 ng/ml



#10 Pentachlorophenol


R.T.: 9.438 min
Delta R.T.: 0.000 min
Response: 4528045122
Conc: 193.00 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.189 min
Delta R.T.: 0.000 min
Response: 4036785566
Conc: 203.41 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC200




#11 2,4,5-TP (SILVEX)

R.T.: 9.814 min
Delta R.T.: 0.000 min
Response: 1826984311
Conc: 192.06 ng/ml




#12 2,4,5-T

R.T.: 9.482 min
Delta R.T.: 0.000 min
Response: 4028786566
Conc: 202.83 ng/ml



#12 2,4,5-T


R.T.: 10.231 min
Delta R.T.: 0.000 min
Response: 1754365931
Conc: 192.50 ng/ml



#13 2,4-DB


R.T.: 10.053 min
Delta R.T.: 0.000 min
Response: 727478115
Conc: 199.80 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC200




#13 2,4-DB

R.T.: 10.796 min
Delta R.T.: 0.000 min
Response: 194762268
Conc: 193.30 ng/ml




#14 DINOSEB

R.T.: 11.255 min
Delta R.T.: 0.000 min
Response: 3520935983
Conc: 203.93 ng/ml



#14 DINOSEB


R.T.: 11.173 min
Delta R.T.: 0.000 min
Response: 1259565479
Conc: 192.83 ng/ml



#15 Picloram


R.T.: 11.067 min
Delta R.T.: 0.000 min
Response: 6433530937
Conc: 198.93 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC200




#15 Picloram

R.T.: 12.257 min
Delta R.T.: 0.000 min
Response: 2426591167
Conc: 182.72 ng/ml



#16 DCPA

R.T.: 11.550 min
Delta R.T.: 0.000 min
Response: 6140119080
Conc: 206.13 ng/ml



#16 DCPA

R.T.: 12.211 min
Delta R.T.: 0.000 min
Response: 2196117913
Conc: 192.33 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
 Data File : PS028902.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2025 10:55
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 11:42:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 11:42:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.198	7.678	1383.1E6	551.8E6	482.190	488.540
Target Compounds						
1) T Dalapon	2.616	2.669	1356.2E6	921.3E6	452.399	448.343
2) T 3,5-DICHL...	6.375	6.643	1853.9E6	763.6E6	451.633	456.823
3) T 4-Nitroph...	6.996	7.207	802.6E6	399.5E6	444.483	442.929
5) T DICAMBA	7.383	7.874	5578.8E6	2603.0E6	461.683	468.364
6) T MCPP	7.563	7.977	325.0E6	141.9E6	48.778	47.471
7) T MCPA	7.712	8.218	456.4E6	195.8E6	46.443	46.130
8) T DICHLORPROP	8.087	8.586	1480.7E6	653.3E6	453.438	460.215
9) T 2,4-D	8.316	8.913	1592.9E6	698.8E6	458.530	459.912
10) T Pentachlo...	8.613	9.436	23342.5E6	11116.8E6	466.389	474.216
11) T 2,4,5-TP ...	9.188	9.813	9128.5E6	4474.4E6	464.881	471.911
12) T 2,4,5-T	9.479	10.230	9171.2E6	4277.0E6	466.059	471.184
13) T 2,4-DB	10.049	10.794	1686.6E6	467.9E6	467.082	467.889
14) T DINOSEB	11.253	11.172	7785.5E6	3019.8E6	457.109	464.855
15) T Picloram	11.064	12.255	15027.1E6	6370.9E6	468.044	478.143
16) T DCPA	11.548	12.210	13845.9E6	5457.5E6	469.769	478.637



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
Data File : PS028902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2025 10:55
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 11:42:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 11:42:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.616 min
Delta R.T.: 0.000 min
Response: 1356183583
Conc: 452.40 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC500




#1 Dalapon

R.T.: 2.669 min
Delta R.T.: 0.000 min
Response: 921269038
Conc: 448.34 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.375 min
Delta R.T.: 0.000 min
Response: 1853941228
Conc: 451.63 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.643 min
Delta R.T.: 0.000 min
Response: 763637203
Conc: 456.82 ng/ml



#3 4-Nitrophenol


R.T.: 6.996 min
Delta R.T.: 0.000 min
Response: 802588060
Conc: 444.48 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC500




#3 4-Nitrophenol

R.T.: 7.207 min
Delta R.T.: 0.000 min
Response: 399515001
Conc: 442.93 ng/ml




#4 2,4-DCAA

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 1383102910
Conc: 482.19 ng/ml



#4 2,4-DCAA


R.T.: 7.678 min
Delta R.T.: 0.000 min
Response: 551807225
Conc: 488.54 ng/ml



#5 DICAMBA


R.T.: 7.383 min
Delta R.T.: 0.000 min
Response: 5578793060
Conc: 461.68 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC500




#5 DICAMBA

R.T.: 7.874 min
Delta R.T.: 0.000 min
Response: 2603003047
Conc: 468.36 ng/ml




#6 MCPP

R.T.: 7.563 min
Delta R.T.: 0.000 min
Response: 324983021
Conc: 48.78 ug/ml



#6 MCPP


R.T.: 7.977 min
Delta R.T.: 0.000 min
Response: 141886506
Conc: 47.47 ug/ml



#7 MCPA


R.T.: 7.712 min
Delta R.T.: 0.000 min
Response: 456415243
Conc: 46.44 ug/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC500




#7 MCPA

R.T.: 8.218 min
Delta R.T.: 0.000 min
Response: 195751789
Conc: 46.13 ug/ml




#8 DICHLORPROP

R.T.: 8.087 min
Delta R.T.: 0.000 min
Response: 1480729748
Conc: 453.44 ng/ml



#8 DICHLORPROP


R.T.: 8.586 min
Delta R.T.: 0.000 min
Response: 653308254
Conc: 460.22 ng/ml



#9 2,4-D


R.T.: 8.316 min
Delta R.T.: 0.000 min
Response: 1592927104
Conc: 458.53 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC500




#9 2,4-D

R.T.: 8.913 min
Delta R.T.: 0.000 min
Response: 698750865
Conc: 459.91 ng/ml




#10 Pentachlorophenol

R.T.: 8.613 min
Delta R.T.: 0.000 min
Response: 23342479435
Conc: 466.39 ng/ml



#10 Pentachlorophenol


R.T.: 9.436 min
Delta R.T.: 0.000 min
Response: 11116848551
Conc: 474.22 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.188 min
Delta R.T.: 0.000 min
Response: 9128474142
Conc: 464.88 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC500




#11 2,4,5-TP (SILVEX)

R.T.: 9.813 min
Delta R.T.: 0.000 min
Response: 4474439764
Conc: 471.91 ng/ml




#12 2,4,5-T

R.T.: 9.479 min
Delta R.T.: 0.000 min
Response: 9171175170
Conc: 466.06 ng/ml



#12 2,4,5-T


R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 4276992400
Conc: 471.18 ng/ml



#13 2,4-DB


R.T.: 10.049 min
Delta R.T.: 0.000 min
Response: 1686623050
Conc: 467.08 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC500




#13 2,4-DB

R.T.: 10.794 min
Delta R.T.: 0.000 min
Response: 467932817
Conc: 467.89 ng/ml




#14 DINOSEB

R.T.: 11.253 min
Delta R.T.: 0.000 min
Response: 7785456867
Conc: 457.11 ng/ml



#14 DINOSEB


R.T.: 11.172 min
Delta R.T.: 0.000 min
Response: 3019847995
Conc: 464.85 ng/ml



#15 Picloram


R.T.: 11.064 min
Delta R.T.: 0.000 min
Response: 15027107162
Conc: 468.04 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC500




#15 Picloram

R.T.: 12.255 min
Delta R.T.: 0.000 min
Response: 6370927225
Conc: 478.14 ng/ml



#16 DCPA

R.T.: 11.548 min
Delta R.T.: 0.000 min
Response: 13845934631
Conc: 469.77 ng/ml



#16 DCPA

R.T.: 12.210 min
Delta R.T.: 0.000 min
Response: 5457469625
Conc: 478.64 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
 Data File : PS028903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2025 11:19
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 11:36:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 11:36:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.197	7.677	1994.8E6	821.5E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.615	2.668	2018.7E6	1376.7E6	682.500	682.500
2) T	3,5-DICHL...	6.374	6.643	2668.0E6	1137.6E6	697.500	697.500
3) T	4-Nitroph...	6.995	7.207	1173.1E6	593.9E6	682.500	682.500
5) T	DICAMBA	7.383	7.874	8146.7E6	3968.0E6	705.000	705.000
6) T	MCPD	7.565	7.979	508.7E6	215.7E6	70.500	70.500
7) T	MCPA	7.714	8.220	694.1E6	295.2E6	69.750	69.750
8) T	DICHLORPROP	8.087	8.586	2118.0E6	980.3E6	705.000	705.000
9) T	2,4-D	8.316	8.913	2282.8E6	1035.6E6	705.000	705.000
10) T	Pentachlo...	8.613	9.436	33385.4E6	16453.1E6	712.500	712.500
11) T	2,4,5-TP ...	9.189	9.813	13141.6E6	6703.9E6	712.500	712.500
12) T	2,4,5-T	9.479	10.230	13197.3E6	6408.0E6	712.500	712.500
13) T	2,4-DB	10.050	10.794	2460.5E6	705.4E6	712.500	712.500
14) T	DINOSEB	11.253	11.172	11140.9E6	4486.6E6	705.000	705.000
15) T	Picloram	11.064	12.255	21960.5E6	9824.6E6	712.500	712.500
16) T	DCPA	11.549	12.210	19869.3E6	8206.9E6	720.000	720.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
Data File : PS028903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2025 11:19
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 11:36:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 11:36:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.615 min
Delta R.T.: 0.000 min
Response: 2018665694
Conc: 682.50 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC750




#1 Dalapon

R.T.: 2.668 min
Delta R.T.: 0.000 min
Response: 1376681116
Conc: 682.50 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.374 min
Delta R.T.: 0.000 min
Response: 2667965312
Conc: 697.50 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.643 min
Delta R.T.: 0.000 min
Response: 1137550883
Conc: 697.50 ng/ml



#3 4-Nitrophenol


R.T.: 6.995 min
Delta R.T.: 0.000 min
Response: 1173142744
Conc: 682.50 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC750




#3 4-Nitrophenol

R.T.: 7.207 min
Delta R.T.: 0.000 min
Response: 593851919
Conc: 682.50 ng/ml




#4 2,4-DCAA

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 1994777262
Conc: 750.00 ng/ml



#4 2,4-DCAA


R.T.: 7.677 min
Delta R.T.: 0.000 min
Response: 821510640
Conc: 750.00 ng/ml



#5 DICAMBA


R.T.: 7.383 min
Delta R.T.: 0.000 min
Response: 8146681260
Conc: 705.00 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC750




#5 DICAMBA

R.T.: 7.874 min
Delta R.T.: 0.000 min
Response: 3967991703
Conc: 705.00 ng/ml




#6 MCPP

R.T.: 7.565 min
Delta R.T.: 0.000 min
Response: 508706510
Conc: 70.50 ug/ml



#6 MCPP


R.T.: 7.979 min
Delta R.T.: 0.000 min
Response: 215723572
Conc: 70.50 ug/ml



#7 MCPA


R.T.: 7.714 min
Delta R.T.: 0.000 min
Response: 694066365
Conc: 69.75 ug/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC750




#7 MCPA

R.T.: 8.220 min
Delta R.T.: 0.000 min
Response: 295249035
Conc: 69.75 ug/ml




#8 DICHLORPROP

R.T.: 8.087 min
Delta R.T.: 0.000 min
Response: 2118008436
Conc: 705.00 ng/ml



#8 DICHLORPROP


R.T.: 8.586 min
Delta R.T.: 0.000 min
Response: 980342169
Conc: 705.00 ng/ml



#9 2,4-D


R.T.: 8.316 min
Delta R.T.: 0.000 min
Response: 2282808841
Conc: 705.00 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC750




#9 2,4-D

R.T.: 8.913 min
Delta R.T.: 0.000 min
Response: 1035595119
Conc: 705.00 ng/ml




#10 Pentachlorophenol

R.T.: 8.613 min
Delta R.T.: 0.000 min
Response: 33385448783
Conc: 712.50 ng/ml



#10 Pentachlorophenol


R.T.: 9.436 min
Delta R.T.: 0.000 min
Response: 16453088564
Conc: 712.50 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.189 min
Delta R.T.: 0.000 min
Response: 13141574024
Conc: 712.50 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC750




#11 2,4,5-TP (SILVEX)

R.T.: 9.813 min
Delta R.T.: 0.000 min
Response: 6703920625
Conc: 712.50 ng/ml




#12 2,4,5-T

R.T.: 9.479 min
Delta R.T.: 0.000 min
Response: 13197337017
Conc: 712.50 ng/ml



#12 2,4,5-T


R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 6407989455
Conc: 712.50 ng/ml



#13 2,4-DB


R.T.: 10.050 min
Delta R.T.: 0.000 min
Response: 2460493983
Conc: 712.50 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC750




#13 2,4-DB

R.T.: 10.794 min
Delta R.T.: 0.000 min
Response: 705441852
Conc: 712.50 ng/ml




#14 DINOSEB

R.T.: 11.253 min
Delta R.T.: 0.000 min
Response: 11140909547
Conc: 705.00 ng/ml



#14 DINOSEB


R.T.: 11.172 min
Delta R.T.: 0.000 min
Response: 4486583360
Conc: 705.00 ng/ml



#15 Picloram


R.T.: 11.064 min
Delta R.T.: 0.000 min
Response: 21960519904
Conc: 712.50 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC750




#15 Picloram

R.T.: 12.255 min
Delta R.T.: 0.000 min
Response: 9824620857
Conc: 712.50 ng/ml



#16 DCPA

R.T.: 11.549 min
Delta R.T.: 0.000 min
Response: 19869334736
Conc: 720.00 ng/ml



#16 DCPA

R.T.: 12.210 min
Delta R.T.: 0.000 min
Response: 8206878898
Conc: 720.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
 Data File : PS028904.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2025 11:43
 Operator : AR\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 12:11:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:11:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.198	7.678	2530.9E6	1074.7E6	909.092	963.192
Target Compounds						
1) T Dalapon	2.615	2.667	2669.7E6	1816.5E6	895.337	890.355
2) T 3,5-DICHL...	6.375	6.643	3415.4E6	1483.9E6	854.524	897.916
3) T 4-Nitroph...	6.996	7.207	1520.9E6	776.4E6	858.259	872.537
5) T DICAMBA	7.384	7.875	10523.3E6	5267.0E6	887.185	945.768
6) T MCPP	7.567	7.981	684.1E6	288.1E6	100.360	95.785
7) T MCPA	7.717	8.223	922.8E6	396.1E6	93.671	93.262
8) T DICHLORPROP	8.088	8.586	2704.9E6	1281.9E6	853.679	911.969
9) T 2,4-D	8.317	8.914	2910.1E6	1353.7E6	861.114	902.769
10) T Pentachlo...	8.614	9.437	40660.1E6	21219.5E6	842.923	915.975
11) T 2,4,5-TP ...	9.189	9.813	16741.1E6	8771.4E6	875.003	931.202
12) T 2,4,5-T	9.480	10.230	16865.6E6	8364.4E6	878.556	928.451
13) T 2,4-DB	10.051	10.795	3188.7E6	933.6E6	898.882	937.587
14) T DINOSEB	11.255	11.172	14187.7E6	5809.5E6	857.405	905.282
15) T Picloram	11.064	12.256	28395.6E6	13020.4E6	899.958	970.250
16) T DCPA	11.550	12.211	25250.3E6	10755.0E6	880.382	947.380



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
Data File : PS028904.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2025 11:43
Operator : AR\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 12:11:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:11:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.615 min
Delta R.T.: 0.000 min
Response: 2669664670
Conc: 895.34 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000




#1 Dalapon

R.T.: 2.667 min
Delta R.T.: 0.000 min
Response: 1816455493
Conc: 890.35 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.375 min
Delta R.T.: 0.000 min
Response: 3415406968
Conc: 854.52 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.643 min
Delta R.T.: 0.000 min
Response: 1483913982
Conc: 897.92 ng/ml



#3 4-Nitrophenol


R.T.: 6.996 min
Delta R.T.: 0.000 min
Response: 1520903644
Conc: 858.26 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000




#3 4-Nitrophenol

R.T.: 7.207 min
Delta R.T.: 0.000 min
Response: 776361144
Conc: 872.54 ng/ml




#4 2,4-DCAA

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 2530922607
Conc: 909.09 ng/ml



#4 2,4-DCAA


R.T.: 7.678 min
Delta R.T.: 0.000 min
Response: 1074742027
Conc: 963.19 ng/ml



#5 DICAMBA


R.T.: 7.384 min
Delta R.T.: 0.000 min
Response: 10523289784
Conc: 887.18 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000




#5 DICAMBA

R.T.: 7.875 min
Delta R.T.: 0.000 min
Response: 5267015305
Conc: 945.77 ng/ml




#6 MCPP

R.T.: 7.567 min
Delta R.T.: 0.000 min
Response: 684083315
Conc: 100.36 ug/ml



#6 MCPP


R.T.: 7.981 min
Delta R.T.: 0.000 min
Response: 288116567
Conc: 95.78 ug/ml



#7 MCPA


R.T.: 7.717 min
Delta R.T.: 0.000 min
Response: 922758739
Conc: 93.67 ug/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000




#7 MCPA

R.T.: 8.223 min
Delta R.T.: 0.000 min
Response: 396124832
Conc: 93.26 ug/ml




#8 DICHLORPROP

R.T.: 8.088 min
Delta R.T.: 0.000 min
Response: 2704941406
Conc: 853.68 ng/ml



#8 DICHLORPROP


R.T.: 8.586 min
Delta R.T.: 0.000 min
Response: 1281862566
Conc: 911.97 ng/ml



#9 2,4-D


R.T.: 8.317 min
Delta R.T.: 0.000 min
Response: 2910090795
Conc: 861.11 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000




#9 2,4-D

R.T.: 8.914 min
Delta R.T.: 0.000 min
Response: 1353717959
Conc: 902.77 ng/ml




#10 Pentachlorophenol

R.T.: 8.614 min
Delta R.T.: 0.000 min
Response: 40660110132
Conc: 842.92 ng/ml



#10 Pentachlorophenol


R.T.: 9.437 min
Delta R.T.: 0.000 min
Response: 21219501611
Conc: 915.98 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.189 min
Delta R.T.: 0.000 min
Response: 16741146063
Conc: 875.00 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000




#11 2,4,5-TP (SILVEX)

R.T.: 9.813 min
Delta R.T.: 0.000 min
Response: 8771366125
Conc: 931.20 ng/ml




#12 2,4,5-T

R.T.: 9.480 min
Delta R.T.: 0.000 min
Response: 16865581213
Conc: 878.56 ng/ml



#12 2,4,5-T


R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 8364414839
Conc: 928.45 ng/ml



#13 2,4-DB


R.T.: 10.051 min
Delta R.T.: 0.000 min
Response: 3188652966
Conc: 898.88 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000




#13 2,4-DB

R.T.: 10.795 min
Delta R.T.: 0.000 min
Response: 933607896
Conc: 937.59 ng/ml




#14 DINOSEB

R.T.: 11.255 min
Delta R.T.: 0.000 min
Response: 14187746813
Conc: 857.41 ng/ml



#14 DINOSEB


R.T.: 11.172 min
Delta R.T.: 0.000 min
Response: 5809479977
Conc: 905.28 ng/ml



#15 Picloram


R.T.: 11.064 min
Delta R.T.: 0.000 min
Response: 28395601774
Conc: 899.96 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000




#15 Picloram

R.T.: 12.256 min
Delta R.T.: 0.000 min
Response: 13020427725
Conc: 970.25 ng/ml



#16 DCPA

R.T.: 11.550 min
Delta R.T.: 0.000 min
Response: 25250250867
Conc: 880.38 ng/ml



#16 DCPA

R.T.: 12.211 min
Delta R.T.: 0.000 min
Response: 10754991695
Conc: 947.38 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
 Data File : PS028905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2025 12:07
 Operator : AR\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 01/14/2025
 Supervised By :Ankita Jodhani 01/15/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 12:24:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:24:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.198	7.678	3620.6E6	1605.1E6	1336.050	1450.411
Target Compounds						
1) T Dalapon	2.615	2.668	4104.6E6	2732.3E6	1373.858m	1344.349
2) T 3,5-DICHL...	6.375	6.643	4927.4E6	2207.8E6	1262.171	1347.366
3) T 4-Nitroph...	6.996	7.207	2257.5E6	1158.9E6	1291.156	1314.550
5) T DICAMBA	7.384	7.875	15288.4E6	7923.7E6	1311.442	1420.235
6) T MCPP	7.571	7.984	1068.2E6	439.2E6	153.296	144.975
7) T MCPA	7.722	8.228	1403.3E6	594.6E6	141.855	139.892
8) T DICHLORPROP	8.088	8.586	3907.5E6	1911.2E6	1264.916	1369.450
9) T 2,4-D	8.317	8.913	4184.2E6	2015.2E6	1269.064	1356.640
10) T Pentachlo...	8.619	9.436	48590.7E6	30432.8E6	1070.058	1334.532
11) T 2,4,5-TP ...	9.190	9.813	23808.0E6	12847.4E6	1276.732	1375.721
12) T 2,4,5-T	9.480	10.230	23961.1E6	12280.1E6	1279.940	1375.041
13) T 2,4-DB	10.050	10.794	4659.8E6	1410.4E6	1334.470	1418.157
14) T DINOSEB	11.255	11.172	20299.9E6	8571.4E6	1259.511	1349.897
15) T Picloram	11.064	12.255	40951.2E6	19407.1E6	1321.466	1441.886
16) T DCPA	11.550	12.210	35612.9E6	15746.0E6	1276.857	1397.307

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
Data File : PS028905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2025 12:07
Operator : AR\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
HSTDICC1500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 12:24:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:24:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 2.615 min
Delta R.T.: 0.000 min
Response: 4104630171
Conc: 1373.86 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#1 Dalapon

R.T.: 2.668 min
Delta R.T.: 0.000 min
Response: 2732337433
Conc: 1344.35 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.375 min
Delta R.T.: 0.000 min
Response: 4927415110
Conc: 1262.17 ng/ml



#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.643 min
Delta R.T.: 0.000 min
Response: 2207837002
Conc: 1347.37 ng/ml




#3 4-Nitrophenol

R.T.: 6.996 min
Delta R.T.: 0.000 min
Response: 2257498653
Conc: 1291.16 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#3 4-Nitrophenol

R.T.: 7.207 min
Delta R.T.: 0.000 min
Response: 1158944307
Conc: 1314.55 ng/ml




#4 2,4-DCAA

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 3620644236
Conc: 1336.05 ng/ml



#4 2,4-DCAA

R.T.: 7.678 min
Delta R.T.: 0.000 min
Response: 1605121752
Conc: 1450.41 ng/ml




#5 DICAMBA

R.T.: 7.384 min
Delta R.T.: 0.000 min
Response: 15288428327
Conc: 1311.44 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#5 DICAMBA

R.T.: 7.875 min
Delta R.T.: 0.000 min
Response: 7923720175
Conc: 1420.23 ng/ml




#6 MCPP

R.T.: 7.571 min
Delta R.T.: 0.000 min
Response: 1068198576
Conc: 153.30 ug/ml



#6 MCPP

R.T.: 7.984 min
Delta R.T.: 0.000 min
Response: 439172090
Conc: 144.97 ug/ml




#7 MCPA

R.T.: 7.722 min
Delta R.T.: 0.000 min
Response: 1403348267
Conc: 141.86 ug/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#7 MCPA

R.T.: 8.228 min
Delta R.T.: 0.000 min
Response: 594601475
Conc: 139.89 ug/ml




#8 DICHLORPROP

R.T.: 8.088 min
Delta R.T.: 0.000 min
Response: 3907459694
Conc: 1264.92 ng/ml



#8 DICHLORPROP

R.T.: 8.586 min
Delta R.T.: 0.000 min
Response: 1911156273
Conc: 1369.45 ng/ml




#9 2,4-D

R.T.: 8.317 min
Delta R.T.: 0.000 min
Response: 4184181925
Conc: 1269.06 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#9 2,4-D

R.T.: 8.913 min
Delta R.T.: 0.000 min
Response: 2015240589
Conc: 1356.64 ng/ml




#10 Pentachlorophenol

R.T.: 8.619 min
Delta R.T.: 0.000 min
Response: 48590682069
Conc: 1070.06 ng/ml



#10 Pentachlorophenol

R.T.: 9.436 min
Delta R.T.: 0.000 min
Response: 30432768138
Conc: 1334.53 ng/ml



#11 2,4,5-TP (SILVEX)

R.T.: 9.190 min
Delta R.T.: 0.000 min
Response: 23807983939
Conc: 1276.73 ng/ml

Instrument :

ECD_S


ClientSampleId :

HSTDICC1500

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#11 2,4,5-TP (SILVEX)

R.T.: 9.813 min
Delta R.T.: 0.000 min
Response: 12847398816
Conc: 1375.72 ng/ml




#12 2,4,5-T

R.T.: 9.480 min
Delta R.T.: 0.000 min
Response: 23961133423
Conc: 1279.94 ng/ml



#12 2,4,5-T

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 12280117089
Conc: 1375.04 ng/ml




#13 2,4-DB

R.T.: 10.050 min
Delta R.T.: 0.000 min
Response: 4659827211
Conc: 1334.47 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#13 2,4-DB

R.T.: 10.794 min
Delta R.T.: 0.000 min
Response: 1410444919
Conc: 1418.16 ng/ml




#14 DINOSEB

R.T.: 11.255 min
Delta R.T.: 0.000 min
Response: 20299877375
Conc: 1259.51 ng/ml



#14 DINOSEB

R.T.: 11.172 min
Delta R.T.: 0.000 min
Response: 8571373262
Conc: 1349.90 ng/ml




#15 Picloram

R.T.: 11.064 min
Delta R.T.: 0.000 min
Response: 40951242678
Conc: 1321.47 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#15 Picloram

R.T.: 12.255 min
Delta R.T.: 0.000 min
Response: 19407128630
Conc: 1441.89 ng/ml



#16 DCPA

R.T.: 11.550 min
Delta R.T.: 0.000 min
Response: 35612880706
Conc: 1276.86 ng/ml



#16 DCPA

R.T.: 12.210 min
Delta R.T.: 0.000 min
Response: 15746015044
Conc: 1397.31 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
 Data File : PS028906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2025 12:31
 Operator : AR\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

ICVPS011425

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 01/14/2025

Supervised By :Ankita Jodhani 01/15/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 12:50:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.198	7.677	1954.1E6	815.5E6	701.913m	730.872
Target Compounds						
1) T Dalapon	2.617	2.667	2021.2E6	1368.1E6	677.864m	670.594
2) T 3,5-DICHL...	6.375	6.643	2643.0E6	1126.0E6	661.269	681.365
3) T 4-Nitroph...	6.996	7.207	1153.1E6	589.2E6	650.699	662.186
5) T DICAMBA	7.383	7.874	8094.6E6	3936.5E6	682.430	706.852
6) T MCPP	7.565	7.979	500.5E6	215.1E6	73.421	71.515
7) T MCPA	7.714	8.220	683.4E6	294.4E6	69.370	69.308
8) T DICHLORPROP	8.087	8.586	2101.2E6	970.0E6	663.140	690.108
9) T 2,4-D	8.317	8.913	2261.2E6	1032.1E6	669.101	688.263
10) T Pentachlo...	8.613	9.436	33194.1E6	16359.2E6	688.146	706.173
11) T 2,4,5-TP ...	9.189	9.813	13049.3E6	6664.0E6	682.045	707.481
12) T 2,4,5-T	9.480	10.230	13104.2E6	6350.6E6	682.619	704.921
13) T 2,4-DB	10.050	10.794	2422.6E6	699.4E6	682.922	702.401
14) T DINOSEB	11.254	11.172	10976.0E6	4397.4E6	663.311	685.233
15) T Picloram	11.064	12.255	21733.8E6	9689.5E6	688.822	722.034
16) T DCPA	11.549	12.209	19741.5E6	8153.3E6	688.312	718.207

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
Data File : PS028906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2025 12:31
Operator : AR\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

ICVPS011425

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 01/14/2025

Supervised By :Ankita Jodhani 01/15/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 12:50:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 2.617 min
Delta R.T.: 0.002 min
Response: 2021216536
Conc: 677.86 ng/ml

Instrument :
ECD_S
ClientSampleId :
ICVPS011425


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#1 Dalapon

R.T.: 2.667 min
Delta R.T.: 0.000 min
Response: 1368111425
Conc: 670.59 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.375 min
Delta R.T.: 0.000 min
Response: 2642993373
Conc: 661.27 ng/ml



#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.643 min
Delta R.T.: 0.000 min
Response: 1126037759
Conc: 681.37 ng/ml



#3 4-Nitrophenol

R.T.: 6.996 min
Delta R.T.: 0.000 min
Response: 1153090162
Conc: 650.70 ng/ml


Instrument :

ECD_S

ClientSampleId :
ICVPS011425


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#3 4-Nitrophenol

R.T.: 7.207 min
Delta R.T.: 0.000 min
Response: 589196510
Conc: 662.19 ng/ml




#4 2,4-DCAA

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 1954135705
Conc: 701.91 ng/ml m



#4 2,4-DCAA

R.T.: 7.677 min
Delta R.T.: 0.000 min
Response: 815516416
Conc: 730.87 ng/ml




#5 DICAMBA

R.T.: 7.383 min
Delta R.T.: 0.000 min
Response: 8094599008
Conc: 682.43 ng/ml

Instrument :
ECD_S
ClientSampleId :
ICVPS011425


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#5 DICAMBA

R.T.: 7.874 min
Delta R.T.: 0.000 min
Response: 3936487030
Conc: 706.85 ng/ml




#6 MCPP

R.T.: 7.565 min
Delta R.T.: -0.002 min
Response: 500456855
Conc: 73.42 ug/ml



#6 MCPP

R.T.: 7.979 min
Delta R.T.: -0.002 min
Response: 215113889
Conc: 71.51 ug/ml




#7 MCPA

R.T.: 7.714 min
Delta R.T.: -0.003 min
Response: 683363222
Conc: 69.37 ug/ml

Instrument :
ECD_S
ClientSampleId :
ICVPS011425


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#7 MCPA

R.T.: 8.220 min
Delta R.T.: -0.003 min
Response: 294383936
Conc: 69.31 ug/ml




#8 DICHLORPROP

R.T.: 8.087 min
Delta R.T.: 0.000 min
Response: 2101207308
Conc: 663.14 ng/ml



#8 DICHLORPROP

R.T.: 8.586 min
Delta R.T.: 0.000 min
Response: 970014347
Conc: 690.11 ng/ml




#9 2,4-D

R.T.: 8.317 min
Delta R.T.: 0.000 min
Response: 2261193949
Conc: 669.10 ng/ml

Instrument :
ECD_S
ClientSampleId :
ICVPS011425


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#9 2,4-D

R.T.: 8.913 min
Delta R.T.: 0.000 min
Response: 1032063315
Conc: 688.26 ng/ml




#10 Pentachlorophenol

R.T.: 8.613 min
Delta R.T.: 0.000 min
Response: 33194118867
Conc: 688.15 ng/ml



#10 Pentachlorophenol

R.T.: 9.436 min
Delta R.T.: 0.000 min
Response: 16359225437
Conc: 706.17 ng/ml



#11 2,4,5-TP (SILVEX)

R.T.: 9.189 min
Delta R.T.: 0.000 min
Response: 13049329632
Conc: 682.04 ng/ml

Instrument :

ECD_S


ClientSampleId :

ICVPS011425

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#11 2,4,5-TP (SILVEX)

R.T.: 9.813 min
Delta R.T.: 0.000 min
Response: 6664044603
Conc: 707.48 ng/ml




#12 2,4,5-T

R.T.: 9.480 min
Delta R.T.: 0.000 min
Response: 13104177427
Conc: 682.62 ng/ml



#12 2,4,5-T

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 6350637897
Conc: 704.92 ng/ml



#13 2,4-DB

R.T.: 10.050 min
Delta R.T.: 0.000 min
Response: 2422567420
Conc: 682.92 ng/ml

Instrument :


ECD_S

ClientSampleId :

ICVPS011425


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#13 2,4-DB

R.T.: 10.794 min
Delta R.T.: 0.000 min
Response: 699420411
Conc: 702.40 ng/ml




#14 DINOSEB

R.T.: 11.254 min
Delta R.T.: 0.000 min
Response: 10976006715
Conc: 663.31 ng/ml



#14 DINOSEB

R.T.: 11.172 min
Delta R.T.: 0.000 min
Response: 4397359740
Conc: 685.23 ng/ml




#15 Picloram

R.T.: 11.064 min
Delta R.T.: 0.000 min
Response: 21733814083
Conc: 688.82 ng/ml

Instrument :
ECD_S
ClientSampleId :
ICVPS011425


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/14/2025
Supervised By :Ankita Jodhani 01/15/2025




#15 Picloram

R.T.: 12.255 min
Delta R.T.: -0.001 min
Response: 9689461634
Conc: 722.03 ng/ml



#16 DCPA

R.T.: 11.549 min
Delta R.T.: 0.000 min
Response: 19741503590
Conc: 688.31 ng/ml



#16 DCPA

R.T.: 12.209 min
Delta R.T.: -0.001 min
Response: 8153339758
Conc: 718.21 ng/ml

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Continuing Calib Date: 01/30/2025 **Initial Calibration Date(s):** 01/14/2025 01/14/2025

Continuing Calib Time: 13:20 **Initial Calibration Time(s):** 10:31 12:07

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.19	7.20	7.10	7.30	0.01
2,4-D	8.31	8.32	8.22	8.42	0.01
2,4,5-TP(Silvex)	9.18	9.19	9.09	9.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/14/2025 01/14/2025

Continuing Calib Time: 13:20 Initial Calibration Time(s): 10:31 12:07

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
2,4-DCAA	7.67	7.68	7.58	7.78	0.01
2,4-D	8.91	8.91	8.81	9.01	0.01
2,4,5-TP(Silvex)	9.80	9.81	9.71	9.91	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 01/30/2025

Lab Sample No.: HSTDCCC750 Data File : PS028989.D Time Analyzed: 13:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.181	9.089	9.289	750.680	712.500	5.4
2,4-D	8.310	8.216	8.416	735.240	705.000	4.3
2,4-DCAA	7.192	7.097	7.297	777.790	750.000	3.7

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 01/30/2025

Lab Sample No.: HSTDCCC750 Data File : PS028989.D Time Analyzed: 13:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.803	9.713	9.913	766.570	712.500	7.6
2,4-D	8.905	8.813	9.013	711.080	705.000	0.9
2,4-DCAA	7.670	7.577	7.777	736.960	750.000	-1.7

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
 Data File : PS028989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2025 13:20
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 05:20:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.192	7.670	2165.4E6	822.3E6	777.786	736.959
Target Compounds							
1) T	Dalapon	2.613	2.665	2223.3E6	1302.6E6	745.627	638.476
2) T	3,5-DICHL...	6.370	6.637	2940.8E6	1137.4E6	735.774	688.221
3) T	4-Nitroph...	6.990	7.200	1301.1E6	634.4E6	734.214	713.021
5) T	DICAMBA	7.377	7.867	9036.6E6	4147.8E6	761.847	744.805
6) T	MCPD	7.560	7.972	530.1E6	205.7E6	77.774	68.381
7) T	MCPA	7.708	8.213	755.5E6	281.9E6	76.696	66.376
8) T	DICHLORPROP	8.081	8.578	2319.8E6	1009.7E6	732.116	718.335
9) T	2,4-D	8.310	8.905	2484.7E6	1066.3E6	735.236	711.082
10) T	Pentachlo...	8.606	9.427	36863.8E6	17810.8E6	764.221	768.832
11) T	2,4,5-TP ...	9.181	9.803	14362.5E6	7220.6E6	750.677	766.568
12) T	2,4,5-T	9.472	10.219	14459.2E6	6819.7E6	753.206	756.985
13) T	2,4-DB	10.042	10.784	2641.6E6	719.5E6	744.655	722.531
14) T	DINOSEB	11.245	11.162	12050.0E6	4752.1E6	728.215	740.508
15) T	Picloram	11.055	12.243	23546.5E6	10381.1E6	746.273	773.575
16) T	DCPA	11.540	12.199	21648.2E6	9068.6E6	754.790	798.830



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 13:20
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:20:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.613 min
Delta R.T.: -0.002 min
Response: 2223269315
Conc: 745.63 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#1 Dalapon

R.T.: 2.665 min
Delta R.T.: -0.002 min
Response: 1302584822
Conc: 638.48 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.370 min
Delta R.T.: -0.005 min
Response: 2940779728
Conc: 735.77 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.637 min
Delta R.T.: -0.006 min
Response: 1137367093
Conc: 688.22 ng/ml



#3 4-Nitrophenol


R.T.: 6.990 min
Delta R.T.: -0.005 min
Response: 1301086368
Conc: 734.21 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#3 4-Nitrophenol

R.T.: 7.200 min
Delta R.T.: -0.007 min
Response: 634428173
Conc: 713.02 ng/ml




#4 2,4-DCAA

R.T.: 7.192 min
Delta R.T.: -0.006 min
Response: 2165367099
Conc: 777.79 ng/ml



#4 2,4-DCAA


R.T.: 7.670 min
Delta R.T.: -0.007 min
Response: 822308544
Conc: 736.96 ng/ml



#5 DICAMBA


R.T.: 7.377 min
Delta R.T.: -0.007 min
Response: 9036598303
Conc: 761.85 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#5 DICAMBA

R.T.: 7.867 min
Delta R.T.: -0.008 min
Response: 4147846100
Conc: 744.80 ng/ml




#6 MCPP

R.T.: 7.560 min
Delta R.T.: -0.008 min
Response: 530126870
Conc: 77.77 ug/ml



#6 MCPP


R.T.: 7.972 min
Delta R.T.: -0.009 min
Response: 205687869
Conc: 68.38 ug/ml



#7 MCPA


R.T.: 7.708 min
Delta R.T.: -0.009 min
Response: 755535303
Conc: 76.70 ug/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#7 MCPA

R.T.: 8.213 min
Delta R.T.: -0.010 min
Response: 281928824
Conc: 66.38 ug/ml




#8 DICHLORPROP

R.T.: 8.081 min
Delta R.T.: -0.007 min
Response: 2319760419
Conc: 732.12 ng/ml



#8 DICHLORPROP


R.T.: 8.578 min
Delta R.T.: -0.009 min
Response: 1009690187
Conc: 718.33 ng/ml



#9 2,4-D


R.T.: 8.310 min
Delta R.T.: -0.007 min
Response: 2484691983
Conc: 735.24 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#9 2,4-D

R.T.: 8.905 min
Delta R.T.: -0.009 min
Response: 1066280668
Conc: 711.08 ng/ml




#10 Pentachlorophenol

R.T.: 8.606 min
Delta R.T.: -0.008 min
Response: 36863756070
Conc: 764.22 ng/ml



#10 Pentachlorophenol


R.T.: 9.427 min
Delta R.T.: -0.010 min
Response: 17810779530
Conc: 768.83 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.181 min
Delta R.T.: -0.008 min
Response: 14362452015
Conc: 750.68 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#11 2,4,5-TP (SILVEX)

R.T.: 9.803 min
Delta R.T.: -0.010 min
Response: 7220611854
Conc: 766.57 ng/ml




#12 2,4,5-T

R.T.: 9.472 min
Delta R.T.: -0.009 min
Response: 14459229727
Conc: 753.21 ng/ml



#12 2,4,5-T


R.T.: 10.219 min
Delta R.T.: -0.011 min
Response: 6819676901
Conc: 756.98 ng/ml



#13 2,4-DB


R.T.: 10.042 min
Delta R.T.: -0.009 min
Response: 2641554760
Conc: 744.65 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#13 2,4-DB

R.T.: 10.784 min
Delta R.T.: -0.011 min
Response: 719464913
Conc: 722.53 ng/ml




#14 DINOSEB

R.T.: 11.245 min
Delta R.T.: -0.010 min
Response: 12049997479
Conc: 728.21 ng/ml



#14 DINOSEB


R.T.: 11.162 min
Delta R.T.: -0.011 min
Response: 4752077478
Conc: 740.51 ng/ml



#15 Picloram


R.T.: 11.055 min
Delta R.T.: -0.010 min
Response: 23546531451
Conc: 746.27 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#15 Picloram

R.T.: 12.243 min
Delta R.T.: -0.013 min
Response: 10381120864
Conc: 773.58 ng/ml



#16 DCPA

R.T.: 11.540 min
Delta R.T.: -0.010 min
Response: 21648153234
Conc: 754.79 ng/ml



#16 DCPA

R.T.: 12.199 min
Delta R.T.: -0.012 min
Response: 9068598738
Conc: 798.83 ng/ml

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Continuing Calib Date: 01/30/2025 **Initial Calibration Date(s):** 01/14/2025 01/14/2025

Continuing Calib Time: 16:55 **Initial Calibration Time(s):** 10:31 12:07

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.19	7.20	7.10	7.30	0.01
2,4-D	8.31	8.32	8.22	8.42	0.01
2,4,5-TP(Silvex)	9.18	9.19	9.09	9.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Continuing Calib Date: 01/30/2025 **Initial Calibration Date(s):** 01/14/2025 01/14/2025

Continuing Calib Time: 16:55 **Initial Calibration Time(s):** 10:31 12:07

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.67	7.68	7.58	7.78	0.01
2,4-D	8.90	8.91	8.81	9.01	0.01
2,4,5-TP(Silvex)	9.80	9.81	9.71	9.91	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 01/30/2025

Lab Sample No.: HSTDCCC750 Data File : PS028998.D Time Analyzed: 16:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.181	9.089	9.289	751.030	712.500	5.4
2,4-D	8.309	8.216	8.416	735.060	705.000	4.3
2,4-DCAA	7.192	7.097	7.297	774.330	750.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 01/30/2025

Lab Sample No.: HSTDCCC750 Data File : PS028998.D Time Analyzed: 16:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.803	9.713	9.913	780.240	712.500	9.5
2,4-D	8.904	8.813	9.013	724.010	705.000	2.7
2,4-DCAA	7.670	7.577	7.777	751.950	750.000	0.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
 Data File : PS028998.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2025 16:55
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 05:22:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.192	7.670	2155.8E6	839.0E6	774.333	751.949
Target Compounds						
1) T Dalapon	2.613	2.665	2178.0E6	1297.7E6	730.445	636.100
2) T 3,5-DICHL...	6.369	6.636	2944.9E6	1158.9E6	736.796	701.259
3) T 4-Nitroph...	6.990	7.200	1291.7E6	645.7E6	728.901	725.658
5) T DICAMBA	7.377	7.866	9031.1E6	4233.1E6	761.382	760.119
6) T MCPP	7.560	7.972	527.0E6	208.6E6	77.320	69.343
7) T MCPA	7.709	8.214	760.3E6	285.2E6	77.178	67.143
8) T DICHLORPROP	8.080	8.578	2331.0E6	1036.6E6	735.656	737.506
9) T 2,4-D	8.309	8.904	2484.1E6	1085.7E6	735.060	724.014
10) T Pentachlo...	8.605	9.426	36915.3E6	18169.0E6	765.290	784.293
11) T 2,4,5-TP ...	9.181	9.803	14369.1E6	7349.4E6	751.025	780.237
12) T 2,4,5-T	9.471	10.220	14466.5E6	6929.0E6	753.587	769.118
13) T 2,4-DB	10.041	10.784	2624.5E6	732.4E6	739.859	735.489
14) T DINOSEB	11.244	11.161	11718.4E6	4648.2E6	708.177	724.327
15) T Picloram	11.054	12.242	22692.6E6	10025.2E6	719.209	747.050
16) T DCPA	11.539	12.199	21762.0E6	9251.5E6	758.759	814.940



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 16:55
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:22:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.613 min
Delta R.T.: -0.002 min
Response: 2177999111
Conc: 730.45 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#1 Dalapon

R.T.: 2.665 min
Delta R.T.: -0.002 min
Response: 1297737505
Conc: 636.10 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.369 min
Delta R.T.: -0.006 min
Response: 2944865208
Conc: 736.80 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.636 min
Delta R.T.: -0.007 min
Response: 1158914776
Conc: 701.26 ng/ml



#3 4-Nitrophenol


R.T.: 6.990 min
Delta R.T.: -0.006 min
Response: 1291669888
Conc: 728.90 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#3 4-Nitrophenol

R.T.: 7.200 min
Delta R.T.: -0.007 min
Response: 645671542
Conc: 725.66 ng/ml




#4 2,4-DCAA

R.T.: 7.192 min
Delta R.T.: -0.006 min
Response: 2155754002
Conc: 774.33 ng/ml



#4 2,4-DCAA


R.T.: 7.670 min
Delta R.T.: -0.007 min
Response: 839034493
Conc: 751.95 ng/ml



#5 DICAMBA


R.T.: 7.377 min
Delta R.T.: -0.007 min
Response: 9031091544
Conc: 761.38 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#5 DICAMBA

R.T.: 7.866 min
Delta R.T.: -0.008 min
Response: 4233132016
Conc: 760.12 ng/ml




#6 MCPP

R.T.: 7.560 min
Delta R.T.: -0.008 min
Response: 527037165
Conc: 77.32 ug/ml



#6 MCPP


R.T.: 7.972 min
Delta R.T.: -0.009 min
Response: 208579405
Conc: 69.34 ug/ml



#7 MCPA


R.T.: 7.709 min
Delta R.T.: -0.009 min
Response: 760283529
Conc: 77.18 ug/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#7 MCPA

R.T.: 8.214 min
Delta R.T.: -0.010 min
Response: 285186075
Conc: 67.14 ug/ml




#8 DICHLORPROP

R.T.: 8.080 min
Delta R.T.: -0.008 min
Response: 2330977691
Conc: 735.66 ng/ml



#8 DICHLORPROP


R.T.: 8.578 min
Delta R.T.: -0.009 min
Response: 1036638169
Conc: 737.51 ng/ml



#9 2,4-D


R.T.: 8.309 min
Delta R.T.: -0.008 min
Response: 2484098048
Conc: 735.06 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#9 2,4-D

R.T.: 8.904 min
Delta R.T.: -0.009 min
Response: 1085672146
Conc: 724.01 ng/ml




#10 Pentachlorophenol

R.T.: 8.605 min
Delta R.T.: -0.009 min
Response: 36915310142
Conc: 765.29 ng/ml



#10 Pentachlorophenol


R.T.: 9.426 min
Delta R.T.: -0.011 min
Response: 18168954826
Conc: 784.29 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.181 min
Delta R.T.: -0.008 min
Response: 14369112152
Conc: 751.03 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#11 2,4,5-TP (SILVEX)

R.T.: 9.803 min
Delta R.T.: -0.011 min
Response: 7349366909
Conc: 780.24 ng/ml




#12 2,4,5-T

R.T.: 9.471 min
Delta R.T.: -0.009 min
Response: 14466544431
Conc: 753.59 ng/ml



#12 2,4,5-T


R.T.: 10.220 min
Delta R.T.: -0.011 min
Response: 6928990880
Conc: 769.12 ng/ml



#13 2,4-DB


R.T.: 10.041 min
Delta R.T.: -0.010 min
Response: 2624541310
Conc: 739.86 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#13 2,4-DB

R.T.: 10.784 min
Delta R.T.: -0.011 min
Response: 732368095
Conc: 735.49 ng/ml




#14 DINOSEB

R.T.: 11.244 min
Delta R.T.: -0.011 min
Response: 11718424366
Conc: 708.18 ng/ml



#14 DINOSEB


R.T.: 11.161 min
Delta R.T.: -0.012 min
Response: 4648234599
Conc: 724.33 ng/ml



#15 Picloram


R.T.: 11.054 min
Delta R.T.: -0.010 min
Response: 22692603090
Conc: 719.21 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#15 Picloram

R.T.: 12.242 min
Delta R.T.: -0.013 min
Response: 10025162624
Conc: 747.05 ng/ml



#16 DCPA

R.T.: 11.539 min
Delta R.T.: -0.011 min
Response: 21761979040
Conc: 758.76 ng/ml



#16 DCPA

R.T.: 12.199 min
Delta R.T.: -0.012 min
Response: 9251482356
Conc: 814.94 ng/ml

Analytical Sequence

Client: RU2 Engineering, LLC	SDG No.: Q1206
Project: NYCDDC SANTWOBR Brooklyn Bridge BI	Instrument ID: ECD_S
GC Column: RTX-CLP	ID: 0.32 (mm) Inst. Calib. Date(s): 01/14/2025 01/14/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCAA RT #	RT #
IBLK	IBLK	01/14/2025	10:07	PS028900.D	7.20	0.00
HSTDICC200	HSTDICC200	01/14/2025	10:31	PS028901.D	7.20	0.00
HSTDICC500	HSTDICC500	01/14/2025	10:55	PS028902.D	7.20	0.00
HSTDICC750	HSTDICC750	01/14/2025	11:19	PS028903.D	7.20	0.00
HSTDICC1000	HSTDICC1000	01/14/2025	11:43	PS028904.D	7.20	0.00
HSTDICC1500	HSTDICC1500	01/14/2025	12:07	PS028905.D	7.20	0.00
IBLK	IBLK	01/30/2025	12:56	PS028988.D	7.19	0.00
HSTDCCC750	HSTDCCC750	01/30/2025	13:20	PS028989.D	7.19	0.00
PB166382BL	PB166382BL	01/30/2025	13:44	PS028990.D	7.19	0.00
PB166382BS	PB166382BS	01/30/2025	14:08	PS028991.D	7.19	0.00
PB166318TB	PB166318TB	01/30/2025	14:32	PS028992.D	7.19	0.00
JPP-20.1-012725	Q1206-04	01/30/2025	14:56	PS028993.D	7.19	0.00
JPP-20.1-012725MS	Q1206-04MS	01/30/2025	15:20	PS028994.D	7.19	0.00
JPP-20.1-012725MSD	Q1206-04MSD	01/30/2025	15:43	PS028995.D	7.19	0.00
JPP-16.3-012725	Q1206-08	01/30/2025	16:07	PS028996.D	7.19	0.00
IBLK	IBLK	01/30/2025	16:31	PS028997.D	7.19	0.00
HSTDCCC750	HSTDCCC750	01/30/2025	16:55	PS028998.D	7.19	0.00

Analytical Sequence

Client: RU2 Engineering, LLC	SDG No.: Q1206
Project: NYCDDC SANTWOBR Brooklyn Bridge BI	Instrument ID: ECD_S
GC Column: RTX-CLP2	ID: 0.32 (mm) Inst. Calib. Date(s): 01/14/2025 01/14/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCAA RT #	RT #
IBLK	IBLK	01/14/2025	10:07	PS028900.D	7.68	0.00
HSTDICC200	HSTDICC200	01/14/2025	10:31	PS028901.D	7.68	0.00
HSTDICC500	HSTDICC500	01/14/2025	10:55	PS028902.D	7.68	0.00
HSTDICC750	HSTDICC750	01/14/2025	11:19	PS028903.D	7.68	0.00
HSTDICC1000	HSTDICC1000	01/14/2025	11:43	PS028904.D	7.68	0.00
HSTDICC1500	HSTDICC1500	01/14/2025	12:07	PS028905.D	7.68	0.00
IBLK	IBLK	01/30/2025	12:56	PS028988.D	7.67	0.00
HSTDCCC750	HSTDCCC750	01/30/2025	13:20	PS028989.D	7.67	0.00
PB166382BL	PB166382BL	01/30/2025	13:44	PS028990.D	7.67	0.00
PB166382BS	PB166382BS	01/30/2025	14:08	PS028991.D	7.67	0.00
PB166318TB	PB166318TB	01/30/2025	14:32	PS028992.D	7.67	0.00
JPP-20.1-012725	Q1206-04	01/30/2025	14:56	PS028993.D	7.67	0.00
JPP-20.1-012725MS	Q1206-04MS	01/30/2025	15:20	PS028994.D	7.67	0.00
JPP-20.1-012725MSD	Q1206-04MSD	01/30/2025	15:43	PS028995.D	7.67	0.00
JPP-16.3-012725	Q1206-08	01/30/2025	16:07	PS028996.D	7.67	0.00
IBLK	IBLK	01/30/2025	16:31	PS028997.D	7.67	0.00
HSTDCCC750	HSTDCCC750	01/30/2025	16:55	PS028998.D	7.67	0.00

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-20.1-012725MS

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1206

SAS No.: Q1206

SDG NO.: Q1206

Lab Sample ID: Q1206-04MS

Date(s) Analyzed: 01/30/2025 01/30/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.31	8.26	8.36	50.4	6.6
	2	8.90	8.85	8.95	47.2	
2,4,5-TP(Silvex)	1	9.18	9.13	9.23	48.1	48.9
	2	9.81	9.76	9.86	79.2	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-20.1-012725MSD

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1206 **SAS No.:** Q1206 **SDG NO.:** Q1206

Lab Sample ID: Q1206-04MSD **Date(s) Analyzed:** 01/30/2025 01/30/2025

Instrument ID (1): ECD_S **Instrument ID (2):** ECD_S

GC Column: (1): RTX-CLP **ID:** 0.32 (mm) **GC Column:(2):** RTX-CLP2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.31	8.26	8.36	49.9	6.6
	2	8.90	8.85	8.95	46.7	
2,4,5-TP(Silvex)	1	9.18	9.13	9.23	47.5	49.4
	2	9.81	9.76	9.86	78.7	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB166382BS

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1206

SAS No.: Q1206

SDG NO.: Q1206

Lab Sample ID: PB166382BS

Date(s) Analyzed: 01/30/2025

01/30/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-TP(Silvex)	1	9.18	9.13	9.23	4.90	2.1
	2	9.80	9.75	9.85	4.80	
2,4-D	1	8.31	8.26	8.36	4.80	6.5
	2	8.91	8.86	8.96	4.50	



QC SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Report of Analysis

Client:	RU2 Engineering, LLC		Date Collected:		
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR		Date Received:		
Client Sample ID:	PB166382BL		SDG No.:	Q1206	
Lab Sample ID:	PB166382BL		Matrix:	TCLP	
Analytical Method:	SW8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028990.D	1	01/29/25 12:09	01/30/25 13:44	PB166382

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	2.00	U	0.49	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	2.00	U	0.45	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	445		39 - 175	89%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 13:44
Operator : AR\AJ
Sample : PB166382BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB166382BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:20:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

4) S	2,4-DCAA	7.192	7.671	1239.7E6	450.8E6	445.278	404.052
------	----------	-------	-------	----------	---------	---------	---------

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 13:44
Operator : AR\AJ
Sample : PB166382BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB166382BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:20:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.192 min
Delta R.T.: -0.005 min
Response: 1239659845
Conc: 445.28 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BL



#4 2,4-DCAA

R.T.: 7.671 min
Delta R.T.: -0.007 min
Response: 450846506
Conc: 404.05 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/14/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/14/25
Client Sample ID:	PIBLK-PS028900.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PS028900.D	Matrix:	TCLP
Analytical Method:	SW8151A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	TCLP Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028900.D	1		01/14/25	PS011425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	2.00	U	0.49	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	2.00	U	0.45	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	492		39 - 175	98%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
Data File : PS028900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2025 10:07
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 12:27:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

4) S	2,4-DCAA	7.198	7.678	1320.2E6	549.1E6	474.225	492.107
------	----------	-------	-------	----------	---------	---------	---------

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011425\
Data File : PS028900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2025 10:07
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 12:27:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 1320247914
Conc: 474.22 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#4 2,4-DCAA

R.T.: 7.678 min
Delta R.T.: 0.000 min
Response: 549099897
Conc: 492.11 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	PIBLK-PS028988.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PS028988.D	Matrix:	TCLP
Analytical Method:	SW8151A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	TCLP Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028988.D	1		01/30/25	PS013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	2.00	U	0.49	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	2.00	U	0.45	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	512		39 - 175	102%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 12:56
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:20:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

4) S	2,4-DCAA	7.192	7.670	1425.0E6	545.8E6	511.861	489.194
------	----------	-------	-------	----------	---------	---------	---------

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 12:56
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:20:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.192 min
Delta R.T.: -0.006 min
Response: 1425027475
Conc: 511.86 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#4 2,4-DCAA

R.T.: 7.670 min
Delta R.T.: -0.008 min
Response: 545849024
Conc: 489.19 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	PIBLK-PS028997.D	SDG No.:	Q1206
Lab Sample ID:	I.BLK-PS028997.D	Matrix:	TCLP
Analytical Method:	SW8151A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	TCLP Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028997.D	1		01/30/25	PS013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	2.00	U	0.49	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	2.00	U	0.45	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	510		39 - 175	102%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 16:31
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

4) S	2,4-DCAA	7.192	7.669	1420.2E6	557.3E6	510.124	499.475
------	----------	-------	-------	----------	---------	---------	---------

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 16:31
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.192 min
Delta R.T.: -0.006 min
Response: 1420192589
Conc: 510.12 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#4 2,4-DCAA

R.T.: 7.669 min
Delta R.T.: -0.008 min
Response: 557320458
Conc: 499.47 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC		Date Collected:		
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR		Date Received:		
Client Sample ID:	PB166382BS		SDG No.:	Q1206	
Lab Sample ID:	PB166382BS		Matrix:	TCLP	
Analytical Method:	SW8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028991.D	1	01/29/25 12:09	01/30/25 14:08	PB166382

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	4.80		0.49	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	4.90		0.45	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	510		39 - 175	102%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
 Data File : PS028991.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2025 14:08
 Operator : AR\AJ
 Sample : PB166382BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB166382BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 05:20:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.192	7.671	1420.7E6	519.3E6	510.299	465.386
Target Compounds						
1) T Dalapon	2.615	2.664	1504.1E6	889.0E6	504.438	435.750
2) T 3,5-DICHL...	6.370	6.637	1915.7E6	716.8E6	479.295	433.712
3) T 4-Nitroph...	6.991	7.201	826.9E6	399.5E6	466.603	448.961
5) T DICAMBA	7.377	7.867	5768.0E6	2533.9E6	486.285	455.002
6) T MCPP	7.557	7.970	316.5E6	124.9E6	46.436	41.535
7) T MCPA	7.705	8.211	460.0E6	173.2E6	46.694	40.766
8) T DICHLOORPROP	8.080	8.578	1517.7E6	634.7E6	478.986	451.565
9) T 2,4-D	8.310	8.905	1628.4E6	674.5E6	481.853	449.817
10) T Pentachlo...	8.606	9.427	24191.9E6	11288.2E6	501.522	487.274
11) T 2,4,5-TP ...	9.181	9.804	9395.0E6	4504.1E6	491.046	478.168
12) T 2,4,5-T	9.472	10.221	9419.2E6	4256.1E6	490.659	472.432
13) T 2,4-DB	10.042	10.785	1678.4E6	443.5E6	473.135	445.398
14) T DINOSEB	11.245	11.162	7815.4E6	2946.5E6	472.305	459.142
15) T Picloram	11.055	12.244	14821.8E6	6093.3E6	469.755	454.055
16) T DCPA	11.539	12.199	14204.3E6	5636.9E6	495.251	496.538



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 14:08
Operator : AR\AJ
Sample : PB166382BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB166382BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:20:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.615 min
Delta R.T.: 0.000 min
Response: 1504103638
Conc: 504.44 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BS




#1 Dalapon

R.T.: 2.664 min
Delta R.T.: -0.003 min
Response: 888993582
Conc: 435.75 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.370 min
Delta R.T.: -0.005 min
Response: 1915671335
Conc: 479.30 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.637 min
Delta R.T.: -0.006 min
Response: 716760619
Conc: 433.71 ng/ml



#3 4-Nitrophenol


R.T.: 6.991 min
Delta R.T.: -0.005 min
Response: 826857755
Conc: 466.60 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BS




#3 4-Nitrophenol

R.T.: 7.201 min
Delta R.T.: -0.006 min
Response: 399474322
Conc: 448.96 ng/ml




#4 2,4-DCAA

R.T.: 7.192 min
Delta R.T.: -0.005 min
Response: 1420679312
Conc: 510.30 ng/ml



#4 2,4-DCAA


R.T.: 7.671 min
Delta R.T.: -0.006 min
Response: 519283855
Conc: 465.39 ng/ml



#5 DICAMBA


R.T.: 7.377 min
Delta R.T.: -0.007 min
Response: 5768040666
Conc: 486.28 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BS




#5 DICAMBA

R.T.: 7.867 min
Delta R.T.: -0.008 min
Response: 2533922735
Conc: 455.00 ng/ml




#6 MCPP

R.T.: 7.557 min
Delta R.T.: -0.010 min
Response: 316519868
Conc: 46.44 ug/ml



#6 MCPP


R.T.: 7.970 min
Delta R.T.: -0.011 min
Response: 124935835
Conc: 41.54 ug/ml



#7 MCPA


R.T.: 7.705 min
Delta R.T.: -0.012 min
Response: 459983527
Conc: 46.69 ug/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BS




#7 MCPA

R.T.: 8.211 min
Delta R.T.: -0.013 min
Response: 173152749
Conc: 40.77 ug/ml




#8 DICHLORPROP

R.T.: 8.080 min
Delta R.T.: -0.007 min
Response: 1517700990
Conc: 478.99 ng/ml



#8 DICHLORPROP


R.T.: 8.578 min
Delta R.T.: -0.008 min
Response: 634719272
Conc: 451.57 ng/ml



#9 2,4-D


R.T.: 8.310 min
Delta R.T.: -0.007 min
Response: 1628398734
Conc: 481.85 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BS




#9 2,4-D

R.T.: 8.905 min
Delta R.T.: -0.008 min
Response: 674508943
Conc: 449.82 ng/ml




#10 Pentachlorophenol

R.T.: 8.606 min
Delta R.T.: -0.009 min
Response: 24191949482
Conc: 501.52 ng/ml



#10 Pentachlorophenol


R.T.: 9.427 min
Delta R.T.: -0.010 min
Response: 11288188163
Conc: 487.27 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.181 min
Delta R.T.: -0.008 min
Response: 9395015703
Conc: 491.05 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BS




#11 2,4,5-TP (SILVEX)

R.T.: 9.804 min
Delta R.T.: -0.009 min
Response: 4504054828
Conc: 478.17 ng/ml




#12 2,4,5-T

R.T.: 9.472 min
Delta R.T.: -0.008 min
Response: 9419151097
Conc: 490.66 ng/ml



#12 2,4,5-T


R.T.: 10.221 min
Delta R.T.: -0.009 min
Response: 4256142852
Conc: 472.43 ng/ml



#13 2,4-DB


R.T.: 10.042 min
Delta R.T.: -0.008 min
Response: 1678377012
Conc: 473.13 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BS




#13 2,4-DB

R.T.: 10.785 min
Delta R.T.: -0.010 min
Response: 443508004
Conc: 445.40 ng/ml




#14 DINOSEB

R.T.: 11.245 min
Delta R.T.: -0.010 min
Response: 7815381011
Conc: 472.31 ng/ml



#14 DINOSEB


R.T.: 11.162 min
Delta R.T.: -0.010 min
Response: 2946459299
Conc: 459.14 ng/ml



#15 Picloram


R.T.: 11.055 min
Delta R.T.: -0.010 min
Response: 14821769096
Conc: 469.75 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB166382BS




#15 Picloram

R.T.: 12.244 min
Delta R.T.: -0.012 min
Response: 6093263718
Conc: 454.05 ng/ml



#16 DCPA

R.T.: 11.539 min
Delta R.T.: -0.010 min
Response: 14204307222
Conc: 495.25 ng/ml



#16 DCPA

R.T.: 12.199 min
Delta R.T.: -0.012 min
Response: 5636877056
Conc: 496.54 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/27/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	JPP-20.1-012725MS	SDG No.:	Q1206
Lab Sample ID:	Q1206-04MS	Matrix:	TCLP
Analytical Method:	SW8151A	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	SW3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028994.D	1	01/29/25 12:09	01/30/25 15:20	PB166382

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	50.4		4.90	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	79.2	P	4.50	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	424		39 - 175	85%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
 Data File : PS028994.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2025 15:20
 Operator : AR\AJ
 Sample : Q1206-04MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 JPP-20.1-012725MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/31/2025
 Supervised By : Ankita Jodhani 01/31/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 05:21:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.193	7.670	1181.5E6	298.5E6	424.394	267.487 #
Target Compounds						
1) T Dalapon	2.618	2.672	1064.4E6	698.4E6	356.971m	342.324m
2) T 3,5-DICHL...	6.369	6.636	1659.4E6	619.8E6	415.180	375.030
3) T 4-Nitroph...	6.991	7.203	21863746	8791881	12.338	9.881m
5) T DICAMBA	7.376	7.866	5041.4E6	2347.2E6	425.026	421.474
6) T MCPP	7.556	7.969	345.4E6	114.5E6	50.678m	38.064
7) T MCPA	7.705	8.210	392.7E6	199.5E6	39.868	46.974
8) T DICHLORPROP	8.079	8.577	1344.2E6	578.6E6	424.232	411.660
9) T 2,4-D	8.308	8.904	1701.6E6	708.2E6	503.504	472.265
10) T Pentachlo...	8.605	9.426	15021.5E6	6845.9E6	311.410	295.516
11) T 2,4,5-TP ...	9.180	9.806	9201.1E6	7459.4E6	480.912	791.923 #
12) T 2,4,5-T	9.471	10.220	8803.6E6	4089.1E6	458.596	453.894
13) T 2,4-DB	10.043	10.785	1136.9E6	379.1E6	320.482	380.745
14) T DINOSEB	11.244	11.161	3407.0E6	1218.6E6	205.892	189.886
15) T Picloram	11.054	12.242	12424.5E6	5208.4E6	393.776	388.120
16) T DCPA	11.538	12.199	11538.0E6	5294.2E6	402.288	466.351

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 15:20
Operator : AR\AJ
Sample : Q1206-04MS
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/31/2025
Supervised By : Ankita Jodhani 01/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Dalapon

R.T.: 2.618 min
Delta R.T.: 0.003 min
Response: 1064396174
Conc: 356.97 ng/ml


Instrument :

ECD_S

ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#1 Dalapon

R.T.: 2.672 min
Delta R.T.: 0.005 min
Response: 698390611
Conc: 342.32 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.369 min
Delta R.T.: -0.006 min
Response: 1659410998
Conc: 415.18 ng/ml



#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.636 min
Delta R.T.: -0.007 min
Response: 619782760
Conc: 375.03 ng/ml




#3 4-Nitrophenol

R.T.: 6.991 min
Delta R.T.: -0.004 min
Response: 21863746
Conc: 12.34 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#3 4-Nitrophenol

R.T.: 7.203 min
Delta R.T.: -0.004 min
Response: 8791881
Conc: 9.88 ng/ml




#4 2,4-DCAA

R.T.: 7.193 min
Delta R.T.: -0.005 min
Response: 1181518228
Conc: 424.39 ng/ml



#4 2,4-DCAA

R.T.: 7.670 min
Delta R.T.: -0.008 min
Response: 298465679
Conc: 267.49 ng/ml




#5 DICAMBA

R.T.: 7.376 min
Delta R.T.: -0.007 min
Response: 5041419952
Conc: 425.03 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#5 DICAMBA

R.T.: 7.866 min
Delta R.T.: -0.009 min
Response: 2347205036
Conc: 421.47 ng/ml




#6 MCPP

R.T.: 7.556 min
Delta R.T.: -0.012 min
Response: 345434449
Conc: 50.68 ug/ml m



#6 MCPP

R.T.: 7.969 min
Delta R.T.: -0.012 min
Response: 114495084
Conc: 38.06 ug/ml




#7 MCPA

R.T.: 7.705 min
Delta R.T.: -0.012 min
Response: 392738712
Conc: 39.87 ug/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#7 MCPA

R.T.: 8.210 min
Delta R.T.: -0.014 min
Response: 199520429
Conc: 46.97 ug/ml




#8 DICHLORPROP

R.T.: 8.079 min
Delta R.T.: -0.009 min
Response: 1344208911
Conc: 424.23 ng/ml



#8 DICHLORPROP

R.T.: 8.577 min
Delta R.T.: -0.009 min
Response: 578628859
Conc: 411.66 ng/ml




#9 2,4-D

R.T.: 8.308 min
Delta R.T.: -0.009 min
Response: 1701566314
Conc: 503.50 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#9 2,4-D

R.T.: 8.904 min
Delta R.T.: -0.010 min
Response: 708169371
Conc: 472.26 ng/ml




#10 Pentachlorophenol

R.T.: 8.605 min
Delta R.T.: -0.010 min
Response: 15021514739
Conc: 311.41 ng/ml



#10 Pentachlorophenol

R.T.: 9.426 min
Delta R.T.: -0.011 min
Response: 6845939571
Conc: 295.52 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.180 min
Delta R.T.: -0.009 min
Response: 9201134277
Conc: 480.91 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#11 2,4,5-TP (SILVEX)

R.T.: 9.806 min
Delta R.T.: -0.008 min
Response: 7459445775
Conc: 791.92 ng/ml




#12 2,4,5-T

R.T.: 9.471 min
Delta R.T.: -0.010 min
Response: 8803633760
Conc: 458.60 ng/ml



#12 2,4,5-T

R.T.: 10.220 min
Delta R.T.: -0.011 min
Response: 4089136184
Conc: 453.89 ng/ml




#13 2,4-DB

R.T.: 10.043 min
Delta R.T.: -0.007 min
Response: 1136863950
Conc: 320.48 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#13 2,4-DB

R.T.: 10.785 min
Delta R.T.: -0.010 min
Response: 379128977
Conc: 380.74 ng/ml




#14 DINOSEB

R.T.: 11.244 min
Delta R.T.: -0.010 min
Response: 3406956478
Conc: 205.89 ng/ml



#14 DINOSEB

R.T.: 11.161 min
Delta R.T.: -0.011 min
Response: 1218560874
Conc: 189.89 ng/ml




#15 Picloram

R.T.: 11.054 min
Delta R.T.: -0.011 min
Response: 12424490250
Conc: 393.78 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#15 Picloram

R.T.: 12.242 min
Delta R.T.: -0.013 min
Response: 5208444458
Conc: 388.12 ng/ml



#16 DCPA

R.T.: 11.538 min
Delta R.T.: -0.012 min
Response: 11538027819
Conc: 402.29 ng/ml



#16 DCPA

R.T.: 12.199 min
Delta R.T.: -0.012 min
Response: 5294179863
Conc: 466.35 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/27/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/28/25
Client Sample ID:	JPP-20.1-012725MSD	SDG No.:	Q1206
Lab Sample ID:	Q1206-04MSD	Matrix:	TCLP
Analytical Method:	SW8151A	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	TCLP Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS028995.D	1	01/29/25 12:09	01/30/25 15:43	PB166382

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	49.9		4.90	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	78.7	P	4.50	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	424		39 - 175	85%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
 Data File : PS028995.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2025 15:43
 Operator : AR\AJ
 Sample : Q1206-04MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 JPP-20.1-012725MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 01/31/2025
 Supervised By :Ankita Jodhani 01/31/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 05:21:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.194	7.671	1179.5E6	296.0E6	423.681	265.280 #
Target Compounds						
1) T Dalapon	2.617	2.668	1033.2E6	685.5E6	346.503m	336.015m
2) T 3,5-DICHL...	6.369	6.637	1691.2E6	619.2E6	423.136m	374.679
3) T 4-Nitroph...	6.986	7.203	29719762	9366163	16.771	10.526m#
5) T DICAMBA	7.377	7.867	4957.9E6	2334.5E6	417.985	419.185
6) T MCPP	7.557	7.970	310.4E6	113.0E6	45.538m	37.583
7) T MCPA	7.705	8.210	382.5E6	200.7E6	38.827	47.241
8) T DICHLORPROP	8.080	8.577	1343.9E6	574.2E6	424.119	408.476
9) T 2,4-D	8.309	8.904	1686.6E6	700.9E6	499.089	467.435
10) T Pentachlo...	8.606	9.427	14766.5E6	6813.0E6	306.124	294.094
11) T 2,4,5-TP ...	9.181	9.808	9080.7E6	7414.5E6	474.620	787.147 #
12) T 2,4,5-T	9.472	10.220	8673.1E6	4076.7E6	451.798	452.510
13) T 2,4-DB	10.044	10.785	1140.5E6	373.2E6	321.508	374.751
14) T DINOSEB	11.245	11.162	3352.0E6	1243.6E6	202.569	193.789m
15) T Picloram	11.055	12.243	12299.2E6	5195.6E6	389.805	387.160
16) T DCPA	11.539	12.200	11433.1E6	5277.9E6	398.628	464.916

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 15:43
Operator : AR\AJ
Sample : Q1206-04MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/31/2025
Supervised By : Ankita Jodhani 01/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:21:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 2.617 min
Delta R.T.: 0.002 min
Response: 1033183208
Conc: 346.50 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#1 Dalapon

R.T.: 2.668 min
Delta R.T.: 0.000 min
Response: 685521048
Conc: 336.02 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.369 min
Delta R.T.: -0.005 min
Response: 1691213013
Conc: 423.14 ng/ml



#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.637 min
Delta R.T.: -0.006 min
Response: 619202584
Conc: 374.68 ng/ml




#3 4-Nitrophenol

R.T.: 6.986 min
Delta R.T.: -0.010 min
Response: 29719762
Conc: 16.77 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#3 4-Nitrophenol

R.T.: 7.203 min
Delta R.T.: -0.004 min
Response: 9366163
Conc: 10.53 ng/ml




#4 2,4-DCAA

R.T.: 7.194 min
Delta R.T.: -0.004 min
Response: 1179534201
Conc: 423.68 ng/ml



#4 2,4-DCAA

R.T.: 7.671 min
Delta R.T.: -0.007 min
Response: 296002929
Conc: 265.28 ng/ml




#5 DICAMBA

R.T.: 7.377 min
Delta R.T.: -0.006 min
Response: 4957901038
Conc: 417.98 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#5 DICAMBA

R.T.: 7.867 min
Delta R.T.: -0.008 min
Response: 2334455244
Conc: 419.18 ng/ml




#6 MCPP

R.T.: 7.557 min
Delta R.T.: -0.010 min
Response: 310397006
Conc: 45.54 ug/ml m



#6 MCPP

R.T.: 7.970 min
Delta R.T.: -0.011 min
Response: 113047325
Conc: 37.58 ug/ml




#7 MCPA

R.T.: 7.705 min
Delta R.T.: -0.012 min
Response: 382482730
Conc: 38.83 ug/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#7 MCPA

R.T.: 8.210 min
Delta R.T.: -0.013 min
Response: 200654252
Conc: 47.24 ug/ml




#8 DICHLORPROP

R.T.: 8.080 min
Delta R.T.: -0.007 min
Response: 1343852640
Conc: 424.12 ng/ml



#8 DICHLORPROP

R.T.: 8.577 min
Delta R.T.: -0.009 min
Response: 574152954
Conc: 408.48 ng/ml




#9 2,4-D

R.T.: 8.309 min
Delta R.T.: -0.008 min
Response: 1686645480
Conc: 499.09 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#9 2,4-D

R.T.: 8.904 min
Delta R.T.: -0.009 min
Response: 700927415
Conc: 467.44 ng/ml




#10 Pentachlorophenol

R.T.: 8.606 min
Delta R.T.: -0.009 min
Response: 14766517826
Conc: 306.12 ng/ml



#10 Pentachlorophenol

R.T.: 9.427 min
Delta R.T.: -0.010 min
Response: 6812985762
Conc: 294.09 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.181 min
Delta R.T.: -0.008 min
Response: 9080749462
Conc: 474.62 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#11 2,4,5-TP (SILVEX)

R.T.: 9.808 min
Delta R.T.: -0.006 min
Response: 7414457519
Conc: 787.15 ng/ml




#12 2,4,5-T

R.T.: 9.472 min
Delta R.T.: -0.008 min
Response: 8673136344
Conc: 451.80 ng/ml



#12 2,4,5-T

R.T.: 10.220 min
Delta R.T.: -0.010 min
Response: 4076661934
Conc: 452.51 ng/ml




#13 2,4-DB

R.T.: 10.044 min
Delta R.T.: -0.007 min
Response: 1140503559
Conc: 321.51 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#13 2,4-DB

R.T.: 10.785 min
Delta R.T.: -0.010 min
Response: 373160817
Conc: 374.75 ng/ml




#14 DINOSEB

R.T.: 11.245 min
Delta R.T.: -0.010 min
Response: 3351965531
Conc: 202.57 ng/ml



#14 DINOSEB

R.T.: 11.162 min
Delta R.T.: -0.010 min
Response: 1243603046
Conc: 193.79 ng/ml m




#15 Picloram

R.T.: 11.055 min
Delta R.T.: -0.010 min
Response: 12299193127
Conc: 389.81 ng/ml

Instrument :
ECD_S
ClientSampleId :
JPP-20.1-012725MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2025
Supervised By :Ankita Jodhani 01/31/2025




#15 Picloram

R.T.: 12.243 min
Delta R.T.: -0.012 min
Response: 5195561651
Conc: 387.16 ng/ml



#16 DCPA

R.T.: 11.539 min
Delta R.T.: -0.011 min
Response: 11433076222
Conc: 398.63 ng/ml



#16 DCPA

R.T.: 12.200 min
Delta R.T.: -0.011 min
Response: 5277889692
Conc: 464.92 ng/ml

Manual Integration Report

Sequence:	PS011425	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
HSTDICC1500	PS028905.D	Dalapon	Abdul	1/14/2025 4:08:21 PM	Ankita	1/15/2025 7:45:19	Peak Integrated by Software
HSTDICV750	PS028906.D	2,4-DCAA	Abdul	1/14/2025 4:08:25 PM	Ankita	1/15/2025 7:45:21	Peak Integrated by Software
HSTDICV750	PS028906.D	Dalapon	Abdul	1/14/2025 4:08:25 PM	Ankita	1/15/2025 7:45:21	Peak Integrated by Software
HSTDCCC750	PS028908.D	Dalapon	Abdul	1/14/2025 4:08:28 PM	Ankita	1/15/2025 7:45:22	Peak Integrated by Software

Manual Integration Report

Sequence:	PS013025	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB166318TB	PS028992.D	2,4-DCAA	Abdul	1/31/2025 10:34:27 AM	Ankita	1/31/2025 11:40:22	Peak Integrated by Software
Q1206-04MS	PS028994.D	4-Nitrophenol #2	Abdul	1/31/2025 10:34:01 AM	Ankita	1/31/2025 11:40:23	Peak Integrated by Software
Q1206-04MS	PS028994.D	Dalapon	Abdul	1/31/2025 10:34:01 AM	Ankita	1/31/2025 11:40:23	Peak Integrated by Software
Q1206-04MS	PS028994.D	Dalapon #2	Abdul	1/31/2025 10:34:01 AM	Ankita	1/31/2025 11:40:23	Peak Integrated by Software
Q1206-04MS	PS028994.D	MCPP	Abdul	1/31/2025 10:34:01 AM	Ankita	1/31/2025 11:40:23	Peak Integrated by Software
Q1206-04MSD	PS028995.D	3,5-DICHLOROBENZOI C ACID	Abdul	1/31/2025 10:34:07 AM	Ankita	1/31/2025 11:40:24	Peak Integrated by Software
Q1206-04MSD	PS028995.D	4-Nitrophenol #2	Abdul	1/31/2025 10:34:07 AM	Ankita	1/31/2025 11:40:24	Peak Integrated by Software
Q1206-04MSD	PS028995.D	Dalapon	Abdul	1/31/2025 10:34:07 AM	Ankita	1/31/2025 11:40:24	Peak Integrated by Software
Q1206-04MSD	PS028995.D	Dalapon #2	Abdul	1/31/2025 10:34:07 AM	Ankita	1/31/2025 11:40:24	Peak Integrated by Software
Q1206-04MSD	PS028995.D	DINOSEB #2	Abdul	1/31/2025 10:34:07 AM	Ankita	1/31/2025 11:40:24	Peak Integrated by Software
Q1206-04MSD	PS028995.D	MCPP	Abdul	1/31/2025 10:34:07 AM	Ankita	1/31/2025 11:40:24	Peak Integrated by Software
HSTDCCC750	PS029005.D	DCPA #2	Abdul	1/31/2025 1:18:42 PM	Ankita	1/31/2025 1:22:37	Peak Integrated by Software
HSTDCCC750	PS029005.D	MCPP	Abdul	1/31/2025 1:18:42 PM	Ankita	1/31/2025 1:22:37	Peak Integrated by Software

Manual Integration Report

Sequence:	PS013025	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
HSTDCCC750	PS029005.D	Picloram #2	Abdul	1/31/2025 1:18:42 PM	Ankita	1/31/2025 1:22:37	Peak Integrated by Software

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS011425

Review By	Abdul	Review On	1/14/2025 4:08:46 PM
Supervise By	Ankita	Supervise On	1/15/2025 7:45:27 AM
SubDirectory	PS011425	HP Acquire Method	HP Processing Method ps011425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS028899.D	14 Jan 2025 09:43	AR\AJ	Ok
2	I.BLK	PS028900.D	14 Jan 2025 10:07	AR\AJ	Ok
3	HSTDICC200	PS028901.D	14 Jan 2025 10:31	AR\AJ	Ok
4	HSTDICC500	PS028902.D	14 Jan 2025 10:55	AR\AJ	Ok
5	HSTDICC750	PS028903.D	14 Jan 2025 11:19	AR\AJ	Ok
6	HSTDICC1000	PS028904.D	14 Jan 2025 11:43	AR\AJ	Ok
7	HSTDICC1500	PS028905.D	14 Jan 2025 12:07	AR\AJ	Ok,M
8	HSTDICV750	PS028906.D	14 Jan 2025 12:31	AR\AJ	Ok,M
9	I.BLK	PS028907.D	14 Jan 2025 12:56	AR\AJ	Ok
10	HSTDCCC750	PS028908.D	14 Jan 2025 13:20	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS013025

Review By	Abdul	Review On	1/31/2025 10:34:57 AM		
Supervise By	Ankita	Supervise On	1/31/2025 11:40:30 AM		
SubDirectory	PS013025	HP Acquire Method	PS011425	HP Processing Method	ps011425 8151
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		PP24064,PP24065,PP24066,PP24067,PP24068			
CCC Internal Standard/PEM		PP24066			
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		PP24069,PP24070			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS028987.D	30 Jan 2025 12:32	AR\AJ	Ok
2	I.BLK	PS028988.D	30 Jan 2025 12:56	AR\AJ	Ok
3	HSTDCCC750	PS028989.D	30 Jan 2025 13:20	AR\AJ	Ok
4	PB166382BL	PS028990.D	30 Jan 2025 13:44	AR\AJ	Ok
5	PB166382BS	PS028991.D	30 Jan 2025 14:08	AR\AJ	Ok
6	PB166318TB	PS028992.D	30 Jan 2025 14:32	AR\AJ	Ok,M
7	Q1206-04	PS028993.D	30 Jan 2025 14:56	AR\AJ	Ok
8	Q1206-04MS	PS028994.D	30 Jan 2025 15:20	AR\AJ	Ok,M
9	Q1206-04MSD	PS028995.D	30 Jan 2025 15:43	AR\AJ	Ok,M
10	Q1206-08	PS028996.D	30 Jan 2025 16:07	AR\AJ	Ok
11	I.BLK	PS028997.D	30 Jan 2025 16:31	AR\AJ	Ok
12	HSTDCCC750	PS028998.D	30 Jan 2025 16:55	AR\AJ	Ok
13	Q1207-04	PS028999.D	30 Jan 2025 17:19	AR\AJ	Ok
14	Q1207-08	PS029000.D	30 Jan 2025 17:43	AR\AJ	Ok
15	Q1207-12	PS029001.D	30 Jan 2025 18:07	AR\AJ	Ok
16	Q1207-16	PS029002.D	30 Jan 2025 18:31	AR\AJ	Ok
17	Q1207-20	PS029003.D	30 Jan 2025 18:55	AR\AJ	Ok
18	I.BLK	PS029004.D	30 Jan 2025 19:19	AR\AJ	Ok
19	HSTDCCC750	PS029005.D	31 Jan 2025 01:01	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS011425

Review By	Abdul	Review On	1/14/2025 4:08:46 PM
Supervise By	Ankita	Supervise On	1/15/2025 7:45:27 AM
SubDirectory	PS011425	HP Acquire Method	HP Processing Method ps011425 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC	PP24066
Internal Standard/PEM	
ICV/I.BLK	PP24069,PP24070
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028899.D	14 Jan 2025 09:43		AR\AJ	Ok
2	I.BLK	I.BLK	PS028900.D	14 Jan 2025 10:07		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS028901.D	14 Jan 2025 10:31		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS028902.D	14 Jan 2025 10:55		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS028903.D	14 Jan 2025 11:19		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS028904.D	14 Jan 2025 11:43		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS028905.D	14 Jan 2025 12:07		AR\AJ	Ok,M
8	HSTDICV750	ICVPS011425	PS028906.D	14 Jan 2025 12:31		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS028907.D	14 Jan 2025 12:56		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS028908.D	14 Jan 2025 13:20		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS013025

Review By	Abdul	Review On	1/31/2025 10:34:57 AM
Supervise By	Ankita	Supervise On	1/31/2025 11:40:30 AM
SubDirectory	PS013025	HP Acquire Method	PS011425 HP Processing Method ps011425 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028987.D	30 Jan 2025 12:32		AR\AJ	Ok
2	I.BLK	I.BLK	PS028988.D	30 Jan 2025 12:56		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS028989.D	30 Jan 2025 13:20		AR\AJ	Ok
4	PB166382BL	PB166382BL	PS028990.D	30 Jan 2025 13:44		AR\AJ	Ok
5	PB166382BS	PB166382BS	PS028991.D	30 Jan 2025 14:08		AR\AJ	Ok
6	PB166318TB	PB166318TB	PS028992.D	30 Jan 2025 14:32		AR\AJ	Ok,M
7	Q1206-04	JPP-20.1-012725	PS028993.D	30 Jan 2025 14:56		AR\AJ	Ok
8	Q1206-04MS	JPP-20.1-012725MS	PS028994.D	30 Jan 2025 15:20	Some compound recovery fail	AR\AJ	Ok,M
9	Q1206-04MSD	JPP-20.1-012725MSD	PS028995.D	30 Jan 2025 15:43	Some compound recovery fail	AR\AJ	Ok,M
10	Q1206-08	JPP-16.3-012725	PS028996.D	30 Jan 2025 16:07		AR\AJ	Ok
11	I.BLK	I.BLK	PS028997.D	30 Jan 2025 16:31		AR\AJ	Ok
12	HSTDCCC750	HSTDCCC750	PS028998.D	30 Jan 2025 16:55		AR\AJ	Ok
13	Q1207-04	JPP-2.1-012725	PS028999.D	30 Jan 2025 17:19		AR\AJ	Ok
14	Q1207-08	JPP-5.1-012725	PS029000.D	30 Jan 2025 17:43		AR\AJ	Ok
15	Q1207-12	JPP-4.5-012725	PS029001.D	30 Jan 2025 18:07		AR\AJ	Ok
16	Q1207-16	JPP-16.2-012725	PS029002.D	30 Jan 2025 18:31		AR\AJ	Ok
17	Q1207-20	JPP-20.2-012725	PS029003.D	30 Jan 2025 18:55		AR\AJ	Ok
18	I.BLK	I.BLK	PS029004.D	30 Jan 2025 19:19		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS013025

Review By	Abdul	Review On	1/31/2025 10:34:57 AM				
Supervise By	Ankita	Supervise On	1/31/2025 11:40:30 AM				
SubDirectory	PS013025	HP Acquire Method	PS011425	HP Processing Method	ps011425 8151		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		PP24064,PP24065,PP24066,PP24067,PP24068					
CCC		PP24066					
Internal Standard/PEM							
ICV/I.BLK		PP24069,PP24070					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

19	HSTDCCC750	HSTDCCC750	PS029005.D	31 Jan 2025 01:01		AR\AJ	Ok,M
----	------------	------------	------------	-------------------	--	-------	------

M : Manual Integration

SOP ID : M1311-TCLP-15
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : T-1
TCLP Filter ID : 114771

Start Prep Date : 01/28/2025 **Time :** 16:30
End Prep Date : 01/29/2025 **Time :** 09:20
Combination Ratio : 20
ZHE Cleaning Batch : N/A
Initial Room Temperature: 24 °C
Final Room Temperature: 22 °C
TCLP Technician Signature : *[Signature]*
Supervisor By : *[Signature]*

Standardized Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP110801
HCL-TCLP,1N	N/A	WP110803
HNO3-TCLP,1N	N/A	WP110804
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1941,W1942	W3166,W1938,W1939,W1940,
1 Liter Amber	N/A	90424-08
120ml Plastic bottle	N/A	405130101
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 checked,30 rpm. q1209-05 is used for MS-MSD. Particle size reduction is not required.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
01/29/25 11:00	<i>[Signature]</i> / TCLP Room	<i>[Signature]</i> / <i>[Signature]</i>
	Preparation Group	Analysis Group

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB166318TB	LEB318	11	N/A	2000	N/A	N/A	N/A	4.94	1.0	T-1
Q1205-02	VNJ-236	01	100.03	2000	N/A	N/A	N/A	4.0	1.5	T-1
Q1206-04	JPP-20.1-012725	02	100.02	2000	N/A	N/A	N/A	6.2	1.0	T-1
Q1206-08	JPP-16.3-012725	03	100.02	2000	N/A	N/A	N/A	7.2	1.5	T-1
Q1207-04	JPP-2.1-012725	04	100.03	2000	N/A	N/A	N/A	7.0	1.0	T-1
Q1207-08	JPP-5.1-012725	05	100.04	2000	N/A	N/A	N/A	7.6	1.5	T-1
Q1207-12	JPP-4.5-012725	06	100.03	2000	N/A	N/A	N/A	7.2	1.5	T-1
Q1207-16	JPP-16.2-012725	07	100.02	2000	N/A	N/A	N/A	7.6	1.0	T-1
Q1207-20	JPP-20.2-012725	08	100.02	2000	N/A	N/A	N/A	7.0	1.5	T-1
Q1209-04	WC-4	09	100.01	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q1209-08	WC-5	10	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-1

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB166318TB	LEB318	N/A	N/A	N/A	N/A	N/A	N/A
Q1205-02	VNJ-236	N/A	N/A	N/A	N/A	100	N/A
Q1206-04	JPP-20.1-012725	N/A	N/A	N/A	N/A	100	N/A
Q1206-08	JPP-16.3-012725	N/A	N/A	N/A	N/A	100	N/A
Q1207-04	JPP-2.1-012725	N/A	N/A	N/A	N/A	100	N/A
Q1207-08	JPP-5.1-012725	N/A	N/A	N/A	N/A	100	N/A
Q1207-12	JPP-4.5-012725	N/A	N/A	N/A	N/A	100	N/A
Q1207-16	JPP-16.2-012725	N/A	N/A	N/A	N/A	100	N/A
Q1207-20	JPP-20.2-012725	N/A	N/A	N/A	N/A	100	N/A
Q1209-04	WC-4	N/A	N/A	N/A	N/A	100	N/A
Q1209-08	WC-5	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 /WC S-2

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB166318TB	LEB318	N/A	N/A	N/A	N/A	#1	4.94
Q1205-02	VNJ-236	5.02	96.5	6.0	2.0	#1	4.94
Q1206-04	JPP-20.1-012725	5.03	96.5	8.6	3.0	#1	4.94
Q1206-08	JPP-16.3-012725	5.02	96.5	9.1	3.5	#1	4.94
Q1207-04	JPP-2.1-012725	5.01	96.5	9.0	4.0	#1	4.94
Q1207-08	JPP-5.1-012725	5.02	96.5	11.0	4.5	#1	4.94
Q1207-12	JPP-4.5-012725	5.03	96.5	11.5	4.5	#1	4.94
Q1207-16	JPP-16.2-012725	5.02	96.5	10.5	4.0	#1	4.94
Q1207-20	JPP-20.2-012725	5.01	96.5	9.1	3.5	#1	4.94
Q1209-04	WC-4	5.02	96.5	6.4	2.5	#1	4.94
Q1209-08	WC-5	5.03	96.5	6.6	2.0	#1	4.94

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tclp q1109

WorkList ID : 187224

Department : TCLP Extraction

Date : 01-28-2025 14:21:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1205-02	VNJ-236	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N31	01/28/2025	1311
Q1206-04	JPP-20.1-012725	Solid	TCLP Extraction	Cool 4 deg C	RUTW01	E11	01/27/2025	1311
Q1206-08	JPP-16.3-012725	Solid	TCLP Extraction	Cool 4 deg C	RUTW01	E11	01/27/2025	1311
Q1207-04	JPP-2.1-012725	Solid	TCLP Extraction	Cool 4 deg C	RUTW01	E11	01/27/2025	1311
Q1207-08	JPP-5.1-012725	Solid	TCLP Extraction	Cool 4 deg C	RUTW01	E11	01/27/2025	1311
Q1207-12	JPP-4.5-012725	Solid	TCLP Extraction	Cool 4 deg C	RUTW01	E11	01/27/2025	1311
Q1207-16	JPP-16.2-012725	Solid	TCLP Extraction	Cool 4 deg C	RUTW01	E11	01/27/2025	1311
Q1207-20	JPP-20.2-012725	Solid	TCLP Extraction	Cool 4 deg C	RUTW01	E11	01/27/2025	1311
Q1209-04	WC-4	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N41	01/28/2025	1311
Q1209-08	WC-5	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N41	01/28/2025	1311

Date/Time 01/28/25 14:30

Raw Sample Received by: SB WEC

Raw Sample Relinquished by: [Signature]

Q1206-TCLP Herbicide

Date/Time 01/28/25 17:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

SOP ID: M8151A-Herbicide-22

Clean Up SOP #: N/A

Matrix : Water

Weigh By: N/A

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: E3574

Extraction By: RS

Filter By: RS

pH Meter ID: N/A

Hood ID: 4,7

Extraction Start Date : 01/29/2025

Extraction Start Time : 12:09

Extraction End Date : 01/30/2025

Extraction End Time : 12:15

Concentration By: EH

Supervisor By : rajesh

Extraction Method: ☒ Separatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhle

Standarded Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5/500 PPM	PP24079
Surrogate	1.0ML	5000 PPB	PP24078
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Ether	N/A	E3370
Acidified Na2SO4	N/A	EP2576
NAOH 6N	N/A	EP2553
1:3 SULPHURIC ACID	N/A	EP2564
NACL	N/A	M4459
ISO OCTANE	N/A	E3554
Diazomethane	N/A	EP2575
Hexane	N/A	E3872
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

pH Adjusted with 6N NaOH>12 prior to Hydrolysis, PH adjusted with cold 12N H2SO4<2 after Hydrolysis, Derivatization procedure is completed and samples are ready to Analyze, 40ml Vial Lot # 03-40BTS721.

KD Bath ID: N/A

Envap ID: NEVAP-02

KD Bath Temperature: N/A

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
01/30/25 12:20	RP (Ext. Lab)	RP (Ext. Lab)
	Preparation Group	Analysis Group

Analytical Method: M8151A-Herbicide-22

Concentration Date: 01/30/2025

Sample ID	Client Sample ID	Test	g /mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Per
					AddedBy	VerifiedBy				
PB166318TB	PB166318TB	TCLP Herbicide	100	6	RUPESH	rajesh	10			SEP-01
PB166382BL	HBLK382	TCLP Herbicide	1000	6	RUPESH	rajesh	10			2
PB166382BS	HLCS382	TCLP Herbicide	1000	6	RUPESH	rajesh	10			3
Q1206-04	JPP-20.1-012725	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		4
Q1206-04MS	JPP-20.1-012725MS	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		5
Q1206-04MS D	JPP-20.1-012725MSD	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		6
Q1206-08	JPP-16.3-012725	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		7
Q1207-04	JPP-2.1-012725	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		8
Q1207-08	JPP-5.1-012725	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		9
Q1207-12	JPP-4.5-012725	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		10
Q1207-16	JPP-16.2-012725	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		11
Q1207-20	JPP-20.2-012725	TCLP Herbicide	100	6	RUPESH	rajesh	10	A		12

* Extracts relinquished on the same date as received.

1/30/25

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Pr P
PB166318TB	LEB318	11	N/A	2000	N/A	N/A	N/A	4.94	1.0	1
Q1205-02	VNJ-236	01	100.03	2000	N/A	N/A	N/A	4.0	1.5	2
Q1206-04	JPP-20.1-012725	02	100.02	2000	N/A	N/A	N/A	6.2	1.0	3
Q1206-08	JPP-16.3-012725	03	100.02	2000	N/A	N/A	N/A	7.2	1.5	4
Q1207-04	JPP-2.1-012725	04	100.03	2000	N/A	N/A	N/A	7.0	1.0	5
Q1207-08	JPP-5.1-012725	05	100.04	2000	N/A	N/A	N/A	7.6	1.5	6
Q1207-12	JPP-4.5-012725	06	100.03	2000	N/A	N/A	N/A	7.2	1.5	7
Q1207-16	JPP-16.2-012725	07	100.02	2000	N/A	N/A	N/A	7.6	1.0	8
Q1207-20	JPP-20.2-012725	08	100.02	2000	N/A	N/A	N/A	7.0	1.5	9
Q1209-04	WC-4	09	100.01	2000	N/A	N/A	N/A	3.5	1.0	10
Q1209-08	WC-5	10	100.02	2000	N/A	N/A	N/A	4.0	1.5	11

111.00
01/29/25

Prep Standard - Chemical Standard Summary

Order ID : Q1206
Test : TCLP Herbicide
Prepbatch ID : PB166382,
Sequence ID/Qc Batch ID: PS013025,

Standard ID :
EP2553,EP2564,EP2576,PP24061,PP24062,PP24064,PP24065,PP24066,PP24067,PP24068,PP24069,PP24070,PP24078,PP24079,

Chemical ID :
E3370,E3551,E3657,E3826,E3843,M4459,M5173,P10549,P11180,P11181,P12619,P12629,P12686,P12708,P12709,P13506,P13507,P13508,P13509,P13523,P13524,P13525,W3112,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3884	6 N NAOH	EP2553	10/21/2024	04/21/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 10/21/2024

FROM 1000.00000ml of W3112 + 240.00000gram of E3657 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1762	1:3 H2SO4 Soln	EP2564	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 250.00000ml of M5173 + 750.00000ml of W3112 = Final Quantity: 1000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
601	Acidified Sodium Sulphate 2	EP2576	01/06/2025	06/02/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 01/06/2025

FROM 100.00000ml of E3370 + 150.00000ml of M5173 + 3000.00000ml of E3551 = Final Quantity: 3000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1321	2/200 PPM Herb Mega Mix	PP24061	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel 11/27/2024

FROM 0.20000ml of P10549 + 1.00000ml of P11180 + 1.00000ml of P12619 + 1.00000ml of P12629 + 1.00000ml of P12686 +
95.80000ml of E3826 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1851	2/200 PPM Herb Mega Mix 2nd Source	PP24062	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 1.00000ml of P11181 + 1.00000ml of P12708 + 1.00000ml of P12709 + 97.00000ml of E3826 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1452	1500 PPB HERB MIX STD	PP24064	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 0.25000ml of E3826 + 0.75000ml of PP24061 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1453	1000 PPB Herb MIX STD	PP24065	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 0.50000ml of E3826 + 0.50000ml of PP24061 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1454	750 PPB Herb MIX STD	PP24066	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 0.25000ml of E3826 + 0.75000ml of PP24065 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1455	500 PPB Herb MIX STD	PP24067	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 0.75000ml of E3826 + 0.25000ml of PP24061 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1456	200 PPB Herb MIX STD	PP24068	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 0.90000ml of E3826 + 0.10000ml of PP24061 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1854	1000 PPB HERB MIX ICV STD	PP24069	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 0.50000ml of E3826 + 0.50000ml of PP24062 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1691	750 PPB ICV HERB STD	PP24070	11/26/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 0.25000ml of E3826 + 0.75000ml of PP24069 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
60	5000 PPB Herbicide Surg Spike (Free Acid)	PP24078	12/10/2024	06/05/2025	Abdul Mirza	None	None	Ankita Jodhani
								12/17/2024

FROM 1.25000ml of P13506 + 1.25000ml of P13507 + 1.25000ml of P13508 + 1.25000ml of P13509 + 195.00000ml of E3843 = Final Quantity: 200.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1848	5000/500000 PPB Herbicide Spike (Free Acid)	PP24079	12/11/2024	06/05/2025	Abdul Mirza	None	None	Ankita Jodhani
								12/17/2024

FROM 0.50000ml of P13525 + 1.00000ml of P13523 + 1.00000ml of P13524 + 47.50000ml of E3843 = Final Quantity: 50.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9244-03 / Ether, Anhydrous, Purified (cs/4x4L)	0000288039	07/17/2025	08/01/2022 / Rajesh	07/13/2022 / Rajesh	E3370

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	05/09/2025	11/09/2024 / Rajesh	11/07/2024 / Rajesh	E3826

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	06/05/2025	12/05/2024 / Rajesh	12/05/2024 / Rajesh	E3843

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3624-05 / Sodium Chloride, Crystal (cs/4x2.5kg)	0000237721	04/13/2026	10/03/2022 / Ankita	10/30/2019 / AMANDEEP	M4459

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32254 / Dalapon Methyl Ester, 1000 ug/ml	A0170243	05/26/2025	11/26/2024 / Ankita	04/06/2021 / dhaval	P10549

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32050 / Herbicide, 8000 series, 515 Surrogate [ester] 2,4-dichlorophenyl acetic acid methyl ester, 1mL, 200ug/mL, Hexane	A0172864	05/26/2025	11/26/2024 / Ankita	11/01/2021 / Abdul	P11180

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32050 / Herbicide, 8000 series, 515 Surrogate [ester] 2,4-dichlorophenyl acetic acid methyl ester, 1mL, 200ug/mL, Hexane	A0172864	05/26/2025	11/26/2024 / Ankita	11/01/2021 / Abdul	P11181

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32062 / Herbicide Mix, 500/8000, Standard #4 [methyl ester] 200ug/mL, hexane, 1mL/ampul	A0155055	05/26/2025	11/26/2024 / Ankita	07/03/2023 / Abdul	P12619

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32055 / Herbicide Mix, 500/8000, Standard #1 [methyl ester] 200ug/mL, hexane, 1mL/ampul	A192429	05/26/2025	11/26/2024 / Ankita	07/03/2023 / Abdul	P12629

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32059 / Herbicide Mix#3 (Methyl Ester), 20000 ug/ml	A0199844	05/26/2025	11/26/2024 / Ankita	07/24/2023 / Abdul	P12686

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151M / Chlorinated Herbicide Mixtures, Methyl Esters	0006752480	05/26/2025	11/26/2024 / Ankita	08/09/2023 / Abdul	P12708

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151M / Chlorinated Herbicide Mixtures, Methyl Esters	0006752480	05/26/2025	11/26/2024 / Ankita	08/09/2023 / Abdul	P12708

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151M / Chlorinated Herbicide Mixtures, Methyl Esters	0006752480	05/26/2025	11/26/2024 / Ankita	08/09/2023 / Abdul	P12709

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151M / Chlorinated Herbicide Mixtures, Methyl Esters	0006752480	05/26/2025	11/26/2024 / Ankita	08/09/2023 / Abdul	P12709

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32049 / Herbicide, 8000 series, 515 Surrogate [free acid] 2,4-dichlorophenyl acetic acid, 1mL, 200ug/mL, MeOH	A0212676	06/10/2025	12/10/2024 / Abdul	08/16/2024 / yogesh	P13506

MeOH

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32049 / Herbicide, 8000 series, 515 Surrogate [free acid] 2,4-dichlorophenyl acetic acid, 1mL, 200ug/mL, MeOH	A0212676	06/10/2025	12/10/2024 / Abdul	08/16/2024 / yogesh	P13507

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32049 / Herbicide, 8000 series, 515 Surrogate [free acid] 2,4-dichlorophenyl acetic acid, 1mL, 200ug/mL, MeOH	A0212676	06/10/2025	12/10/2024 / Abdul	08/16/2024 / yogesh	P13508

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32049 / Herbicide, 8000 series, 515 Surrogate [free acid] 2,4-dichlorophenyl acetic acid, 1mL, 200ug/mL, MeOH	A0212676	06/10/2025	12/10/2024 / Abdul	08/16/2024 / yogesh	P13509

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151A / Chlorinated Herbicide Mixtures, Free Acids	0006810955	06/11/2025	12/11/2024 / Abdul	09/03/2024 / Abdul	P13523

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151A / Chlorinated Herbicide Mixtures, Free Acids	0006810955	06/11/2025	12/11/2024 / Abdul	09/03/2024 / Abdul	P13523

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151A / Chlorinated Herbicide Mixtures, Free Acids	0006810955	06/11/2025	12/11/2024 / Abdul	09/03/2024 / Abdul	P13524

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151A / Chlorinated Herbicide Mixtures, Free Acids	0006810955	06/11/2025	12/11/2024 / Abdul	09/03/2024 / Abdul	P13524

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151A / Chlorinated Herbicide Mixtures, Free Acids	0006810955	06/11/2025	12/11/2024 / Abdul	09/03/2024 / Abdul	P13525

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	HBM-8151A / Chlorinated Herbicide Mixtures, Free Acids	0006810955	06/11/2025	12/11/2024 / Abdul	09/03/2024 / Abdul	P13525

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Sodium Chloride, Crystal
BAKER ANALYZED® A.C.S. Reagent



From M4452 to M4459

Received on : 10/30/2019

Received by : AK

Material No.: 3624-05

Batch No.: 0000237721

Manufactured Date: 2019/04/15

Retest Date: 2026/04/13

Revision No: 1

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,


Test	Specification	Result
Assay (NaCl) (by Ag titrn)	$\geq 99.0 \%$	100.3
pH of 5% Solution at 25°C	5.0 - 9.0	6.0
ACS - Insoluble Matter	$\leq 0.005 \%$	< 0.001
Iodide (I)	$\leq 0.002 \%$	< 0.002
Bromide (Br)	$\leq 0.01 \%$	< 0.01
Chlorate and Nitrate (as NO_3)	$\leq 0.003 \%$	< 0.001
ACS - Phosphate (PO_4)	$\leq 5 \text{ ppm}$	< 5
Sulfate (SO_4)	$\leq 0.004 \%$	< 0.004
Barium (Ba)	Passes Test	PT
ACS - Heavy Metals (as Pb)	$\leq 5 \text{ ppm}$	< 5
Iron (Fe)	$\leq 2 \text{ ppm}$	< 2
Calcium (Ca)	$\leq 0.002 \%$	< 0.001
Magnesium (Mg)	$\leq 0.001 \%$	< 0.001
Potassium (K)	$\leq 0.005 \%$	0.002

For Laboratory, Research or Manufacturing Use

Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US

Packaging Site: Paris Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Ether, Anhydrous
BAKER ANALYZED® A.C.S. Reagent
Contains BHT as a Preservative
Suitable for Fat Extraction



Material No.: 9244-03
Batch No.: 0000288039
Manufactured Date: 2021/07/22
Expiration Date: 2023/07/22
Revision No: 1

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,


Test	Specification	Result
Assay ((C ₂ H ₅) ₂ O) (by GC, corrected for water)	>= 99.0 %	100.0
Alcohol (C ₂ H ₅ OH)	Passes Test	PT
Carbonyl Compounds (as HCHO) (by polarography)	<= 0.001 %	< 0.001
Color (APHA)	<= 10	< 5
Peroxide (as H ₂ O ₂)	<= 1 ppm	< 1
Preservative (BHT)	>= 7 ppm	9
Residue after Evaporation	<= 0.0010 %	< 0.0010
Titration Acid (μeq/g)	<= 0.2	< 0.2
Water (by KF, coulometric)	<= 0.01 %	0.01

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US

Recd. by RP on 7/13/22

E 3370


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %
COMMENTS		
QC: PhC Irma Belmares		

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3826

Rec'd by RP on 11/7/24

Jamie Croak
Director Quality Operations, Bioscience Production

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 12/5/24

E 3843

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis



Material No.: 9530-33
Batch No.: 0000281827
Manufactured Date: 2021/03/30
Retest Date: 2026/03/29
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Katelyn McGinnis
Katelyn McGinnis - Operations Tech I

Date Mixed: 28-May-2021

Balance: B345965662

Marlene Cowan
Marlene Cowan - Operations Tech I

Date Passed: 02-Jun-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

1911177
70
P 111 86
AR
11/02/21



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32050 **Lot No.:** A0172864

Description : 2,4-Dichlorophenylacetic Acid Methyl Ester Standard

515 Surrogate (ester) 2, 4-dichlorophenyl Acetic Acid Methyl Ester
200µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 29, 2028 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2,4-Dichlorophenyl acetic acid methyl ester	202.0 µg/mL	+/- 1.4323 µg/mL Gravimetric
	CAS # 55954-23-9 (Lot CSC42194-01)		+/- 6.8182 µg/mL Unstressed
	Purity 99%		+/- 6.8182 µg/mL Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P11177
↓
P11186
—
AR
01/02/21

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Katelyn McGinnis - Operations Tech I

Date Mixed: 28-May-2021

Balance: B345965662


Marlene Cowan - Operations Tech I

Date Passed: 02-Jun-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

1911177
70
P 111 86
AR
11/02/21



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32050 **Lot No.:** A0172864
Description : 2,4-Dichlorophenylacetic Acid Methyl Ester Standard
515 Surrogate (ester) 2, 4-dichlorophenyl Acetic Acid Methyl Ester
200µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : February 29, 2028 **Storage:** 10°C or colder
Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2,4-Dichlorophenyl acetic acid methyl ester	202.0 µg/mL	+/- 1.4323 µg/mL Gravimetric
	CAS # 55954-23-9 (Lot CSC42194-01)		+/- 6.8182 µg/mL Unstressed
	Purity 99%		+/- 6.8182 µg/mL Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P11177
↓
P11186
—
AR
01/02/21



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32062 **Lot No.:** A0155055

Description : Herbicide Mix #4/ME (Methyl Ester)
Herbicide Mix #4/ME (Methyl Ester) 200µg/mL,
Hexane/Methyl-tert-butyl-ether, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 10°C or colder

P12616 / (5)
↓
P12620
✓ *Paul* = 7/5/2023

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)			
1	3,5-Dichlorobenzoic acid methyl ester	200.0 µg/mL	+/-	1.4182	µg/mL	Gravimetric
	CAS # 2905-67-1 (Lot 3903900)		+/-	6.7507	µg/mL	Unstressed
	Purity 99%		+/-	6.7507	µg/mL	Stressed
2	4-Nitroanisole	200.0 µg/mL	+/-	1.4182	µg/mL	Gravimetric
	CAS # 100-17-4 (Lot 24765/7)		+/-	6.7507	µg/mL	Unstressed
	Purity 99%		+/-	6.7507	µg/mL	Stressed
3	Pentachloroanisole	200.0 µg/mL	+/-	1.4182	µg/mL	Gravimetric
	CAS # 1825-21-4 (Lot 7921100)		+/-	6.7507	µg/mL	Unstressed
	Purity 99%		+/-	6.7507	µg/mL	Stressed
4	Chloramben methyl ester	199.9 µg/mL	+/-	1.4176	µg/mL	Gravimetric
	CAS # 7286-84-2 (Lot 6487100)		+/-	6.7480	µg/mL	Unstressed
	Purity 98%		+/-	6.7480	µg/mL	Stressed
5	Bentazon methyl ester	200.0 µg/mL	+/-	1.4182	µg/mL	Gravimetric
	CAS # 61592-45-8 (Lot 817100)		+/-	6.7507	µg/mL	Unstressed
	Purity 99%		+/-	6.7507	µg/mL	Stressed
6	Picloram methyl ester	201.9 µg/mL	+/-	1.4315	µg/mL	Gravimetric
	CAS # 14143-55-6 (Lot 386-21B)		+/-	6.8141	µg/mL	Unstressed
	Purity 98%		+/-	6.8141	µg/mL	Stressed
7	DCPA methyl ester (Chlorthal-dimethyl)	200.0 µg/mL	+/-	1.4182	µg/mL	Gravimetric
	CAS # 1861-32-1 (Lot 8008700)		+/-	6.7507	µg/mL	Unstressed
	Purity 99%		+/-	6.7507	µg/mL	Stressed

8	Acifluorfen methyl ester	200.0 µg/mL	+/- 1.4182	µg/mL	Gravimetric
	CAS # 50594-67-7 (Lot 6282300)		+/- 6.7507	µg/mL	Unstressed
	Purity 99%		+/- 6.7507	µg/mL	Stressed

Solvent: Hexane/Methyl-tert-butyl-ether
CAS # 110-54-3/1634-04-4
Purity 99%

Column:
 30m x 0.25mm x 0.25µm
 Rtx-5 (cat.#10223)


Carrier Gas:
 hydrogen-constant pressure 10 psi.

Temp. Program:
 75°C (hold 1 min.) to 330°C
 @ 20°C/min. (hold 10 min.)

Inj. Temp:
 250°C

Det. Temp:
 330°C

Det. Type:
 FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maje

Date Mixed: 14-Nov-2019 **Balance:** 1128353505

Justine Albertson
 Justine Albertson - Operations Tech-ARM QC

Date Passed: 18-Nov-2019

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32055 **Lot No.:** A0192429

Description : Herbicide Mix #1/ME (Methyl Ester)
Herbicide Mix #1/ME (Methyl Ester) 200 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

P12626 / 5
P12630
7/5/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dicamba methyl ester	6597-78-0	11705400	99%	201.6 µg/mL	+/- 3.4204
2	Dichlorprop methyl ester	57153-17-0	11672100	99%	201.4 µg/mL	+/- 3.4170
3	2,4-D methyl ester	1928-38-7	10048000	99%	201.2 µg/mL	+/- 3.4136
4	2,4,5-TP (silvex) methyl ester	4841-20-7	6364900	99%	201.2 µg/mL	+/- 3.4136
5	2,4,5-T methyl ester	1928-37-6	6875800	98%	200.7 µg/mL	+/- 3.4052
6	Dinoseb methyl ether	6099-79-2	12914300	99%	200.8 µg/mL	+/- 3.4068
7	2,4-DB methyl ester	18625-12-2	12542000	99%	201.0 µg/mL	+/- 3.4102

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin
Penelope Riglin - Operations Tech I

Date Mixed: 09-Dec-2022

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 12-Dec-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32059 **Lot No.:** A0199844

Description : Herbicide Mix #3/ME (Methyl Ester)
Herbicide Mix #3/ME (Methyl Ester) 20,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

P 12685 / (S)
P 12689 /
RAU = 7/24/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	MCP (Mecoprop) methyl ester	23844-56-6	14546400	99%	20,035.0 µg/mL	+/- 360.1907
2	MCPA methyl ester	2436-73-9	SL201209	99%	20,055.0 µg/mL	+/- 360.5503

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician

Date Mixed: 12-Jul-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 19-Jul-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P12706
↓
P12715
/ (10)
/ 1
/ 8/15/23

ISO 17034

Reference Material Certificate
Product Information Sheet

Product Name: Chlorinated Methylated Herbicides Standard

Lot Number: 0006752480

Product Number: HBM-8151M-1

Lot Issue Date: 18-Jul-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Aug-2025

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
acifluorfen methyl ester	100.3 ±	0.5 µg/mL	050594-67-7	RM03058
bentazon methyl derivative	100.2 ±	0.5 µg/mL	061592-45-8	RM13829
chloramben methyl ester	100.4 ±	0.5 µg/mL	007286-84-2	RM03055
2,4-D methyl ester	100.2 ±	0.5 µg/mL	001928-38-7	RM03040
dalapon methyl ester	100.4 ±	0.5 µg/mL	017640-02-7	RM14219
2,4-DB methyl ester	100.2 ±	0.5 µg/mL	018625-12-2	RM03029
DCPA	100.2 ±	0.5 µg/mL	001861-32-1	RM13426
dicamba methyl ester	100.4 ±	0.5 µg/mL	006597-78-0	RM03039
methyl-3,5-dichlorobenzoate	100.1 ±	0.5 µg/mL	002905-67-1	RM03048
dichlorprop methyl ester	100.4 ±	0.5 µg/mL	057153-17-0	NT02086
dinoseb methyl ether	100.5 ±	0.5 µg/mL	006099-79-2	RM03051
MCPA methyl ester	10031 ±	50 µg/mL	002436-73-9	RM12863
MCPP methyl ester	10031 ±	50 µg/mL	023844-56-6	RM20060
4-nitroanisole	100.3 ±	0.5 µg/mL	000100-17-4	RM02806
pentachloroanisole	100.4 ±	0.5 µg/mL	001825-21-4	RM02457
picloram methyl ester	100.2 ±	0.5 µg/mL	014143-55-6	RM03044
silvex methyl ester	100.2 ±	0.5 µg/mL	004841-20-7	RM03799
2,4,5-T methyl ester	100.4 ±	0.5 µg/mL	001928-37-6	RM03033

Matrix: methanol (methyl alcohol)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.


Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative

P 12706 / (10)
↓
P 12715
✓ RAUF
8.15.23



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937

P12706
↓
P12715
/ (10)
/ 1
/ 8/15/23

ISO 17034

Reference Material Certificate
Product Information Sheet

Product Name: Chlorinated Methylated Herbicides Standard

Lot Number: 0006752480

Product Number: HBM-8151M-1

Lot Issue Date: 18-Jul-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Aug-2025

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
acifluorfen methyl ester	100.3 ±	0.5 µg/mL	050594-67-7	RM03058
bentazon methyl derivative	100.2 ±	0.5 µg/mL	061592-45-8	RM13829
chloramben methyl ester	100.4 ±	0.5 µg/mL	007286-84-2	RM03055
2,4-D methyl ester	100.2 ±	0.5 µg/mL	001928-38-7	RM03040
dalapon methyl ester	100.4 ±	0.5 µg/mL	017640-02-7	RM14219
2,4-DB methyl ester	100.2 ±	0.5 µg/mL	018625-12-2	RM03029
DCPA	100.2 ±	0.5 µg/mL	001861-32-1	RM13426
dicamba methyl ester	100.4 ±	0.5 µg/mL	006597-78-0	RM03039
methyl-3,5-dichlorobenzoate	100.1 ±	0.5 µg/mL	002905-67-1	RM03048
dichlorprop methyl ester	100.4 ±	0.5 µg/mL	057153-17-0	NT02086
dinoseb methyl ether	100.5 ±	0.5 µg/mL	006099-79-2	RM03051
MCPA methyl ester	10031 ±	50 µg/mL	002436-73-9	RM12863
MCPP methyl ester	10031 ±	50 µg/mL	023844-56-6	RM20060
4-nitroanisole	100.3 ±	0.5 µg/mL	000100-17-4	RM02806
pentachloroanisole	100.4 ±	0.5 µg/mL	001825-21-4	RM02457
picloram methyl ester	100.2 ±	0.5 µg/mL	014143-55-6	RM03044
silvex methyl ester	100.2 ±	0.5 µg/mL	004841-20-7	RM03799
2,4,5-T methyl ester	100.4 ±	0.5 µg/mL	001928-37-6	RM03033

Matrix: methanol (methyl alcohol)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.


Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative

P 12706 / (10)
↓
P 12715
✓ RAUF
8.15.23



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32049 **Lot No.:** A0212676
Description : 2,4-Dichlorophenylacetic Acid Standard
2, 4-Dichlorophenyl Acetic Acid 200µg/mL, Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2027 **Storage:** 10°C or colder
Handling: This product is photosensitive. **Ship:** Ambient

P13697 } Y.P.
↓
P13515 } 08/15/26

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4-dichlorophenylacetic acid	19719-28-9	STBK3827	99%	200.0 µg/mL	+/- 2.7154

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Specific Reference Material Notes:

Failure to derivatize this standard will lead to incorrect quantitative results.

Quality Confirmation Test

Column:
150mm x 4.6mm
Allure C18 Cat. (#9164565)

Flow Rate:
1.0 ml/min.


Mobile Phase A:
0.14% H₃PO₄ in water

Mobile Phase B:
acetonitrile

Mobile Phase Composition:
90%B Isocratic

Det. Type:
Wavelength: 220 & 254 nm

Inj. Vol
5µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 11-Jun-2024

Balance Serial # B345965662


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Jun-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32049 **Lot No.:** A0212676
Description : 2,4-Dichlorophenylacetic Acid Standard
2, 4-Dichlorophenyl Acetic Acid 200µg/mL, Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2027 **Storage:** 10°C or colder
Handling: This product is photosensitive. **Ship:** Ambient

P13697 } Y.P.
↓
P13515 } 08/15/26

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4-dichlorophenylacetic acid	19719-28-9	STBK3827	99%	200.0 µg/mL	+/- 2.7154

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Specific Reference Material Notes:

Failure to derivatize this standard will lead to incorrect quantitative results.

Quality Confirmation Test

Column:
150mm x 4.6mm
Allure C18 Cat. (#9164565)

Flow Rate:
1.0 ml/min.


Mobile Phase A:
0.14% H₃PO₄ in water

Mobile Phase B:
acetonitrile

Mobile Phase Composition:
90%B Isocratic

Det. Type:
Wavelength: 220 & 254 nm

Inj. Vol
5µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 11-Jun-2024

Balance Serial # B345965662


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Jun-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32049 **Lot No.:** A0212676
Description : 2,4-Dichlorophenylacetic Acid Standard
2, 4-Dichlorophenyl Acetic Acid 200µg/mL, Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2027 **Storage:** 10°C or colder
Handling: This product is photosensitive. **Ship:** Ambient

P13697 } Y.P.
↓
P13515 } 08/15/26

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4-dichlorophenylacetic acid	19719-28-9	STBK3827	99%	200.0 µg/mL	+/- 2.7154

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Specific Reference Material Notes:

Failure to derivatize this standard will lead to incorrect quantitative results.

Quality Confirmation Test

Column:
150mm x 4.6mm
Allure C18 Cat. (#9164565)

Flow Rate:
1.0 ml/min.


Mobile Phase A:
0.14% H₃PO₄ in water

Mobile Phase B:
acetonitrile

Mobile Phase Composition:
90%B Isocratic

Det. Type:
Wavelength: 220 & 254 nm

Inj. Vol
5µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 11-Jun-2024

Balance Serial # B345965662


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Jun-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32049 **Lot No.:** A0212676
Description : 2,4-Dichlorophenylacetic Acid Standard
2, 4-Dichlorophenyl Acetic Acid 200µg/mL, Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2027 **Storage:** 10°C or colder
Handling: This product is photosensitive. **Ship:** Ambient

P13697 } Y.P.
↓
P13515 } 08/15/26

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4-dichlorophenylacetic acid	19719-28-9	STBK3827	99%	200.0 µg/mL	+/- 2.7154

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Specific Reference Material Notes:

Failure to derivatize this standard will lead to incorrect quantitative results.

Quality Confirmation Test

Column:
150mm x 4.6mm
Allure C18 Cat. (#9164565)

Flow Rate:
1.0 ml/min.


Mobile Phase A:
0.14% H₃PO₄ in water

Mobile Phase B:
acetonitrile

Mobile Phase Composition:
90%B Isocratic

Det. Type:
Wavelength: 220 & 254 nm

Inj. Vol
5µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 11-Jun-2024

Balance Serial # B345965662


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Jun-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Reference Material Certificate
Product Information Sheet

Product Name: Chlorinated Herbicides Standard
Product Number: HBM-8151A-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006810955
Lot Issue Date: 20-Aug-2024
Expiration Date: 30-Sep-2026

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
acifluorfen	100.2 ±	0.5 µg/mL	050594-66-6	NT02057
bentazon	100.4 ±	0.5 µg/mL	025057-89-0	RM21359
chloramben	100.3 ±	0.5 µg/mL	000133-90-4	RM02698
2,4-D	100.4 ±	0.5 µg/mL	000094-75-7	RM17172
dalapon	100.4 ±	0.5 µg/mL	000075-99-0	RM19677
2,4-DB	100.1 ±	0.5 µg/mL	000094-82-6	RM02866
tetrachloroterephthalic acid	100.4 ±	0.5 µg/mL	002136-79-0	RM15140
dicamba	100.3 ±	0.5 µg/mL	001918-00-9	RM22113
3,5-dichlorobenzoic acid	100.4 ±	0.5 µg/mL	000051-36-5	RM02768
dichlorprop	100.2 ±	0.5 µg/mL	000120-36-5	RM21688
dinoseb	100.3 ±	0.5 µg/mL	000088-85-7	RM22275
MCPA	10019 ±	50 µg/mL	000094-74-6	RM12220
MCPP (mecoprop)	10011 ±	50 µg/mL	000093-65-2	RM09273
4-nitrophenol	100.4 ±	0.5 µg/mL	000100-02-7	RM02391
pentachlorophenol	100.2 ±	0.5 µg/mL	000087-86-5	RM02474
picloram	100.4 ±	0.5 µg/mL	001918-02-1	RM20442
silvex	100.5 ±	0.5 µg/mL	000093-72-1	RM22116
2,4,5-T	100.3 ±	0.5 µg/mL	000093-76-5	RM19314

Matrix: methanol (methyl alcohol)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Page: 1 of 2

CSD-QA-015.2

250 Smith Street North Kingstown, Rhode Island 02852

P13520
↓
P13536
18
BALT
9/4/2024

www.agilent.com/quality

ISO 17025
Cert No. AT-1937

Reference Material Certificate
Product Information Sheet

Product Name: Chlorinated Herbicides Standard
Product Number: HBM-8151A-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006810955
Lot Issue Date: 20-Aug-2024
Expiration Date: 30-Sep-2026

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
acifluorfen	100.2 ±	0.5 µg/mL	050594-66-6	NT02057
bentazon	100.4 ±	0.5 µg/mL	025057-89-0	RM21359
chloramben	100.3 ±	0.5 µg/mL	000133-90-4	RM02698
2,4-D	100.4 ±	0.5 µg/mL	000094-75-7	RM17172
dalapon	100.4 ±	0.5 µg/mL	000075-99-0	RM19677
2,4-DB	100.1 ±	0.5 µg/mL	000094-82-6	RM02866
tetrachloroterephthalic acid	100.4 ±	0.5 µg/mL	002136-79-0	RM15140
dicamba	100.3 ±	0.5 µg/mL	001918-00-9	RM22113
3,5-dichlorobenzoic acid	100.4 ±	0.5 µg/mL	000051-36-5	RM02768
dichlorprop	100.2 ±	0.5 µg/mL	000120-36-5	RM21688
dinoseb	100.3 ±	0.5 µg/mL	000088-85-7	RM22275
MCPA	10019 ±	50 µg/mL	000094-74-6	RM12220
MCPP (mecoprop)	10011 ±	50 µg/mL	000093-65-2	RM09273
4-nitrophenol	100.4 ±	0.5 µg/mL	000100-02-7	RM02391
pentachlorophenol	100.2 ±	0.5 µg/mL	000087-86-5	RM02474
picloram	100.4 ±	0.5 µg/mL	001918-02-1	RM20442
silvex	100.5 ±	0.5 µg/mL	000093-72-1	RM22116
2,4,5-T	100.3 ±	0.5 µg/mL	000093-76-5	RM19314

Matrix: methanol (methyl alcohol)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Page: 1 of 2

CSD-QA-015.2

250 Smith Street North Kingstown, Rhode Island 02852

P13520
↓
P13536
18
BALT
9/4/2024

www.agilent.com/quality

ISO 17025
Cert No. AT-1937

Reference Material Certificate
Product Information Sheet

Product Name: Chlorinated Herbicides Standard
Product Number: HBM-8151A-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006810955
Lot Issue Date: 20-Aug-2024
Expiration Date: 30-Sep-2026

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
acifluorfen	100.2 ±	0.5 µg/mL	050594-66-6	NT02057
bentazon	100.4 ±	0.5 µg/mL	025057-89-0	RM21359
chloramben	100.3 ±	0.5 µg/mL	000133-90-4	RM02698
2,4-D	100.4 ±	0.5 µg/mL	000094-75-7	RM17172
dalapon	100.4 ±	0.5 µg/mL	000075-99-0	RM19677
2,4-DB	100.1 ±	0.5 µg/mL	000094-82-6	RM02866
tetrachloroterephthalic acid	100.4 ±	0.5 µg/mL	002136-79-0	RM15140
dicamba	100.3 ±	0.5 µg/mL	001918-00-9	RM22113
3,5-dichlorobenzoic acid	100.4 ±	0.5 µg/mL	000051-36-5	RM02768
dichlorprop	100.2 ±	0.5 µg/mL	000120-36-5	RM21688
dinoseb	100.3 ±	0.5 µg/mL	000088-85-7	RM22275
MCPA	10019 ±	50 µg/mL	000094-74-6	RM12220
MCPP (mecoprop)	10011 ±	50 µg/mL	000093-65-2	RM09273
4-nitrophenol	100.4 ±	0.5 µg/mL	000100-02-7	RM02391
pentachlorophenol	100.2 ±	0.5 µg/mL	000087-86-5	RM02474
picloram	100.4 ±	0.5 µg/mL	001918-02-1	RM20442
silvex	100.5 ±	0.5 µg/mL	000093-72-1	RM22116
2,4,5-T	100.3 ±	0.5 µg/mL	000093-76-5	RM19314

Matrix: methanol (methyl alcohol)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Page: 1 of 2

CSD-QA-015.2

250 Smith Street North Kingstown, Rhode Island 02852

P13520
↓
P13536 } (18)

BACF
9/4/2024

www.agilent.com/quality

ISO 17025
Cert No. AT-1937



SHIPPING DOCUMENTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

247 of 250

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1206	RUTW01	Order Date : 1/28/2025 11:18:51 AM	YG	Project Mgr : Kiran
Client Name : RU2 Engineering, LLC		Project Name : SANTWOBR-BMCR-Bro 02/03/25		Report Type : NYS ASP B
Client Contact : Rutu Manani		Receive DateTime : 1/28/2025 12:59:00 PM		EDD Type : Excel NY
Invoice Name : RU2 Engineering, LLC		Purchase Order :		Hard Copy Date :
Invoice Contact : Rutu Manani				Date Signoff : 1/28/2025 2:56:10 PM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1206-01	JPP-20.1-012725	Solid	01/27/2025	14:15					
					VOCMS Group1		8260D	10 Bus. Days	
Q1206-05	JPP-16.3-012725	Solid	01/27/2025	15:10					
					VOCMS Group1		8260D	10 Bus. Days	


Relinquished By :

Date / Time :


1-28-25 1447

Received By :

Date / Time :


1/28/25 14:47 

Storage Area : VOA Refridgerator Room

LOGIN REPORT/SAMPLE TRANSFER


Order ID : Q1206	RUTW01	Order Date : 1/28/2025 11:18:51 AM	Project Mgr : Kiran
Client Name : RU2 Engineering, LLC	Project Name : SANTWOBR-BMCR Bro NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Report Type : NYS ASP B	
Client Contact : Rutu Manani	Receive DateTime : 1/28/2025 12:59:00 PM	EDD Type : Excel NY	
Invoice Name : RU2 Engineering, LLC	Purchase Order :	Hard Copy Date :	
Invoice Contact : Rutu Manani		Date Signoff : 1/28/2025 2:56:10 PM	

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1206-0301	JPP-20.1-012725	Solid	01/27/2025	14:18 14:15	Gasoline Range Organics		8015D		10 Bus. Days
Q1206-0705	JPP-16.3-012725	Solid	01/27/2025	15:17 15:10	Gasoline Range Organics		8015D		10 Bus. Days
		YG	02/03/25						

Relinquished By: 

Date / Time :


1-28-25 1447

Received By: 

Date / Time :

1/28/25 14.47

Storage Area : VOA Refridgerator Room


P22